

metals
- 5a -
MATRIX SPIKE SUMMARY

client: <u>Tully Environmental, Inc</u>	level: <u>low</u>	sdg no.: <u>Q3283</u>
contract: <u>TULL01</u>		lab code: <u>ACE</u>
matrix: <u>Water</u>	sample id: <u>Q3283-02</u>	client id: <u>002-35TH-AVE(SEP)MS</u>
Percent Solids for Sample: <u>NA</u>	Spiked ID: <u>Q3283-02MS</u>	Percent Solids for Spike Sample: <u>NA</u>

Analyte	Units	Acceptance Limit %R	Spiked Result	C	Sample Result	C	Spike Added	% Recovery	Qual	M
Copper	ug/L	75 - 125	152		2.83	J	150	100		P
Iron	ug/L	75 - 125	2390		859		1500	102		P
Lead	ug/L	75 - 125	468		1.88	J	500	93		P

metals
- 5a -
MATRIX SPIKE DUPLICATE SUMMARY

client: <u>Tully Environmental, Inc</u>	level: <u>low</u>	sdg no.: <u>Q3283</u>
contract: <u>TULL01</u>		lab code: <u>ACE</u>
matrix: <u>Water</u>	sample id: <u>Q3283-02</u>	client id: <u>002-35TH-AVE(SEP)MSD</u>
Percent Solids for Sample: <u>NA</u>	Spiked ID: <u>Q3283-02MSD</u>	Percent Solids for Spike Sample: <u>NA</u>

Analyte	Units	Acceptance Limit %R	MSD Result	C	Sample Result	C	Spike Added	% Recovery	Qual	M
Copper	ug/L	75 - 125	154		2.83	J	150	101		P
Iron	ug/L	75 - 125	2450		859		1500	106		P
Lead	ug/L	75 - 125	471		1.88	J	500	94		P