

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** CLOU03
Lab Code: ACE **SDG NO.:** Q3284
GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 09/23/2025 09/23/2025

Client Sample No.: CCAL01 **Date Analyzed:** 10/09/2025
Lab Sample No.: AR1660CCC500 **Data File :** PP075684.D **Time Analyzed:** 20:22

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	5.791	5.708	5.908	566.800	500.000	13.4
Aroclor-1016-2	5.815	5.729	5.929	475.710	500.000	-4.9
Aroclor-1016-3	5.878	5.792	5.992	464.110	500.000	-7.2
Aroclor-1016-4	5.974	5.889	6.089	503.270	500.000	0.7
Aroclor-1016-5	6.267	6.181	6.381	499.930	500.000	0.0
Aroclor-1260-1	7.384	7.299	7.499	450.080	500.000	-10.0
Aroclor-1260-2	7.637	7.551	7.751	439.090	500.000	-12.2
Aroclor-1260-3	7.995	7.909	8.109	471.330	500.000	-5.7
Aroclor-1260-4	8.222	8.137	8.337	488.890	500.000	-2.2
Aroclor-1260-5	8.548	8.464	8.664	488.020	500.000	-2.4
Decachlorobiphenyl	10.413	10.334	10.534	46.790	50.000	-6.4
Tetrachloro-m-xylene	4.643	4.555	4.755	43.010	50.000	-14.0

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** CLOU03
Lab Code: ACE **SDG NO.:** Q3284
GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 09/23/2025 09/23/2025

Client Sample No.: CCAL01 **Date Analyzed:** 10/09/2025
Lab Sample No.: AR1660CCC500 **Data File :** PP075684.D **Time Analyzed:** 20:22

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.885	4.793	4.993	562.380	500.000	12.5
Aroclor-1016-2	4.942	4.851	5.051	602.230	500.000	20.4
Aroclor-1016-3	5.063	4.971	5.171	540.510	500.000	8.1
Aroclor-1016-4	5.103	5.011	5.211	557.870	500.000	11.6
Aroclor-1016-5	5.317	5.225	5.425	564.790	500.000	13.0
Aroclor-1260-1	6.344	6.255	6.455	450.620	500.000	-9.9
Aroclor-1260-2	6.531	6.441	6.641	455.090	500.000	-9.0
Aroclor-1260-3	6.685	6.595	6.795	438.720	500.000	-12.3
Aroclor-1260-4	7.153	7.065	7.265	399.670	500.000	-20.1
Aroclor-1260-5	7.393	7.304	7.504	420.930	500.000	-15.8
Decachlorobiphenyl	8.789	8.703	8.903	51.280	50.000	2.6
Tetrachloro-m-xylene	3.789	3.696	3.896	46.710	50.000	-6.6