

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** CLOU03
Lab Code: ACE **SDG NO.:** Q3284
GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 09/23/2025 09/23/2025

Client Sample No.: CCAL02 **Date Analyzed:** 10/10/2025
Lab Sample No.: AR1660CCC500 **Data File :** PP075698.D **Time Analyzed:** 01:30

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	5.801	5.708	5.908	507.960	500.000	1.6
Aroclor-1016-2	5.823	5.729	5.929	504.130	500.000	0.8
Aroclor-1016-3	5.885	5.792	5.992	489.210	500.000	-2.2
Aroclor-1016-4	5.982	5.889	6.089	529.650	500.000	5.9
Aroclor-1016-5	6.274	6.181	6.381	538.510	500.000	7.7
Aroclor-1260-1	7.392	7.299	7.499	495.670	500.000	-0.9
Aroclor-1260-2	7.644	7.551	7.751	486.500	500.000	-2.7
Aroclor-1260-3	8.002	7.909	8.109	524.410	500.000	4.9
Aroclor-1260-4	8.230	8.137	8.337	543.930	500.000	8.8
Aroclor-1260-5	8.555	8.464	8.664	534.490	500.000	6.9
Decachlorobiphenyl	10.421	10.334	10.534	51.070	50.000	2.1
Tetrachloro-m-xylene	4.648	4.555	4.755	44.410	50.000	-11.2

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** CLOU03
Lab Code: ACE **SDG NO.:** Q3284
GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 09/23/2025 09/23/2025

Client Sample No.: CCAL02 **Date Analyzed:** 10/10/2025
Lab Sample No.: AR1660CCC500 **Data File :** PP075698.D **Time Analyzed:** 01:30

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.888	4.793	4.993	587.780	500.000	17.6
Aroclor-1016-2	4.944	4.851	5.051	576.750	500.000	15.4
Aroclor-1016-3	5.064	4.971	5.171	577.650	500.000	15.5
Aroclor-1016-4	5.106	5.011	5.211	574.140	500.000	14.8
Aroclor-1016-5	5.320	5.225	5.425	548.320	500.000	9.7
Aroclor-1260-1	6.347	6.255	6.455	505.490	500.000	1.1
Aroclor-1260-2	6.534	6.441	6.641	515.250	500.000	3.1
Aroclor-1260-3	6.687	6.595	6.795	502.890	500.000	0.6
Aroclor-1260-4	7.155	7.065	7.265	476.570	500.000	-4.7
Aroclor-1260-5	7.395	7.304	7.504	510.360	500.000	2.1
Decachlorobiphenyl	8.790	8.703	8.903	56.330	50.000	12.7
Tetrachloro-m-xylene	3.793	3.696	3.896	47.410	50.000	-5.2