



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Tetra Tech, EMI

SDG No.: Q3291

Contract: TETR16

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
ICB61	Mercury	0.076	+/-0.2	U	0.20	CV	10/07/2025	10:06	LB137441

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Tetra Tech, EMI

SDG No.: Q3291

Contract: TETR16

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB90	Mercury	0.076	+/-0.2	U	0.20	CV	10/07/2025	10:11	LB137441
CCB91	Mercury	0.076	+/-0.2	U	0.20	CV	10/07/2025	10:47	LB137441
CCB92	Mercury	0.076	+/-0.2	U	0.20	CV	10/07/2025	11:16	LB137441
CCB93	Mercury	0.076	+/-0.2	U	0.20	CV	10/07/2025	11:39	LB137441
CCB94	Mercury	0.076	+/-0.2	U	0.20	CV	10/07/2025	11:57	LB137441

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Tetra Tech, EMI

SDG No.: Q3291

Contract: TETR16

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
ICB01	Aluminum	1.94	+/-10	U	20.0	P	10/07/2025	17:33	LB137449
	Antimony	0.11	+/-1	U	2.00	P	10/07/2025	17:33	LB137449
	Arsenic	0.089	+/-0.5	U	1.00	P	10/07/2025	17:33	LB137449
	Barium	0.21	+/-5	U	10.0	P	10/07/2025	17:33	LB137449
	Beryllium	0.32	+/-0.5	U	1.00	P	10/07/2025	17:33	LB137449
	Cadmium	0.34	+/-0.5	U	1.00	P	10/07/2025	17:33	LB137449
	Calcium	45.7	+/-250	U	500	P	10/07/2025	17:33	LB137449
	Chromium	0.21	+/-1	U	2.00	P	10/07/2025	17:33	LB137449
	Cobalt	0.070	+/-0.5	U	1.00	P	10/07/2025	17:33	LB137449
	Copper	0.30	+/-1	U	2.00	P	10/07/2025	17:33	LB137449
	Iron	7.81	+/-25	U	50.0	P	10/07/2025	17:33	LB137449
	Lead	0.21	+/-0.5	U	1.00	P	10/07/2025	17:33	LB137449
	Magnesium	19.5	+/-250	U	500	P	10/07/2025	17:33	LB137449
	Manganese	0.43	+/-0.5	U	1.00	P	10/07/2025	17:33	LB137449
	Nickel	0.27	+/-0.5	U	1.00	P	10/07/2025	17:33	LB137449
	Potassium	36.4	+/-250	U	500	P	10/07/2025	17:33	LB137449
	Selenium	2.90	+/-2.5	U	5.00	P	10/07/2025	17:33	LB137449
	Silver	0.060	+/-0.5	U	1.00	P	10/07/2025	17:33	LB137449
	Sodium	128	+/-250	U	500	P	10/07/2025	17:33	LB137449
	Thallium	0.060	+/-0.5	U	1.00	P	10/07/2025	17:33	LB137449
	Vanadium	0.077	+/-2.5	U	5.00	P	10/07/2025	17:33	LB137449
	Zinc	1.25	+/-2.5	U	5.00	P	10/07/2025	17:33	LB137449

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Tetra Tech, EMI

SDG No.: Q3291

Contract: TETR16

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB01	Aluminum	1.94	+/-10	U	20.0	P	10/07/2025	17:45	LB137449
	Antimony	0.17	+/-1	J	2.00	P	10/07/2025	17:45	LB137449
	Arsenic	0.089	+/-0.5	U	1.00	P	10/07/2025	17:45	LB137449
	Barium	0.21	+/-5	U	10.0	P	10/07/2025	17:45	LB137449
	Beryllium	0.32	+/-0.5	U	1.00	P	10/07/2025	17:45	LB137449
	Cadmium	0.34	+/-0.5	U	1.00	P	10/07/2025	17:45	LB137449
	Calcium	45.7	+/-250	U	500	P	10/07/2025	17:45	LB137449
	Chromium	0.21	+/-1	U	2.00	P	10/07/2025	17:45	LB137449
	Cobalt	0.070	+/-0.5	U	1.00	P	10/07/2025	17:45	LB137449
	Copper	0.30	+/-1	U	2.00	P	10/07/2025	17:45	LB137449
	Iron	7.81	+/-25	U	50.0	P	10/07/2025	17:45	LB137449
	Lead	0.21	+/-0.5	U	1.00	P	10/07/2025	17:45	LB137449
	Magnesium	19.5	+/-250	U	500	P	10/07/2025	17:45	LB137449
	Manganese	0.43	+/-0.5	U	1.00	P	10/07/2025	17:45	LB137449
	Nickel	0.27	+/-0.5	U	1.00	P	10/07/2025	17:45	LB137449
	Potassium	36.4	+/-250	U	500	P	10/07/2025	17:45	LB137449
	Selenium	2.90	+/-2.5	U	5.00	P	10/07/2025	17:45	LB137449
	Silver	0.060	+/-0.5	J	1.00	P	10/07/2025	17:45	LB137449
	Sodium	128	+/-250	U	500	P	10/07/2025	17:45	LB137449
	Thallium	0.11	+/-0.5	J	1.00	P	10/07/2025	17:45	LB137449
	Vanadium	0.077	+/-2.5	U	5.00	P	10/07/2025	17:45	LB137449
	Zinc	1.25	+/-2.5	U	5.00	P	10/07/2025	17:45	LB137449
CCB02	Aluminum	1.94	+/-10	U	20.0	P	10/07/2025	18:22	LB137449
	Antimony	0.17	+/-1	J	2.00	P	10/07/2025	18:22	LB137449
	Arsenic	0.089	+/-0.5	U	1.00	P	10/07/2025	18:22	LB137449
	Barium	0.21	+/-5	U	10.0	P	10/07/2025	18:22	LB137449
	Beryllium	0.32	+/-0.5	U	1.00	P	10/07/2025	18:22	LB137449
	Cadmium	0.34	+/-0.5	U	1.00	P	10/07/2025	18:22	LB137449
	Calcium	45.7	+/-250	U	500	P	10/07/2025	18:22	LB137449
	Chromium	0.21	+/-1	U	2.00	P	10/07/2025	18:22	LB137449
	Cobalt	0.070	+/-0.5	U	1.00	P	10/07/2025	18:22	LB137449
	Copper	0.31	+/-1	J	2.00	P	10/07/2025	18:22	LB137449
	Iron	7.81	+/-25	U	50.0	P	10/07/2025	18:22	LB137449
	Lead	0.21	+/-0.5	U	1.00	P	10/07/2025	18:22	LB137449
	Magnesium	19.5	+/-250	U	500	P	10/07/2025	18:22	LB137449
	Manganese	0.43	+/-0.5	U	1.00	P	10/07/2025	18:22	LB137449
	Nickel	0.27	+/-0.5	U	1.00	P	10/07/2025	18:22	LB137449
	Potassium	36.4	+/-250	U	500	P	10/07/2025	18:22	LB137449
	Selenium	2.90	+/-2.5	U	5.00	P	10/07/2025	18:22	LB137449

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Tetra Tech, EMI

SDG No.: Q3291

Contract: TETR16

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB02	Silver	0.17	+/-0.5	J	1.00	P	10/07/2025	18:22	LB137449
	Sodium	146	+/-250	J	500	P	10/07/2025	18:22	LB137449
	Thallium	0.090	+/-0.5	J	1.00	P	10/07/2025	18:22	LB137449
	Vanadium	0.077	+/-2.5	U	5.00	P	10/07/2025	18:22	LB137449
	Zinc	1.25	+/-2.5	U	5.00	P	10/07/2025	18:22	LB137449
CCB03	Aluminum	1.94	+/-10	U	20.0	P	10/07/2025	19:09	LB137449
	Antimony	0.11	+/-1	U	2.00	P	10/07/2025	19:09	LB137449
	Arsenic	0.089	+/-0.5	U	1.00	P	10/07/2025	19:09	LB137449
	Barium	0.21	+/-5	U	10.0	P	10/07/2025	19:09	LB137449
	Beryllium	0.32	+/-0.5	U	1.00	P	10/07/2025	19:09	LB137449
	Cadmium	0.34	+/-0.5	U	1.00	P	10/07/2025	19:09	LB137449
	Calcium	45.7	+/-250	U	500	P	10/07/2025	19:09	LB137449
	Chromium	0.21	+/-1	U	2.00	P	10/07/2025	19:09	LB137449
	Cobalt	0.070	+/-0.5	U	1.00	P	10/07/2025	19:09	LB137449
	Copper	0.30	+/-1	U	2.00	P	10/07/2025	19:09	LB137449
	Iron	7.81	+/-25	U	50.0	P	10/07/2025	19:09	LB137449
	Lead	0.21	+/-0.5	U	1.00	P	10/07/2025	19:09	LB137449
	Magnesium	19.5	+/-250	U	500	P	10/07/2025	19:09	LB137449
	Manganese	0.43	+/-0.5	U	1.00	P	10/07/2025	19:09	LB137449
	Nickel	0.27	+/-0.5	U	1.00	P	10/07/2025	19:09	LB137449
	Potassium	36.4	+/-250	U	500	P	10/07/2025	19:09	LB137449
	Selenium	2.90	+/-2.5	U	5.00	P	10/07/2025	19:09	LB137449
	Silver	0.060	+/-0.5	U	1.00	P	10/07/2025	19:09	LB137449
	Sodium	128	+/-250	U	500	P	10/07/2025	19:09	LB137449
	Thallium	0.060	+/-0.5	U	1.00	P	10/07/2025	19:09	LB137449
	Vanadium	0.077	+/-2.5	U	5.00	P	10/07/2025	19:09	LB137449
	Zinc	1.25	+/-2.5	U	5.00	P	10/07/2025	19:09	LB137449