

Metals

- 7 -

LABORATORY CONTROL SAMPLE SUMMARY

Client: Tetra Tech, EMI

SDG No.: Q3291

Contract: TETR16

Lab Code: ACE

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
PB170000BS							
Aluminum	ug/L	10000	10000		100	80 - 120	P
Antimony	ug/L	500	532		106	80 - 120	P
Arsenic	ug/L	500	503		101	80 - 120	P
Barium	ug/L	2500	2600		104	80 - 120	P
Beryllium	ug/L	500	513		103	80 - 120	P
Cadmium	ug/L	500	531		106	80 - 120	P
Calcium	ug/L	50000	48800		98	80 - 120	P
Chromium	ug/L	500	506		101	80 - 120	P
Cobalt	ug/L	500	511		102	80 - 120	P
Copper	ug/L	5000	5110		102	80 - 120	P
Iron	ug/L	25000	27000		108	80 - 120	P
Lead	ug/L	2500	2540		102	80 - 120	P
Magnesium	ug/L	50000	49000		98	80 - 120	P
Manganese	ug/L	5000	5100		102	80 - 120	P
Nickel	ug/L	500	511		102	80 - 120	P
Potassium	ug/L	25000	25400		102	80 - 120	P
Selenium	ug/L	500	518		104	80 - 120	P
Silver	ug/L	500	538		108	80 - 120	P
Sodium	ug/L	50000	50600		101	80 - 120	P
Thallium	ug/L	500	507		101	80 - 120	P
Vanadium	ug/L	500	502		100	80 - 120	P
Zinc	ug/L	5000	5100		102	80 - 120	P

Metals

- 7 -

LABORATORY CONTROL SAMPLE SUMMARY

Client: Tetra Tech, EMI

SDG No.: Q3291

Contract: TETR16

Lab Code: ACE

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
PB170005BS Mercury	ug/L	4.0	3.79		95	80 - 120	CV