

Duplicate Sample Summary

Client:	Tetra Tech NUS, Inc.	SDG No.:	Q3292
Project:	NWIRP Bethpage 112G08005-WE13	Sample ID:	Q3281-02
Client ID:	002-35TH-AVE(OCT)DUP	Percent Solids for Spike Sample:	0

Analyte	Units	Acceptance Limit	Sample Result	Conc. Qualifier	Duplicate Result	Conc. Qualifier	Dilution Factor	RPD/ AD	Qual	Analysis Date
TSS	mg/L	+/-5	26.4		26.2		1	0.76		10/08/2025

Duplicate Sample Summary

Client: Tetra Tech NUS, Inc.	SDG No.: Q3292
Project: NWIRP Bethpage 112G08005-WE13	Sample ID: Q3292-01
Client ID: RW8-SP100-20251003DUP	Percent Solids for Spike Sample: 0

Analyte	Units	Acceptance Limit	Sample Result	Conc. Qualifier	Duplicate Result	Conc. Qualifier	Dilution Factor	RPD/ AD	Qual	Analysis Date
TDS	mg/L	+/-5	4.00	J	4.00	J	1	0		10/06/2025