

Laboratory Control Sample Summary

Client: Tetra Tech NUS, Inc.

SDG No.: Q3292

Project: NWIRP Bethpage 112G08005-WE13

Run No.: LB137440

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB137440BS							
TDS	mg/L	100	95.0		95	1	90-110	10/06/2025

Laboratory Control Sample Summary

Client: Tetra Tech NUS, Inc.

SDG No.: Q3292

Project: NWIRP Bethpage 112G08005-WE13

Run No.: LB137455

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB137455BS							
TSS	mg/L	550	531		96	1	90-110	10/08/2025