

Cover Page

Order ID : Q3292

Project ID : NWIRP Bethpage 112G08005-WE13

Client : Tetra Tech NUS, Inc.

Lab Sample Number

Q3292-01
Q3292-02

Client Sample Number

RW8-SP100-20251003
RW8-SP303-20251003

I certify that the data package is in compliance with the terms and conditions of the contract, both technically and for completeness, for other than the conditions detailed above. Release of the data contained in this hard copy data package has been authorized by the laboratory manager or his designee, as verified by the following signature.

Signature : _____

Date: 10/15/2025

NYDOH CERTIFICATION NO - 11376

NJDEP CERTIFICATION NO - 20012

CASE NARRATIVE

Tetra Tech NUS, Inc.

Project Name: NWIRP Bethpage 112G08005-WE13

Project Manager : Ernie Wu

Order ID # Q3292

Test Name: SVOC-SIMGroup1

A. Number of Samples and Date of Receipt:

2 Water samples were received on 10/06/2025.

B. Parameters

According to the Chain of Custody document, the following analyses were requested: SVOC-SIMGroup1. This data package contains results for SVOC-SIMGroup1.

C. Analytical Techniques:

The samples were analyzed on instrument BNA_N using GC Column ZB-SemiVolatiles Guardian which is 30 meters, 0.25 mm ID, 0.5 um df, Catalog # 7HG-G027-17-GGA. The analysis of SVOC-SIMGroup1 was based on method 8270-Modified and extraction was done based on method 3510.

D. QA/ QC Samples:

The Holding Times were met for all analysis.

The Surrogate recoveries were met for all analysis except for, RW8-SP100-20251003 [Terphenyl-d14 - 143%], BP-VPB-184-GW-740-742MS [2-Fluorobiphenyl - 183%, Terphenyl-d14 - 269%] and BP-VPB-184-GW-740-742MSD [2-Fluorobiphenyl - 218%, Terphenyl-d14 - 553%]. Failed surrogates are not associated with DOD, therefor no further corrective action was taken.

The Internal Standards Areas were met for all analysis except for PB170022BL, PB170022BS, BP-VPB-184-GW-740-742MS and BP-VPB-184-GW-740-742MSD. Failed internal standard is not associated with DOD, therefor no further corrective action was taken.

The Retention Times were met for all analysis.

The MS {Q3298-06MS} with File ID: BN038086.D recoveries met the requirements for all compounds except for 1,4-Dioxane[168%], Recovery failed due to matrix interference, therefor no further corrective action was taken.

The MSD {Q3298-07MSD} with File ID: BN038087.D recoveries met the requirements for all compounds except for 1,4-Dioxane[173%], Recovery failed due to matrix interference , therefor no further corrective action was taken.

The RPD were met for all analysis.

The Blank Spike met requirements for all compounds.

The Blank analysis did not indicate the presence of lab contamination.

The %RSD is greater than 20% in the Method 8270-SIM-BN092325.M for 4,6-Dinitro-2-methylphenol. This is passing on Quadratic regression.Compound#24 removed from 5.0 ppm.

The Continuous Calibration File ID BN038066.D met the requirements except for Nitrobenzene-d5 and Phenol-d6. Failed surrogates are not associated with DOD, Therefor no further corrective action was taken.

The Continuous Calibration File ID BN038083.D met the requirements except for 2,4,6-Tribromophenol and Phenol-d6. Failed surrogates are not associated with DOD, Therefor no further corrective action was taken.

The Tuning criteria met requirements.

E. Additional Comments:

The Form 6 is not included in the data package because the Initial Calibration was performed using 7 points.

The not QT review data is reported in the Miscellaneous.

The laboratory certifies that the all-electronic diskette deliverable exactly match the data Summary forms (i.e. Form Is)."

F. Manual Integration Comments:

Please refer to the Manual integration Report included with the Run Logs for information on the manual integrations performed.

I certify that the data package is in compliance with the terms and conditions of the contract, both technically and for completeness, for other than the conditions detailed above. The laboratory manager or his designee, as verified by the following signature has authorized release of the data contained in this hard copy data package.

Signature_____

DATA REPORTING QUALIFIERS- ORGANIC

For reporting results, the following “ Results Qualifiers” are used:

Value	If the result is a value greater than or equal to the detection limit, report the value
U	Indicates the compound was analyzed for but was not detected. Report the minimum detection limit for the sample with the U, i.e. “10 U”. This is not necessarily the instrument detection limit attainable for this particular sample based on any concentration or dilution that may have been required.
ND	Indicates the analyte was analyzed for, but not detected
J	Indicates an estimated value. This flag is used: (1) When estimating a concentration for a tentatively identified compound (library search hits, where a 1:1 response is assumed.) (2) When the mass spectral data indicated the identification, however the result was less than the specified detection limit greater than zero. If the detection limit was 10ug/L and a concentration of 3 ug/L was calculated report as 3 J. This flag is used when similar situation arise on any organic parameter i.e. Pest, PCB and others.
B	Indicates the analyte was found in the blank as well as the sample report as “12 B”.
E	Indicates the analyte ‘s concentration exceeds the calibrated range of the instrument for that specific analysis.
D	This flag identifies all compounds identified in an analysis at a secondary dilution factor.
P	This flag is used for Pesticide/PCB target analyte when there is >25% difference for detected concentrations between the two GC columns. The lower of the two values is reported on Form 1 and flagged with a “P”.
N	This flag indicates presumptive evidence of a compound. This is only used for tentatively identified compounds (TICs), where the identification is based on a mass spectral library search. It applies to all TIC results. For generic characterization of a TIC, such as chlorinated hydrocarbon, the flag is not used.
A	This flag indicates that a Tentatively Identified Compound is a suspected aldol-condensation product.
Q	Indicates the LCS did not meet the control limits requirements

GC/MS SEMI-VOLATILE ANALYSIS CONFORMANCE/NON-CONFORMANCE SUMMARY

ORDER ID: Q3292

MATRIX: Water

METHOD: 8270-Modified/3510

		NA	NO	YES
1.	Chromatograms Labeled/Compounds Identified. (Field samples and Method Blanks)			✓
2.	GC/MS Tuning Specifications. DFTPP Meet Criteria. (NOTE THAT THERE ARE DIFFERENT CRITERIA FOR NY ASP CLP, CLP AND NJ)			✓
3.	GC/MS Tuning Frequency - Performed every 24 hours for 600 series and 12 hours for 8000 Series.			✓
4.	GC/MS Calibration - Initial Calibration performed within 30 days before sample analysis and continuing calibration performed within 24 hours of sample analysis for 600 series and 12 hours for 8000 series.			✓
5.	GC/MS Calibration Requirements.		✓	
The %RSD is greater than 20% in the Method 8270-SIM-BN092325.M for 4,6-Dinitro-2-methylphenol. This is passing on Quadratic regression.Compound#24 removed from 5.0 ppm. The Continuous Calibration File ID BN038066.D met the requirements except for Nitrobenzene-d5 and Phenol-d6. Failed surrogates are not associated with DOD, Therefor no further corrective action was taken. The Continuous Calibration File ID BN038083.D met the requirements except for 2,4,6-Tribromophenol and Phenol-d6. Failed surrogates are not associated with DOD, Therefor no further corrective action was taken.				
6.	Blank Contamination - If yes, list compounds and concentrations in each blank:			✓

ALLIANCE 284 Sheffield Street, Mountainside New Jersey 07092

NEW JERSEY LAB ID#: 20012; NEW YORK LAB ID#: 11376

GC/MS SEMI-VOLATILE ANALYSIS CONFORMANCE/NON-CONFORMANCE SUMMARY**(CONTINUED)**

		NA	NO	YES
7.	Surrogate Recoveries Meet Criteria		✓	
	If not met, list those compounds and their recoveries which fall outside the acceptable ranges.			
	The Surrogate recoveries were met for all analysis except for, RW8-SP100-20251003 [Terphenyl-d14 - 143%], BP-VPB-184-GW-740-742MS [2-Fluorobiphenyl - 183%, Terphenyl-d14 - 269%] and BP-VPB-184-GW-740-742MSD [2-Fluorobiphenyl - 218%, Terphenyl-d14 - 553%]. Failed surrogates are not associated with DOD, therefor no further corrective action was taken.			
8.	Matrix Spike/Matrix Spike Duplicate Recoveries Meet Criteria		✓	
	If not met, list those compounds and their recoveries which fall outside the acceptable range.			
	The MS {Q3298-06MS} with File ID: BN038086.D recoveries met the requirements for all compounds except for 1,4-Dioxane[168%], Recovery failed due to matrix interference , therefor no further corrective action was taken.			
	The MSD {Q3298-07MSD} with File ID: BN038087.D recoveries met the requirements for all compounds except for 1,4-Dioxane[173%], Recovery failed due to matrix interference , therefor no further corrective action was taken.			
	The Blank Spike met requirements for all compounds.			
9.	Internal Standard Area/Retention Time Shift Meet Criteria		✓	
	Comments: The Internal Standards Areas were met for all analysis except for PB170022BL, PB170022BS, BP-VPB-184-GW-740-742MS and BP-VPB-184-GW-740-742MSD. Failed internal standard is not associated with DOD, therefor no further corrective action was taken.			
10.	Extraction Holding Time Met			✓
	If not met, list number of days exceeded for each sample:			
11.	Analysis Holding Time Met			✓
	If not met, list number of days exceeded for each sample:			

ALLIANCE 284 Sheffield Street, Mountainside New Jersey 07092

NEW JERSEY LAB ID#: 20012; NEW YORK LAB ID#: 11376

GC/MS SEMI-VOLATILE ANALYSIS CONFORMANCE/NON-CONFORMANCE SUMMARY

(CONTINUED)

NA NO YES

ADDITIONAL COMMENTS:

The Form 6 is not included in the data package because the Initial Calibration was performed using 7 points.

The not QT review data is reported in the Miscellaneous.

The laboratory certifies that the all-electronic diskette deliverable exactly match the data Summary forms (i.e. Form Is)."

QA REVIEW

Date

APPENDIX A

QA REVIEW GENERAL DOCUMENTATION

Project #: Q3292

Completed

For thorough review, the report must have the following:

GENERAL:

Are all original paperwork present (chain of custody, record of communication,airbill, sample management lab chronicle, login page)

✓

Check chain-of-custody for proper relinquish/return of samples

✓

Is the chain of custody signed and complete

✓

Check internal chain-of-custody for proper relinquish/return of samples /sample extracts

✓

Collect information for each project id from server. Were all requirements followed

✓

COVER PAGE:

Do numbers of samples correspond to the number of samples in the Chain of Custody on login page

✓

Do lab numbers and client Ids on cover page agree with the Chain of Custody

✓

CHAIN OF CUSTODY:

Do requested analyses on Chain of Custody agree with form I results

✓

Do requested analyses on Chain of Custody agree with the log-in page

✓

Were the correct method log-in for analysis according to the Analytical Request and Chain of Custody

✓

Were the samples received within hold time

✓

Were any problems found with the samples at arrival recorded in the Sample Management Laboratory Chronicle

✓

ANALYTICAL:

Was method requirement followed?

✓

Was client requirement followed?

✓

Does the case narrative summarize all QC failure?

✓

All runlogs and manual integration are reviewed for requirements

✓

All manual calculations and /or hand notations verified

✓

QA Review Signature: SOHIL JODHANI

Date: 10/15/2025

LAB CHRONICLE

OrderID:	Q3292	OrderDate:	10/6/2025 10:37:00 AM
Client:	Tetra Tech NUS, Inc.	Project:	NWIRP Bethpage 112G08005-WE13
Contact:	Ernie Wu	Location:	D31

LabID	ClientID	Matrix	Test	Method	Sample Date	Prep Date	Anal Date	Received
Q3292-01	RW8-SP100-2025100 3	Water			10/03/25			10/06/25
			SVOC-SIMGroup1	8270-Modified		10/08/25	10/13/25	
Q3292-02	RW8-SP303-2025100 3	Water			10/03/25			10/06/25
			SVOC-SIMGroup1	8270-Modified		10/08/25	10/13/25	



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Hit Summary Sheet
SW-846

SDG No.: Q3292

Client: Tetra Tech NUS, Inc.

Sample ID	Client ID	Parameter	Concentration	C	MDL	LOD	RDL	Units
Client ID : RW8-SP100-20251003								
Q3292-01	RW8-SP100-20251003	WATER	1,4-Dioxane	0.120	J	0.07	0.2	0.2 ug/L
Total Svoc :				0.12				
Total Concentration:				0.12				



QC SUMMARY

Surrogate Summary

SW-846

SDG No.: Q3292

Client: Tetra Tech NUS, Inc.

Analytical Method: 8270-Modified

Lab Sample ID	Client ID	Parameter	Spike (PPM)	Result (PPM)	Recovery (%)	Qual	Limits (%)	
							Low	High
PB170022BL	PB170022BL	2-Methylnaphthalene-d10	0.4	0.32	80		30	150
		Fluoranthene-d10	0.4	0.35	88		30	150
		Nitrobenzene-d5	0.4	0.37	92		55	111
		2-Fluorobiphenyl	0.4	0.38	95		53	106
		Terphenyl-d14	0.4	0.49	123		58	132
PB170022BS	PB170022BS	2-Methylnaphthalene-d10	0.4	0.36	89		30	150
		Fluoranthene-d10	0.4	0.32	80		30	150
		Nitrobenzene-d5	0.4	0.37	92		55	111
		2-Fluorobiphenyl	0.4	0.41	103		53	106
		Terphenyl-d14	0.4	0.49	122		58	132
Q3292-01	RW8-SP100-20251003	2-Methylnaphthalene-d10	0.4	0.32	80		30	150
		Fluoranthene-d10	0.4	0.44	110		30	150
		Nitrobenzene-d5	0.4	0.37	91		55	111
		2-Fluorobiphenyl	0.4	0.39	96		53	106
		Terphenyl-d14	0.4	0.57	143	*	58	132
Q3292-02	RW8-SP303-20251003	2-Methylnaphthalene-d10	0.4	0.34	85		30	150
		Fluoranthene-d10	0.4	0.42	104		30	150
		Nitrobenzene-d5	0.4	0.37	92		55	111
		2-Fluorobiphenyl	0.4	0.37	92		53	106
		Terphenyl-d14	0.4	0.52	131		58	132
Q3298-06MS	BP-VPB-184-GW-740-742MS	2-Methylnaphthalene-d10	0.4	0.26	64		30	150
		Fluoranthene-d10	0.4	0.19	48		30	150
		Nitrobenzene-d5	0.4	0.29	72		55	111
		2-Fluorobiphenyl	0.4	0.73	183	*	53	106
		Terphenyl-d14	0.4	1.08	269	*	58	132
Q3298-07MSD	BP-VPB-184-GW-740-742MSD	2-Methylnaphthalene-d10	0.4	0.26	65		30	150
		Fluoranthene-d10	0.4	0.14	36		30	150
		Nitrobenzene-d5	0.4	0.32	79		55	111
		2-Fluorobiphenyl	0.4	0.87	218	*	53	106
		Terphenyl-d14	0.4	2.21	553	*	58	132

Matrix Spike/Matrix Spike Duplicate Summary

SW-846

SDG No.: Q3292

Analytical Method: SW8270-Modified

Client: Tetra Tech NUS, Inc.

DataFile: BN038086.D

Parameter	Spike	Sample Result	Result	Units	Rec	Rec Qual	RPD	RPD Qual	Low	Limits High	RPD
Lab Sample ID:	Q3298-06MS	Client Sample ID:	BP-VPB-184-GW-740-742MS								
1,4-Dioxane	0.95	0	1.60	ug/L	168	*			70	130	

Matrix Spike/Matrix Spike Duplicate Summary

SW-846

SDG No.: Q3292

Analytical Method: SW8270-Modified

Client: Tetra Tech NUS, Inc.

DataFile: BN038087.D

Parameter	Spike	Sample Result	Result	Units	Rec	Rec Qual	RPD	RPD Qual	Low	Limits High	RPD
Lab Sample ID:	Q3298-07MSD	Client Sample ID:	BP-VPB-184-GW-740-742MSD								
1,4-Dioxane	1.1	0	1.90	ug/L	173	*	3		70	130	20

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: Q3292

Analytical Method: 8270-Modified

Client: Tetra Tech NUS, Inc.

DataFile: BN038071.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	RPD	Limits		RPD
								Qual	Low	High	
PB170022BS	1,4-Dioxane	0.4	0.35	ug/L	88				70	130	

4B

SEMIVOLATILE METHOD BLANK SUMMARY

Client ID

PB170022BL

Lab Name: Alliance

Contract: TETR06

Lab Code: ACE

SDG NO.: Q3292

Lab File ID: BN038070.D

Lab Sample ID: PB170022BL

Instrument ID: BNA_N

Date Extracted: 10/08/2025

Matrix: (soil/water) Water

Date Analyzed: 10/13/2025

Level: (low/med) LOW

Time Analyzed: 16:12

THIS METHOD BLANK APPLIES TO THE FOLLOWING SAMPLES, MS AND MSD:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED
PB170022BS	PB170022BS	BN038071.D	10/13/2025
RW8-SP100-20251003	Q3292-01	BN038072.D	10/13/2025
RW8-SP303-20251003	Q3292-02	BN038073.D	10/13/2025
BP-VPB-184-GW-740-742MS	Q3298-06MS	BN038086.D	10/14/2025
BP-VPB-184-GW-740-742MSD	Q3298-07MSD	BN038087.D	10/14/2025

COMMENTS: _____



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

5B

SEMIVOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
DECAFLUOROTRIPHENYLPHOSPHINE (DFTPP)

Lab Name: Alliance
Lab Code: ACE
Lab File ID: BN037860.D
Instrument ID: BNA_N

Contract: TETR06
SDG NO.: Q3292
DFTPP Injection Date: 09/23/2025
DFTPP Injection Time: 11:33

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
68	Less than 2.0% of mass 69	0.7 (1.8) 1
69	Mass 69 relative abundance	100
70	Less than 2.0% of mass 69	0.2 (0.6) 1
197	Less than 2.0% of mass 198	0.0
198	Base Peak, 100% relative abundance	100
199	5.0 to 9.0% of mass 198	6.9
365	Greater than 1% of mass 198	4.2
441	Present, but less than mass 443	72.6
442	Greater than 50% of mass 198	100
443	15.0 - 24.0% of mass 442	17.7 (20.8) 2

1-Value is % mass 69

2-Value is % mass 442

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
SSTDICC0.1	SSTDICC0.1	BN037861.D	09/23/2025	12:13
SSTDICC0.2	SSTDICC0.2	BN037862.D	09/23/2025	12:49
SSTDICCC0.4	SSTDICCC0.4	BN037863.D	09/23/2025	13:25
SSTDICC0.8	SSTDICC0.8	BN037864.D	09/23/2025	14:02
SSTDICC1.6	SSTDICC1.6	BN037865.D	09/23/2025	14:38
SSTDICC3.2	SSTDICC3.2	BN037866.D	09/23/2025	15:15
SSTDICC5.0	SSTDICC5.0	BN037867.D	09/23/2025	15:51



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

5B

SEMIVOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
DECAFLUOROTRIPHENYLPHOSPHINE (DFTPP)

Lab Name: Alliance
Lab Code: ACE
Lab File ID: BN038065.D
Instrument ID: BNA_N

Contract: TETR06
SDG NO.: Q3292
DFTPP Injection Date: 10/13/2025
DFTPP Injection Time: 13:08

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
68	Less than 2.0% of mass 69	0.5 (1.2) 1
69	Mass 69 relative abundance	100
70	Less than 2.0% of mass 69	0.2 (0.5) 1
197	Less than 2.0% of mass 198	0.2
198	Base Peak, 100% relative abundance	100
199	5.0 to 9.0% of mass 198	6.7
365	Greater than 1% of mass 198	3.4
441	Present, but less than mass 443	70.3
442	Greater than 50% of mass 198	100
443	15.0 - 24.0% of mass 442	19.2 (21) 2

1-Value is % mass 69

2-Value is % mass 442

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
SSTDCCC0.4	SSTDCCC0.4	BN038066.D	10/13/2025	13:47
PB170022BL	PB170022BL	BN038070.D	10/13/2025	16:12
PB170022BS	PB170022BS	BN038071.D	10/13/2025	16:48
RW8-SP100-20251003	Q3292-01	BN038072.D	10/13/2025	17:24
RW8-SP303-20251003	Q3292-02	BN038073.D	10/13/2025	18:00
SSTDCCC0.4EC	SSTDCCC0.4	BN038081.D	10/13/2025	23:24



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

5B

SEMIVOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
DECAFLUOROTRIPHENYLPHOSPHINE (DFTPP)

Lab Name: Alliance
Lab Code: ACE
Lab File ID: BN038082.D
Instrument ID: BNA_N

Contract: TETR06
SDG NO.: Q3292
DFTPP Injection Date: 10/14/2025
DFTPP Injection Time: 00:40

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
68	Less than 2.0% of mass 69	0.7 (1.7) 1
69	Mass 69 relative abundance	100
70	Less than 2.0% of mass 69	0.2 (0.5) 1
197	Less than 2.0% of mass 198	0.4
198	Base Peak, 100% relative abundance	100
199	5.0 to 9.0% of mass 198	6.8
365	Greater than 1% of mass 198	3.5
441	Present, but less than mass 443	83.1
442	Greater than 50% of mass 198	100
443	15.0 - 24.0% of mass 442	17 (18) 2

1-Value is % mass 69

2-Value is % mass 442

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
SSTDCCC0.4	SSTDCCC0.4	BN038083.D	10/14/2025	01:19
BP-VPB-184-GW-740-742MS	Q3298-06MS	BN038086.D	10/14/2025	03:43
BP-VPB-184-GW-740-742MSD	Q3298-07MSD	BN038087.D	10/14/2025	04:19
SSTDCCC0.4EC	SSTDCCC0.4	BN038088.D	10/14/2025	04:56

8B

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance

Lab Code: ACE

SDG NO.: Q3292

Client ID : SSTDCCC0.4

Date Analyzed: 10/13/2025

Lab File ID: BN038066.D

Time Analyzed: 13:47

Instrument ID: BNA_N

GC Column: ZB-GR ID: 0.25 (mm)

	IS1 (DCB) AREA #	RT #	IS2 (NPT) AREA #	RT #	IS3 (ANT) AREA #	RT #
12 HOUR STD	7166	7.797	18651	10.59	10244	14.45
UPPER LIMIT	14332	8.297	37302	11.094	20488	14.952
LOWER LIMIT	3583	7.297	9325.5	10.094	5122	13.952
EPA SAMPLE NO.						
01 PB170022BL	6807	7.80	16619	10.61	7880	14.45
02 PB170022BS	6425	7.80	16580	10.59	7629	14.45
03 RW8-SP100-20251003	5762	7.80	13937	10.61	7066	14.45
04 RW8-SP303-20251003	6016	7.80	14258	10.61	7203	14.45

IS1 (DCB) = 1,4-Dichlorobenzene-d4

IS2 (NPT) = Naphthalene-d8

IS3 (ANT) = Acenaphthene-d10

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.

8C

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance

Lab Code: ACE

SDG NO.: Q3292

Client ID: SSTDCCC0.4

Date Analyzed: 10/13/2025

Lab File ID: BN038066.D

Time Analyzed: 13:47

Instrument ID: BNA_N

GC Column: ZB-GR ID: 0.25 (mm)

	IS4 (PHN) AREA #	RT #	IS5 (CRY) AREA #	RT #	IS6 (PRY) AREA #	RT #
12 HOUR STD	18901	17.198	16827	21.393	17326	23.712
UPPER LIMIT	37802	17.698	33654	21.893	34652	24.212
LOWER LIMIT	9450.5	16.698	8413.5	20.893	8663	23.212
EPA SAMPLE NO.						
01 PB170022BL	12265	17.21	8413 *	21.39	7507 *	23.72
02 PB170022BS	12726	17.20	8201 *	21.39	7302 *	23.71
03 RW8-SP100-20251003	12532	17.21	10826	21.39	10606	23.72
04 RW8-SP303-20251003	12433	17.21	10491	21.39	10411	23.72

IS4 (PHN) = Phenanthrene-d10

IS5 (CRY) = Chrysene-d12

IS6 (PRY) = Perylene-d12

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.

8B

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance

Lab Code: ACE

SDG NO.: Q3292

Client ID : SSTDCCC0.4

Date Analyzed: 10/14/2025

Lab File ID: BN038083.D

Time Analyzed: 01:19

Instrument ID: BNA_N

GC Column: ZB-GR ID: 0.25 (mm)

	IS1 (DCB) AREA #	RT #	IS2 (NPT) AREA #	RT #	IS3 (ANT) AREA #	RT #
12 HOUR STD	7061	7.797	18151	10.59	9016	14.45
UPPER LIMIT	14122	8.297	36302	11.094	18032	14.952
LOWER LIMIT	3530.5	7.297	9075.5	10.094	4508	13.952
EPA SAMPLE NO.						
01 BP-VPB-184-GW-740-742MS	5765	7.80	14968	10.59	3442 *	14.45
02 BP-VPB-184-GW-740-742MSD	6052	7.80	15943	10.59	3279 *	14.45

IS1 (DCB) = 1,4-Dichlorobenzene-d4

IS2 (NPT) = Naphthalene-d8

IS3 (ANT) = Acenaphthene-d10

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.

8C

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance

Lab Code: ACE

SDG NO.: Q3292

Client ID: SSTDCCC0.4

Date Analyzed: 10/14/2025

Lab File ID: BN038083.D

Time Analyzed: 01:19

Instrument ID: BNA_N

GC Column: ZB-GR ID: 0.25 (mm)

	IS4 (PHN) AREA #	RT #	IS5 (CRY) AREA #	RT #	IS6 (PRY) AREA #	RT #
12 HOUR STD	16225	17.198	13579	21.384	13727	23.709
UPPER LIMIT	32450	17.698	27158	21.884	27454	24.209
LOWER LIMIT	8112.5	16.698	6789.5	20.884	6863.5	23.209
EPA SAMPLE NO.						
01 BP-VPB-184-GW-740-742MS	10343	17.20	3302 *	21.38	80 *	23.71
02 BP-VPB-184-GW-740-742MSD	9259	17.20	1395 *	21.38	59 *	23.71

IS4 (PHN) = Phenanthrene-d10

IS5 (CRY) = Chrysene-d12

IS6 (PRY) = Perylene-d12

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.



SAMPLE DATA



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Report of Analysis

Client:	Tetra Tech NUS, Inc.		Date Collected:	10/03/25
Project:	NWIRP Bethpage 112G08005-WE13		Date Received:	10/06/25
Client Sample ID:	RW8-SP100-20251003		SDG No.:	Q3292
Lab Sample ID:	Q3292-01		Matrix:	Water
Analytical Method:	SW8270ESIM	Level : LOW	% Solid:	0
Sample Wt/Vol:	1000 mL	Final Vol: 1000 uL	Test:	SVOC-SIMGroup1
Prep Method :	Prep Date: 10/08/25			

CAS Number	Parameter	Conc.	Qua.	DF	MDL	LOD	LOQ / CRQL	Units	Date Ana.	Prep BatchID
TARGETS										
123-91-1	1,4-Dioxane	0.12	J	1	0.070	0.20	0.20	ug/L	10/13/25 17:24	PB170022
SURROGATES										
7297-45-2	2-Methylnaphthalene-d10	0.32			30 - 150		80%	SPK: 0.4		
93951-69-0	Fluoranthene-d10	0.44			30 - 150		110%	SPK: 0.4		
4165-60-0	Nitrobenzene-d5	0.37			55 - 111		91%	SPK: 0.4		
321-60-8	2-Fluorobiphenyl	0.39			53 - 106		96%	SPK: 0.4		
1718-51-0	Terphenyl-d14	0.57	*		58 - 132		143%	SPK: 0.4		
INTERNAL STANDARDS										
		Area								
3855-82-1	1,4-Dichlorobenzene-d4	5760								
1146-65-2	Naphthalene-d8	13900								
15067-26-2	Acenaphthene-d10	7070								
1517-22-2	Phenanthrene-d10	12500								
1719-03-5	Chrysene-d12	10800								
1520-96-3	Perylene-d12	10600								

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN101325\
 Data File : BN038072.D
 Acq On : 13 Oct 2025 17:24
 Operator : RC/JU
 Sample : Q3292-01
 Misc :
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 RW8-SP100-20251003

Quant Time: Oct 13 17:53:47 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN092325.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Sep 24 11:00:01 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.797	152	5762	0.400	ng	-0.03
7) Naphthalene-d8	10.605	136	13937	0.400	ng	-0.02
13) Acenaphthene-d10	14.452	164	7066	0.400	ng	-0.02
19) Phenanthrene-d10	17.210	188	12532	0.400	ng	-0.01
29) Chrysene-d12	21.393	240	10826	0.400	ng	# 0.00
35) Perylene-d12	23.715	264	10606	0.400	ng	#-0.01
System Monitoring Compounds						
4) 2-Fluorophenol	5.370	112	163	0.012	ng	-0.01
5) Phenol-d6	6.959	99	565	0.033	ng	-0.01
8) Nitrobenzene-d5	8.950	82	3889	0.366	ng	-0.02
11) 2-Methylnaphthalene-d10	12.197	152	6174	0.319	ng	-0.02
14) 2,4,6-Tribromophenol	15.957	330	16	0.006	ng	-0.01
15) 2-Fluorobiphenyl	13.073	172	11132	0.385	ng	-0.02
27) Fluoranthene-d10	19.238	212	13875	0.440	ng	-0.01
31) Terphenyl-d14	19.838	244	12328	0.572	ng	-0.01
Target Compounds						
2) 1,4-Dioxane	3.283	88	705	0.121	ng	# 77
34) Bis(2-ethylhexyl)phtha...	21.313	149	2809	0.188	ng	# 97


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN101325\
Data File : BN038072.D
Acq On : 13 Oct 2025 17:24
Operator : RC/JU
Sample : Q3292-01
Misc :
ALS Vial : 8 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
RW8-SP100-20251003


Quant Time: Oct 13 17:53:47 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN092325.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Sep 24 11:00:01 2025
Response via : Initial Calibration








#1
1,4-Dichlorobenzene-d4
Concen: 0.400 ng
RT: 7.797 min Scan# 61
Delta R.T. -0.029 min
Lab File: BN038072.D
Acq: 13 Oct 2025 17:24

Instrument :
BNA_N
ClientSampleId :
RW8-SP100-20251003







Tgt Ion: 152 Resp: 5762
Ion Ratio Lower Upper
152 100
150 155.4 121.0 181.4
115 62.1 47.4 71.0



#2
1,4-Dioxane
Concen: 0.121 ng
RT: 3.283 min Scan# 28
Delta R.T. -0.015 min
Lab File: BN038072.D
Acq: 13 Oct 2025 17:24

Tgt Ion: 88 Resp: 705
Ion Ratio Lower Upper
88 100
43 0.0 26.6 39.8#
58 68.2 58.9 88.3









#4
2-Fluorophenol
Concen: 0.012 ng
RT: 5.370 min Scan# 31
Delta R.T. -0.015 min
Lab File: BN038072.D
Acq: 13 Oct 2025 17:24




Instrument :
BNA_N
ClientSampleId :
RW8-SP100-20251003


Tgt Ion:112 Resp: 163
Ion Ratio Lower Upper
112 100
64 71.8 44.6 66.8#
63 61.3 24.9 37.3#



#5
Phenol-d6
Concen: 0.033 ng
RT: 6.959 min Scan# 537
Delta R.T. -0.015 min
Lab File: BN038072.D
Acq: 13 Oct 2025 17:24

Tgt Ion: 99 Resp: 565
Ion Ratio Lower Upper
99 100
42 27.4 17.2 25.8#
71 34.2 27.3 40.9









#7
Naphthalene-d8
Concen: 0.400 ng
RT: 10.605 min Scan# 941
Delta R.T. -0.021 min
Lab File: BN038072.D
Acq: 13 Oct 2025 17:24




Instrument :
BNA_N
ClientSampleId :
RW8-SP100-20251003


Tgt Ion:	136	Resp:	13937
Ion	Ratio	Lower	Upper
136	100		
137	11.6	9.0	13.4
54	6.7	5.8	8.8
68	5.4	4.3	6.5



#8
Nitrobenzene-d5
Concen: 0.366 ng
RT: 8.950 min Scan# 784
Delta R.T. -0.021 min
Lab File: BN038072.D
Acq: 13 Oct 2025 17:24

Tgt Ion:	82	Resp:	3889
Ion	Ratio	Lower	Upper
82	100		
128	42.9	34.8	52.2
54	54.7	42.2	63.2









#11
2-Methylnaphthalene-d10
Concen: 0.319 ng
RT: 12.197 min Scan# 1163
Delta R.T. -0.020 min
Lab File: BN038072.D
Acq: 13 Oct 2025 17:24




Instrument :
BNA_N
ClientSampleId :
RW8-SP100-20251003


Tgt Ion:152 Resp: 6174
Ion Ratio Lower Upper
152 100
151 21.3 17.4 26.2



#13
Acenaphthene-d10
Concen: 0.400 ng
RT: 14.452 min Scan# 1454
Delta R.T. -0.021 min
Lab File: BN038072.D
Acq: 13 Oct 2025 17:24

Tgt Ion:164 Resp: 7066
Ion Ratio Lower Upper
164 100
162 105.4 84.8 127.2
160 57.7 46.1 69.1









#14
2,4,6-Tribromophenol
Concen: 0.006 ng
RT: 15.957 min Scan# 1589
Delta R.T. -0.012 min
Lab File: BN038072.D
Acq: 13 Oct 2025 17:24




Instrument :
BNA_N
ClientSampleId :
RW8-SP100-20251003


Tgt Ion: 330 Resp: 16
Ion Ratio Lower Upper
330 100
332 112.5 77.2 115.8
141 0.0 37.2 55.8#



#15
2-Fluorobiphenyl
Concen: 0.385 ng
RT: 13.073 min Scan# 1325
Delta R.T. -0.021 min
Lab File: BN038072.D
Acq: 13 Oct 2025 17:24

Tgt Ion: 172 Resp: 11132
Ion Ratio Lower Upper
172 100
171 36.1 28.6 43.0
170 25.4 19.7 29.5









#19
Phenanthrene-d10
Concen: 0.400 ng
RT: 17.210 min Scan# 1691
Delta R.T. -0.013 min
Lab File: BN038072.D
Acq: 13 Oct 2025 17:24




Instrument :
BNA_N
ClientSampleId :
RW8-SP100-20251003


Tgt Ion:	188	Resp:	12532
Ion Ratio	Lower	Upper	
188	100		
94	0.0	0.0	0.0
80	9.1	7.6	11.4



#27
Fluoranthene-d10
Concen: 0.440 ng
RT: 19.238 min Scan# 1951
Delta R.T. -0.014 min
Lab File: BN038072.D
Acq: 13 Oct 2025 17:24

Tgt Ion:	212	Resp:	13875
Ion Ratio	Lower	Upper	
212	100		
106	16.6	12.6	19.0
104	9.1	7.4	11.0





#29

Chrysene-d12

Concen: 0.400 ng

RT: 21.393 min Scan# 21

Delta R.T. -0.009 min

Lab File: BN038072.D


Acq: 13 Oct 2025 17:24

Instrument :

BNA_N

ClientSampleId :

RW8-SP100-20251003






Tgt Ion:240 Resp: 10826

Ion Ratio Lower Upper

240 100

120 18.3 11.4 17.2#

236 29.3 23.0 34.6



#31

Terphenyl-d14

Concen: 0.572 ng

RT: 19.838 min Scan# 2080

Delta R.T. -0.014 min

Lab File: BN038072.D

Acq: 13 Oct 2025 17:24




Tgt Ion:244 Resp: 12328


Ion Ratio Lower Upper

244 100

212 9.5 7.6 11.4

122 18.0 13.9 20.9









#34
Bis(2-ethylhexyl)phthalate
Concen: 0.188 ng
RT: 21.313 min Scan# 21
Delta R.T. -0.009 min
Lab File: BN038072.D
Acq: 13 Oct 2025 17:24




Instrument :
BNA_N
ClientSampleId :
RW8-SP100-20251003

Tgt Ion:149 Resp: 2809
Ion Ratio Lower Upper
149 100
167 23.8 20.4 30.6
279 4.0 2.4 3.6#



#35
Perylene-d12
Concen: 0.400 ng
RT: 23.715 min Scan# 2832
Delta R.T. -0.012 min
Lab File: BN038072.D
Acq: 13 Oct 2025 17:24

Tgt Ion:264 Resp: 10606
Ion Ratio Lower Upper
264 100
260 27.8 21.5 32.3
265 116.4 53.8 80.6#



Report of Analysis

Client: Tetra Tech NUS, Inc.	Date Collected: 10/03/25	
Project: NWIRP Bethpage 112G08005-WE13	Date Received: 10/06/25	
Client Sample ID: RW8-SP303-20251003	SDG No.: Q3292	
Lab Sample ID: Q3292-02	Matrix: Water	
Analytical Method: SW8270ESIM Level : LOW	% Solid: 0	
Sample Wt/Vol: 1000 mL Final Vol: 1000 uL	Test: SVOC-SIMGroup1	
Prep Method :	Prep Date: 10/08/25	

CAS Number	Parameter	Conc.	Qua.	DF	MDL	LOD	LOQ / CRQL	Units	Date Ana.	Prep BatchID
TARGETS										
123-91-1	1,4-Dioxane	0.20	U	1	0.070	0.20	0.20	ug/L	10/13/25 18:00	PB170022
SURROGATES										
7297-45-2	2-Methylnaphthalene-d10	0.34			30 - 150		85%	SPK: 0.4		
93951-69-0	Fluoranthene-d10	0.42			30 - 150		104%	SPK: 0.4		
4165-60-0	Nitrobenzene-d5	0.37			55 - 111		92%	SPK: 0.4		
321-60-8	2-Fluorobiphenyl	0.37			53 - 106		92%	SPK: 0.4		
1718-51-0	Terphenyl-d14	0.52			58 - 132		131%	SPK: 0.4		
INTERNAL STANDARDS										
		Area								
3855-82-1	1,4-Dichlorobenzene-d4	6020								
1146-65-2	Naphthalene-d8	14300								
15067-26-2	Acenaphthene-d10	7200								
1517-22-2	Phenanthrene-d10	12400								
1719-03-5	Chrysene-d12	10500								
1520-96-3	Perylene-d12	10400								

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN101325\
Data File : BN038073.D
Acq On : 13 Oct 2025 18:00
Operator : RC/JU
Sample : Q3292-02
Misc :
ALS Vial : 9 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
RW8-SP303-20251003

Quant Time: Oct 13 18:29:07 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN092325.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Sep 24 11:00:01 2025
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.797	152	6016	0.400	ng	-0.03
7) Naphthalene-d8	10.605	136	14258	0.400	ng	-0.02
13) Acenaphthene-d10	14.452	164	7203	0.400	ng	-0.02
19) Phenanthrene-d10	17.210	188	12433	0.400	ng	-0.01
29) Chrysene-d12	21.393	240	10491	0.400	ng	# 0.00
35) Perylene-d12	23.715	264	10411	0.400	ng	#-0.01
System Monitoring Compounds						
4) 2-Fluorophenol	5.370	112	2027	0.148	ng	-0.01
5) Phenol-d6	6.959	99	1235	0.068	ng	-0.01
8) Nitrobenzene-d5	8.950	82	3997	0.368	ng	-0.02
11) 2-Methylnaphthalene-d10	12.197	152	6712	0.339	ng	-0.02
14) 2,4,6-Tribromophenol	15.957	330	649	0.238	ng	-0.01
15) 2-Fluorobiphenyl	13.073	172	10881	0.369	ng	-0.02
27) Fluoranthene-d10	19.238	212	13072	0.418	ng	-0.01
31) Terphenyl-d14	19.838	244	10940	0.524	ng	-0.01
Target Compounds						Qvalue
34) Bis(2-ethylhexyl)phtha...	21.313	149	2111	0.146	ng	# 97


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN101325\
Data File : BN038073.D
Acq On : 13 Oct 2025 18:00
Operator : RC/JU
Sample : Q3292-02
Misc :
ALS Vial : 9 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
RW8-SP303-20251003


Quant Time: Oct 13 18:29:07 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN092325.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Sep 24 11:00:01 2025
Response via : Initial Calibration








#1
1,4-Dichlorobenzene-d4
Concen: 0.400 ng
RT: 7.797 min Scan# 61
Delta R.T. -0.029 min
Lab File: BN038073.D
Acq: 13 Oct 2025 18:00

Instrument :
BNA_N
ClientSampleId :
RW8-SP303-20251003







Tgt Ion:152 Resp: 6016
Ion Ratio Lower Upper
152 100
150 151.5 121.0 181.4
115 60.1 47.4 71.0



#4
2-Fluorophenol
Concen: 0.148 ng
RT: 5.370 min Scan# 317
Delta R.T. -0.015 min
Lab File: BN038073.D
Acq: 13 Oct 2025 18:00


Tgt Ion:112 Resp: 2027
Ion Ratio Lower Upper
112 100
64 57.5 44.6 66.8
63 31.9 24.9 37.3








#5
Phenol-d6
Concen: 0.068 ng
RT: 6.959 min Scan# 531
Delta R.T. -0.015 min
Lab File: BN038073.D
Acq: 13 Oct 2025 18:00

Instrument :
BNA_N
ClientSampleId :
RW8-SP303-20251003







Tgt Ion	Ratio	Lower	Upper
99	100		
42	26.1	17.2	25.8#
71	29.3	27.3	40.9



#7
Naphthalene-d8
Concen: 0.400 ng
RT: 10.605 min Scan# 939
Delta R.T. -0.021 min
Lab File: BN038073.D
Acq: 13 Oct 2025 18:00

Tgt Ion	Ratio	Lower	Upper
136	100		
137	11.7	9.0	13.4
54	6.6	5.8	8.8
68	5.5	4.3	6.5









#8
Nitrobenzene-d5
Concen: 0.368 ng
RT: 8.950 min Scan# 786
Delta R.T. -0.021 min
Lab File: BN038073.D
Acq: 13 Oct 2025 18:00




Instrument :
BNA_N
ClientSampleId :
RW8-SP303-20251003


Tgt Ion:	82	Resp:	3997
Ion	Ratio	Lower	Upper
82	100		
128	43.6	34.8	52.2
54	53.7	42.2	63.2



#11
2-Methylnaphthalene-d10
Concen: 0.339 ng
RT: 12.197 min Scan# 1159
Delta R.T. -0.020 min
Lab File: BN038073.D
Acq: 13 Oct 2025 18:00

Tgt Ion:	152	Resp:	6712
Ion	Ratio	Lower	Upper
152	100		
151	21.2	17.4	26.2









#13
Acenaphthene-d10
Concen: 0.400 ng
RT: 14.452 min Scan# 1456
Delta R.T. -0.021 min
Lab File: BN038073.D
Acq: 13 Oct 2025 18:00




Instrument :
BNA_N
ClientSampleId :
RW8-SP303-20251003


Tgt Ion:164 Resp: 7203
Ion Ratio Lower Upper
164 100
162 107.1 84.8 127.2
160 57.2 46.1 69.1



#14
2,4,6-Tribromophenol
Concen: 0.238 ng
RT: 15.957 min Scan# 1589
Delta R.T. -0.012 min
Lab File: BN038073.D
Acq: 13 Oct 2025 18:00

Tgt Ion:330 Resp: 649
Ion Ratio Lower Upper
330 100
332 95.4 77.2 115.8
141 46.4 37.2 55.8









#15
2-Fluorobiphenyl
Concen: 0.369 ng
RT: 13.073 min Scan# 1327
Delta R.T. -0.021 min
Lab File: BN038073.D
Acq: 13 Oct 2025 18:00




Instrument :
BNA_N
ClientSampleId :
RW8-SP303-20251003


Tgt Ion	Ratio	Lower	Upper
172	100		
171	36.4	28.6	43.0
170	25.0	19.7	29.5



#19
Phenanthrene-d10
Concen: 0.400 ng
RT: 17.210 min Scan# 1690
Delta R.T. -0.013 min
Lab File: BN038073.D
Acq: 13 Oct 2025 18:00

Tgt Ion	Ratio	Lower	Upper
188	100		
94	0.0	0.0	0.0
80	9.5	7.6	11.4









#27
Fluoranthene-d10
Concen: 0.418 ng
RT: 19.238 min Scan# 1951
Delta R.T. -0.014 min
Lab File: BN038073.D
Acq: 13 Oct 2025 18:00




Instrument :
BNA_N
ClientSampleId :
RW8-SP303-20251003


Tgt Ion	Ratio	Lower	Upper
212	100		
106	16.7	12.6	19.0
104	9.6	7.4	11.0



#29
Chrysene-d12
Concen: 0.400 ng
RT: 21.393 min Scan# 2271
Delta R.T. -0.009 min
Lab File: BN038073.D
Acq: 13 Oct 2025 18:00

Tgt Ion	Ratio	Lower	Upper
240	100		
120	17.5	11.4	17.2
236	30.0	23.0	34.6









#31
Terphenyl-d14
Concen: 0.524 ng
RT: 19.838 min Scan# 2080
Delta R.T. -0.014 min
Lab File: BN038073.D
Acq: 13 Oct 2025 18:00




Instrument :
BNA_N
ClientSampleId :
RW8-SP303-20251003


Tgt Ion	Ratio	Lower	Upper
244	100		
212	9.6	7.6	11.4
122	17.6	13.9	20.9



#34
Bis(2-ethylhexyl)phthalate
Concen: 0.146 ng
RT: 21.313 min Scan# 2262
Delta R.T. -0.009 min
Lab File: BN038073.D
Acq: 13 Oct 2025 18:00


Tgt Ion	Ratio	Lower	Upper
149	100		
167	24.3	20.4	30.6
279	7.0	2.4	3.6







#35
Perylene-d12
Concen: 0.400 ng
RT: 23.715 min Scan# 2832
Delta R.T. -0.012 min
Lab File: BN038073.D
Acq: 13 Oct 2025 18:00

Instrument :
BNA_N
ClientSampleId :
RW8-SP303-20251003



Tgt Ion:264 Resp: 10411
Ion Ratio Lower Upper
264 100
260 26.8 21.5 32.3
265 113.3 53.8 80.6#





CALIBRATION SUMMARY

Method Path : Z:\svoasrv\HPCHEM1\BNA_N\Methods\
 Method File : 8270-SIM-BN092325.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Wed Sep 24 11:00:01 2025
 Response Via : Initial Calibration

Calibration Files

0.1 =BN037861.D 0.2 =BN037862.D 0.4 =BN037863.D 0.8 =BN037864.D 1.6 =BN037865.D 3.2 =BN037866.D 5 =BN037867.D

	Compound	0.1	0.2	0.4	0.8	1.6	3.2	5	Avg	%RSD

1) I	1,4-Dichlorobenzen...	-----ISTD-----								
2)	1,4-Dioxane	0.413	0.410	0.403	0.425	0.410	0.376	0.406		4.04
3)	n-Nitrosodimet...	0.541	0.517	0.526	0.561	0.549	0.547	0.540		3.03
4) S	2-Fluorophenol	0.944	0.921	0.856	0.839	0.908	0.923	0.998	0.913	5.84
5) S	Phenol-d6	1.190	1.200	1.118	1.088	1.211	1.236	1.353	1.199	7.15
6)	bis(2-Chloroet...	1.133	1.096	1.082	1.066	1.150	1.124	1.143	1.113	2.89
7) I	Naphthalene-d8	-----ISTD-----								
8) S	Nitrobenzene-d5	0.299	0.289	0.289	0.290	0.310	0.320	0.338	0.305	6.14
9)	Naphthalene	1.105	1.089	1.048	1.051	1.123	1.137	1.160	1.102	3.84
10)	Hexachlorobuta...	0.198	0.184	0.182	0.183	0.192	0.188	0.188	0.188	2.94
11) SURR2	Methylnaphth...	0.515	0.518	0.499	0.504	0.553	0.587	0.714	0.556	13.74
12)	2-Methylnaphth...	0.678	0.681	0.668	0.682	0.754	0.775	0.802	0.720	7.68
13) I	Acenaphthene-d10	-----ISTD-----								
14) S	2,4,6-Tribromo...	0.148	0.142	0.135	0.139	0.163	0.184		0.152	12.15
15) S	2-Fluorobiphenyl	1.660	1.631	1.559	1.547	1.654	1.650	1.762	1.638	4.36
16)	Acenaphthylene	1.698	1.721	1.663	1.680	1.907	1.951	2.094	1.816	9.23
17)	Acenaphthene	1.262	1.230	1.198	1.208	1.351	1.371	1.437	1.294	7.16
18)	Fluorene	1.455	1.455	1.452	1.466	1.664	1.717	1.748	1.565	8.77
19) I	Phenanthrene-d10	-----ISTD-----								
20)	4,6-Dinitro-2-...	0.041	0.044	0.046	0.058	0.071	0.084	0.057		29.75
21)	4-Bromophenyl-...	0.252	0.247	0.244	0.244	0.272	0.275	0.291	0.261	7.07
22)	Hexachlorobenzene	0.317	0.312	0.305	0.300	0.324	0.323	0.332	0.316	3.53
23)	Atrazine	0.180	0.181	0.172	0.178	0.205	0.224	0.249	0.199	14.70
24)	Pentachlorophenol	0.101	0.092	0.097	0.116	0.136		0.108		16.50
25)	Phenanthrene	1.301	1.271	1.230	1.224	1.358	1.379	1.452	1.316	6.37
26)	Anthracene	1.034	1.035	1.023	1.057	1.209	1.275	1.361	1.142	12.10
27) SURR	Fluoranthene-d10	0.880	0.918	0.887	0.900	1.004	1.085	1.370	1.006	17.59
28)	Fluoranthene	1.278	1.315	1.311	1.358	1.555	1.617	1.713	1.450	12.07
29) I	Chrysene-d12	-----ISTD-----								
30)	Pyrene	1.576	1.600	1.521	1.476	1.607	1.595	1.679	1.579	4.12
31) S	Terphenyl-d14	0.799	0.791	0.753	0.735	0.798	0.809	0.893	0.797	6.34
32)	Benzo(a)anthra...	1.376	1.359	1.294	1.285	1.423	1.483	1.567	1.398	7.27
33)	Chrysene	1.508	1.526	1.462	1.446	1.559	1.512	1.569	1.512	3.03
34)	Bis(2-ethylhex...	0.592	0.487	0.476	0.518	0.569	0.676	0.553		13.57
35) I	Perylene-d12	-----ISTD-----								

Method Path : Z:\svoasrv\HPCHEM1\BNA_N\Methods\

Method File : 8270-SIM-BN092325.M

36)	Indeno(1,2,3-c...	1.715	1.672	1.658	1.707	1.887	1.996	2.159	1.828	10.52
37)	Benzo(b)fluora...	1.599	1.606	1.641	1.590	1.810	1.848	1.981	1.725	8.95
38)	Benzo(k)fluora...	1.620	1.586	1.597	1.628	1.875	1.917	1.926	1.736	9.26
39) C	Benzo(a)pyrene	1.254	1.296	1.231	1.254	1.413	1.492	1.610	1.364	10.61
40)	Dibenzo(a,h)an...	1.285	1.283	1.303	1.333	1.495	1.587	1.705	1.427	11.87
41)	Benzo(g,h,i)pe...	1.407	1.381	1.404	1.397	1.521	1.613	1.773	1.499	9.85

(#) = Out of Range

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN092325\
 Data File : BN037861.D
 Acq On : 23 Sep 2025 12:13
 Operator : RC/JU
 Sample : SSTDICC0.1
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC0.1

Quant Time: Sep 23 16:21:01 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN092325.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Sep 23 16:17:53 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.818	152	5856	0.400	ng	0.00
7) Naphthalene-d8	10.626	136	16223	0.400	ng	0.00
13) Acenaphthene-d10	14.473	164	8379	0.400	ng	0.00
19) Phenanthrene-d10	17.223	188	17197	0.400	ng	0.00
29) Chrysene-d12	21.402	240	13854	0.400	ng	0.00
35) Perylene-d12	23.736	264	13928	0.400	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.384	112	1382	0.103	ng	0.00
5) Phenol-d6	6.973	99	1742	0.099	ng	0.00
8) Nitrobenzene-d5	8.971	82	1212	0.098	ng	0.00
11) 2-Methylnaphthalene-d10	12.217	152	2089	0.093	ng	0.00
14) 2,4,6-Tribromophenol	15.969	330	309	0.091	ng	0.00
15) 2-Fluorobiphenyl	13.094	172	3478	0.101	ng	0.00
27) Fluoranthene-d10	19.253	212	3783	0.087	ng	0.00
31) Terphenyl-d14	19.852	244	2766	0.100	ng	0.00
Target Compounds						Qvalue
6) bis(2-Chloroethyl)ether	7.240	93	1658	0.102	ng	99
9) Naphthalene	10.669	128	4483	0.100	ng	96
10) Hexachlorobutadiene	10.979	225	802	0.105	ng	# 96
12) 2-Methylnaphthalene	12.293	142	2749	0.094	ng	98
16) Acenaphthylene	14.195	152	3557	0.093	ng	98
17) Acenaphthene	14.537	154	2644	0.098	ng	99
18) Fluorene	15.523	166	3048	0.093	ng	99
21) 4-Bromophenyl-phenylether	16.416	248	1084	0.097	ng	96
22) Hexachlorobenzene	16.540	284	1361	0.100	ng	99
23) Atrazine	16.689	200	772	0.090	ng	# 95
25) Phenanthrene	17.260	178	5595	0.099	ng	99
26) Anthracene	17.360	178	4446	0.091	ng	99
28) Fluoranthene	19.285	202	5493	0.088	ng	99
30) Pyrene	19.647	202	5458	0.100	ng	99
32) Benzo(a)anthracene	21.393	228	4766	0.098	ng	99
33) Chrysene	21.438	228	5223	0.100	ng	99
36) Indeno(1,2,3-cd)pyrene	26.113	276	5972	0.094	ng	98
37) Benzo(b)fluoranthene	23.031	252	5567	0.093	ng	# 78
38) Benzo(k)fluoranthene	23.075	252	5642	0.093	ng	# 80
39) Benzo(a)pyrene	23.630	252	4368	0.092	ng	# 71
40) Dibenzo(a,h)anthracene	26.127	278	4476	0.090	ng	# 85
41) Benzo(g,h,i)perylene	26.832	276	4900	0.094	ng	93


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN092325\
Data File : BN037861.D
Acq On : 23 Sep 2025 12:13
Operator : RC/JU
Sample : SSTDICC0.1
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
SSTDICC0.1

Quant Time: Sep 23 16:21:01 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN092325.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Sep 23 16:17:53 2025
Response via : Initial Calibration









#1
1,4-Dichlorobenzene-d4
Concen: 0.400 ng
RT: 7.818 min Scan# 61
Delta R.T. -0.007 min
Lab File: BN037861.D
Acq: 23 Sep 2025 12:13




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.1


Tgt Ion:152 Resp: 5856
Ion Ratio Lower Upper
152 100
150 154.8 121.0 181.4
115 59.9 47.4 71.0



#4
2-Fluorophenol
Concen: 0.103 ng
RT: 5.384 min Scan# 319
Delta R.T. -0.000 min
Lab File: BN037861.D
Acq: 23 Sep 2025 12:13

Tgt Ion:112 Resp: 1382
Ion Ratio Lower Upper
112 100
64 55.3 44.6 66.8
63 28.7 24.9 37.3





#5

Phenol-d6

Concen: 0.099 ng

RT: 6.973 min Scan# 539

Delta R.T. 0.000 min

Lab File: BN037861.D

Acq: 23 Sep 2025 12:13

Instrument :

BNA_N

ClientSampleId :

SSTDICC0.1





Tgt Ion: 99 Resp: 1742

Ion Ratio Lower Upper

99 100

42 21.0 17.2 25.8

71 33.5 27.3 40.9



#6

bis(2-Chloroethyl)ether

Concen: 0.102 ng

RT: 7.240 min Scan# 576

Delta R.T. 0.000 min

Lab File: BN037861.D

Acq: 23 Sep 2025 12:13




Tgt Ion: 93 Resp: 1658


Ion Ratio Lower Upper

93 100

63 75.2 60.1 90.1

95 31.0 25.4 38.2









#7
Naphthalene-d8
Concen: 0.400 ng
RT: 10.626 min Scan# 941
Delta R.T. 0.000 min
Lab File: BN037861.D
Acq: 23 Sep 2025 12:13




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.1


Tgt Ion:	136	Resp:	16223
Ion	Ratio	Lower	Upper
136	100		
137	11.6	9.0	13.4
54	7.3	5.8	8.8
68	5.5	4.3	6.5



#8
Nitrobenzene-d5
Concen: 0.098 ng
RT: 8.971 min Scan# 786
Delta R.T. -0.000 min
Lab File: BN037861.D
Acq: 23 Sep 2025 12:13

Tgt Ion:	82	Resp:	1212
Ion	Ratio	Lower	Upper
82	100		
128	48.1	34.8	52.2
54	56.9	42.2	63.2









#9
Naphthalene
Concen: 0.100 ng
RT: 10.669 min Scan# 945
Delta R.T. 0.000 min
Lab File: BN037861.D
Acq: 23 Sep 2025 12:13




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.1


Tgt Ion:128 Resp: 4483
Ion Ratio Lower Upper
128 100
129 13.2 9.2 13.8
127 15.2 11.0 16.6



#10
Hexachlorobutadiene
Concen: 0.105 ng
RT: 10.979 min Scan# 974
Delta R.T. 0.000 min
Lab File: BN037861.D
Acq: 23 Sep 2025 12:13

Tgt Ion:225 Resp: 802
Ion Ratio Lower Upper
225 100
223 0.0 0.0 0.0
227 60.8 51.4 77.2









#11
2-Methylnaphthalene-d10
Concen: 0.093 ng
RT: 12.217 min Scan# 1163
Delta R.T. -0.000 min
Lab File: BN037861.D
Acq: 23 Sep 2025 12:13




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.1


Tgt Ion:152 Resp: 2089
Ion Ratio Lower Upper
152 100
151 22.0 17.4 26.2



#12
2-Methylnaphthalene
Concen: 0.094 ng
RT: 12.293 min Scan# 1178
Delta R.T. 0.000 min
Lab File: BN037861.D
Acq: 23 Sep 2025 12:13

Tgt Ion:142 Resp: 2749
Ion Ratio Lower Upper
142 100
141 91.5 71.2 106.8
115 37.8 29.7 44.5





#13

Acenaphthene-d10

Concen: 0.400 ng

RT: 14.473 min Scan# 1456

Delta R.T. -0.000 min

Lab File: BN037861.D

Acq: 23 Sep 2025 12:13

Instrument :





BNA_N

ClientSampleId :

SSTDICC0.1

Tgt Ion:164 Resp: 8379

Ion	Ratio	Lower	Upper
164	100		
162	107.3	84.8	127.2
160	57.6	46.1	69.1



#14

2,4,6-Tribromophenol

Concen: 0.091 ng

RT: 15.969 min Scan# 1590




Delta R.T. 0.000 min


Lab File: BN037861.D

Acq: 23 Sep 2025 12:13

Tgt Ion:330 Resp: 309

Ion	Ratio	Lower	Upper
330	100		
332	100.3	77.2	115.8
141	50.8	37.2	55.8









#15
2-Fluorobiphenyl
Concen: 0.101 ng
RT: 13.094 min Scan# 1327
Delta R.T. 0.000 min
Lab File: BN037861.D
Acq: 23 Sep 2025 12:13




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.1


Tgt Ion	Ratio	Lower	Upper
172	100		
171	35.4	28.6	43.0
170	25.6	19.7	29.5



#16
Acenaphthylene
Concen: 0.093 ng
RT: 14.195 min Scan# 1430
Delta R.T. -0.000 min
Lab File: BN037861.D
Acq: 23 Sep 2025 12:13

Tgt Ion	Ratio	Lower	Upper
152	100		
151	21.2	15.8	23.8
153	13.2	10.5	15.7





#17

Acenaphthene

Concen: 0.098 ng

RT: 14.537 min Scan# 1462

Delta R.T. 0.000 min

Lab File: BN037861.D

Acq: 23 Sep 2025 12:13

Instrument :

BNA_N

ClientSampleId :

SSTDICC0.1





Tgt Ion:154 Resp: 2644

Ion Ratio Lower Upper

154 100

153 111.7 90.5 135.7

152 64.5 51.8 77.8



#18

Fluorene

Concen: 0.093 ng

RT: 15.523 min Scan# 1554

Delta R.T. -0.000 min

Lab File: BN037861.D

Acq: 23 Sep 2025 12:13




Tgt Ion:166 Resp: 3048


Ion Ratio Lower Upper

166 100

165 100.0 79.0 118.4

167 13.7 10.6 16.0





#19

Phenanthrene-d10

Concen: 0.400 ng

RT: 17.223 min Scan# 1691

Delta R.T. 0.000 min

Lab File: BN037861.D

Acq: 23 Sep 2025 12:13

Instrument :

BNA_N

ClientSampleId :

SSTDICC0.1





Tgt Ion:188 Resp: 17197

Ion Ratio Lower Upper

188 100

94 0.0 0.0 0.0

80 8.9 7.6 11.4



#21

4-Bromophenyl-phenylether

Concen: 0.097 ng

RT: 16.416 min Scan# 1626

Delta R.T. -0.000 min

Lab File: BN037861.D

Acq: 23 Sep 2025 12:13




Tgt Ion:248 Resp: 1084


Ion Ratio Lower Upper

248 100

250 99.3 77.6 116.4

141 65.3 48.8 73.2





#22

Hexachlorobenzene

Concen: 0.100 ng

RT: 16.540 min Scan# 1636

Delta R.T. 0.000 min

Lab File: BN037861.D

Acq: 23 Sep 2025 12:13

Instrument :

BNA_N

ClientSampleId :

SSTDICC0.1





Tgt Ion:284 Resp: 1361

Ion Ratio Lower Upper

284 100

142 38.9 30.9 46.3

249 30.6 23.9 35.9



#23

Atrazine

Concen: 0.090 ng

RT: 16.689 min Scan# 1648

Delta R.T. 0.000 min

Lab File: BN037861.D

Acq: 23 Sep 2025 12:13




Tgt Ion:200 Resp: 772


Ion Ratio Lower Upper

200 100

173 28.4 18.3 27.5#

215 49.4 38.6 58.0









#25
Phenanthrene
Concen: 0.099 ng
RT: 17.260 min Scan# 1694
Delta R.T. 0.000 min
Lab File: BN037861.D
Acq: 23 Sep 2025 12:13




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.1


Tgt Ion:178 Resp: 5595
Ion Ratio Lower Upper
178 100
176 19.3 15.7 23.5
179 16.2 12.3 18.5



#26
Anthracene
Concen: 0.091 ng
RT: 17.360 min Scan# 1702
Delta R.T. 0.000 min
Lab File: BN037861.D
Acq: 23 Sep 2025 12:13

Tgt Ion:178 Resp: 4446
Ion Ratio Lower Upper
178 100
176 18.5 15.0 22.4
179 14.9 12.3 18.5





#27

Fluoranthene-d10

Concen: 0.087 ng

RT: 19.253 min Scan# 1954

Delta R.T. -0.000 min

Lab File: BN037861.D

Acq: 23 Sep 2025 12:13

Instrument :

BNA_N

ClientSampleId :

SSTDICC0.1





Tgt Ion:212 Resp: 3783

Ion Ratio Lower Upper

212 100

106 16.8 12.6 19.0

104 9.7 7.4 11.0



#28

Fluoranthene

Concen: 0.088 ng

RT: 19.285 min Scan# 1961

Delta R.T. -0.000 min

Lab File: BN037861.D

Acq: 23 Sep 2025 12:13




Tgt Ion:202 Resp: 5493


Ion Ratio Lower Upper

202 100

101 11.3 9.3 13.9

203 16.8 13.6 20.4





#29

Chrysene-d12

Concen: 0.400 ng

RT: 21.402 min Scan# 21

Delta R.T. -0.000 min

Lab File: BN037861.D

Acq: 23 Sep 2025 12:13

Instrument :

BNA_N

ClientSampleId :

SSTDICC0.1





Tgt Ion:240 Resp: 13854

Ion Ratio Lower Upper

240 100

120 16.9 11.4 17.2

236 29.4 23.0 34.6



#30

Pyrene

Concen: 0.100 ng

RT: 19.647 min Scan# 2039

Delta R.T. 0.000 min

Lab File: BN037861.D

Acq: 23 Sep 2025 12:13




Tgt Ion:202 Resp: 5458


Ion Ratio Lower Upper

202 100

200 21.1 16.7 25.1


203 18.2 14.2 21.4







#31
Terphenyl-d14
Concen: 0.100 ng
RT: 19.852 min Scan# 2083
Delta R.T. 0.000 min
Lab File: BN037861.D
Acq: 23 Sep 2025 12:13

Instrument :
BNA_N
ClientSampleId :
SSTDICC0.1







Tgt Ion:244 Resp: 2766
Ion Ratio Lower Upper
244 100
212 10.7 7.6 11.4
122 19.4 13.9 20.9



#32
Benzo(a)anthracene
Concen: 0.098 ng
RT: 21.393 min Scan# 2271
Delta R.T. 0.009 min
Lab File: BN037861.D
Acq: 23 Sep 2025 12:13

Tgt Ion:228 Resp: 4766
Ion Ratio Lower Upper
228 100
226 27.9 22.6 33.8
229 20.8 15.6 23.4









#33
Chrysene
Concen: 0.100 ng
RT: 21.438 min Scan# 21
Delta R.T. 0.000 min
Lab File: BN037861.D
Acq: 23 Sep 2025 12:13




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.1


Tgt Ion	Ratio	Lower	Upper
228	100		
226	31.3	24.6	37.0
229	19.9	15.6	23.4



#35
Perylene-d12
Concen: 0.400 ng
RT: 23.736 min Scan# 2839
Delta R.T. 0.009 min
Lab File: BN037861.D
Acq: 23 Sep 2025 12:13

Tgt Ion	Ratio	Lower	Upper
264	100		
260	26.6	21.5	32.3
265	63.0	53.8	80.6





#36

Indeno(1,2,3-cd)pyrene

Concen: 0.094 ng

RT: 26.113 min Scan# 3652

Delta R.T. 0.006 min

Lab File: BN037861.D

Acq: 23 Sep 2025 12:13

Instrument :

BNA_N

ClientSampleId :

SSTDIC0.1





Tgt Ion:276 Resp: 5972

Ion Ratio Lower Upper

276 100

138 27.9 21.4 32.0

277 24.6 20.3 30.5



#37

Benzo(b)fluoranthene

Concen: 0.093 ng

RT: 23.031 min Scan# 2598

Delta R.T. 0.006 min

Lab File: BN037861.D

Acq: 23 Sep 2025 12:13




Tgt Ion:252 Resp: 5567


Ion Ratio Lower Upper

252 100

253 31.0 19.4 29.0#

125 31.3 13.0 19.4#





#38

Benzo(k)fluoranthene

Concen: 0.093 ng

RT: 23.075 min Scan# 2612

Delta R.T. 0.003 min

Lab File: BN037861.D

Acq: 23 Sep 2025 12:13

Instrument :

BNA_N

ClientSampleId :

SSTDICC0.1





Tgt Ion:252 Resp: 5642

Ion Ratio Lower Upper

252 100

253 31.4 19.5 29.3#

125 28.8 13.0 19.4#



#39

Benzo(a)pyrene

Concen: 0.092 ng

RT: 23.630 min Scan# 2803

Delta R.T. 0.006 min

Lab File: BN037861.D

Acq: 23 Sep 2025 12:13




Tgt Ion:252 Resp: 4368


Ion Ratio Lower Upper

252 100

253 35.2 20.6 30.8#

125 40.6 16.7 25.1#





#40

Dibenzo(a,h)anthracene

Concen: 0.090 ng

RT: 26.127 min Scan# 30

Delta R.T. 0.009 min

Lab File: BN037861.D

Acq: 23 Sep 2025 12:13

Instrument :

BNA_N

ClientSampleId :

SSTDICC0.1





Tgt Ion:278 Resp: 4476

Ion Ratio Lower Upper

278 100

139 25.9 15.4 23.2#

279 33.7 20.8 31.2#



#41

Benzo(g,h,i)perylene

Concen: 0.094 ng

RT: 26.832 min Scan# 3898

Delta R.T. 0.003 min

Lab File: BN037861.D

Acq: 23 Sep 2025 12:13




Tgt Ion:276 Resp: 4900

Ion Ratio Lower Upper

276 100

277 28.9 20.2 30.4

138 28.5 19.8 29.8



Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN092325\
 Data File : BN037862.D
 Acq On : 23 Sep 2025 12:49
 Operator : RC/JU
 Sample : SSTDICC0.2
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC0.2

Quant Time: Sep 23 16:21:29 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN092325.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Sep 23 16:17:53 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.826	152	5544	0.400	ng	0.00
7) Naphthalene-d8	10.626	136	15739	0.400	ng	0.00
13) Acenaphthene-d10	14.473	164	8069	0.400	ng	0.00
19) Phenanthrene-d10	17.223	188	15912	0.400	ng	0.00
29) Chrysene-d12	21.402	240	13158	0.400	ng	0.00
35) Perylene-d12	23.730	264	12974	0.400	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.384	112	2553	0.202	ng	0.00
5) Phenol-d6	6.973	99	3326	0.200	ng	0.00
8) Nitrobenzene-d5	8.971	82	2278	0.190	ng	0.00
11) 2-Methylnaphthalene-d10	12.217	152	4078	0.187	ng	0.00
14) 2,4,6-Tribromophenol	15.969	330	573	0.176	ng	0.00
15) 2-Fluorobiphenyl	13.094	172	6579	0.199	ng	0.00
27) Fluoranthene-d10	19.253	212	7304	0.182	ng	0.00
31) Terphenyl-d14	19.852	244	5201	0.198	ng	0.00
Target Compounds						Qvalue
2) 1,4-Dioxane	3.297	88	1144	0.203	ng	97
3) n-Nitrosodimethylamine	3.601	42	1500	0.200	ng	# 94
6) bis(2-Chloroethyl)ether	7.241	93	3039	0.197	ng	98
9) Naphthalene	10.669	128	8573	0.198	ng	99
10) Hexachlorobutadiene	10.979	225	1450	0.196	ng	# 99
12) 2-Methylnaphthalene	12.293	142	5361	0.189	ng	100
16) Acenaphthylene	14.195	152	6943	0.190	ng	99
17) Acenaphthene	14.537	154	4963	0.190	ng	98
18) Fluorene	15.523	166	5870	0.186	ng	99
20) 4,6-Dinitro-2-methylph...	15.597	198	325	0.230	ng	# 64
21) 4-Bromophenyl-phenylether	16.416	248	1965	0.190	ng	97
22) Hexachlorobenzene	16.540	284	2483	0.198	ng	99
23) Atrazine	16.677	200	1442	0.183	ng	# 89
24) Pentachlorophenol	16.876	266	804	0.186	ng	99
25) Phenanthrene	17.260	178	10109	0.193	ng	99
26) Anthracene	17.347	178	8235	0.181	ng	99
28) Fluoranthene	19.280	202	10466	0.181	ng	100
30) Pyrene	19.643	202	10528	0.203	ng	100
32) Benzo(a)anthracene	21.384	228	8943	0.194	ng	99
33) Chrysene	21.438	228	10042	0.202	ng	99
34) Bis(2-ethylhexyl)phtha...	21.322	149	3892	0.214	ng	# 100
36) Indeno(1,2,3-cd)pyrene	26.110	276	10845	0.183	ng	99
37) Benzo(b)fluoranthene	23.025	252	10420	0.186	ng	# 91
38) Benzo(k)fluoranthene	23.069	252	10290	0.183	ng	# 93
39) Benzo(a)pyrene	23.628	252	8405	0.190	ng	# 88
40) Dibenzo(a,h)anthracene	26.121	278	8321	0.180	ng	92
41) Benzo(g,h,i)perylene	26.826	276	8956	0.184	ng	97


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN092325\
Data File : BN037862.D
Acq On : 23 Sep 2025 12:49
Operator : RC/JU
Sample : SSTDICC0.2
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
SSTDICC0.2


Quant Time: Sep 23 16:21:29 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN092325.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Sep 23 16:17:53 2025
Response via : Initial Calibration








#1
1,4-Dichlorobenzene-d4
Concen: 0.400 ng
RT: 7.826 min Scan# 61
Delta R.T. 0.001 min
Lab File: BN037862.D
Acq: 23 Sep 2025 12:49

Instrument :
BNA_N
ClientSampleId :
SSTDICC0.2







Tgt Ion: 152 Resp: 5544
Ion Ratio Lower Upper
152 100
150 152.7 121.0 181.4
115 59.1 47.4 71.0



#2
1,4-Dioxane
Concen: 0.203 ng
RT: 3.297 min Scan# 30
Delta R.T. 0.000 min
Lab File: BN037862.D
Acq: 23 Sep 2025 12:49

Tgt Ion: 88 Resp: 1144
Ion Ratio Lower Upper
88 100
43 34.1 26.6 39.8
58 76.4 58.9 88.3









#3
n-Nitrosodimethylamine
Concen: 0.200 ng
RT: 3.601 min Scan# 72
Delta R.T. 0.000 min
Lab File: BN037862.D
Acq: 23 Sep 2025 12:49




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.2


Tgt Ion: 42 Resp: 1500
Ion Ratio Lower Upper
42 100
74 112.1 93.4 140.0
44 20.9 22.2 33.4#



#4
2-Fluorophenol
Concen: 0.202 ng
RT: 5.384 min Scan# 319
Delta R.T. 0.000 min
Lab File: BN037862.D
Acq: 23 Sep 2025 12:49

Tgt Ion: 112 Resp: 2553
Ion Ratio Lower Upper
112 100
64 54.6 44.6 66.8
63 30.9 24.9 37.3





#5

Phenol-d6

Concen: 0.200 ng

RT: 6.973 min Scan# 539

Delta R.T. 0.000 min

Lab File: BN037862.D

Acq: 23 Sep 2025 12:49

Instrument :

BNA_N

ClientSampleId :

SSTDICC0.2





Tgt Ion: 99 Resp: 3326

Ion Ratio Lower Upper

99 100

42 21.3 17.2 25.8

71 33.3 27.3 40.9



#6

bis(2-Chloroethyl)ether

Concen: 0.197 ng

RT: 7.241 min Scan# 576

Delta R.T. 0.000 min

Lab File: BN037862.D

Acq: 23 Sep 2025 12:49




Tgt Ion: 93 Resp: 3039


Ion Ratio Lower Upper

93 100

63 77.1 60.1 90.1

95 33.0 25.4 38.2









#7
Naphthalene-d8
Concen: 0.400 ng
RT: 10.626 min Scan# 941
Delta R.T. 0.000 min
Lab File: BN037862.D
Acq: 23 Sep 2025 12:49




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.2


Tgt Ion:	136	Resp:	15739
Ion	Ratio	Lower	Upper
136	100		
137	11.4	9.0	13.4
54	7.1	5.8	8.8
68	5.7	4.3	6.5



#8
Nitrobenzene-d5
Concen: 0.190 ng
RT: 8.971 min Scan# 786
Delta R.T. 0.000 min
Lab File: BN037862.D
Acq: 23 Sep 2025 12:49

Tgt Ion:	82	Resp:	2278
Ion	Ratio	Lower	Upper
82	100		
128	45.6	34.8	52.2
54	53.7	42.2	63.2









#9
Naphthalene
Concen: 0.198 ng
RT: 10.669 min Scan# 945
Delta R.T. 0.000 min
Lab File: BN037862.D
Acq: 23 Sep 2025 12:49




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.2


Tgt Ion	Ratio	Lower	Upper
128	100		
129	12.1	9.2	13.8
127	14.3	11.0	16.6



#10
Hexachlorobutadiene
Concen: 0.196 ng
RT: 10.979 min Scan# 974
Delta R.T. 0.000 min
Lab File: BN037862.D
Acq: 23 Sep 2025 12:49

Tgt Ion	Ratio	Lower	Upper
225	100		
223	0.0	0.0	0.0
227	65.0	51.4	77.2









#11
2-Methylnaphthalene-d10
Concen: 0.187 ng
RT: 12.217 min Scan# 1163
Delta R.T. 0.000 min
Lab File: BN037862.D
Acq: 23 Sep 2025 12:49




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.2


Tgt Ion:152 Resp: 4078
Ion Ratio Lower Upper
152 100
151 21.8 17.4 26.2



#12
2-Methylnaphthalene
Concen: 0.189 ng
RT: 12.293 min Scan# 1178
Delta R.T. 0.000 min
Lab File: BN037862.D
Acq: 23 Sep 2025 12:49

Tgt Ion:142 Resp: 5361
Ion Ratio Lower Upper
142 100
141 88.6 71.2 106.8
115 37.3 29.7 44.5









#13
Acenaphthene-d10
Concen: 0.400 ng
RT: 14.473 min Scan# 1456
Delta R.T. 0.000 min
Lab File: BN037862.D
Acq: 23 Sep 2025 12:49




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.2


Tgt Ion:164 Resp: 8069
Ion Ratio Lower Upper
164 100
162 106.7 84.8 127.2
160 57.3 46.1 69.1



#14
2,4,6-Tribromophenol
Concen: 0.176 ng
RT: 15.969 min Scan# 1590
Delta R.T. 0.000 min
Lab File: BN037862.D
Acq: 23 Sep 2025 12:49

Tgt Ion:330 Resp: 573
Ion Ratio Lower Upper
330 100
332 94.2 77.2 115.8
141 44.0 37.2 55.8









#15
2-Fluorobiphenyl
Concen: 0.199 ng
RT: 13.094 min Scan# 1327
Delta R.T. 0.000 min
Lab File: BN037862.D
Acq: 23 Sep 2025 12:49




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.2


Tgt Ion	Ratio	Lower	Upper
172	100		
171	36.6	28.6	43.0
170	25.1	19.7	29.5



#16
Acenaphthylene
Concen: 0.190 ng
RT: 14.195 min Scan# 1430
Delta R.T. 0.000 min
Lab File: BN037862.D
Acq: 23 Sep 2025 12:49

Tgt Ion	Ratio	Lower	Upper
152	100		
151	20.4	15.8	23.8
153	13.1	10.5	15.7









#17
Acenaphthene
Concen: 0.190 ng
RT: 14.537 min Scan# 1462
Delta R.T. 0.000 min
Lab File: BN037862.D
Acq: 23 Sep 2025 12:49




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.2


Tgt Ion	154	Resp	4963
Ion Ratio	Lower	Upper	
154	100		
153	115.4	90.5	135.7
152	65.5	51.8	77.8



#18
Fluorene
Concen: 0.186 ng
RT: 15.523 min Scan# 1554
Delta R.T. 0.000 min
Lab File: BN037862.D
Acq: 23 Sep 2025 12:49

Tgt Ion	166	Resp	5870
Ion Ratio	Lower	Upper	
166	100		
165	99.4	79.0	118.4
167	13.9	10.6	16.0





#19

Phenanthrene-d10

Concen: 0.400 ng

RT: 17.223 min Scan# 1691

Delta R.T. 0.000 min

Lab File: BN037862.D

Acq: 23 Sep 2025 12:49

Instrument :

BNA_N

ClientSampleId :

SSTDICC0.2





Tgt Ion:188 Resp: 15912

Ion Ratio Lower Upper

188 100

94 0.0 0.0 0.0

80 9.4 7.6 11.4



#20

4,6-Dinitro-2-methylphenol

Concen: 0.230 ng

RT: 15.597 min Scan# 1560

Delta R.T. 0.000 min

Lab File: BN037862.D

Acq: 23 Sep 2025 12:49




Tgt Ion:198 Resp: 325


Ion Ratio Lower Upper

198 100

51 111.2 59.1 88.7#

105 68.2 41.3 61.9#









#21
4-Bromophenyl-phenylether
Concen: 0.190 ng
RT: 16.416 min Scan# 1626
Delta R.T. 0.000 min
Lab File: BN037862.D
Acq: 23 Sep 2025 12:49




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.2


Tgt Ion	Ratio	Lower	Upper
248	100		
250	98.8	77.6	116.4
141	64.5	48.8	73.2



#22
Hexachlorobenzene
Concen: 0.198 ng
RT: 16.540 min Scan# 1636
Delta R.T. 0.000 min
Lab File: BN037862.D
Acq: 23 Sep 2025 12:49

Tgt Ion	Ratio	Lower	Upper
284	100		
142	39.2	30.9	46.3
249	30.3	23.9	35.9









#23
Atrazine
Concen: 0.183 ng
RT: 16.677 min Scan# 1647
Delta R.T. -0.012 min
Lab File: BN037862.D
Acq: 23 Sep 2025 12:49




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.2


Tgt Ion:200 Resp: 1442
Ion Ratio Lower Upper
200 100
173 37.6 18.3 27.5#
215 46.8 38.6 58.0



#24
Pentachlorophenol
Concen: 0.186 ng
RT: 16.876 min Scan# 1663
Delta R.T. 0.001 min
Lab File: BN037862.D
Acq: 23 Sep 2025 12:49

Tgt Ion:266 Resp: 804
Ion Ratio Lower Upper
266 100
264 63.3 50.9 76.3
268 66.4 54.3 81.5









#25
Phenanthrene
Concen: 0.193 ng
RT: 17.260 min Scan# 1694
Delta R.T. 0.000 min
Lab File: BN037862.D
Acq: 23 Sep 2025 12:49




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.2


Tgt Ion:178 Resp: 10109
Ion Ratio Lower Upper
178 100
176 19.1 15.7 23.5
179 15.6 12.3 18.5



#26
Anthracene
Concen: 0.181 ng
RT: 17.347 min Scan# 1701
Delta R.T. -0.012 min
Lab File: BN037862.D
Acq: 23 Sep 2025 12:49

Tgt Ion:178 Resp: 8235
Ion Ratio Lower Upper
178 100
176 18.1 15.0 22.4
179 15.7 12.3 18.5





#27

Fluoranthene-d10

Concen: 0.182 ng

RT: 19.253 min Scan# 1954

Delta R.T. 0.000 min

Lab File: BN037862.D

Acq: 23 Sep 2025 12:49

Instrument :

BNA_N

ClientSampleId :

SSTDICC0.2





Tgt Ion:212 Resp: 7304

Ion Ratio Lower Upper

212 100

106 16.1 12.6 19.0

104 9.3 7.4 11.0



#28

Fluoranthene

Concen: 0.181 ng

RT: 19.280 min Scan# 1960

Delta R.T. -0.005 min

Lab File: BN037862.D

Acq: 23 Sep 2025 12:49




Tgt Ion:202 Resp: 10466


Ion Ratio Lower Upper

202 100

101 11.8 9.3 13.9

203 17.2 13.6 20.4





#29

Chrysene-d12

Concen: 0.400 ng

RT: 21.402 min Scan# 21

Delta R.T. 0.000 min

Lab File: BN037862.D

Acq: 23 Sep 2025 12:49

Instrument :

BNA_N

ClientSampleId :

SSTDICC0.2





Tgt Ion:240 Resp: 13158

Ion Ratio Lower Upper

240 100

120 14.8 11.4 17.2

236 29.3 23.0 34.6



#30

Pyrene

Concen: 0.203 ng

RT: 19.643 min Scan# 2038

Delta R.T. -0.005 min

Lab File: BN037862.D

Acq: 23 Sep 2025 12:49




Tgt Ion:202 Resp: 10528


Ion Ratio Lower Upper

202 100

200 20.8 16.7 25.1

203 17.9 14.2 21.4









#31
Terphenyl-d14
Concen: 0.198 ng
RT: 19.852 min Scan# 2083
Delta R.T. 0.000 min
Lab File: BN037862.D
Acq: 23 Sep 2025 12:49




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.2


Tgt Ion:244 Resp: 5201
Ion Ratio Lower Upper
244 100
212 9.8 7.6 11.4
122 18.2 13.9 20.9



#32
Benzo(a)anthracene
Concen: 0.194 ng
RT: 21.384 min Scan# 2270
Delta R.T. 0.000 min
Lab File: BN037862.D
Acq: 23 Sep 2025 12:49

Tgt Ion:228 Resp: 8943
Ion Ratio Lower Upper
228 100
226 28.0 22.6 33.8
229 19.9 15.6 23.4





#33

Chrysene

Concen: 0.202 ng

RT: 21.438 min Scan# 21

Delta R.T. 0.000 min

Lab File: BN037862.D

Acq: 23 Sep 2025 12:49

Instrument :

BNA_N

ClientSampleId :

SSTDICC0.2





Tgt Ion:228 Resp: 10042

Ion Ratio Lower Upper

228 100

226 30.7 24.6 37.0

229 20.3 15.6 23.4



#34

Bis(2-ethylhexyl)phthalate

Concen: 0.214 ng

RT: 21.322 min Scan# 2263

Delta R.T. 0.000 min

Lab File: BN037862.D

Acq: 23 Sep 2025 12:49




Tgt Ion:149 Resp: 3892


Ion Ratio Lower Upper

149 100

167 25.5 20.4 30.6

279 3.6 2.4 3.6#









#35
Perylene-d12
Concen: 0.400 ng
RT: 23.730 min Scan# 2837
Delta R.T. 0.003 min
Lab File: BN037862.D
Acq: 23 Sep 2025 12:49




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.2


Tgt Ion:264 Resp: 12974
Ion Ratio Lower Upper
264 100
260 26.7 21.5 32.3
265 67.2 53.8 80.6



#36
Indeno(1,2,3-cd)pyrene
Concen: 0.183 ng
RT: 26.110 min Scan# 3651
Delta R.T. 0.003 min
Lab File: BN037862.D
Acq: 23 Sep 2025 12:49

Tgt Ion:276 Resp: 10845
Ion Ratio Lower Upper
276 100
138 27.3 21.4 32.0
277 25.3 20.3 30.5





#37

Benzo(b)fluoranthene

Concen: 0.186 ng

RT: 23.025 min Scan# 21

Delta R.T. 0.000 min

Lab File: BN037862.D

Acq: 23 Sep 2025 12:49

Instrument :

BNA_N

ClientSampleId :

SSTDICC0.2





Tgt Ion:252 Resp: 10420

Ion Ratio Lower Upper

252 100

253 27.2 19.4 29.0

125 21.9 13.0 19.4#



#38

Benzo(k)fluoranthene

Concen: 0.183 ng

RT: 23.069 min Scan# 2611

Delta R.T. -0.003 min

Lab File: BN037862.D

Acq: 23 Sep 2025 12:49




Tgt Ion:252 Resp: 10290


Ion Ratio Lower Upper

252 100

253 26.7 19.5 29.3

125 20.9 13.0 19.4#





#39

Benzo(a)pyrene

Concen: 0.190 ng

RT: 23.628 min Scan# 2801

Delta R.T. 0.003 min

Lab File: BN037862.D

Acq: 23 Sep 2025 12:49

Instrument :

BNA_N

ClientSampleId :

SSTDICC0.2





Tgt Ion:252 Resp: 8405

Ion Ratio Lower Upper

252 100

253 29.9 20.6 30.8

125 28.7 16.7 25.1#



#40

Dibenzo(a,h)anthracene

Concen: 0.180 ng

RT: 26.121 min Scan# 3655

Delta R.T. 0.003 min

Lab File: BN037862.D

Acq: 23 Sep 2025 12:49




Tgt Ion:278 Resp: 8321


Ion Ratio Lower Upper

278 100

139 22.4 15.4 23.2

279 30.3 20.8 31.2





#41

Benzo(g,h,i)perylene

Concen: 0.184 ng

RT: 26.826 min Scan# 38

Delta R.T. -0.003 min

Lab File: BN037862.D


Acq: 23 Sep 2025 12:49

Instrument :

BNA_N

ClientSampleId :

SSTDICC0.2





Tgt Ion:276 Resp: 8956

Ion Ratio Lower Upper

276 100

277 26.1 20.2 30.4

138 26.7 19.8 29.8



Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN092325\
 Data File : BN037863.D
 Acq On : 23 Sep 2025 13:25
 Operator : RC/JU
 Sample : SSTDICCC0.4
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SSTDICCC0.4

Quant Time: Sep 23 16:21:54 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN092325.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Sep 23 16:17:53 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.825	152	5584	0.400	ng	0.00
7) Naphthalene-d8	10.626	136	15792	0.400	ng	0.00
13) Acenaphthene-d10	14.473	164	8247	0.400	ng	0.00
19) Phenanthrene-d10	17.223	188	15939	0.400	ng	0.00
29) Chrysene-d12	21.402	240	13579	0.400	ng	0.00
35) Perylene-d12	23.727	264	12579	0.400	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.384	112	4779	0.375	ng	0.00
5) Phenol-d6	6.973	99	6244	0.373	ng	0.00
8) Nitrobenzene-d5	8.971	82	4566	0.379	ng	0.00
11) 2-Methylnaphthalene-d10	12.217	152	7877	0.359	ng	0.00
14) 2,4,6-Tribromophenol	15.969	330	1117	0.336	ng	0.00
15) 2-Fluorobiphenyl	13.094	172	12861	0.381	ng	0.00
27) Fluoranthene-d10	19.253	212	14130	0.352	ng	0.00
31) Terphenyl-d14	19.852	244	10222	0.378	ng	0.00
Target Compounds						Qvalue
2) 1,4-Dioxane	3.297	88	2288	0.404	ng	100
3) n-Nitrosodimethylamine	3.601	42	2885	0.383	ng	100
6) bis(2-Chloroethyl)ether	7.240	93	6041	0.389	ng	100
9) Naphthalene	10.669	128	16547	0.380	ng	100
10) Hexachlorobutadiene	10.979	225	2874	0.387	ng	# 100
12) 2-Methylnaphthalene	12.293	142	10542	0.371	ng	100
16) Acenaphthylene	14.195	152	13712	0.366	ng	100
17) Acenaphthene	14.537	154	9880	0.370	ng	100
18) Fluorene	15.523	166	11975	0.371	ng	100
20) 4,6-Dinitro-2-methylph...	15.597	198	696	0.406	ng	100
21) 4-Bromophenyl-phenylether	16.416	248	3896	0.375	ng	100
22) Hexachlorobenzene	16.540	284	4854	0.385	ng	100
23) Atrazine	16.689	200	2739	0.346	ng	100
24) Pentachlorophenol	16.875	266	1466	0.339	ng	100
25) Phenanthrene	17.260	178	19609	0.374	ng	100
26) Anthracene	17.360	178	16310	0.358	ng	100
28) Fluoranthene	19.285	202	20902	0.362	ng	100
30) Pyrene	19.647	202	20652	0.385	ng	100
32) Benzo(a)anthracene	21.384	228	17571	0.370	ng	100
33) Chrysene	21.438	228	19852	0.387	ng	100
34) Bis(2-ethylhexyl)phtha...	21.322	149	6617	0.352	ng	99
36) Indeno(1,2,3-cd)pyrene	26.107	276	20858	0.363	ng	100
37) Benzo(b)fluoranthene	23.025	252	20638	0.380	ng	100
38) Benzo(k)fluoranthene	23.072	252	20090	0.368	ng	100
39) Benzo(a)pyrene	23.625	252	15479	0.361	ng	100
40) Dibenzo(a,h)anthracene	26.118	278	16392	0.365	ng	100
41) Benzo(g,h,i)perylene	26.829	276	17656	0.374	ng	100


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN092325\
Data File : BN037863.D
Acq On : 23 Sep 2025 13:25
Operator : RC/JU
Sample : SSTDICC0.4
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
SSTDICC0.4


Quant Time: Sep 23 16:21:54 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN092325.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Sep 23 16:17:53 2025
Response via : Initial Calibration








#1
1,4-Dichlorobenzene-d4
Concen: 0.400 ng
RT: 7.825 min Scan# 61
Delta R.T. 0.000 min
Lab File: BN037863.D
Acq: 23 Sep 2025 13:25


Instrument :
BNA_N
ClientSampleId :
SSTDICCC0.4





Tgt Ion:152 Resp: 5584
Ion Ratio Lower Upper
152 100
150 151.2 121.0 181.4
115 59.2 47.4 71.0




#2
1,4-Dioxane
Concen: 0.404 ng
RT: 3.297 min Scan# 30
Delta R.T. 0.000 min
Lab File: BN037863.D
Acq: 23 Sep 2025 13:25



Tgt Ion: 88 Resp: 2288
Ion Ratio Lower Upper
88 100
43 33.2 26.6 39.8
58 73.6 58.9 88.3









#3
n-Nitrosodimethylamine
Concen: 0.383 ng
RT: 3.601 min Scan# 72
Delta R.T. 0.000 min
Lab File: BN037863.D
Acq: 23 Sep 2025 13:25




Instrument :
BNA_N
ClientSampleId :
SSTDICCC0.4


Tgt Ion: 42 Resp: 2885
Ion Ratio Lower Upper
42 100
74 116.7 93.4 140.0
44 27.8 22.2 33.4



#4
2-Fluorophenol
Concen: 0.375 ng
RT: 5.384 min Scan# 319
Delta R.T. 0.000 min
Lab File: BN037863.D
Acq: 23 Sep 2025 13:25

Tgt Ion: 112 Resp: 4779
Ion Ratio Lower Upper
112 100
64 55.7 44.6 66.8
63 31.1 24.9 37.3





#5

Phenol-d6

Concen: 0.373 ng

RT: 6.973 min Scan# 539

Delta R.T. 0.000 min

Lab File: BN037863.D


Acq: 23 Sep 2025 13:25

Instrument :

BNA_N

ClientSampleId :

SSTDICCC0.4





Tgt Ion: 99 Resp: 6244

Ion Ratio Lower Upper

99 100

42 21.5 17.2 25.8

71 34.1 27.3 40.9



#6

bis(2-Chloroethyl)ether

Concen: 0.389 ng

RT: 7.240 min Scan# 576

Delta R.T. 0.000 min

Lab File: BN037863.D

Acq: 23 Sep 2025 13:25




Tgt Ion: 93 Resp: 6041


Ion Ratio Lower Upper

93 100

63 75.1 60.1 90.1

95 31.8 25.4 38.2









#7
Naphthalene-d8
Concen: 0.400 ng
RT: 10.626 min Scan# 941
Delta R.T. 0.000 min
Lab File: BN037863.D
Acq: 23 Sep 2025 13:25




Instrument :
BNA_N
ClientSampleId :
SSTDICCC0.4


Tgt Ion:	136	Resp:	15792
Ion	Ratio	Lower	Upper
136	100		
137	11.2	9.0	13.4
54	7.3	5.8	8.8
68	5.4	4.3	6.5



#8
Nitrobenzene-d5
Concen: 0.379 ng
RT: 8.971 min Scan# 786
Delta R.T. 0.000 min
Lab File: BN037863.D
Acq: 23 Sep 2025 13:25

Tgt Ion:	82	Resp:	4566
Ion	Ratio	Lower	Upper
82	100		
128	43.5	34.8	52.2
54	52.7	42.2	63.2









#9
Naphthalene
Concen: 0.380 ng
RT: 10.669 min Scan# 945
Delta R.T. 0.000 min
Lab File: BN037863.D
Acq: 23 Sep 2025 13:25




Instrument :
BNA_N
ClientSampleId :
SSTDICCC0.4


Tgt Ion	128	129	127
Resp:	16547		
Ion Ratio	100	11.5	13.8
Lower		9.2	11.0
Upper		13.8	16.6



#10
Hexachlorobutadiene
Concen: 0.387 ng
RT: 10.979 min Scan# 974
Delta R.T. 0.000 min
Lab File: BN037863.D
Acq: 23 Sep 2025 13:25

Tgt Ion	225	223	227
Resp:	2874		
Ion Ratio	100	0.0	64.3
Lower		0.0	51.4
Upper		0.0	77.2









#11
2-Methylnaphthalene-d10
Concen: 0.359 ng
RT: 12.217 min Scan# 1163
Delta R.T. 0.000 min
Lab File: BN037863.D
Acq: 23 Sep 2025 13:25




Instrument :
BNA_N
ClientSampleId :
SSTDICCC0.4


Tgt Ion:152 Resp: 7877
Ion Ratio Lower Upper
152 100
151 21.8 17.4 26.2



#12
2-Methylnaphthalene
Concen: 0.371 ng
RT: 12.293 min Scan# 1178
Delta R.T. 0.000 min
Lab File: BN037863.D
Acq: 23 Sep 2025 13:25

Tgt Ion:142 Resp: 10542
Ion Ratio Lower Upper
142 100
141 89.0 71.2 106.8
115 37.1 29.7 44.5









#13
Acenaphthene-d10
Concen: 0.400 ng
RT: 14.473 min Scan# 1456
Delta R.T. 0.000 min
Lab File: BN037863.D
Acq: 23 Sep 2025 13:25




Instrument :
BNA_N
ClientSampleId :
SSTDICCC0.4


Tgt Ion:164 Resp: 8247
Ion Ratio Lower Upper
164 100
162 106.0 84.8 127.2
160 57.6 46.1 69.1



#14
2,4,6-Tribromophenol
Concen: 0.336 ng
RT: 15.969 min Scan# 1590
Delta R.T. 0.000 min
Lab File: BN037863.D
Acq: 23 Sep 2025 13:25

Tgt Ion:330 Resp: 1117
Ion Ratio Lower Upper
330 100
332 96.4 77.2 115.8
141 46.4 37.2 55.8









#15
2-Fluorobiphenyl
Concen: 0.381 ng
RT: 13.094 min Scan# 1327
Delta R.T. 0.000 min
Lab File: BN037863.D
Acq: 23 Sep 2025 13:25




Instrument :
BNA_N
ClientSampleId :
SSTDICCC0.4


Tgt Ion	Ratio	Lower	Upper
172	100		
171	35.8	28.6	43.0
170	24.6	19.7	29.5



#16
Acenaphthylene
Concen: 0.366 ng
RT: 14.195 min Scan# 1430
Delta R.T. 0.000 min
Lab File: BN037863.D
Acq: 23 Sep 2025 13:25

Tgt Ion	Ratio	Lower	Upper
152	100		
151	19.8	15.8	23.8
153	13.1	10.5	15.7









#17
Acenaphthene
Concen: 0.370 ng
RT: 14.537 min Scan# 1462
Delta R.T. 0.000 min
Lab File: BN037863.D
Acq: 23 Sep 2025 13:25




Instrument :
BNA_N
ClientSampleId :
SSTDICCC0.4


Tgt Ion:154 Resp: 9880
Ion Ratio Lower Upper
154 100
153 113.1 90.5 135.7
152 64.8 51.8 77.8



#18
Fluorene
Concen: 0.371 ng
RT: 15.523 min Scan# 1554
Delta R.T. 0.000 min
Lab File: BN037863.D
Acq: 23 Sep 2025 13:25

Tgt Ion:166 Resp: 11975
Ion Ratio Lower Upper
166 100
165 99.0 79.0 118.4
167 13.3 10.6 16.0





#19

Phenanthrene-d10

Concen: 0.400 ng

RT: 17.223 min Scan# 1691

Delta R.T. 0.000 min

Lab File: BN037863.D

Acq: 23 Sep 2025 13:25

Instrument :

BNA_N

ClientSampleId :

SSTDICCC0.4





Tgt Ion:188 Resp: 15939

Ion Ratio Lower Upper

188 100

94 0.0 0.0 0.0

80 9.5 7.6 11.4



#20

4,6-Dinitro-2-methylphenol

Concen: 0.406 ng

RT: 15.597 min Scan# 1560

Delta R.T. 0.000 min

Lab File: BN037863.D

Acq: 23 Sep 2025 13:25




Tgt Ion:198 Resp: 696


Ion Ratio Lower Upper

198 100

51 73.9 59.1 88.7

105 51.6 41.3 61.9









#21
4-Bromophenyl-phenylether
Concen: 0.375 ng
RT: 16.416 min Scan# 1626
Delta R.T. 0.000 min
Lab File: BN037863.D
Acq: 23 Sep 2025 13:25




Instrument :
BNA_N
ClientSampleId :
SSTDICCC0.4


Tgt Ion	248	Ratio	100	Lower	Upper
248	100				
250	97.0		77.6	116.4	
141	61.0		48.8	73.2	



#22
Hexachlorobenzene
Concen: 0.385 ng
RT: 16.540 min Scan# 1636
Delta R.T. 0.000 min
Lab File: BN037863.D
Acq: 23 Sep 2025 13:25

Tgt Ion	284	Ratio	100	Lower	Upper
284	100				
142	38.6		30.9	46.3	
249	29.9		23.9	35.9	









#23
Atrazine
Concen: 0.346 ng
RT: 16.689 min Scan# 1648
Delta R.T. 0.000 min
Lab File: BN037863.D
Acq: 23 Sep 2025 13:25




Instrument :
BNA_N
ClientSampleId :
SSTDICCC0.4


Tgt Ion	200	Resp:	2739
Ion Ratio	Lower	Upper	
200	100		
173	22.9	18.3	27.5
215	48.3	38.6	58.0



#24
Pentachlorophenol
Concen: 0.339 ng
RT: 16.875 min Scan# 1663
Delta R.T. 0.000 min
Lab File: BN037863.D
Acq: 23 Sep 2025 13:25

Tgt Ion	266	Resp:	1466
Ion Ratio	Lower	Upper	
266	100		
264	63.6	50.9	76.3
268	67.9	54.3	81.5





#25

Phenanthrene

Concen: 0.374 ng

RT: 17.260 min Scan# 1694

Delta R.T. 0.000 min

Lab File: BN037863.D

Acq: 23 Sep 2025 13:25

Instrument :

BNA_N

ClientSampleId :

SSTDICCC0.4





Tgt Ion:178 Resp: 19609

Ion Ratio Lower Upper

178 100

176 19.6 15.7 23.5

179 15.4 12.3 18.5



#26

Anthracene

Concen: 0.358 ng

RT: 17.360 min Scan# 1702

Delta R.T. 0.000 min

Lab File: BN037863.D

Acq: 23 Sep 2025 13:25




Tgt Ion:178 Resp: 16310


Ion Ratio Lower Upper

178 100

176 18.7 15.0 22.4

179 15.4 12.3 18.5





#27

Fluoranthene-d10

Concen: 0.352 ng

RT: 19.253 min Scan# 1954

Delta R.T. 0.000 min

Lab File: BN037863.D

Acq: 23 Sep 2025 13:25

Instrument :

BNA_N

ClientSampleId :

SSTDICCC0.4





Tgt Ion:212 Resp: 14130

Ion Ratio Lower Upper

212 100

106 15.8 12.6 19.0

104 9.2 7.4 11.0



#28

Fluoranthene

Concen: 0.362 ng

RT: 19.285 min Scan# 1961

Delta R.T. 0.000 min

Lab File: BN037863.D

Acq: 23 Sep 2025 13:25




Tgt Ion:202 Resp: 20902


Ion Ratio Lower Upper

202 100

101 11.6 9.3 13.9

203 17.0 13.6 20.4





#29

Chrysene-d12

Concen: 0.400 ng

RT: 21.402 min Scan# 21

Delta R.T. 0.000 min

Lab File: BN037863.D

Acq: 23 Sep 2025 13:25

Instrument :

BNA_N

ClientSampleId :

SSTDICCC0.4





Tgt Ion:240 Resp: 13579

Ion Ratio Lower Upper

240 100

120 14.3 11.4 17.2

236 28.8 23.0 34.6



#30

Pyrene

Concen: 0.385 ng

RT: 19.647 min Scan# 2039

Delta R.T. 0.000 min

Lab File: BN037863.D

Acq: 23 Sep 2025 13:25




Tgt Ion:202 Resp: 20652


Ion Ratio Lower Upper

202 100

200 20.9 16.7 25.1

203 17.8 14.2 21.4









#31
Terphenyl-d14
Concen: 0.378 ng
RT: 19.852 min Scan# 2083
Delta R.T. 0.000 min
Lab File: BN037863.D
Acq: 23 Sep 2025 13:25




Instrument :
BNA_N
ClientSampleId :
SSTDICCC0.4


Tgt Ion	Ratio	Lower	Upper
244	100		
212	9.5	7.6	11.4
122	17.4	13.9	20.9



#32
Benzo(a)anthracene
Concen: 0.370 ng
RT: 21.384 min Scan# 2270
Delta R.T. 0.000 min
Lab File: BN037863.D
Acq: 23 Sep 2025 13:25

Tgt Ion	Ratio	Lower	Upper
228	100		
226	28.2	22.6	33.8
229	19.5	15.6	23.4





#33

Chrysene

Concen: 0.387 ng

RT: 21.438 min Scan# 21

Delta R.T. 0.000 min

Lab File: BN037863.D

Acq: 23 Sep 2025 13:25

Instrument :

BNA_N

ClientSampleId :

SSTDICCC0.4





Tgt Ion:228 Resp: 19852

Ion Ratio Lower Upper

228 100

226 30.8 24.6 37.0

229 19.5 15.6 23.4



#34

Bis(2-ethylhexyl)phthalate

Concen: 0.352 ng

RT: 21.322 min Scan# 2263

Delta R.T. 0.000 min

Lab File: BN037863.D

Acq: 23 Sep 2025 13:25




Tgt Ion:149 Resp: 6617


Ion Ratio Lower Upper

149 100

167 26.2 20.4 30.6

279 3.1 2.4 3.6





#35

Perylene-d12

Concen: 0.400 ng

RT: 23.727 min Scan# 2836

Delta R.T. 0.000 min

Lab File: BN037863.D


Acq: 23 Sep 2025 13:25

Instrument :

BNA_N




ClientSampleId :

SSTDICCC0.4



Tgt Ion:264 Resp: 12579

Ion	Ratio	Lower	Upper
264	100		
260	26.9	21.5	32.3
265	67.2	53.8	80.6



#36

Indeno(1,2,3-cd)pyrene

Concen: 0.363 ng

RT: 26.107 min Scan# 3650




Delta R.T. 0.000 min


Lab File: BN037863.D

Acq: 23 Sep 2025 13:25

Tgt Ion:276 Resp: 20858

Ion	Ratio	Lower	Upper
276	100		
138	26.7	21.4	32.0
277	25.4	20.3	30.5





#37

Benzo(b)fluoranthene

Concen: 0.380 ng

RT: 23.025 min Scan# 2596

Delta R.T. 0.000 min

Lab File: BN037863.D

Acq: 23 Sep 2025 13:25

Instrument :

BNA_N

ClientSampleId :

SSTDICCC0.4





Tgt Ion:252 Resp: 20638

Ion Ratio Lower Upper

252 100

253 24.2 19.4 29.0

125 16.2 13.0 19.4



#38

Benzo(k)fluoranthene

Concen: 0.368 ng

RT: 23.072 min Scan# 2612

Delta R.T. 0.000 min

Lab File: BN037863.D

Acq: 23 Sep 2025 13:25




Tgt Ion:252 Resp: 20090


Ion Ratio Lower Upper

252 100

253 24.4 19.5 29.3

125 16.2 13.0 19.4





#39

Benzo(a)pyrene

Concen: 0.361 ng

RT: 23.625 min Scan# 2801

Delta R.T. 0.000 min

Lab File: BN037863.D

Acq: 23 Sep 2025 13:25

Instrument :

BNA_N

ClientSampleId :

SSTDICCC0.4





Tgt Ion:252 Resp: 15479

Ion Ratio Lower Upper

252 100

253 25.7 20.6 30.8

125 20.9 16.7 25.1



#40

Dibenzo(a,h)anthracene

Concen: 0.365 ng

RT: 26.118 min Scan# 3654

Delta R.T. 0.000 min

Lab File: BN037863.D

Acq: 23 Sep 2025 13:25




Tgt Ion:278 Resp: 16392


Ion Ratio Lower Upper

278 100

139 19.3 15.4 23.2

279 26.0 20.8 31.2





#41

Benzo(g,h,i)perylene

Concen: 0.374 ng

RT: 26.829 min Scan# 38

Delta R.T. 0.000 min

Lab File: BN037863.D


Acq: 23 Sep 2025 13:25

Instrument :

BNA_N

ClientSampleId :

SSTDICCC0.4





Tgt Ion:276 Resp: 17656

Ion Ratio Lower Upper

276 100

277 25.3 20.2 30.4

138 24.8 19.8 29.8



Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN092325\
 Data File : BN037864.D
 Acq On : 23 Sep 2025 14:02
 Operator : RC/JU
 Sample : SSTDICC0.8
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC0.8

Quant Time: Sep 23 16:22:19 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN092325.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Sep 23 16:17:53 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.825	152	5204	0.400	ng	0.00
7) Naphthalene-d8	10.626	136	14715	0.400	ng	0.00
13) Acenaphthene-d10	14.473	164	7889	0.400	ng	0.00
19) Phenanthrene-d10	17.223	188	15207	0.400	ng	0.00
29) Chrysene-d12	21.402	240	13674	0.400	ng	0.00
35) Perylene-d12	23.730	264	12464	0.400	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.384	112	8737	0.736	ng	0.00
5) Phenol-d6	6.973	99	11319	0.725	ng	0.00
8) Nitrobenzene-d5	8.971	82	8536	0.761	ng	0.00
11) 2-Methylnaphthalene-d10	12.217	152	14822	0.725	ng	0.00
14) 2,4,6-Tribromophenol	15.969	330	2186	0.687	ng	0.00
15) 2-Fluorobiphenyl	13.094	172	24413	0.756	ng	0.00
27) Fluoranthene-d10	19.252	212	27365	0.715	ng	0.00
31) Terphenyl-d14	19.852	244	20095	0.738	ng	0.00
Target Compounds						Qvalue
2) 1,4-Dioxane	3.297	88	4191	0.793	ng	98
3) n-Nitrosodimethylamine	3.600	42	5472	0.779	ng	# 89
6) bis(2-Chloroethyl)ether	7.240	93	11096	0.766	ng	99
9) Naphthalene	10.669	128	30935	0.763	ng	99
10) Hexachlorobutadiene	10.979	225	5393	0.780	ng	# 99
12) 2-Methylnaphthalene	12.293	142	20065	0.758	ng	99
16) Acenaphthylene	14.195	152	26510	0.740	ng	99
17) Acenaphthene	14.537	154	19057	0.747	ng	100
18) Fluorene	15.522	166	23130	0.749	ng	99
20) 4,6-Dinitro-2-methylph...	15.597	198	1414	0.755	ng	# 77
21) 4-Bromophenyl-phenylether	16.416	248	7412	0.748	ng	98
22) Hexachlorobenzene	16.528	284	9138	0.761	ng	99
23) Atrazine	16.677	200	5405	0.716	ng	# 87
24) Pentachlorophenol	16.875	266	2945	0.714	ng	97
25) Phenanthrene	17.260	178	37218	0.744	ng	99
26) Anthracene	17.347	178	32138	0.740	ng	100
28) Fluoranthene	19.285	202	41292	0.749	ng	100
30) Pyrene	19.643	202	40377	0.748	ng	100
32) Benzo(a)anthracene	21.384	228	35130	0.735	ng	100
33) Chrysene	21.438	228	39541	0.765	ng	100
34) Bis(2-ethylhexyl)phtha...	21.322	149	13026	0.689	ng	99
36) Indeno(1,2,3-cd)pyrene	26.107	276	42560	0.747	ng	100
37) Benzo(b)fluoranthene	23.025	252	39646	0.738	ng	96
38) Benzo(k)fluoranthene	23.075	252	40574	0.750	ng	96
39) Benzo(a)pyrene	23.627	252	31265	0.735	ng	94
40) Dibenzo(a,h)anthracene	26.121	278	33235	0.747	ng	98
41) Benzo(g,h,i)perylene	26.835	276	34817	0.745	ng	98


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN092325\
Data File : BN037864.D
Acq On : 23 Sep 2025 14:02
Operator : RC/JU
Sample : SSTDICC0.8
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
SSTDICC0.8


Quant Time: Sep 23 16:22:19 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN092325.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Sep 23 16:17:53 2025
Response via : Initial Calibration








#1
1,4-Dichlorobenzene-d4
Concen: 0.400 ng
RT: 7.825 min Scan# 61
Delta R.T. 0.000 min
Lab File: BN037864.D
Acq: 23 Sep 2025 14:02

Instrument :
BNA_N
ClientSampleId :
SSTDICC0.8







Tgt Ion:152 Resp: 5204
Ion Ratio Lower Upper
152 100
150 155.6 121.0 181.4
115 60.6 47.4 71.0



#2
1,4-Dioxane
Concen: 0.793 ng
RT: 3.297 min Scan# 30
Delta R.T. -0.000 min
Lab File: BN037864.D
Acq: 23 Sep 2025 14:02

Tgt Ion: 88 Resp: 4191
Ion Ratio Lower Upper
88 100
43 30.7 26.6 39.8
58 74.2 58.9 88.3









#3
n-Nitrosodimethylamine
Concen: 0.779 ng
RT: 3.600 min Scan# 71
Delta R.T. -0.000 min
Lab File: BN037864.D
Acq: 23 Sep 2025 14:02




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.8


Tgt Ion: 42 Resp: 5472
Ion Ratio Lower Upper
42 100
74 111.5 93.4 140.0
44 7.3 22.2 33.4#



#4
2-Fluorophenol
Concen: 0.736 ng
RT: 5.384 min Scan# 319
Delta R.T. -0.000 min
Lab File: BN037864.D
Acq: 23 Sep 2025 14:02

Tgt Ion: 112 Resp: 8737
Ion Ratio Lower Upper
112 100
64 54.9 44.6 66.8
63 29.8 24.9 37.3









#5
Phenol-d6
Concen: 0.725 ng
RT: 6.973 min Scan# 539
Delta R.T. -0.000 min
Lab File: BN037864.D
Acq: 23 Sep 2025 14:02




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.8


Tgt Ion	Ratio	Lower	Upper
99	100		
42	21.0	17.2	25.8
71	33.5	27.3	40.9



#6
bis(2-Chloroethyl)ether
Concen: 0.766 ng
RT: 7.240 min Scan# 576
Delta R.T. -0.000 min
Lab File: BN037864.D
Acq: 23 Sep 2025 14:02

Tgt Ion	Ratio	Lower	Upper
93	100		
63	75.5	60.1	90.1
95	32.2	25.4	38.2









#7
Naphthalene-d8
Concen: 0.400 ng
RT: 10.626 min Scan# 941
Delta R.T. -0.000 min
Lab File: BN037864.D
Acq: 23 Sep 2025 14:02




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.8


Tgt Ion:	136	Resp:	14715
Ion	Ratio	Lower	Upper
136	100		
137	11.3	9.0	13.4
54	7.2	5.8	8.8
68	5.6	4.3	6.5



#8
Nitrobenzene-d5
Concen: 0.761 ng
RT: 8.971 min Scan# 786
Delta R.T. -0.000 min
Lab File: BN037864.D
Acq: 23 Sep 2025 14:02

Tgt Ion:	82	Resp:	8536
Ion	Ratio	Lower	Upper
82	100		
128	41.8	34.8	52.2
54	52.1	42.2	63.2









#9
Naphthalene
Concen: 0.763 ng
RT: 10.669 min Scan# 945
Delta R.T. -0.000 min
Lab File: BN037864.D
Acq: 23 Sep 2025 14:02




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.8


Tgt Ion:128 Resp: 30935
Ion Ratio Lower Upper
128 100
129 11.2 9.2 13.8
127 13.6 11.0 16.6



#10
Hexachlorobutadiene
Concen: 0.780 ng
RT: 10.979 min Scan# 974
Delta R.T. -0.000 min
Lab File: BN037864.D
Acq: 23 Sep 2025 14:02

Tgt Ion:225 Resp: 5393
Ion Ratio Lower Upper
225 100
223 0.0 0.0 0.0
227 63.7 51.4 77.2









#11
2-Methylnaphthalene-d10
Concen: 0.725 ng
RT: 12.217 min Scan# 1163
Delta R.T. -0.000 min
Lab File: BN037864.D
Acq: 23 Sep 2025 14:02




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.8


Tgt Ion:152 Resp: 14822
Ion Ratio Lower Upper
152 100
151 21.7 17.4 26.2



#12
2-Methylnaphthalene
Concen: 0.758 ng
RT: 12.293 min Scan# 1178
Delta R.T. -0.000 min
Lab File: BN037864.D
Acq: 23 Sep 2025 14:02

Tgt Ion:142 Resp: 20065
Ion Ratio Lower Upper
142 100
141 89.5 71.2 106.8
115 36.0 29.7 44.5





#13

Acenaphthene-d10

Concen: 0.400 ng

RT: 14.473 min Scan# 1456

Delta R.T. -0.000 min

Lab File: BN037864.D

Acq: 23 Sep 2025 14:02

Instrument :

BNA_N

ClientSampleId :

SSTDICC0.8





Tgt Ion:164 Resp: 7889

Ion Ratio Lower Upper

164 100

162 107.2 84.8 127.2

160 57.1 46.1 69.1



#14

2,4,6-Tribromophenol

Concen: 0.687 ng

RT: 15.969 min Scan# 1590

Delta R.T. -0.000 min

Lab File: BN037864.D

Acq: 23 Sep 2025 14:02




Tgt Ion:330 Resp: 2186


Ion Ratio Lower Upper

330 100

332 97.2 77.2 115.8

141 45.1 37.2 55.8









#15
2-Fluorobiphenyl
Concen: 0.756 ng
RT: 13.094 min Scan# 1327
Delta R.T. -0.000 min
Lab File: BN037864.D
Acq: 23 Sep 2025 14:02




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.8


Tgt Ion	Ratio	Lower	Upper
172	100		
171	36.0	28.6	43.0
170	25.1	19.7	29.5



#16
Acenaphthylene
Concen: 0.740 ng
RT: 14.195 min Scan# 1430
Delta R.T. -0.000 min
Lab File: BN037864.D
Acq: 23 Sep 2025 14:02

Tgt Ion	Ratio	Lower	Upper
152	100		
151	20.2	15.8	23.8
153	12.9	10.5	15.7









#17
Acenaphthene
Concen: 0.747 ng
RT: 14.537 min Scan# 1462
Delta R.T. -0.000 min
Lab File: BN037864.D
Acq: 23 Sep 2025 14:02




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.8


Tgt Ion:154 Resp: 19057
Ion Ratio Lower Upper
154 100
153 112.9 90.5 135.7
152 64.2 51.8 77.8



#18
Fluorene
Concen: 0.749 ng
RT: 15.522 min Scan# 1554
Delta R.T. -0.000 min
Lab File: BN037864.D
Acq: 23 Sep 2025 14:02

Tgt Ion:166 Resp: 23130
Ion Ratio Lower Upper
166 100
165 99.4 79.0 118.4
167 13.4 10.6 16.0





#19

Phenanthrene-d10

Concen: 0.400 ng

RT: 17.223 min Scan# 1691

Delta R.T. -0.000 min

Lab File: BN037864.D

Acq: 23 Sep 2025 14:02

Instrument :

BNA_N

ClientSampleId :

SSTDICC0.8





Tgt Ion:188 Resp: 15207

Ion Ratio Lower Upper

188 100

94 0.0 0.0 0.0

80 9.0 7.6 11.4



#20

4,6-Dinitro-2-methylphenol

Concen: 0.755 ng

RT: 15.597 min Scan# 1560

Delta R.T. -0.000 min

Lab File: BN037864.D

Acq: 23 Sep 2025 14:02




Tgt Ion:198 Resp: 1414


Ion Ratio Lower Upper

198 100

51 49.5 59.1 88.7#

105 41.0 41.3 61.9#





#21

4-Bromophenyl-phenylether

Concen: 0.748 ng

RT: 16.416 min Scan# 1626

Delta R.T. -0.000 min

Lab File: BN037864.D

Acq: 23 Sep 2025 14:02

Instrument :

BNA_N

ClientSampleId :

SSTDICC0.8





Tgt Ion:248 Resp: 7412

Ion Ratio Lower Upper

248 100

250 95.7 77.6 116.4

141 58.0 48.8 73.2



#22

Hexachlorobenzene

Concen: 0.761 ng

RT: 16.528 min Scan# 1635

Delta R.T. -0.012 min

Lab File: BN037864.D

Acq: 23 Sep 2025 14:02




Tgt Ion:284 Resp: 9138


Ion Ratio Lower Upper

284 100

142 39.7 30.9 46.3

249 30.2 23.9 35.9









#23
Atrazine
Concen: 0.716 ng
RT: 16.677 min Scan# 1647
Delta R.T. -0.012 min
Lab File: BN037864.D
Acq: 23 Sep 2025 14:02




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.8


Tgt Ion:200 Resp: 5405
Ion Ratio Lower Upper
200 100
173 34.5 18.3 27.5#
215 42.7 38.6 58.0



#24
Pentachlorophenol
Concen: 0.714 ng
RT: 16.875 min Scan# 1663
Delta R.T. 0.000 min
Lab File: BN037864.D
Acq: 23 Sep 2025 14:02

Tgt Ion:266 Resp: 2945
Ion Ratio Lower Upper
266 100
264 62.0 50.9 76.3
268 64.2 54.3 81.5









#25
Phenanthrene
Concen: 0.744 ng
RT: 17.260 min Scan# 1694
Delta R.T. -0.000 min
Lab File: BN037864.D
Acq: 23 Sep 2025 14:02




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.8


Tgt Ion:178 Resp: 37218
Ion Ratio Lower Upper
178 100
176 19.2 15.7 23.5
179 15.1 12.3 18.5



#26
Anthracene
Concen: 0.740 ng
RT: 17.347 min Scan# 1701
Delta R.T. -0.012 min
Lab File: BN037864.D
Acq: 23 Sep 2025 14:02

Tgt Ion:178 Resp: 32138
Ion Ratio Lower Upper
178 100
176 18.6 15.0 22.4
179 15.3 12.3 18.5





#27

Fluoranthene-d10

Concen: 0.715 ng

RT: 19.252 min Scan# 1954

Delta R.T. -0.000 min

Lab File: BN037864.D

Acq: 23 Sep 2025 14:02

Instrument :

BNA_N

ClientSampleId :

SSTDICC0.8





Tgt Ion:212 Resp: 27365

Ion Ratio Lower Upper

212 100

106 15.8 12.6 19.0

104 9.1 7.4 11.0



#28

Fluoranthene

Concen: 0.749 ng

RT: 19.285 min Scan# 1961

Delta R.T. -0.000 min

Lab File: BN037864.D

Acq: 23 Sep 2025 14:02




Tgt Ion:202 Resp: 41292


Ion Ratio Lower Upper

202 100

101 11.9 9.3 13.9

203 17.1 13.6 20.4





#29

Chrysene-d12

Concen: 0.400 ng

RT: 21.402 min Scan# 21

Delta R.T. -0.000 min

Lab File: BN037864.D

Acq: 23 Sep 2025 14:02

Instrument :

BNA_N

ClientSampleId :

SSTDICC0.8





Tgt Ion:240 Resp: 13674

Ion Ratio Lower Upper

240 100

120 14.9 11.4 17.2

236 28.9 23.0 34.6



#30

Pyrene

Concen: 0.748 ng

RT: 19.643 min Scan# 2038

Delta R.T. -0.005 min

Lab File: BN037864.D

Acq: 23 Sep 2025 14:02




Tgt Ion:202 Resp: 40377


Ion Ratio Lower Upper

202 100

200 21.1 16.7 25.1

203 17.7 14.2 21.4









#31
Terphenyl-d14
Concen: 0.738 ng
RT: 19.852 min Scan# 2083
Delta R.T. -0.000 min
Lab File: BN037864.D
Acq: 23 Sep 2025 14:02




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.8


Tgt Ion	Ratio	Lower	Upper
244	100		
212	9.3	7.6	11.4
122	17.0	13.9	20.9



#32
Benzo(a)anthracene
Concen: 0.735 ng
RT: 21.384 min Scan# 2270
Delta R.T. -0.000 min
Lab File: BN037864.D
Acq: 23 Sep 2025 14:02

Tgt Ion	Ratio	Lower	Upper
228	100		
226	28.3	22.6	33.8
229	19.5	15.6	23.4





#33

Chrysene

Concen: 0.765 ng

RT: 21.438 min Scan# 21

Delta R.T. -0.000 min

Lab File: BN037864.D

Acq: 23 Sep 2025 14:02

Instrument :

BNA_N

ClientSampleId :

SSTDICC0.8





Tgt Ion:228 Resp: 39541

Ion Ratio Lower Upper

228 100

226 30.6 24.6 37.0

229 19.7 15.6 23.4



#34

Bis(2-ethylhexyl)phthalate

Concen: 0.689 ng

RT: 21.322 min Scan# 2263

Delta R.T. -0.000 min

Lab File: BN037864.D

Acq: 23 Sep 2025 14:02




Tgt Ion:149 Resp: 13026


Ion Ratio Lower Upper

149 100

167 26.1 20.4 30.6

279 3.2 2.4 3.6









#35
Perylene-d12
Concen: 0.400 ng
RT: 23.730 min Scan# 2837
Delta R.T. 0.003 min
Lab File: BN037864.D
Acq: 23 Sep 2025 14:02




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.8


Tgt Ion:264 Resp: 12464
Ion Ratio Lower Upper
264 100
260 27.0 21.5 32.3
265 67.5 53.8 80.6



#36
Indeno(1,2,3-cd)pyrene
Concen: 0.747 ng
RT: 26.107 min Scan# 3650
Delta R.T. -0.000 min
Lab File: BN037864.D
Acq: 23 Sep 2025 14:02

Tgt Ion:276 Resp: 42560
Ion Ratio Lower Upper
276 100
138 26.8 21.4 32.0
277 25.3 20.3 30.5





#37

Benzo(b)fluoranthene

Concen: 0.738 ng

RT: 23.025 min Scan# 2596

Delta R.T. -0.000 min

Lab File: BN037864.D


Acq: 23 Sep 2025 14:02

Instrument :

BNA_N

ClientSampleId :

SSTDICC0.8





Tgt Ion:252 Resp: 39646

Ion Ratio Lower Upper

252 100

253 22.9 19.4 29.0

125 13.7 13.0 19.4



#38

Benzo(k)fluoranthene


Concen: 0.750 ng

RT: 23.075 min Scan# 2613

Delta R.T. 0.003 min

Lab File: BN037864.D

Acq: 23 Sep 2025 14:02





Tgt Ion:252 Resp: 40574


Ion Ratio Lower Upper

252 100

253 23.0 19.5 29.3

125 13.3 13.0 19.4





#39

Benzo(a)pyrene

Concen: 0.735 ng

RT: 23.627 min Scan# 2801

Delta R.T. 0.003 min

Lab File: BN037864.D


Acq: 23 Sep 2025 14:02

Instrument :

BNA_N

ClientSampleId :

SSTDICC0.8






Tgt Ion:252 Resp: 31265

Ion Ratio Lower Upper

252 100

253 23.7 20.6 30.8

125 16.9 16.7 25.1



#40

Dibenzo(a,h)anthracene

Concen: 0.747 ng

RT: 26.121 min Scan# 3655

Delta R.T. 0.003 min

Lab File: BN037864.D

Acq: 23 Sep 2025 14:02




Tgt Ion:278 Resp: 33235


Ion Ratio Lower Upper

278 100

139 18.7 15.4 23.2

279 25.1 20.8 31.2





#41

Benzo(g,h,i)perylene

Concen: 0.745 ng

RT: 26.835 min Scan# 38

Delta R.T. 0.006 min

Lab File: BN037864.D


Acq: 23 Sep 2025 14:02

Instrument :

BNA_N

ClientSampleId :

SSTDICC0.8





Tgt Ion:276 Resp: 34817

Ion Ratio Lower Upper

276 100

277 24.2 20.2 30.4

138 23.9 19.8 29.8



Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN092325\
 Data File : BN037865.D
 Acq On : 23 Sep 2025 14:38
 Operator : RC/JU
 Sample : SSTDICC1.6
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC1.6

Quant Time: Sep 23 16:22:44 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN092325.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Sep 23 16:17:53 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.825	152	5367	0.400	ng	0.00
7) Naphthalene-d8	10.626	136	15494	0.400	ng	0.00
13) Acenaphthene-d10	14.473	164	8404	0.400	ng	0.00
19) Phenanthrene-d10	17.223	188	15875	0.400	ng	0.00
29) Chrysene-d12	21.402	240	14963	0.400	ng	0.00
35) Perylene-d12	23.735	264	12917	0.400	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.391	112	19493	1.592	ng	0.00
5) Phenol-d6	6.973	99	26005	1.616	ng	0.00
8) Nitrobenzene-d5	8.971	82	19230	1.627	ng	0.00
11) 2-Methylnaphthalene-d10	12.217	152	34256	1.592	ng	0.00
14) 2,4,6-Tribromophenol	15.969	330	5471	1.613	ng	0.00
15) 2-Fluorobiphenyl	13.094	172	55598	1.616	ng	0.00
27) Fluoranthene-d10	19.252	212	63771	1.597	ng	0.00
31) Terphenyl-d14	19.852	244	47765	1.603	ng	0.00
Target Compounds						
						Qvalue
2) 1,4-Dioxane	3.297	88	9123	1.674	ng	98
3) n-Nitrosodimethylamine	3.600	42	12048	1.662	ng	# 89
6) bis(2-Chloroethyl)ether	7.240	93	24687	1.653	ng	100
9) Naphthalene	10.679	128	69620	1.631	ng	98
10) Hexachlorobutadiene	10.978	225	11899	1.634	ng	# 100
12) 2-Methylnaphthalene	12.293	142	46704	1.675	ng	99
16) Acenaphthylene	14.195	152	64093	1.680	ng	100
17) Acenaphthene	14.537	154	45411	1.670	ng	97
18) Fluorene	15.522	166	55922	1.701	ng	100
20) 4,6-Dinitro-2-methylph...	15.597	198	3669	1.602	ng	# 64
21) 4-Bromophenyl-phenylether	16.416	248	17278	1.670	ng	98
22) Hexachlorobenzene	16.540	284	20548	1.638	ng	99
23) Atrazine	16.689	200	13043	1.656	ng	97
24) Pentachlorophenol	16.875	266	7393	1.718	ng	98
25) Phenanthrene	17.260	178	86207	1.650	ng	99
26) Anthracene	17.359	178	76744	1.693	ng	100
28) Fluoranthene	19.285	202	98772	1.717	ng	99
30) Pyrene	19.647	202	96189	1.628	ng	100
32) Benzo(a)anthracene	21.384	228	85145	1.628	ng	99
33) Chrysene	21.438	228	93309	1.650	ng	100
34) Bis(2-ethylhexyl)phtha...	21.330	149	31032	1.500	ng	98
36) Indeno(1,2,3-cd)pyrene	26.112	276	97476	1.652	ng	99
37) Benzo(b)fluoranthene	23.028	252	93525	1.679	ng	# 94
38) Benzo(k)fluoranthene	23.075	252	96868	1.728	ng	# 94
39) Benzo(a)pyrene	23.630	252	73009	1.657	ng	# 90
40) Dibenzo(a,h)anthracene	26.127	278	77237	1.676	ng	97
41) Benzo(g,h,i)perylene	26.835	276	78576	1.623	ng	98


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN092325\
Data File : BN037865.D
Acq On : 23 Sep 2025 14:38
Operator : RC/JU
Sample : SSTDICC1.6
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
SSTDICC1.6

Quant Time: Sep 23 16:22:44 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN092325.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Sep 23 16:17:53 2025
Response via : Initial Calibration









#1
1,4-Dichlorobenzene-d4
Concen: 0.400 ng
RT: 7.825 min Scan# 61
Delta R.T. 0.000 min
Lab File: BN037865.D
Acq: 23 Sep 2025 14:38




Instrument :
BNA_N
ClientSampleId :
SSTDICC1.6


Tgt Ion:152 Resp: 5367
Ion Ratio Lower Upper
152 100
150 156.4 121.0 181.4
115 61.0 47.4 71.0



#2
1,4-Dioxane
Concen: 1.674 ng
RT: 3.297 min Scan# 30
Delta R.T. -0.000 min
Lab File: BN037865.D
Acq: 23 Sep 2025 14:38

Tgt Ion: 88 Resp: 9123
Ion Ratio Lower Upper
88 100
43 31.3 26.6 39.8
58 74.6 58.9 88.3









#3
n-Nitrosodimethylamine
Concen: 1.662 ng
RT: 3.600 min Scan# 71
Delta R.T. -0.000 min
Lab File: BN037865.D
Acq: 23 Sep 2025 14:38




Instrument :
BNA_N
ClientSampleId :
SSTDICC1.6


Tgt Ion: 42 Resp: 12048
Ion Ratio Lower Upper
42 100
74 110.9 93.4 140.0
44 8.4 22.2 33.4#



#4
2-Fluorophenol
Concen: 1.592 ng
RT: 5.391 min Scan# 320
Delta R.T. 0.007 min
Lab File: BN037865.D
Acq: 23 Sep 2025 14:38

Tgt Ion: 112 Resp: 19493
Ion Ratio Lower Upper
112 100
64 55.9 44.6 66.8
63 30.3 24.9 37.3









#5
Phenol-d6
Concen: 1.616 ng
RT: 6.973 min Scan# 539
Delta R.T. -0.000 min
Lab File: BN037865.D
Acq: 23 Sep 2025 14:38




Instrument :
BNA_N
ClientSampleId :
SSTDICC1.6


Tgt Ion	Ratio	Lower	Upper
99	100		
42	21.3	17.2	25.8
71	33.5	27.3	40.9



#6
bis(2-Chloroethyl)ether
Concen: 1.653 ng
RT: 7.240 min Scan# 576
Delta R.T. -0.000 min
Lab File: BN037865.D
Acq: 23 Sep 2025 14:38

Tgt Ion	Ratio	Lower	Upper
93	100		
63	75.4	60.1	90.1
95	32.1	25.4	38.2









#7
Naphthalene-d8
Concen: 0.400 ng
RT: 10.626 min Scan# 941
Delta R.T. -0.000 min
Lab File: BN037865.D
Acq: 23 Sep 2025 14:38




Instrument :
BNA_N
ClientSampleId :
SSTDICC1.6


Tgt Ion:	136	Resp:	15494
Ion	Ratio	Lower	Upper
136	100		
137	11.1	9.0	13.4
54	7.3	5.8	8.8
68	5.8	4.3	6.5



#8
Nitrobenzene-d5
Concen: 1.627 ng
RT: 8.971 min Scan# 786
Delta R.T. -0.000 min
Lab File: BN037865.D
Acq: 23 Sep 2025 14:38

Tgt Ion:	82	Resp:	19230
Ion	Ratio	Lower	Upper
82	100		
128	42.0	34.8	52.2
54	51.9	42.2	63.2









#9
Naphthalene
Concen: 1.631 ng
RT: 10.679 min Scan# 945
Delta R.T. 0.010 min
Lab File: BN037865.D
Acq: 23 Sep 2025 14:38




Instrument :
BNA_N
ClientSampleId :
SSTDICC1.6


Tgt Ion:128 Resp: 69620
Ion Ratio Lower Upper
128 100
129 10.9 9.2 13.8
127 13.0 11.0 16.6



#10
Hexachlorobutadiene
Concen: 1.634 ng
RT: 10.978 min Scan# 974
Delta R.T. -0.000 min
Lab File: BN037865.D
Acq: 23 Sep 2025 14:38

Tgt Ion:225 Resp: 11899
Ion Ratio Lower Upper
225 100
223 0.0 0.0 0.0
227 64.6 51.4 77.2









#11
2-Methylnaphthalene-d10
Concen: 1.592 ng
RT: 12.217 min Scan# 1163
Delta R.T. -0.000 min
Lab File: BN037865.D
Acq: 23 Sep 2025 14:38




Instrument :
BNA_N
ClientSampleId :
SSTDICC1.6


Tgt Ion:152 Resp: 34256
Ion Ratio Lower Upper
152 100
151 21.7 17.4 26.2



#12
2-Methylnaphthalene
Concen: 1.675 ng
RT: 12.293 min Scan# 1178
Delta R.T. -0.000 min
Lab File: BN037865.D
Acq: 23 Sep 2025 14:38

Tgt Ion:142 Resp: 46704
Ion Ratio Lower Upper
142 100
141 88.2 71.2 106.8
115 35.4 29.7 44.5









#13
Acenaphthene-d10
Concen: 0.400 ng
RT: 14.473 min Scan# 1456
Delta R.T. -0.000 min
Lab File: BN037865.D
Acq: 23 Sep 2025 14:38




Instrument :
BNA_N
ClientSampleId :
SSTDICC1.6


Tgt Ion	Ratio	Lower	Upper
164	100		
162	106.4	84.8	127.2
160	57.0	46.1	69.1



#14
2,4,6-Tribromophenol
Concen: 1.613 ng
RT: 15.969 min Scan# 1590
Delta R.T. -0.000 min
Lab File: BN037865.D
Acq: 23 Sep 2025 14:38

Tgt Ion	Ratio	Lower	Upper
330	100		
332	98.1	77.2	115.8
141	44.0	37.2	55.8





#15

2-Fluorobiphenyl

Concen: 1.616 ng

RT: 13.094 min Scan# 1327

Delta R.T. -0.000 min

Lab File: BN037865.D

Acq: 23 Sep 2025 14:38

Instrument :

BNA_N

ClientSampleId :

SSTDICC1.6





Tgt Ion:172 Resp: 55598

Ion Ratio Lower Upper

172 100

171 35.8 28.6 43.0

170 24.7 19.7 29.5



#16

Acenaphthylene

Concen: 1.680 ng

RT: 14.195 min Scan# 1430

Delta R.T. -0.000 min

Lab File: BN037865.D

Acq: 23 Sep 2025 14:38




Tgt Ion:152 Resp: 64093


Ion Ratio Lower Upper

152 100

151 19.9 15.8 23.8

153 12.9 10.5 15.7









#17
Acenaphthene
Concen: 1.670 ng
RT: 14.537 min Scan# 1462
Delta R.T. -0.000 min
Lab File: BN037865.D
Acq: 23 Sep 2025 14:38




Instrument :
BNA_N
ClientSampleId :
SSTDICC1.6


Tgt Ion:154 Resp: 45411
Ion Ratio Lower Upper
154 100
153 110.4 90.5 135.7
152 61.0 51.8 77.8



#18
Fluorene
Concen: 1.701 ng
RT: 15.522 min Scan# 1554
Delta R.T. -0.000 min
Lab File: BN037865.D
Acq: 23 Sep 2025 14:38

Tgt Ion:166 Resp: 55922
Ion Ratio Lower Upper
166 100
165 99.1 79.0 118.4
167 13.1 10.6 16.0





#19

Phenanthrene-d10

Concen: 0.400 ng

RT: 17.223 min Scan# 1691

Delta R.T. -0.000 min

Lab File: BN037865.D

Acq: 23 Sep 2025 14:38

Instrument :

BNA_N

ClientSampleId :

SSTDICC1.6





Tgt Ion:188 Resp: 15875

Ion Ratio Lower Upper

188 100

94 0.0 0.0 0.0

80 9.8 7.6 11.4



#20

4,6-Dinitro-2-methylphenol

Concen: 1.602 ng

RT: 15.597 min Scan# 1560

Delta R.T. -0.000 min

Lab File: BN037865.D

Acq: 23 Sep 2025 14:38




Tgt Ion:198 Resp: 3669


Ion Ratio Lower Upper

198 100

51 36.7 59.1 88.7#

105 34.9 41.3 61.9#









#21
4-Bromophenyl-phenylether
Concen: 1.670 ng
RT: 16.416 min Scan# 1626
Delta R.T. -0.000 min
Lab File: BN037865.D
Acq: 23 Sep 2025 14:38




Instrument :
BNA_N
ClientSampleId :
SSTDICC1.6


Tgt Ion:248 Resp: 17278
Ion Ratio Lower Upper
248 100
250 95.0 77.6 116.4
141 62.0 48.8 73.2



#22
Hexachlorobenzene
Concen: 1.638 ng
RT: 16.540 min Scan# 1636
Delta R.T. -0.000 min
Lab File: BN037865.D
Acq: 23 Sep 2025 14:38

Tgt Ion:284 Resp: 20548
Ion Ratio Lower Upper
284 100
142 39.6 30.9 46.3
249 29.9 23.9 35.9





#23

Atrazine

Concen: 1.656 ng

RT: 16.689 min Scan# 1648

Delta R.T. -0.000 min

Lab File: BN037865.D

Acq: 23 Sep 2025 14:38

Instrument :

BNA_N

ClientSampleId :

SSTDICC1.6





Tgt Ion:200 Resp: 13043

Ion Ratio Lower Upper

200 100

173 21.0 18.3 27.5

215 46.4 38.6 58.0



#24

Pentachlorophenol

Concen: 1.718 ng

RT: 16.875 min Scan# 1663

Delta R.T. 0.000 min

Lab File: BN037865.D

Acq: 23 Sep 2025 14:38




Tgt Ion:266 Resp: 7393


Ion Ratio Lower Upper

266 100

264 63.1 50.9 76.3

268 64.6 54.3 81.5









#25
Phenanthrene
Concen: 1.650 ng
RT: 17.260 min Scan# 1694
Delta R.T. -0.000 min
Lab File: BN037865.D
Acq: 23 Sep 2025 14:38




Instrument :
BNA_N
ClientSampleId :
SSTDICC1.6


Tgt Ion:178 Resp: 86207
Ion Ratio Lower Upper
178 100
176 19.3 15.7 23.5
179 15.3 12.3 18.5



#26
Anthracene
Concen: 1.693 ng
RT: 17.359 min Scan# 1702
Delta R.T. -0.000 min
Lab File: BN037865.D
Acq: 23 Sep 2025 14:38

Tgt Ion:178 Resp: 76744
Ion Ratio Lower Upper
178 100
176 18.6 15.0 22.4
179 15.3 12.3 18.5





#27

Fluoranthene-d10

Concen: 1.597 ng

RT: 19.252 min Scan# 1954

Delta R.T. -0.000 min

Lab File: BN037865.D

Acq: 23 Sep 2025 14:38

Instrument :

BNA_N

ClientSampleId :

SSTDICC1.6





Tgt Ion:212 Resp: 63771

Ion Ratio Lower Upper

212 100

106 15.8 12.6 19.0

104 9.1 7.4 11.0



#28

Fluoranthene

Concen: 1.717 ng

RT: 19.285 min Scan# 1961

Delta R.T. -0.000 min

Lab File: BN037865.D

Acq: 23 Sep 2025 14:38




Tgt Ion:202 Resp: 98772


Ion Ratio Lower Upper

202 100

101 11.8 9.3 13.9

203 17.2 13.6 20.4





#29

Chrysene-d12

Concen: 0.400 ng

RT: 21.402 min Scan# 21

Delta R.T. -0.000 min

Lab File: BN037865.D

Acq: 23 Sep 2025 14:38

Instrument :

BNA_N

ClientSampleId :

SSTDICC1.6





Tgt Ion:240 Resp: 14963

Ion Ratio Lower Upper

240 100

120 15.4 11.4 17.2

236 29.1 23.0 34.6



#30

Pyrene

Concen: 1.628 ng

RT: 19.647 min Scan# 2039

Delta R.T. -0.000 min

Lab File: BN037865.D

Acq: 23 Sep 2025 14:38




Tgt Ion:202 Resp: 96189


Ion Ratio Lower Upper

202 100

200 21.1 16.7 25.1

203 17.6 14.2 21.4









#31
Terphenyl-d14
Concen: 1.603 ng
RT: 19.852 min Scan# 2083
Delta R.T. -0.000 min
Lab File: BN037865.D
Acq: 23 Sep 2025 14:38




Instrument :
BNA_N
ClientSampleId :
SSTDICC1.6


Tgt Ion:244 Resp: 47765
Ion Ratio Lower Upper
244 100
212 9.0 7.6 11.4
122 16.9 13.9 20.9



#32
Benzo(a)anthracene
Concen: 1.628 ng
RT: 21.384 min Scan# 2270
Delta R.T. -0.000 min
Lab File: BN037865.D
Acq: 23 Sep 2025 14:38

Tgt Ion:228 Resp: 85145
Ion Ratio Lower Upper
228 100
226 28.5 22.6 33.8
229 19.1 15.6 23.4





#33

Chrysene

Concen: 1.650 ng

RT: 21.438 min Scan# 21

Delta R.T. -0.000 min

Lab File: BN037865.D

Acq: 23 Sep 2025 14:38

Instrument :

BNA_N

ClientSampleId :

SSTDICC1.6





Tgt Ion:228 Resp: 93309

Ion Ratio Lower Upper

228 100

226 30.9 24.6 37.0

229 19.3 15.6 23.4



#34

Bis(2-ethylhexyl)phthalate

Concen: 1.500 ng

RT: 21.330 min Scan# 2264

Delta R.T. 0.009 min

Lab File: BN037865.D

Acq: 23 Sep 2025 14:38




Tgt Ion:149 Resp: 31032


Ion Ratio Lower Upper

149 100

167 26.4 20.4 30.6

279 3.4 2.4 3.6









#35
Perylene-d12
Concen: 0.400 ng
RT: 23.735 min Scan# 2839
Delta R.T. 0.009 min
Lab File: BN037865.D
Acq: 23 Sep 2025 14:38




Instrument :
BNA_N
ClientSampleId :
SSTDICC1.6


Tgt Ion:264 Resp: 12917
Ion Ratio Lower Upper
264 100
260 27.4 21.5 32.3
265 67.5 53.8 80.6



#36
Indeno(1,2,3-cd)pyrene
Concen: 1.652 ng
RT: 26.112 min Scan# 3652
Delta R.T. 0.006 min
Lab File: BN037865.D
Acq: 23 Sep 2025 14:38

Tgt Ion:276 Resp: 97476
Ion Ratio Lower Upper
276 100
138 27.1 21.4 32.0
277 25.1 20.3 30.5





#37

Benzo(b)fluoranthene

Concen: 1.679 ng

RT: 23.028 min Scan# 21

Delta R.T. 0.003 min

Lab File: BN037865.D

Acq: 23 Sep 2025 14:38

Instrument :

BNA_N

ClientSampleId :

SSTDICC1.6





Tgt Ion:252 Resp: 93525

Ion Ratio Lower Upper

252 100

253 22.2 19.4 29.0

125 12.3 13.0 19.4#



#38

Benzo(k)fluoranthene

Concen: 1.728 ng

RT: 23.075 min Scan# 2613

Delta R.T. 0.003 min

Lab File: BN037865.D

Acq: 23 Sep 2025 14:38




Tgt Ion:252 Resp: 96868


Ion Ratio Lower Upper

252 100

253 22.1 19.5 29.3

125 12.2 13.0 19.4#





#39

Benzo(a)pyrene

Concen: 1.657 ng

RT: 23.630 min Scan# 2803

Delta R.T. 0.006 min

Lab File: BN037865.D


Acq: 23 Sep 2025 14:38

Instrument :

BNA_N

ClientSampleId :

SSTDICC1.6






Tgt Ion:252 Resp: 73009

Ion Ratio Lower Upper

252 100

253 22.5 20.6 30.8

125 14.2 16.7 25.1#



#40

Dibenzo(a,h)anthracene

Concen: 1.676 ng

RT: 26.127 min Scan# 3657

Delta R.T. 0.009 min

Lab File: BN037865.D

Acq: 23 Sep 2025 14:38




Tgt Ion:278 Resp: 77237


Ion Ratio Lower Upper

278 100

139 18.1 15.4 23.2

279 23.9 20.8 31.2





#41

Benzo(g,h,i)perylene

Concen: 1.623 ng

RT: 26.835 min Scan# 38

Delta R.T. 0.006 min

Lab File: BN037865.D


Acq: 23 Sep 2025 14:38

Instrument :

BNA_N

ClientSampleId :

SSTDICC1.6





Tgt Ion:276 Resp: 78576

Ion Ratio Lower Upper

276 100

277 24.1 20.2 30.4

138 23.5 19.8 29.8



Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN092325\
 Data File : BN037866.D
 Acq On : 23 Sep 2025 15:15
 Operator : RC/JU
 Sample : SSTDICC3.2
 Misc :
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC3.2

Quant Time: Sep 23 16:23:09 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN092325.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Sep 23 16:17:53 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.825	152	5486	0.400	ng	0.00
7) Naphthalene-d8	10.626	136	15583	0.400	ng	0.00
13) Acenaphthene-d10	14.473	164	8623	0.400	ng	0.00
19) Phenanthrene-d10	17.223	188	16225	0.400	ng	0.00
29) Chrysene-d12	21.402	240	16309	0.400	ng	# 0.00
35) Perylene-d12	23.736	264	13684	0.400	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.392	112	40521	3.237	ng	0.00
5) Phenol-d6	6.980	99	54237	3.297	ng	0.00
8) Nitrobenzene-d5	8.971	82	39859	3.353	ng	0.00
11) 2-Methylnaphthalene-d10	12.222	152	73211	3.382	ng	0.00
14) 2,4,6-Tribromophenol	15.969	330	12677	3.642	ng	0.00
15) 2-Fluorobiphenyl	13.094	172	113791	3.223	ng	0.00
27) Fluoranthene-d10	19.257	212	140845	3.451	ng	0.00
31) Terphenyl-d14	19.852	244	105582	3.250	ng	0.00
Target Compounds						Qvalue
2) 1,4-Dioxane	3.297	88	18009	3.234	ng	99
3) n-Nitrosodimethylamine	3.593	42	24113	3.255	ng	# 89
6) bis(2-Chloroethyl)ether	7.240	93	49318	3.230	ng	100
9) Naphthalene	10.680	128	141732	3.301	ng	98
10) Hexachlorobutadiene	10.979	225	23493	3.208	ng	# 99
12) 2-Methylnaphthalene	12.293	142	96667	3.447	ng	99
16) Acenaphthylene	14.195	152	134565	3.437	ng	100
17) Acenaphthene	14.537	154	94611	3.392	ng	95
18) Fluorene	15.522	166	118468	3.511	ng	100
20) 4,6-Dinitro-2-methylph...	15.597	198	9172	3.211	ng	# 62
21) 4-Bromophenyl-phenylether	16.416	248	35637	3.370	ng	96
22) Hexachlorobenzene	16.540	284	41902	3.269	ng	100
23) Atrazine	16.689	200	29112	3.616	ng	97
24) Pentachlorophenol	16.875	266	17645	4.011	ng	97
25) Phenanthrene	17.260	178	178985	3.352	ng	99
26) Anthracene	17.360	178	165527	3.573	ng	100
28) Fluoranthene	19.285	202	209920	3.570	ng	99
30) Pyrene	19.647	202	208135	3.232	ng	100
32) Benzo(a)anthracene	21.393	228	193464	3.394	ng	98
33) Chrysene	21.438	228	197283	3.201	ng	99
34) Bis(2-ethylhexyl)phtha...	21.331	149	74195	3.291	ng	98
36) Indeno(1,2,3-cd)pyrene	26.118	276	218487	3.494	ng	100
37) Benzo(b)fluoranthene	23.031	252	202282	3.428	ng	# 93
38) Benzo(k)fluoranthene	23.078	252	209819	3.534	ng	# 93
39) Benzo(a)pyrene	23.636	252	163321	3.499	ng	# 88
40) Dibenzo(a,h)anthracene	26.133	278	173724	3.558	ng	96
41) Benzo(g,h,i)perylene	26.841	276	176617	3.443	ng	97


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN092325\
Data File : BN037866.D
Acq On : 23 Sep 2025 15:15
Operator : RC/JU
Sample : SSTDICC3.2
Misc :
ALS Vial : 7 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2


Quant Time: Sep 23 16:23:09 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN092325.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Sep 23 16:17:53 2025
Response via : Initial Calibration








#1
1,4-Dichlorobenzene-d4
Concen: 0.400 ng
RT: 7.825 min Scan# 61
Delta R.T. 0.000 min
Lab File: BN037866.D
Acq: 23 Sep 2025 15:15

Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2







Tgt Ion:152 Resp: 5486
Ion Ratio Lower Upper
152 100
150 151.9 121.0 181.4
115 59.6 47.4 71.0



#2
1,4-Dioxane
Concen: 3.234 ng
RT: 3.297 min Scan# 30
Delta R.T. -0.000 min
Lab File: BN037866.D
Acq: 23 Sep 2025 15:15

Tgt Ion: 88 Resp: 18009
Ion Ratio Lower Upper
88 100
43 30.7 26.6 39.8
58 73.6 58.9 88.3





#3

n-Nitrosodimethylamine

Concen: 3.255 ng

RT: 3.593 min Scan# 71

Delta R.T. -0.007 min

Lab File: BN037866.D

Acq: 23 Sep 2025 15:15

Instrument :

BNA_N

ClientSampleId :

SSTDICC3.2





Tgt Ion: 42 Resp: 24113

Ion Ratio Lower Upper

42 100

74 112.5 93.4 140.0

44 6.3 22.2 33.4#



#4

2-Fluorophenol

Concen: 3.237 ng

RT: 5.392 min Scan# 320

Delta R.T. 0.007 min

Lab File: BN037866.D

Acq: 23 Sep 2025 15:15




Tgt Ion: 112 Resp: 40521


Ion Ratio Lower Upper

112 100

64 55.6 44.6 66.8

63 30.1 24.9 37.3





#5

Phenol-d6

Concen: 3.297 ng

RT: 6.980 min Scan# 54

Delta R.T. 0.007 min

Lab File: BN037866.D

Acq: 23 Sep 2025 15:15

Instrument :

BNA_N

ClientSampleId :

SSTDICC3.2





Tgt Ion: 99 Resp: 54237

Ion Ratio Lower Upper

99 100

42 21.2 17.2 25.8

71 33.7 27.3 40.9



#6

bis(2-Chloroethyl)ether

Concen: 3.230 ng

RT: 7.240 min Scan# 576

Delta R.T. -0.000 min

Lab File: BN037866.D

Acq: 23 Sep 2025 15:15




Tgt Ion: 93 Resp: 49318


Ion Ratio Lower Upper

93 100

63 74.8 60.1 90.1

95 32.1 25.4 38.2









#7
Naphthalene-d8
Concen: 0.400 ng
RT: 10.626 min Scan# 941
Delta R.T. -0.000 min
Lab File: BN037866.D
Acq: 23 Sep 2025 15:15




Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2


Tgt Ion:	136	Resp:	15583
Ion	Ratio	Lower	Upper
136	100		
137	11.3	9.0	13.4
54	7.5	5.8	8.8
68	5.7	4.3	6.5



#8
Nitrobenzene-d5
Concen: 3.353 ng
RT: 8.971 min Scan# 786
Delta R.T. -0.000 min
Lab File: BN037866.D
Acq: 23 Sep 2025 15:15

Tgt Ion:	82	Resp:	39859
Ion	Ratio	Lower	Upper
82	100		
128	41.7	34.8	52.2
54	52.0	42.2	63.2









#9
Naphthalene
Concen: 3.301 ng
RT: 10.680 min Scan# 945
Delta R.T. 0.011 min
Lab File: BN037866.D
Acq: 23 Sep 2025 15:15




Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2


Tgt Ion:128 Resp: 141732
Ion Ratio Lower Upper
128 100
129 10.9 9.2 13.8
127 12.7 11.0 16.6



#10
Hexachlorobutadiene
Concen: 3.208 ng
RT: 10.979 min Scan# 974
Delta R.T. -0.000 min
Lab File: BN037866.D
Acq: 23 Sep 2025 15:15

Tgt Ion:225 Resp: 23493
Ion Ratio Lower Upper
225 100
223 0.0 0.0 0.0
227 63.8 51.4 77.2









#11
2-Methylnaphthalene-d10
Concen: 3.382 ng
RT: 12.222 min Scan# 1164
Delta R.T. 0.005 min
Lab File: BN037866.D
Acq: 23 Sep 2025 15:15




Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2


Tgt Ion:152 Resp: 73211
Ion Ratio Lower Upper
152 100
151 21.3 17.4 26.2



#12
2-Methylnaphthalene
Concen: 3.447 ng
RT: 12.293 min Scan# 1178
Delta R.T. -0.000 min
Lab File: BN037866.D
Acq: 23 Sep 2025 15:15

Tgt Ion:142 Resp: 96667
Ion Ratio Lower Upper
142 100
141 88.9 71.2 106.8
115 35.7 29.7 44.5









#13
Acenaphthene-d10
Concen: 0.400 ng
RT: 14.473 min Scan# 1456
Delta R.T. -0.000 min
Lab File: BN037866.D
Acq: 23 Sep 2025 15:15




Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2


Tgt Ion:164 Resp: 8623
Ion Ratio Lower Upper
164 100
162 105.8 84.8 127.2
160 56.4 46.1 69.1



#14
2,4,6-Tribromophenol
Concen: 3.642 ng
RT: 15.969 min Scan# 1590
Delta R.T. -0.000 min
Lab File: BN037866.D
Acq: 23 Sep 2025 15:15

Tgt Ion:330 Resp: 12677
Ion Ratio Lower Upper
330 100
332 97.5 77.2 115.8
141 44.0 37.2 55.8









#15
2-Fluorobiphenyl
Concen: 3.223 ng
RT: 13.094 min Scan# 1327
Delta R.T. -0.000 min
Lab File: BN037866.D
Acq: 23 Sep 2025 15:15




Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2


Tgt Ion:172 Resp: 113791
Ion Ratio Lower Upper
172 100
171 36.1 28.6 43.0
170 24.6 19.7 29.5



#16
Acenaphthylene
Concen: 3.437 ng
RT: 14.195 min Scan# 1430
Delta R.T. -0.000 min
Lab File: BN037866.D
Acq: 23 Sep 2025 15:15

Tgt Ion:152 Resp: 134565
Ion Ratio Lower Upper
152 100
151 19.7 15.8 23.8
153 13.0 10.5 15.7









#17
Acenaphthene
Concen: 3.392 ng
RT: 14.537 min Scan# 1462
Delta R.T. -0.000 min
Lab File: BN037866.D
Acq: 23 Sep 2025 15:15




Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2


Tgt Ion:154 Resp: 94611
Ion Ratio Lower Upper
154 100
153 109.6 90.5 135.7
152 58.7 51.8 77.8



#18
Fluorene
Concen: 3.511 ng
RT: 15.522 min Scan# 1554
Delta R.T. -0.000 min
Lab File: BN037866.D
Acq: 23 Sep 2025 15:15

Tgt Ion:166 Resp: 118468
Ion Ratio Lower Upper
166 100
165 98.2 79.0 118.4
167 13.1 10.6 16.0





#19

Phenanthrene-d10

Concen: 0.400 ng

RT: 17.223 min Scan# 1691

Delta R.T. -0.000 min

Lab File: BN037866.D

Acq: 23 Sep 2025 15:15

Instrument :

BNA_N

ClientSampleId :

SSTDICC3.2





Tgt Ion:188 Resp: 16225

Ion Ratio Lower Upper

188 100

94 0.0 0.0 0.0

80 10.3 7.6 11.4



#20

4,6-Dinitro-2-methylphenol

Concen: 3.211 ng

RT: 15.597 min Scan# 1560

Delta R.T. -0.000 min

Lab File: BN037866.D

Acq: 23 Sep 2025 15:15




Tgt Ion:198 Resp: 9172


Ion Ratio Lower Upper

198 100

51 33.8 59.1 88.7#

105 34.5 41.3 61.9#





#21

4-Bromophenyl-phenylether

Concen: 3.370 ng

RT: 16.416 min Scan# 1626

Delta R.T. -0.000 min

Lab File: BN037866.D

Acq: 23 Sep 2025 15:15

Instrument :

BNA_N

ClientSampleId :

SSTDICC3.2





Tgt Ion:248 Resp: 35637

Ion Ratio Lower Upper

248 100

250 96.5 77.6 116.4

141 68.8 48.8 73.2



#22

Hexachlorobenzene

Concen: 3.269 ng

RT: 16.540 min Scan# 1636

Delta R.T. -0.000 min

Lab File: BN037866.D

Acq: 23 Sep 2025 15:15




Tgt Ion:284 Resp: 41902


Ion Ratio Lower Upper

284 100

142 38.5 30.9 46.3

249 29.5 23.9 35.9









#23
Atrazine
Concen: 3.616 ng
RT: 16.689 min Scan# 1648
Delta R.T. -0.000 min
Lab File: BN037866.D
Acq: 23 Sep 2025 15:15




Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2


Tgt Ion:200 Resp: 29112
Ion Ratio Lower Upper
200 100
173 22.0 18.3 27.5
215 45.5 38.6 58.0



#24
Pentachlorophenol
Concen: 4.011 ng
RT: 16.875 min Scan# 1663
Delta R.T. 0.000 min
Lab File: BN037866.D
Acq: 23 Sep 2025 15:15

Tgt Ion:266 Resp: 17645
Ion Ratio Lower Upper
266 100
264 62.5 50.9 76.3
268 64.0 54.3 81.5









#25
Phenanthrene
Concen: 3.352 ng
RT: 17.260 min Scan# 1694
Delta R.T. -0.000 min
Lab File: BN037866.D
Acq: 23 Sep 2025 15:15




Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2


Tgt Ion:178 Resp: 178985
Ion Ratio Lower Upper
178 100
176 19.2 15.7 23.5
179 15.1 12.3 18.5



#26
Anthracene
Concen: 3.573 ng
RT: 17.360 min Scan# 1702
Delta R.T. -0.000 min
Lab File: BN037866.D
Acq: 23 Sep 2025 15:15

Tgt Ion:178 Resp: 165527
Ion Ratio Lower Upper
178 100
176 18.6 15.0 22.4
179 15.2 12.3 18.5





#27

Fluoranthene-d10

Concen: 3.451 ng

RT: 19.257 min Scan# 1954

Delta R.T. 0.005 min

Lab File: BN037866.D

Acq: 23 Sep 2025 15:15

Instrument :

BNA_N

ClientSampleId :

SSTDICC3.2





Tgt Ion:212 Resp: 140845

Ion Ratio Lower Upper

212 100

106 15.7 12.6 19.0

104 9.0 7.4 11.0



#28

Fluoranthene

Concen: 3.570 ng

RT: 19.285 min Scan# 1961

Delta R.T. -0.000 min

Lab File: BN037866.D

Acq: 23 Sep 2025 15:15




Tgt Ion:202 Resp: 209920


Ion Ratio Lower Upper

202 100

101 11.7 9.3 13.9

203 17.3 13.6 20.4





#29

Chrysene-d12

Concen: 0.400 ng

RT: 21.402 min Scan# 21

Delta R.T. -0.000 min

Lab File: BN037866.D

Acq: 23 Sep 2025 15:15

Instrument :

BNA_N

ClientSampleId :

SSTDICC3.2





Tgt Ion:240 Resp: 16309

Ion Ratio Lower Upper

240 100

120 18.9 11.4 17.2#

236 29.3 23.0 34.6



#30

Pyrene

Concen: 3.232 ng

RT: 19.647 min Scan# 2039

Delta R.T. -0.000 min

Lab File: BN037866.D

Acq: 23 Sep 2025 15:15




Tgt Ion:202 Resp: 208135


Ion Ratio Lower Upper

202 100

200 21.0 16.7 25.1

203 17.7 14.2 21.4









#31
Terphenyl-d14
Concen: 3.250 ng
RT: 19.852 min Scan# 2083
Delta R.T. -0.000 min
Lab File: BN037866.D
Acq: 23 Sep 2025 15:15




Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2


Tgt Ion:244 Resp: 105582
Ion Ratio Lower Upper
244 100
212 9.1 7.6 11.4
122 17.9 13.9 20.9



#32
Benzo(a)anthracene
Concen: 3.394 ng
RT: 21.393 min Scan# 2271
Delta R.T. 0.009 min
Lab File: BN037866.D
Acq: 23 Sep 2025 15:15

Tgt Ion:228 Resp: 193464
Ion Ratio Lower Upper
228 100
226 26.7 22.6 33.8
229 19.8 15.6 23.4





#33

Chrysene

Concen: 3.201 ng

RT: 21.438 min Scan# 21

Delta R.T. -0.000 min

Lab File: BN037866.D

Acq: 23 Sep 2025 15:15

Instrument :

BNA_N

ClientSampleId :

SSTDICC3.2





Tgt Ion:228 Resp: 197283

Ion Ratio Lower Upper

228 100

226 31.3 24.6 37.0

229 19.2 15.6 23.4



#34

Bis(2-ethylhexyl)phthalate

Concen: 3.291 ng

RT: 21.331 min Scan# 2264

Delta R.T. 0.009 min

Lab File: BN037866.D

Acq: 23 Sep 2025 15:15




Tgt Ion:149 Resp: 74195


Ion Ratio Lower Upper

149 100

167 26.4 20.4 30.6

279 3.2 2.4 3.6





#35

Perylene-d12

Concen: 0.400 ng

RT: 23.736 min Scan# 2839

Delta R.T. 0.009 min

Lab File: BN037866.D

Acq: 23 Sep 2025 15:15

Instrument :

BNA_N

ClientSampleId :

SSTDICC3.2





Tgt Ion:264 Resp: 13684

Ion Ratio Lower Upper

264 100

260 27.8 21.5 32.3

265 66.7 53.8 80.6



#36

Indeno(1,2,3-cd)pyrene

Concen: 3.494 ng

RT: 26.118 min Scan# 3654

Delta R.T. 0.012 min

Lab File: BN037866.D

Acq: 23 Sep 2025 15:15




Tgt Ion:276 Resp: 218487


Ion Ratio Lower Upper

276 100

138 26.7 21.4 32.0

277 25.1 20.3 30.5





#37

Benzo(b)fluoranthene

Concen: 3.428 ng

RT: 23.031 min Scan# 21

Delta R.T. 0.006 min

Lab File: BN037866.D


Acq: 23 Sep 2025 15:15

Instrument :

BNA_N

ClientSampleId :

SSTDICC3.2





Tgt Ion:252 Resp: 202282

Ion Ratio Lower Upper

252 100

253 21.7 19.4 29.0

125 11.7 13.0 19.4#



#38

Benzo(k)fluoranthene

Concen: 3.534 ng

RT: 23.078 min Scan# 2614

Delta R.T. 0.006 min

Lab File: BN037866.D

Acq: 23 Sep 2025 15:15




Tgt Ion:252 Resp: 209819


Ion Ratio Lower Upper

252 100

253 21.9 19.5 29.3

125 11.5 13.0 19.4#





#39

Benzo(a)pyrene

Concen: 3.499 ng

RT: 23.636 min Scan# 2805

Delta R.T. 0.012 min

Lab File: BN037866.D


Acq: 23 Sep 2025 15:15

Instrument :

BNA_N

ClientSampleId :

SSTDICC3.2





Tgt Ion:252 Resp: 163321

Ion Ratio Lower Upper

252 100

253 21.8 20.6 30.8

125 12.9 16.7 25.1#



#40

Dibenzo(a,h)anthracene

Concen: 3.558 ng

RT: 26.133 min Scan# 3659

Delta R.T. 0.015 min

Lab File: BN037866.D

Acq: 23 Sep 2025 15:15




Tgt Ion:278 Resp: 173724


Ion Ratio Lower Upper

278 100

139 17.8 15.4 23.2

279 23.6 20.8 31.2





#41

Benzo(g,h,i)perylene

Concen: 3.443 ng

RT: 26.841 min Scan# 3901

Delta R.T. 0.012 min

Lab File: BN037866.D


Acq: 23 Sep 2025 15:15

Instrument :

BNA_N

ClientSampleId :

SSTDICC3.2





Tgt Ion:276 Resp: 176617

Ion Ratio Lower Upper

276 100

277 23.8 20.2 30.4

138 23.0 19.8 29.8



Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN092325\
 Data File : BN037867.D
 Acq On : 23 Sep 2025 15:51
 Operator : RC/JU
 Sample : SSTDICC5.0
 Misc :
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC5.0

Quant Time: Sep 23 16:23:34 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN092325.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Sep 23 16:17:53 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.825	152	6080	0.400	ng	0.00
7) Naphthalene-d8	10.626	136	17301	0.400	ng	0.00
13) Acenaphthene-d10	14.473	164	9570	0.400	ng	0.00
19) Phenanthrene-d10	17.223	188	17948	0.400	ng	0.00
29) Chrysene-d12	21.411	240	18531	0.400	ng	# 0.00
35) Perylene-d12	23.736	264	15649	0.400	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.392	112	75850	5.467	ng	0.00
5) Phenol-d6	6.980	99	102802	5.639	ng	0.00
8) Nitrobenzene-d5	8.971	82	73146	5.543	ng	0.00
11) 2-Methylnaphthalene-d10	12.222	152	154332	6.422	ng	0.00
14) 2,4,6-Tribromophenol	0.000	330	0d	0.000	ng	
15) 2-Fluorobiphenyl	13.094	172	210748	5.379	ng	0.00
27) Fluoranthene-d10	19.257	212	307332	6.807	ng	0.00
31) Terphenyl-d14	19.856	244	206917	5.606	ng	0.00
Target Compounds						Qvalue
2) 1,4-Dioxane	3.290	88	28576	4.630	ng	98
3) n-Nitrosodimethylamine	3.593	42	41574	5.063	ng	# 92
6) bis(2-Chloroethyl)ether	7.240	93	86881	5.134	ng	99
9) Naphthalene	10.680	128	250883	5.264	ng	98
10) Hexachlorobutadiene	10.979	225	40670	5.003	ng	# 100
12) 2-Methylnaphthalene	12.293	142	173335	5.567	ng	99
16) Acenaphthylene	14.195	152	250496	5.765	ng	100
17) Acenaphthene	14.537	154	171955	5.554	ng	94
18) Fluorene	15.522	166	209070	5.583	ng	97
20) 4,6-Dinitro-2-methylph...	15.597	198	18771	4.997	ng	# 59
21) 4-Bromophenyl-phenylether	16.416	248	65230	5.577	ng	92
22) Hexachlorobenzene	16.540	284	74475	5.252	ng	99
23) Atrazine	16.689	200	55972	6.284	ng	96
25) Phenanthrene	17.260	178	325691	5.514	ng	99
26) Anthracene	17.359	178	305295	5.958	ng	100
28) Fluoranthene	19.285	202	384292	5.908	ng	99
30) Pyrene	19.647	202	388849	5.315	ng	100
32) Benzo(a)anthracene	21.393	228	362927	5.604	ng	98
33) Chrysene	21.447	228	363421	5.189	ng	99
34) Bis(2-ethylhexyl)phtha...	21.331	149	156492	6.109	ng	98
36) Indeno(1,2,3-cd)pyrene	26.115	276	422337	5.906	ng	99
37) Benzo(b)fluoranthene	23.034	252	387512	5.742	ng	# 93
38) Benzo(k)fluoranthene	23.081	252	376732	5.548	ng	# 92
39) Benzo(a)pyrene	23.636	252	314913	5.900	ng	# 88
40) Dibenzo(a,h)anthracene	26.133	278	333427	5.971	ng	96
41) Benzo(g,h,i)perylene	26.840	276	346868	5.913	ng	96


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN092325\
Data File : BN037867.D
Acq On : 23 Sep 2025 15:51
Operator : RC/JU
Sample : SSTDICC5.0
Misc :
ALS Vial : 8 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
SSTDICC5.0

Quant Time: Sep 23 16:23:34 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN092325.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Sep 23 16:17:53 2025
Response via : Initial Calibration









#1
1,4-Dichlorobenzene-d4
Concen: 0.400 ng
RT: 7.825 min Scan# 61
Delta R.T. 0.000 min
Lab File: BN037867.D
Acq: 23 Sep 2025 15:51




Instrument :
BNA_N
ClientSampleId :
SSTDICC5.0


Tgt Ion:152 Resp: 6080
Ion Ratio Lower Upper
152 100
150 155.7 121.0 181.4
115 58.9 47.4 71.0



#2
1,4-Dioxane
Concen: 4.630 ng
RT: 3.290 min Scan# 29
Delta R.T. -0.007 min
Lab File: BN037867.D
Acq: 23 Sep 2025 15:51

Tgt Ion: 88 Resp: 28576
Ion Ratio Lower Upper
88 100
43 29.8 26.6 39.8
58 73.3 58.9 88.3









#3
n-Nitrosodimethylamine
Concen: 5.063 ng
RT: 3.593 min Scan# 71
Delta R.T. -0.007 min
Lab File: BN037867.D
Acq: 23 Sep 2025 15:51




Instrument :
BNA_N
ClientSampleId :
SSTDICC5.0


Tgt Ion: 42 Resp: 41574
Ion Ratio Lower Upper
42 100
74 117.0 93.4 140.0
44 6.5 22.2 33.4#



#4
2-Fluorophenol
Concen: 5.467 ng
RT: 5.392 min Scan# 320
Delta R.T. 0.007 min
Lab File: BN037867.D
Acq: 23 Sep 2025 15:51

Tgt Ion: 112 Resp: 75850
Ion Ratio Lower Upper
112 100
64 55.3 44.6 66.8
63 29.7 24.9 37.3









#5
Phenol-d6
Concen: 5.639 ng
RT: 6.980 min Scan# 54
Delta R.T. 0.007 min
Lab File: BN037867.D
Acq: 23 Sep 2025 15:51




Instrument :
BNA_N
ClientSampleId :
SSTDICC5.0


Tgt Ion: 99 Resp: 102802
Ion Ratio Lower Upper
99 100
42 20.6 17.2 25.8
71 33.3 27.3 40.9



#6
bis(2-Chloroethyl)ether
Concen: 5.134 ng
RT: 7.240 min Scan# 576
Delta R.T. -0.000 min
Lab File: BN037867.D
Acq: 23 Sep 2025 15:51

Tgt Ion: 93 Resp: 86881
Ion Ratio Lower Upper
93 100
63 74.6 60.1 90.1
95 32.2 25.4 38.2









#7
Naphthalene-d8
Concen: 0.400 ng
RT: 10.626 min Scan# 941
Delta R.T. -0.000 min
Lab File: BN037867.D
Acq: 23 Sep 2025 15:51




Instrument :
BNA_N
ClientSampleId :
SSTDICC5.0


Tgt Ion:	136	Resp:	17301
Ion	Ratio	Lower	Upper
136	100		
137	11.0	9.0	13.4
54	7.8	5.8	8.8
68	5.8	4.3	6.5



#8
Nitrobenzene-d5
Concen: 5.543 ng
RT: 8.971 min Scan# 786
Delta R.T. -0.000 min
Lab File: BN037867.D
Acq: 23 Sep 2025 15:51

Tgt Ion:	82	Resp:	73146
Ion	Ratio	Lower	Upper
82	100		
128	41.2	34.8	52.2
54	52.4	42.2	63.2









#9
Naphthalene
Concen: 5.264 ng
RT: 10.680 min Scan# 945
Delta R.T. 0.011 min
Lab File: BN037867.D
Acq: 23 Sep 2025 15:51




Instrument :
BNA_N
ClientSampleId :
SSTDICC5.0


Tgt Ion:128 Resp: 250883
Ion Ratio Lower Upper
128 100
129 10.8 9.2 13.8
127 12.8 11.0 16.6



#10
Hexachlorobutadiene
Concen: 5.003 ng
RT: 10.979 min Scan# 974
Delta R.T. -0.000 min
Lab File: BN037867.D
Acq: 23 Sep 2025 15:51

Tgt Ion:225 Resp: 40670
Ion Ratio Lower Upper
225 100
223 0.0 0.0 0.0
227 64.0 51.4 77.2









#11
2-Methylnaphthalene-d10
Concen: 6.422 ng
RT: 12.222 min Scan# 1164
Delta R.T. 0.005 min
Lab File: BN037867.D
Acq: 23 Sep 2025 15:51




Instrument :
BNA_N
ClientSampleId :
SSTDICC5.0


Tgt Ion:152 Resp: 154332
Ion Ratio Lower Upper
152 100
151 21.2 17.4 26.2



#12
2-Methylnaphthalene
Concen: 5.567 ng
RT: 12.293 min Scan# 1178
Delta R.T. -0.000 min
Lab File: BN037867.D
Acq: 23 Sep 2025 15:51

Tgt Ion:142 Resp: 173335
Ion Ratio Lower Upper
142 100
141 88.7 71.2 106.8
115 35.2 29.7 44.5









#13
Acenaphthene-d10
Concen: 0.400 ng
RT: 14.473 min Scan# 14
Delta R.T. -0.000 min
Lab File: BN037867.D
Acq: 23 Sep 2025 15:51

Instrument :
BNA_N
ClientSampleId :
SSTDICC5.0


Tgt Ion:164 Resp: 9570
Ion Ratio Lower Upper
164 100
162 105.9 84.8 127.2
160 55.3 46.1 69.1




#14
2,4,6-Tribromophenol
Concen: 0.000 ng
Expected RT: 15.97 min

Lab File: BN037867.D
Acq: 23 Sep 2025 15:51

Tgt Ion: 330
Sig Exp Ratio
330 100
332 96.5
141 46.5









#15
2-Fluorobiphenyl
Concen: 5.379 ng
RT: 13.094 min Scan# 1327
Delta R.T. -0.000 min
Lab File: BN037867.D
Acq: 23 Sep 2025 15:51




Instrument :
BNA_N
ClientSampleId :
SSTDICC5.0


Tgt Ion:172 Resp: 210748
Ion Ratio Lower Upper
172 100
171 35.8 28.6 43.0
170 24.4 19.7 29.5



#16
Acenaphthylene
Concen: 5.765 ng
RT: 14.195 min Scan# 1430
Delta R.T. -0.000 min
Lab File: BN037867.D
Acq: 23 Sep 2025 15:51

Tgt Ion:152 Resp: 250496
Ion Ratio Lower Upper
152 100
151 19.8 15.8 23.8
153 13.1 10.5 15.7









#17
Acenaphthene
Concen: 5.554 ng
RT: 14.537 min Scan# 1462
Delta R.T. -0.000 min
Lab File: BN037867.D
Acq: 23 Sep 2025 15:51




Instrument :
BNA_N
ClientSampleId :
SSTDICC5.0


Tgt Ion:154 Resp: 171955
Ion Ratio Lower Upper
154 100
153 108.8 90.5 135.7
152 56.5 51.8 77.8



#18
Fluorene
Concen: 5.583 ng
RT: 15.522 min Scan# 1554
Delta R.T. -0.000 min
Lab File: BN037867.D
Acq: 23 Sep 2025 15:51

Tgt Ion:166 Resp: 209070
Ion Ratio Lower Upper
166 100
165 95.7 79.0 118.4
167 12.9 10.6 16.0





#19

Phenanthrene-d10

Concen: 0.400 ng

RT: 17.223 min Scan# 1

Delta R.T. -0.000 min

Lab File: BN037867.D

Acq: 23 Sep 2025 15:51

Instrument :

BNA_N

ClientSampleId :

SSTDICC5.0





Tgt Ion:188 Resp: 17948

Ion Ratio Lower Upper

188 100

94 0.0 0.0 0.0

80 10.8 7.6 11.4



#20

4,6-Dinitro-2-methylphenol

Concen: 4.997 ng

RT: 15.597 min Scan# 1560

Delta R.T. -0.000 min

Lab File: BN037867.D

Acq: 23 Sep 2025 15:51




Tgt Ion:198 Resp: 18771


Ion Ratio Lower Upper

198 100

51 30.7 59.1 88.7#

105 32.7 41.3 61.9#









#21
4-Bromophenyl-phenylether
Concen: 5.577 ng
RT: 16.416 min Scan# 1626
Delta R.T. -0.000 min
Lab File: BN037867.D
Acq: 23 Sep 2025 15:51




Instrument :
BNA_N
ClientSampleId :
SSTDICC5.0


Tgt Ion:248 Resp: 65230
Ion Ratio Lower Upper
248 100
250 94.2 77.6 116.4
141 72.6 48.8 73.2



#22
Hexachlorobenzene
Concen: 5.252 ng
RT: 16.540 min Scan# 1636
Delta R.T. -0.000 min
Lab File: BN037867.D
Acq: 23 Sep 2025 15:51

Tgt Ion:284 Resp: 74475
Ion Ratio Lower Upper
284 100
142 38.4 30.9 46.3
249 29.6 23.9 35.9









#23
Atrazine
Concen: 6.284 ng
RT: 16.689 min Scan# 1648
Delta R.T. -0.000 min
Lab File: BN037867.D
Acq: 23 Sep 2025 15:51




Instrument :
BNA_N
ClientSampleId :
SSTDICC5.0


Tgt Ion	200	Resp	55972
Ion Ratio	Lower	Upper	
200	100		
173	21.8	18.3	27.5
215	45.5	38.6	58.0



#25
Phenanthrene
Concen: 5.514 ng
RT: 17.260 min Scan# 1694
Delta R.T. -0.000 min
Lab File: BN037867.D
Acq: 23 Sep 2025 15:51

Tgt Ion	178	Resp	325691
Ion Ratio	Lower	Upper	
178	100		
176	19.1	15.7	23.5
179	15.1	12.3	18.5









#26
Anthracene
Concen: 5.958 ng
RT: 17.359 min Scan# 1702
Delta R.T. -0.000 min
Lab File: BN037867.D
Acq: 23 Sep 2025 15:51




Instrument :
BNA_N
ClientSampleId :
SSTDICC5.0


Tgt Ion:178 Resp: 305295
Ion Ratio Lower Upper
178 100
176 18.6 15.0 22.4
179 15.2 12.3 18.5



#27
Fluoranthene-d10
Concen: 6.807 ng
RT: 19.257 min Scan# 1955
Delta R.T. 0.005 min
Lab File: BN037867.D
Acq: 23 Sep 2025 15:51

Tgt Ion:212 Resp: 307332
Ion Ratio Lower Upper
212 100
106 15.7 12.6 19.0
104 9.0 7.4 11.0









#28
Fluoranthene
Concen: 5.908 ng
RT: 19.285 min Scan# 1961
Delta R.T. -0.000 min
Lab File: BN037867.D
Acq: 23 Sep 2025 15:51




Instrument :
BNA_N
ClientSampleId :
SSTDICC5.0


Tgt Ion:202 Resp: 384292
Ion Ratio Lower Upper
202 100
101 11.7 9.3 13.9
203 17.3 13.6 20.4



#29
Chrysene-d12
Concen: 0.400 ng
RT: 21.411 min Scan# 2273
Delta R.T. 0.009 min
Lab File: BN037867.D
Acq: 23 Sep 2025 15:51

Tgt Ion:240 Resp: 18531
Ion Ratio Lower Upper
240 100
120 10.0 11.4 17.2#
236 27.5 23.0 34.6





#30

Pyrene

Concen: 5.315 ng

RT: 19.647 min Scan# 2039

Delta R.T. -0.000 min

Lab File: BN037867.D

Acq: 23 Sep 2025 15:51

Instrument :

BNA_N

ClientSampleId :

SSTDICC5.0




Tgt Ion:202 Resp: 388849

Ion Ratio Lower Upper

202 100

200 21.0 16.7 25.1

203 17.7 14.2 21.4



#31

Terphenyl-d14

Concen: 5.606 ng

RT: 19.856 min Scan# 2084

Delta R.T. 0.005 min

Lab File: BN037867.D

Acq: 23 Sep 2025 15:51




Tgt Ion:244 Resp: 206917


Ion Ratio Lower Upper

244 100

212 8.8 7.6 11.4

122 15.3 13.9 20.9









#32
Benzo(a)anthracene
Concen: 5.604 ng
RT: 21.393 min Scan# 21393
Delta R.T. 0.009 min
Lab File: BN037867.D
Acq: 23 Sep 2025 15:51




Instrument :
BNA_N
ClientSampleId :
SSTDICC5.0


Tgt Ion:228 Resp: 362927
Ion Ratio Lower Upper
228 100
226 26.9 22.6 33.8
229 19.7 15.6 23.4



#33
Chrysene
Concen: 5.189 ng
RT: 21.447 min Scan# 2277
Delta R.T. 0.009 min
Lab File: BN037867.D
Acq: 23 Sep 2025 15:51

Tgt Ion:228 Resp: 363421
Ion Ratio Lower Upper
228 100
226 29.6 24.6 37.0
229 19.5 15.6 23.4









#34
Bis(2-ethylhexyl)phthalate
Concen: 6.109 ng
RT: 21.331 min Scan# 21
Delta R.T. 0.009 min
Lab File: BN037867.D
Acq: 23 Sep 2025 15:51




Instrument :
BNA_N
ClientSampleId :
SSTDICC5.0


Tgt Ion:149 Resp: 156492
Ion Ratio Lower Upper
149 100
167 26.8 20.4 30.6
279 3.2 2.4 3.6



#35
Perylene-d12
Concen: 0.400 ng
RT: 23.736 min Scan# 2839
Delta R.T. 0.009 min
Lab File: BN037867.D
Acq: 23 Sep 2025 15:51

Tgt Ion:264 Resp: 15649
Ion Ratio Lower Upper
264 100
260 26.8 21.5 32.3
265 63.5 53.8 80.6





#36

Indeno(1,2,3-cd)pyrene

Concen: 5.906 ng

RT: 26.115 min Scan# 30

Delta R.T. 0.009 min

Lab File: BN037867.D


Acq: 23 Sep 2025 15:51

Instrument :

BNA_N

ClientSampleId :

SSTDICC5.0






Tgt Ion:276 Resp: 422337

Ion Ratio Lower Upper

276 100

138 26.3 21.4 32.0

277 25.1 20.3 30.5



#37

Benzo(b)fluoranthene

Concen: 5.742 ng

RT: 23.034 min Scan# 2599

Delta R.T. 0.009 min

Lab File: BN037867.D

Acq: 23 Sep 2025 15:51




Tgt Ion:252 Resp: 387512


Ion Ratio Lower Upper

252 100

253 22.0 19.4 29.0

125 11.1 13.0 19.4#





#38

Benzo(k)fluoranthene

Concen: 5.548 ng

RT: 23.081 min Scan# 20

Delta R.T. 0.009 min

Lab File: BN037867.D

Acq: 23 Sep 2025 15:51

Instrument :

BNA_N

ClientSampleId :

SSTDICC5.0





Tgt Ion:252 Resp: 376732

Ion Ratio Lower Upper

252 100

253 21.8 19.5 29.3

125 11.2 13.0 19.4#



#39

Benzo(a)pyrene

Concen: 5.900 ng

RT: 23.636 min Scan# 2805

Delta R.T. 0.012 min

Lab File: BN037867.D

Acq: 23 Sep 2025 15:51




Tgt Ion:252 Resp: 314913


Ion Ratio Lower Upper

252 100

253 22.0 20.6 30.8

125 12.3 16.7 25.1#









#40
Dibenzo(a,h)anthracene
Concen: 5.971 ng
RT: 26.133 min Scan# 301
Delta R.T. 0.015 min
Lab File: BN037867.D
Acq: 23 Sep 2025 15:51




Instrument :
BNA_N
ClientSampleId :
SSTDICC5.0

Tgt Ion:278 Resp: 333427
Ion Ratio Lower Upper
278 100
139 17.5 15.4 23.2
279 23.6 20.8 31.2



#41
Benzo(g,h,i)perylene
Concen: 5.913 ng
RT: 26.840 min Scan# 3901
Delta R.T. 0.012 min
Lab File: BN037867.D
Acq: 23 Sep 2025 15:51

Tgt Ion:276 Resp: 346868
Ion Ratio Lower Upper
276 100
277 23.8 20.2 30.4
138 22.7 19.8 29.8



Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN092325\
 Data File : BN037868.D
 Acq On : 23 Sep 2025 16:28
 Operator : RC/JU
 Sample : SSTDICV0.4
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 ICVBN092325

Quant Time: Sep 24 11:02:35 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN092325.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Sep 24 11:00:01 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.825	152	5572	0.400	ng	0.00
7) Naphthalene-d8	10.626	136	16158	0.400	ng	0.00
13) Acenaphthene-d10	14.473	164	8429	0.400	ng	0.00
19) Phenanthrene-d10	17.223	188	16279	0.400	ng	# 0.00
29) Chrysene-d12	21.402	240	14204	0.400	ng	# 0.00
35) Perylene-d12	23.736	264	13265	0.400	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.392	112	5005	0.394	ng	0.00
5) Phenol-d6	6.980	99	6375	0.382	ng	0.00
8) Nitrobenzene-d5	8.971	82	5453	0.442	ng	0.00
11) 2-Methylnaphthalene-d10	12.222	152	9492	0.423	ng	0.00
14) 2,4,6-Tribromophenol	15.969	330	1107	0.346	ng	0.00
15) 2-Fluorobiphenyl	13.105	172	16723	0.485	ng	0.01
27) Fluoranthene-d10	19.257	212	16910	0.413	ng	0.00
31) Terphenyl-d14	19.856	244	13646	0.482	ng	0.00
Target Compounds						
						Qvalue
2) 1,4-Dioxane	3.297	88	2698	0.477	ng	98
3) n-Nitrosodimethylamine	3.600	42	3070	0.408	ng	# 89
6) bis(2-Chloroethyl)ether	7.240	93	6823	0.440	ng	99
9) Naphthalene	10.680	128	18392	0.413	ng	99
10) Hexachlorobutadiene	10.979	225	3253	0.428	ng	# 100
12) 2-Methylnaphthalene	12.293	142	10904	0.375	ng	99
16) Acenaphthylene	14.195	152	16721	0.437	ng	99
17) Acenaphthene	14.537	154	11059	0.406	ng	99
18) Fluorene	15.522	166	13327	0.404	ng	97
20) 4,6-Dinitro-2-methylph...	15.597	198	737	0.418	ng	99
21) 4-Bromophenyl-phenylether	16.416	248	4157	0.392	ng	# 89
22) Hexachlorobenzene	16.540	284	5455	0.424	ng	100
23) Atrazine	16.689	200	3196	0.396	ng	98
24) Pentachlorophenol	16.875	266	1396	0.316	ng	99
25) Phenanthrene	17.260	178	22020	0.411	ng	100
26) Anthracene	17.360	178	18967	0.408	ng	99
28) Fluoranthene	19.285	202	23044	0.391	ng	100
30) Pyrene	19.647	202	22563	0.402	ng	100
32) Benzo(a)anthracene	21.393	228	20398	0.411	ng	99
33) Chrysene	21.438	228	22204	0.414	ng	99
34) Bis(2-ethylhexyl)phtha...	21.331	149	7939	0.404	ng	98
36) Indeno(1,2,3-cd)pyrene	26.115	276	25340	0.418	ng	99
37) Benzo(b)fluoranthene	23.031	252	22193	0.388	ng	99
38) Benzo(k)fluoranthene	23.078	252	23886	0.415	ng	100
39) Benzo(a)pyrene	23.633	252	19091	0.422	ng	100
40) Dibenzo(a,h)anthracene	26.130	278	19441	0.411	ng	100
41) Benzo(g,h,i)perylene	26.841	276	19570	0.394	ng	99


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN092325\
Data File : BN037868.D
Acq On : 23 Sep 2025 16:28
Operator : RC/JU
Sample : SSTDICV0.4
Misc :
ALS Vial : 9 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
ICVBN092325


Quant Time: Sep 24 11:02:35 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN092325.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Sep 24 11:00:01 2025
Response via : Initial Calibration








#1
1,4-Dichlorobenzene-d4
Concen: 0.400 ng
RT: 7.825 min Scan# 61
Delta R.T. 0.000 min
Lab File: BN037868.D
Acq: 23 Sep 2025 16:28

Instrument :
BNA_N
ClientSampleId :
ICVBN092325







Tgt Ion:152 Resp: 5572
Ion Ratio Lower Upper
152 100
150 154.8 121.0 181.4
115 61.0 47.4 71.0



#2
1,4-Dioxane
Concen: 0.477 ng
RT: 3.297 min Scan# 30
Delta R.T. -0.000 min
Lab File: BN037868.D
Acq: 23 Sep 2025 16:28

Tgt Ion: 88 Resp: 2698
Ion Ratio Lower Upper
88 100
43 30.6 26.6 39.8
58 72.8 58.9 88.3









#3
n-Nitrosodimethylamine
Concen: 0.408 ng
RT: 3.600 min Scan# 71
Delta R.T. -0.000 min
Lab File: BN037868.D
Acq: 23 Sep 2025 16:28




Instrument :
BNA_N
ClientSampleId :
ICVBN092325


Tgt Ion: 42 Resp: 3070
Ion Ratio Lower Upper
42 100
74 121.8 93.4 140.0
44 9.2 22.2 33.4#



#4
2-Fluorophenol
Concen: 0.394 ng
RT: 5.392 min Scan# 320
Delta R.T. 0.007 min
Lab File: BN037868.D
Acq: 23 Sep 2025 16:28

Tgt Ion: 112 Resp: 5005
Ion Ratio Lower Upper
112 100
64 54.6 44.6 66.8
63 29.8 24.9 37.3









#5
Phenol-d6
Concen: 0.382 ng
RT: 6.980 min Scan# 54
Delta R.T. 0.007 min
Lab File: BN037868.D
Acq: 23 Sep 2025 16:28




Instrument :
BNA_N
ClientSampleId :
ICVBN092325


Tgt Ion: 99 Resp: 6375
Ion Ratio Lower Upper
99 100
42 20.3 17.2 25.8
71 32.5 27.3 40.9



#6
bis(2-Chloroethyl)ether
Concen: 0.440 ng
RT: 7.240 min Scan# 576
Delta R.T. -0.000 min
Lab File: BN037868.D
Acq: 23 Sep 2025 16:28

Tgt Ion: 93 Resp: 6823
Ion Ratio Lower Upper
93 100
63 74.7 60.1 90.1
95 32.3 25.4 38.2









#7
Naphthalene-d8
Concen: 0.400 ng
RT: 10.626 min Scan# 941
Delta R.T. -0.000 min
Lab File: BN037868.D
Acq: 23 Sep 2025 16:28




Instrument :
BNA_N
ClientSampleId :
ICVBN092325


Tgt Ion:	136	Resp:	16158
Ion Ratio	Lower	Upper	
136	100		
137	11.0	9.0	13.4
54	7.7	5.8	8.8
68	6.0	4.3	6.5



#8
Nitrobenzene-d5
Concen: 0.442 ng
RT: 8.971 min Scan# 786
Delta R.T. -0.000 min
Lab File: BN037868.D
Acq: 23 Sep 2025 16:28

Tgt Ion:	82	Resp:	5453
Ion Ratio	Lower	Upper	
82	100		
128	41.4	34.8	52.2
54	52.9	42.2	63.2









#9
Naphthalene
Concen: 0.413 ng
RT: 10.680 min Scan# 945
Delta R.T. 0.011 min
Lab File: BN037868.D
Acq: 23 Sep 2025 16:28




Instrument :
BNA_N
ClientSampleId :
ICVBN092325


Tgt Ion:128 Resp: 18392
Ion Ratio Lower Upper
128 100
129 11.4 9.2 13.8
127 13.1 11.0 16.6



#10
Hexachlorobutadiene
Concen: 0.428 ng
RT: 10.979 min Scan# 974
Delta R.T. -0.000 min
Lab File: BN037868.D
Acq: 23 Sep 2025 16:28

Tgt Ion:225 Resp: 3253
Ion Ratio Lower Upper
225 100
223 0.0 0.0 0.0
227 64.5 51.4 77.2









#11
2-Methylnaphthalene-d10
Concen: 0.423 ng
RT: 12.222 min Scan# 1164
Delta R.T. 0.005 min
Lab File: BN037868.D
Acq: 23 Sep 2025 16:28




Instrument :
BNA_N
ClientSampleId :
ICVBN092325


Tgt Ion:152 Resp: 9492
Ion Ratio Lower Upper
152 100
151 21.7 17.4 26.2



#12
2-Methylnaphthalene
Concen: 0.375 ng
RT: 12.293 min Scan# 1178
Delta R.T. -0.000 min
Lab File: BN037868.D
Acq: 23 Sep 2025 16:28

Tgt Ion:142 Resp: 10904
Ion Ratio Lower Upper
142 100
141 88.6 71.2 106.8
115 36.5 29.7 44.5









#13
Acenaphthene-d10
Concen: 0.400 ng
RT: 14.473 min Scan# 1456
Delta R.T. -0.000 min
Lab File: BN037868.D
Acq: 23 Sep 2025 16:28




Instrument :
BNA_N
ClientSampleId :
ICVBN092325


Tgt Ion:164 Resp: 8429
Ion Ratio Lower Upper
164 100
162 106.3 84.8 127.2
160 57.3 46.1 69.1



#14
2,4,6-Tribromophenol
Concen: 0.346 ng
RT: 15.969 min Scan# 1590
Delta R.T. 0.000 min
Lab File: BN037868.D
Acq: 23 Sep 2025 16:28

Tgt Ion:330 Resp: 1107
Ion Ratio Lower Upper
330 100
332 94.1 77.2 115.8
141 42.0 37.2 55.8









#15
2-Fluorobiphenyl
Concen: 0.485 ng
RT: 13.105 min Scan# 1328
Delta R.T. 0.011 min
Lab File: BN037868.D
Acq: 23 Sep 2025 16:28




Instrument :
BNA_N
ClientSampleId :
ICVBN092325


Tgt Ion:172 Resp: 16723
Ion Ratio Lower Upper
172 100
171 34.6 28.6 43.0
170 22.9 19.7 29.5



#16
Acenaphthylene
Concen: 0.437 ng
RT: 14.195 min Scan# 1430
Delta R.T. -0.000 min
Lab File: BN037868.D
Acq: 23 Sep 2025 16:28

Tgt Ion:152 Resp: 16721
Ion Ratio Lower Upper
152 100
151 20.1 15.8 23.8
153 13.0 10.5 15.7









#17
Acenaphthene
Concen: 0.406 ng
RT: 14.537 min Scan# 1462
Delta R.T. -0.000 min
Lab File: BN037868.D
Acq: 23 Sep 2025 16:28




Instrument :
BNA_N
ClientSampleId :
ICVBN092325


Tgt Ion	154	Resp	11059
Ion Ratio	Lower	Upper	
154	100		
153	112.6	90.5	135.7
152	63.5	51.8	77.8



#18
Fluorene
Concen: 0.404 ng
RT: 15.522 min Scan# 1554
Delta R.T. -0.000 min
Lab File: BN037868.D
Acq: 23 Sep 2025 16:28

Tgt Ion	166	Resp	13327
Ion Ratio	Lower	Upper	
166	100		
165	95.9	79.0	118.4
167	13.7	10.6	16.0





#19

Phenanthrene-d10

Concen: 0.400 ng

RT: 17.223 min Scan# 1

Delta R.T. -0.000 min

Lab File: BN037868.D

Acq: 23 Sep 2025 16:28

Instrument :

BNA_N

ClientSampleId :

ICVBN092325





Tgt Ion:188 Resp: 16279

Ion Ratio Lower Upper

188 100

94 0.0 0.0 0.0

80 11.7 7.6 11.4#



#20

4,6-Dinitro-2-methylphenol

Concen: 0.418 ng

RT: 15.597 min Scan# 1560

Delta R.T. -0.000 min

Lab File: BN037868.D

Acq: 23 Sep 2025 16:28




Tgt Ion:198 Resp: 737


Ion Ratio Lower Upper

198 100

51 73.3 59.1 88.7

105 49.9 41.3 61.9









#21
4-Bromophenyl-phenylether
Concen: 0.392 ng
RT: 16.416 min Scan# 1626
Delta R.T. -0.000 min
Lab File: BN037868.D
Acq: 23 Sep 2025 16:28




Instrument :
BNA_N
ClientSampleId :
ICVBN092325


Tgt Ion:248 Resp: 4157
Ion Ratio Lower Upper
248 100
250 94.6 77.6 116.4
141 79.1 48.8 73.2#



#22
Hexachlorobenzene
Concen: 0.424 ng
RT: 16.540 min Scan# 1636
Delta R.T. -0.000 min
Lab File: BN037868.D
Acq: 23 Sep 2025 16:28

Tgt Ion:284 Resp: 5455
Ion Ratio Lower Upper
284 100
142 38.9 30.9 46.3
249 30.0 23.9 35.9









#23
Atrazine
Concen: 0.396 ng
RT: 16.689 min Scan# 1648
Delta R.T. -0.000 min
Lab File: BN037868.D
Acq: 23 Sep 2025 16:28




Instrument :
BNA_N
ClientSampleId :
ICVBN092325


Tgt Ion	Ratio	Lower	Upper
200	100		
173	23.5	18.3	27.5
215	46.6	38.6	58.0



#24
Pentachlorophenol
Concen: 0.316 ng
RT: 16.875 min Scan# 1663
Delta R.T. 0.000 min
Lab File: BN037868.D
Acq: 23 Sep 2025 16:28

Tgt Ion	Ratio	Lower	Upper
266	100		
264	63.8	50.9	76.3
268	66.0	54.3	81.5









#25
Phenanthrene
Concen: 0.411 ng
RT: 17.260 min Scan# 1694
Delta R.T. -0.000 min
Lab File: BN037868.D
Acq: 23 Sep 2025 16:28




Instrument :
BNA_N
ClientSampleId :
ICVBN092325


Tgt Ion:178 Resp: 22020
Ion Ratio Lower Upper
178 100
176 19.3 15.7 23.5
179 15.4 12.3 18.5



#26
Anthracene
Concen: 0.408 ng
RT: 17.360 min Scan# 1702
Delta R.T. -0.000 min
Lab File: BN037868.D
Acq: 23 Sep 2025 16:28

Tgt Ion:178 Resp: 18967
Ion Ratio Lower Upper
178 100
176 18.9 15.0 22.4
179 15.0 12.3 18.5





#27

Fluoranthene-d10

Concen: 0.413 ng

RT: 19.257 min Scan# 1954

Delta R.T. 0.005 min

Lab File: BN037868.D

Acq: 23 Sep 2025 16:28

Instrument :

BNA_N

ClientSampleId :

ICVBN092325





Tgt Ion:212 Resp: 16910

Ion Ratio Lower Upper

212 100

106 15.8 12.6 19.0

104 9.1 7.4 11.0



#28

Fluoranthene

Concen: 0.391 ng

RT: 19.285 min Scan# 1961

Delta R.T. -0.000 min

Lab File: BN037868.D

Acq: 23 Sep 2025 16:28




Tgt Ion:202 Resp: 23044


Ion Ratio Lower Upper

202 100

101 12.0 9.3 13.9

203 17.0 13.6 20.4





#29

Chrysene-d12

Concen: 0.400 ng

RT: 21.402 min Scan# 21

Delta R.T. -0.000 min

Lab File: BN037868.D

Acq: 23 Sep 2025 16:28

Instrument :

BNA_N

ClientSampleId :

ICVBN092325





Tgt Ion:240 Resp: 14204

Ion Ratio Lower Upper

240 100

120 20.0 11.4 17.2#

236 29.8 23.0 34.6



#30

Pyrene

Concen: 0.402 ng

RT: 19.647 min Scan# 2039

Delta R.T. -0.000 min

Lab File: BN037868.D

Acq: 23 Sep 2025 16:28




Tgt Ion:202 Resp: 22563


Ion Ratio Lower Upper

202 100

200 21.0 16.7 25.1


203 17.7 14.2 21.4








#31
Terphenyl-d14
Concen: 0.482 ng
RT: 19.856 min Scan# 2083
Delta R.T. 0.005 min
Lab File: BN037868.D
Acq: 23 Sep 2025 16:28

Instrument :
BNA_N
ClientSampleId :
ICVBN092325







Tgt Ion:244 Resp: 13646
Ion Ratio Lower Upper
244 100
212 9.0 7.6 11.4
122 15.9 13.9 20.9



#32
Benzo(a)anthracene
Concen: 0.411 ng
RT: 21.393 min Scan# 2271
Delta R.T. 0.009 min
Lab File: BN037868.D
Acq: 23 Sep 2025 16:28

Tgt Ion:228 Resp: 20398
Ion Ratio Lower Upper
228 100
226 27.3 22.6 33.8
229 19.8 15.6 23.4





#33

Chrysene

Concen: 0.414 ng

RT: 21.438 min Scan# 21

Delta R.T. -0.000 min

Lab File: BN037868.D

Acq: 23 Sep 2025 16:28

Instrument :

BNA_N

ClientSampleId :

ICVBN092325





Tgt Ion:228 Resp: 22204

Ion Ratio Lower Upper

228 100

226 31.0 24.6 37.0

229 19.1 15.6 23.4



#34

Bis(2-ethylhexyl)phthalate

Concen: 0.404 ng

RT: 21.331 min Scan# 2264

Delta R.T. 0.009 min

Lab File: BN037868.D

Acq: 23 Sep 2025 16:28




Tgt Ion:149 Resp: 7939


Ion Ratio Lower Upper

149 100

167 26.4 20.4 30.6

279 3.2 2.4 3.6









#35
Perylene-d12
Concen: 0.400 ng
RT: 23.736 min Scan# 21
Delta R.T. 0.009 min
Lab File: BN037868.D
Acq: 23 Sep 2025 16:28




Instrument :
BNA_N
ClientSampleId :
ICVBN092325


Tgt Ion	264	260	265
Resp:	13265		
Ion Ratio	100	26.9	71.6
Lower		21.5	53.8
Upper		32.3	80.6



#36
Indeno(1,2,3-cd)pyrene
Concen: 0.418 ng
RT: 26.115 min Scan# 3653
Delta R.T. 0.009 min
Lab File: BN037868.D
Acq: 23 Sep 2025 16:28

Tgt Ion	276	138	277
Resp:	25340		
Ion Ratio	100	27.5	25.0
Lower		21.4	20.3
Upper		32.0	30.5





#37

Benzo(b)fluoranthene

Concen: 0.388 ng

RT: 23.031 min Scan# 2596

Delta R.T. 0.006 min

Lab File: BN037868.D


Acq: 23 Sep 2025 16:28

Instrument :

BNA_N

ClientSampleId :

ICVBN092325





Tgt Ion:252 Resp: 22193

Ion Ratio Lower Upper

252 100

253 24.6 19.4 29.0

125 16.8 13.0 19.4



#38

Benzo(k)fluoranthene


Concen: 0.415 ng

RT: 23.078 min Scan# 2614

Delta R.T. 0.006 min

Lab File: BN037868.D

Acq: 23 Sep 2025 16:28





Tgt Ion:252 Resp: 23886


Ion Ratio Lower Upper

252 100

253 24.7 19.5 29.3

125 16.1 13.0 19.4





#39

Benzo(a)pyrene

Concen: 0.422 ng

RT: 23.633 min Scan# 21

Delta R.T. 0.009 min

Lab File: BN037868.D


Acq: 23 Sep 2025 16:28

Instrument :

BNA_N

ClientSampleId :

ICVBN092325






Tgt Ion:252 Resp: 19091

Ion Ratio Lower Upper

252 100

253 25.9 20.6 30.8

125 20.9 16.7 25.1



#40

Dibenzo(a,h)anthracene

Concen: 0.411 ng

RT: 26.130 min Scan# 3658

Delta R.T. 0.012 min

Lab File: BN037868.D

Acq: 23 Sep 2025 16:28




Tgt Ion:278 Resp: 19441


Ion Ratio Lower Upper

278 100

139 19.6 15.4 23.2

279 25.8 20.8 31.2





#41

Benzo(g,h,i)perylene

Concen: 0.394 ng

RT: 26.841 min Scan# 3901

Delta R.T. 0.012 min

Lab File: BN037868.D


Acq: 23 Sep 2025 16:28

Instrument :

BNA_N

ClientSampleId :

ICVBN092325





Tgt Ion:276 Resp: 19570

Ion Ratio Lower Upper

276 100

277 24.5 20.2 30.4

138 24.1 19.8 29.8



Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN092325\
 Data File : BN037868.D
 Acq On : 23 Sep 2025 16:28
 Operator : RC/JU
 Sample : SSTDICV0.4
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 ICVBN092325

Quant Time: Sep 24 11:02:35 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN092325.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Sep 24 11:00:01 2025
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	100	0.00
2	1,4-Dioxane	0.406	0.484	-19.2	118	0.00
3	n-Nitrosodimethylamine	0.540	0.551	-2.0	106	0.00
4 S	2-Fluorophenol	0.913	0.898	1.6	105	0.00
5 S	Phenol-d6	1.199	1.144	4.6	102	0.00
6	bis(2-Chloroethyl)ether	1.113	1.225	-10.1	113	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	102	0.00
8 S	Nitrobenzene-d5	0.305	0.337	-10.5	119	0.00
9	Naphthalene	1.102	1.138	-3.3	111	0.01
10	Hexachlorobutadiene	0.188	0.201	-6.9	113	0.00
11 SURR	2-Methylnaphthalene-d10	0.556	0.587	-5.6	121	0.00
12	2-Methylnaphthalene	0.720	0.675	6.2	103	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	102	0.00
14 S	2,4,6-Tribromophenol	0.152	0.131	13.8	99	0.00
15 S	2-Fluorobiphenyl	1.638	1.984	-21.1	130	0.01
16	Acenaphthylene	1.816	1.984	-9.3	122	0.00
17	Acenaphthene	1.294	1.312	-1.4	112	0.00
18	Fluorene	1.565	1.581	-1.0	111	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	102	0.00
20	4,6-Dinitro-2-methylphenol	0.057	0.045	21.1	106	0.00
21	4-Bromophenyl-phenylether	0.261	0.255	2.3	107	0.00
22	Hexachlorobenzene	0.316	0.335	-6.0	112	0.00
23	Atrazine	0.199	0.196	1.5	117	0.00
24	Pentachlorophenol	0.108	0.086	20.4	95	0.00
25	Phenanthrene	1.316	1.353	-2.8	112	0.00
26	Anthracene	1.142	1.165	-2.0	116	0.00
27 SURR	Fluoranthene-d10	1.006	1.039	-3.3	120	0.00
28	Fluoranthene	1.450	1.416	2.3	110	0.00
29 I	Chrysene-d12	1.000	1.000	0.0	105	0.00
30	Pyrene	1.579	1.588	-0.6	109	0.00
31 S	Terphenyl-d14	0.797	0.961	-20.6	133	0.00
32	Benzo(a)anthracene	1.398	1.436	-2.7	116	0.00
33	Chrysene	1.512	1.563	-3.4	112	0.00
34	Bis(2-ethylhexyl)phthalate	0.553	0.559	-1.1	120	0.00
35 I	Perylene-d12	1.000	1.000	0.0	105	0.00
36	Indeno(1,2,3-cd)pyrene	1.828	1.910	-4.5	121	0.00
37	Benzo(b)fluoranthene	1.725	1.673	3.0	108	0.00
38	Benzo(k)fluoranthene	1.736	1.801	-3.7	119	0.00
39 C	Benzo(a)pyrene	1.364	1.439	-5.5	123	0.00
40	Dibenzo(a,h)anthracene	1.427	1.466	-2.7	119	0.01
41	Benzo(g,h,i)perylene	1.499	1.475	1.6	111	0.01

(#) = Out of Range

SPCC's out = 0 CCC's out = 0

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN092325\
 Data File : BN037868.D
 Acq On : 23 Sep 2025 16:28
 Operator : RC/JU
 Sample : SSTDICV0.4
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 ICVBN092325

Quant Time: Sep 24 11:02:35 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN092325.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Sep 24 11:00:01 2025
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	100	0.00
2	1,4-Dioxane	0.400	0.477	-19.2	118	0.00
3	n-Nitrosodimethylamine	0.400	0.408	-2.0	106	0.00
4 S	2-Fluorophenol	0.400	0.394	1.5	105	0.00
5 S	Phenol-d6	0.400	0.382	4.5	102	0.00
6	bis(2-Chloroethyl)ether	0.400	0.440	-10.0	113	0.00
7 I	Naphthalene-d8	0.400	0.400	0.0	102	0.00
8 S	Nitrobenzene-d5	0.400	0.442	-10.5	119	0.00
9	Naphthalene	0.400	0.413	-3.2	111	0.01
10	Hexachlorobutadiene	0.400	0.428	-7.0	113	0.00
11 SURR	2-Methylnaphthalene-d10	0.400	0.423	-5.7	121	0.00
12	2-Methylnaphthalene	0.400	0.375	6.3	103	0.00
13 I	Acenaphthene-d10	0.400	0.400	0.0	102	0.00
14 S	2,4,6-Tribromophenol	0.400	0.346	13.5	99	0.00
15 S	2-Fluorobiphenyl	0.400	0.485	-21.2	130	0.01
16	Acenaphthylene	0.400	0.437	-9.2	122	0.00
17	Acenaphthene	0.400	0.406	-1.5	112	0.00
18	Fluorene	0.400	0.404	-1.0	111	0.00
19 I	Phenanthrene-d10	0.400	0.400	0.0	102	0.00
20	4,6-Dinitro-2-methylphenol	0.400	0.418	-4.5	106	0.00
21	4-Bromophenyl-phenylether	0.400	0.392	2.0	107	0.00
22	Hexachlorobenzene	0.400	0.424	-6.0	112	0.00
23	Atrazine	0.400	0.396	1.0	117	0.00
24	Pentachlorophenol	0.400	0.316	21.0	95	0.00
25	Phenanthrene	0.400	0.411	-2.7	112	0.00
26	Anthracene	0.400	0.408	-2.0	116	0.00
27 SURR	Fluoranthene-d10	0.400	0.413	-3.2	120	0.00
28	Fluoranthene	0.400	0.391	2.3	110	0.00
29 I	Chrysene-d12	0.400	0.400	0.0	105	0.00
30	Pyrene	0.400	0.402	-0.5	109	0.00
31 S	Terphenyl-d14	0.400	0.482	-20.5	133	0.00
32	Benzo(a)anthracene	0.400	0.411	-2.7	116	0.00
33	Chrysene	0.400	0.414	-3.5	112	0.00
34	Bis(2-ethylhexyl)phthalate	0.400	0.404	-1.0	120	0.00
35 I	Perylene-d12	0.400	0.400	0.0	105	0.00
36	Indeno(1,2,3-cd)pyrene	0.400	0.418	-4.5	121	0.00
37	Benzo(b)fluoranthene	0.400	0.388	3.0	108	0.00
38	Benzo(k)fluoranthene	0.400	0.415	-3.7	119	0.00
39 C	Benzo(a)pyrene	0.400	0.422	-5.5	123	0.00
40	Dibenzo(a,h)anthracene	0.400	0.411	-2.7	119	0.01
41	Benzo(g,h,i)perylene	0.400	0.394	1.5	111	0.01

(#) = Out of Range

SPCC's out = 0 CCC's out = 0



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

7C

SEMIVOLATILE CONTINUING CALIBRATION CHECK

Lab Name: Alliance Contract: TETR06
 Lab Code: ACE SDG No.: Q3292
 Instrument ID: BNA_N Calibration Date/Time: 10/13/2025 13:47
 Lab File ID: BN038066.D Init. Calib. Date(s): 09/23/2025 09/23/2025
 EPA Sample No.: SSTDCCC0.4 Init. Calib. Time(s): 12:13 15:51
 GC Column: ZB-GR ID: 0.25 (mm)

COMPOUND	RRF	RRF0.4	MIN RRF	%D	MAX%D
2-Methylnaphthalene-d10	0.556	0.571		2.7	20.0
Fluoranthene-d10	1.006	1.065		5.9	20.0
2-Fluorophenol	0.913	0.893		-2.2	20.0
Phenol-d6	1.199	0.939		-21.7	20.0
Nitrobenzene-d5	0.305	0.238		-22.0	20.0
2-Fluorobiphenyl	1.638	1.428		-12.8	20.0
2,4,6-Tribromophenol	0.152	0.137		-9.9	20.0
Terphenyl-d14	0.797	0.724		-9.2	20.0
1,4-Dioxane	0.406	0.465		14.5	20.0

All other compounds must meet a minimum RRF of 0.010.

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN101325\
 Data File : BN038066.D
 Acq On : 13 Oct 2025 13:47
 Operator : RC/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :

BNA_N

ClientSampleId :

SSTDCCC0.4

Manual Integrations

APPROVED

Reviewed By :Jagrut Upadhyay 10/14/2025
 Supervised By :mohammad ahmed 10/15/2025

Quant Time: Oct 13 14:16:09 2025

Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN092325.M

Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION

QLast Update : Wed Sep 24 11:00:01 2025

Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.797	152	7166	0.400	ng	-0.03
7) Naphthalene-d8	10.594	136	18651	0.400	ng	#-0.03
13) Acenaphthene-d10	14.452	164	10244	0.400	ng	-0.02
19) Phenanthrene-d10	17.198	188	18901	0.400	ng	#-0.02
29) Chrysene-d12	21.393	240	16827	0.400	ng	0.00
35) Perylene-d12	23.712	264	17326	0.400	ng	-0.01
System Monitoring Compounds						
4) 2-Fluorophenol	5.370	112	6402	0.391	ng	-0.01
5) Phenol-d6	6.959	99	6728	0.313	ng	-0.01
8) Nitrobenzene-d5	8.950	82	4437	0.312	ng	-0.02
11) 2-Methylnaphthalene-d10	12.192	152	10642	0.411	ng	-0.03
14) 2,4,6-Tribromophenol	15.945	330	1404	0.361	ng	-0.02
15) 2-Fluorobiphenyl	13.073	172	14631	0.349	ng	-0.02
27) Fluoranthene-d10	19.239	212	20125	0.423	ng	-0.01
31) Terphenyl-d14	19.838	244	12185	0.364	ng	-0.01
Target Compounds						Qvalue
2) 1,4-Dioxane	3.283	88	3335	0.458	ng	94
3) n-Nitrosodimethylamine	3.586	42	3720	0.384	ng	# 92
6) bis(2-Chloroethyl)ether	7.219	93	7145	0.358	ng	97
9) Naphthalene	10.648	128	19331	0.376	ng	100
10) Hexachlorobutadiene	10.947	225	3527	0.402	ng	# 100
12) 2-Methylnaphthalene	12.268	142	12367	0.368	ng	99
16) Acenaphthylene	14.174	152	16983	0.365	ng	100
17) Acenaphthene	14.516	154	11898	0.359	ng	98
18) Fluorene	15.499	166	12971	0.324	ng	99
20) 4,6-Dinitro-2-methylph...	15.572	198	741	0.373	ng	# 10
21) 4-Bromophenyl-phenylether	16.404	248	4416	0.359	ng	# 82
22) Hexachlorobenzene	16.515	284	5456	0.365	ng	99
23) Atrazine	16.664	200	3115	0.332	ng	# 92
24) Pentachlorophenol	16.851	266	2441	0.476	ng	95
25) Phenanthrene	17.248	178	21937	0.353	ng	100
26) Anthracene	17.335	178	18915	0.351	ng	100
28) Fluoranthene	19.266	202	24986	0.365	ng	100
30) Pyrene	19.629	202	25549	0.385	ng	100
32) Benzo(a)anthracene	21.375	228	21568	0.367	ng	99
33) Chrysene	21.429	228	24503	0.385	ng	99
34) Bis(2-ethylhexyl)phtha...	21.313	149	15324	0.659	ng	97
36) Indeno(1,2,3-cd)pyrene	26.080	276	30553	0.386	ng	98
37) Benzo(b)fluoranthene	23.013	252	25246m	0.338	ng	
38) Benzo(k)fluoranthene	23.057	252	25627m	0.341	ng	
39) Benzo(a)pyrene	23.613	252	21668	0.367	ng	# 89
40) Dibenzo(a,h)anthracene	26.098	278	22721	0.368	ng	98
41) Benzo(g,h,i)perylene	26.800	276	27477	0.423	ng	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN101325\
 Data File : BN038066.D
 Acq On : 13 Oct 2025 13:47
 Operator : RC/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1


Instrument :
 BNA_N
 ClientSampleId :
 SSTDCCC0.4

Quant Time: Oct 13 14:16:09 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN092325.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Sep 24 11:00:01 2025
 Response via : Initial Calibration

Manual Integrations APPROVED

Reviewed By :Jagrut Upadhyay 10/14/2025
 Supervised By :mohammad ahmed 10/15/2025






#1
1,4-Dichlorobenzene-d4
Concen: 0.400 ng
RT: 7.797 min Scan# 61
Delta R.T. -0.028 min
Lab File: BN038066.D
Acq: 13 Oct 2025 13:47

Instrument :

BNA_N

ClientSampleId :

SSTDCCC0.4






Tgt Ion: 152 Resp: 7166
Ion Ratio Lower Upper
152 100
150 152.5 121.0 181.4
115 61.0 47.4 71.0

Manual Integrations




APPROVED


Reviewed By :Jagrut Upadhyay 10/14/2025
Supervised By :mohammad ahmed 10/15/2025



#2
1,4-Dioxane
Concen: 0.458 ng
RT: 3.283 min Scan# 28
Delta R.T. -0.014 min
Lab File: BN038066.D
Acq: 13 Oct 2025 13:47


Tgt Ion: 88 Resp: 3335
Ion Ratio Lower Upper
88 100
43 30.0 26.6 39.8
58 68.4 58.9 88.3





#3
n-Nitrosodimethylamine
Concen: 0.384 ng
RT: 3.586 min Scan# 70
Delta R.T. -0.014 min
Lab File: BN038066.D
Acq: 13 Oct 2025 13:47




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4



Tgt Ion: 42 Resp: 3720
Ion Ratio Lower Upper
42 100
74 119.3 93.4 140.0
44 11.7 22.2 33.4




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 10/14/2025
Supervised By :mohammad ahmed 10/15/2025



#4
2-Fluorophenol
Concen: 0.391 ng
RT: 5.370 min Scan# 317
Delta R.T. -0.014 min
Lab File: BN038066.D
Acq: 13 Oct 2025 13:47

Tgt Ion:112 Resp: 6402
Ion Ratio Lower Upper
112 100
64 55.7 44.6 66.8
63 29.4 24.9 37.3





#5
Phenol-d6
Concen: 0.313 ng
RT: 6.959 min Scan# 531
Delta R.T. -0.014 min
Lab File: BN038066.D
Acq: 13 Oct 2025 13:47

Instrument :

BNA_N

ClientSampleId :





SSTDCCC0.4

Tgt Ion	Ratio	Lower	Upper
99	100		
42	23.3	17.2	25.8
71	33.2	27.3	40.9

Manual Integrations




APPROVED


Reviewed By :Jagrut Upadhyay 10/14/2025
Supervised By :mohammad ahmed 10/15/2025



#6
bis(2-Chloroethyl)ether
Concen: 0.358 ng
RT: 7.219 min Scan# 573
Delta R.T. -0.022 min
Lab File: BN038066.D
Acq: 13 Oct 2025 13:47


Tgt Ion	Ratio	Lower	Upper
93	100		
63	78.0	60.1	90.1
95	33.0	25.4	38.2





#7
Naphthalene-d8
Concen: 0.400 ng
RT: 10.594 min Scan# 941
Delta R.T. -0.032 min
Lab File: BN038066.D
Acq: 13 Oct 2025 13:47




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4



Tgt Ion:136 Resp: 1865
Ion Ratio Lower Upper
136 100
137 11.4 9.0 13.4
54 10.1 5.8 8.8
68 7.9 4.3 6.5#




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 10/14/2025
Supervised By :mohammad ahmed 10/15/2025



#8
Nitrobenzene-d5
Concen: 0.312 ng
RT: 8.950 min Scan# 784
Delta R.T. -0.021 min
Lab File: BN038066.D
Acq: 13 Oct 2025 13:47


Tgt Ion: 82 Resp: 4437
Ion Ratio Lower Upper
82 100
128 45.7 34.8 52.2
54 54.6 42.2 63.2





#9
Naphthalene
Concen: 0.376 ng
RT: 10.648 min Scan# 945
Delta R.T. -0.021 min
Lab File: BN038066.D
Acq: 13 Oct 2025 13:47




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4



Tgt Ion:128 Resp: 1933
Ion Ratio Lower Upper
128 100
129 11.6 9.2 13.8
127 13.6 11.0 16.6




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 10/14/2025
Supervised By :mohammad ahmed 10/15/2025



#10
Hexachlorobutadiene
Concen: 0.402 ng
RT: 10.947 min Scan# 971
Delta R.T. -0.032 min
Lab File: BN038066.D
Acq: 13 Oct 2025 13:47


Tgt Ion:225 Resp: 3527
Ion Ratio Lower Upper
225 100
223 0.0 0.0 0.0
227 64.5 51.4 77.2





#11
2-Methylnaphthalene-d10
Concen: 0.411 ng
RT: 12.192 min Scan# 1163
Delta R.T. -0.025 min
Lab File: BN038066.D
Acq: 13 Oct 2025 13:47




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4



Tgt Ion:152 Resp: 10642
Ion Ratio Lower Upper
152 100
151 21.2 17.4 26.2




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 10/14/2025
Supervised By :mohammad ahmed 10/15/2025



#12
2-Methylnaphthalene
Concen: 0.368 ng
RT: 12.268 min Scan# 1173
Delta R.T. -0.025 min
Lab File: BN038066.D
Acq: 13 Oct 2025 13:47


Tgt Ion:142 Resp: 12367
Ion Ratio Lower Upper
142 100
141 88.5 71.2 106.8
115 35.9 29.7 44.5





#13
Acenaphthene-d10
Concen: 0.400 ng
RT: 14.452 min Scan# 1456
Delta R.T. -0.021 min
Lab File: BN038066.D
Acq: 13 Oct 2025 13:47




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4




Tgt Ion:164 Resp: 10244
Ion Ratio Lower Upper
164 100
162 105.5 84.8 127.2
160 53.4 46.1 69.1

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 10/14/2025
Supervised By :mohammad ahmed 10/15/2025




#14
2,4,6-Tribromophenol
Concen: 0.361 ng
RT: 15.945 min Scan# 1588
Delta R.T. -0.024 min
Lab File: BN038066.D
Acq: 13 Oct 2025 13:47



Tgt Ion:330 Resp: 1404
Ion Ratio Lower Upper
330 100
332 98.0 77.2 115.8
141 40.4 37.2 55.8





#15
2-Fluorobiphenyl
Concen: 0.349 ng
RT: 13.073 min Scan# 1327
Delta R.T. -0.021 min
Lab File: BN038066.D
Acq: 13 Oct 2025 13:47

Instrument :

BNA_N

ClientSampleId :





SSTDCCC0.4

Tgt Ion	Ratio	Lower	Upper
172	100		
171	35.7	28.6	43.0
170	23.8	19.7	29.5

Manual Integrations




APPROVED


Reviewed By :Jagrut Upadhyay 10/14/2025
Supervised By :mohammad ahmed 10/15/2025



#16
Acenaphthylene
Concen: 0.365 ng
RT: 14.174 min Scan# 1428
Delta R.T. -0.021 min
Lab File: BN038066.D
Acq: 13 Oct 2025 13:47


Tgt Ion	Ratio	Lower	Upper
152	100		
151	19.7	15.8	23.8
153	12.8	10.5	15.7





#17
Acenaphthene
Concen: 0.359 ng
RT: 14.516 min Scan# 1460
Delta R.T. -0.021 min
Lab File: BN038066.D
Acq: 13 Oct 2025 13:47




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4



Tgt Ion:154 Resp: 11898
Ion Ratio Lower Upper
154 100
153 112.0 90.5 135.7
152 61.7 51.8 77.8




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 10/14/2025
Supervised By :mohammad ahmed 10/15/2025



#18
Fluorene
Concen: 0.324 ng
RT: 15.499 min Scan# 1552
Delta R.T. -0.023 min
Lab File: BN038066.D
Acq: 13 Oct 2025 13:47

Tgt Ion:166 Resp: 12971
Ion Ratio Lower Upper
166 100
165 99.3 79.0 118.4
167 13.3 10.6 16.0





#19

Phenanthrene-d10

Concen: 0.400 ng

RT: 17.198 min Scan# 1689

Delta R.T. -0.025 min

Lab File: BN038066.D

Acq: 13 Oct 2025 13:47

Instrument :

BNA_N

ClientSampleId :

SSTDCCC0.4

Tgt Ion:188 Resp: 1890

Ion Ratio Lower Upper





188 100

94 0.0 0.0 0.0

80 13.9 7.6 11.4

Manual Integrations

APPROVED

Reviewed By :Jagrut Upadhyay 10/14/2025
Supervised By :mohammad ahmed 10/15/2025

#20

4,6-Dinitro-2-methylphenol

Concen: 0.373 ng

RT: 15.572 min Scan# 1558

Delta R.T. -0.025 min

Lab File: BN038066.D

Acq: 13 Oct 2025 13:47




Tgt Ion:198 Resp: 741


Ion Ratio Lower Upper

198 100

51 176.1 59.1 88.7#


105 83.2 41.3 61.9#





#21
4-Bromophenyl-phenylether
Concen: 0.359 ng
RT: 16.404 min Scan# 1625
Delta R.T. -0.012 min
Lab File: BN038066.D
Acq: 13 Oct 2025 13:47




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4



Tgt Ion:248 Resp: 4416
Ion Ratio Lower Upper
248 100
250 105.0 77.6 116.4
141 35.8 48.8 73.2




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 10/14/2025
Supervised By :mohammad ahmed 10/15/2025



#22
Hexachlorobenzene
Concen: 0.365 ng
RT: 16.515 min Scan# 1634
Delta R.T. -0.025 min
Lab File: BN038066.D
Acq: 13 Oct 2025 13:47

Tgt Ion:284 Resp: 5456
Ion Ratio Lower Upper
284 100
142 38.2 30.9 46.3
249 30.1 23.9 35.9









#23
Atrazine
Concen: 0.332 ng
RT: 16.664 min Scan# 1646
Delta R.T. -0.025 min
Lab File: BN038066.D
Acq: 13 Oct 2025 13:47

Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4

Tgt Ion:200 Resp: 3115
Ion Ratio Lower Upper
200 100
173 28.7 18.3 27.5
215 44.3 38.6 58.0




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 10/14/2025
Supervised By :mohammad ahmed 10/15/2025



#24
Pentachlorophenol
Concen: 0.476 ng
RT: 16.851 min Scan# 1661
Delta R.T. -0.024 min
Lab File: BN038066.D
Acq: 13 Oct 2025 13:47


Tgt Ion:266 Resp: 2441
Ion Ratio Lower Upper
266 100
264 60.9 50.9 76.3
268 62.2 54.3 81.5





#25
Phenanthrene
Concen: 0.353 ng
RT: 17.248 min Scan# 1693
Delta R.T. -0.012 min
Lab File: BN038066.D
Acq: 13 Oct 2025 13:47




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4



Tgt Ion:178 Resp: 21937
Ion Ratio Lower Upper
178 100
176 19.5 15.7 23.5
179 15.6 12.3 18.5




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 10/14/2025
Supervised By :mohammad ahmed 10/15/2025



#26
Anthracene
Concen: 0.351 ng
RT: 17.335 min Scan# 1700
Delta R.T. -0.025 min
Lab File: BN038066.D
Acq: 13 Oct 2025 13:47

Tgt Ion:178 Resp: 18915
Ion Ratio Lower Upper
178 100
176 18.6 15.0 22.4
179 15.2 12.3 18.5





#27

Fluoranthene-d10

Concen: 0.423 ng

RT: 19.239 min Scan# 1951

Delta R.T. -0.014 min

Lab File: BN038066.D

Acq: 13 Oct 2025 13:47

Instrument :

BNA_N

ClientSampleId :





SSTDCCC0.4

Tgt Ion:	212	Resp:	20125
Ion Ratio	Lower	Upper	
212	100		
106	15.7	12.6	19.0
104	9.1	7.4	11.0

Manual Integrations

APPROVED

Reviewed By :Jagrut Upadhyay 10/14/2025
Supervised By :mohammad ahmed 10/15/2025



#28

Fluoranthene

Concen: 0.365 ng




RT: 19.266 min Scan# 1957


Delta R.T. -0.019 min

Lab File: BN038066.D

Acq: 13 Oct 2025 13:47

Tgt Ion:	202	Resp:	24986
Ion Ratio	Lower	Upper	
202	100		
101	11.3	9.3	13.9
203	16.9	13.6	20.4






#29
Chrysene-d12
Concen: 0.400 ng
RT: 21.393 min Scan# 21
Delta R.T. -0.009 min
Lab File: BN038066.D
Acq: 13 Oct 2025 13:47

Instrument :

BNA_N

ClientSampleId :

SSTDCCC0.4






Tgt Ion:240 Resp: 1682
Ion Ratio Lower Upper
240 100
120 14.2 11.4 17.2
236 29.0 23.0 34.6


Manual Integrations

APPROVED



Reviewed By :Jagrut Upadhyay 10/14/2025
Supervised By :mohammad ahmed 10/15/2025




#30
Pyrene
Concen: 0.385 ng
RT: 19.629 min Scan# 2035
Delta R.T. -0.019 min
Lab File: BN038066.D
Acq: 13 Oct 2025 13:47



Tgt Ion:202 Resp: 25549
Ion Ratio Lower Upper
202 100
200 21.0 16.7 25.1
203 17.6 14.2 21.4






#31
Terphenyl-d14
Concen: 0.364 ng
RT: 19.838 min Scan# 2080
Delta R.T. -0.014 min
Lab File: BN038066.D
Acq: 13 Oct 2025 13:47

Instrument :

BNA_N

ClientSampleId :

SSTDCCC0.4






Tgt Ion:244 Resp: 12185
Ion Ratio Lower Upper
244 100
212 9.0 7.6 11.4
122 16.7 13.9 20.9


Manual Integrations

APPROVED



Reviewed By :Jagrut Upadhyay 10/14/2025
Supervised By :mohammad ahmed 10/15/2025




#32
Benzo(a)anthracene
Concen: 0.367 ng
RT: 21.375 min Scan# 2269
Delta R.T. -0.009 min
Lab File: BN038066.D
Acq: 13 Oct 2025 13:47




Tgt Ion:228 Resp: 21568
Ion Ratio Lower Upper
228 100
226 27.7 22.6 33.8
229 19.6 15.6 23.4





#33
Chrysene
Concen: 0.385 ng
RT: 21.429 min Scan# 2275
Delta R.T. -0.009 min
Lab File: BN038066.D
Acq: 13 Oct 2025 13:47




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4



Tgt Ion:228 Resp: 24503
Ion Ratio Lower Upper
228 100
226 30.4 24.6 37.0
229 19.8 15.6 23.4




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 10/14/2025
Supervised By :mohammad ahmed 10/15/2025



#34
Bis(2-ethylhexyl)phthalate
Concen: 0.659 ng
RT: 21.313 min Scan# 2262
Delta R.T. -0.009 min
Lab File: BN038066.D
Acq: 13 Oct 2025 13:47

Tgt Ion:149 Resp: 15324
Ion Ratio Lower Upper
149 100
167 27.0 20.4 30.6
279 3.3 2.4 3.6





#35

Perylene-d12

Concen: 0.400 ng

RT: 23.712 min Scan# 2831

Delta R.T. -0.015 min

Lab File: BN038066.D


Acq: 13 Oct 2025 13:47

Instrument :

BNA_N

ClientSampleId :

SSTDCCC0.4



Tgt Ion:264 Resp: 17326

Ion Ratio Lower Upper




264 100

260 26.6 21.5 32.3

265 79.8 53.8 80.6

Manual Integrations

APPROVED

Reviewed By :Jagrut Upadhyay 10/14/2025
Supervised By :mohammad ahmed 10/15/2025

#36

Indeno(1,2,3-cd)pyrene


Concen: 0.386 ng

RT: 26.080 min Scan# 3641

Delta R.T. -0.026 min

Lab File: BN038066.D

Acq: 13 Oct 2025 13:47





Tgt Ion:276 Resp: 30553


Ion Ratio Lower Upper

276 100

138 26.3 21.4 32.0


277 23.7 20.3 30.5





#37
Benzo(b)fluoranthene
Concen: 0.338 ng m
RT: 23.013 min Scan# 2592
Delta R.T. -0.012 min
Lab File: BN038066.D
Acq: 13 Oct 2025 13:47




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4



Tgt Ion:252 Resp: 25246
Ion Ratio Lower Upper
252 100
253 25.5 19.4 29.0
125 20.5 13.0 19.4




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 10/14/2025
Supervised By :mohammad ahmed 10/15/2025



#38
Benzo(k)fluoranthene
Concen: 0.341 ng m
RT: 23.057 min Scan# 2607
Delta R.T. -0.015 min
Lab File: BN038066.D
Acq: 13 Oct 2025 13:47

Tgt Ion:252 Resp: 25627
Ion Ratio Lower Upper
252 100
253 25.7 19.5 29.3
125 18.7 13.0 19.4





#39

Benzo(a)pyrene

Concen: 0.367 ng

RT: 23.613 min Scan# 21

Delta R.T. -0.012 min

Lab File: BN038066.D

Acq: 13 Oct 2025 13:47

Instrument :

BNA_N

ClientSampleId :





SSTDCCC0.4

Tgt Ion:	252	Resp:	21668
Ion Ratio	Lower	Upper	
252	100		
253	28.7	20.6	30.8
125	29.2	16.7	25.1

Manual Integrations

APPROVED

Reviewed By :Jagrut Upadhyay 10/14/2025
Supervised By :mohammad ahmed 10/15/2025



#40

Dibenzo(a,h)anthracene

Concen: 0.368 ng




RT: 26.098 min Scan# 3647


Delta R.T. -0.020 min

Lab File: BN038066.D

Acq: 13 Oct 2025 13:47

Tgt Ion:	278	Resp:	22721
Ion Ratio	Lower	Upper	
278	100		
139	20.6	15.4	23.2
279	26.4	20.8	31.2





#41

Benzo(g,h,i)perylene

Concen: 0.423 ng

RT: 26.800 min Scan# 38

Delta R.T. -0.029 min

Lab File: BN038066.D


Acq: 13 Oct 2025 13:47

Instrument :

BNA_N

ClientSampleId :

SSTDCCC0.4



Tgt Ion:276 Resp: 2747

Ion Ratio Lower Upper

276 100

277 24.8 20.2 30.4



138 25.5 19.8 29.8

Manual Integrations

APPROVED

Reviewed By :Jagrut Upadhyay 10/14/2025

Supervised By :mohammad ahmed 10/15/2025



Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN101325\
 Data File : BN038066.D
 Acq On : 13 Oct 2025 13:47
 Operator : RC/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: Oct 13 14:16:09 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN092325.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Sep 24 11:00:01 2025
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	128	-0.03
2	1,4-Dioxane	0.406	0.465	-14.5	146	-0.01
3	n-Nitrosodimethylamine	0.540	0.519	3.9	129	-0.01
4 S	2-Fluorophenol	0.913	0.893	2.2	134	-0.01
5 S	Phenol-d6	1.199	0.939	21.7	108	-0.01
6	bis(2-Chloroethyl)ether	1.113	0.997	10.4	118	-0.02
7 I	Naphthalene-d8	1.000	1.000	0.0	118	-0.03
8 S	Nitrobenzene-d5	0.305	0.238	22.0	97	-0.02
9	Naphthalene	1.102	1.036	6.0	117	-0.02
10	Hexachlorobutadiene	0.188	0.189	-0.5	123	-0.03
11 SURR	2-Methylnaphthalene-d10	0.556	0.571	-2.7	135	-0.03
12	2-Methylnaphthalene	0.720	0.663	7.9	117	-0.03
13 I	Acenaphthene-d10	1.000	1.000	0.0	124	-0.02
14 S	2,4,6-Tribromophenol	0.152	0.137	9.9	126	-0.02
15 S	2-Fluorobiphenyl	1.638	1.428	12.8	114	-0.02
16	Acenaphthylene	1.816	1.658	8.7	124	-0.02
17	Acenaphthene	1.294	1.161	10.3	120	-0.02
18	Fluorene	1.565	1.266	19.1	108	-0.02
19 I	Phenanthrene-d10	1.000	1.000	0.0	119	-0.02
20	4,6-Dinitro-2-methylphenol	0.057	0.039	31.6#	106	-0.02
21	4-Bromophenyl-phenylether	0.261	0.234	10.3	113	-0.01
22	Hexachlorobenzene	0.316	0.289	8.5	112	-0.02
23	Atrazine	0.199	0.165	17.1	114	-0.02
24	Pentachlorophenol	0.108	0.129	-19.4	167#	-0.02
25	Phenanthrene	1.316	1.161	11.8	112	-0.01
26	Anthracene	1.142	1.001	12.3	116	-0.02
27 SURR	Fluoranthene-d10	1.006	1.065	-5.9	142	-0.01
28	Fluoranthene	1.450	1.322	8.8	120	-0.02
29 I	Chrysene-d12	1.000	1.000	0.0	124	0.00
30	Pyrene	1.579	1.518	3.9	124	-0.02
31 S	Terphenyl-d14	0.797	0.724	9.2	119	-0.01
32	Benzo(a)anthracene	1.398	1.282	8.3	123	0.00
33	Chrysene	1.512	1.456	3.7	123	0.00
34	Bis(2-ethylhexyl)phthalate	0.553	0.911	-64.7#	232#	0.00
35 I	Perylene-d12	1.000	1.000	0.0	138	-0.01
36	Indeno(1,2,3-cd)pyrene	1.828	1.763	3.6	146	-0.03
37	Benzo(b)fluoranthene	1.725	1.457	15.5	122	-0.01
38	Benzo(k)fluoranthene	1.736	1.479	14.8	128	-0.01
39 C	Benzo(a)pyrene	1.364	1.251	8.3	140	-0.01
40	Dibenzo(a,h)anthracene	1.427	1.311	8.1	139	-0.02
41	Benzo(g,h,i)perylene	1.499	1.586	-5.8	156#	-0.03

(#) = Out of Range

SPCC's out = 0 CCC's out = 0

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN101325\
 Data File : BN038066.D
 Acq On : 13 Oct 2025 13:47
 Operator : RC/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: Oct 13 14:16:09 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN092325.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Sep 24 11:00:01 2025
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	128	-0.03
2	1,4-Dioxane	0.400	0.458	-14.5	146	-0.01
3	n-Nitrosodimethylamine	0.400	0.384	4.0	129	-0.01
4 S	2-Fluorophenol	0.400	0.391	2.3	134	-0.01
5 S	Phenol-d6	0.400	0.313	21.8	108	-0.01
6	bis(2-Chloroethyl)ether	0.400	0.358	10.5	118	-0.02
7 I	Naphthalene-d8	0.400	0.400	0.0	118	-0.03
8 S	Nitrobenzene-d5	0.400	0.312	22.0	97	-0.02
9	Naphthalene	0.400	0.376	6.0	117	-0.02
10	Hexachlorobutadiene	0.400	0.402	-0.5	123	-0.03
11 SURR	2-Methylnaphthalene-d10	0.400	0.411	-2.7	135	-0.03
12	2-Methylnaphthalene	0.400	0.368	8.0	117	-0.03
13 I	Acenaphthene-d10	0.400	0.400	0.0	124	-0.02
14 S	2,4,6-Tribromophenol	0.400	0.361	9.8	126	-0.02
15 S	2-Fluorobiphenyl	0.400	0.349	12.8	114	-0.02
16	Acenaphthylene	0.400	0.365	8.8	124	-0.02
17	Acenaphthene	0.400	0.359	10.3	120	-0.02
18	Fluorene	0.400	0.324	19.0	108	-0.02
19 I	Phenanthrene-d10	0.400	0.400	0.0	119	-0.02
20	4,6-Dinitro-2-methylphenol	0.400	0.373	6.8	106	-0.02
21	4-Bromophenyl-phenylether	0.400	0.359	10.3	113	-0.01
22	Hexachlorobenzene	0.400	0.365	8.8	112	-0.02
23	Atrazine	0.400	0.332	17.0	114	-0.02
24	Pentachlorophenol	0.400	0.476	-19.0	167	-0.02
25	Phenanthrene	0.400	0.353	11.8	112	-0.01
26	Anthracene	0.400	0.351	12.3	116	-0.02
27 SURR	Fluoranthene-d10	0.400	0.423	-5.7	142	-0.01
28	Fluoranthene	0.400	0.365	8.8	120	-0.02
29 I	Chrysene-d12	0.400	0.400	0.0	124	0.00
30	Pyrene	0.400	0.385	3.8	124	-0.02
31 S	Terphenyl-d14	0.400	0.364	9.0	119	-0.01
32	Benzo(a)anthracene	0.400	0.367	8.3	123	0.00
33	Chrysene	0.400	0.385	3.8	123	0.00
34	Bis(2-ethylhexyl)phthalate	0.400	0.659	-64.8#	232	0.00
35 I	Perylene-d12	0.400	0.400	0.0	138	-0.01
36	Indeno(1,2,3-cd)pyrene	0.400	0.386	3.5	146	-0.03
37	Benzo(b)fluoranthene	0.400	0.338	15.5	122	-0.01
38	Benzo(k)fluoranthene	0.400	0.341	14.8	128	-0.01
39 C	Benzo(a)pyrene	0.400	0.367	8.3	140	-0.01
40	Dibenzo(a,h)anthracene	0.400	0.368	8.0	139	-0.02
41	Benzo(g,h,i)perylene	0.400	0.423	-5.7	156	-0.03

(#) = Out of Range

SPCC's out = 0 CCC's out = 0



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

7C

SEMIVOLATILE CONTINUING CALIBRATION CHECK

Lab Name: Alliance Contract: TETR06
 Lab Code: ACE SDG No.: Q3292
 Instrument ID: BNA_N Calibration Date/Time: 10/13/2025 23:24
 Lab File ID: BN038081.D Init. Calib. Date(s): 09/23/2025 09/23/2025
 EPA Sample No.: SSTDCCC0.4EC Init. Calib. Time(s): 12:13 15:51
 GC Column: ZB-GR ID: 0.25 (mm)

COMPOUND	RRF	RRF0.4	MIN RRF	%D	MAX%D
2-Methylnaphthalene-d10	0.556	0.494		-11.2	50.0
Fluoranthene-d10	1.006	0.861		-14.4	50.0
2-Fluorophenol	0.913	0.864		-5.4	50.0
Phenol-d6	1.199	0.927		-22.7	50.0
Nitrobenzene-d5	0.305	0.270		-11.5	50.0
2-Fluorobiphenyl	1.638	1.596		-2.6	50.0
2,4,6-Tribromophenol	0.152	0.107		-29.6	50.0
Terphenyl-d14	0.797	0.757		-5.0	50.0
1,4-Dioxane	0.406	0.485		19.5	50.0

All other compounds must meet a minimum RRF of 0.010.

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN101325\
 Data File : BN038081.D
 Acq On : 13 Oct 2025 23:24
 Operator : RC/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SSTDCCC0.4EC

Quant Time: Oct 14 02:41:04 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN092325.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Sep 24 11:00:01 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.796	152	6102	0.400	ng	-0.03
7) Naphthalene-d8	10.594	136	16304	0.400	ng	#-0.03
13) Acenaphthene-d10	14.452	164	8122	0.400	ng	-0.02
19) Phenanthrene-d10	17.198	188	14166	0.400	ng	#-0.02
29) Chrysene-d12	21.393	240	10220	0.400	ng	0.00
35) Perylene-d12	23.709	264	9438	0.400	ng	#-0.02
System Monitoring Compounds						
4) 2-Fluorophenol	5.363	112	5272	0.379	ng	-0.02
5) Phenol-d6	6.951	99	5654	0.309	ng	-0.02
8) Nitrobenzene-d5	8.950	82	4404	0.354	ng	-0.02
11) 2-Methylnaphthalene-d10	12.197	152	8062	0.356	ng	-0.02
14) 2,4,6-Tribromophenol	15.944	330	869	0.282	ng	-0.02
15) 2-Fluorobiphenyl	13.072	172	12964	0.390	ng	-0.02
27) Fluoranthene-d10	19.238	212	12203	0.342	ng	-0.01
31) Terphenyl-d14	19.838	244	7736	0.380	ng	-0.01
Target Compounds					Qvalue	
2) 1,4-Dioxane	3.283	88	2958	0.478	ng	99
3) n-Nitrosodimethylamine	3.586	42	3747	0.455	ng	# 87
6) bis(2-Chloroethyl)ether	7.219	93	6694	0.394	ng	99
9) Naphthalene	10.648	128	17982	0.400	ng	99
10) Hexachlorobutadiene	10.946	225	3142	0.410	ng	# 99
12) 2-Methylnaphthalene	12.268	142	10665	0.363	ng	100
16) Acenaphthylene	14.174	152	14258	0.387	ng	100
17) Acenaphthene	14.516	154	9994	0.380	ng	99
18) Fluorene	15.499	166	10564	0.332	ng	96
20) 4,6-Dinitro-2-methylph...	15.584	198	515	0.352	ng	# 1
21) 4-Bromophenyl-phenylether	16.391	248	3312	0.359	ng	# 81
22) Hexachlorobenzene	16.515	284	4363	0.390	ng	97
23) Atrazine	16.664	200	2133	0.303	ng	# 93
24) Pentachlorophenol	16.863	266	1349	0.351	ng	98
25) Phenanthrene	17.248	178	17038	0.365	ng	100
26) Anthracene	17.335	178	14291	0.353	ng	100
28) Fluoranthene	19.266	202	17495	0.341	ng	99
30) Pyrene	19.629	202	17496	0.434	ng	100
32) Benzo(a)anthracene	21.375	228	12446	0.348	ng	98
33) Chrysene	21.420	228	16053	0.416	ng	99
34) Bis(2-ethylhexyl)phtha...	21.304	149	5282	0.374	ng	99
36) Indeno(1,2,3-cd)pyrene	26.074	276	16710	0.387	ng	98
37) Benzo(b)fluoranthene	23.008	252	14457	0.355	ng	# 90
38) Benzo(k)fluoranthene	23.051	252	15321	0.374	ng	# 90
39) Benzo(a)pyrene	23.607	252	12298	0.382	ng	# 83
40) Dibenzo(a,h)anthracene	26.095	278	12118	0.360	ng	# 90
41) Benzo(g,h,i)perylene	26.791	276	15568	0.440	ng	98


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN101325\
Data File : BN038081.D
Acq On : 13 Oct 2025 23:24
Operator : RC/JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 17 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC

Quant Time: Oct 14 02:41:04 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN092325.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Sep 24 11:00:01 2025
Response via : Initial Calibration









#1
1,4-Dichlorobenzene-d4
Concen: 0.400 ng
RT: 7.796 min Scan# 61
Delta R.T. -0.029 min
Lab File: BN038081.D
Acq: 13 Oct 2025 23:24




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC


Tgt Ion:152 Resp: 6102
Ion Ratio Lower Upper
152 100
150 156.0 121.0 181.4
115 61.2 47.4 71.0



#2
1,4-Dioxane
Concen: 0.478 ng
RT: 3.283 min Scan# 28
Delta R.T. -0.015 min
Lab File: BN038081.D
Acq: 13 Oct 2025 23:24

Tgt Ion: 88 Resp: 2958
Ion Ratio Lower Upper
88 100
43 32.3 26.6 39.8
58 73.2 58.9 88.3









#3
n-Nitrosodimethylamine
Concen: 0.455 ng
RT: 3.586 min Scan# 70
Delta R.T. -0.015 min
Lab File: BN038081.D
Acq: 13 Oct 2025 23:24




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC


Tgt Ion: 42 Resp: 3747
Ion Ratio Lower Upper
42 100
74 106.9 93.4 140.0
44 11.2 22.2 33.4#



#4
2-Fluorophenol
Concen: 0.379 ng
RT: 5.363 min Scan# 316
Delta R.T. -0.022 min
Lab File: BN038081.D
Acq: 13 Oct 2025 23:24

Tgt Ion: 112 Resp: 5272
Ion Ratio Lower Upper
112 100
64 56.4 44.6 66.8
63 31.5 24.9 37.3









#5
Phenol-d6
Concen: 0.309 ng
RT: 6.951 min Scan# 531
Delta R.T. -0.022 min
Lab File: BN038081.D
Acq: 13 Oct 2025 23:24




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC


Tgt Ion	Ratio	Lower	Upper
99	100		
42	24.1	17.2	25.8
71	33.5	27.3	40.9



#6
bis(2-Chloroethyl)ether
Concen: 0.394 ng
RT: 7.219 min Scan# 573
Delta R.T. -0.022 min
Lab File: BN038081.D
Acq: 13 Oct 2025 23:24

Tgt Ion	Ratio	Lower	Upper
93	100		
63	76.5	60.1	90.1
95	31.2	25.4	38.2









#7
Naphthalene-d8
Concen: 0.400 ng
RT: 10.594 min Scan# 941
Delta R.T. -0.032 min
Lab File: BN038081.D
Acq: 13 Oct 2025 23:24




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC


Tgt Ion:136 Resp: 16304
Ion Ratio Lower Upper
136 100
137 11.3 9.0 13.4
54 9.7 5.8 8.8#
68 7.4 4.3 6.5#



#8
Nitrobenzene-d5
Concen: 0.354 ng
RT: 8.950 min Scan# 784
Delta R.T. -0.021 min
Lab File: BN038081.D
Acq: 13 Oct 2025 23:24

Tgt Ion: 82 Resp: 4404
Ion Ratio Lower Upper
82 100
128 43.9 34.8 52.2
54 52.5 42.2 63.2









#9
Naphthalene
Concen: 0.400 ng
RT: 10.648 min Scan# 94
Delta R.T. -0.021 min
Lab File: BN038081.D
Acq: 13 Oct 2025 23:24




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC


Tgt Ion:	128	Resp:	17982
Ion	Ratio	Lower	Upper
128	100		
129	11.3	9.2	13.8
127	13.4	11.0	16.6



#10
Hexachlorobutadiene
Concen: 0.410 ng
RT: 10.946 min Scan# 971
Delta R.T. -0.032 min
Lab File: BN038081.D
Acq: 13 Oct 2025 23:24

Tgt Ion:	225	Resp:	3142
Ion	Ratio	Lower	Upper
225	100		
223	0.0	0.0	0.0
227	63.7	51.4	77.2









#11
2-Methylnaphthalene-d10
Concen: 0.356 ng
RT: 12.197 min Scan# 1163
Delta R.T. -0.020 min
Lab File: BN038081.D
Acq: 13 Oct 2025 23:24




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC


Tgt Ion:152 Resp: 8062
Ion Ratio Lower Upper
152 100
151 21.3 17.4 26.2



#12
2-Methylnaphthalene
Concen: 0.363 ng
RT: 12.268 min Scan# 1173
Delta R.T. -0.025 min
Lab File: BN038081.D
Acq: 13 Oct 2025 23:24

Tgt Ion:142 Resp: 10665
Ion Ratio Lower Upper
142 100
141 89.4 71.2 106.8
115 37.1 29.7 44.5









#13
Acenaphthene-d10
Concen: 0.400 ng
RT: 14.452 min Scan# 1456
Delta R.T. -0.022 min
Lab File: BN038081.D
Acq: 13 Oct 2025 23:24




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC


Tgt Ion:164 Resp: 8122
Ion Ratio Lower Upper
164 100
162 104.2 84.8 127.2
160 56.4 46.1 69.1



#14
2,4,6-Tribromophenol
Concen: 0.282 ng
RT: 15.944 min Scan# 1588
Delta R.T. -0.025 min
Lab File: BN038081.D
Acq: 13 Oct 2025 23:24

Tgt Ion:330 Resp: 869
Ion Ratio Lower Upper
330 100
332 97.7 77.2 115.8
141 50.3 37.2 55.8









#15
2-Fluorobiphenyl
Concen: 0.390 ng
RT: 13.072 min Scan# 1327
Delta R.T. -0.022 min
Lab File: BN038081.D
Acq: 13 Oct 2025 23:24




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC


Tgt Ion	Ratio	Lower	Upper
172	100		
171	35.9	28.6	43.0
170	24.6	19.7	29.5



#16
Acenaphthylene
Concen: 0.387 ng
RT: 14.174 min Scan# 1428
Delta R.T. -0.022 min
Lab File: BN038081.D
Acq: 13 Oct 2025 23:24

Tgt Ion	Ratio	Lower	Upper
152	100		
151	19.6	15.8	23.8
153	13.0	10.5	15.7









#17
Acenaphthene
Concen: 0.380 ng
RT: 14.516 min Scan# 1460
Delta R.T. -0.022 min
Lab File: BN038081.D
Acq: 13 Oct 2025 23:24




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC


Tgt Ion:154 Resp: 9994
Ion Ratio Lower Upper
154 100
153 114.8 90.5 135.7
152 64.2 51.8 77.8



#18
Fluorene
Concen: 0.332 ng
RT: 15.499 min Scan# 1552
Delta R.T. -0.023 min
Lab File: BN038081.D
Acq: 13 Oct 2025 23:24

Tgt Ion:166 Resp: 10564
Ion Ratio Lower Upper
166 100
165 103.3 79.0 118.4
167 13.6 10.6 16.0









#19
Phenanthrene-d10
Concen: 0.400 ng
RT: 17.198 min Scan# 1689
Delta R.T. -0.025 min
Lab File: BN038081.D
Acq: 13 Oct 2025 23:24




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC


Tgt Ion	Ratio	Lower	Upper
188	100		
94	0.0	0.0	0.0
80	14.1	7.6	11.4#



#20
4,6-Dinitro-2-methylphenol
Concen: 0.352 ng
RT: 15.584 min Scan# 1559
Delta R.T. -0.013 min
Lab File: BN038081.D
Acq: 13 Oct 2025 23:24

Tgt Ion	Ratio	Lower	Upper
198	100		
51	219.4	59.1	88.7#
105	83.4	41.3	61.9#









#21
4-Bromophenyl-phenylether
Concen: 0.359 ng
RT: 16.391 min Scan# 1624
Delta R.T. -0.025 min
Lab File: BN038081.D
Acq: 13 Oct 2025 23:24




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC


Tgt Ion:248 Resp: 3312
Ion Ratio Lower Upper
248 100
250 93.3 77.6 116.4
141 94.0 48.8 73.2#



#22
Hexachlorobenzene
Concen: 0.390 ng
RT: 16.515 min Scan# 1634
Delta R.T. -0.025 min
Lab File: BN038081.D
Acq: 13 Oct 2025 23:24

Tgt Ion:284 Resp: 4363
Ion Ratio Lower Upper
284 100
142 41.4 30.9 46.3
249 30.8 23.9 35.9









#23
Atrazine
Concen: 0.303 ng
RT: 16.664 min Scan# 1646
Delta R.T. -0.025 min
Lab File: BN038081.D
Acq: 13 Oct 2025 23:24




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC


Tgt Ion	Ratio	Lower	Upper
200	100		
173	29.7	18.3	27.5#
215	46.2	38.6	58.0



#24
Pentachlorophenol
Concen: 0.351 ng
RT: 16.863 min Scan# 1662
Delta R.T. -0.012 min
Lab File: BN038081.D
Acq: 13 Oct 2025 23:24

Tgt Ion	Ratio	Lower	Upper
266	100		
264	60.8	50.9	76.3
268	68.8	54.3	81.5









#25
Phenanthrene
Concen: 0.365 ng
RT: 17.248 min Scan# 1694
Delta R.T. -0.013 min
Lab File: BN038081.D
Acq: 13 Oct 2025 23:24




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC


Tgt Ion:178 Resp: 17038
Ion Ratio Lower Upper
178 100
176 19.8 15.7 23.5
179 15.2 12.3 18.5



#26
Anthracene
Concen: 0.353 ng
RT: 17.335 min Scan# 1700
Delta R.T. -0.025 min
Lab File: BN038081.D
Acq: 13 Oct 2025 23:24

Tgt Ion:178 Resp: 14291
Ion Ratio Lower Upper
178 100
176 18.6 15.0 22.4
179 15.2 12.3 18.5





#27

Fluoranthene-d10

Concen: 0.342 ng

RT: 19.238 min Scan# 1951

Delta R.T. -0.014 min

Lab File: BN038081.D

Acq: 13 Oct 2025 23:24

Instrument :

BNA_N

ClientSampleId :

SSTDCCC0.4EC





Tgt Ion:212 Resp: 12203

Ion Ratio Lower Upper

212 100

106 16.5 12.6 19.0

104 9.4 7.4 11.0



#28

Fluoranthene

Concen: 0.341 ng

RT: 19.266 min Scan# 1957

Delta R.T. -0.019 min

Lab File: BN038081.D

Acq: 13 Oct 2025 23:24




Tgt Ion:202 Resp: 17495


Ion Ratio Lower Upper

202 100

101 12.4 9.3 13.9

203 17.0 13.6 20.4





#29

Chrysene-d12

Concen: 0.400 ng

RT: 21.393 min Scan# 21

Delta R.T. -0.009 min

Lab File: BN038081.D

Acq: 13 Oct 2025 23:24

Instrument :

BNA_N

ClientSampleId :

SSTDCCC0.4EC





Tgt Ion:240 Resp: 10220

Ion Ratio Lower Upper

240 100

120 15.7 11.4 17.2

236 29.1 23.0 34.6



#30

Pyrene

Concen: 0.434 ng

RT: 19.629 min Scan# 2035

Delta R.T. -0.019 min

Lab File: BN038081.D

Acq: 13 Oct 2025 23:24




Tgt Ion:202 Resp: 17496


Ion Ratio Lower Upper

202 100

200 20.8 16.7 25.1

203 17.8 14.2 21.4









#31
Terphenyl-d14
Concen: 0.380 ng
RT: 19.838 min Scan# 2080
Delta R.T. -0.014 min
Lab File: BN038081.D
Acq: 13 Oct 2025 23:24




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC


Tgt Ion:244 Resp: 7736
Ion Ratio Lower Upper
244 100
212 9.9 7.6 11.4
122 18.0 13.9 20.9



#32
Benzo(a)anthracene
Concen: 0.348 ng
RT: 21.375 min Scan# 2269
Delta R.T. -0.009 min
Lab File: BN038081.D
Acq: 13 Oct 2025 23:24

Tgt Ion:228 Resp: 12446
Ion Ratio Lower Upper
228 100
226 27.3 22.6 33.8
229 20.1 15.6 23.4









#33
Chrysene
Concen: 0.416 ng
RT: 21.420 min Scan# 21
Delta R.T. -0.018 min
Lab File: BN038081.D
Acq: 13 Oct 2025 23:24




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC


Tgt Ion	Ratio	Lower	Upper
228	100		
226	31.5	24.6	37.0
229	19.8	15.6	23.4



#34
Bis(2-ethylhexyl)phthalate
Concen: 0.374 ng
RT: 21.304 min Scan# 2261
Delta R.T. -0.018 min
Lab File: BN038081.D
Acq: 13 Oct 2025 23:24

Tgt Ion	Ratio	Lower	Upper
149	100		
167	24.9	20.4	30.6
279	3.5	2.4	3.6









#35
Perylene-d12
Concen: 0.400 ng
RT: 23.709 min Scan# 2830
Delta R.T. -0.018 min
Lab File: BN038081.D
Acq: 13 Oct 2025 23:24




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC


Tgt Ion:264 Resp: 9438
Ion Ratio Lower Upper
264 100
260 27.1 21.5 32.3
265 119.2 53.8 80.6#



#36
Indeno(1,2,3-cd)pyrene
Concen: 0.387 ng
RT: 26.074 min Scan# 3639
Delta R.T. -0.032 min
Lab File: BN038081.D
Acq: 13 Oct 2025 23:24

Tgt Ion:276 Resp: 16710
Ion Ratio Lower Upper
276 100
138 27.4 21.4 32.0
277 23.6 20.3 30.5









#37
Benzo(b)fluoranthene
Concen: 0.355 ng
RT: 23.008 min Scan# 21
Delta R.T. -0.018 min
Lab File: BN038081.D
Acq: 13 Oct 2025 23:24




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC


Tgt Ion:252 Resp: 14457
Ion Ratio Lower Upper
252 100
253 27.5 19.4 29.0
125 22.8 13.0 19.4#



#38
Benzo(k)fluoranthene
Concen: 0.374 ng
RT: 23.051 min Scan# 2605
Delta R.T. -0.021 min
Lab File: BN038081.D
Acq: 13 Oct 2025 23:24

Tgt Ion:252 Resp: 15321
Ion Ratio Lower Upper
252 100
253 27.7 19.5 29.3
125 22.4 13.0 19.4#





#39

Benzo(a)pyrene

Concen: 0.382 ng

RT: 23.607 min Scan# 21

Delta R.T. -0.018 min

Lab File: BN038081.D

Acq: 13 Oct 2025 23:24

Instrument :

BNA_N

ClientSampleId :

SSTDCCC0.4EC





Tgt Ion:252 Resp: 12298

Ion Ratio Lower Upper

252 100

253 31.2 20.6 30.8#

125 32.8 16.7 25.1#



#40

Dibenzo(a,h)anthracene

Concen: 0.360 ng

RT: 26.095 min Scan# 3646

Delta R.T. -0.024 min

Lab File: BN038081.D

Acq: 13 Oct 2025 23:24




Tgt Ion:278 Resp: 12118


Ion Ratio Lower Upper

278 100

139 23.7 15.4 23.2#

279 31.3 20.8 31.2#





#41

Benzo(g,h,i)perylene

Concen: 0.440 ng

RT: 26.791 min Scan# 38

Delta R.T. -0.038 min

Lab File: BN038081.D


Acq: 13 Oct 2025 23:24

Instrument :

BNA_N

ClientSampleId :

SSTDCCC0.4EC





Tgt Ion:276 Resp: 15568

Ion Ratio Lower Upper

276 100

277 25.7 20.2 30.4

138 26.7 19.8 29.8



Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN101325\
 Data File : BN038081.D
 Acq On : 13 Oct 2025 23:24
 Operator : RC/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: Oct 14 02:41:04 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN092325.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Sep 24 11:00:01 2025
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	109	-0.03
2	1,4-Dioxane	0.406	0.485	-19.5	129	-0.01
3	n-Nitrosodimethylamine	0.540	0.614	-13.7	130	-0.01
4 S	2-Fluorophenol	0.913	0.864	5.4	110	-0.02
5 S	Phenol-d6	1.199	0.927	22.7	91	-0.02
6	bis(2-Chloroethyl)ether	1.113	1.097	1.4	111	-0.02
7 I	Naphthalene-d8	1.000	1.000	0.0	103	-0.03
8 S	Nitrobenzene-d5	0.305	0.270	11.5	96	-0.02
9	Naphthalene	1.102	1.103	-0.1	109	-0.02
10	Hexachlorobutadiene	0.188	0.193	-2.7	109	-0.03
11 SURR	2-Methylnaphthalene-d10	0.556	0.494	11.2	102	-0.02
12	2-Methylnaphthalene	0.720	0.654	9.2	101	-0.03
13 I	Acenaphthene-d10	1.000	1.000	0.0	98	-0.02
14 S	2,4,6-Tribromophenol	0.152	0.107	29.6#	78	-0.02
15 S	2-Fluorobiphenyl	1.638	1.596	2.6	101	-0.02
16	Acenaphthylene	1.816	1.755	3.4	104	-0.02
17	Acenaphthene	1.294	1.230	4.9	101	-0.02
18	Fluorene	1.565	1.301	16.9	88	-0.02
19 I	Phenanthrene-d10	1.000	1.000	0.0	89	-0.02
20	4,6-Dinitro-2-methylphenol	0.057	0.036	36.8#	74	-0.01
21	4-Bromophenyl-phenylether	0.261	0.234	10.3	85	-0.02
22	Hexachlorobenzene	0.316	0.308	2.5	90	-0.02
23	Atrazine	0.199	0.151	24.1	78	-0.02
24	Pentachlorophenol	0.108	0.095	12.0	92	-0.01
25	Phenanthrene	1.316	1.203	8.6	87	-0.01
26	Anthracene	1.142	1.009	11.6	88	-0.02
27 SURR	Fluoranthene-d10	1.006	0.861	14.4	86	-0.01
28	Fluoranthene	1.450	1.235	14.8	84	-0.02
29 I	Chrysene-d12	1.000	1.000	0.0	75	0.00
30	Pyrene	1.579	1.712	-8.4	85	-0.02
31 S	Terphenyl-d14	0.797	0.757	5.0	76	-0.01
32	Benzo(a)anthracene	1.398	1.218	12.9	71	0.00
33	Chrysene	1.512	1.571	-3.9	81	-0.02
34	Bis(2-ethylhexyl)phthalate	0.553	0.517	6.5	80	-0.02
35 I	Perylene-d12	1.000	1.000	0.0	75	-0.02
36	Indeno(1,2,3-cd)pyrene	1.828	1.771	3.1	80	-0.03
37	Benzo(b)fluoranthene	1.725	1.532	11.2	70	-0.02
38	Benzo(k)fluoranthene	1.736	1.623	6.5	76	-0.02
39 C	Benzo(a)pyrene	1.364	1.303	4.5	79	-0.02
40	Dibenzo(a,h)anthracene	1.427	1.284	10.0	74	-0.02
41	Benzo(g,h,i)perylene	1.499	1.650	-10.1	88	-0.04

(#) = Out of Range

SPCC's out = 0 CCC's out = 0

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN101325\
 Data File : BN038081.D
 Acq On : 13 Oct 2025 23:24
 Operator : RC/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: Oct 14 02:41:04 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN092325.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Sep 24 11:00:01 2025
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	109	-0.03
2	1,4-Dioxane	0.400	0.478	-19.5	129	-0.01
3	n-Nitrosodimethylamine	0.400	0.455	-13.7	130	-0.01
4 S	2-Fluorophenol	0.400	0.379	5.3	110	-0.02
5 S	Phenol-d6	0.400	0.309	22.8	91	-0.02
6	bis(2-Chloroethyl)ether	0.400	0.394	1.5	111	-0.02
7 I	Naphthalene-d8	0.400	0.400	0.0	103	-0.03
8 S	Nitrobenzene-d5	0.400	0.354	11.5	96	-0.02
9	Naphthalene	0.400	0.400	0.0	109	-0.02
10	Hexachlorobutadiene	0.400	0.410	-2.5	109	-0.03
11 SURR	2-Methylnaphthalene-d10	0.400	0.356	11.0	102	-0.02
12	2-Methylnaphthalene	0.400	0.363	9.3	101	-0.03
13 I	Acenaphthene-d10	0.400	0.400	0.0	98	-0.02
14 S	2,4,6-Tribromophenol	0.400	0.282	29.5#	78	-0.02
15 S	2-Fluorobiphenyl	0.400	0.390	2.5	101	-0.02
16	Acenaphthylene	0.400	0.387	3.3	104	-0.02
17	Acenaphthene	0.400	0.380	5.0	101	-0.02
18	Fluorene	0.400	0.332	17.0	88	-0.02
19 I	Phenanthrene-d10	0.400	0.400	0.0	89	-0.02
20	4,6-Dinitro-2-methylphenol	0.400	0.352	12.0	74	-0.01
21	4-Bromophenyl-phenylether	0.400	0.359	10.3	85	-0.02
22	Hexachlorobenzene	0.400	0.390	2.5	90	-0.02
23	Atrazine	0.400	0.303	24.3	78	-0.02
24	Pentachlorophenol	0.400	0.351	12.3	92	-0.01
25	Phenanthrene	0.400	0.365	8.8	87	-0.01
26	Anthracene	0.400	0.353	11.8	88	-0.02
27 SURR	Fluoranthene-d10	0.400	0.342	14.5	86	-0.01
28	Fluoranthene	0.400	0.341	14.8	84	-0.02
29 I	Chrysene-d12	0.400	0.400	0.0	75	0.00
30	Pyrene	0.400	0.434	-8.5	85	-0.02
31 S	Terphenyl-d14	0.400	0.380	5.0	76	-0.01
32	Benzo(a)anthracene	0.400	0.348	13.0	71	0.00
33	Chrysene	0.400	0.416	-4.0	81	-0.02
34	Bis(2-ethylhexyl)phthalate	0.400	0.374	6.5	80	-0.02
35 I	Perylene-d12	0.400	0.400	0.0	75	-0.02
36	Indeno(1,2,3-cd)pyrene	0.400	0.387	3.3	80	-0.03
37	Benzo(b)fluoranthene	0.400	0.355	11.3	70	-0.02
38	Benzo(k)fluoranthene	0.400	0.374	6.5	76	-0.02
39 C	Benzo(a)pyrene	0.400	0.382	4.5	79	-0.02
40	Dibenzo(a,h)anthracene	0.400	0.360	10.0	74	-0.02
41	Benzo(g,h,i)perylene	0.400	0.440	-10.0	88	-0.04

(#) = Out of Range

SPCC's out = 0 CCC's out = 0



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

7C

SEMIVOLATILE CONTINUING CALIBRATION CHECK

Lab Name: Alliance Contract: TETR06
 Lab Code: ACE SDG No.: Q3292
 Instrument ID: BNA_N Calibration Date/Time: 10/14/2025 01:19
 Lab File ID: BN038083.D Init. Calib. Date(s): 09/23/2025 09/23/2025
 EPA Sample No.: SSTDCCC0.4 Init. Calib. Time(s): 12:13 15:51
 GC Column: ZB-GR ID: 0.25 (mm)

COMPOUND	RRF	RRF0.4	MIN RRF	%D	MAX%D
2-Methylnaphthalene-d10	0.556	0.546		-1.8	20.0
Fluoranthene-d10	1.006	1.016		1.0	20.0
2-Fluorophenol	0.913	0.823		-9.9	20.0
Phenol-d6	1.199	0.940		-21.6	20.0
Nitrobenzene-d5	0.305	0.248		-18.7	20.0
2-Fluorobiphenyl	1.638	1.500		-8.4	20.0
2,4,6-Tribromophenol	0.152	0.116		-23.7	20.0
Terphenyl-d14	0.797	0.709		-11.0	20.0
1,4-Dioxane	0.406	0.473		16.5	20.0

All other compounds must meet a minimum RRF of 0.010.

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN101325\
 Data File : BN038083.D
 Acq On : 14 Oct 2025 01:19
 Operator : RC/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 18 Sample Multiplier: 1

Instrument :

BNA_N

ClientSampleId :

SSTDCCC0.4

Manual Integrations**APPROVED**

Reviewed By :Jagrut Upadhyay 10/14/2025
 Supervised By :mohammad ahmed 10/15/2025

Quant Time: Oct 14 03:52:02 2025

Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN092325.M

Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION

QLast Update : Wed Sep 24 11:00:01 2025

Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.797	152	7061	0.400	ng	-0.03
7) Naphthalene-d8	10.594	136	18151	0.400	ng	#-0.03
13) Acenaphthene-d10	14.452	164	9016	0.400	ng	-0.02
19) Phenanthrene-d10	17.198	188	16225	0.400	ng	#-0.02
29) Chrysene-d12	21.384	240	13579	0.400	ng	#-0.02
35) Perylene-d12	23.709	264	13727	0.400	ng	#-0.02
System Monitoring Compounds						
4) 2-Fluorophenol	5.363	112	5814	0.361	ng	-0.02
5) Phenol-d6	6.952	99	6635	0.313	ng	-0.02
8) Nitrobenzene-d5	8.950	82	4510	0.326	ng	-0.02
11) 2-Methylnaphthalene-d10	12.192	152	9902	0.393	ng	-0.03
14) 2,4,6-Tribromophenol	15.944	330	1046	0.306	ng	-0.02
15) 2-Fluorobiphenyl	13.073	172	13525	0.366	ng	-0.02
27) Fluoranthene-d10	19.234	212	16484	0.404	ng	-0.02
31) Terphenyl-d14	19.838	244	9629	0.356	ng	-0.01
Target Compounds						Qvalue
2) 1,4-Dioxane	3.283	88	3339	0.466	ng	94
3) n-Nitrosodimethylamine	3.586	42	3735	0.392	ng	# 91
6) bis(2-Chloroethyl)ether	7.212	93	7427	0.378	ng	99
9) Naphthalene	10.648	128	19092	0.382	ng	99
10) Hexachlorobutadiene	10.947	225	3405	0.399	ng	# 99
12) 2-Methylnaphthalene	12.268	142	11497	0.352	ng	99
16) Acenaphthylene	14.174	152	15325	0.374	ng	100
17) Acenaphthene	14.516	154	10578	0.363	ng	99
18) Fluorene	15.499	166	11554	0.327	ng	98
20) 4,6-Dinitro-2-methylph...	15.585	198	580	0.347	ng	# 10
21) 4-Bromophenyl-phenylether	16.391	248	3637	0.344	ng	# 85
22) Hexachlorobenzene	16.515	284	4887	0.381	ng	99
23) Atrazine	16.664	200	2436	0.303	ng	# 93
24) Pentachlorophenol	16.863	266	1846	0.420	ng	95
25) Phenanthrene	17.248	178	19133	0.358	ng	99
26) Anthracene	17.335	178	15982	0.345	ng	100
28) Fluoranthene	19.266	202	20674	0.352	ng	99
30) Pyrene	19.629	202	20709	0.386	ng	100
32) Benzo(a)anthracene	21.375	228	17445	0.368	ng	99
33) Chrysene	21.420	228	20081	0.391	ng	100
34) Bis(2-ethylhexyl)phtha...	21.304	149	8392	0.447	ng	98
36) Indeno(1,2,3-cd)pyrene	26.072	276	26580	0.424	ng	97
37) Benzo(b)fluoranthene	23.005	252	21481m	0.363	ng	
38) Benzo(k)fluoranthene	23.049	252	22684	0.381	ng	96
39) Benzo(a)pyrene	23.607	252	18395	0.393	ng	# 83
40) Dibenzo(a,h)anthracene	26.092	278	19499	0.398	ng	98
41) Benzo(g,h,i)perylene	26.788	276	22628	0.440	ng	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN101325\
Data File : BN038083.D
Acq On : 14 Oct 2025 01:19
Operator : RC/JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 18 Sample Multiplier: 1


Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4

Quant Time: Oct 14 03:52:02 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN092325.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Sep 24 11:00:01 2025
Response via : Initial Calibration

Manual Integrations
APPROVED

Reviewed By :Jagrut Upadhyay 10/14/2025
Supervised By :mohammad ahmed 10/15/2025






#1
1,4-Dichlorobenzene-d4
Concen: 0.400 ng
RT: 7.797 min Scan# 61
Delta R.T. -0.028 min
Lab File: BN038083.D
Acq: 14 Oct 2025 01:19

Instrument :

BNA_N

ClientSampleId :

SSTDCCC0.4






Tgt Ion: 152 Resp: 7061
Ion Ratio Lower Upper
152 100
150 155.0 121.0 181.4
115 60.3 47.4 71.0

Manual Integrations




APPROVED


Reviewed By :Jagrut Upadhyay 10/14/2025
Supervised By :mohammad ahmed 10/15/2025



#2
1,4-Dioxane
Concen: 0.466 ng
RT: 3.283 min Scan# 28
Delta R.T. -0.015 min
Lab File: BN038083.D
Acq: 14 Oct 2025 01:19


Tgt Ion: 88 Resp: 3339
Ion Ratio Lower Upper
88 100
43 28.6 26.6 39.8
58 68.9 58.9 88.3





#3
n-Nitrosodimethylamine
Concen: 0.392 ng
RT: 3.586 min Scan# 70
Delta R.T. -0.015 min
Lab File: BN038083.D
Acq: 14 Oct 2025 01:19




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4



Tgt Ion: 42 Resp: 3735
Ion Ratio Lower Upper
42 100
74 112.6 93.4 140.0
44 10.4 22.2 33.4




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 10/14/2025
Supervised By :mohammad ahmed 10/15/2025



#4
2-Fluorophenol
Concen: 0.361 ng
RT: 5.363 min Scan# 316
Delta R.T. -0.022 min
Lab File: BN038083.D
Acq: 14 Oct 2025 01:19


Tgt Ion:112 Resp: 5814
Ion Ratio Lower Upper
112 100
64 55.5 44.6 66.8
63 30.8 24.9 37.3





#5
Phenol-d6
Concen: 0.313 ng
RT: 6.952 min Scan# 539
Delta R.T. -0.022 min
Lab File: BN038083.D
Acq: 14 Oct 2025 01:19




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4



Tgt Ion: 99 Resp: 6635
Ion Ratio Lower Upper
99 100
42 22.9 17.2 25.8
71 33.4 27.3 40.9




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 10/14/2025
Supervised By :mohammad ahmed 10/15/2025



#6
bis(2-Chloroethyl)ether
Concen: 0.378 ng
RT: 7.212 min Scan# 572
Delta R.T. -0.029 min
Lab File: BN038083.D
Acq: 14 Oct 2025 01:19


Tgt Ion: 93 Resp: 7427
Ion Ratio Lower Upper
93 100
63 76.0 60.1 90.1
95 31.2 25.4 38.2





#7
Naphthalene-d8
Concen: 0.400 ng
RT: 10.594 min Scan# 941
Delta R.T. -0.032 min
Lab File: BN038083.D
Acq: 14 Oct 2025 01:19




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4



Tgt Ion: 136 Resp: 1815
Ion Ratio Lower Upper
136 100
137 11.5 9.0 13.4
54 9.3 5.8 8.8
68 7.0 4.3 6.5




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 10/14/2025
Supervised By :mohammad ahmed 10/15/2025



#8
Nitrobenzene-d5
Concen: 0.326 ng
RT: 8.950 min Scan# 784
Delta R.T. -0.021 min
Lab File: BN038083.D
Acq: 14 Oct 2025 01:19


Tgt Ion: 82 Resp: 4510
Ion Ratio Lower Upper
82 100
128 46.8 34.8 52.2
54 53.6 42.2 63.2





#9
Naphthalene
Concen: 0.382 ng
RT: 10.648 min Scan# 943
Delta R.T. -0.021 min
Lab File: BN038083.D
Acq: 14 Oct 2025 01:19




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4



Tgt Ion:128 Resp: 19092
Ion Ratio Lower Upper
128 100
129 11.4 9.2 13.8
127 13.4 11.0 16.6




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 10/14/2025
Supervised By :mohammad ahmed 10/15/2025



#10
Hexachlorobutadiene
Concen: 0.399 ng
RT: 10.947 min Scan# 971
Delta R.T. -0.032 min
Lab File: BN038083.D
Acq: 14 Oct 2025 01:19


Tgt Ion:225 Resp: 3405
Ion Ratio Lower Upper
225 100
223 0.0 0.0 0.0
227 63.3 51.4 77.2





#11
2-Methylnaphthalene-d10
Concen: 0.393 ng
RT: 12.192 min Scan# 1158
Delta R.T. -0.025 min
Lab File: BN038083.D
Acq: 14 Oct 2025 01:19




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4



Tgt Ion:152 Resp: 9902
Ion Ratio Lower Upper
152 100
151 21.8 17.4 26.2




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 10/14/2025
Supervised By :mohammad ahmed 10/15/2025



#12
2-Methylnaphthalene
Concen: 0.352 ng
RT: 12.268 min Scan# 1173
Delta R.T. -0.025 min
Lab File: BN038083.D
Acq: 14 Oct 2025 01:19


Tgt Ion:142 Resp: 11497
Ion Ratio Lower Upper
142 100
141 88.6 71.2 106.8
115 36.0 29.7 44.5





#13
Acenaphthene-d10
Concen: 0.400 ng
RT: 14.452 min Scan# 1456
Delta R.T. -0.021 min
Lab File: BN038083.D
Acq: 14 Oct 2025 01:19




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4



Tgt Ion:164 Resp: 9016
Ion Ratio Lower Upper
164 100
162 105.1 84.8 127.2
160 55.0 46.1 69.1




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 10/14/2025
Supervised By :mohammad ahmed 10/15/2025



#14
2,4,6-Tribromophenol
Concen: 0.306 ng
RT: 15.944 min Scan# 1588
Delta R.T. -0.025 min
Lab File: BN038083.D
Acq: 14 Oct 2025 01:19

Tgt Ion:330 Resp: 1046
Ion Ratio Lower Upper
330 100
332 98.3 77.2 115.8
141 46.3 37.2 55.8





#15
2-Fluorobiphenyl
Concen: 0.366 ng
RT: 13.073 min Scan# 1327
Delta R.T. -0.021 min
Lab File: BN038083.D
Acq: 14 Oct 2025 01:19

Instrument :

BNA_N

ClientSampleId :





SSTDCCC0.4

Tgt Ion	Ratio	Lower	Upper
172	100		
171	36.0	28.6	43.0
170	24.1	19.7	29.5

Manual Integrations




APPROVED


Reviewed By :Jagrut Upadhyay 10/14/2025
Supervised By :mohammad ahmed 10/15/2025



#16
Acenaphthylene
Concen: 0.374 ng
RT: 14.174 min Scan# 1428
Delta R.T. -0.021 min
Lab File: BN038083.D
Acq: 14 Oct 2025 01:19


Tgt Ion	Ratio	Lower	Upper
152	100		
151	19.6	15.8	23.8
153	13.1	10.5	15.7





#17
Acenaphthene
Concen: 0.363 ng
RT: 14.516 min Scan# 1460
Delta R.T. -0.021 min
Lab File: BN038083.D
Acq: 14 Oct 2025 01:19




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4



Tgt Ion:154 Resp: 10578
Ion Ratio Lower Upper
154 100
153 114.3 90.5 135.7
152 65.5 51.8 77.8




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 10/14/2025
Supervised By :mohammad ahmed 10/15/2025



#18
Fluorene
Concen: 0.327 ng
RT: 15.499 min Scan# 1552
Delta R.T. -0.023 min
Lab File: BN038083.D
Acq: 14 Oct 2025 01:19

Tgt Ion:166 Resp: 11554
Ion Ratio Lower Upper
166 100
165 101.2 79.0 118.4
167 12.9 10.6 16.0





#19

Phenanthrene-d10

Concen: 0.400 ng

RT: 17.198 min Scan# 1689

Delta R.T. -0.025 min

Lab File: BN038083.D

Acq: 14 Oct 2025 01:19

Instrument :

BNA_N

ClientSampleId :





SSTDCCC0.4

Tgt Ion:	188	Resp:	16225
Ion Ratio	Lower	Upper	
188	100		
94	0.0	0.0	0.0
80	13.4	7.6	11.4

Manual Integrations

APPROVED

Reviewed By :Jagrut Upadhyay 10/14/2025
Supervised By :mohammad ahmed 10/15/2025



#20

4,6-Dinitro-2-methylphenol

Concen: 0.347 ng




RT: 15.585 min Scan# 1559


Delta R.T. -0.012 min

Lab File: BN038083.D

Acq: 14 Oct 2025 01:19


Tgt Ion:	198	Resp:	580
Ion Ratio	Lower	Upper	
198	100		
51	187.8	59.1	88.7#
105	70.5	41.3	61.9#





#21
4-Bromophenyl-phenylether
Concen: 0.344 ng
RT: 16.391 min Scan# 1624
Delta R.T. -0.025 min
Lab File: BN038083.D
Acq: 14 Oct 2025 01:19




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4



Tgt Ion:248 Resp: 363
Ion Ratio Lower Upper
248 100
250 92.5 77.6 116.4
141 85.5 48.8 73.2




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 10/14/2025
Supervised By :mohammad ahmed 10/15/2025



#22
Hexachlorobenzene
Concen: 0.381 ng
RT: 16.515 min Scan# 1634
Delta R.T. -0.025 min
Lab File: BN038083.D
Acq: 14 Oct 2025 01:19


Tgt Ion:284 Resp: 4887
Ion Ratio Lower Upper
284 100
142 39.1 30.9 46.3
249 30.2 23.9 35.9





#23
Atrazine
Concen: 0.303 ng
RT: 16.664 min Scan# 1646
Delta R.T. -0.025 min
Lab File: BN038083.D
Acq: 14 Oct 2025 01:19




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4



Tgt Ion:200 Resp: 2436
Ion Ratio Lower Upper
200 100
173 29.6 18.3 27.5
215 45.4 38.6 58.0




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 10/14/2025
Supervised By :mohammad ahmed 10/15/2025



#24
Pentachlorophenol
Concen: 0.420 ng
RT: 16.863 min Scan# 1662
Delta R.T. -0.012 min
Lab File: BN038083.D
Acq: 14 Oct 2025 01:19


Tgt Ion:266 Resp: 1846
Ion Ratio Lower Upper
266 100
264 59.0 50.9 76.3
268 64.0 54.3 81.5





#25
Phenanthrene
Concen: 0.358 ng
RT: 17.248 min Scan# 1693
Delta R.T. -0.012 min
Lab File: BN038083.D
Acq: 14 Oct 2025 01:19




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4



Tgt Ion:178 Resp: 19133
Ion Ratio Lower Upper
178 100
176 19.3 15.7 23.5
179 15.1 12.3 18.5




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 10/14/2025
Supervised By :mohammad ahmed 10/15/2025



#26
Anthracene
Concen: 0.345 ng
RT: 17.335 min Scan# 1700
Delta R.T. -0.025 min
Lab File: BN038083.D
Acq: 14 Oct 2025 01:19


Tgt Ion:178 Resp: 15982
Ion Ratio Lower Upper
178 100
176 18.6 15.0 22.4
179 15.1 12.3 18.5





#27
Fluoranthene-d10
Concen: 0.404 ng
RT: 19.234 min Scan# 1950
Delta R.T. -0.019 min
Lab File: BN038083.D
Acq: 14 Oct 2025 01:19




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4



Tgt Ion:212 Resp: 16484
Ion Ratio Lower Upper
212 100
106 16.2 12.6 19.0
104 9.4 7.4 11.0




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 10/14/2025
Supervised By :mohammad ahmed 10/15/2025



#28
Fluoranthene
Concen: 0.352 ng
RT: 19.266 min Scan# 1957
Delta R.T. -0.019 min
Lab File: BN038083.D
Acq: 14 Oct 2025 01:19

Tgt Ion:202 Resp: 20674
Ion Ratio Lower Upper
202 100
101 12.1 9.3 13.9
203 17.2 13.6 20.4






#29
Chrysene-d12
Concen: 0.400 ng
RT: 21.384 min Scan# 2135
Delta R.T. -0.018 min
Lab File: BN038083.D
Acq: 14 Oct 2025 01:19

Instrument :

BNA_N

ClientSampleId :

SSTDCCC0.4






Tgt Ion:240 Resp: 13579
Ion Ratio Lower Upper
240 100
120 20.2 11.4 17.2
236 29.4 23.0 34.6


Manual Integrations

APPROVED



Reviewed By :Jagrut Upadhyay 10/14/2025
Supervised By :mohammad ahmed 10/15/2025




#30
Pyrene
Concen: 0.386 ng
RT: 19.629 min Scan# 2035
Delta R.T. -0.019 min
Lab File: BN038083.D
Acq: 14 Oct 2025 01:19



Tgt Ion:202 Resp: 20709
Ion Ratio Lower Upper
202 100
200 21.0 16.7 25.1
203 18.1 14.2 21.4






#31
Terphenyl-d14
Concen: 0.356 ng
RT: 19.838 min Scan# 2080
Delta R.T. -0.014 min
Lab File: BN038083.D
Acq: 14 Oct 2025 01:19

Instrument :

BNA_N

ClientSampleId :

SSTDCCC0.4






Tgt Ion:244 Resp: 9629
Ion Ratio Lower Upper
244 100
212 9.5 7.6 11.4
122 17.2 13.9 20.9

Manual Integrations




APPROVED


Reviewed By :Jagrut Upadhyay 10/14/2025
Supervised By :mohammad ahmed 10/15/2025



#32
Benzo(a)anthracene
Concen: 0.368 ng
RT: 21.375 min Scan# 2269
Delta R.T. -0.009 min
Lab File: BN038083.D
Acq: 14 Oct 2025 01:19


Tgt Ion:228 Resp: 17445
Ion Ratio Lower Upper
228 100
226 27.3 22.6 33.8
229 19.6 15.6 23.4





#33
Chrysene
Concen: 0.391 ng
RT: 21.420 min Scan# 2276
Delta R.T. -0.018 min
Lab File: BN038083.D
Acq: 14 Oct 2025 01:19




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4



Tgt Ion: 228 Resp: 2008
Ion Ratio Lower Upper
228 100
226 30.6 24.6 37.0
229 19.3 15.6 23.4




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 10/14/2025
Supervised By :mohammad ahmed 10/15/2025



#34
Bis(2-ethylhexyl)phthalate
Concen: 0.447 ng
RT: 21.304 min Scan# 2261
Delta R.T. -0.018 min
Lab File: BN038083.D
Acq: 14 Oct 2025 01:19


Tgt Ion: 149 Resp: 8392
Ion Ratio Lower Upper
149 100
167 26.4 20.4 30.6
279 3.2 2.4 3.6





#35
Perylene-d12
Concen: 0.400 ng
RT: 23.709 min Scan# 2830
Delta R.T. -0.018 min
Lab File: BN038083.D
Acq: 14 Oct 2025 01:19




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4



Tgt Ion:264 Resp: 1372
Ion Ratio Lower Upper
264 100
260 26.5 21.5 32.3
265 86.3 53.8 80.6




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 10/14/2025
Supervised By :mohammad ahmed 10/15/2025



#36
Indeno(1,2,3-cd)pyrene
Concen: 0.424 ng
RT: 26.072 min Scan# 3638
Delta R.T. -0.035 min
Lab File: BN038083.D
Acq: 14 Oct 2025 01:19


Tgt Ion:276 Resp: 26580
Ion Ratio Lower Upper
276 100
138 28.5 21.4 32.0
277 23.8 20.3 30.5





#37
Benzo(b)fluoranthene
Concen: 0.363 ng m
RT: 23.005 min Scan# 2148
Delta R.T. -0.021 min
Lab File: BN038083.D
Acq: 14 Oct 2025 01:19




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4



Tgt Ion:252 Resp: 2148
Ion Ratio Lower Upper
252 100
253 25.6 19.4 29.0
125 22.5 13.0 19.4




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 10/14/2025
Supervised By :mohammad ahmed 10/15/2025



#38
Benzo(k)fluoranthene
Concen: 0.381 ng
RT: 23.049 min Scan# 2604
Delta R.T. -0.023 min
Lab File: BN038083.D
Acq: 14 Oct 2025 01:19

Tgt Ion:252 Resp: 22684
Ion Ratio Lower Upper
252 100
253 25.7 19.5 29.3
125 19.3 13.0 19.4





#39

Benzo(a)pyrene

Concen: 0.393 ng

RT: 23.607 min Scan# 21

Delta R.T. -0.018 min

Lab File: BN038083.D

Acq: 14 Oct 2025 01:19

Instrument :

BNA_N

ClientSampleId :





SSTDCCC0.4

Tgt Ion:	252	Resp:	18395
Ion Ratio	Lower	Upper	
252	100		
253	28.4	20.6	30.8
125	35.4	16.7	25.1

Manual Integrations

APPROVED

Reviewed By :Jagrut Upadhyay 10/14/2025
Supervised By :mohammad ahmed 10/15/2025



#40

Dibenzo(a,h)anthracene

Concen: 0.398 ng




RT: 26.092 min Scan# 3645


Delta R.T. -0.026 min

Lab File: BN038083.D

Acq: 14 Oct 2025 01:19

Tgt Ion:	278	Resp:	19499
Ion Ratio	Lower	Upper	
278	100		
139	20.3	15.4	23.2
279	26.9	20.8	31.2





#41

Benzo(g,h,i)perylene

Concen: 0.440 ng

RT: 26.788 min Scan# 38

Delta R.T. -0.041 min

Lab File: BN038083.D


Acq: 14 Oct 2025 01:19

Instrument :

BNA_N

ClientSampleId :

SSTDCCC0.4



Tgt Ion:276 Resp: 22628

Ion Ratio Lower Upper

276 100

277 24.3 20.2 30.4



138 25.7 19.8 29.8

Manual Integrations

APPROVED

Reviewed By :Jagrut Upadhyay 10/14/2025

Supervised By :mohammad ahmed 10/15/2025



Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN101325\
 Data File : BN038083.D
 Acq On : 14 Oct 2025 01:19
 Operator : RC/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: Oct 14 03:52:02 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN092325.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Sep 24 11:00:01 2025
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	126	-0.03
2	1,4-Dioxane	0.406	0.473	-16.5	146	-0.01
3	n-Nitrosodimethylamine	0.540	0.529	2.0	129	-0.01
4 S	2-Fluorophenol	0.913	0.823	9.9	122	-0.02
5 S	Phenol-d6	1.199	0.940	21.6	106	-0.02
6	bis(2-Chloroethyl)ether	1.113	1.052	5.5	123	-0.03
7 I	Naphthalene-d8	1.000	1.000	0.0	115	-0.03
8 S	Nitrobenzene-d5	0.305	0.248	18.7	99	-0.02
9	Naphthalene	1.102	1.052	4.5	115	-0.02
10	Hexachlorobutadiene	0.188	0.188	0.0	118	-0.03
11 SURR	2-Methylnaphthalene-d10	0.556	0.546	1.8	126	-0.03
12	2-Methylnaphthalene	0.720	0.633	12.1	109	-0.03
13 I	Acenaphthene-d10	1.000	1.000	0.0	109	-0.02
14 S	2,4,6-Tribromophenol	0.152	0.116	23.7	94	-0.02
15 S	2-Fluorobiphenyl	1.638	1.500	8.4	105	-0.02
16	Acenaphthylene	1.816	1.700	6.4	112	-0.02
17	Acenaphthene	1.294	1.173	9.4	107	-0.02
18	Fluorene	1.565	1.281	18.1	96	-0.02
19 I	Phenanthrene-d10	1.000	1.000	0.0	102	-0.02
20	4,6-Dinitro-2-methylphenol	0.057	0.036	36.8#	83	-0.01
21	4-Bromophenyl-phenylether	0.261	0.224	14.2	93	-0.02
22	Hexachlorobenzene	0.316	0.301	4.7	101	-0.02
23	Atrazine	0.199	0.150	24.6	89	-0.02
24	Pentachlorophenol	0.108	0.114	-5.6	126	-0.01
25	Phenanthrene	1.316	1.179	10.4	98	-0.01
26	Anthracene	1.142	0.985	13.7	98	-0.02
27 SURR	Fluoranthene-d10	1.006	1.016	-1.0	117	-0.02
28	Fluoranthene	1.450	1.274	12.1	99	-0.02
29 I	Chrysene-d12	1.000	1.000	0.0	100	-0.02
30	Pyrene	1.579	1.525	3.4	100	-0.02
31 S	Terphenyl-d14	0.797	0.709	11.0	94	-0.01
32	Benzo(a)anthracene	1.398	1.285	8.1	99	0.00
33	Chrysene	1.512	1.479	2.2	101	-0.02
34	Bis(2-ethylhexyl)phthalate	0.553	0.618	-11.8	127	-0.02
35 I	Perylene-d12	1.000	1.000	0.0	109	-0.02
36	Indeno(1,2,3-cd)pyrene	1.828	1.936	-5.9	127	-0.04
37	Benzo(b)fluoranthene	1.725	1.565	9.3	104	-0.02
38	Benzo(k)fluoranthene	1.736	1.653	4.8	113	-0.02
39 C	Benzo(a)pyrene	1.364	1.340	1.8	119	-0.02
40	Dibenzo(a,h)anthracene	1.427	1.420	0.5	119	-0.03
41	Benzo(g,h,i)perylene	1.499	1.648	-9.9	128	-0.04

(#) = Out of Range

SPCC's out = 0 CCC's out = 0

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN101325\
 Data File : BN038083.D
 Acq On : 14 Oct 2025 01:19
 Operator : RC/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: Oct 14 03:52:02 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN092325.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Sep 24 11:00:01 2025
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	126	-0.03
2	1,4-Dioxane	0.400	0.466	-16.5	146	-0.01
3	n-Nitrosodimethylamine	0.400	0.392	2.0	129	-0.01
4 S	2-Fluorophenol	0.400	0.361	9.8	122	-0.02
5 S	Phenol-d6	0.400	0.313	21.8	106	-0.02
6	bis(2-Chloroethyl)ether	0.400	0.378	5.5	123	-0.03
7 I	Naphthalene-d8	0.400	0.400	0.0	115	-0.03
8 S	Nitrobenzene-d5	0.400	0.326	18.5	99	-0.02
9	Naphthalene	0.400	0.382	4.5	115	-0.02
10	Hexachlorobutadiene	0.400	0.399	0.3	118	-0.03
11 SURR	2-Methylnaphthalene-d10	0.400	0.393	1.8	126	-0.03
12	2-Methylnaphthalene	0.400	0.352	12.0	109	-0.03
13 I	Acenaphthene-d10	0.400	0.400	0.0	109	-0.02
14 S	2,4,6-Tribromophenol	0.400	0.306	23.5	94	-0.02
15 S	2-Fluorobiphenyl	0.400	0.366	8.5	105	-0.02
16	Acenaphthylene	0.400	0.374	6.5	112	-0.02
17	Acenaphthene	0.400	0.363	9.3	107	-0.02
18	Fluorene	0.400	0.327	18.3	96	-0.02
19 I	Phenanthrene-d10	0.400	0.400	0.0	102	-0.02
20	4,6-Dinitro-2-methylphenol	0.400	0.347	13.3	83	-0.01
21	4-Bromophenyl-phenylether	0.400	0.344	14.0	93	-0.02
22	Hexachlorobenzene	0.400	0.381	4.8	101	-0.02
23	Atrazine	0.400	0.303	24.3	89	-0.02
24	Pentachlorophenol	0.400	0.420	-5.0	126	-0.01
25	Phenanthrene	0.400	0.358	10.5	98	-0.01
26	Anthracene	0.400	0.345	13.8	98	-0.02
27 SURR	Fluoranthene-d10	0.400	0.404	-1.0	117	-0.02
28	Fluoranthene	0.400	0.352	12.0	99	-0.02
29 I	Chrysene-d12	0.400	0.400	0.0	100	-0.02
30	Pyrene	0.400	0.386	3.5	100	-0.02
31 S	Terphenyl-d14	0.400	0.356	11.0	94	-0.01
32	Benzo(a)anthracene	0.400	0.368	8.0	99	0.00
33	Chrysene	0.400	0.391	2.3	101	-0.02
34	Bis(2-ethylhexyl)phthalate	0.400	0.447	-11.7	127	-0.02
35 I	Perylene-d12	0.400	0.400	0.0	109	-0.02
36	Indeno(1,2,3-cd)pyrene	0.400	0.424	-6.0	127	-0.04
37	Benzo(b)fluoranthene	0.400	0.363	9.3	104	-0.02
38	Benzo(k)fluoranthene	0.400	0.381	4.8	113	-0.02
39 C	Benzo(a)pyrene	0.400	0.393	1.8	119	-0.02
40	Dibenzo(a,h)anthracene	0.400	0.398	0.5	119	-0.03
41	Benzo(g,h,i)perylene	0.400	0.440	-10.0	128	-0.04

(#) = Out of Range

SPCC's out = 0 CCC's out = 0



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

7C

SEMIVOLATILE CONTINUING CALIBRATION CHECK

Lab Name:	<u>Alliance</u>	Contract:	<u>TETR06</u>
Lab Code:	<u>ACE</u>	SDG No.:	<u>Q3292</u>
Instrument ID:	<u>BNA_N</u>	Calibration Date/Time:	<u>10/14/2025 04:56</u>
Lab File ID:	<u>BN038088.D</u>	Init. Calib. Date(s):	<u>09/23/2025 09/23/2025</u>
EPA Sample No.:	<u>SSTDCCC0.4EC</u>	Init. Calib. Time(s):	<u>12:13 15:51</u>
GC Column:	<u>ZB-GR</u>	ID:	<u>0.25</u> (mm)

COMPOUND	RRF	RRF0.4	MIN RRF	%D	MAX%D
2-Methylnaphthalene-d10	0.556	0.537		-3.4	50.0
Fluoranthene-d10	1.006	0.979		-2.7	50.0
2-Fluorophenol	0.913	0.880		-3.6	50.0
Phenol-d6	1.199	0.975		-18.7	50.0
Nitrobenzene-d5	0.305	0.255		-16.4	50.0
2-Fluorobiphenyl	1.638	1.577		-3.7	50.0
2,4,6-Tribromophenol	0.152	0.119		-21.7	50.0
Terphenyl-d14	0.797	0.748		-6.1	50.0
1,4-Dioxane	0.406	0.494		21.7	50.0

All other compounds must meet a minimum RRF of 0.010.

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN101325\
 Data File : BN038088.D
 Acq On : 14 Oct 2025 04:56
 Operator : RC/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SSTDCCC0.4EC

Manual Integrations
 APPROVED

Reviewed By :Jagrut Upadhyay 10/14/2025
 Supervised By :mohammad ahmed 10/15/2025

Quant Time: Oct 14 05:20:56 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN092325.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Sep 24 11:00:01 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.797	152	6905	0.400	ng	-0.03
7) Naphthalene-d8	10.594	136	18816	0.400	ng	#-0.03
13) Acenaphthene-d10	14.452	164	9558	0.400	ng	-0.02
19) Phenanthrene-d10	17.198	188	17543	0.400	ng	#-0.02
29) Chrysene-d12	21.384	240	14165	0.400	ng	#-0.02
35) Perylene-d12	23.706	264	13844	0.400	ng	#-0.02
System Monitoring Compounds						
4) 2-Fluorophenol	5.363	112	6075	0.386	ng	-0.02
5) Phenol-d6	6.952	99	6734	0.325	ng	-0.02
8) Nitrobenzene-d5	8.950	82	4804	0.335	ng	-0.02
11) 2-Methylnaphthalene-d10	12.192	152	10106	0.387	ng	-0.03
14) 2,4,6-Tribromophenol	15.945	330	1142	0.315	ng	-0.02
15) 2-Fluorobiphenyl	13.073	172	15070	0.385	ng	-0.02
27) Fluoranthene-d10	19.239	212	17173	0.389	ng	-0.01
31) Terphenyl-d14	19.833	244	10595	0.376	ng	-0.02
Target Compounds						Qvalue
2) 1,4-Dioxane	3.283	88	3409	0.486	ng	94
3) n-Nitrosodimethylamine	3.586	42	3947	0.423	ng	# 88
6) bis(2-Chloroethyl)ether	7.219	93	7442	0.387	ng	99
9) Naphthalene	10.648	128	20288	0.391	ng	99
10) Hexachlorobutadiene	10.947	225	3539	0.400	ng	# 97
12) 2-Methylnaphthalene	12.268	142	12369	0.365	ng	99
16) Acenaphthylene	14.174	152	16454	0.379	ng	100
17) Acenaphthene	14.516	154	11586	0.375	ng	99
18) Fluorene	15.499	166	13064	0.349	ng	100
20) 4,6-Dinitro-2-methylph...	15.585	198	690	0.374	ng	# 26
21) 4-Bromophenyl-phenylether	16.391	248	3981	0.348	ng	# 83
22) Hexachlorobenzene	16.515	284	5224	0.377	ng	100
23) Atrazine	16.664	200	2825	0.324	ng	# 90
24) Pentachlorophenol	16.863	266	1613	0.339	ng	97
25) Phenanthrene	17.248	178	20782	0.360	ng	100
26) Anthracene	17.335	178	17015	0.340	ng	100
28) Fluoranthene	19.266	202	22691	0.357	ng	99
30) Pyrene	19.629	202	23294	0.417	ng	100
32) Benzo(a)anthracene	21.375	228	18122	0.366	ng	98
33) Chrysene	21.420	228	21548	0.403	ng	99
34) Bis(2-ethylhexyl)phtha...	21.304	149	7685	0.392	ng	# 99
36) Indeno(1,2,3-cd)pyrene	26.072	276	25477	0.403	ng	98
37) Benzo(b)fluoranthene	23.005	252	22201m	0.372	ng	
38) Benzo(k)fluoranthene	23.049	252	22592	0.376	ng	95
39) Benzo(a)pyrene	23.607	252	18632	0.395	ng	# 87
40) Dibenzo(a,h)anthracene	26.089	278	18798	0.381	ng	97
41) Benzo(g,h,i)perylene	26.788	276	22396	0.432	ng	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN101325\
 Data File : BN038088.D
 Acq On : 14 Oct 2025 04:56
 Operator : RC/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 23 Sample Multiplier: 1


Instrument :
 BNA_N
 ClientSampleId :
 SSTDCCC0.4EC

Manual Integrations APPROVED

Reviewed By :Jagrut Upadhyay 10/14/2025
 Supervised By :mohammad ahmed 10/15/2025


Quant Time: Oct 14 05:20:56 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN092325.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Sep 24 11:00:01 2025
 Response via : Initial Calibration





#1
1,4-Dichlorobenzene-d4
Concen: 0.400 ng
RT: 7.797 min Scan# 61
Delta R.T. -0.028 min
Lab File: BN038088.D
Acq: 14 Oct 2025 04:56




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC



Tgt Ion:152 Resp: 6905
Ion Ratio Lower Upper
152 100
150 154.8 121.0 181.4
115 62.2 47.4 71.0




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 10/14/2025
Supervised By :mohammad ahmed 10/15/2025



#2
1,4-Dioxane
Concen: 0.486 ng
RT: 3.283 min Scan# 28
Delta R.T. -0.014 min
Lab File: BN038088.D
Acq: 14 Oct 2025 04:56


Tgt Ion: 88 Resp: 3409
Ion Ratio Lower Upper
88 100
43 27.4 26.6 39.8
58 70.3 58.9 88.3





#3
n-Nitrosodimethylamine
Concen: 0.423 ng
RT: 3.586 min Scan# 70
Delta R.T. -0.014 min
Lab File: BN038088.D
Acq: 14 Oct 2025 04:56




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC



Tgt Ion: 42 Resp: 394
Ion Ratio Lower Upper
42 100
74 111.5 93.4 140.0
44 6.5 22.2 33.4




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 10/14/2025
Supervised By :mohammad ahmed 10/15/2025



#4
2-Fluorophenol
Concen: 0.386 ng
RT: 5.363 min Scan# 316
Delta R.T. -0.022 min
Lab File: BN038088.D
Acq: 14 Oct 2025 04:56


Tgt Ion:112 Resp: 6075
Ion Ratio Lower Upper
112 100
64 57.2 44.6 66.8
63 30.8 24.9 37.3





#5
Phenol-d6
Concen: 0.325 ng
RT: 6.952 min Scan# 531
Delta R.T. -0.022 min
Lab File: BN038088.D
Acq: 14 Oct 2025 04:56




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC



Tgt Ion: 99 Resp: 6734
Ion Ratio Lower Upper
99 100
42 20.6 17.2 25.8
71 31.6 27.3 40.9




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 10/14/2025
Supervised By :mohammad ahmed 10/15/2025



#6
bis(2-Chloroethyl)ether
Concen: 0.387 ng
RT: 7.219 min Scan# 573
Delta R.T. -0.022 min
Lab File: BN038088.D
Acq: 14 Oct 2025 04:56


Tgt Ion: 93 Resp: 7442
Ion Ratio Lower Upper
93 100
63 74.5 60.1 90.1
95 31.9 25.4 38.2





#7
Naphthalene-d8
Concen: 0.400 ng
RT: 10.594 min Scan# 941
Delta R.T. -0.032 min
Lab File: BN038088.D
Acq: 14 Oct 2025 04:56




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC



Tgt Ion: 136 Resp: 18816
Ion Ratio Lower Upper
136 100
137 11.3 9.0 13.4
54 9.5 5.8 8.8
68 6.8 4.3 6.5




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 10/14/2025
Supervised By :mohammad ahmed 10/15/2025



#8
Nitrobenzene-d5
Concen: 0.335 ng
RT: 8.950 min Scan# 784
Delta R.T. -0.021 min
Lab File: BN038088.D
Acq: 14 Oct 2025 04:56


Tgt Ion: 82 Resp: 4804
Ion Ratio Lower Upper
82 100
128 46.7 34.8 52.2
54 51.8 42.2 63.2





#9
Naphthalene
Concen: 0.391 ng
RT: 10.648 min Scan# 945
Delta R.T. -0.021 min
Lab File: BN038088.D
Acq: 14 Oct 2025 04:56




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC



Tgt Ion:128 Resp: 20288
Ion Ratio Lower Upper
128 100
129 11.3 9.2 13.8
127 13.6 11.0 16.6




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 10/14/2025
Supervised By :mohammad ahmed 10/15/2025



#10
Hexachlorobutadiene
Concen: 0.400 ng
RT: 10.947 min Scan# 971
Delta R.T. -0.032 min
Lab File: BN038088.D
Acq: 14 Oct 2025 04:56


Tgt Ion:225 Resp: 3539
Ion Ratio Lower Upper
225 100
223 0.0 0.0 0.0
227 62.1 51.4 77.2





#11
2-Methylnaphthalene-d10
Concen: 0.387 ng
RT: 12.192 min Scan# 1163
Delta R.T. -0.025 min
Lab File: BN038088.D
Acq: 14 Oct 2025 04:56




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC



Tgt Ion:152 Resp: 10106
Ion Ratio Lower Upper
152 100
151 20.9 17.4 26.2




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 10/14/2025
Supervised By :mohammad ahmed 10/15/2025



#12
2-Methylnaphthalene
Concen: 0.365 ng
RT: 12.268 min Scan# 1173
Delta R.T. -0.025 min
Lab File: BN038088.D
Acq: 14 Oct 2025 04:56


Tgt Ion:142 Resp: 12369
Ion Ratio Lower Upper
142 100
141 88.9 71.2 106.8
115 35.7 29.7 44.5





#13
Acenaphthene-d10
Concen: 0.400 ng
RT: 14.452 min Scan# 1456
Delta R.T. -0.021 min
Lab File: BN038088.D
Acq: 14 Oct 2025 04:56




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC



Tgt Ion:164 Resp: 9558
Ion Ratio Lower Upper
164 100
162 105.4 84.8 127.2
160 55.4 46.1 69.1




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 10/14/2025
Supervised By :mohammad ahmed 10/15/2025



#14
2,4,6-Tribromophenol
Concen: 0.315 ng
RT: 15.945 min Scan# 1588
Delta R.T. -0.024 min
Lab File: BN038088.D
Acq: 14 Oct 2025 04:56

Tgt Ion:330 Resp: 1142
Ion Ratio Lower Upper
330 100
332 97.7 77.2 115.8
141 45.8 37.2 55.8






#15
2-Fluorobiphenyl
Concen: 0.385 ng
RT: 13.073 min Scan# 1327
Delta R.T. -0.021 min
Lab File: BN038088.D
Acq: 14 Oct 2025 04:56

Instrument :

BNA_N

ClientSampleId :

SSTDCCC0.4EC






Tgt Ion:172 Resp: 15070
Ion Ratio Lower Upper
172 100
171 35.2 28.6 43.0
170 24.3 19.7 29.5


Manual Integrations

APPROVED



Reviewed By :Jagrut Upadhyay 10/14/2025
Supervised By :mohammad ahmed 10/15/2025




#16
Acenaphthylene
Concen: 0.379 ng
RT: 14.174 min Scan# 1428
Delta R.T. -0.021 min
Lab File: BN038088.D
Acq: 14 Oct 2025 04:56




Tgt Ion:152 Resp: 16454
Ion Ratio Lower Upper
152 100
151 19.5 15.8 23.8
153 13.1 10.5 15.7





#17
Acenaphthene
Concen: 0.375 ng
RT: 14.516 min Scan# 1460
Delta R.T. -0.021 min
Lab File: BN038088.D
Acq: 14 Oct 2025 04:56




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC



Tgt Ion:154 Resp: 11586
Ion Ratio Lower Upper
154 100
153 113.4 90.5 135.7
152 63.0 51.8 77.8




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 10/14/2025
Supervised By :mohammad ahmed 10/15/2025



#18
Fluorene
Concen: 0.349 ng
RT: 15.499 min Scan# 1552
Delta R.T. -0.023 min
Lab File: BN038088.D
Acq: 14 Oct 2025 04:56

Tgt Ion:166 Resp: 13064
Ion Ratio Lower Upper
166 100
165 98.7 79.0 118.4
167 13.2 10.6 16.0









#19
Phenanthrene-d10
Concen: 0.400 ng
RT: 17.198 min Scan# 1689
Delta R.T. -0.025 min
Lab File: BN038088.D
Acq: 14 Oct 2025 04:56

Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC

Tgt Ion:188 Resp: 17543
Ion Ratio Lower Upper
188 100
94 0.0 0.0 0.0
80 13.4 7.6 11.4




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 10/14/2025
Supervised By :mohammad ahmed 10/15/2025



#20
4,6-Dinitro-2-methylphenol
Concen: 0.374 ng
RT: 15.585 min Scan# 1559
Delta R.T. -0.012 min
Lab File: BN038088.D
Acq: 14 Oct 2025 04:56


Tgt Ion:198 Resp: 690
Ion Ratio Lower Upper
198 100
51 165.8 59.1 88.7#
105 67.8 41.3 61.9#





#21
4-Bromophenyl-phenylether
Concen: 0.348 ng
RT: 16.391 min Scan# 1624
Delta R.T. -0.025 min
Lab File: BN038088.D
Acq: 14 Oct 2025 04:56




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC



Tgt Ion:248 Resp: 3981
Ion Ratio Lower Upper
248 100
250 92.5 77.6 116.4
141 89.3 48.8 73.2




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 10/14/2025
Supervised By :mohammad ahmed 10/15/2025



#22
Hexachlorobenzene
Concen: 0.377 ng
RT: 16.515 min Scan# 1634
Delta R.T. -0.025 min
Lab File: BN038088.D
Acq: 14 Oct 2025 04:56


Tgt Ion:284 Resp: 5224
Ion Ratio Lower Upper
284 100
142 38.4 30.9 46.3
249 30.2 23.9 35.9





#23
Atrazine
Concen: 0.324 ng
RT: 16.664 min Scan# 1646
Delta R.T. -0.025 min
Lab File: BN038088.D
Acq: 14 Oct 2025 04:56




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC



Tgt Ion:200 Resp: 2825
Ion Ratio Lower Upper
200 100
173 33.5 18.3 27.5
215 45.1 38.6 58.0




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 10/14/2025
Supervised By :mohammad ahmed 10/15/2025



#24
Pentachlorophenol
Concen: 0.339 ng
RT: 16.863 min Scan# 1662
Delta R.T. -0.012 min
Lab File: BN038088.D
Acq: 14 Oct 2025 04:56


Tgt Ion:266 Resp: 1613
Ion Ratio Lower Upper
266 100
264 64.8 50.9 76.3
268 64.7 54.3 81.5





#25
Phenanthrene
Concen: 0.360 ng
RT: 17.248 min Scan# 1694
Delta R.T. -0.012 min
Lab File: BN038088.D
Acq: 14 Oct 2025 04:56




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC



Tgt Ion:178 Resp: 20782
Ion Ratio Lower Upper
178 100
176 19.4 15.7 23.5
179 15.5 12.3 18.5




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 10/14/2025
Supervised By :mohammad ahmed 10/15/2025



#26
Anthracene
Concen: 0.340 ng
RT: 17.335 min Scan# 1700
Delta R.T. -0.025 min
Lab File: BN038088.D
Acq: 14 Oct 2025 04:56


Tgt Ion:178 Resp: 17015
Ion Ratio Lower Upper
178 100
176 18.6 15.0 22.4
179 15.2 12.3 18.5





#27
Fluoranthene-d10
Concen: 0.389 ng
RT: 19.239 min Scan# 1951
Delta R.T. -0.014 min
Lab File: BN038088.D
Acq: 14 Oct 2025 04:56




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC




Tgt Ion:212 Resp: 17173
Ion Ratio Lower Upper
212 100
106 16.4 12.6 19.0
104 9.2 7.4 11.0

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 10/14/2025
Supervised By :mohammad ahmed 10/15/2025




#28
Fluoranthene
Concen: 0.357 ng
RT: 19.266 min Scan# 1957
Delta R.T. -0.019 min
Lab File: BN038088.D
Acq: 14 Oct 2025 04:56




Tgt Ion:202 Resp: 22691
Ion Ratio Lower Upper
202 100
101 12.2 9.3 13.9
203 17.2 13.6 20.4





#29
Chrysene-d12
Concen: 0.400 ng
RT: 21.384 min Scan# 21
Delta R.T. -0.018 min
Lab File: BN038088.D
Acq: 14 Oct 2025 04:56




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC



Tgt Ion:240 Resp: 14165
Ion Ratio Lower Upper
240 100
120 20.2 11.4 17.2
236 29.2 23.0 34.6




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 10/14/2025
Supervised By :mohammad ahmed 10/15/2025



#30
Pyrene
Concen: 0.417 ng
RT: 19.629 min Scan# 2035
Delta R.T. -0.019 min
Lab File: BN038088.D
Acq: 14 Oct 2025 04:56

Tgt Ion:202 Resp: 23294
Ion Ratio Lower Upper
202 100
200 21.0 16.7 25.1
203 17.6 14.2 21.4






#31
Terphenyl-d14
Concen: 0.376 ng
RT: 19.833 min Scan# 2083
Delta R.T. -0.019 min
Lab File: BN038088.D
Acq: 14 Oct 2025 04:56

Instrument :

BNA_N

ClientSampleId :

SSTDCCC0.4EC






Tgt Ion:244 Resp: 10595
Ion Ratio Lower Upper
244 100
212 9.5 7.6 11.4
122 18.1 13.9 20.9


Manual Integrations

APPROVED



Reviewed By :Jagrut Upadhyay 10/14/2025
Supervised By :mohammad ahmed 10/15/2025




#32
Benzo(a)anthracene
Concen: 0.366 ng
RT: 21.375 min Scan# 2269
Delta R.T. -0.009 min
Lab File: BN038088.D
Acq: 14 Oct 2025 04:56




Tgt Ion:228 Resp: 18122
Ion Ratio Lower Upper
228 100
226 27.1 22.6 33.8
229 19.9 15.6 23.4





#33
Chrysene
Concen: 0.403 ng
RT: 21.420 min Scan# 2148
Delta R.T. -0.018 min
Lab File: BN038088.D
Acq: 14 Oct 2025 04:56




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC



Tgt Ion:228 Resp: 21548
Ion Ratio Lower Upper
228 100
226 30.4 24.6 37.0
229 19.1 15.6 23.4




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 10/14/2025
Supervised By :mohammad ahmed 10/15/2025



#34
Bis(2-ethylhexyl)phthalate
Concen: 0.392 ng
RT: 21.304 min Scan# 2261
Delta R.T. -0.018 min
Lab File: BN038088.D
Acq: 14 Oct 2025 04:56


Tgt Ion:149 Resp: 7685
Ion Ratio Lower Upper
149 100
167 26.2 20.4 30.6
279 3.6 2.4 3.6#





#35
Perylene-d12
Concen: 0.400 ng
RT: 23.706 min Scan# 2829
Delta R.T. -0.020 min
Lab File: BN038088.D
Acq: 14 Oct 2025 04:56




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC



Tgt Ion:264 Resp: 13844
Ion Ratio Lower Upper
264 100
260 26.7 21.5 32.3
265 86.9 53.8 80.6




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 10/14/2025
Supervised By :mohammad ahmed 10/15/2025



#36
Indeno(1,2,3-cd)pyrene
Concen: 0.403 ng
RT: 26.072 min Scan# 3638
Delta R.T. -0.035 min
Lab File: BN038088.D
Acq: 14 Oct 2025 04:56


Tgt Ion:276 Resp: 25477
Ion Ratio Lower Upper
276 100
138 27.1 21.4 32.0
277 24.1 20.3 30.5





#37
Benzo(b)fluoranthene
Concen: 0.372 ng m
RT: 23.005 min Scan# 21
Delta R.T. -0.021 min
Lab File: BN038088.D
Acq: 14 Oct 2025 04:56




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC



Tgt Ion:252 Resp: 2220
Ion Ratio Lower Upper
252 100
253 25.7 19.4 29.0
125 20.4 13.0 19.4




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 10/14/2025
Supervised By :mohammad ahmed 10/15/2025



#38
Benzo(k)fluoranthene
Concen: 0.376 ng
RT: 23.049 min Scan# 2604
Delta R.T. -0.023 min
Lab File: BN038088.D
Acq: 14 Oct 2025 04:56

Tgt Ion:252 Resp: 22592
Ion Ratio Lower Upper
252 100
253 26.0 19.5 29.3
125 19.0 13.0 19.4





#39

Benzo(a)pyrene

Concen: 0.395 ng

RT: 23.607 min Scan# 21

Delta R.T. -0.018 min

Lab File: BN038088.D


Acq: 14 Oct 2025 04:56

Instrument :

BNA_N

ClientSampleId :

SSTDCCC0.4EC



Tgt Ion:252 Resp: 18632

Ion Ratio Lower Upper

252 100

253 28.2 20.6 30.8




125 31.4 16.7 25.1

Manual Integrations

APPROVED

Reviewed By :Jagrut Upadhyay 10/14/2025

Supervised By :mohammad ahmed 10/15/2025



#40

Dibenzo(a,h)anthracene


Concen: 0.381 ng

RT: 26.089 min Scan# 3644

Delta R.T. -0.029 min

Lab File: BN038088.D

Acq: 14 Oct 2025 04:56





Tgt Ion:278 Resp: 18798


Ion Ratio Lower Upper

278 100

139 20.7 15.4 23.2

279 27.4 20.8 31.2





#41

Benzo(g,h,i)perylene

Concen: 0.432 ng

RT: 26.788 min Scan# 38

Delta R.T. -0.041 min

Lab File: BN038088.D


Acq: 14 Oct 2025 04:56

Instrument :

BNA_N

ClientSampleId :

SSTDCCC0.4EC



Tgt Ion:276 Resp: 22396

Ion Ratio Lower Upper

276 100

277 24.0 20.2 30.4



138 25.6 19.8 29.8

Manual Integrations

APPROVED

Reviewed By :Jagrut Upadhyay 10/14/2025

Supervised By :mohammad ahmed 10/15/2025



Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN101325\
 Data File : BN038088.D
 Acq On : 14 Oct 2025 04:56
 Operator : RC/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: Oct 14 05:20:56 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN092325.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Sep 24 11:00:01 2025
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	124	-0.03
2	1,4-Dioxane	0.406	0.494	-21.7	149	-0.01
3	n-Nitrosodimethylamine	0.540	0.572	-5.9	137	-0.01
4 S	2-Fluorophenol	0.913	0.880	3.6	127	-0.02
5 S	Phenol-d6	1.199	0.975	18.7	108	-0.02
6	bis(2-Chloroethyl)ether	1.113	1.078	3.1	123	-0.02
7 I	Naphthalene-d8	1.000	1.000	0.0	119	-0.03
8 S	Nitrobenzene-d5	0.305	0.255	16.4	105	-0.02
9	Naphthalene	1.102	1.078	2.2	123	-0.02
10	Hexachlorobutadiene	0.188	0.188	0.0	123	-0.03
11 SURR	2-Methylnaphthalene-d10	0.556	0.537	3.4	128	-0.03
12	2-Methylnaphthalene	0.720	0.657	8.7	117	-0.03
13 I	Acenaphthene-d10	1.000	1.000	0.0	116	-0.02
14 S	2,4,6-Tribromophenol	0.152	0.119	21.7	102	-0.02
15 S	2-Fluorobiphenyl	1.638	1.577	3.7	117	-0.02
16	Acenaphthylene	1.816	1.721	5.2	120	-0.02
17	Acenaphthene	1.294	1.212	6.3	117	-0.02
18	Fluorene	1.565	1.367	12.7	109	-0.02
19 I	Phenanthrene-d10	1.000	1.000	0.0	110	-0.02
20	4,6-Dinitro-2-methylphenol	0.057	0.039	31.6#	99	-0.01
21	4-Bromophenyl-phenylether	0.261	0.227	13.0	102	-0.02
22	Hexachlorobenzene	0.316	0.298	5.7	108	-0.02
23	Atrazine	0.199	0.161	19.1	103	-0.02
24	Pentachlorophenol	0.108	0.092	14.8	110	-0.01
25	Phenanthrene	1.316	1.185	10.0	106	-0.01
26	Anthracene	1.142	0.970	15.1	104	-0.02
27 SURR	Fluoranthene-d10	1.006	0.979	2.7	122	-0.01
28	Fluoranthene	1.450	1.293	10.8	109	-0.02
29 I	Chrysene-d12	1.000	1.000	0.0	104	-0.02
30	Pyrene	1.579	1.644	-4.1	113	-0.02
31 S	Terphenyl-d14	0.797	0.748	6.1	104	-0.02
32	Benzo(a)anthracene	1.398	1.279	8.5	103	0.00
33	Chrysene	1.512	1.521	-0.6	109	-0.02
34	Bis(2-ethylhexyl)phthalate	0.553	0.543	1.8	116	-0.02
35 I	Perylene-d12	1.000	1.000	0.0	110	-0.02
36	Indeno(1,2,3-cd)pyrene	1.828	1.840	-0.7	122	-0.04
37	Benzo(b)fluoranthene	1.725	1.604	7.0	108	-0.02
38	Benzo(k)fluoranthene	1.736	1.632	6.0	112	-0.02
39 C	Benzo(a)pyrene	1.364	1.346	1.3	120	-0.02
40	Dibenzo(a,h)anthracene	1.427	1.358	4.8	115	-0.03
41	Benzo(g,h,i)perylene	1.499	1.618	-7.9	127	-0.04

(#) = Out of Range

SPCC's out = 0 CCC's out = 0

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN101325\
 Data File : BN038088.D
 Acq On : 14 Oct 2025 04:56
 Operator : RC/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: Oct 14 05:20:56 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN092325.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Sep 24 11:00:01 2025
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	124	-0.03
2	1,4-Dioxane	0.400	0.486	-21.5	149	-0.01
3	n-Nitrosodimethylamine	0.400	0.423	-5.7	137	-0.01
4 S	2-Fluorophenol	0.400	0.386	3.5	127	-0.02
5 S	Phenol-d6	0.400	0.325	18.8	108	-0.02
6	bis(2-Chloroethyl)ether	0.400	0.387	3.3	123	-0.02
7 I	Naphthalene-d8	0.400	0.400	0.0	119	-0.03
8 S	Nitrobenzene-d5	0.400	0.335	16.3	105	-0.02
9	Naphthalene	0.400	0.391	2.3	123	-0.02
10	Hexachlorobutadiene	0.400	0.400	0.0	123	-0.03
11 SURR	2-Methylnaphthalene-d10	0.400	0.387	3.3	128	-0.03
12	2-Methylnaphthalene	0.400	0.365	8.8	117	-0.03
13 I	Acenaphthene-d10	0.400	0.400	0.0	116	-0.02
14 S	2,4,6-Tribromophenol	0.400	0.315	21.3	102	-0.02
15 S	2-Fluorobiphenyl	0.400	0.385	3.8	117	-0.02
16	Acenaphthylene	0.400	0.379	5.3	120	-0.02
17	Acenaphthene	0.400	0.375	6.3	117	-0.02
18	Fluorene	0.400	0.349	12.8	109	-0.02
19 I	Phenanthrene-d10	0.400	0.400	0.0	110	-0.02
20	4,6-Dinitro-2-methylphenol	0.400	0.374	6.5	99	-0.01
21	4-Bromophenyl-phenylether	0.400	0.348	13.0	102	-0.02
22	Hexachlorobenzene	0.400	0.377	5.8	108	-0.02
23	Atrazine	0.400	0.324	19.0	103	-0.02
24	Pentachlorophenol	0.400	0.339	15.3	110	-0.01
25	Phenanthrene	0.400	0.360	10.0	106	-0.01
26	Anthracene	0.400	0.340	15.0	104	-0.02
27 SURR	Fluoranthene-d10	0.400	0.389	2.8	122	-0.01
28	Fluoranthene	0.400	0.357	10.8	109	-0.02
29 I	Chrysene-d12	0.400	0.400	0.0	104	-0.02
30	Pyrene	0.400	0.417	-4.2	113	-0.02
31 S	Terphenyl-d14	0.400	0.376	6.0	104	-0.02
32	Benzo(a)anthracene	0.400	0.366	8.5	103	0.00
33	Chrysene	0.400	0.403	-0.8	109	-0.02
34	Bis(2-ethylhexyl)phthalate	0.400	0.392	2.0	116	-0.02
35 I	Perylene-d12	0.400	0.400	0.0	110	-0.02
36	Indeno(1,2,3-cd)pyrene	0.400	0.403	-0.8	122	-0.04
37	Benzo(b)fluoranthene	0.400	0.372	7.0	108	-0.02
38	Benzo(k)fluoranthene	0.400	0.376	6.0	112	-0.02
39 C	Benzo(a)pyrene	0.400	0.395	1.3	120	-0.02
40	Dibenzo(a,h)anthracene	0.400	0.381	4.8	115	-0.03
41	Benzo(g,h,i)perylene	0.400	0.432	-8.0	127	-0.04

(#) = Out of Range

SPCC's out = 0 CCC's out = 0





QC SAMPLE DATA

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN092325\
Data File : BN037860.D
Acq On : 23 Sep 2025 11:33
Operator : RC/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
DFTPP

Integration File: rteint.p

Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN092325.M
Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
Last Update : Tue Sep 23 17:30:29 2025






AutoFind: Scans 2481, 2482, 2483; Background Corrected with Scan 2474

Target	Rel. to	Lower	Upper	Rel.	Raw	Result
Mass	Mass	Limit%	Limit%	Abn%	Abn	Pass/Fail
68	69	0.00	2	1.8	7592	PASS
69	69	100	100	100.0	410472	PASS
70	69	0.00	2	0.6	2535	PASS
197	198	0.00	2	0.0	0	PASS
198	198	100	100	100.0	1043541	PASS
199	198	5	9	6.9	71928	PASS
365	198	1	100	4.2	43627	PASS
441	443	0.01	150	72.6	133917	PASS
442	442	100	100	100.0	886933	PASS
443	442	15	24	20.8	184525	PASS

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN092325\
Data File : BN037860.D
Acq On : 23 Sep 2025 11:33
Operator : RC/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
DFTPP

Quant Time: Sep 23 19:20:36 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270E-Tune.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Sat Sep 20 01:53:13 2025
Response via : Initial Calibration



TIC: BN037860.D\data.ms

(70) Pentachlorophenol (C)

16.886min (+ 0.001) 30744.23 ng




response 1679260

Ion	Exp%	Act%
265.70	100.00	100.00
268.00	62.20	67.46
264.00	61.60	66.17
0.00	0.00	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN092325\
Data File : BN037860.D
Acq On : 23 Sep 2025 11:33
Operator : RC/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
DFTPP

Quant Time: Sep 23 19:20:36 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270E-Tune.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Sat Sep 20 01:53:13 2025
Response via : Initial Calibration



TIC: BN037860.D\data.ms

(77) Benzdine

19.486min (+ 0.012) 0.00 ng

response 9413593

Ion	Exp%	Act%
184.00	100.00	100.00
185.00	15.50	14.50
183.00	13.20	12.37
0.00	0.00	0.00

Instrument :
BNA_N
ClientSampleId :
DFTPP

DDT Breakdown


Date	Instrument Name	DFTPP Data File
9/23/2025	BNA_N	<u>BN037860.D</u>
Compound Name	Response	Retention Time
DDT	4305996	20.716
DDD	11703	20.327
DDE	258	19.769
SUM(DDD+DDE)	SUM(DDT+DDD+DDE)	% Breakdown Of DDT
11961	4317957	0.28

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN101325\
Data File : BN038065.D
Acq On : 13 Oct 2025 13:08
Operator : RC/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
DFTPP

Integration File: rteint.p

Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN092325.M
Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
Last Update : Wed Sep 24 11:00:01 2025






AutoFind: Scans 2476, 2477, 2478; Background Corrected with Scan 2469

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
68	69	0.00	2	1.2	2414	PASS
69	69	100	100	100.0	207810	PASS
70	69	0.00	2	0.5	1031	PASS
197	198	0.00	2	0.2	1009	PASS
198	198	100	100	100.0	519680	PASS
199	198	5	9	6.7	34837	PASS
365	198	1	100	3.4	17780	PASS
441	443	0.01	150	70.3	70149	PASS
442	442	100	100	100.0	475904	PASS
443	442	15	24	21.0	99800	PASS

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN101325\
Data File : BN038065.D
Acq On : 13 Oct 2025 13:08
Operator : RC/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
DFTPP

Quant Time: Oct 13 13:56:12 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270E-Tune.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Oct 13 13:56:00 2025
Response via : Initial Calibration



TIC: BN038065.D\data.ms

(70) Pentachlorophenol (C)

16.857min (0.000) 0.00 ng




response 981737

Ion	Exp%	Act%
265.70	100.00	100.00
268.00	62.20	66.35
264.00	61.60	61.83
0.00	0.00	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN101325\
Data File : BN038065.D
Acq On : 13 Oct 2025 13:08
Operator : RC/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
DFTPP

Quant Time: Oct 13 13:56:12 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270E-Tune.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Oct 13 13:56:00 2025
Response via : Initial Calibration



TIC: BN038065.D\data.ms

(77) Benizidine

19.457min (0.000) 0.00 ng

response 4930116

Ion	Exp%	Act%
184.00	100.00	100.00
185.00	15.50	13.95
183.00	13.20	12.60
0.00	0.00	0.00

Instrument :
 BNA_N
 ClientSampleId :
 DFTPP

DDT Breakdown


Date	Instrument Name	DFTPP Data File
10/13/2025	BNA_N	<u>BN038065.D</u>
Compound Name	Response	Retention Time
DDT	1986801	20.692
DDD	17211	20.251
DDE	771	19.745
SUM(DDD+DDE)	SUM(DDT+DDD+DDE)	% Breakdown Of DDT
17982	2004783	0.90

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN101325\
Data File : BN038082.D
Acq On : 14 Oct 2025 00:40
Operator : RC/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
DFTPP

Integration File: rteint.p

Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN092325.M
Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
Last Update : Wed Sep 24 11:00:01 2025






AutoFind: Scans 2476, 2477, 2478; Background Corrected with Scan 2469

Target	Rel. to	Lower	Upper	Rel.	Raw	Result
Mass	Mass	Limit%	Limit%	Abn%	Abn	Pass/Fail
68	69	0.00	2	1.7	3299	PASS
69	69	100	100	100.0	194987	PASS
70	69	0.00	2	0.5	982	PASS
197	198	0.00	2	0.4	1825	PASS
198	198	100	100	100.0	456107	PASS
199	198	5	9	6.8	31077	PASS
365	198	1	100	3.5	16041	PASS
441	443	0.01	150	83.1	64464	PASS
442	442	100	100	100.0	430080	PASS
443	442	15	24	18.0	77528	PASS

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN101325\
Data File : BN038082.D
Acq On : 14 Oct 2025 00:40
Operator : RC/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
DFTPP

Quant Time: Oct 14 04:05:57 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270E-Tune.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Oct 13 13:56:00 2025
Response via : Initial Calibration



TIC: BN038082.D\data.ms

(70) Pentachlorophenol (C)

16.851min (-0.035) 0.00 ng




response 862228

Ion	Exp%	Act%
265.70	100.00	100.00
268.00	62.20	66.16
264.00	61.60	58.53
0.00	0.00	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN101325\
Data File : BN038082.D
Acq On : 14 Oct 2025 00:40
Operator : RC/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
DFTPP

Quant Time: Oct 14 04:05:57 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270E-Tune.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Oct 13 13:56:00 2025
Response via : Initial Calibration



TIC: BN038082.D\data.ms

(77) Benzidine

19.451min (-0.023) 0.00 ng

response 4359890

Ion	Exp%	Act%
184.00	100.00	100.00
185.00	15.50	13.86
183.00	13.20	12.55
0.00	0.00	0.00

Instrument :
BNA_N
ClientSampleId :
DFTPP

DDT Breakdown

Date	Instrument Name	DFTPP Data File
10/13/2025	BNA_N	<u>BN038082.D</u>
Compound Name	Response	Retention Time
DDT	1777716	20.692
DDD	13147	20.245
DDE	608	19.745
SUM(DDD+DDE)	SUM(DDT+DDD+DDE)	% Breakdown Of DDT
13755	1791471	0.77



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Report of Analysis

Client:	Tetra Tech NUS, Inc.		Date Collected:	
Project:	NWIRP Bethpage 112G08005-WE13		Date Received:	
Client Sample ID:	PB170022BL		SDG No.:	Q3292
Lab Sample ID:	PB170022BL		Matrix:	Water
Analytical Method:	SW8270ESIM	Level : LOW	% Solid:	0
Sample Wt/Vol:	1000 mL	Final Vol: 1000 uL	Test:	SVOC-SIMGroup1
Prep Method :		Prep Date: 10/08/25		

CAS Number	Parameter	Conc.	Qua.	DF	MDL	LOD	LOQ / CRQL	Units	Date Ana.	Prep BatchID
TARGETS										
123-91-1	1,4-Dioxane	0.20	U	1	0.070	0.20	0.20	ug/L	10/13/25 16:12	PB170022
SURROGATES										
7297-45-2	2-Methylnaphthalene-d10	0.32			30 - 150		80%	SPK: 0.4		
93951-69-0	Fluoranthene-d10	0.35			30 - 150		88%	SPK: 0.4		
4165-60-0	Nitrobenzene-d5	0.37			55 - 111		92%	SPK: 0.4		
321-60-8	2-Fluorobiphenyl	0.38			53 - 106		95%	SPK: 0.4		
1718-51-0	Terphenyl-d14	0.49			58 - 132		123%	SPK: 0.4		
INTERNAL STANDARDS										
		Area								
3855-82-1	1,4-Dichlorobenzene-d4	6810								
1146-65-2	Naphthalene-d8	16600								
15067-26-2	Acenaphthene-d10	7880								
1517-22-2	Phenanthrene-d10	12300								
1719-03-5	Chrysene-d12	8410								
1520-96-3	Perylene-d12	7510								

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN101325\
Data File : BN038070.D
Acq On : 13 Oct 2025 16:12
Operator : RC/JU
Sample : PB170022BL
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
PB170022BL

Quant Time: Oct 13 17:24:28 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN092325.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Sep 24 11:00:01 2025
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.797	152	6807	0.400	ng	-0.03
7) Naphthalene-d8	10.605	136	16619	0.400	ng	-0.02
13) Acenaphthene-d10	14.452	164	7880	0.400	ng	-0.02
19) Phenanthrene-d10	17.211	188	12265	0.400	ng	-0.01
29) Chrysene-d12	21.393	240	8413	0.400	ng	# 0.00
35) Perylene-d12	23.715	264	7507	0.400	ng	#-0.01
System Monitoring Compounds						
4) 2-Fluorophenol	5.370	112	5160	0.332	ng	-0.01
5) Phenol-d6	6.959	99	5208	0.255	ng	-0.01
8) Nitrobenzene-d5	8.950	82	4655	0.367	ng	-0.02
11) 2-Methylnaphthalene-d10	12.197	152	7410	0.321	ng	-0.02
14) 2,4,6-Tribromophenol	15.957	330	518	0.173	ng	-0.01
15) 2-Fluorobiphenyl	13.073	172	12302	0.381	ng	-0.02
27) Fluoranthene-d10	19.239	212	10813	0.350	ng	-0.01
31) Terphenyl-d14	19.838	244	8262	0.493	ng	-0.01


Target Compounds Qvalue


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN101325\
Data File : BN038070.D
Acq On : 13 Oct 2025 16:12
Operator : RC/JU
Sample : PB170022BL
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
PB170022BL


Quant Time: Oct 13 17:24:28 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN092325.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Sep 24 11:00:01 2025
Response via : Initial Calibration








#1
1,4-Dichlorobenzene-d4
Concen: 0.400 ng
RT: 7.797 min Scan# 61
Delta R.T. -0.028 min
Lab File: BN038070.D
Acq: 13 Oct 2025 16:12

Instrument :
BNA_N
ClientSampleId :
PB170022BL







Tgt Ion:152 Resp: 6807
Ion Ratio Lower Upper
152 100
150 157.2 121.0 181.4
115 62.6 47.4 71.0



#4
2-Fluorophenol
Concen: 0.332 ng
RT: 5.370 min Scan# 317
Delta R.T. -0.015 min
Lab File: BN038070.D
Acq: 13 Oct 2025 16:12

Tgt Ion:112 Resp: 5160
Ion Ratio Lower Upper
112 100
64 55.8 44.6 66.8
63 30.0 24.9 37.3









#5
Phenol-d6
Concen: 0.255 ng
RT: 6.959 min Scan# 51
Delta R.T. -0.015 min
Lab File: BN038070.D
Acq: 13 Oct 2025 16:12




Instrument :
BNA_N
ClientSampleId :
PB170022BL


Tgt Ion:	99	Resp:	5208
Ion	Ratio	Lower	Upper
99	100		
42	20.5	17.2	25.8
71	33.0	27.3	40.9



#7
Naphthalene-d8
Concen: 0.400 ng
RT: 10.605 min Scan# 939
Delta R.T. -0.021 min
Lab File: BN038070.D
Acq: 13 Oct 2025 16:12

Tgt Ion:	136	Resp:	16619
Ion	Ratio	Lower	Upper
136	100		
137	11.3	9.0	13.4
54	6.9	5.8	8.8
68	5.3	4.3	6.5









#8
Nitrobenzene-d5
Concen: 0.367 ng
RT: 8.950 min Scan# 786
Delta R.T. -0.021 min
Lab File: BN038070.D
Acq: 13 Oct 2025 16:12




Instrument :
BNA_N
ClientSampleId :
PB170022BL


Tgt Ion:	82	Resp:	4655
Ion	Ratio	Lower	Upper
82	100		
128	43.1	34.8	52.2
54	56.1	42.2	63.2



#11
2-Methylnaphthalene-d10
Concen: 0.321 ng
RT: 12.197 min Scan# 1159
Delta R.T. -0.020 min
Lab File: BN038070.D
Acq: 13 Oct 2025 16:12

Tgt Ion:	152	Resp:	7410
Ion	Ratio	Lower	Upper
152	100		
151	21.3	17.4	26.2









#13
Acenaphthene-d10
Concen: 0.400 ng
RT: 14.452 min Scan# 1456
Delta R.T. -0.021 min
Lab File: BN038070.D
Acq: 13 Oct 2025 16:12




Instrument :
BNA_N
ClientSampleId :
PB170022BL


Tgt Ion:164 Resp: 7880
Ion Ratio Lower Upper
164 100
162 105.9 84.8 127.2
160 56.6 46.1 69.1



#14
2,4,6-Tribromophenol
Concen: 0.173 ng
RT: 15.957 min Scan# 1589
Delta R.T. -0.012 min
Lab File: BN038070.D
Acq: 13 Oct 2025 16:12

Tgt Ion:330 Resp: 518
Ion Ratio Lower Upper
330 100
332 95.0 77.2 115.8
141 47.1 37.2 55.8









#15
2-Fluorobiphenyl
Concen: 0.381 ng
RT: 13.073 min Scan# 1327
Delta R.T. -0.021 min
Lab File: BN038070.D
Acq: 13 Oct 2025 16:12




Instrument :
BNA_N
ClientSampleId :
PB170022BL


Tgt Ion	Ratio	Lower	Upper
172	100		
171	36.7	28.6	43.0
170	24.9	19.7	29.5



#19
Phenanthrene-d10
Concen: 0.400 ng
RT: 17.211 min Scan# 1690
Delta R.T. -0.013 min
Lab File: BN038070.D
Acq: 13 Oct 2025 16:12

Tgt Ion	Ratio	Lower	Upper
188	100		
94	0.0	0.0	0.0
80	10.7	7.6	11.4





#27

Fluoranthene-d10

Concen: 0.350 ng

RT: 19.239 min Scan# 1951

Delta R.T. -0.014 min

Lab File: BN038070.D

Acq: 13 Oct 2025 16:12

Instrument :

BNA_N

ClientSampleId :

PB170022BL





Tgt Ion:212 Resp: 10813

Ion Ratio Lower Upper

212 100

106 15.8 12.6 19.0

104 9.1 7.4 11.0



#29

Chrysene-d12

Concen: 0.400 ng

RT: 21.393 min Scan# 2271

Delta R.T. -0.009 min

Lab File: BN038070.D

Acq: 13 Oct 2025 16:12




Tgt Ion:240 Resp: 8413


Ion Ratio Lower Upper

240 100

120 19.3 11.4 17.2#

236 29.2 23.0 34.6









#31
Terphenyl-d14
Concen: 0.493 ng
RT: 19.838 min Scan# 2080
Delta R.T. -0.014 min
Lab File: BN038070.D
Acq: 13 Oct 2025 16:12




Instrument :
BNA_N
ClientSampleId :
PB170022BL

Tgt Ion:244 Resp: 8262
Ion Ratio Lower Upper
244 100
212 9.6 7.6 11.4
122 18.4 13.9 20.9



#35
Perylene-d12
Concen: 0.400 ng
RT: 23.715 min Scan# 2832
Delta R.T. -0.012 min
Lab File: BN038070.D
Acq: 13 Oct 2025 16:12

Tgt Ion:264 Resp: 7507
Ion Ratio Lower Upper
264 100
260 27.3 21.5 32.3
265 166.5 53.8 80.6#





284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Report of Analysis

Client:	Tetra Tech NUS, Inc.		Date Collected:	
Project:	NWIRP Bethpage 112G08005-WE13		Date Received:	
Client Sample ID:	PB170022BS		SDG No.:	Q3292
Lab Sample ID:	PB170022BS		Matrix:	Water
Analytical Method:	SW8270ESIM	Level : LOW	% Solid:	0
Sample Wt/Vol:	1000 mL	Final Vol: 1000 uL	Test:	SVOC-SIMGroup1
Prep Method :		Prep Date: 10/08/25		

CAS Number	Parameter	Conc.	Qua.	DF	MDL	LOD	LOQ / CRQL	Units	Date Ana.	Prep BatchID
TARGETS										
123-91-1	1,4-Dioxane	0.35		1	0.070	0.20	0.20	ug/L	10/13/25 16:48	PB170022
SURROGATES										
7297-45-2	2-Methylnaphthalene-d10	0.36			30 - 150		89%	SPK: 0.4		
93951-69-0	Fluoranthene-d10	0.32			30 - 150		80%	SPK: 0.4		
4165-60-0	Nitrobenzene-d5	0.37			55 - 111		92%	SPK: 0.4		
321-60-8	2-Fluorobiphenyl	0.41			53 - 106		103%	SPK: 0.4		
1718-51-0	Terphenyl-d14	0.49			58 - 132		122%	SPK: 0.4		
INTERNAL STANDARDS										
		Area								
3855-82-1	1,4-Dichlorobenzene-d4	6430								
1146-65-2	Naphthalene-d8	16600								
15067-26-2	Acenaphthene-d10	7630								
1517-22-2	Phenanthrene-d10	12700								
1719-03-5	Chrysene-d12	8200								
1520-96-3	Perylene-d12	7300								

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN101325\
 Data File : BN038071.D
 Acq On : 13 Oct 2025 16:48
 Operator : RC/JU
 Sample : PB170022BS
 Misc :
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 PB170022BS

Quant Time: Oct 13 17:24:42 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN092325.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Sep 24 11:00:01 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.797	152	6425	0.400	ng	-0.03
7) Naphthalene-d8	10.594	136	16580	0.400	ng	#-0.03
13) Acenaphthene-d10	14.452	164	7629	0.400	ng	-0.02
19) Phenanthrene-d10	17.198	188	12726	0.400	ng	#-0.02
29) Chrysene-d12	21.393	240	8201	0.400	ng	# 0.00
35) Perylene-d12	23.712	264	7302	0.400	ng	#-0.01
System Monitoring Compounds						
4) 2-Fluorophenol	5.370	112	5044	0.344	ng	-0.01
5) Phenol-d6	6.959	99	5224	0.271	ng	-0.01
8) Nitrobenzene-d5	8.950	82	4635	0.367	ng	-0.02
11) 2-Methylnaphthalene-d10	12.192	152	8170	0.355	ng	-0.03
14) 2,4,6-Tribromophenol	15.944	330	763	0.264	ng	-0.02
15) 2-Fluorobiphenyl	13.073	172	12858	0.412	ng	-0.02
27) Fluoranthene-d10	19.239	212	10197	0.319	ng	-0.01
31) Terphenyl-d14	19.838	244	7936	0.486	ng	-0.01
Target Compounds						Qvalue
2) 1,4-Dioxane	3.283	88	2277	0.349	ng	# 43
3) n-Nitrosodimethylamine	3.586	42	3145	0.362	ng	# 92
6) bis(2-Chloroethyl)ether	7.219	93	6235	0.349	ng	99
9) Naphthalene	10.648	128	16453	0.360	ng	99
10) Hexachlorobutadiene	10.947	225	3069	0.394	ng	# 98
12) 2-Methylnaphthalene	12.268	142	8956	0.300	ng	99
16) Acenaphthylene	14.174	152	13502	0.390	ng	100
17) Acenaphthene	14.516	154	8445	0.342	ng	98
18) Fluorene	15.499	166	9320	0.312	ng	99
20) 4,6-Dinitro-2-methylph...	15.585	198	362	0.292	ng	# 1
21) 4-Bromophenyl-phenylether	16.391	248	2791	0.337	ng	# 80
22) Hexachlorobenzene	16.515	284	3789	0.377	ng	98
23) Atrazine	16.664	200	1769	0.280	ng	# 93
24) Pentachlorophenol	16.863	266	2006	0.581	ng	97
25) Phenanthrene	17.248	178	14128	0.337	ng	100
26) Anthracene	17.335	178	12363	0.340	ng	99
28) Fluoranthene	19.266	202	13814	0.300	ng	100
30) Pyrene	19.629	202	14038	0.434	ng	100
32) Benzo(a)anthracene	21.375	228	9985	0.348	ng	99
33) Chrysene	21.429	228	12239	0.395	ng	100
34) Bis(2-ethylhexyl)phtha...	21.313	149	4879	0.430	ng	98
36) Indeno(1,2,3-cd)pyrene	26.080	276	12743	0.382	ng	98
37) Benzo(b)fluoranthene	23.011	252	10405	0.330	ng	# 84
38) Benzo(k)fluoranthene	23.057	252	11365	0.359	ng	# 83
39) Benzo(a)pyrene	23.610	252	9245	0.371	ng	# 79
40) Dibenzo(a,h)anthracene	26.104	278	9200	0.353	ng	# 87
41) Benzo(g,h,i)perylene	26.797	276	11736	0.429	ng	97


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN101325\
 Data File : BN038071.D
 Acq On : 13 Oct 2025 16:48
 Operator : RC/JU
 Sample : PB170022BS
 Misc :
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 PB170022BS

Quant Time: Oct 13 17:24:42 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN092325.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Sep 24 11:00:01 2025
 Response via : Initial Calibration









#1
1,4-Dichlorobenzene-d4
Concen: 0.400 ng
RT: 7.797 min Scan# 61
Delta R.T. -0.028 min
Lab File: BN038071.D
Acq: 13 Oct 2025 16:48




Instrument :
BNA_N
ClientSampleId :
PB170022BS


Tgt Ion:152 Resp: 6425
Ion Ratio Lower Upper
152 100
150 154.6 121.0 181.4
115 60.0 47.4 71.0



#2
1,4-Dioxane
Concen: 0.349 ng
RT: 3.283 min Scan# 28
Delta R.T. -0.015 min
Lab File: BN038071.D
Acq: 13 Oct 2025 16:48

Tgt Ion: 88 Resp: 2277
Ion Ratio Lower Upper
88 100
43 95.4 26.6 39.8#
58 101.0 58.9 88.3#









#3
n-Nitrosodimethylamine
Concen: 0.362 ng
RT: 3.586 min Scan# 70
Delta R.T. -0.015 min
Lab File: BN038071.D
Acq: 13 Oct 2025 16:48




Instrument :
BNA_N
ClientSampleId :
PB170022BS


Tgt Ion: 42 Resp: 3145
Ion Ratio Lower Upper
42 100
74 120.6 93.4 140.0
44 14.1 22.2 33.4#



#4
2-Fluorophenol
Concen: 0.344 ng
RT: 5.370 min Scan# 317
Delta R.T. -0.015 min
Lab File: BN038071.D
Acq: 13 Oct 2025 16:48

Tgt Ion: 112 Resp: 5044
Ion Ratio Lower Upper
112 100
64 55.6 44.6 66.8
63 29.9 24.9 37.3









#5
Phenol-d6
Concen: 0.271 ng
RT: 6.959 min Scan# 51
Delta R.T. -0.015 min
Lab File: BN038071.D
Acq: 13 Oct 2025 16:48




Instrument :
BNA_N
ClientSampleId :
PB170022BS


Tgt Ion: 99 Resp: 5224
Ion Ratio Lower Upper
99 100
42 23.5 17.2 25.8
71 32.1 27.3 40.9



#6
bis(2-Chloroethyl)ether
Concen: 0.349 ng
RT: 7.219 min Scan# 573
Delta R.T. -0.022 min
Lab File: BN038071.D
Acq: 13 Oct 2025 16:48

Tgt Ion: 93 Resp: 6235
Ion Ratio Lower Upper
93 100
63 76.5 60.1 90.1
95 31.8 25.4 38.2









#7
Naphthalene-d8
Concen: 0.400 ng
RT: 10.594 min Scan# 941
Delta R.T. -0.032 min
Lab File: BN038071.D
Acq: 13 Oct 2025 16:48




Instrument :
BNA_N
ClientSampleId :
PB170022BS


Tgt Ion:	136	Resp:	16580
Ion	Ratio	Lower	Upper
136	100		
137	11.3	9.0	13.4
54	10.3	5.8	8.8#
68	7.5	4.3	6.5#



#8
Nitrobenzene-d5
Concen: 0.367 ng
RT: 8.950 min Scan# 784
Delta R.T. -0.021 min
Lab File: BN038071.D
Acq: 13 Oct 2025 16:48

Tgt Ion:	82	Resp:	4635
Ion	Ratio	Lower	Upper
82	100		
128	44.2	34.8	52.2
54	53.0	42.2	63.2









#9
Naphthalene
Concen: 0.360 ng
RT: 10.648 min Scan# 94
Delta R.T. -0.021 min
Lab File: BN038071.D
Acq: 13 Oct 2025 16:48




Instrument :
BNA_N
ClientSampleId :
PB170022BS


Tgt Ion:	128	Resp:	16453
Ion	Ratio	Lower	Upper
128	100		
129	11.9	9.2	13.8
127	13.6	11.0	16.6



#10
Hexachlorobutadiene
Concen: 0.394 ng
RT: 10.947 min Scan# 971
Delta R.T. -0.032 min
Lab File: BN038071.D
Acq: 13 Oct 2025 16:48

Tgt Ion:	225	Resp:	3069
Ion	Ratio	Lower	Upper
225	100		
223	0.0	0.0	0.0
227	62.7	51.4	77.2









#11
2-Methylnaphthalene-d10
Concen: 0.355 ng
RT: 12.192 min Scan# 1163
Delta R.T. -0.025 min
Lab File: BN038071.D
Acq: 13 Oct 2025 16:48




Instrument :
BNA_N
ClientSampleId :
PB170022BS


Tgt Ion:152 Resp: 8170
Ion Ratio Lower Upper
152 100
151 20.7 17.4 26.2



#12
2-Methylnaphthalene
Concen: 0.300 ng
RT: 12.268 min Scan# 1173
Delta R.T. -0.025 min
Lab File: BN038071.D
Acq: 13 Oct 2025 16:48

Tgt Ion:142 Resp: 8956
Ion Ratio Lower Upper
142 100
141 90.0 71.2 106.8
115 36.8 29.7 44.5









#13
Acenaphthene-d10
Concen: 0.400 ng
RT: 14.452 min Scan# 1456
Delta R.T. -0.021 min
Lab File: BN038071.D
Acq: 13 Oct 2025 16:48




Instrument :
BNA_N
ClientSampleId :
PB170022BS


Tgt Ion:164 Resp: 7629
Ion Ratio Lower Upper
164 100
162 106.2 84.8 127.2
160 55.7 46.1 69.1



#14
2,4,6-Tribromophenol
Concen: 0.264 ng
RT: 15.944 min Scan# 1588
Delta R.T. -0.025 min
Lab File: BN038071.D
Acq: 13 Oct 2025 16:48

Tgt Ion:330 Resp: 763
Ion Ratio Lower Upper
330 100
332 95.4 77.2 115.8
141 44.4 37.2 55.8









#15
2-Fluorobiphenyl
Concen: 0.412 ng
RT: 13.073 min Scan# 1327
Delta R.T. -0.021 min
Lab File: BN038071.D
Acq: 13 Oct 2025 16:48




Instrument :
BNA_N
ClientSampleId :
PB170022BS


Tgt Ion:	172	Resp:	12858
Ion	Ratio	Lower	Upper
172	100		
171	36.0	28.6	43.0
170	24.7	19.7	29.5



#16
Acenaphthylene
Concen: 0.390 ng
RT: 14.174 min Scan# 1428
Delta R.T. -0.021 min
Lab File: BN038071.D
Acq: 13 Oct 2025 16:48

Tgt Ion:	152	Resp:	13502
Ion	Ratio	Lower	Upper
152	100		
151	19.7	15.8	23.8
153	13.2	10.5	15.7









#17
Acenaphthene
Concen: 0.342 ng
RT: 14.516 min Scan# 1460
Delta R.T. -0.021 min
Lab File: BN038071.D
Acq: 13 Oct 2025 16:48




Instrument :
BNA_N
ClientSampleId :
PB170022BS


Tgt Ion:154 Resp: 8445
Ion Ratio Lower Upper
154 100
153 116.2 90.5 135.7
152 65.6 51.8 77.8



#18
Fluorene
Concen: 0.312 ng
RT: 15.499 min Scan# 1552
Delta R.T. -0.023 min
Lab File: BN038071.D
Acq: 13 Oct 2025 16:48

Tgt Ion:166 Resp: 9320
Ion Ratio Lower Upper
166 100
165 99.5 79.0 118.4
167 13.2 10.6 16.0









#19
Phenanthrene-d10
Concen: 0.400 ng
RT: 17.198 min Scan# 1689
Delta R.T. -0.025 min
Lab File: BN038071.D
Acq: 13 Oct 2025 16:48




Instrument :
BNA_N
ClientSampleId :
PB170022BS


Tgt Ion:188 Resp: 12726
Ion Ratio Lower Upper
188 100
94 0.0 0.0 0.0
80 14.4 7.6 11.4#



#20
4,6-Dinitro-2-methylphenol
Concen: 0.292 ng
RT: 15.585 min Scan# 1559
Delta R.T. -0.012 min
Lab File: BN038071.D
Acq: 13 Oct 2025 16:48

Tgt Ion:198 Resp: 362
Ion Ratio Lower Upper
198 100
51 231.3 59.1 88.7#
105 106.6 41.3 61.9#









#21
4-Bromophenyl-phenylether
Concen: 0.337 ng
RT: 16.391 min Scan# 1624
Delta R.T. -0.025 min
Lab File: BN038071.D
Acq: 13 Oct 2025 16:48




Instrument :
BNA_N
ClientSampleId :
PB170022BS


Tgt Ion:248 Resp: 2791
Ion Ratio Lower Upper
248 100
250 92.6 77.6 116.4
141 94.3 48.8 73.2#



#22
Hexachlorobenzene
Concen: 0.377 ng
RT: 16.515 min Scan# 1634
Delta R.T. -0.025 min
Lab File: BN038071.D
Acq: 13 Oct 2025 16:48

Tgt Ion:284 Resp: 3789
Ion Ratio Lower Upper
284 100
142 40.1 30.9 46.3
249 31.0 23.9 35.9









#23
Atrazine
Concen: 0.280 ng
RT: 16.664 min Scan# 1646
Delta R.T. -0.025 min
Lab File: BN038071.D
Acq: 13 Oct 2025 16:48




Instrument :
BNA_N
ClientSampleId :
PB170022BS


Tgt Ion:200 Resp: 1769
Ion Ratio Lower Upper
200 100
173 30.9 18.3 27.5#
215 46.3 38.6 58.0



#24
Pentachlorophenol
Concen: 0.581 ng
RT: 16.863 min Scan# 1662
Delta R.T. -0.012 min
Lab File: BN038071.D
Acq: 13 Oct 2025 16:48

Tgt Ion:266 Resp: 2006
Ion Ratio Lower Upper
266 100
264 62.3 50.9 76.3
268 64.0 54.3 81.5









#25
Phenanthrene
Concen: 0.337 ng
RT: 17.248 min Scan# 1694
Delta R.T. -0.012 min
Lab File: BN038071.D
Acq: 13 Oct 2025 16:48




Instrument :
BNA_N
ClientSampleId :
PB170022BS


Tgt Ion:178 Resp: 14128
Ion Ratio Lower Upper
178 100
176 19.8 15.7 23.5
179 15.3 12.3 18.5



#26
Anthracene
Concen: 0.340 ng
RT: 17.335 min Scan# 1700
Delta R.T. -0.025 min
Lab File: BN038071.D
Acq: 13 Oct 2025 16:48

Tgt Ion:178 Resp: 12363
Ion Ratio Lower Upper
178 100
176 19.1 15.0 22.4
179 15.5 12.3 18.5









#27
Fluoranthene-d10
Concen: 0.319 ng
RT: 19.239 min Scan# 1951
Delta R.T. -0.014 min
Lab File: BN038071.D
Acq: 13 Oct 2025 16:48




Instrument :
BNA_N
ClientSampleId :
PB170022BS


Tgt Ion	Ratio	Lower	Upper
212	100		
106	16.0	12.6	19.0
104	9.6	7.4	11.0



#28
Fluoranthene
Concen: 0.300 ng
RT: 19.266 min Scan# 1957
Delta R.T. -0.019 min
Lab File: BN038071.D
Acq: 13 Oct 2025 16:48

Tgt Ion	Ratio	Lower	Upper
202	100		
101	11.9	9.3	13.9
203	17.1	13.6	20.4









#29
Chrysene-d12
Concen: 0.400 ng
RT: 21.393 min Scan# 21
Delta R.T. -0.009 min
Lab File: BN038071.D
Acq: 13 Oct 2025 16:48




Instrument :
BNA_N
ClientSampleId :
PB170022BS


Tgt Ion	Ratio	Lower	Upper
240	100		
120	19.0	11.4	17.2#
236	30.6	23.0	34.6



#30
Pyrene
Concen: 0.434 ng
RT: 19.629 min Scan# 2035
Delta R.T. -0.019 min
Lab File: BN038071.D
Acq: 13 Oct 2025 16:48

Tgt Ion	Ratio	Lower	Upper
202	100		
200	20.8	16.7	25.1
203	17.7	14.2	21.4









#31
Terphenyl-d14
Concen: 0.486 ng
RT: 19.838 min Scan# 2080
Delta R.T. -0.014 min
Lab File: BN038071.D
Acq: 13 Oct 2025 16:48




Instrument :
BNA_N
ClientSampleId :
PB170022BS


Tgt Ion	Ratio	Lower	Upper
244	100		
212	9.9	7.6	11.4
122	17.8	13.9	20.9



#32
Benzo(a)anthracene
Concen: 0.348 ng
RT: 21.375 min Scan# 2269
Delta R.T. -0.009 min
Lab File: BN038071.D
Acq: 13 Oct 2025 16:48

Tgt Ion	Ratio	Lower	Upper
228	100		
226	28.9	22.6	33.8
229	19.6	15.6	23.4









#33
Chrysene
Concen: 0.395 ng
RT: 21.429 min Scan# 21
Delta R.T. -0.009 min
Lab File: BN038071.D
Acq: 13 Oct 2025 16:48




Instrument :
BNA_N
ClientSampleId :
PB170022BS


Tgt Ion	Ratio	Lower	Upper
228	100		
226	30.8	24.6	37.0
229	19.5	15.6	23.4



#34
Bis(2-ethylhexyl)phthalate
Concen: 0.430 ng
RT: 21.313 min Scan# 2262
Delta R.T. -0.009 min
Lab File: BN038071.D
Acq: 13 Oct 2025 16:48

Tgt Ion	Ratio	Lower	Upper
149	100		
167	24.7	20.4	30.6
279	3.4	2.4	3.6









#35
Perylene-d12
Concen: 0.400 ng
RT: 23.712 min Scan# 2831
Delta R.T. -0.015 min
Lab File: BN038071.D
Acq: 13 Oct 2025 16:48




Instrument :
BNA_N
ClientSampleId :
PB170022BS


Tgt Ion:264 Resp: 7302
Ion Ratio Lower Upper
264 100
260 26.9 21.5 32.3
265 151.7 53.8 80.6#



#36
Indeno(1,2,3-cd)pyrene
Concen: 0.382 ng
RT: 26.080 min Scan# 3641
Delta R.T. -0.026 min
Lab File: BN038071.D
Acq: 13 Oct 2025 16:48

Tgt Ion:276 Resp: 12743
Ion Ratio Lower Upper
276 100
138 25.7 21.4 32.0
277 24.3 20.3 30.5





#37

Benzo(b)fluoranthene

Concen: 0.330 ng

RT: 23.011 min Scan# 21

Delta R.T. -0.015 min

Lab File: BN038071.D

Acq: 13 Oct 2025 16:48

Instrument :

BNA_N

ClientSampleId :

PB170022BS





Tgt Ion:252 Resp: 10405

Ion Ratio Lower Upper

252 100

253 29.1 19.4 29.0#

125 26.7 13.0 19.4#



#38

Benzo(k)fluoranthene

Concen: 0.359 ng

RT: 23.057 min Scan# 2607

Delta R.T. -0.015 min

Lab File: BN038071.D

Acq: 13 Oct 2025 16:48




Tgt Ion:252 Resp: 11365


Ion Ratio Lower Upper

252 100

253 29.9 19.5 29.3#

125 26.7 13.0 19.4#





#39

Benzo(a)pyrene

Concen: 0.371 ng

RT: 23.610 min Scan# 21

Delta R.T. -0.015 min

Lab File: BN038071.D


Acq: 13 Oct 2025 16:48

Instrument :



BNA_N

ClientSampleId :

PB170022BS



Tgt Ion:	252	Resp:	9245
Ion Ratio	Lower	Upper	
252	100		
253	33.6	20.6	30.8#
125	34.3	16.7	25.1#



#40

Dibenzo(a,h)anthracene


Concen: 0.353 ng

RT: 26.104 min Scan# 3649



Delta R.T. -0.015 min


Lab File: BN038071.D

Acq: 13 Oct 2025 16:48



Tgt Ion:	278	Resp:	9200
Ion Ratio	Lower	Upper	
278	100		
139	23.9	15.4	23.2#
279	33.9	20.8	31.2#





#41

Benzo(g,h,i)perylene

Concen: 0.429 ng

RT: 26.797 min Scan# 38

Delta R.T. -0.032 min

Lab File: BN038071.D


Acq: 13 Oct 2025 16:48

Instrument :

BNA_N

ClientSampleId :

PB170022BS





Tgt Ion:276 Resp: 11736

Ion Ratio Lower Upper

276 100

277 26.6 20.2 30.4

138 26.5 19.8 29.8



Report of Analysis

Client:	Tetra Tech NUS, Inc.		Date Collected:	10/06/25
Project:	NWIRP Bethpage 112G08005-WE13		Date Received:	10/06/25
Client Sample ID:	BP-VPB-184-GW-740-742MS		SDG No.:	Q3292
Lab Sample ID:	Q3298-06MS		Matrix:	Water
Analytical Method:	SW8270ESIM	Level : LOW	% Solid:	0
Sample Wt/Vol:	420 mL	Final Vol: 1000 uL	Test:	SVOC-SIMGroup1
Prep Method :	Prep Date: 10/08/25			

CAS Number	Parameter	Conc.	Qua.	DF	MDL	LOD	LOQ / CRQL	Units	Date Ana.	Prep BatchID
TARGETS										
123-91-1	1,4-Dioxane	1.60		1	0.16	0.48	0.48	ug/L	10/14/25 03:43	PB170022
SURROGATES										
7297-45-2	2-Methylnaphthalene-d10	0.26			30 - 150		64%	SPK: 0.4		
93951-69-0	Fluoranthene-d10	0.19			30 - 150		48%	SPK: 0.4		
4165-60-0	Nitrobenzene-d5	0.29			55 - 111		72%	SPK: 0.4		
321-60-8	2-Fluorobiphenyl	0.73	*		53 - 106		183%	SPK: 0.4		
1718-51-0	Terphenyl-d14	1.08	*		58 - 132		269%	SPK: 0.4		
INTERNAL STANDARDS										
		Area								
3855-82-1	1,4-Dichlorobenzene-d4	5770								
1146-65-2	Naphthalene-d8	15000								
15067-26-2	Acenaphthene-d10	3440								
1517-22-2	Phenanthrene-d10	10300								
1719-03-5	Chrysene-d12	3300								
1520-96-3	Perylene-d12	80.0								

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN101325\
 Data File : BN038086.D
 Acq On : 14 Oct 2025 03:43
 Operator : RC/JU
 Sample : Q3298-06MS
 Misc :
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 BP-VPB-184-GW-740-742MS

Quant Time: Oct 14 04:13:18 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN092325.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Sep 24 11:00:01 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.797	152	5765	0.400	ng	-0.03
7) Naphthalene-d8	10.594	136	14968	0.400	ng	#-0.03
13) Acenaphthene-d10	14.452	164	3442	0.400	ng	-0.02
19) Phenanthrene-d10	17.198	188	10343	0.400	ng	#-0.02
29) Chrysene-d12	21.384	240	3302	0.400	ng	#-0.02
35) Perylene-d12	23.706	264	80	0.400	ng	#-0.02
System Monitoring Compounds						
4) 2-Fluorophenol	5.370	112	2575	0.196	ng	-0.01
5) Phenol-d6	6.952	99	2419	0.140	ng	-0.02
8) Nitrobenzene-d5	8.950	82	3312	0.290	ng	-0.02
11) 2-Methylnaphthalene-d10	12.192	152	5360	0.258	ng	-0.03
14) 2,4,6-Tribromophenol	15.945	330	742	0.569	ng	-0.02
15) 2-Fluorobiphenyl	13.073	172	10321	0.732	ng	-0.02
27) Fluoranthene-d10	19.234	212	4978	0.191	ng	-0.02
31) Terphenyl-d14	19.833	244	7090	1.078	ng	-0.02
Target Compounds					Qvalue	
2) 1,4-Dioxane	3.283	88	3930	0.672	ng	# 60
3) n-Nitrosodimethylamine	3.586	42	948	0.122	ng	# 69
6) bis(2-Chloroethyl)ether	7.212	93	4808	0.300	ng	# 88
9) Naphthalene	10.648	128	12172	0.295	ng	100
10) Hexachlorobutadiene	10.947	225	2034	0.289	ng	# 99
12) 2-Methylnaphthalene	12.268	142	6809	0.253	ng	98
16) Acenaphthylene	14.174	152	3188	0.204	ng	99
17) Acenaphthene	14.516	154	3893	0.350	ng	98
18) Fluorene	15.499	166	9372	0.696	ng	# 79
20) 4,6-Dinitro-2-methylph...	15.572	198	496	0.438	ng	# 1
21) 4-Bromophenyl-phenylether	16.391	248	2435	0.361	ng	# 89
22) Hexachlorobenzene	16.515	284	3120	0.382	ng	97
24) Pentachlorophenol	16.851	266	2618	0.934	ng	91
25) Phenanthrene	17.235	178	12213	0.359	ng	100
26) Anthracene	17.335	178	5577	0.189	ng	100
28) Fluoranthene	19.266	202	8095	0.216	ng	98
30) Pyrene	19.629	202	2468	0.189	ng	98
32) Benzo(a)anthracene	21.375	228	3656	0.317	ng	97
33) Chrysene	21.420	228	6744	0.540	ng	97
34) Bis(2-ethylhexyl)phtha...	21.304	149	3033	0.664	ng	99
36) Indeno(1,2,3-cd)pyrene	26.083	276	1016	2.779	ng	# 76
37) Benzo(b)fluoranthene	23.005	252	4538	13.154	ng	# 50
38) Benzo(k)fluoranthene	23.051	252	1889	5.442	ng	# 1
39) Benzo(a)pyrene	23.607	252	604	2.214	ng	# 1
40) Dibenzo(a,h)anthracene	26.098	278	2154	7.546	ng	# 49
41) Benzo(g,h,i)perylene	26.797	276	809	2.698	ng	# 34


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN101325\
Data File : BN038086.D
Acq On : 14 Oct 2025 03:43
Operator : RC/JU
Sample : Q3298-06MS
Misc :
ALS Vial : 21 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
BP-VPB-184-GW-740-742MS

Quant Time: Oct 14 04:13:18 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN092325.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Sep 24 11:00:01 2025
Response via : Initial Calibration









#1
1,4-Dichlorobenzene-d4
Concen: 0.400 ng
RT: 7.797 min Scan# 61
Delta R.T. -0.028 min
Lab File: BN038086.D
Acq: 14 Oct 2025 03:43




Instrument :
BNA_N
ClientSampleId :
BP-VPB-184-GW-740-742MS


Tgt Ion:152 Resp: 5765
Ion Ratio Lower Upper
152 100
150 151.0 121.0 181.4
115 61.0 47.4 71.0



#2
1,4-Dioxane
Concen: 0.672 ng
RT: 3.283 min Scan# 28
Delta R.T. -0.014 min
Lab File: BN038086.D
Acq: 14 Oct 2025 03:43

Tgt Ion: 88 Resp: 3930
Ion Ratio Lower Upper
88 100
43 101.2 26.6 39.8#
58 76.4 58.9 88.3









#3
n-Nitrosodimethylamine
Concen: 0.122 ng
RT: 3.586 min Scan# 70
Delta R.T. -0.014 min
Lab File: BN038086.D
Acq: 14 Oct 2025 03:43




Instrument :
BNA_N
ClientSampleId :
BP-VPB-184-GW-740-742MS


Tgt Ion: 42 Resp: 948
Ion Ratio Lower Upper
42 100
74 88.6 93.4 140.0#
44 0.0 22.2 33.4#



#4
2-Fluorophenol
Concen: 0.196 ng
RT: 5.370 min Scan# 317
Delta R.T. -0.014 min
Lab File: BN038086.D
Acq: 14 Oct 2025 03:43

Tgt Ion: 112 Resp: 2575
Ion Ratio Lower Upper
112 100
64 61.5 44.6 66.8
63 32.6 24.9 37.3









#5
Phenol-d6
Concen: 0.140 ng
RT: 6.952 min Scan# 531
Delta R.T. -0.022 min
Lab File: BN038086.D
Acq: 14 Oct 2025 03:43




Instrument :
BNA_N
ClientSampleId :
BP-VPB-184-GW-740-742MS


Tgt Ion: 99 Resp: 2419
Ion Ratio Lower Upper
99 100
42 66.6 17.2 25.8#
71 183.7 27.3 40.9#



#6
bis(2-Chloroethyl)ether
Concen: 0.300 ng
RT: 7.212 min Scan# 572
Delta R.T. -0.029 min
Lab File: BN038086.D
Acq: 14 Oct 2025 03:43

Tgt Ion: 93 Resp: 4808
Ion Ratio Lower Upper
93 100
63 79.2 60.1 90.1
95 48.3 25.4 38.2#









#7
Naphthalene-d8
Concen: 0.400 ng
RT: 10.594 min Scan# 941
Delta R.T. -0.032 min
Lab File: BN038086.D
Acq: 14 Oct 2025 03:43




Instrument :
BNA_N
ClientSampleId :
BP-VPB-184-GW-740-742MS


Tgt Ion:	136	Resp:	14968
Ion	Ratio	Lower	Upper
136	100		
137	12.4	9.0	13.4
54	9.6	5.8	8.8#
68	7.2	4.3	6.5#



#8
Nitrobenzene-d5
Concen: 0.290 ng
RT: 8.950 min Scan# 784
Delta R.T. -0.021 min
Lab File: BN038086.D
Acq: 14 Oct 2025 03:43

Tgt Ion:	82	Resp:	3312
Ion	Ratio	Lower	Upper
82	100		
128	47.0	34.8	52.2
54	51.3	42.2	63.2









#9
Naphthalene
Concen: 0.295 ng
RT: 10.648 min Scan# 945
Delta R.T. -0.021 min
Lab File: BN038086.D
Acq: 14 Oct 2025 03:43




Instrument :
BNA_N
ClientSampleId :
BP-VPB-184-GW-740-742MS


Tgt Ion:128 Resp: 12172
Ion Ratio Lower Upper
128 100
129 11.7 9.2 13.8
127 13.6 11.0 16.6



#10
Hexachlorobutadiene
Concen: 0.289 ng
RT: 10.947 min Scan# 971
Delta R.T. -0.032 min
Lab File: BN038086.D
Acq: 14 Oct 2025 03:43

Tgt Ion:225 Resp: 2034
Ion Ratio Lower Upper
225 100
223 0.0 0.0 0.0
227 64.9 51.4 77.2









#11
2-Methylnaphthalene-d10
Concen: 0.258 ng
RT: 12.192 min Scan# 1163
Delta R.T. -0.025 min
Lab File: BN038086.D
Acq: 14 Oct 2025 03:43




Instrument :
BNA_N
ClientSampleId :
BP-VPB-184-GW-740-742MS


Tgt Ion:152 Resp: 5360
Ion Ratio Lower Upper
152 100
151 27.8 17.4 26.2#



#12
2-Methylnaphthalene
Concen: 0.253 ng
RT: 12.268 min Scan# 1173
Delta R.T. -0.025 min
Lab File: BN038086.D
Acq: 14 Oct 2025 03:43

Tgt Ion:142 Resp: 6809
Ion Ratio Lower Upper
142 100
141 87.1 71.2 106.8
115 38.7 29.7 44.5









#13
Acenaphthene-d10
Concen: 0.400 ng
RT: 14.452 min Scan# 14
Delta R.T. -0.021 min
Lab File: BN038086.D
Acq: 14 Oct 2025 03:43




Instrument :
BNA_N
ClientSampleId :
BP-VPB-184-GW-740-742MS


Tgt Ion:164 Resp: 3442
Ion Ratio Lower Upper
164 100
162 110.8 84.8 127.2
160 63.2 46.1 69.1



#14
2,4,6-Tribromophenol
Concen: 0.569 ng
RT: 15.945 min Scan# 1588
Delta R.T. -0.024 min
Lab File: BN038086.D
Acq: 14 Oct 2025 03:43

Tgt Ion:330 Resp: 742
Ion Ratio Lower Upper
330 100
332 96.1 77.2 115.8
141 49.3 37.2 55.8









#15
2-Fluorobiphenyl
Concen: 0.732 ng
RT: 13.073 min Scan# 1327
Delta R.T. -0.021 min
Lab File: BN038086.D
Acq: 14 Oct 2025 03:43




Instrument :
BNA_N
ClientSampleId :
BP-VPB-184-GW-740-742MS


Tgt Ion:	172	Resp:	10321
Ion Ratio	Lower	Upper	
172	100		
171	34.8	28.6	43.0
170	23.9	19.7	29.5



#16
Acenaphthylene
Concen: 0.204 ng
RT: 14.174 min Scan# 1428
Delta R.T. -0.021 min
Lab File: BN038086.D
Acq: 14 Oct 2025 03:43


Tgt Ion:	152	Resp:	3188
Ion Ratio	Lower	Upper	
152	100		
151	20.0	15.8	23.8
153	13.8	10.5	15.7








#17
Acenaphthene
Concen: 0.350 ng
RT: 14.516 min Scan# 1460
Delta R.T. -0.021 min
Lab File: BN038086.D
Acq: 14 Oct 2025 03:43

Instrument :
BNA_N
ClientSampleId :
BP-VPB-184-GW-740-742MS







Tgt Ion:154 Resp: 3893
Ion Ratio Lower Upper
154 100
153 112.7 90.5 135.7
152 60.8 51.8 77.8



#18
Fluorene
Concen: 0.696 ng
RT: 15.499 min Scan# 1552
Delta R.T. -0.023 min
Lab File: BN038086.D
Acq: 14 Oct 2025 03:43

Tgt Ion:166 Resp: 9372
Ion Ratio Lower Upper
166 100
165 75.2 79.0 118.4#
167 14.1 10.6 16.0









#19
Phenanthrene-d10
Concen: 0.400 ng
RT: 17.198 min Scan# 1689
Delta R.T. -0.025 min
Lab File: BN038086.D
Acq: 14 Oct 2025 03:43




Instrument :
BNA_N
ClientSampleId :
BP-VPB-184-GW-740-742MS


Tgt Ion:	188	Resp:	10343
Ion Ratio	Lower	Upper	
188	100		
94	0.0	0.0	0.0
80	13.7	7.6	11.4#



#20
4,6-Dinitro-2-methylphenol
Concen: 0.438 ng
RT: 15.572 min Scan# 1558
Delta R.T. -0.025 min
Lab File: BN038086.D
Acq: 14 Oct 2025 03:43

Tgt Ion:	198	Resp:	496
Ion Ratio	Lower	Upper	
198	100		
51	213.1	59.1	88.7#
105	107.3	41.3	61.9#









#21
4-Bromophenyl-phenylether
Concen: 0.361 ng
RT: 16.391 min Scan# 1624
Delta R.T. -0.025 min
Lab File: BN038086.D
Acq: 14 Oct 2025 03:43




Instrument :
BNA_N
ClientSampleId :
BP-VPB-184-GW-740-742MS


Tgt Ion	Ratio	Lower	Upper
248	100		
250	95.0	77.6	116.4
141	79.4	48.8	73.2#



#22
Hexachlorobenzene
Concen: 0.382 ng
RT: 16.515 min Scan# 1634
Delta R.T. -0.025 min
Lab File: BN038086.D
Acq: 14 Oct 2025 03:43

Tgt Ion	Ratio	Lower	Upper
284	100		
142	40.7	30.9	46.3
249	30.6	23.9	35.9









#24
Pentachlorophenol
Concen: 0.934 ng
RT: 16.851 min Scan# 1661
Delta R.T. -0.024 min
Lab File: BN038086.D
Acq: 14 Oct 2025 03:43




Instrument :
BNA_N
ClientSampleId :
BP-VPB-184-GW-740-742MS


Tgt Ion	266	Ratio	100	Lower	Upper
266	100				
264	59.7		50.9	76.3	
268	58.0		54.3	81.5	



#25
Phenanthrene
Concen: 0.359 ng
RT: 17.235 min Scan# 1692
Delta R.T. -0.025 min
Lab File: BN038086.D
Acq: 14 Oct 2025 03:43

Tgt Ion	178	Ratio	100	Lower	Upper
178	100				
176	19.4		15.7	23.5	
179	15.6		12.3	18.5	









#26
Anthracene
Concen: 0.189 ng
RT: 17.335 min Scan# 1700
Delta R.T. -0.025 min
Lab File: BN038086.D
Acq: 14 Oct 2025 03:43




Instrument :
BNA_N
ClientSampleId :
BP-VPB-184-GW-740-742MS


Tgt Ion:178 Resp: 5577
Ion Ratio Lower Upper
178 100
176 18.6 15.0 22.4
179 15.2 12.3 18.5



#27
Fluoranthene-d10
Concen: 0.191 ng
RT: 19.234 min Scan# 1950
Delta R.T. -0.019 min
Lab File: BN038086.D
Acq: 14 Oct 2025 03:43

Tgt Ion:212 Resp: 4978
Ion Ratio Lower Upper
212 100
106 16.4 12.6 19.0
104 10.1 7.4 11.0









#28
Fluoranthene
Concen: 0.216 ng
RT: 19.266 min Scan# 1957
Delta R.T. -0.019 min
Lab File: BN038086.D
Acq: 14 Oct 2025 03:43




Instrument :
BNA_N
ClientSampleId :
BP-VPB-184-GW-740-742MS


Tgt Ion:202 Resp: 8095
Ion Ratio Lower Upper
202 100
101 12.8 9.3 13.9
203 16.7 13.6 20.4



#29
Chrysene-d12
Concen: 0.400 ng
RT: 21.384 min Scan# 2270
Delta R.T. -0.018 min
Lab File: BN038086.D
Acq: 14 Oct 2025 03:43

Tgt Ion:240 Resp: 3302
Ion Ratio Lower Upper
240 100
120 34.5 11.4 17.2#
236 39.0 23.0 34.6#









#30
Pyrene
Concen: 0.189 ng
RT: 19.629 min Scan# 2035
Delta R.T. -0.019 min
Lab File: BN038086.D
Acq: 14 Oct 2025 03:43




Instrument :
BNA_N
ClientSampleId :
BP-VPB-184-GW-740-742MS


Tgt Ion:	202	Resp:	2468
Ion	Ratio	Lower	Upper
202	100		
200	21.8	16.7	25.1
203	18.9	14.2	21.4



#31
Terphenyl-d14
Concen: 1.078 ng
RT: 19.833 min Scan# 2079
Delta R.T. -0.019 min
Lab File: BN038086.D
Acq: 14 Oct 2025 03:43

Tgt Ion:	244	Resp:	7090
Ion	Ratio	Lower	Upper
244	100		
212	10.2	7.6	11.4
122	19.2	13.9	20.9









#32
Benzo(a)anthracene
Concen: 0.317 ng
RT: 21.375 min Scan# 21
Delta R.T. -0.009 min
Lab File: BN038086.D
Acq: 14 Oct 2025 03:43




Instrument :
BNA_N
ClientSampleId :
BP-VPB-184-GW-740-742MS


Tgt Ion	228	226	229
Resp:	3656		
Ratio	100	28.9	22.2
Lower		22.6	15.6
Upper		33.8	23.4



#33
Chrysene
Concen: 0.540 ng
RT: 21.420 min Scan# 2274
Delta R.T. -0.018 min
Lab File: BN038086.D
Acq: 14 Oct 2025 03:43

Tgt Ion	228	226	229
Resp:	6744		
Ratio	100	32.6	20.8
Lower		24.6	15.6
Upper		37.0	23.4









#34
Bis(2-ethylhexyl)phthalate
Concen: 0.664 ng
RT: 21.304 min Scan# 21
Delta R.T. -0.018 min
Lab File: BN038086.D
Acq: 14 Oct 2025 03:43




Instrument :
BNA_N
ClientSampleId :
BP-VPB-184-GW-740-742MS


Tgt Ion:149 Resp: 3033
Ion Ratio Lower Upper
149 100
167 25.8 20.4 30.6
279 3.4 2.4 3.6



#35
Perylene-d12
Concen: 0.400 ng
RT: 23.706 min Scan# 2829
Delta R.T. -0.020 min
Lab File: BN038086.D
Acq: 14 Oct 2025 03:43

Tgt Ion:264 Resp: 80
Ion Ratio Lower Upper
264 100
260 58.2 21.5 32.3#
265 3161.9 53.8 80.6#





#36

Indeno(1,2,3-cd)pyrene

Concen: 2.779 ng

RT: 26.083 min Scan# 30

Delta R.T. -0.023 min

Lab File: BN038086.D


Acq: 14 Oct 2025 03:43

Instrument :

BNA_N

ClientSampleId :

BP-VPB-184-GW-740-742MS






Tgt Ion:276 Resp: 1016

Ion Ratio Lower Upper

276 100

138 17.3 21.4 32.0#

277 10.6 20.3 30.5#



#37

Benzo(b)fluoranthene

Concen: 13.154 ng

RT: 23.005 min Scan# 2589

Delta R.T. -0.020 min

Lab File: BN038086.D

Acq: 14 Oct 2025 03:43




Tgt Ion:252 Resp: 4538


Ion Ratio Lower Upper

252 100

253 42.8 19.4 29.0#

125 45.3 13.0 19.4#





#38

Benzo(k)fluoranthene

Concen: 5.442 ng

RT: 23.051 min Scan# 20

Delta R.T. -0.021 min

Lab File: BN038086.D

Acq: 14 Oct 2025 03:43

Instrument :

BNA_N

ClientSampleId :

BP-VPB-184-GW-740-742MS





Tgt Ion:252 Resp: 1889

Ion Ratio Lower Upper

252 100

253 63.3 19.5 29.3#

125 73.9 13.0 19.4#



#39

Benzo(a)pyrene

Concen: 2.214 ng

RT: 23.607 min Scan# 2795

Delta R.T. -0.018 min

Lab File: BN038086.D

Acq: 14 Oct 2025 03:43




Tgt Ion:252 Resp: 604


Ion Ratio Lower Upper

252 100

253 99.4 20.6 30.8#

125 157.6 16.7 25.1#









#40
Dibenzo(a,h)anthracene
Concen: 7.546 ng
RT: 26.098 min Scan# 30
Delta R.T. -0.020 min
Lab File: BN038086.D
Acq: 14 Oct 2025 03:43




Instrument :
BNA_N
ClientSampleId :
BP-VPB-184-GW-740-742MS

Tgt Ion:278 Resp: 2154
Ion Ratio Lower Upper
278 100
139 35.7 15.4 23.2#
279 58.0 20.8 31.2#



#41
Benzo(g,h,i)perylene
Concen: 2.698 ng
RT: 26.797 min Scan# 3886
Delta R.T. -0.032 min
Lab File: BN038086.D
Acq: 14 Oct 2025 03:43

Tgt Ion:276 Resp: 809
Ion Ratio Lower Upper
276 100
277 54.0 20.2 30.4#
138 62.0 19.8 29.8#





284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Report of Analysis

Client:	Tetra Tech NUS, Inc.		Date Collected:	10/06/25
Project:	NWIRP Bethpage 112G08005-WE13		Date Received:	10/06/25
Client Sample ID:	BP-VPB-184-GW-740-742MSD		SDG No.:	Q3292
Lab Sample ID:	Q3298-07MSD		Matrix:	Water
Analytical Method:	SW8270ESIM	Level : LOW	% Solid:	0
Sample Wt/Vol:	360 mL	Final Vol: 1000 uL	Test:	SVOC-SIMGroup1
Prep Method :	Prep Date: 10/08/25			

CAS Number	Parameter	Conc.	Qua.	DF	MDL	LOD	LOQ / CRQL	Units	Date Ana.	Prep BatchID
TARGETS										
123-91-1	1,4-Dioxane	1.90		1	0.18	0.56	0.56	ug/L	10/14/25 04:19	PB170022
SURROGATES										
7297-45-2	2-Methylnaphthalene-d10	0.26			30 - 150		65%	SPK: 0.4		
93951-69-0	Fluoranthene-d10	0.14			30 - 150		36%	SPK: 0.4		
4165-60-0	Nitrobenzene-d5	0.32			55 - 111		79%	SPK: 0.4		
321-60-8	2-Fluorobiphenyl	0.87	*		53 - 106		218%	SPK: 0.4		
1718-51-0	Terphenyl-d14	2.21	*		58 - 132		553%	SPK: 0.4		
INTERNAL STANDARDS										
		Area								
3855-82-1	1,4-Dichlorobenzene-d4	6050								
1146-65-2	Naphthalene-d8	15900								
15067-26-2	Acenaphthene-d10	3280								
1517-22-2	Phenanthrene-d10	9260								
1719-03-5	Chrysene-d12	1400								
1520-96-3	Perylene-d12	59.0								

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN101325\
 Data File : BN038087.D
 Acq On : 14 Oct 2025 04:19
 Operator : RC/JU
 Sample : Q3298-07MSD
 Misc :
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 BP-VPB-184-GW-740-742MSD

Quant Time: Oct 14 04:44:06 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN092325.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Sep 24 11:00:01 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.796	152	6052	0.400	ng	-0.03
7) Naphthalene-d8	10.594	136	15943	0.400	ng	#-0.03
13) Acenaphthene-d10	14.452	164	3279	0.400	ng	-0.02
19) Phenanthrene-d10	17.198	188	9259	0.400	ng	#-0.02
29) Chrysene-d12	21.384	240	1395	0.400	ng	#-0.02
35) Perylene-d12	23.706	264	59	0.400	ng	#-0.02
System Monitoring Compounds						
4) 2-Fluorophenol	5.370	112	2791	0.202	ng	-0.01
5) Phenol-d6	6.951	99	1922	0.106	ng	-0.02
8) Nitrobenzene-d5	8.950	82	3825	0.315	ng	-0.02
11) 2-Methylnaphthalene-d10	12.192	152	5786	0.261	ng	-0.03
14) 2,4,6-Tribromophenol	15.944	330	792	0.637	ng	-0.02
15) 2-Fluorobiphenyl	13.072	172	11734	0.874	ng	-0.02
27) Fluoranthene-d10	19.238	212	3375	0.145	ng	-0.01
31) Terphenyl-d14	19.833	244	6153	2.214	ng	-0.02
Target Compounds						Qvalue
2) 1,4-Dioxane	3.283	88	4280	0.697	ng	# 76
3) n-Nitrosodimethylamine	3.586	42	1205	0.147	ng	# 74
6) bis(2-Chloroethyl)ether	7.211	93	5399	0.321	ng	# 87
9) Naphthalene	10.648	128	13896	0.316	ng	99
10) Hexachlorobutadiene	10.946	225	2302	0.307	ng	# 100
12) 2-Methylnaphthalene	12.268	142	7718	0.269	ng	99
16) Acenaphthylene	14.174	152	3670	0.247	ng	100
17) Acenaphthene	14.516	154	4085	0.385	ng	93
18) Fluorene	15.499	166	12500	0.974	ng	# 62
20) 4,6-Dinitro-2-methylph...	15.572	198	558	0.527	ng	# 1
21) 4-Bromophenyl-phenylether	16.391	248	2615	0.433	ng	# 87
22) Hexachlorobenzene	16.515	284	3494	0.478	ng	98
24) Pentachlorophenol	16.851	266	2534	1.009	ng	93
25) Phenanthrene	17.235	178	12812	0.420	ng	99
26) Anthracene	17.335	178	5302	0.201	ng	98
28) Fluoranthene	19.266	202	6037	0.180	ng	99
30) Pyrene	19.629	202	3813	0.692	ng	98
32) Benzo(a)anthracene	21.375	228	2019	0.414	ng	94
33) Chrysene	21.420	228	4643	0.881	ng	97
34) Bis(2-ethylhexyl)phtha...	21.304	149	1679	0.871	ng	# 97
36) Indeno(1,2,3-cd)pyrene	26.086	276	656	2.433	ng	# 64
37) Benzo(b)fluoranthene	23.005	252	2316	9.102	ng	# 8
38) Benzo(k)fluoranthene	23.051	252	809	3.160	ng	# 1
39) Benzo(a)pyrene	23.607	252	800	3.976	ng	# 1
40) Dibenzo(a,h)anthracene	26.098	278	978	4.646	ng	# 1
41) Benzo(g,h,i)perylene	26.799	276	1194	5.399	ng	# 57


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN101325\
Data File : BN038087.D
Acq On : 14 Oct 2025 04:19
Operator : RC/JU
Sample : Q3298-07MSD
Misc :
ALS Vial : 22 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
BP-VPB-184-GW-740-742MSD

Quant Time: Oct 14 04:44:06 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN092325.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Sep 24 11:00:01 2025
Response via : Initial Calibration









#1
1,4-Dichlorobenzene-d4
Concen: 0.400 ng
RT: 7.796 min Scan# 61
Delta R.T. -0.029 min
Lab File: BN038087.D
Acq: 14 Oct 2025 04:19




Instrument :
BNA_N
ClientSampleId :
BP-VPB-184-GW-740-742MSD


Tgt Ion:152 Resp: 6052
Ion Ratio Lower Upper
152 100
150 156.4 121.0 181.4
115 62.2 47.4 71.0



#2
1,4-Dioxane
Concen: 0.697 ng
RT: 3.283 min Scan# 28
Delta R.T. -0.015 min
Lab File: BN038087.D
Acq: 14 Oct 2025 04:19

Tgt Ion: 88 Resp: 4280
Ion Ratio Lower Upper
88 100
43 0.0 26.6 39.8#
58 80.7 58.9 88.3









#3
n-Nitrosodimethylamine
Concen: 0.147 ng
RT: 3.586 min Scan# 70
Delta R.T. -0.015 min
Lab File: BN038087.D
Acq: 14 Oct 2025 04:19




Instrument :
BNA_N
ClientSampleId :
BP-VPB-184-GW-740-742MSD


Tgt Ion	Ratio	Lower	Upper
42	100		
74	95.4	93.4	140.0
44	0.0	22.2	33.4



#4
2-Fluorophenol
Concen: 0.202 ng
RT: 5.370 min Scan# 317
Delta R.T. -0.015 min
Lab File: BN038087.D
Acq: 14 Oct 2025 04:19

Tgt Ion	Ratio	Lower	Upper
112	100		
64	65.6	44.6	66.8
63	35.7	24.9	37.3









#5
Phenol-d6
Concen: 0.106 ng
RT: 6.951 min Scan# 531
Delta R.T. -0.022 min
Lab File: BN038087.D
Acq: 14 Oct 2025 04:19




Instrument :
BNA_N
ClientSampleId :
BP-VPB-184-GW-740-742MSD


Tgt Ion: 99 Resp: 1922
Ion Ratio Lower Upper
99 100
42 43.3 17.2 25.8#
71 146.6 27.3 40.9#



#6
bis(2-Chloroethyl)ether
Concen: 0.321 ng
RT: 7.211 min Scan# 572
Delta R.T. -0.029 min
Lab File: BN038087.D
Acq: 14 Oct 2025 04:19

Tgt Ion: 93 Resp: 5399
Ion Ratio Lower Upper
93 100
63 80.0 60.1 90.1
95 49.1 25.4 38.2#









#7
Naphthalene-d8
Concen: 0.400 ng
RT: 10.594 min Scan# 941
Delta R.T. -0.032 min
Lab File: BN038087.D
Acq: 14 Oct 2025 04:19




Instrument :
BNA_N
ClientSampleId :
BP-VPB-184-GW-740-742MSD


Tgt Ion	136	137	54	68
Resp:	15943	12.5	9.4	7.2
Lower		9.0	5.8	4.3
Upper		13.4	8.8#	6.5#



#8
Nitrobenzene-d5
Concen: 0.315 ng
RT: 8.950 min Scan# 784
Delta R.T. -0.021 min
Lab File: BN038087.D
Acq: 14 Oct 2025 04:19

Tgt Ion	82	128	54
Resp:	3825	47.7	54.2
Lower		34.8	42.2
Upper		52.2	63.2









#9
Naphthalene
Concen: 0.316 ng
RT: 10.648 min Scan# 94
Delta R.T. -0.021 min
Lab File: BN038087.D
Acq: 14 Oct 2025 04:19




Instrument :
BNA_N
ClientSampleId :
BP-VPB-184-GW-740-742MSD


Tgt Ion:128 Resp: 13896
Ion Ratio Lower Upper
128 100
129 11.7 9.2 13.8
127 13.5 11.0 16.6



#10
Hexachlorobutadiene
Concen: 0.307 ng
RT: 10.946 min Scan# 971
Delta R.T. -0.032 min
Lab File: BN038087.D
Acq: 14 Oct 2025 04:19

Tgt Ion:225 Resp: 2302
Ion Ratio Lower Upper
225 100
223 0.0 0.0 0.0
227 64.3 51.4 77.2









#11
2-Methylnaphthalene-d10
Concen: 0.261 ng
RT: 12.192 min Scan# 1163
Delta R.T. -0.026 min
Lab File: BN038087.D
Acq: 14 Oct 2025 04:19




Instrument :
BNA_N
ClientSampleId :
BP-VPB-184-GW-740-742MSD


Tgt Ion:152 Resp: 5786
Ion Ratio Lower Upper
152 100
151 34.5 17.4 26.2#



#12
2-Methylnaphthalene
Concen: 0.269 ng
RT: 12.268 min Scan# 1173
Delta R.T. -0.025 min
Lab File: BN038087.D
Acq: 14 Oct 2025 04:19

Tgt Ion:142 Resp: 7718
Ion Ratio Lower Upper
142 100
141 88.7 71.2 106.8
115 38.5 29.7 44.5









#13
Acenaphthene-d10
Concen: 0.400 ng
RT: 14.452 min Scan# 1456
Delta R.T. -0.022 min
Lab File: BN038087.D
Acq: 14 Oct 2025 04:19




Instrument :
BNA_N
ClientSampleId :
BP-VPB-184-GW-740-742MSD


Tgt Ion:164 Resp: 3279
Ion Ratio Lower Upper
164 100
162 111.9 84.8 127.2
160 62.7 46.1 69.1



#14
2,4,6-Tribromophenol
Concen: 0.637 ng
RT: 15.944 min Scan# 1588
Delta R.T. -0.025 min
Lab File: BN038087.D
Acq: 14 Oct 2025 04:19

Tgt Ion:330 Resp: 792
Ion Ratio Lower Upper
330 100
332 99.4 77.2 115.8
141 54.7 37.2 55.8









#15
2-Fluorobiphenyl
Concen: 0.874 ng
RT: 13.072 min Scan# 1327
Delta R.T. -0.022 min
Lab File: BN038087.D
Acq: 14 Oct 2025 04:19




Instrument :
BNA_N
ClientSampleId :
BP-VPB-184-GW-740-742MSD


Tgt Ion:	172	Resp:	11734
Ion	Ratio	Lower	Upper
172	100		
171	35.6	28.6	43.0
170	24.3	19.7	29.5



#16
Acenaphthylene
Concen: 0.247 ng
RT: 14.174 min Scan# 1428
Delta R.T. -0.022 min
Lab File: BN038087.D
Acq: 14 Oct 2025 04:19

Tgt Ion:	152	Resp:	3670
Ion	Ratio	Lower	Upper
152	100		
151	19.7	15.8	23.8
153	12.8	10.5	15.7









#17
Acenaphthene
Concen: 0.385 ng
RT: 14.516 min Scan# 1460
Delta R.T. -0.022 min
Lab File: BN038087.D
Acq: 14 Oct 2025 04:19




Instrument :
BNA_N
ClientSampleId :
BP-VPB-184-GW-740-742MSD


Tgt Ion	Ratio	Lower	Upper
154	100		
153	106.6	90.5	135.7
152	58.1	51.8	77.8



#18
Fluorene
Concen: 0.974 ng
RT: 15.499 min Scan# 1552
Delta R.T. -0.023 min
Lab File: BN038087.D
Acq: 14 Oct 2025 04:19

Tgt Ion	Ratio	Lower	Upper
166	100		
165	56.0	79.0	118.4#
167	11.5	10.6	16.0





#19

Phenanthrene-d10

Concen: 0.400 ng

RT: 17.198 min Scan# 1

Delta R.T. -0.025 min

Lab File: BN038087.D

Acq: 14 Oct 2025 04:19

Instrument :

BNA_N

ClientSampleId :

BP-VPB-184-GW-740-742MSD





Tgt Ion:188 Resp: 9259

Ion Ratio Lower Upper

188 100

94 0.0 0.0 0.0

80 14.6 7.6 11.4#



#20

4,6-Dinitro-2-methylphenol

Concen: 0.527 ng

RT: 15.572 min Scan# 1558

Delta R.T. -0.025 min

Lab File: BN038087.D

Acq: 14 Oct 2025 04:19




Tgt Ion:198 Resp: 558


Ion Ratio Lower Upper

198 100

51 188.7 59.1 88.7#

105 92.5 41.3 61.9#









#21
4-Bromophenyl-phenylether
Concen: 0.433 ng
RT: 16.391 min Scan# 1624
Delta R.T. -0.025 min
Lab File: BN038087.D
Acq: 14 Oct 2025 04:19




Instrument :
BNA_N
ClientSampleId :
BP-VPB-184-GW-740-742MSD


Tgt Ion:248 Resp: 2615
Ion Ratio Lower Upper
248 100
250 96.5 77.6 116.4
141 86.3 48.8 73.2#



#22
Hexachlorobenzene
Concen: 0.478 ng
RT: 16.515 min Scan# 1634
Delta R.T. -0.025 min
Lab File: BN038087.D
Acq: 14 Oct 2025 04:19

Tgt Ion:284 Resp: 3494
Ion Ratio Lower Upper
284 100
142 40.5 30.9 46.3
249 30.0 23.9 35.9









#24
Pentachlorophenol
Concen: 1.009 ng
RT: 16.851 min Scan# 1661
Delta R.T. -0.024 min
Lab File: BN038087.D
Acq: 14 Oct 2025 04:19




Instrument :
BNA_N
ClientSampleId :
BP-VPB-184-GW-740-742MSD


Tgt Ion:266 Resp: 2534
Ion Ratio Lower Upper
266 100
264 60.8 50.9 76.3
268 59.0 54.3 81.5



#25
Phenanthrene
Concen: 0.420 ng
RT: 17.235 min Scan# 1692
Delta R.T. -0.025 min
Lab File: BN038087.D
Acq: 14 Oct 2025 04:19

Tgt Ion:178 Resp: 12812
Ion Ratio Lower Upper
178 100
176 19.8 15.7 23.5
179 15.7 12.3 18.5









#26
Anthracene
Concen: 0.201 ng
RT: 17.335 min Scan# 1700
Delta R.T. -0.025 min
Lab File: BN038087.D
Acq: 14 Oct 2025 04:19




Instrument :
BNA_N
ClientSampleId :
BP-VPB-184-GW-740-742MSD


Tgt Ion:178 Resp: 5302
Ion Ratio Lower Upper
178 100
176 20.2 15.0 22.4
179 16.0 12.3 18.5



#27
Fluoranthene-d10
Concen: 0.145 ng
RT: 19.238 min Scan# 1951
Delta R.T. -0.014 min
Lab File: BN038087.D
Acq: 14 Oct 2025 04:19

Tgt Ion:212 Resp: 3375
Ion Ratio Lower Upper
212 100
106 17.3 12.6 19.0
104 10.6 7.4 11.0









#28
Fluoranthene
Concen: 0.180 ng
RT: 19.266 min Scan# 1957
Delta R.T. -0.019 min
Lab File: BN038087.D
Acq: 14 Oct 2025 04:19




Instrument :
BNA_N
ClientSampleId :
BP-VPB-184-GW-740-742MSD


Tgt Ion:202 Resp: 6037
Ion Ratio Lower Upper
202 100
101 12.0 9.3 13.9
203 16.7 13.6 20.4



#29
Chrysene-d12
Concen: 0.400 ng
RT: 21.384 min Scan# 2270
Delta R.T. -0.018 min
Lab File: BN038087.D
Acq: 14 Oct 2025 04:19

Tgt Ion:240 Resp: 1395
Ion Ratio Lower Upper
240 100
120 56.8 11.4 17.2#
236 54.4 23.0 34.6#









#30
Pyrene
Concen: 0.692 ng
RT: 19.629 min Scan# 2035
Delta R.T. -0.019 min
Lab File: BN038087.D
Acq: 14 Oct 2025 04:19




Instrument :
BNA_N
ClientSampleId :
BP-VPB-184-GW-740-742MSD


Tgt Ion:	202	Resp:	3813
Ion	Ratio	Lower	Upper
202	100		
200	20.6	16.7	25.1
203	18.9	14.2	21.4



#31
Terphenyl-d14
Concen: 2.214 ng
RT: 19.833 min Scan# 2079
Delta R.T. -0.019 min
Lab File: BN038087.D
Acq: 14 Oct 2025 04:19

Tgt Ion:	244	Resp:	6153
Ion	Ratio	Lower	Upper
244	100		
212	10.5	7.6	11.4
122	20.5	13.9	20.9









#32
Benzo(a)anthracene
Concen: 0.414 ng
RT: 21.375 min Scan# 21
Delta R.T. -0.009 min
Lab File: BN038087.D
Acq: 14 Oct 2025 04:19




Instrument :
BNA_N
ClientSampleId :
BP-VPB-184-GW-740-742MSD


Tgt Ion	Ratio	Lower	Upper
228	100		
226	30.3	22.6	33.8
229	23.2	15.6	23.4



#33
Chrysene
Concen: 0.881 ng
RT: 21.420 min Scan# 2274
Delta R.T. -0.018 min
Lab File: BN038087.D
Acq: 14 Oct 2025 04:19

Tgt Ion	Ratio	Lower	Upper
228	100		
226	32.3	24.6	37.0
229	21.2	15.6	23.4









#34
Bis(2-ethylhexyl)phthalate
Concen: 0.871 ng
RT: 21.304 min Scan# 21
Delta R.T. -0.018 min
Lab File: BN038087.D
Acq: 14 Oct 2025 04:19




Instrument :
BNA_N
ClientSampleId :
BP-VPB-184-GW-740-742MSD


Tgt Ion:149 Resp: 1679
Ion Ratio Lower Upper
149 100
167 24.1 20.4 30.6
279 4.9 2.4 3.6#



#35
Perylene-d12
Concen: 0.400 ng
RT: 23.706 min Scan# 2829
Delta R.T. -0.021 min
Lab File: BN038087.D
Acq: 14 Oct 2025 04:19

Tgt Ion:264 Resp: 59
Ion Ratio Lower Upper
264 100
260 79.5 21.5 32.3#
265 3483.6 53.8 80.6#









#36
Indeno(1,2,3-cd)pyrene
Concen: 2.433 ng
RT: 26.086 min Scan# 3643
Delta R.T. -0.021 min
Lab File: BN038087.D
Acq: 14 Oct 2025 04:19




Instrument :
BNA_N
ClientSampleId :
BP-VPB-184-GW-740-742MSD


Tgt Ion:276 Resp: 656
Ion Ratio Lower Upper
276 100
138 3.7 21.4 32.0#
277 11.4 20.3 30.5#



#37
Benzo(b)fluoranthene
Concen: 9.102 ng
RT: 23.005 min Scan# 2589
Delta R.T. -0.021 min
Lab File: BN038087.D
Acq: 14 Oct 2025 04:19

Tgt Ion:252 Resp: 2316
Ion Ratio Lower Upper
252 100
253 57.6 19.4 29.0#
125 71.0 13.0 19.4#





#38

Benzo(k)fluoranthene

Concen: 3.160 ng

RT: 23.051 min Scan# 20

Delta R.T. -0.021 min

Lab File: BN038087.D

Acq: 14 Oct 2025 04:19

Instrument :

BNA_N

ClientSampleId :

BP-VPB-184-GW-740-742MSD





Tgt Ion:252 Resp: 809

Ion Ratio Lower Upper

252 100

253 78.4 19.5 29.3#

125 102.9 13.0 19.4#



#39

Benzo(a)pyrene

Concen: 3.976 ng

RT: 23.607 min Scan# 2795

Delta R.T. -0.018 min

Lab File: BN038087.D

Acq: 14 Oct 2025 04:19




Tgt Ion:252 Resp: 800


Ion Ratio Lower Upper

252 100

253 88.9 20.6 30.8#

125 142.0 16.7 25.1#





#40

Dibenzo(a,h)anthracene

Concen: 4.646 ng

RT: 26.098 min Scan# 3647

Delta R.T. -0.021 min

Lab File: BN038087.D


Acq: 14 Oct 2025 04:19

Instrument :

BNA_N




ClientSampleId :

BP-VPB-184-GW-740-742MSD



Tgt Ion:278 Resp: 978

Ion	Ratio	Lower	Upper
278	100		
139	51.9	15.4	23.2#
279	93.3	20.8	31.2#



#41

Benzo(g,h,i)perylene

Concen: 5.399 ng

RT: 26.799 min Scan# 3887




Delta R.T. -0.029 min

Lab File: BN038087.D

Acq: 14 Oct 2025 04:19

Tgt Ion:276 Resp: 1194

Ion	Ratio	Lower	Upper
276	100		
277	45.6	20.2	30.4#
138	48.1	19.8	29.8#





284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900, Fax : 908 789 8922

Manual Integration Report

Sequence:

BN092325

Instrument

BNA_n

Sample ID

File ID

Parameter

Review By

Review On

Supervised
By

Supervised On

Reason

Manual Integration Report

Sequence:

BN101325

Instrument

BNA_n

Sample ID	File ID	Parameter	Review By	Review On	Supervised By	Supervised On	Reason
SSTDCCC0.4	BN038066.D	Benzo(b)fluoranthene	Jagrut	10/14/2025 2:48:09 PM	mohammad	10/15/2025 4:43:39 AM	Peak Integrated by Software
SSTDCCC0.4	BN038066.D	Benzo(k)fluoranthene	Jagrut	10/14/2025 2:48:09 PM	mohammad	10/15/2025 4:43:39 AM	Peak Integrated by Software
SSTDCCC0.4	BN038083.D	Benzo(b)fluoranthene	Jagrut	10/14/2025 2:48:14 PM	mohammad	10/15/2025 4:43:39 AM	Peak Integrated by Software
SSTDCCC0.4	BN038088.D	Benzo(b)fluoranthene	Jagrut	10/14/2025 2:48:16 PM	mohammad	10/15/2025 4:43:39 AM	Peak Integrated by Software

Instrument ID: **BNA_N**

Daily Analysis Runlog For Sequence/QC Batch ID # BN092325

Review By	Rahul	Review On	9/24/2025 8:51:58 AM		
Supervise By	Jagrut	Supervise On	9/24/2025 9:44:29 AM		
SubDirectory	BN092325	HP Acquire Method	BNA_N	HP Processing Method	bn092325
STD. NAME		STD REF.#			
Tune/Reschk	SP6757				
Initial Calibration Stds	SP6842,SP6843,SP6844,SP6845,SP6846,SP6847,SP6848				
CCC	SP6846				
Internal Standard/PEM	SP6830,1ul/100ul sample				
ICV/I.BLK	SP6854				
Surrogate Standard					
MS/MSD Standard					
LCS Standard					

Sr#	SampleId	Data File Name	Date-Time	Operator	Status
1	DFTPP	BN037860.D	23 Sep 2025 11:33	RC/JU	Ok
2	SSTDICC0.1	BN037861.D	23 Sep 2025 12:13	RC/JU	Ok
3	SSTDICC0.2	BN037862.D	23 Sep 2025 12:49	RC/JU	Ok
4	SSTDICCC0.4	BN037863.D	23 Sep 2025 13:25	RC/JU	Ok
5	SSTDICC0.8	BN037864.D	23 Sep 2025 14:02	RC/JU	Ok
6	SSTDICC1.6	BN037865.D	23 Sep 2025 14:38	RC/JU	Ok
7	SSTDICC3.2	BN037866.D	23 Sep 2025 15:15	RC/JU	Ok
8	SSTDICC5.0	BN037867.D	23 Sep 2025 15:51	RC/JU	Ok
9	SSTDICV0.4	BN037868.D	23 Sep 2025 16:28	RC/JU	Ok
10	PB169743BL	BN037869.D	23 Sep 2025 17:42	RC/JU	Not Ok

M : Manual Integration

Instrument ID: **BNA_N**

Daily Analysis Runlog For Sequence/QC Batch ID # BN101325

Review By	Jagrut	Review On	10/14/2025 2:55:47 PM
Supervise By	mohammad	Supervise On	10/15/2025 4:43:39 AM
SubDirectory	BN101325	HP Acquire Method	BNA_N, 8270_SiM HP Processing Method BN092325
STD. NAME	STD REF.#		
Tune/Reschk	SP6875		
Initial Calibration Stds	SP6842,SP6843,SP6844,SP6845,SP6846,SP6847,SP6848		
CCC	SP6846		
Internal Standard/PEM	SP6830,10ul/1000ul sample		
ICV/I.BLK	SP6854		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

Sr#	SampleId	Data File Name	Date-Time	Operator	Status
1	DFTPP	BN038065.D	13 Oct 2025 13:08	RC/JU	Ok
2	SSTDCCC0.4	BN038066.D	13 Oct 2025 13:47	RC/JU	Ok,M
3	PB169992BL	BN038067.D	13 Oct 2025 14:23	RC/JU	Ok
4	PB169992BS	BN038068.D	13 Oct 2025 14:59	RC/JU	Ok
5	PB169992BSD	BN038069.D	13 Oct 2025 15:35	RC/JU	Ok
6	PB170022BL	BN038070.D	13 Oct 2025 16:12	RC/JU	Ok
7	PB170022BS	BN038071.D	13 Oct 2025 16:48	RC/JU	Ok
8	Q3292-01	BN038072.D	13 Oct 2025 17:24	RC/JU	Ok
9	Q3292-02	BN038073.D	13 Oct 2025 18:00	RC/JU	Ok
10	Q3298-02	BN038074.D	13 Oct 2025 18:36	RC/JU	Ok
11	Q3298-03	BN038075.D	13 Oct 2025 19:12	RC/JU	Ok
12	Q3324-03	BN038076.D	13 Oct 2025 19:48	RC/JU	Ok
13	Q3324-05	BN038077.D	13 Oct 2025 20:24	RC/JU	Ok
14	Q3324-06	BN038078.D	13 Oct 2025 21:00	RC/JU	Ok
15	PB170061BS	BN038079.D	13 Oct 2025 21:36	RC/JU	Ok
16	PB170061BSD	BN038080.D	13 Oct 2025 22:12	RC/JU	Ok,M
17	SSTDCCC0.4	BN038081.D	13 Oct 2025 23:24	RC/JU	Ok
18	DFTPP	BN038082.D	14 Oct 2025 00:40	RC/JU	Ok
19	SSTDCCC0.4	BN038083.D	14 Oct 2025 01:19	RC/JU	Ok,M
20	PB170061BL	BN038084.D	14 Oct 2025 02:31	RC/JU	Ok
21	Q3298-05	BN038085.D	14 Oct 2025 03:07	RC/JU	Ok

Instrument ID: **BNA_N**

Daily Analysis Runlog For Sequence/QC Batch ID # BN101325

Review By	Jagrut	Review On	10/14/2025 2:55:47 PM
Supervise By	mohammad	Supervise On	10/15/2025 4:43:39 AM
SubDirectory	BN101325	HP Acquire Method	BNA_N, 8270_SiM HP Processing Method BN092325
STD. NAME	STD REF.#		
Tune/Reschk	SP6875		
Initial Calibration Stds	SP6842,SP6843,SP6844,SP6845,SP6846,SP6847,SP6848		
CCC	SP6846		
Internal Standard/PEM	SP6830,10ul/1000ul sample		
ICV/I.BLK	SP6854		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

22	Q3298-06MS	BN038086.D	14 Oct 2025 03:43	RC/JU	Ok
23	Q3298-07MSD	BN038087.D	14 Oct 2025 04:19	RC/JU	Ok
24	SSTDCCC0.4	BN038088.D	14 Oct 2025 04:56	RC/JU	Ok,M

M : Manual Integration

Instrument ID: BNA_N

Daily Analysis Runlog For Sequence/QC Batch ID # BN092325

Review By	Rahul	Review On	9/24/2025 8:51:58 AM		
Supervise By	Jagrut	Supervise On	9/24/2025 9:44:29 AM		
SubDirectory	BN092325	HP Acquire Method	BNA_N	HP Processing Method	bn092325

STD. NAME	STD REF.#
Tune/Reschk	SP6757
Initial Calibration Stds	SP6842, SP6843, SP6844, SP6845, SP6846, SP6847, SP6848
CCC	SP6846
Internal Standard/PEM	SP6830, 1ul/100ul sample
ICV/I.BLK	SP6854
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	DFTPP	DFTPP	BN037860.D	23 Sep 2025 11:33		RC/JU	Ok
2	SSTDICC0.1	SSTDICC0.1	BN037861.D	23 Sep 2025 12:13		RC/JU	Ok
3	SSTDICC0.2	SSTDICC0.2	BN037862.D	23 Sep 2025 12:49		RC/JU	Ok
4	SSTDICCC0.4	SSTDICCC0.4	BN037863.D	23 Sep 2025 13:25	Compound #20 kept on QR	RC/JU	Ok
5	SSTDICC0.8	SSTDICC0.8	BN037864.D	23 Sep 2025 14:02		RC/JU	Ok
6	SSTDICC1.6	SSTDICC1.6	BN037865.D	23 Sep 2025 14:38		RC/JU	Ok
7	SSTDICC3.2	SSTDICC3.2	BN037866.D	23 Sep 2025 15:15		RC/JU	Ok
8	SSTDICC5.0	SSTDICC5.0	BN037867.D	23 Sep 2025 15:51	Compound#14,24 removed from 5.0 ppm	RC/JU	Ok
9	SSTDICV0.4	ICVBN092325	BN037868.D	23 Sep 2025 16:28		RC/JU	Ok
10	PB169743BL	PB169743BL	BN037869.D	23 Sep 2025 17:42	END CCC missing, Analyzed for Contamination check.	RC/JU	Not Ok

M : Manual Integration

Instrument ID: BNA_N

Daily Analysis Runlog For Sequence/QC Batch ID # BN101325

Review By	Jagrut	Review On	10/14/2025 2:55:47 PM
Supervise By	mohammad	Supervise On	10/15/2025 4:43:39 AM
SubDirectory	BN101325	HP Acquire Method	BNA_N, 8270_HP Processing Method BN092325

STD. NAME	STD REF.#
Tune/Reschk	SP6875
Initial Calibration Stds	SP6842,SP6843,SP6844,SP6845,SP6846,SP6847,SP6848
CCC	SP6846
Internal Standard/PEM	SP6830,10ul/1000ul sample
ICV/I.BLK	SP6854
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	DFTPP	DFTPP	BN038065.D	13 Oct 2025 13:08		RC/JU	Ok
2	SSTDCCC0.4	SSTDCCC0.4	BN038066.D	13 Oct 2025 13:47		RC/JU	Ok,M
3	PB169992BL	PB169992BL	BN038067.D	13 Oct 2025 14:23		RC/JU	Ok
4	PB169992BS	PB169992BS	BN038068.D	13 Oct 2025 14:59		RC/JU	Ok
5	PB169992BSD	PB169992BSD	BN038069.D	13 Oct 2025 15:35		RC/JU	Ok
6	PB170022BL	PB170022BL	BN038070.D	13 Oct 2025 16:12		RC/JU	Ok
7	PB170022BS	PB170022BS	BN038071.D	13 Oct 2025 16:48		RC/JU	Ok
8	Q3292-01	RW8-SP100-20251003	BN038072.D	13 Oct 2025 17:24		RC/JU	Ok
9	Q3292-02	RW8-SP303-20251003	BN038073.D	13 Oct 2025 18:00		RC/JU	Ok
10	Q3298-02	BP-VPB-184-EB-20251	BN038074.D	13 Oct 2025 18:36		RC/JU	Ok
11	Q3298-03	VPB184-HYD-2025100	BN038075.D	13 Oct 2025 19:12	Surrogate Fail	RC/JU	Ok
12	Q3324-03	BP-VPB-184-GW-EB-2	BN038076.D	13 Oct 2025 19:48		RC/JU	Ok
13	Q3324-05	BP-VPB-184-GW-780-7	BN038077.D	13 Oct 2025 20:24		RC/JU	Ok
14	Q3324-06	BP-VPB-184-GW-840-8	BN038078.D	13 Oct 2025 21:00		RC/JU	Ok
15	PB170061BS	PB170061BS	BN038079.D	13 Oct 2025 21:36		RC/JU	Ok
16	PB170061BSD	PB170061BSD	BN038080.D	13 Oct 2025 22:12		RC/JU	Ok,M
17	SSTDCCC0.4	SSTDCCC0.4EC	BN038081.D	13 Oct 2025 23:24		RC/JU	Ok
18	DFTPP	DFTPP	BN038082.D	14 Oct 2025 00:40		RC/JU	Ok

Instrument ID: BNA_N

Daily Analysis Runlog For Sequence/QC Batch ID # BN101325

Review By	Jagrut	Review On	10/14/2025 2:55:47 PM
Supervise By	mohammad	Supervise On	10/15/2025 4:43:39 AM
SubDirectory	BN101325	HP Acquire Method	BNA_N, 8270_HP Processing Method BN092325
STD. NAME	STD REF.#		
Tune/Reschk	SP6875		
Initial Calibration Stds	SP6842,SP6843,SP6844,SP6845,SP6846,SP6847,SP6848		
CCC	SP6846		
Internal Standard/PEM	SP6830,10ul/1000ul sample		
ICV/I.BLK	SP6854		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

19	SSTDCCC0.4	SSTDCCC0.4	BN038083.D	14 Oct 2025 01:19		RC/JU	Ok,M
20	PB170061BL	PB170061BL	BN038084.D	14 Oct 2025 02:31		RC/JU	Ok
21	Q3298-05	BP-VPB-184-GW-740-7	BN038085.D	14 Oct 2025 03:07		RC/JU	Ok
22	Q3298-06MS	BP-VPB-184-GW-740-7	BN038086.D	14 Oct 2025 03:43		RC/JU	Ok
23	Q3298-07MSD	BP-VPB-184-GW-740-7	BN038087.D	14 Oct 2025 04:19		RC/JU	Ok
24	SSTDCCC0.4	SSTDCCC0.4EC	BN038088.D	14 Oct 2025 04:56		RC/JU	Ok,M

M : Manual Integration

SOP ID: M3510C,3580A-Extraction SVOC-21

Clean Up SOP #: N/A

Matrix : Water

Weigh By: N/A **Extraction By:** RJ

Balance check: N/A **Filter By:** RJ

Balance ID: N/A **pH Meter ID:** N/A

pH Strip Lot#: E3953 **Hood ID:** 4,5,6,7

Extraction Method: ☒ Separatory Funnel ☐ Continuous Liquid/Liquid ☐ Sonication ☐ Waste Dilution ☐ Soxhlet

Extraction Start Date : 10/08/2025

Extraction Start Time : 08:50

Extraction End Date : 10/08/2025

Extraction End Time : 14:05

Concentration By: EH

Supervisor By : ritesh

Standard Name	MLS USED	Concentration ug/mL	STD REF. # FROM LOG
Spike Sol 1	1.0ML	0.4 PPM	SP6876
Surrogate	1.0ML	0.4 PPM	SP6873
N/A	N/A	N/A	N/A
N/A	N/A	N/A	N/A
N/A	N/A	N/A	N/A

Chemical Used	ML/SAMPLE USED	Lot Number
Methylene Chloride	N/A	E3973
Baked Na2SO4	N/A	EP2646
H2SO4 1:1	N/A	EP2610
10N NaoH	N/A	EP2609
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A

Extraction Conformance/Non-Conformance Comments:

PH Adjusted TO <2 with 1:1 H2SO4 & >11with 10N NAOH ,1.5ML Vial Lot #2210443.

KD Bath ID: WATER BATH-1,2 **Envap ID:** NEVAP-02


KD Bath Temperature: 60 °C **Envap Temperature:** 40 °C

Date / Time	Prepped Sample Relinquished By/Location	Received By/Location
10/08/25	RJ (EXT-106)	R/SVOC
14:10	Preparation Group	Analysis Group

Analytical Method: M3510C,3580A-Extraction SVOC-21

Concentration Date: 10/08/2025

Sample ID	Client Sample ID	Test	g / mL	PH	Surr/Spike By:		Final Vol. (mL)	JarID	Comments	Prep Pos
					AddedBy	VerifiedBy				
PB170022BL	SBLK022	SVOC-SIMGrou p1	1000	6	ritesh	Evelyn	1			SEP-1
PB170022BS	SLCS022	SVOC-SIMGrou p1	1000	6	ritesh	Evelyn	1			2
Q3292-01	RW8-SP100-20251003	SVOC-SIMGrou p1	1000	6	ritesh	Evelyn	1	D		3
Q3292-02	RW8-SP303-251003	SVOC-SIMGrou p1	1000	6	ritesh	Evelyn	1	D		4
Q3298-02	BP-VPB-EB-20251003	SVOC-SIMGrou p1	750	6	ritesh	Evelyn	1	C		5
Q3298-03	VPB184-HYD-20251003	SVOC-SIMGrou p1	1000	6	ritesh	Evelyn	1	C		6
Q3298-05	BP-VPB-GW-740-742	SVOC-SIMGrou p1	400	6	ritesh	Evelyn	1	C	Muddy	7
Q3298-06	Q3298-05MS	SVOC-SIMGrou p1	420	6	ritesh	Evelyn	1		Muddy	8
Q3298-07	Q3298-05MSD	SVOC-SIMGrou p1	360	6	ritesh	Evelyn	1		Muddy	9



130023
8:57

WORKLIST(Hardcopy Internal Chain)

WorkList Name : Q3292

WorkList ID : 192345

Department : Extraction

Date : 10-08-2025 08:45:50

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q3292-01	RW8-SP100-20251003	Water	SVOC-SIMGroup1	Cool 4 deg C	TETR06	D31	10/03/2025	8270-Modified
Q3292-02	RW8-SP303-251003	Water	SVOC-SIMGroup1	Cool 4 deg C	TETR06	D31	10/03/2025	8270-Modified
Q3298-02	BP-VPB-EB-20251003	Water	SVOC-SIMGroup1	Cool 4 deg C	TETR06	D31	10/03/2025	8270-Modified
Q3298-03	VPB184-HYD-20251003	Water	SVOC-SIMGroup1	Cool 4 deg C	TETR06	D31	10/03/2025	8270-Modified
Q3298-05	BP-VPB-GW-740-742	Water	SVOC-SIMGroup1	Cool 4 deg C	TETR06	D31	10/06/2025	8270-Modified
Q3298-06	Q3298-05MS	Water	SVOC-SIMGroup1	Cool 4 deg C	TETR06	D31	10/06/2025	8270-Modified
Q3298-07	Q3298-05MSD	Water	SVOC-SIMGroup1	Cool 4 deg C	TETR06	D31	10/06/2025	8270-Modified

Date/Time 10/08/25 8:45
Raw Sample Received by: RS CETA-1206)
Raw Sample Relinquished by: CP SM

Date/Time 10/08/25 9:15
Raw Sample Received by: CP SM
Raw Sample Relinquished by: RS CETA-1206)

Prep Standard - Chemical Standard Summary

Order ID : Q3292

Test : SVOC-SIMGroup1

Prepbatch ID : PB170022,

Sequence ID/Qc Batch ID: BN101325,

Standard ID :

EP2609,EP2610,EP2646,SP6830,SP6841,SP6842,SP6843,SP6844,SP6845,SP6846,SP6847,SP6848,SP6853,SP6854,SP6873,SP6875,SP6876,

Chemical ID :

10ul/1000ul

sample,E3657,E3875,E3942,E3943,E3972,E3973,M6157,S10105,S11073,S11496,S11652,S11828,S12115,S12197,S12201,S12220,S12273,S12499,S12552,S12560,S12577,S12670,S12905,S12986,S13058,S13097,S13120,S13245,W3112,



<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
1874	10 N SODIUM HYDROXIDE SOLN	EP2609	05/07/2025	11/07/2025	RUPESHKUMAR SHAH	Extraction_SC ALE_2 (EX-SC-2)	None	Riteshkumar Patel 05/07/2025
<u>FROM</u>	1000.00000ml of W3112 + 400.00000gram of E3657 = Final Quantity: 1000.000 ml							

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
314	1.1 H2SO4 SOLN	EP2610	05/07/2025	11/07/2025	RUPESHKUMAR SHAH	Extraction_SCALE_2 (EX-SC-2)	None	Riteshkumar Patel 05/07/2025
<u>FROM</u>	1000.00000ml of M6157 + 1000.00000ml of W3112 = Final Quantity: 2000.000 ml							



<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
3923	Baked Sodium Sulfate	EP2646	09/26/2025	01/28/2026	Evelyn Huang	Extraction_SC ALE_2 (EX-SC-2)	None	Riteshkumar Patel 09/26/2025
<u>FROM</u> 4000.00000gram of E3875 = Final Quantity: 4000.000 gram								

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
3493	Internal Standard 0.4 PPM	SP6830	06/17/2025	12/13/2025	Jagrut Upadhyay	None	None	Rahul Chavli 06/19/2025
<u>FROM</u>	0.10000ml of S12670 + 4.90000ml of E3942 = Final Quantity: 5.000 ml							



<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
3339	8270 sim calibration stock 10ppm (CPI)	SP6841	06/25/2025	10/28/2025	Jagrut Upadhyay	None	None	Rahul Chavli 06/25/2025
<u>FROM</u>	0.03350ml of S10105 + 0.05000ml of S11496 + 0.12500ml of S11828 + 0.12500ml of S12115 + 0.25000ml of S12273 + 0.25000ml of S13058 + 24.16650ml of E3943 = Final Quantity: 25.000 ml							

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
3361	8270-SIM MDL-5PPM CALIBRATION SOLUTION	SP6842	06/25/2025	10/28/2025	Jagrut Upadhyay	None	None	Rahul Chavli 06/25/2025
<u>FROM</u> 0.50000ml of E3943 + 0.01000ml of SP6830 + 0.50000ml of SP6841 = Final Quantity: 1.010 ml								

SVOC STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
3341	8270-SIM MDL-3.2PPM CALIBRATION SOLUTION	SP6843	06/25/2025	10/28/2025	Jagrut Upadhyay	None	None	Rahul Chavli
								06/25/2025

FROM 0.68000ml of E3943 + 0.01000ml of SP6830 + 0.32000ml of SP6841 = Final Quantity: 1.010 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
3344	8270-SIM MDL-1.6PPM CALIBRATION SOLUTION	SP6844	06/25/2025	10/28/2025	Jagrut Upadhyay	None	None	Rahul Chavli
								06/25/2025

FROM 0.84000ml of E3943 + 0.01000ml of SP6830 + 0.16000ml of SP6841 = Final Quantity: 1.010 ml

SVOC STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
3342	8270-SIM MDL-0.8PPM CALIBRATION SOLUTION	SP6845	06/25/2025	10/28/2025	Jagrut Upadhyay	None	None	Rahul Chavli
								06/25/2025

FROM 0.92000ml of E3943 + 0.01000ml of SP6830 + 0.08000ml of SP6841 = Final Quantity: 1.010 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
3343	8270-SIM MDL-0.4PPM CALIBRATION SOLUTION	SP6846	06/25/2025	10/28/2025	Jagrut Upadhyay	None	None	Rahul Chavli
								06/25/2025

FROM 0.96000ml of E3943 + 0.01000ml of SP6830 + 0.04000ml of SP6841 = Final Quantity: 1.010 ml

SVOC STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
3345	8270-SIM MDL-0.2PPM CALIBRATION SOLUTION	SP6847	06/25/2025	10/28/2025	Jagrut Upadhyay	None	None	Rahul Chavli
								06/25/2025

FROM 0.50000ml of E3943 + 0.01000ml of SP6830 + 0.50000ml of SP6846 = Final Quantity: 1.010 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
3346	8270-SIM MDL-0.1PPM CALIBRATION SOLUTION	SP6848	06/25/2025	10/28/2025	Jagrut Upadhyay	None	None	Rahul Chavli
								06/25/2025

FROM 0.75000ml of E3943 + 0.01000ml of SP6830 + 0.25000ml of SP6846 = Final Quantity: 1.010 ml



<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
3355	8270-SIM MDL-3.2PPM CALIBRATION STOCK SOL- 2ND	SP6853	07/02/2025	09/30/2025	Jagrut Upadhyay	None	None	Rahul Chavli
SOURCE 0.00630ml of S12197 + 0.01280ml of S12220 + 0.03200ml of S11073 + 0.03200ml of S11828 + 0.06400ml of S12499 + 0.06400ml of S12552 + 0.06400ml of S12986 + 19.72490ml of E3943 = Final Quantity: 20.000 ml								
FROM								

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
3356	8270-SIM MDL-0.4PPM CALIBRATION SOL ICV-2ND	SP6854	07/02/2025	09/30/2025	Jagrut Upadhyay	None	None	Rahul Chavli
SOURCE FROM 0.87500ml of E3943 + 0.01000ml of SP6830 + 0.12500ml of SP6853 = Final Quantity: 1.010 ml								

SVOC STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
3491	8270-SIM-Surrogate 0.4 PPM	SP6873	09/30/2025	12/09/2025	Jagrut Upadhyay	None	None	Rahul Chavli
								09/30/2025

FROM 0.00400ml of S12201 + 0.00800ml of S12905 + 0.02000ml of S11828 + 99.96800ml of E3972 = Final Quantity: 100.000 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
3895	50 ug/ml DFTPP 8270E	SP6875	09/30/2025	03/30/2026	Rahul Chavli	None	None	Jagrut Upadhyay
								09/30/2025

FROM 1.00000ml of S12577 + 19.00000ml of E3973 = Final Quantity: 20.000 ml

SVOC STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
3492	8270-SIM-Spike 0.4 PPM	SP6876	09/30/2025	11/30/2025	Jagrut Upadhyay	None	None	Rahul Chavli 09/30/2025
<u>FROM</u>	0.00080ml of S11652 + 0.01000ml of S13120 + 0.02000ml of S12560 + 0.02000ml of S13097 + 0.02000ml of S13245 + 49.92920ml of E3972 = Final Quantity: 50.000 ml							

CHEMICAL RECEIPT LOG BOOK

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
PCI Scientific Supply, Inc.	PC19510-5 / Sodium Hydroxide Pellets 2.5 Kg, Pk of 4	23B1556310	12/31/2025	12/04/2023 / Rajesh	12/01/2023 / Rajesh	E3657

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
PCI Scientific Supply, Inc.	PC19631-100 / SODIUM SULFATE, ANHYDROUS, PEST GRADE, 1	417203	01/28/2026	07/28/2025 / RUPESH	01/29/2025 / Rajesh	E3875

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Seidler Chemical	BA-9644-A4 / Methylene Chloride,U-Resi, Cycle-Tainer (215L)	25A2862010	12/13/2025	06/13/2025 / Rajesh	02/28/2025 / Rajesh	E3942

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Seidler Chemical	BA-9644-A4 / Methylene Chloride,U-Resi, Cycle-Tainer (215L)	25A2862010	12/13/2025	06/13/2025 / Rajesh	02/28/2025 / Rajesh	E3943

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Seidler Chemical	BA-9254-03 / Acetone, Ultra Resi (cs/4x4L)	24H1462005	05/24/2027	09/16/2025 / Evelyn	09/04/2025 / Riteshkumar	E3972

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Seidler Chemical	BA-9644-A4 / Methylene Chloride,U-Resi, Cycle-Tainer (215L)	25C1262005	04/16/2026	09/15/2025 / Riteshkumar	09/15/2025 / Riteshkumar	E3973

CHEMICAL RECEIPT LOG BOOK

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Seidler Chemical	BA-9673-33 / Sulfuric Acid, Instra-Analyzed (cs/6c2.5L)	24i1262013	11/07/2025	05/07/2025 / RUPESH	02/18/2025 / Mohan	M6157

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
CPI International	Z-112090-04 / CLP Acid Surrogate Solution, 7500 mg/L, 1ml	440246	12/19/2025	06/19/2025 / Jagrut	12/09/2021 / Christian	S10105

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	31853 / 1,4-Dioxane, 2000 ug/ml , Solvent: Methylene Chloride	A0187043	11/16/2025	05/16/2025 / Jagrut	02/06/2023 / Christian	S11073

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
CPI International	Z-110094-02 / CLP Base/Neutral Surrogate Solution, 5000 mg/L, 1ml	506889	10/28/2025	04/28/2025 / Jagrut	08/11/2023 / Yogesh	S11496

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	555872 / Custom Standard, pentachlorophenol Std [CS 5328-5]	A0201728	01/30/2026	07/30/2025 / Rahul	11/09/2023 / Yogesh	S11652

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	33913 / SOM01.0 SIM Analysis Standard (Surrogate), 2000 PPM	A0201976	12/09/2025	06/09/2025 / Jagrut	11/21/2023 / rahul	S11828

CHEMICAL RECEIPT LOG BOOK

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
CPI International	z-010223-01 / 1,4-Dioxane Solution, 2,000mg/L, 1ml	454157	10/28/2025	04/28/2025 / Jagrut	03/08/2024 / Rahul	S12115

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	31087 / Acid Surrogate 10,000ug/ml,methanol,5ml/ ampul	A0206206	01/02/2026	07/02/2025 / Jagrut	03/15/2024 / Rahul	S12197

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	31087 / Acid Surrogate 10,000ug/ml,methanol,5ml/ ampul	A0206206	03/22/2026	09/22/2025 / Jagrut	03/15/2024 / Rahul	S12201

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	31086 / Base Neutral Surrogate 5000ug/ml,CH ₂ Cl ₂ ,5ml	A0206381	01/02/2026	07/02/2025 / Jagrut	03/15/2024 / Rahul	S12220

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
CPI International	z-110381-01 / 8270 Calibration Solution, 76-1, 500 & 1,000 mg/L, 1ml	520963	12/25/2025	06/25/2025 / Jagrut	05/24/2024 / Rahul	S12273

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	555223 / Custom 8270 Plus Std #1 [2nd lot at \$100 per ampul if requested - contact ARM with Request]	A0214021	01/01/2026	07/01/2025 / Rahul	07/23/2024 / RAHUL	S12499

[CS 4978-1]

CHEMICAL RECEIPT LOG BOOK

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	555224 / Custom 8270 Plus Std #2 [2nd lot at \$85 per ampul if requested - contact ARM with Request]	A0214017	01/01/2026	07/01/2025 / Rahul	07/23/2024 / RAHUL	S12552

[CS 4978-2]

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	555224 / Custom 8270 Plus Std #2 [2nd lot at \$85 per ampul if requested - contact ARM with Request]	A0214017	03/22/2026	09/22/2025 / Jagrut	07/23/2024 / RAHUL	S12560

[CS 4978-2]

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	31615 / SV Mixture, GC/MS Tuning Mixture, CH ₂ Cl ₂ , 1mL,	A0212955	06/30/2027	03/31/2025 / Rahul	08/01/2024 / Rahul	S12577

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	31206 / SV Mix, CLP method, Internal Std, 2000ug/mL, CH ₂ Cl ₂ , 1mL	A0212266	12/16/2025	06/16/2025 / anahy	09/20/2024 / anahy	S12670

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	31086 / Base Neutral Surrogate 5000ug/ml, CH ₂ Cl ₂ , 5ml	A0216785	03/22/2026	09/22/2025 / Jagrut	12/09/2024 / anahy	S12905

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	31850 / 8270 SV Mix, 8270 Mega Mix 1mL, 1000ug/mL, CH ₂ Cl ₂ [New Solvent 100% CH ₂ Cl ₂]	A0219438	09/30/2025	06/04/2025 / Jagrut	12/11/2024 / anahy	S12986

CHEMICAL RECEIPT LOG BOOK

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
CPI International	Z-110816-01 / Custom 8270 Mix, 4-79, 1000 mg/L, 1 mL, (Maximum Expiration: 180 Days)	531243	12/25/2025	06/25/2025 / Jagrut	01/16/2025 / anahy	S13058

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	31850 / 8270 SV Mix, 8270 Mega Mix 1mL, 1000ug/mL, CH ₂ Cl ₂ [New Solvent 100% CH ₂ Cl ₂]	A0221014	11/30/2025	09/03/2025 / Jagrut	05/20/2025 / Rahul	S13097

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	31853 / 1,4-Dioxane, 2000 ug/ml, Solvent: Methylene Chloride	A0218894	02/25/2026	08/25/2025 / Jagrut	05/20/2025 / Rahul	S13120

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	555223 / Custom 8270 Plus Std #1 [2nd lot at \$100 per ampul if requested - contact ARM with Request]	A0228451	03/23/2026	09/22/2025 / Jagrut	08/06/2025 / Rahul	S13245

[CS 4978-1]

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Seidler Chemical	DIW / DI Water	Daily Lab-Certified	07/03/2029	07/03/2024 / lwona	07/03/2024 / lwona	W3112



5580 Skylane Blvd
Santa Rosa, CA 95403

(707)525-5788
(800)878-7654 Toll Free
(707)545-7901 Fax

Manufacturer's Quality System
Audited & Registered
by TUV USA to ISO 9001:2015

Date Received: _____

Certificate of Analysis

Rev 0

Page 1 of 1

Catalog No.: Lot No.: **Storage:** **Solvent:** **Exp. Date:** **Description:**
Z-112090 440246 $\leq -10^{\circ}\text{C}$ Methylene Chloride 2/16/2026 CLP Acid Surrogate Solution, 7,500 mg/L, 1 mL
-04

Compound	CAS No.	Purity (%)	Compound Lot No.	Concentration, mg/L
2-chlorophenol-d ₄	93951-73-6	99.3	248.12.7P	7487 $\pm$ 17.2
2-fluorophenol	367-12-4	99.8	10.7.3.3P	7513 $\pm$ 17.26
phenol-d ₆	13127-88-3	99.9	949.120.8P	7481 $\pm$ 17.19
2,4,6-tribromophenol	118-79-6	99.8	12.1.6P	7469 $\pm$ 17.17

Received on

02/25/21

by
CG

S9236
to

S9240

*Not a certified value

Manufactured by o2si smart solutions, Accredited to ISO 9001:2008 by NSF and ISO/IEC 17025:2005 (Certification No. 3031.01) and ISO Guide 34:2009 (Certification No. 3031.02) by A2LA

All weights are traceable through N. I. S. T. Test No. 822/264157-00.
Concentration (correct for purity) and uncertainty (95% confidence) values listed are determined gravimetrically.

Certified By: _____

Erica Castiglione
Chemist



CERTIFIED REFERENCE MATERIAL

110 Benner Circle
Bellefonte, PA 16823-8812
Tel: (800)356-1688
Fax: (814)353-1309

www.restek.com

Certificate of Analysis



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Received on
02/08/23

b1

CG

S 11071

to

S 11075

Catalog No. : 31853

Lot No.: A0187043

Description : 1,4-dioxane

1,4-Dioxane 2,000µg/mL, Methylene Chloride, 1mL/ampul

Container Size : 2 mL

Pkg Amt: > 1 mL

Expiration Date : July 31, 2027

Storage: 0°C or colder

Ship: Ambient

CERTIFIED VALUES

Elution Order	Compound	Grav. Conc. (weight/volume)	Expanded Uncertainty (95% C.L.; K=2)
1	1,4-Dioxane CAS # 123-91-1 Purity 99% (Lot SHBN5929)	2,019.0 µg/mL	+/- 11.8486 µg/mL Gravimetric +/- 43.2570 µg/mL Unstressed +/- 44.5129 µg/mL Stressed

Solvent: Methylene chloride

CAS # 75-09-2

Purity 99%

Column:

105m x 0.53mm x 3.0µm
Rtx-502.2 (cat.#10910)

Carrier Gas:

hydrogen-constant pressure 11.0 psi.

Temp. Program:

40°C (hold 2 min.) to 240°C
@ 8°C/min. (hold 5 min.)

Inj. Temp:


200°C

Det. Temp:

250°C

Det. Type:

FID



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.

Brittany Federinko - Operations Tech I

Date Mixed: 07-Jul-2022

Balance: 1128360905

Marlina Cowan - Operations Tech II ARM QC

Date Passed: 12-Jul-2022

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397



Certificate of Analysis

Sodium Hydroxide (Pellets)

Material: 0583
Grade: ACS GRADE
Batch Number: 23B1556310

Chemical Formula: NaOH
Molecular Weight: 40
CAS #: 1310-73-2
Appearance:

Manufacture Date: 12/14/2022
Expiration Date: 12/31/2025

Storage: Room Temperature

Pellets

TEST	SPECIFICATION	ANALYSIS	DISPOSITION
Calcium	$\leq 0.005 \%$	$< 0.005 \%$	PASS
Chloride	$\leq 0.005 \%$	0.002 %	PASS
Heavy Metals	$\leq 0.002 \%$	$< 0.002 \%$	PASS
Iron	$\leq 0.001 \%$	$< 0.001 \%$	PASS
Magnesium	$\leq 0.002 \%$	$< 0.002 \%$	PASS
Mercury	$\leq 0.1 \text{ ppm}$	$< 0.1 \text{ ppm}$	PASS
Nickel	$\leq 0.001 \%$	$< 0.001 \%$	PASS
Nitrogen Compounds	$\leq 0.001 \%$	$< 0.001 \%$	PASS
Phosphate	$\leq 0.001 \%$	$< 0.001 \%$	PASS
Potassium	$\leq 0.02 \%$	$< 0.02 \%$	PASS
Purity	$\geq 97.0 \%$	99.2 %	PASS
Sodium Carbonate	$\leq 1.0 \%$	0.5 %	PASS
Sulfate	$\leq 0.003 \%$	$< 0.003 \%$	PASS

Internal ID #: 710

Signature

We certify that this batch conforms to the specifications listed.


This document has been electronically produced and is valid without a signature.

Leona Edwardson, Quality Control Sr. Manager - Solon
VWR Chemicals, LLC.
28600 Fountain Parkway, Solon OH 44139 USA

Additional Information

Analysis may have been rounded to significant digits in specification limits.

Product meets analytical specifications of the grades listed.





**PRODUCTOS
QUÍMICOS
MONTERREY, S.A. DE C.V.**

MIRADOR 201, COL. MIRADOR
MONTERREY, N.L. MÉXICO
CP 64070
TEL +52 81 13 52 67 67
www.pqm.com.mx

CERTIFICATE OF ANALYSIS

PRODUCT :	SODIUM SULFATE CRYSTALS ANHYDROUS		
QUALITY :	ACS (CODE RMB3375)	FORMULA :	Na ₂ SO ₄
SPECIFICATION NUMBER:	6399	RELEASE DATE:	MAY/23/2024
LOT NUMBER :	417203		

TEST	SPECIFICATIONS	LOT VALUES
Assay (Na ₂ SO ₄)	Min. 99.0%	99.8 %
pH of a 5% solution at 25°C	5.2 - 9.2	6.2
Insoluble matter	Max. 0.01%	0.001 %
Loss on ignition	Max. 0.5%	0.1 %
Chloride (Cl)	Max. 0.001%	<0.001 %
Nitrogen compounds (as N)	Max. 5 ppm	<5 ppm
Phosphate (PO ₄)	Max. 0.001%	<0.001 %
Heavy metals (as Pb)	Max. 5 ppm	<5 ppm
Iron (Fe)	Max. 0.001%	<0.001 %
Calcium (Ca)	Max. 0.01%	0.001 %
Magnesium (Mg)	Max. 0.005%	0.001 %
Potassium (K)	Max. 0.008%	0.001 %
Extraction-concentration suitability	Passes test	Passes test
Appearance	Passes test	Passes test
Identification	Passes test	Passes test
Solubility and foreign matter	Passes test	Passes test
Retained on US Standard No. 10 sieve	Max. 1%	0.2 %
Retained on US Standard No. 60 sieve	Min. 94%	96.2 %
Through US Standard No. 60 sieve	Max. 5%	3.5 %
Through US Standard No. 100 sieve	Max. 10%	0.1 %

COMMENTS

QC: PhC Irma Belmares

If you need further details, please call our factory or contact our local distributor.

RE-02-01, Ed. 3

E 3875

Methylene Chloride
ULTRA RESI-ANALYZED
For Organic Residue Analysis
(dichloromethane)



Material No.: 9266-A4
Batch No.: 25A2862010
Manufactured Date: 2024-12-18
Expiration Date: 2026-03-19
Revision No.: 0

Certificate of Analysis

Test	Specification	Result
FID-Sensitive Impurities (as 2-Octanol) Single Impurity Peak (ng/mL)	≤ 5	<1
ECD Sensitive Impurities (as HeptachlorEpoxide) Single Peak (pg/mL)	≤ 10	2
Assay (CH ₂ Cl ₂) (by GC, exclusive of preservative, corrected for water)	$\geq 99.8 \%$	99.9 %
Color (APHA)	≤ 10	5
Residue after Evaporation	≤ 1.0 ppm	0.3 ppm
Titration Acid (μ eq/g)	≤ 0.3	<0.1
Chloride (Cl)	≤ 10 ppm	<5 ppm
Water (by KF, coulometric)	$\leq 0.02 \%$	<0.01 %

For Laboratory, Research, or Manufacturing Use
MEETS SPECIFICATIONS WITHIN THE EXPIRATION PERIOD

Country of Origin: United States
Packaging Site: Phillipsburg Mfg Ctr & DC

E3942

Jamie Croak
Director Quality Operations, Bioscience Production

For questions on this Certificate of Analysis please contact Technical Services at 855.282.6867 or +1.610.386.1700

Avantor Performance Materials LLC

Methylene Chloride
ULTRA RESI-ANALYZED
For Organic Residue Analysis
(dichloromethane)

avantor™



Material No.: 9266-A4

Batch No.: 25A2862010

Manufactured Date: 2024-12-18

Expiration Date: 2026-03-19

Revision No.: 0

Certificate of Analysis

Test	Specification	Result
FID-Sensitive Impurities (as 2-Octanol) Single Impurity Peak (ng/mL)	≤ 5	<1
ECD Sensitive Impurities (as HeptachlorEpoxide) Single Peak (pg/mL)	≤ 10	2
Assay (CH ₂ Cl ₂) (by GC, exclusive of preservative, corrected for water)	$\geq 99.8 \%$	99.9 %
Color (APHA)	≤ 10	5
Residue after Evaporation	≤ 1.0 ppm	0.3 ppm
Titration Acid (μ eq/g)	≤ 0.3	<0.1
Chloride (Cl)	≤ 10 ppm	<5 ppm
Water (by KF, coulometric)	$\leq 0.02 \%$	<0.01 %

For Laboratory, Research, or Manufacturing Use

MEETS SPECIFICATIONS WITHIN THE EXPIRATION PERIOD

Country of Origin: United States

Packaging Site: Phillipsburg Mfg Ctr & DC

E3942

Jamie Croak
Director Quality Operations, Bioscience Production

For questions on this Certificate of Analysis please contact Technical Services at 855.282.6867 or +1.610.386.1700

Avantor Performance Materials LLC

Acetone
BAKER RESI-ANALYZED® Reagent
For Organic Residue Analysis

avantor™



Material No.: 9254-03

Batch No.: 24H1462005

Manufactured Date: 2024-05-24

Expiration Date: 2027-05-24

Revision No.: 0

Certificate of Analysis

Test	Specification	Result
Assay ((CH ₃) ₂ CO) (by GC, corrected for water)	>= 99.4 %	99.8 %
Color (APHA)	<= 10	5
Residue after Evaporation	<= 1.0 ppm	0.2 ppm
Substances Reducing Permanganate	Passes Test	Passes Test
Titration Acid (µeq/g)	<= 0.3	0.2
Titration Base (µeq/g)	<= 0.6	<0.1
Water (H ₂ O)	<= 0.5 %	0.2 %
FID-Sensitive Impurities (as 2-Octanol) Single Impurity Peak (ng/mL)	<= 5	<1
ECD Sensitive Impurities (as Heptachlor Epoxide) Single Peak (pg/mL)	<= 10	1

For Laboratory, Research, or Manufacturing Use
MEETS SPECIFICATIONS WITHIN THE EXPIRATION PERIOD

Country of Origin: United States
Packaging Site: Phillipsburg Mfg Ctr & DC

E3972

Jamie Croak
Director Quality Operations, Bioscience Production

For questions on this Certificate of Analysis please contact Technical Services at 855.282.6867 or +1.610.386.1700

Avantor Performance Materials LLC

Sulfuric Acid
BAKER INSTRA-ANALYZED® Reagent
For Trace Metal Analysis
Low Selenium

avantor™



MG157
MS

Material No.: 9673-33

Batch No.: 24I1262013

Manufactured Date: 2024-08-07

Retest Date: 2029-08-06

Revision No.: 0

Certificate of Analysis

Test	Specification	Result
ACS – Assay (H ₂ SO ₄)	95.0 – 98.0 %	96.2 %
Appearance	Passes Test	Passes Test
ACS – Color (APHA)	<= 10	5
ACS – Residue after Ignition	<= 3 ppm	<1 ppm
ACS – Substances Reducing Permanganate(as SO ₂)	<= 2 ppm	<2 ppm
Ammonium (NH ₄)	<= 1 ppm	<1 ppm
Chloride (Cl)	<= 0.1 ppm	<0.1 ppm
Nitrate (NO ₃)	<= 0.2 ppm	0.1 ppm
Phosphate (PO ₄)	<= 0.5 ppm	<0.1 ppm
Trace Impurities – Aluminum (Al)	<= 30.0 ppb	<5.0 ppb
Arsenic & Antimony (as As)	<= 4.0 ppb	<2.0 ppb
Trace Impurities – Boron (B)	<= 10.0 ppb	<5.0 ppb
Trace Impurities – Cadmium (Cd)	<= 2.0 ppb	<1.0 ppb
Trace Impurities – Chromium (Cr)	<= 6.0 ppb	<1.0 ppb
Trace Impurities – Cobalt (Co)	<= 0.5 ppb	<0.3 ppb
Trace Impurities – Copper (Cu)	<= 1.0 ppb	<1.0 ppb
Trace Impurities – Gold (Au)	<= 10.0 ppb	<5.0 ppb
Heavy Metals (as Pb)	<= 500.0 ppb	<100.0 ppb
Trace Impurities – Iron (Fe)	<= 50.0 ppb	<1.0 ppb
Trace Impurities – Lead (Pb)	<= 0.5 ppb	<0.5 ppb
Trace Impurities – Magnesium (Mg)	<= 7.0 ppb	<1.0 ppb
Trace Impurities – Manganese (Mn)	<= 1.0 ppb	<1.0 ppb
Trace Impurities – Mercury (Hg)	<= 0.5 ppb	<0.1 ppb
Trace Impurities – Nickel (Ni)	<= 2.0 ppb	<0.3 ppb
Trace Impurities – Potassium (K)	<= 500.0 ppb	<10.0 ppb
Trace Impurities – Selenium (Se)	<= 50.0 ppb	7.2 ppb
Trace Impurities – Silicon (Si)	<= 100.0 ppb	12.8 ppb
Trace Impurities – Silver (Ag)	<= 1.0 ppb	<1.0 ppb

For questions on this Certificate of Analysis please contact Technical Services at 855.282.6867 or +1.610.386.1700

Avantor Performance Materials LLC

Sulfuric Acid

BAKER INSTRA-ANALYZED® Reagent

For Trace Metal Analysis

Low Selenium

 **avantor**™



Material No.: 9673-33

Batch No.: 2411262013

Test	Specification	Result
Trace Impurities – Sodium (Na)	≤ 500.0 ppb	< 5.0 ppb
Trace Impurities – Strontium (Sr)	≤ 5.0 ppb	< 1.0 ppb
Trace Impurities – Tin (Sn)	≤ 5.0 ppb	1.1 ppb
Trace Impurities – Zinc (Zn)	≤ 5.0 ppb	< 1.0 ppb

For Laboratory, Research, or Manufacturing Use

Country of Origin: United States

Packaging Site: Phillipsburg Mfg Ctr & DC



Jamie Croak

Director Quality Operations, Bioscience Production

For questions on this Certificate of Analysis please contact Technical Services at 855.282.6867 or +1.610.386.1700

Avantor Performance Materials LLC



5580 Skylane Blvd
Santa Rosa, CA 95403

(707)525-5788
(800)878-7654 Toll Free
(707)545-7901 Fax

Manufacturer's Quality System
Audited & Registered
by TUV USA to ISO 9001:2015

Date Received: _____

Certificate of Analysis

Rev 0

Page 1 of 1

Catalog No.:	Lot No.:	Storage:	Solvent:	Exp. Date:	Description:
Z-110094-02	506889	≤ -10 °C	Methylene Chloride	7/25/2028	CLP Base/Neutral Surrogate Solution, 5,000 mg/L, 1 ml

Compound	CAS No.	Purity (%)	Compound Lot No.	Concentration, mg/L
1,2-dichlorobenzene-d ₄	2199-69-1	99.7	247.29.3P	5035 ± 28.02
2-fluorobiphenyl	321-60-8	99.69	8.286.1.1P	4999 ± 103.66
nitrobenzene-d5	4165-60-0	99.67	7.9.3P	4988 ± 27.32
p-terphenyl-d14	1718-51-0	99.3	9.120.8P	5005 ± 27.85

511494 } Y.P.
↓
511498 } 08/11/2023

*Not a certified value

Certified By: _____

Clint Tipton
Chemist

All weights are traceable through N. I. S. T. Test No. 822/264157-00.
Concentration (correct for purity) and uncertainty (95% confidence) values
listed are determined gravimetrically.



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

gravimetric



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 555872 **Lot No.:** A0201728

Description : Custom Pentachlorophenol Standard
Custom Pentachlorophenol Standard 25,000µg/mL, Methanol, 1mL/ampul

Container Size : 2 mL **Pkg Amt:** > 1 mL

Expiration Date : September 30, 2026 **Storage:** 10°C or colder

Ship: Ambient

511649
↓
511658 } Y.P.
11/13/23

CERTIFIED VALUES

Component #	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	Pentachlorophenol	87-86-5	RP230530RSR	99%	25,000.0 µg/mL	+/- 777.0837

Solvent: Methanol
CAS # 67-56-1
Purity 99%

Josh McCloskey - Operations Technician I

Date Mixed: 05-Sep-2023

Balance: B251644995

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397

General Certified Reference Material Notes

Expiration Notes:

- Expiration date valid for unopened ampul stored in compliance with the recommended conditions.
- Uncertainty, concentration, and expiration of the CRM are based on the unopened product being stored according to the recommended condition found in the storage field.

Purity Notes:

- Purity and/or chemical identity are determined by one or more of the following techniques: GC/FID, HPLC, GC/μECD, GC/MS, LC/MS, RI, and/or melting point.
- Compounds with a listed purity of less than 99% have been weight corrected to compensate for impurities and/or salts. A correction factor is used to calculate the amount of compound necessary to achieve the desired concentration of the parent compound in solution.
- Purity of isomeric compounds is reported as the sum of the isomers.
- Purity values are rounded to the nearest whole number.

Certified Uncertainty Value Notes:

- The uncertainties are determined in accordance with ISO 17034 and Guide 35. The certified expanded uncertainty value includes gravimetric uncertainty, homogeneity between-ampul uncertainty, storage stability uncertainty and shipping stability uncertainty and were combined using the following formula:

$$U_{combined\ uncertainty} = k \sqrt{u_{gravimetric}^2 + u_{homogeneity}^2 + u_{storage\ stability}^2 + u_{shipping\ stability}^2}$$

k is a coverage factor of 2, which gives a level of confidence of approximately 95%.

- The packaged amount is the minimum sample size for which uncertainty is valid. The ampuls are over-filled to ensure that the minimum packaged amount can be sufficiently transferred.

Manufacturing Notes:

- Concentration is based upon gravimetric preparation using either a balance whose calibration has been verified daily using NIST traceable weights, and/or dilutions with Class A glassware.

Handling Notes:

- Stability of the unopened product, when stored in compliance with the recommended conditions, is guaranteed through the expiration displayed on the product label and certificate. Contact Restek for additional opened product stability information, with the knowledge/understanding that open product stability is subject to the specific handling and environmental conditions to which the product is exposed. For your convenience Restek supplies deactivated vials with most standards packed in 2mL ampuls. Larger volume deactivated vials are available through Restek as a custom ordered item. Additionally, Restek sells DMDCS for the purpose of glassware deactivation as catalog number 31861, which includes complete instructions.
- If any undissolved material is visible inside the ampul, sonicate the unopened ampul until the material is completely dissolved.



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

chromatographic plus



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 33913 **Lot No.:** A0201976

Description : SOM01.0 SIM Analysis Standard
SOM01.0 SIM Analysis Standard 2000µg/mL, Methylene chloride, 1mL /ampul

Container Size : 2 mL **Pkg Amt:** > 1 mL

Expiration Date : August 31, 2029 **Storage:** 10°C or colder

Handling: Sonication required. Mix is photosensitive. **Ship:** Ambient

S11828
↓
S11832 } RC/
11/30/23

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	2-Methylnaphthalene-d10	7297-45-2	EF-135	98%	2,015.9 µg/mL	+/- 90.8098
2	Fluoranthene-d10	93951-69-0	PR-32557	99%	2,020.0 µg/mL	+/- 90.9963

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: Methylene chloride
CAS # 75-09-2
Purity 99%

Quality Confirmation Test

Column:

30m x 0.25mm x 0.25µm
Rtx-5 (cat.#10223)

Carrier Gas:

hydrogen-constant pressure 10 psi.

Temp. Program:

75°C (hold 1 min.) to 330°C
@ 20°C/min. (hold 10 min.)

Inj. Temp:

250°C

Det. Temp:

330°C

Det. Type:


FID

Split Vent:

10 ml/min.

Inj. Vol

1µl




This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.


Dakota Parson - Operations Technician I

Date Mixed: 13-Sep-2023

Balance Serial # B442140311


Jennifer Pollino - Operations Tech III - ARM QC

Date Passed: 28-Sep-2023

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397

General Certified Reference Material Notes

Expiration Notes:

- Expiration date valid for unopened ampul stored in compliance with the recommended conditions.
- Uncertainty, concentration, and expiration of the CRM are based on the unopened product being stored according to the recommended condition found in the storage field.

Purity Notes:

- Purity and/or chemical identity are determined by one or more of the following techniques: GC/FID, HPLC, GC/μECD, GC/MS, LC/MS, RI, and/or melting point.
- Compounds with a listed purity of less than 99% have been weight corrected to compensate for impurities and/or salts. A correction factor is used to calculate the amount of compound necessary to achieve the desired concentration of the parent compound in solution.
- Purity of isomeric compounds is reported as the sum of the isomers.
- Purity values are rounded to the nearest whole number.

Certified Uncertainty Value Notes:

- The uncertainties are determined in accordance with ISO 17034 and Guide 35. The certified expanded uncertainty value includes gravimetric uncertainty, homogeneity between-ampul uncertainty, storage stability uncertainty and shipping stability uncertainty and were combined using the following formula:

$$U_{combined\ uncertainty} = k \sqrt{u_{gravimetric}^2 + u_{homogeneity}^2 + u_{storage\ stability}^2 + u_{shipping\ stability}^2}$$

k is a coverage factor of 2, which gives a level of confidence of approximately 95%.

- The packaged amount is the minimum sample size for which uncertainty is valid. The ampuls are over-filled to ensure that the minimum packaged amount can be sufficiently transferred.

Manufacturing Notes:

- Concentration is based upon gravimetric preparation using either a balance whose calibration has been verified daily using NIST traceable weights, and/or dilutions with Class A glassware.

Handling Notes:

- Stability of the unopened product, when stored in compliance with the recommended conditions, is guaranteed through the expiration displayed on the product label and certificate. Contact Restek for additional opened product stability information, with the knowledge/understanding that open product stability is subject to the specific handling and environmental conditions to which the product is exposed. For your convenience Restek supplies deactivated vials with most standards packed in 2mL ampuls. Larger volume deactivated vials are available through Restek as a custom ordered item. Additionally, Restek sells DMDCS for the purpose of glassware deactivation as catalog number 31861, which includes complete instructions.
- If any undissolved material is visible inside the ampul, sonicate the unopened ampul until the material is completely dissolved.



5580 Skylane Blvd
Santa Rosa, CA 95403

(707)525-5788
(800)878-7654 Toll Free
(707)545-7901 Fax

Manufacturer's Quality System
Audited & Registered
by TUV USA to ISO 9001:2015

Date Received: _____

Certificate of Analysis

Rev 0

Page 1 of 1

Catalog No.:	Lot No.:	Storage:	Solvent:	Exp. Date:	Description:
Z-020223-01	454157	≤ -10 °C	P/T Methanol	6/10/2026	1,4-Dioxane Solution, 2000 mg/L, 1 mL

Compound	CAS No.	Purity (%)	Compound Lot No.	Concentration, mg/L
1,4-dioxane	123-91-1	100	223.1.3P	1997 ± 57.08

512112 } RC/
↓
912116 } 03/08/24

*Not a certified value

Certified By: _____

Melissa Workoff
Melissa Workoff
Chemist

All weights are traceable through N. I. S. T. Test No. 822/264157-00.
Concentration (correct for purity) and uncertainty (95% confidence) values
listed are determined gravimetrically.



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

chromatographic plus



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 31087 **Lot No.:** A0206206

Description : Acid Surrogate Mix (4/89 SOW)

Acid Surrogate 10, 000µg/mL, Methanol, 5mL/ampul

Container Size : 5 mL **Pkg Amt:** > 5 mL

Expiration Date : January 31, 2032 **Storage:** 10°C or colder

Ship: Ambient

512187 } RC/
↓ } 03/18/24
512206 }

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	2-Fluorophenol	367-12-4	STBK1705	99%	10,005.3 µg/mL	+/- 302.5390
2	Phenol-d6	13127-88-3	PR-33287A	99%	10,005.5 µg/mL	+/- 302.5475
3	2,4,6-Tribromophenol	118-79-6	RP230831RSR	99%	10,006.6 µg/mL	+/- 302.5783

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: Methanol
CAS # 67-56-1
Purity 99%

Quality Confirmation Test

Column:

30m x 0.25mm x 0.25µm
Rtx-5 (cat.#10223)

Carrier Gas:

hydrogen-constant pressure 10 psi.

Temp. Program:

40°C (hold 2 min.) to 330°C
@ 10°C/min. (hold 10 min.)

Inj. Temp:

250°C

Det. Temp:

330°C

Det. Type:


FID

Split Vent:

2 ml/min.

Inj. Vol

1µl



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.

Penelope S. Riglin

Penelope Riglin - Operations Tech I

Date Mixed: 04-Jan-2024

Balance Serial # 1128360905

Christie Mills

Christie Mills - Operations Lead Tech - ARM QC

Date Passed: 08-Jan-2024

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

chromatographic plus



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 31086 **Lot No.:** A0206381
Description : B/N Surrogate Mix (4/89 SOW)
Base Neutral Surrogate 5000µg/mL, Methylene Chloride, 5mL/ampul
Container Size : 5 mL **Pkg Amt:** > 5 mL
Expiration Date : December 31, 2029 **Storage:** 10°C or colder
Handling: Sonicate prior to use. **Ship:** Ambient

S12207 } RC/
↓
S12221 } 03/18/24

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	Nitrobenzene-d5	4165-60-0	I-25158	99%	5,029.3 µg/mL	+/- 226.5204
2	2-Fluorobiphenyl	321-60-8	00021384	99%	5,030.9 µg/mL	+/- 226.5936
3	p-Terphenyl-d14	1718-51-0	PR-32599	99%	5,026.4 µg/mL	+/- 226.3909

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: Methylene chloride
CAS # 75-09-2
Purity 99%

Tech Tips:

Due to the limited solubility of p-terphenyl-d14 in methanol, we do not recommend that this mixture be diluted in methanol.

Quality Confirmation Test

Column:

30m x 0.25mm x 0.25µm
Rtx-S (cat.#10223)

Carrier Gas:

hydrogen-constant pressure 10 psi.

Temp. Program:

40°C (hold 2 min.) to 330°C
@ 10°C/min. (hold 10 min.)

Inj. Temp:

250°C

Det. Temp:

330°C

Det. Type:


FID

Split Vent:

2 ml/min.

Inj. Vol

1µl



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.

Jess Hoy - Operations Tech I

Date Mixed: 09-Jan-2024 Balance Serial # 1128360905

Jennifer Pollino - Operations Tech III - ARM QC

Date Passed: 11-Jan-2024

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397



5580 Skylane Blvd
Santa Rosa, CA 95403

(707)525-5788
(800)878-7654 Toll Free
(707)545-7901 Fax

Manufacturer's Quality System
Audited & Registered
by TUV USA to ISO 9001:2015

Date Received: _____

Certificate of Analysis

Rev 0

Page 1 of 4

Catalog No.:	Lot No.:	Storage:	Solvent:	Exp. Date:	Description:
Z-110381-01	520963	≤ -10 °C	Methylene Chloride	10/10/2028	Method 8270 Calibration Solution, 76-1, 500 & 1,000 mg/L, 1 mL

Compound	CAS No.	Purity (%)	Compound Lot No.	Concentration, mg/L
acenaphthene	83-32-9	99.9	13.1.5P	1010 ± 9.89
acenaphthylene	208-96-8	97.6	14.290.1P	1014 ± 9.93
aniline	62-53-3	99.97	64.1.4P	1001 ± 9.8
anthracene	120-12-7	99.5	15.7.1P	999.6 ± 9.79
azobenzene	103-33-3	98.1	252.7.2P	999.1 ± 9.8
benzo[a]anthracene	56-55-3	100	16.7.3P	1007 ± 9.86
benzo[b]fluoranthene	205-99-2	99.8	17.421.3P	1011 ± 14.11
benzo[k]fluoranthene	207-08-9	98.9	18.421.4P	1001 ± 10.96
benzo[ghi]perylene	191-24-2	93	19.286.4P	999.6 ± 13.95
benzo[a]pyrene	50-32-8	97	20.286.2P	999.9 ± 22.24
benzyl alcohol	100-51-6	99.9	65.18.1P	1001 ± 9.82
bis(2-chloroethoxy)methane	111-91-1	99.1	31.3.15P	1000 ± 14.69
bis(2-chloroethyl)ether	111-44-4	99.8	32.7.1P	1003 ± 13.89
bis(2-chloro-1-methylethyl) ether	108-60-1	99.5	34.3.15P	999.4 ± 14.68
bis(2-ethylhexyl)adipate	103-23-1	99.5	874.7.1P	999.5 ± 9.8
bis(2-ethylhexyl)phthalate	117-81-7	99.4	33.29.1P	998.8 ± 17.03
4-bromophenyl phenyl ether	101-55-3	99.4	35.7.1.1P	1000 ± 13.85
butyl benzyl phthalate	85-68-7	98.4	36.1.6P	984.7 ± 16.79
carbazole	86-74-8	99.4	239.7.2P	1000 ± 9.8

*Not a certified value

512270 } RC/
↓
512274 } 05/24/24

Kerry Kane

Certified By: _____

Kerry Kane
Chemist

All weights are traceable through N. I. S. T. Test No. 822/264157-00.
Concentration (correct for purity) and uncertainty (95% confidence) values
listed are determined gravimetrically.

Certificate of Analysis

Page 2 of 4

Catalog No.: Z-110381-01

Lot No.: 520963

Expiration Date: 10/10/2028

Compound	CAS No.	Purity (%)	Compound Lot No.	Concentration, mg/L
4-chloroaniline	106-47-8	100	66.7.1P	1000 ± 9.79
4-chlorophenylphenyl ether	7005-72-3	98	37.158.2P	1001 ± 17.07
4-chloro-3-methylphenol	59-50-7	99	102.1.2P	1006 ± 17.16
2-chloronaphthalene	91-58-7	99.9	42.7.6P	1000 ± 9.79
2-chlorophenol	95-57-8	99.8	103.7.1P	1007 ± 13.96
chrysene	218-01-9	96	21.286.2P	998.4 ± 12.85
dibenz[a,h]anthracene	53-70-3	99.44	22.286.3P	1000 ± 9.74
dibenzofuran	132-64-9	100	67.7.2.1P	1002 ± 9.77
di-n-butyl phthalate	84-74-2	99.84	40.286.1P	1007 ± 24.48
1,2-dichlorobenzene	95-50-1	99.8	43.7.1P	1000 ± 9.79
1,3-dichlorobenzene	541-73-1	99.5	44.1.3P	999.4 ± 9.79
1,4-dichlorobenzene	106-46-7	99.9	45.29.2P	1000 ± 9.79
2,4-dichlorophenol	120-83-2	99.6	104.7.1.1P	1005 ± 13.93
diethyl phthalate	84-66-2	99.8	38.7.1P	1011 ± 14
2,4-dimethylphenol	105-67-9	99.6	105.7.1.1P	1009 ± 13.98
dimethyl phthalate	131-11-3	99.9	39.9.2P	996.5 ± 13.8
1,2-dinitrobenzene	528-29-0	99.86	86.7.3.1P	999.5 ± 9.75
1,3-dinitrobenzene	99-65-0	100	313.7.2P	998 ± 9.79
1,4-dinitrobenzene	100-25-4	100	907.7.1P	999.5 ± 9.8
2,4-dinitrophenol	51-28-5	99.9	106.1.6DP	1002 ± 13.89
2,4-dinitrotoluene	121-14-2	100	87.7.3P	999.8 ± 13.85
2,6-dinitrotoluene	606-20-2	99.4	88.7.2.1P	999.6 ± 13.85
di-n-octyl phthalate	117-84-0	99.1	41.7.5P	991.6 ± 13.74
diphenylamine	122-39-4	100	78.1.6P	998 ± 13.79
2,3,5,6-tetrachlorophenol	935-95-5	97	1112.286.1P	1004 ± 14.02
fluoranthene	206-44-0	98.6	23.7.4P	999.6 ± 9.79
fluorene	86-73-7	98.4	24.7.1P	999.7 ± 9.79

*Not a certified value

Kerry Kane

Certified By:

Kerry Kane
Chemist

All weights are traceable through N. I. S. T. Test No. 822/264157-00.
Concentration (correct for purity) and uncertainty (95% confidence) values listed are determined gravimetrically.

Certificate of Analysis

Page 3 of 4

Catalog No.: Z-110381-01

Lot No.: 520963

Expiration Date: 10/10/2028

Compound	CAS No.	Purity (%)	Compound Lot No.	Concentration, mg/L
hexachlorobenzene	118-74-1	99	46.158.4P	999.9 ± 13.96
hexachlorobutadiene	87-68-3	97.4	47.1.4P	1000 ± 9.79
hexachlorocyclopentadiene	77-47-4	99.2	48.2.2P	1001 ± 9.8
hexachloroethane	67-72-1	99.9	49.1.4P	1003 ± 9.82
indeno[1,2,3-cd]pyrene	193-39-5	98	25.286.4P	999.4 ± 22.23
isophorone	78-59-1	98.9	90.1.4P	999.9 ± 13.85
2-methyl-4,6-dinitrophenol	534-52-1	99.6	107.421.2DP	991 ± 24.09
1-methylnaphthalene	90-12-0	97.1	249.7.5P	999.2 ± 13.95
2-methylnaphthalene	91-57-6	97.4	68.7.2P	1006 ± 22.38
2-methylphenol	95-48-7	99.6	114.7.3P	1001 ± 13.87
3-methylphenol	108-39-4	99.1	115.7.4P	499.7 ± 6.92
4-methylphenol	106-44-5	99.5	116.7.1P	501.2 ± 6.94
naphthalene	91-20-3	99.8	26.9.1P	1018 ± 9.97
2-nitroaniline	88-74-4	99.7	69.29.1P	999.6 ± 9.79
3-nitroaniline	99-09-2	100	70.7.3P	1000 ± 9.74
4-nitroaniline	100-01-6	99.7	71.29.1P	1001 ± 9.8
nitrobenzene	98-95-3	100	94.7.1P	1000 ± 13.85
2-nitrophenol	88-75-5	99.1	108.29.1P	996.5 ± 13.81
4-nitrophenol	100-02-7	100	109.7.1P	1000 ± 13.82
N-nitrosodimethylamine	62-75-9	99.5	57.3.19P	998.5 ± 14.67
N-nitrosodi-n-propylamine	621-64-7	99.8	59.286.1P	996.8 ± 17
pentachlorophenol	87-86-5	99	110.1.7P	1004 ± 13.92
phenanthrene	85-01-8	99.7	27.1.5P	999 ± 12.87
phenol	108-95-2	100	112.7.1P	998.5 ± 13.8
pyrene	129-00-0	99.2	28.9.2P	998.9 ± 9.78
pyridine	110-86-1	100	101.24.1P	999 ± 9.73
2,3,4,6-Tetrachlorophenol	58-90-2	91.8	120.421.1P	996.5 ± 13.92

*Not a certified value

Kerry Kane

Certified By:

Kerry Kane
Chemist

All weights are traceable through N. I. S. T. Test No. 822/264157-00.
Concentration (correct for purity) and uncertainty (95% confidence) values listed are determined gravimetrically.

Certificate of Analysis

Page 4 of 4

Catalog No.: Z-110381-01

Lot No.: 520963

Expiration Date: 10/10/2028

Compound	CAS No.	Purity (%)	Compound Lot No.	Concentration, mg/L
1,2,4-trichlorobenzene	120-82-1	99.6	54.29.1P	999.6 ± 9.79
2,4,5-trichlorophenol	95-95-4	96.5	121.7.1.1P	999.5 ± 13.85
2,4,6-trichlorophenol	88-06-2	99.6	113.7.1P	996 ± 13.8

*Not a certified value



Certified By:

Kerry Kane
Chemist

All weights are traceable through N. I. S. T. Test No. 822/264157-00.
Concentration (correct for purity) and uncertainty (95% confidence) values
listed are determined gravimetrically.



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

gravimetric



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 555223 **Lot No.:** A0214021

Description : Custom 8270 Plus Standard #1

Custom 8270 Plus Standard #1 1,000µg/mL, Methylene Chloride, 1mL/ampul

Container Size : 2 mL **Pkg Amt:** > 1 mL

Expiration Date : July 31, 2026 **Storage:** 10°C or colder

Handling: This product is photosensitive. **Ship:** Ambient

CERTIFIED VALUES

Component #	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	3,3'-Dichlorobenzidine	91-94-1	S240326RSR	99%	1,004.0 µg/mL	+/- 23.0487
2	Atrazine	1912-24-9	5FYWL	99%	1,005.0 µg/mL	+/- 23.0717
3	Benzidine	92-87-5	S240430RSR	99%	1,006.0 µg/mL	+/- 23.0947
4	epsilon-Caprolactam	105-60-2	Y16H012	99%	1,000.0 µg/mL	+/- 22.9569

Solvent: Methylene chloride
CAS # 75-09-2
Purity 99%

S12449 } RC/
↓
S12508 } 7/24/24

Rebecca Gingerich - Operations Tech II

Date Mixed: 18-Jul-2024

Balance: 1128353505

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397

General Certified Reference Material Notes

Expiration Notes:

- Expiration date valid for unopened ampul stored in compliance with the recommended conditions.
- Uncertainty, concentration, and expiration of the CRM are based on the unopened product being stored according to the recommended condition found in the storage field.

Purity Notes:

- Purity and/or chemical identity are determined by one or more of the following techniques: GC/FID, HPLC, GC/μECD, GC/MS, LC/MS, RI, and/or melting point.
- Compounds with a listed purity of less than 99% have been weight corrected to compensate for impurities and/or salts. A correction factor is used to calculate the amount of compound necessary to achieve the desired concentration of the parent compound in solution.
- Purity of isomeric compounds is reported as the sum of the isomers.
- Purity values are rounded to the nearest whole number.

Certified Uncertainty Value Notes:

- The uncertainties are determined in accordance with ISO 17034 and Guide 35. The certified expanded uncertainty value includes gravimetric uncertainty, homogeneity between-ampul uncertainty, storage stability uncertainty and shipping stability uncertainty and were combined using the following formula:

$$U_{combined\ uncertainty} = k \sqrt{u_{gravimetric}^2 + u_{homogeneity}^2 + u_{storage\ stability}^2 + u_{shipping\ stability}^2}$$

k is a coverage factor of 2, which gives a level of confidence of approximately 95%.

- The packaged amount is the minimum sample size for which uncertainty is valid. The ampuls are over-filled to ensure that the minimum packaged amount can be sufficiently transferred.

Manufacturing Notes:

- Concentration is based upon gravimetric preparation using either a balance whose calibration has been verified daily using NIST traceable weights, and/or dilutions with Class A glassware.

Handling Notes:

- Stability of the unopened product, when stored in compliance with the recommended conditions, is guaranteed through the expiration displayed on the product label and certificate. Contact Restek for additional opened product stability information, with the knowledge/understanding that open product stability is subject to the specific handling and environmental conditions to which the product is exposed. For your convenience Restek supplies deactivated vials with most standards packed in 2mL ampuls. Larger volume deactivated vials are available through Restek as a custom ordered item. Additionally, Restek sells DMDCS for the purpose of glassware deactivation as catalog number 31861, which includes complete instructions.
- If any undissolved material is visible inside the ampul, sonicate the unopened ampul until the material is completely dissolved.



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

gravimetric



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 555224 **Lot No.:** A0214017

Description : Custom 8270 Plus Standard #2

Custom 8270 Plus Standard #2 1,000µg/mL, Methylene Chloride, 1mL/ampul

Container Size : 2 mL **Pkg Amt:** > 1 mL

Expiration Date : July 31, 2026 **Storage:** 10°C or colder

Ship: Ambient

CERTIFIED VALUES

Component #	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	1,2,4,5-Tetrachlorobenzene	95-94-3	MKCT9480	99%	1,005.0 µg/mL	+/- 29.541899
2	Acetophenone	98-86-2	STBH8205	99%	1,005.0 µg/mL	+/- 29.541899
3	Benzaldehyde	100-52-7	RD231129RSRA	99%	1,008.0 µg/mL	+/- 29.630084
4	Benzoic acid	65-85-0	MKCR2694	99%	1,010.0 µg/mL	+/- 29.688874
5	Biphenyl	92-52-4	MKCS5928	99%	1,008.0 µg/mL	+/- 29.630084

Solvent: Methylene chloride
CAS # 75-09-2
Purity 99%

S12509 } RC/
↓
S12568 } 7/24/24


Jess Hoy - Operations Tech I

Date Mixed: 18-Jul-2024

Balance: 1128360905

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397

General Certified Reference Material Notes

Expiration Notes:

- Expiration date valid for unopened ampul stored in compliance with the recommended conditions.
- Uncertainty, concentration, and expiration of the CRM are based on the unopened product being stored according to the recommended condition found in the storage field.

Purity Notes:

- Purity and/or chemical identity are determined by one or more of the following techniques: GC/FID, HPLC, GC/μECD, GC/MS, LC/MS, RI, and/or melting point.
- Compounds with a listed purity of less than 99% have been weight corrected to compensate for impurities and/or salts. A correction factor is used to calculate the amount of compound necessary to achieve the desired concentration of the parent compound in solution.
- Purity of isomeric compounds is reported as the sum of the isomers.
- Purity values are rounded to the nearest whole number.

Certified Uncertainty Value Notes:

- The uncertainties are determined in accordance with ISO 17034 and Guide 35. The certified expanded uncertainty value includes gravimetric uncertainty, homogeneity between-ampul uncertainty, storage stability uncertainty and shipping stability uncertainty and were combined using the following formula:

$$U_{combined\ uncertainty} = k \sqrt{u_{gravimetric}^2 + u_{homogeneity}^2 + u_{storage\ stability}^2 + u_{shipping\ stability}^2}$$

k is a coverage factor of 2, which gives a level of confidence of approximately 95%.

- The packaged amount is the minimum sample size for which uncertainty is valid. The ampuls are over-filled to ensure that the minimum packaged amount can be sufficiently transferred.

Manufacturing Notes:

- Concentration is based upon gravimetric preparation using either a balance whose calibration has been verified daily using NIST traceable weights, and/or dilutions with Class A glassware.

Handling Notes:

- Stability of the unopened product, when stored in compliance with the recommended conditions, is guaranteed through the expiration displayed on the product label and certificate. Contact Restek for additional opened product stability information, with the knowledge/understanding that open product stability is subject to the specific handling and environmental conditions to which the product is exposed. For your convenience Restek supplies deactivated vials with most standards packed in 2mL ampuls. Larger volume deactivated vials are available through Restek as a custom ordered item. Additionally, Restek sells DMDCS for the purpose of glassware deactivation as catalog number 31861, which includes complete instructions.
- If any undissolved material is visible inside the ampul, sonicate the unopened ampul until the material is completely dissolved.



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

chromatographic plus



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 31615 **Lot No.:** A0212955
Description : GC/MS Tuning Mixture
GC/MS Tuning Mixture 1,000µg/mL, Methylene Chloride, 1mL/ampul
Container Size : 2 mL **Pkg Amt:** > 1 mL
Expiration Date : June 30, 2027 **Storage:** 10°C or colder
Handling: Contains carcinogen/reproductive toxin. **Ship:** Ambient

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	Pentachlorophenol	87-86-5	RP240517RSR	99%	1,004.5 µg/mL	+/- 44.8902
2	DFTPP (Decafluorotriphenylphosphine)	5074-71-5	Q117-147	99%	1,004.5 µg/mL	+/- 44.8902
3	Benzidine	92-87-5	S240430RSR	99%	1,006.0 µg/mL	+/- 44.9572
4	4,4'-DDT	50-29-3	S240530RSR	97%	1,000.1 µg/mL	+/- 44.6922

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: Methylene chloride
CAS # 75-09-2
Purity 99%

S12577 } RC
↓
S12579 } 8/2/24

Quality Confirmation Test

Column:

30m x 0.25mm x 0.25µm
Rtx-5 (cat.#10223)

Carrier Gas:

hydrogen-constant pressure 10 psi.

Temp. Program:

75°C (hold 1 min.) to 330°C
@ 20°C/min. (hold 10 min.)

Inj. Temp:

250°C

Det. Temp:

330°C

Det. Type:


FID

Split Vent:

10 ml/min.

Inj. Vol

1µl




This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.


Ethan Winiarski - Operations Tech I

Date Mixed: 19-Jun-2024

Balance Serial # 1128353505


Jennifer Pollino - Operations Tech III - ARM QC

Date Passed: 26-Jun-2024

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

chromatographic plus



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 31206 **Lot No.:** A0212266

Description : SV Internal Standard Mix 2mg/ml

SV Internal Standard Mix 2mg/ml 2000 µg/ml, Methylene Chloride, 1mL/ampul

Container Size : 2 mL **Pkg Amt:** > 1 mL

Expiration Date : April 30, 2030 **Storage:** 10°C or colder

Handling: Sonication required. Mix is photosensitive. **Ship:** Ambient

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	1,4-Dichlorobenzene-d4	3855-82-1	PR-30447	99%	2,000.6 µg/mL	+/- 90.1075
2	Naphthalene-d8	1146-65-2	M-2180	99%	2,000.3 µg/mL	+/- 90.0925
3	Acenaphthene-d10	15067-26-2	PR-33507	99%	2,000.4 µg/mL	+/- 90.1000
4	Phenanthrene-d10	1517-22-2	PR-34099	99%	2,000.5 µg/mL	+/- 90.1037
5	Chrysene-d12	1719-03-5	PR-33506	99%	2,000.7 µg/mL	+/- 90.1112
6	Perylene-d12	1520-96-3	PR-33205	99%	2,000.6 µg/mL	+/- 90.1075

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: Methylene chloride
CAS # 75-09-2
Purity 99%

S12645
↓
S12674 } AC
10/1/24



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

chromatographic plus



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 31850 **Lot No.:** A0219438

Description : 8270 MegaMix®
8270 MegaMix® 500-1000 µg/mL, Methylene Chloride, 1mL/ampul

Container Size : 2 mL **Pkg Amt:** > 1 mL

Expiration Date : September 30, 2025 **Storage:** 0°C or colder

Handling: Sonication required. Mix is photosensitive. **Ship:** Ambient

S12963
↓
S12992 } AC
12/17/24

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	Pyridine	110-86-1	SHBP6240	99%	1,008.3 µg/mL	+/- 36.6849
2	N-Nitrosodimethylamine	62-75-9	S240313RSR	99%	1,008.6 µg/mL	+/- 36.6985
3	Phenol	108-95-2	MKCK1120	99%	1,003.5 µg/mL	+/- 36.5120
4	Aniline	62-53-3	X22F726	99%	1,002.9 µg/mL	+/- 36.4893
5	Bis(2-chloroethyl)ether	111-44-4	002891T24M	99%	1,003.0 µg/mL	+/- 36.4938
6	2-Chlorophenol	95-57-8	STBJ3909	99%	1,005.6 µg/mL	+/- 36.5894
7	1,3-Dichlorobenzene	541-73-1	BCCD5315	99%	1,004.1 µg/mL	+/- 36.5348
8	1,4-Dichlorobenzene	106-46-7	MKBS7929V	99%	1,002.1 µg/mL	+/- 36.4620
9	Benzyl alcohol	100-51-6	SHBK5469	99%	1,003.5 µg/mL	+/- 36.5120
10	1,2-Dichlorobenzene	95-50-1	SHBL6287	99%	1,005.3 µg/mL	+/- 36.5757
11	2-Methylphenol (o-cresol)	95-48-7	SHBN7598	99%	1,008.4 µg/mL	+/- 36.6894
12	2,2'-oxybis(1-chloropropane)	108-60-1	29-MAR-45-5	99%	1,004.6 µg/mL	+/- 36.5530
13	3-Methylphenol (m-cresol)	108-39-4	STBJ0710	99%	502.1 µg/mL	+/- 18.2697
14	4-Methylphenol (p-cresol)	106-44-5	SHBN3411	99%	503.8 µg/mL	+/- 18.3288
15	N-Nitroso-di-n-propylamine	621-64-7	N63MG	99%	1,006.5 µg/mL	+/- 36.6212
16	Hexachloroethane	67-72-1	DAXRI	99%	1,004.5 µg/mL	+/- 36.5484
17	Nitrobenzene	98-95-3	10224044	99%	1,002.5 µg/mL	+/- 36.4757

18	Isophorone	78-59-1	MKCR3249	99%	1,003.4	µg/mL	+/- 36.5075
19	2-Nitrophenol	88-75-5	RP230710	99%	1,002.5	µg/mL	+/- 36.4757
20	2,4-Dimethylphenol	105-67-9	XW5GK	99%	1,006.5	µg/mL	+/- 36.6212
21	Bis(2-chloroethoxy)methane	111-91-1	15705100	99%	1,006.6	µg/mL	+/- 36.6257
22	2,4-Dichlorophenol	120-83-2	BCKK6969	99%	1,001.5	µg/mL	+/- 36.4393
23	1,2,4-Trichlorobenzene	120-82-1	SHBP5900	99%	1,006.4	µg/mL	+/- 36.6166
24	Naphthalene	91-20-3	STBL1057	99%	1,002.1	µg/mL	+/- 36.4620
25	4-Chloroaniline	106-47-8	BCCJ3217	99%	1,004.4	µg/mL	+/- 36.5439
26	Hexachlorobutadiene	87-68-3	X05J	98%	1,002.5	µg/mL	+/- 36.4771
27	4-Chloro-3-methylphenol	59-50-7	BCCD4461	99%	1,004.5	µg/mL	+/- 36.5484
28	2-Methylnaphthalene	91-57-6	STBL3028	99%	1,000.0	µg/mL	+/- 36.3847
29	1-Methylnaphthalene	90-12-0	5234.00-8	98%	990.2	µg/mL	+/- 36.0269
30	Hexachlorocyclopentadiene	77-47-4	099063I14L	98%	1,001.3	µg/mL	+/- 36.4325
31	2,4,6-Trichlorophenol	88-06-2	STBK8870	99%	1,006.4	µg/mL	+/- 36.6166
32	2,4,5-Trichlorophenol	95-95-4	3YFRE	97%	1,004.6	µg/mL	+/- 36.5505
33	2-Chloronaphthalene	91-58-7	RPN7O	99%	1,004.3	µg/mL	+/- 36.5393
34	2-Nitroaniline	88-74-4	RP240715RSR	99%	1,004.4	µg/mL	+/- 36.5439
35	1,4-Dinitrobenzene	100-25-4	RP240703RSR	99%	1,002.8	µg/mL	+/- 36.4847
36	Acenaphthylene	208-96-8	RP241029RSR	98%	1,000.0	µg/mL	+/- 36.3835
37	1,3-Dinitrobenzene	99-65-0	TRC3-1075941-2-1	99%	1,006.3	µg/mL	+/- 36.6121
38	Dimethylphthalate	131-11-3	358221L17K	99%	1,008.9	µg/mL	+/- 36.7076
39	2,6-Dinitrotoluene	606-20-2	BCCG1833	99%	1,006.6	µg/mL	+/- 36.6257
40	1,2-Dinitrobenzene	528-29-0	RP240701RSR	99%	1,002.5	µg/mL	+/- 36.4757
41	Acenaphthene	83-32-9	MKCR7169	99%	1,000.0	µg/mL	+/- 36.3847
42	3-Nitroaniline	99-09-2	RP240708RSR	99%	1,004.6	µg/mL	+/- 36.5530
43	2,4-Dinitrophenol	51-28-5	D240927RSR	----%	1,005.6	µg/mL	+/- 36.5894
44	Dibenzofuran	132-64-9	MKCN1772	99%	1,003.5	µg/mL	+/- 36.5120
45	2,4-Dinitrotoluene	121-14-2	102869V26E	99%	1,008.3	µg/mL	+/- 36.6849
46	4-Nitrophenol	100-02-7	20241029-2-AN	99%	1,004.8	µg/mL	+/- 36.5575
47	2,3,4,6-Tetrachlorophenol	58-90-2	PR-34476	99%	1,005.8	µg/mL	+/- 36.5939
48	2,3,5,6-Tetrachlorophenol	935-95-5	RP231219RSR	99%	1,006.4	µg/mL	+/- 36.6166
49	Fluorene	86-73-7	10246250	98%	1,000.7	µg/mL	+/- 36.4102
50	4-Chlorophenyl phenyl ether	7005-72-3	MKCT7248	99%	1,004.9	µg/mL	+/- 36.5621
51	Diethylphthalate	84-66-2	BCCJ6241	99%	1,003.9	µg/mL	+/- 36.5257
52	4-Nitroaniline	100-01-6	RP230111	99%	1,006.6	µg/mL	+/- 36.6257
53	4,6-Dinitro-2-methylphenol (Dinitro-o-cresol)	534-52-1	S241008RSR	99%	1,001.3	µg/mL	+/- 36.4302

54	Diphenylamine	122-39-4	MKCT1512	99%	1,003.0	µg/mL	+/- 36.4938
55	Azobenzene	103-33-3	BCKK0887	99%	1,002.4	µg/mL	+/- 36.4711
56	4-Bromophenyl phenyl ether	101-55-3	STBH6361	99%	1,008.8	µg/mL	+/- 36.7031
57	Hexachlorobenzene	118-74-1	15458400	99%	1,005.1	µg/mL	+/- 36.5712
58	Pentachlorophenol	87-86-5	RP240517RSR	99%	1,005.9	µg/mL	+/- 36.5984
59	Phenanthrene	85-01-8	MKCT3391	99%	1,004.9	µg/mL	+/- 36.5621
60	Anthracene	120-12-7	101492T18R	99%	1,005.1	µg/mL	+/- 36.5712
61	Carbazole	86-74-8	15276700	99%	1,005.4	µg/mL	+/- 36.5803
62	Di-n-butylphthalate	84-74-2	MKCN4337	99%	1,006.3	µg/mL	+/- 36.6121
63	Fluoranthene	206-44-0	MKCQ4728	99%	1,003.5	µg/mL	+/- 36.5120
64	Pyrene	129-00-0	BCKK2592	99%	1,002.0	µg/mL	+/- 36.4575
65	Benzyl butyl phthalate	85-68-7	X12I018	99%	1,007.5	µg/mL	+/- 36.6576
66	Bis(2-ethylhexyl)adipate	103-23-1	MKCM1988	99%	1,005.9	µg/mL	+/- 36.5984
67	Benz(a)anthracene	56-55-3	I70012022BAA	99%	1,005.5	µg/mL	+/- 36.5848
68	Chrysene	218-01-9	RP241007RSR	99%	1,005.3	µg/mL	+/- 36.5757
69	Bis(2-ethylhexyl)phthalate	117-81-7	MKCS8065	99%	1,007.5	µg/mL	+/- 36.6576
70	Di-n-octyl phthalate	117-84-0	15566400	99%	1,002.3	µg/mL	+/- 36.4666
71	Benzo(b)fluoranthene	205-99-2	052013B	99%	1,004.1	µg/mL	+/- 36.5348
72	Benzo(k)fluoranthene	207-08-9	012022K	99%	1,002.8	µg/mL	+/- 36.4847
73	Benzo(a)pyrene	50-32-8	NQLXA	98%	1,006.2	µg/mL	+/- 36.6108
74	Indeno(1,2,3-cd)pyrene	193-39-5	12-JKL-118-9	97%	1,001.8	µg/mL	+/- 36.4490
75	Dibenz(a,h)anthracene	53-70-3	2-ASA-59-1	99%	1,003.3	µg/mL	+/- 36.5029
76	Benzo(g,h,i)perylene	191-24-2	RP241014RSR	98%	1,003.8	µg/mL	+/- 36.5217

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: Methylene chloride
CAS # 75-09-2
Purity 99%

Tech Tips:

N-Nitrosodiphenylamine (86-30-6) is prone to breakdown in the injection port and will be converted to Diphenylamine (122-39-4). When comparing the response of Diphenylamine to mixtures manufactured using N-Nitrosodiphenylamine, a difference in response will be observed. The ratio of the MW can be used to calculate the theoretical concentration of the N-Nitrosodiphenylamine.



5580 Skyline Blvd
Santa Rosa, CA 95403

(707)525-5788
(800)878-7654 Toll Free
(707)545-7901 Fax

Manufacturer's Quality System
Audited & Registered
by TUV USA to ISO 9001:2015

Date Received: _____

Certificate of Analysis

Rev 0

Page 1 of 1

Catalog No.:	Lot No.:	Storage:	Solvent:	Exp. Date:	Description:
Z-110816-01	531243	$\leq -10^{\circ}\text{C}$	Methylene Chloride	1/2/2030	Custom 8270 Mix, 4-79, 1000 mg/L, 1 mL

Compound	CAS No.	Purity (%)	Compound Lot No.	Concentration, mg/L
atrazine	1912-24-9	99.5	337.7.4P	997 $\pm$ 5.81
benzidine	92-87-5	99.9	124.18.6.2P	993.8 $\pm$ 5.78
caprolactam	105-60-2	99.9	271.1.6P	999 $\pm$ 5.82

513057
↓
513061 } AC
1/16/25

*Not a certified value

Manufactured by o2si smart solutions, Accredited to ISO 9001:2008 by NSF and ISO/IEC 17025:2005 (Certification No. 3031.01) and ISO Guide 34:2009 (Certification No. 3031.02) by A2LA

Certified By: _____

Melissa Workoff
Chemist

All weights are traceable through N. I. S. T. Test No. 822/264157-00.
Concentration (correct for purity) and uncertainty (95% confidence) values listed are determined gravimetrically.



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

chromatographic plus



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 31850 **Lot No.:** A0221014

Description : 8270 MegaMix®

8270 MegaMix® 500-1000 µg/mL, Methylene Chloride, 1mL/ampul

Container Size : 2 mL **Pkg Amt:** > 1 mL

Expiration Date : November 30, 2025 **Storage:** 0°C or colder

Handling: Sonication required. Mix is photosensitive. **Ship:** Ambient

513088 }
↓
513117 } 24
5/20/25.

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	Pyridine	110-86-1	SHBR4811	99%	1,005.2 µg/mL	+/- 36.5730
2	N-Nitrosodimethylamine	62-75-9	S241226RSR	99%	1,005.5 µg/mL	+/- 36.5848
3	Phenol	108-95-2	MKCK1120	99%	1,005.3 µg/mL	+/- 36.5780
4	Aniline	62-53-3	X22F726	99%	1,005.4 µg/mL	+/- 36.5816
5	Bis(2-chloroethyl)ether	111-44-4	002891T24M	99%	1,004.7 µg/mL	+/- 36.5562
6	2-Chlorophenol	95-57-8	STBJ3909	99%	1,005.3 µg/mL	+/- 36.5766
7	1,3-Dichlorobenzene	541-73-1	BCCD5315	99%	1,004.6 µg/mL	+/- 36.5530
8	1,4-Dichlorobenzene	106-46-7	MKBS7929V	99%	1,006.3 µg/mL	+/- 36.6125
9	Benzyl alcohol	100-51-6	SHBK5469	99%	1,005.5 µg/mL	+/- 36.5853
10	1,2-Dichlorobenzene	95-50-1	SHBL6287	99%	1,004.6 µg/mL	+/- 36.5507
11	2-Methylphenol (o-cresol)	95-48-7	SHBN7598	99%	1,004.8 µg/mL	+/- 36.5575
12	2,2'-oxybis(1-chloropropane)	108-60-1	29-MAR-45-5	99%	1,005.4 µg/mL	+/- 36.5803
13	3-Methylphenol (m-cresol)	108-39-4	STBL3873	99%	502.7 µg/mL	+/- 18.2888
14	4-Methylphenol (p-cresol)	106-44-5	SHBN3411	99%	502.6 µg/mL	+/- 18.2869
15	N-Nitroso-di-n-propylamine	621-64-7	N63MG	99%	1,004.6 µg/mL	+/- 36.5502
16	Hexachloroethane	67-72-1	DAXRI	99%	1,004.6 µg/mL	+/- 36.5530
17	Nitrobenzene	98-95-3	10224044	99%	1,005.3 µg/mL	+/- 36.5780

18	Isophorone	78-59-1	MKCR3249	99%	1,004.6	µg/mL	+/- 36.5511
19	2-Nitrophenol	88-75-5	RP230710	99%	1,005.7	µg/mL	+/- 36.5916
20	2,4-Dimethylphenol	105-67-9	DIRAF	99%	1,004.9	µg/mL	+/- 36.5612
21	Bis(2-chloroethoxy)methane	111-91-1	15705100	99%	1,004.3	µg/mL	+/- 36.5402
22	2,4-Dichlorophenol	120-83-2	BCCK6969	99%	1,004.7	µg/mL	+/- 36.5571
23	1,2,4-Trichlorobenzene	120-82-1	SHBP5900	99%	1,004.6	µg/mL	+/- 36.5516
24	Naphthalene	91-20-3	STBL1057	99%	1,006.8	µg/mL	+/- 36.6335
25	4-Chloroaniline	106-47-8	BCCJ3217	99%	1,005.9	µg/mL	+/- 36.5980
26	Hexachlorobutadiene	87-68-3	X05J	98%	1,004.1	µg/mL	+/- 36.5328
27	4-Chloro-3-methylphenol	59-50-7	BCCD4461	99%	1,005.4	µg/mL	+/- 36.5793
28	2-Methylnaphthalene	91-57-6	STBL3028	99%	1,006.3	µg/mL	+/- 36.6144
29	1-Methylnaphthalene	90-12-0	5234.00-8	98%	990.2	µg/mL	+/- 36.0269
30	Hexachlorocyclopentadiene	77-47-4	099063P13G	99%	1,005.9	µg/mL	+/- 36.5975
31	2,4,6-Trichlorophenol	88-06-2	STBK8870	99%	1,004.7	µg/mL	+/- 36.5566
32	2,4,5-Trichlorophenol	95-95-4	3YFRE	97%	1,004.7	µg/mL	+/- 36.5571
33	2-Chloronaphthalene	91-58-7	RPN7O	99%	1,005.3	µg/mL	+/- 36.5757
34	2-Nitroaniline	88-74-4	RP240715RSR	99%	1,004.8	µg/mL	+/- 36.5584
35	1,4-Dinitrobenzene	100-25-4	RP240703RSR	99%	1,004.5	µg/mL	+/- 36.5475
36	Acenaphthylene	208-96-8	214935V18H	95%	1,000.1	µg/mL	+/- 36.3888
37	1,3-Dinitrobenzene	99-65-0	TRC3-1075941-2-1	99%	1,004.9	µg/mL	+/- 36.5616
38	Dimethylphthalate	131-11-3	358221L17K	99%	1,004.5	µg/mL	+/- 36.5489
39	2,6-Dinitrotoluene	606-20-2	BCCG1833	99%	1,005.2	µg/mL	+/- 36.5734
40	1,2-Dinitrobenzene	528-29-0	RP240701RSR	99%	1,005.9	µg/mL	+/- 36.6003
41	Acenaphthene	83-32-9	MKCR7169	99%	1,000.0	µg/mL	+/- 36.3847
42	3-Nitroaniline	99-09-2	RP240708RSR	99%	1,004.6	µg/mL	+/- 36.5525
43	2,4-Dinitrophenol	51-28-5	D240927RSR	99%	1,005.0	µg/mL	+/- 36.5657
44	Dibenzofuran	132-64-9	MKCN1772	99%	1,005.4	µg/mL	+/- 36.5803
45	2,4-Dinitrotoluene	121-14-2	102869V26E	99%	1,005.4	µg/mL	+/- 36.5803
46	4-Nitrophenol	100-02-7	20241120-1-AN	99%	1,005.7	µg/mL	+/- 36.5930
47	2,3,4,6-Tetrachlorophenol	58-90-2	PR-34476	99%	1,004.9	µg/mL	+/- 36.5616
48	2,3,5,6-Tetrachlorophenol	935-95-5	RP240523RSR	99%	1,005.2	µg/mL	+/- 36.5739
49	Fluorene	86-73-7	10246250	98%	1,005.8	µg/mL	+/- 36.5948
50	4-Chlorophenyl phenyl ether	7005-72-3	002531K02D	99%	1,004.8	µg/mL	+/- 36.5584
51	Diethylphthalate	84-66-2	STBL3611	99%	1,005.3	µg/mL	+/- 36.5762
52	4-Nitroaniline	100-01-6	RP240830RSR	99%	1,005.5	µg/mL	+/- 36.5844
53	4,6-Dinitro-2-methylphenol (Dinitro-o-cresol)	534-52-1	S241008RSR	99%	1,004.8	µg/mL	+/- 36.5589

54	Diphenylamine	122-39-4	MKCT1512	99%	1,005.2	µg/mL	+/- 36.5725
55	Azobenzene	103-33-3	BCCL3292	99%	1,004.7	µg/mL	+/- 36.5566
56	4-Bromophenyl phenyl ether	101-55-3	STBH6361	99%	1,005.6	µg/mL	+/- 36.5875
57	Hexachlorobenzene	118-74-1	15828800	99%	1,005.3	µg/mL	+/- 36.5789
58	Pentachlorophenol	87-86-5	RP240517RSR	99%	1,005.4	µg/mL	+/- 36.5825
59	Phenanthrene	85-01-8	MKCT3391	99%	1,004.8	µg/mL	+/- 36.5584
60	Anthracene	120-12-7	MKCW9141	99%	1,005.5	µg/mL	+/- 36.5834
61	Carbazole	86-74-8	15630800	99%	1,005.3	µg/mL	+/- 36.5789
62	Di-n-butylphthalate	84-74-2	MKCN4337	99%	1,005.5	µg/mL	+/- 36.5848
63	Fluoranthene	206-44-0	A0458721	99%	1,005.4	µg/mL	+/- <0.0001
64	Pyrene	129-00-0	BCCL8032	99%	1,005.2	µg/mL	+/- 36.5734
65	Benzyl butyl phthalate	85-68-7	X12I018	99%	1,004.8	µg/mL	+/- 36.5584
66	Bis(2-ethylhexyl)adipate	103-23-1	MKCR8567	99%	1,004.5	µg/mL	+/- 36.5480
67	Benz(a)anthracene	56-55-3	I60012022BAA	99%	1,005.5	µg/mL	+/- 36.5844
68	Chrysene	218-01-9	RP241212RSR	99%	1,005.5	µg/mL	+/- 36.5848
69	Bis(2-ethylhexyl)phthalate	117-81-7	MKCS8065	99%	1,004.7	µg/mL	+/- 36.5548
70	Di-n-octyl phthalate	117-84-0	15817300	99%	1,004.8	µg/mL	+/- 36.5607
71	Benzo(b)fluoranthene	205-99-2	022013B	99%	1,005.1	µg/mL	+/- 36.5716
72	Benzo(k)fluoranthene	207-08-9	012022K	98%	1,004.6	µg/mL	+/- 36.5511
73	Benzo(a)pyrene	50-32-8	NQLXA	98%	1,004.2	µg/mL	+/- 36.5377
74	Indeno(1,2,3-cd)pyrene	193-39-5	12-JKL-118-9	97%	1,004.1	µg/mL	+/- 36.5337
75	Dibenz(a,h)anthracene	53-70-3	712061504-1-1	99%	1,005.7	µg/mL	+/- 36.5930
76	Benzo(g,h,i)perylene	191-24-2	RP241014RSR	98%	1,005.4	µg/mL	+/- 36.5814

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: Methylene chloride
CAS # 75-09-2
Purity 99%

Tech Tips:

N-Nitrosodiphenylamine (86-30-6) is prone to breakdown in the injection port and will be converted to Diphenylamine (122-39-4). When comparing the response of Diphenylamine to mixtures manufactured using N-Nitrosodiphenylamine, a difference in response will be observed. The ratio of the MW can be used to calculate the theoretical concentration of the N-Nitrosodiphenylamine.

Quality Confirmation Test

Column:

30m x 0.25mm x 0.25µm
Rtx-5 (cat.#10223)

Carrier Gas:

hydrogen-constant flow 1.8 mL/min.

Temp. Program:

80°C (hold 0.1 min.) to 330°C
@ 9.6°C/min. (hold 2.86 min.)

Inj. Temp:

250°C

Det. Temp:

340°C

Det. Type:


FID

Split Vent:


100 ml/min.

Inj. Vol

1µl




This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.


Penelope Riglin - Operations Tech I

Date Mixed: 12-Jan-2025

Balance Serial # 1128360905


Dillan Murphy - Operations Technician I

Date Passed: 21-Jan-2025

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

chromatographic plus



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 31853 **Lot No.:** A0218894
Description : 1,4-dioxane
1,4-Dioxane 2,000µg/mL, Methylene Chloride, 1mL/ampul
Container Size : 2 mL **Pkg Amt:** > 1 mL
Expiration Date : November 30, 2029 **Storage:** 0°C or colder
Ship: Ambient

513118 } RC/
↓
513147 } 5/20/25.

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	1,4-Dioxane	123-91-1	SHBQ1693	99%	2,002.4 µg/mL	+/- 24.9202

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: Methylene chloride
CAS # 75-09-2
Purity 99%

Quality Confirmation Test

Column:

30m x 0.25mm x 0.25µm
Rtx-5 (cat.#10223)

Carrier Gas:

hydrogen-constant flow 1.8 mL/min.

Temp. Program:

80°C (hold 0.1 min.) to 330°C
@ 9.6°C/min. (hold 2.86 min.)

Inj. Temp:

250°C

Det. Temp:

340°C

Det. Type:


FID

Split Vent:

100 mL/min.

Inj. Vol

1µl



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.

Penelope Riglin - Operations Tech I

Date Mixed: 07-Nov-2024

Balance Serial # 1128360905

Dillan Murphy - Operations Technician I

Date Passed: 11-Nov-2024

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

gravimetric



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 555223 **Lot No.:** A0228451

Description : Custom 8270 Plus Standard #1
Custom 8270 Plus Standard #1 1,000µg/mL, Methylene Chloride, 1mL/ampul

Container Size : 2 mL **Pkg Amt:** > 1 mL

Expiration Date : August 31, 2027 **Storage:** 10°C or colder

Handling: This product is photosensitive. **Ship:** Ambient

S13239 } RC/
↓
S13268 } 08/06/25

CERTIFIED VALUES

Component #	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	3,3'-Dichlorobenzidine	91-94-1	S250226RSR	99%	1,006.0 µg/mL	+/- 23.0947
2	Atrazine	1912-24-9	5FYWL	99%	1,007.0 µg/mL	+/- 23.1176
3	Benzidine	92-87-5	S250227ECS	99%	1,003.0 µg/mL	+/- 23.0258
4	epsilon-Caprolactam	105-60-2	Y16H012	99%	1,003.0 µg/mL	+/- 23.0258

Solvent: Methylene chloride
CAS # 75-09-2
Purity 99%

Tom Sucka - Mix Technician

Date Mixed: 01-Aug-2025

Balance: 1128360905

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397



SHIPPING DOCUMENTS

Q3292

BILLING INFORMATION

STATE: ZIP:
PHONE:

DATA DELIVERABLE INFORMATION

STANDARD TURNAROUND TIME IS 10 BUSINESS DAYS

1	1,4-Dioxane SW846 8270 SIM
2	Iron, Total
3	TSS
4	TDS
5	
6	
7	
8	
9	

COMMENTS

```
<-- Specify Preservatives
A-HCl      B-HNO3
```

1

7.1

1

1001

1

1

Figure 1

of the

☐ YES ☐ NO

WHITE - CHEMTECH COPY FOR RETURN TO CLIENT YELLOW - CHEMTECH COPY PINK - SAMPLER COPY

☐ YES ☐ NO

Laboratory Certification

Certified By	License No.
Connecticut	PH-0830
DOD ELAP (ANAB)	L2219
Maine	2024021
Maryland	296
New Hampshire	255425
New Jersey	20012
New York	11376
Pennsylvania	68-00548
Soil Permit	525-24-234-08441
Texas	TX-C25-00189
Virginia	460312

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN101325\
 Data File : BN038066.D
 Acq On : 13 Oct 2025 13:47
 Operator : RC/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SSTDCCC0.4

Quant Time: Oct 13 14:16:09 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN092325.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Sep 24 11:00:01 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.797	152	7166	0.400	ng	-0.03
7) Naphthalene-d8	10.594	136	18651	0.400	ng	#-0.03
13) Acenaphthene-d10	14.452	164	10244	0.400	ng	-0.02
19) Phenanthrene-d10	17.198	188	18901	0.400	ng	#-0.02
29) Chrysene-d12	21.393	240	16827	0.400	ng	0.00
35) Perylene-d12	23.712	264	17326	0.400	ng	-0.01
System Monitoring Compounds						
4) 2-Fluorophenol	5.370	112	6402	0.391	ng	-0.01
5) Phenol-d6	6.959	99	6728	0.313	ng	-0.01
8) Nitrobenzene-d5	8.950	82	4437	0.312	ng	-0.02
11) 2-Methylnaphthalene-d10	12.192	152	10642	0.411	ng	-0.03
14) 2,4,6-Tribromophenol	15.945	330	1404	0.361	ng	-0.02
15) 2-Fluorobiphenyl	13.073	172	14631	0.349	ng	-0.02
27) Fluoranthene-d10	19.239	212	20125	0.423	ng	-0.01
31) Terphenyl-d14	19.838	244	12185	0.364	ng	-0.01
Target Compounds						
						Qvalue
2) 1,4-Dioxane	3.283	88	3335	0.458	ng	94
3) n-Nitrosodimethylamine	3.586	42	3720	0.384	ng	# 92
6) bis(2-Chloroethyl)ether	7.219	93	7145	0.358	ng	97
9) Naphthalene	10.648	128	19331	0.376	ng	100
10) Hexachlorobutadiene	10.947	225	3527	0.402	ng	# 100
12) 2-Methylnaphthalene	12.268	142	12367	0.368	ng	99
16) Acenaphthylene	14.174	152	16983	0.365	ng	100
17) Acenaphthene	14.516	154	11898	0.359	ng	98
18) Fluorene	15.499	166	12971	0.324	ng	99
20) 4,6-Dinitro-2-methylph...	15.572	198	741	0.373	ng	# 10
21) 4-Bromophenyl-phenylether	16.404	248	4416	0.359	ng	# 82
22) Hexachlorobenzene	16.515	284	5456	0.365	ng	99
23) Atrazine	16.664	200	3115	0.332	ng	# 92
24) Pentachlorophenol	16.851	266	2441	0.476	ng	95
25) Phenanthrene	17.248	178	21937	0.353	ng	100
26) Anthracene	17.335	178	18915	0.351	ng	100
28) Fluoranthene	19.266	202	24986	0.365	ng	100
30) Pyrene	19.629	202	25549	0.385	ng	100
32) Benzo(a)anthracene	21.375	228	21568	0.367	ng	99
33) Chrysene	21.429	228	24503	0.385	ng	99
34) Bis(2-ethylhexyl)phtha...	21.313	149	15324	0.659	ng	97
36) Indeno(1,2,3-cd)pyrene	26.080	276	30553	0.386	ng	98
37) Benzo(b)fluoranthene	23.057	252	26813	0.359	ng	96
38) Benzo(k)fluoranthene	23.057	252	26813	0.357	ng	96
39) Benzo(a)pyrene	23.613	252	21668	0.367	ng	# 89
40) Dibenzo(a,h)anthracene	26.098	278	22721	0.368	ng	98
41) Benzo(g,h,i)perylene	26.800	276	27477	0.423	ng	99


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN101325\
Data File : BN038066.D
Acq On : 13 Oct 2025 13:47
Operator : RC/JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4

Quant Time: Oct 13 14:16:09 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN092325.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Sep 24 11:00:01 2025
Response via : Initial Calibration









#1
1,4-Dichlorobenzene-d4
Concen: 0.400 ng
RT: 7.797 min Scan# 61
Delta R.T. -0.028 min
Lab File: BN038066.D
Acq: 13 Oct 2025 13:47




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion:152 Resp: 7166
Ion Ratio Lower Upper
152 100
150 152.5 121.0 181.4
115 61.0 47.4 71.0



#2
1,4-Dioxane
Concen: 0.458 ng
RT: 3.283 min Scan# 28
Delta R.T. -0.014 min
Lab File: BN038066.D
Acq: 13 Oct 2025 13:47

Tgt Ion: 88 Resp: 3335
Ion Ratio Lower Upper
88 100
43 30.0 26.6 39.8
58 68.4 58.9 88.3









#3
 n-Nitrosodimethylamine
 Concen: 0.384 ng
 RT: 3.586 min Scan# 70
 Delta R.T. -0.014 min
 Lab File: BN038066.D
 Acq: 13 Oct 2025 13:47




Instrument :
 BNA_N
 ClientSampleId :
 SSTDCCC0.4


Tgt Ion: 42 Resp: 3720
 Ion Ratio Lower Upper
 42 100
 74 119.3 93.4 140.0
 44 11.7 22.2 33.4#



#4
 2-Fluorophenol
 Concen: 0.391 ng
 RT: 5.370 min Scan# 317
 Delta R.T. -0.014 min
 Lab File: BN038066.D
 Acq: 13 Oct 2025 13:47


Tgt Ion: 112 Resp: 6402
 Ion Ratio Lower Upper
 112 100
 64 55.7 44.6 66.8
 63 29.4 24.9 37.3








#5
Phenol-d6
Concen: 0.313 ng
RT: 6.959 min Scan# 51
Delta R.T. -0.014 min
Lab File: BN038066.D
Acq: 13 Oct 2025 13:47

Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4



Tgt Ion: 99 Resp: 6728




Ion	Ratio	Lower	Upper
99	100		
42	23.3	17.2	25.8
71	33.2	27.3	40.9




#6
bis(2-Chloroethyl)ether
Concen: 0.358 ng
RT: 7.219 min Scan# 573
Delta R.T. -0.022 min
Lab File: BN038066.D
Acq: 13 Oct 2025 13:47

Tgt Ion: 93 Resp: 7145

Ion	Ratio	Lower	Upper
93	100		
63	78.0	60.1	90.1
95	33.0	25.4	38.2









#7
Naphthalene-d8
Concen: 0.400 ng
RT: 10.594 min Scan# 91
Delta R.T. -0.032 min
Lab File: BN038066.D
Acq: 13 Oct 2025 13:47




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion:	136	Resp:	18651
Ion Ratio	Lower	Upper	
136	100		
137	11.4	9.0	13.4
54	10.1	5.8	8.8#
68	7.9	4.3	6.5#



#8
Nitrobenzene-d5
Concen: 0.312 ng
RT: 8.950 min Scan# 784
Delta R.T. -0.021 min
Lab File: BN038066.D
Acq: 13 Oct 2025 13:47

Tgt Ion:	82	Resp:	4437
Ion Ratio	Lower	Upper	
82	100		
128	45.7	34.8	52.2
54	54.6	42.2	63.2









#9
Naphthalene
Concen: 0.376 ng
RT: 10.648 min Scan# 945
Delta R.T. -0.021 min
Lab File: BN038066.D
Acq: 13 Oct 2025 13:47




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion:128	Resp:	19331
Ion Ratio	Lower	Upper
128	100	
129	11.6	9.2 13.8
127	13.6	11.0 16.6



#10
Hexachlorobutadiene
Concen: 0.402 ng
RT: 10.947 min Scan# 971
Delta R.T. -0.032 min
Lab File: BN038066.D
Acq: 13 Oct 2025 13:47

Tgt Ion:225	Resp:	3527
Ion Ratio	Lower	Upper
225	100	
223	0.0	0.0 0.0
227	64.5	51.4 77.2









#11
2-Methylnaphthalene-d10
Concen: 0.411 ng
RT: 12.192 min Scan# 1163
Delta R.T. -0.025 min
Lab File: BN038066.D
Acq: 13 Oct 2025 13:47




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion:152 Resp: 10642
Ion Ratio Lower Upper
152 100
151 21.2 17.4 26.2



#12
2-Methylnaphthalene
Concen: 0.368 ng
RT: 12.268 min Scan# 1173
Delta R.T. -0.025 min
Lab File: BN038066.D
Acq: 13 Oct 2025 13:47

Tgt Ion:142 Resp: 12367
Ion Ratio Lower Upper
142 100
141 88.5 71.2 106.8
115 35.9 29.7 44.5









#13
Acenaphthene-d10
Concen: 0.400 ng
RT: 14.452 min Scan# 1456
Delta R.T. -0.021 min
Lab File: BN038066.D
Acq: 13 Oct 2025 13:47




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion:164 Resp: 10244
Ion Ratio Lower Upper
164 100
162 105.5 84.8 127.2
160 53.4 46.1 69.1



#14
2,4,6-Tribromophenol
Concen: 0.361 ng
RT: 15.945 min Scan# 1588
Delta R.T. -0.024 min
Lab File: BN038066.D
Acq: 13 Oct 2025 13:47

Tgt Ion:330 Resp: 1404
Ion Ratio Lower Upper
330 100
332 98.0 77.2 115.8
141 40.4 37.2 55.8









#15
2-Fluorobiphenyl
Concen: 0.349 ng
RT: 13.073 min Scan# 11
Delta R.T. -0.021 min
Lab File: BN038066.D
Acq: 13 Oct 2025 13:47




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion:172	Resp:	14631
Ion Ratio	Lower	Upper
172	100	
171	35.7	28.6 43.0
170	23.8	19.7 29.5



#16
Acenaphthylene
Concen: 0.365 ng
RT: 14.174 min Scan# 1428
Delta R.T. -0.021 min
Lab File: BN038066.D
Acq: 13 Oct 2025 13:47

Tgt Ion:152	Resp:	16983
Ion Ratio	Lower	Upper
152	100	
151	19.7	15.8 23.8
153	12.8	10.5 15.7









#17
Acenaphthene
Concen: 0.359 ng
RT: 14.516 min Scan# 1460
Delta R.T. -0.021 min
Lab File: BN038066.D
Acq: 13 Oct 2025 13:47




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion:154 Resp: 11898
Ion Ratio Lower Upper
154 100
153 112.0 90.5 135.7
152 61.7 51.8 77.8



#18
Fluorene
Concen: 0.324 ng
RT: 15.499 min Scan# 1552
Delta R.T. -0.023 min
Lab File: BN038066.D
Acq: 13 Oct 2025 13:47

Tgt Ion:166 Resp: 12971
Ion Ratio Lower Upper
166 100
165 99.3 79.0 118.4
167 13.3 10.6 16.0









#19
Phenanthrene-d10
Concen: 0.400 ng
RT: 17.198 min Scan# 11
Delta R.T. -0.025 min
Lab File: BN038066.D
Acq: 13 Oct 2025 13:47




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion:188 Resp: 18901
Ion Ratio Lower Upper
188 100
94 0.0 0.0 0.0
80 13.9 7.6 11.4#



#20
4,6-Dinitro-2-methylphenol
Concen: 0.373 ng
RT: 15.572 min Scan# 1558
Delta R.T. -0.025 min
Lab File: BN038066.D
Acq: 13 Oct 2025 13:47

Tgt Ion:198 Resp: 741
Ion Ratio Lower Upper
198 100
51 176.1 59.1 88.7#
105 83.2 41.3 61.9#









#21
4-Bromophenyl-phenylether
Concen: 0.359 ng
RT: 16.404 min Scan# 1625
Delta R.T. -0.012 min
Lab File: BN038066.D
Acq: 13 Oct 2025 13:47




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion:248 Resp: 4416
Ion Ratio Lower Upper
248 100
250 105.0 77.6 116.4
141 35.8 48.8 73.2#



#22
Hexachlorobenzene
Concen: 0.365 ng
RT: 16.515 min Scan# 1634
Delta R.T. -0.025 min
Lab File: BN038066.D
Acq: 13 Oct 2025 13:47

Tgt Ion:284 Resp: 5456
Ion Ratio Lower Upper
284 100
142 38.2 30.9 46.3
249 30.1 23.9 35.9









#23
Atrazine
Concen: 0.332 ng
RT: 16.664 min Scan# 1646
Delta R.T. -0.025 min
Lab File: BN038066.D
Acq: 13 Oct 2025 13:47




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion:200 Resp: 3115
Ion Ratio Lower Upper
200 100
173 28.7 18.3 27.5#
215 44.3 38.6 58.0



#24
Pentachlorophenol
Concen: 0.476 ng
RT: 16.851 min Scan# 1661
Delta R.T. -0.024 min
Lab File: BN038066.D
Acq: 13 Oct 2025 13:47

Tgt Ion:266 Resp: 2441
Ion Ratio Lower Upper
266 100
264 60.9 50.9 76.3
268 62.2 54.3 81.5









#25
Phenanthrene
Concen: 0.353 ng
RT: 17.248 min Scan# 1700
Delta R.T. -0.012 min
Lab File: BN038066.D
Acq: 13 Oct 2025 13:47




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion:178 Resp: 21937
Ion Ratio Lower Upper
178 100
176 19.5 15.7 23.5
179 15.6 12.3 18.5



#26
Anthracene
Concen: 0.351 ng
RT: 17.335 min Scan# 1700
Delta R.T. -0.025 min
Lab File: BN038066.D
Acq: 13 Oct 2025 13:47

Tgt Ion:178 Resp: 18915
Ion Ratio Lower Upper
178 100
176 18.6 15.0 22.4
179 15.2 12.3 18.5





#27

Fluoranthene-d10

Concen: 0.423 ng

RT: 19.239 min Scan# 1951

Delta R.T. -0.014 min

Lab File: BN038066.D

Acq: 13 Oct 2025 13:47

Instrument :

BNA_N

ClientSampleId :

SSTDCCC0.4





Tgt Ion:212 Resp: 20125

Ion Ratio Lower Upper

212 100

106 15.7 12.6 19.0

104 9.1 7.4 11.0



#28

Fluoranthene

Concen: 0.365 ng

RT: 19.266 min Scan# 1957

Delta R.T. -0.019 min

Lab File: BN038066.D

Acq: 13 Oct 2025 13:47




Tgt Ion:202 Resp: 24986


Ion Ratio Lower Upper

202 100

101 11.3 9.3 13.9

203 16.9 13.6 20.4









#29
Chrysene-d12
Concen: 0.400 ng
RT: 21.393 min Scan# 21
Delta R.T. -0.009 min
Lab File: BN038066.D
Acq: 13 Oct 2025 13:47




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion:240 Resp: 16827
Ion Ratio Lower Upper
240 100
120 14.2 11.4 17.2
236 29.0 23.0 34.6



#30
Pyrene
Concen: 0.385 ng
RT: 19.629 min Scan# 2035
Delta R.T. -0.019 min
Lab File: BN038066.D
Acq: 13 Oct 2025 13:47

Tgt Ion:202 Resp: 25549
Ion Ratio Lower Upper
202 100
200 21.0 16.7 25.1
203 17.6 14.2 21.4









#31
Terphenyl-d14
Concen: 0.364 ng
RT: 19.838 min Scan# 2080
Delta R.T. -0.014 min
Lab File: BN038066.D
Acq: 13 Oct 2025 13:47




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion:244 Resp: 12185
Ion Ratio Lower Upper
244 100
212 9.0 7.6 11.4
122 16.7 13.9 20.9



#32
Benzo(a)anthracene
Concen: 0.367 ng
RT: 21.375 min Scan# 2269
Delta R.T. -0.009 min
Lab File: BN038066.D
Acq: 13 Oct 2025 13:47

Tgt Ion:228 Resp: 21568
Ion Ratio Lower Upper
228 100
226 27.7 22.6 33.8
229 19.6 15.6 23.4









#33
Chrysene
Concen: 0.385 ng
RT: 21.429 min Scan# 21
Delta R.T. -0.009 min
Lab File: BN038066.D
Acq: 13 Oct 2025 13:47




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion:228 Resp: 24503
Ion Ratio Lower Upper
228 100
226 30.4 24.6 37.0
229 19.8 15.6 23.4



#34
Bis(2-ethylhexyl)phthalate
Concen: 0.659 ng
RT: 21.313 min Scan# 2262
Delta R.T. -0.009 min
Lab File: BN038066.D
Acq: 13 Oct 2025 13:47


Tgt Ion:149 Resp: 15324
Ion Ratio Lower Upper
149 100
167 27.0 20.4 30.6
279 3.3 2.4 3.6








#35
Perylene-d12
Concen: 0.400 ng
RT: 23.712 min Scan# 21
Delta R.T. -0.015 min
Lab File: BN038066.D
Acq: 13 Oct 2025 13:47

Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4







Tgt Ion:264 Resp: 17326
Ion Ratio Lower Upper
264 100
260 26.6 21.5 32.3
265 79.8 53.8 80.6



#36
Indeno(1,2,3-cd)pyrene
Concen: 0.386 ng
RT: 26.080 min Scan# 3641
Delta R.T. -0.026 min
Lab File: BN038066.D
Acq: 13 Oct 2025 13:47


Tgt Ion:276 Resp: 30553
Ion Ratio Lower Upper
276 100
138 26.3 21.4 32.0
277 23.7 20.3 30.5








#37
Benzo(b)fluoranthene
Concen: 0.359 ng
RT: 23.057 min Scan# 2607
Delta R.T. 0.032 min
Lab File: BN038066.D
Acq: 13 Oct 2025 13:47

Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4







Tgt Ion:252 Resp: 26813
Ion Ratio Lower Upper
252 100
253 25.7 19.4 29.0
125 18.7 13.0 19.4



#38
Benzo(k)fluoranthene
Concen: 0.357 ng
RT: 23.057 min Scan# 2607
Delta R.T. -0.015 min
Lab File: BN038066.D
Acq: 13 Oct 2025 13:47

Tgt Ion:252 Resp: 26813
Ion Ratio Lower Upper
252 100
253 25.7 19.5 29.3
125 18.7 13.0 19.4









#39
Benzo(a)pyrene
Concen: 0.367 ng
RT: 23.613 min Scan# 21
Delta R.T. -0.012 min
Lab File: BN038066.D
Acq: 13 Oct 2025 13:47




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion:252 Resp: 21668
Ion Ratio Lower Upper
252 100
253 28.7 20.6 30.8
125 29.2 16.7 25.1#



#40
Dibenzo(a,h)anthracene
Concen: 0.368 ng
RT: 26.098 min Scan# 3647
Delta R.T. -0.020 min
Lab File: BN038066.D
Acq: 13 Oct 2025 13:47

Tgt Ion:278 Resp: 22721
Ion Ratio Lower Upper
278 100
139 20.6 15.4 23.2
279 26.4 20.8 31.2








#41
 Benzo(g,h,i)perylene
 Concen: 0.423 ng
 RT: 26.800 min Scan# 31
 Delta R.T. -0.029 min
 Lab File: BN038066.D
 Acq: 13 Oct 2025 13:47

Instrument :
 BNA_N
 ClientSampleId :
 SSTDCCC0.4

Tgt Ion:	276	Resp:	27477
Ion	Ratio	Lower	Upper
276	100		
277	24.8	20.2	30.4
138	25.5	19.8	29.8



Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN101325\
 Data File : BN038083.D
 Acq On : 14 Oct 2025 01:19
 Operator : RC/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SSTDCCC0.4

Quant Time: Oct 14 03:52:02 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN092325.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Sep 24 11:00:01 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.797	152	7061	0.400	ng	-0.03
7) Naphthalene-d8	10.594	136	18151	0.400	ng	#-0.03
13) Acenaphthene-d10	14.452	164	9016	0.400	ng	-0.02
19) Phenanthrene-d10	17.198	188	16225	0.400	ng	#-0.02
29) Chrysene-d12	21.384	240	13579	0.400	ng	#-0.02
35) Perylene-d12	23.709	264	13727	0.400	ng	#-0.02
System Monitoring Compounds						
4) 2-Fluorophenol	5.363	112	5814	0.361	ng	-0.02
5) Phenol-d6	6.952	99	6635	0.313	ng	-0.02
8) Nitrobenzene-d5	8.950	82	4510	0.326	ng	-0.02
11) 2-Methylnaphthalene-d10	12.192	152	9902	0.393	ng	-0.03
14) 2,4,6-Tribromophenol	15.944	330	1046	0.306	ng	-0.02
15) 2-Fluorobiphenyl	13.073	172	13525	0.366	ng	-0.02
27) Fluoranthene-d10	19.234	212	16484	0.404	ng	-0.02
31) Terphenyl-d14	19.838	244	9629	0.356	ng	-0.01
Target Compounds						
						Qvalue
2) 1,4-Dioxane	3.283	88	3339	0.466	ng	94
3) n-Nitrosodimethylamine	3.586	42	3735	0.392	ng	# 91
6) bis(2-Chloroethyl)ether	7.212	93	7427	0.378	ng	99
9) Naphthalene	10.648	128	19092	0.382	ng	99
10) Hexachlorobutadiene	10.947	225	3405	0.399	ng	# 99
12) 2-Methylnaphthalene	12.268	142	11497	0.352	ng	99
16) Acenaphthylene	14.174	152	15325	0.374	ng	100
17) Acenaphthene	14.516	154	10578	0.363	ng	99
18) Fluorene	15.499	166	11554	0.327	ng	98
20) 4,6-Dinitro-2-methylph...	15.585	198	580	0.347	ng	# 10
21) 4-Bromophenyl-phenylether	16.391	248	3637	0.344	ng	# 85
22) Hexachlorobenzene	16.515	284	4887	0.381	ng	99
23) Atrazine	16.664	200	2436	0.303	ng	# 93
24) Pentachlorophenol	16.863	266	1846	0.420	ng	95
25) Phenanthrene	17.248	178	19133	0.358	ng	99
26) Anthracene	17.335	178	15982	0.345	ng	100
28) Fluoranthene	19.266	202	20674	0.352	ng	99
30) Pyrene	19.629	202	20709	0.386	ng	100
32) Benzo(a)anthracene	21.375	228	17445	0.368	ng	99
33) Chrysene	21.420	228	20081	0.391	ng	100
34) Bis(2-ethylhexyl)phtha...	21.304	149	8392	0.447	ng	98
36) Indeno(1,2,3-cd)pyrene	26.072	276	26580	0.424	ng	97
37) Benzo(b)fluoranthene	23.049	252	22684	0.383	ng	95
38) Benzo(k)fluoranthene	23.049	252	22684	0.381	ng	96
39) Benzo(a)pyrene	23.607	252	18395	0.393	ng	# 83
40) Dibenzo(a,h)anthracene	26.092	278	19499	0.398	ng	98
41) Benzo(g,h,i)perylene	26.788	276	22628	0.440	ng	98


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN101325\
Data File : BN038083.D
Acq On : 14 Oct 2025 01:19
Operator : RC/JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 18 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4

Quant Time: Oct 14 03:52:02 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN092325.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Sep 24 11:00:01 2025
Response via : Initial Calibration









#1
 1,4-Dichlorobenzene-d4
 Concen: 0.400 ng
 RT: 7.797 min Scan# 61
 Delta R.T. -0.028 min
 Lab File: BN038083.D
 Acq: 14 Oct 2025 01:19




Instrument :
 BNA_N
 ClientSampleId :
 SSTDCCC0.4


Tgt Ion:152 Resp: 7061
 Ion Ratio Lower Upper
 152 100
 150 155.0 121.0 181.4
 115 60.3 47.4 71.0



#2
 1,4-Dioxane
 Concen: 0.466 ng
 RT: 3.283 min Scan# 28
 Delta R.T. -0.015 min
 Lab File: BN038083.D
 Acq: 14 Oct 2025 01:19

Tgt Ion: 88 Resp: 3339
 Ion Ratio Lower Upper
 88 100
 43 28.6 26.6 39.8
 58 68.9 58.9 88.3









#3
 n-Nitrosodimethylamine
 Concen: 0.392 ng
 RT: 3.586 min Scan# 70
 Delta R.T. -0.015 min
 Lab File: BN038083.D
 Acq: 14 Oct 2025 01:19




Instrument :
 BNA_N
 ClientSampleId :
 SSTDCCC0.4


Tgt Ion: 42 Resp: 3735
 Ion Ratio Lower Upper
 42 100
 74 112.6 93.4 140.0
 44 10.4 22.2 33.4#



#4
 2-Fluorophenol
 Concen: 0.361 ng
 RT: 5.363 min Scan# 316
 Delta R.T. -0.022 min
 Lab File: BN038083.D
 Acq: 14 Oct 2025 01:19

Tgt Ion: 112 Resp: 5814
 Ion Ratio Lower Upper
 112 100
 64 55.5 44.6 66.8
 63 30.8 24.9 37.3









#5
Phenol-d6
Concen: 0.313 ng
RT: 6.952 min Scan# 51
Delta R.T. -0.022 min
Lab File: BN038083.D
Acq: 14 Oct 2025 01:19




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion: 99 Resp: 6635
Ion Ratio Lower Upper
99 100
42 22.9 17.2 25.8
71 33.4 27.3 40.9



#6
bis(2-Chloroethyl)ether
Concen: 0.378 ng
RT: 7.212 min Scan# 572
Delta R.T. -0.029 min
Lab File: BN038083.D
Acq: 14 Oct 2025 01:19

Tgt Ion: 93 Resp: 7427
Ion Ratio Lower Upper
93 100
63 76.0 60.1 90.1
95 31.2 25.4 38.2





#7

Naphthalene-d8

Concen: 0.400 ng

RT: 10.594 min Scan# 91

Delta R.T. -0.032 min

Lab File: BN038083.D

Acq: 14 Oct 2025 01:19

Instrument :

BNA_N

ClientSampleId :

SSTDCCC0.4

Tgt Ion:136 Resp: 18151





Ion Ratio Lower Upper

136 100

137 11.5 9.0 13.4

54 9.3 5.8 8.8#

68 7.0 4.3 6.5#



#8

Nitrobenzene-d5

Concen: 0.326 ng

RT: 8.950 min Scan# 784

Delta R.T. -0.021 min

Lab File: BN038083.D

Acq: 14 Oct 2025 01:19




Tgt Ion: 82 Resp: 4510


Ion Ratio Lower Upper

82 100

128 46.8 34.8 52.2

54 53.6 42.2 63.2





#9

Naphthalene

Concen: 0.382 ng

RT: 10.648 min Scan# 94

Delta R.T. -0.021 min

Lab File: BN038083.D

Acq: 14 Oct 2025 01:19

Instrument :

BNA_N

ClientSampleId :

SSTDCCC0.4





Tgt Ion:128 Resp: 19092

Ion Ratio Lower Upper

128 100

129 11.4 9.2 13.8

127 13.4 11.0 16.6



#10

Hexachlorobutadiene

Concen: 0.399 ng

RT: 10.947 min Scan# 971

Delta R.T. -0.032 min

Lab File: BN038083.D

Acq: 14 Oct 2025 01:19




Tgt Ion:225 Resp: 3405


Ion Ratio Lower Upper

225 100

223 0.0 0.0 0.0

227 63.3 51.4 77.2









#11
2-Methylnaphthalene-d10
Concen: 0.393 ng
RT: 12.192 min Scan# 1163
Delta R.T. -0.025 min
Lab File: BN038083.D
Acq: 14 Oct 2025 01:19




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion:152 Resp: 9902
Ion Ratio Lower Upper
152 100
151 21.8 17.4 26.2



#12
2-Methylnaphthalene
Concen: 0.352 ng
RT: 12.268 min Scan# 1173
Delta R.T. -0.025 min
Lab File: BN038083.D
Acq: 14 Oct 2025 01:19

Tgt Ion:142 Resp: 11497
Ion Ratio Lower Upper
142 100
141 88.6 71.2 106.8
115 36.0 29.7 44.5









#13
 Acenaphthene-d10
 Concen: 0.400 ng
 RT: 14.452 min Scan# 14
 Delta R.T. -0.021 min
 Lab File: BN038083.D
 Acq: 14 Oct 2025 01:19




Instrument :
 BNA_N
 ClientSampleId :
 SSTDCCC0.4


Tgt Ion:164 Resp: 9016
 Ion Ratio Lower Upper
 164 100
 162 105.1 84.8 127.2
 160 55.0 46.1 69.1



#14
 2,4,6-Tribromophenol
 Concen: 0.306 ng
 RT: 15.944 min Scan# 1588
 Delta R.T. -0.025 min
 Lab File: BN038083.D
 Acq: 14 Oct 2025 01:19

Tgt Ion:330 Resp: 1046
 Ion Ratio Lower Upper
 330 100
 332 98.3 77.2 115.8
 141 46.3 37.2 55.8









#15
2-Fluorobiphenyl
Concen: 0.366 ng
RT: 13.073 min Scan# 11
Delta R.T. -0.021 min
Lab File: BN038083.D
Acq: 14 Oct 2025 01:19




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion:172 Resp: 13525
Ion Ratio Lower Upper
172 100
171 36.0 28.6 43.0
170 24.1 19.7 29.5



#16
Acenaphthylene
Concen: 0.374 ng
RT: 14.174 min Scan# 1428
Delta R.T. -0.021 min
Lab File: BN038083.D
Acq: 14 Oct 2025 01:19

Tgt Ion:152 Resp: 15325
Ion Ratio Lower Upper
152 100
151 19.6 15.8 23.8
153 13.1 10.5 15.7









#17
Acenaphthene
Concen: 0.363 ng
RT: 14.516 min Scan# 1460
Delta R.T. -0.021 min
Lab File: BN038083.D
Acq: 14 Oct 2025 01:19




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion:154 Resp: 10578
Ion Ratio Lower Upper
154 100
153 114.3 90.5 135.7
152 65.5 51.8 77.8



#18
Fluorene
Concen: 0.327 ng
RT: 15.499 min Scan# 1552
Delta R.T. -0.023 min
Lab File: BN038083.D
Acq: 14 Oct 2025 01:19

Tgt Ion:166 Resp: 11554
Ion Ratio Lower Upper
166 100
165 101.2 79.0 118.4
167 12.9 10.6 16.0









#19
Phenanthrene-d10
Concen: 0.400 ng
RT: 17.198 min Scan# 1689
Delta R.T. -0.025 min
Lab File: BN038083.D
Acq: 14 Oct 2025 01:19




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion:188 Resp: 16225
Ion Ratio Lower Upper
188 100
94 0.0 0.0 0.0
80 13.4 7.6 11.4#



#20
4,6-Dinitro-2-methylphenol
Concen: 0.347 ng
RT: 15.585 min Scan# 1559
Delta R.T. -0.012 min
Lab File: BN038083.D
Acq: 14 Oct 2025 01:19

Tgt Ion:198 Resp: 580
Ion Ratio Lower Upper
198 100
51 187.8 59.1 88.7#
105 70.5 41.3 61.9#









#21
4-Bromophenyl-phenylether
Concen: 0.344 ng
RT: 16.391 min Scan# 1624
Delta R.T. -0.025 min
Lab File: BN038083.D
Acq: 14 Oct 2025 01:19




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion:248 Resp: 3637
Ion Ratio Lower Upper
248 100
250 92.5 77.6 116.4
141 85.5 48.8 73.2#



#22
Hexachlorobenzene
Concen: 0.381 ng
RT: 16.515 min Scan# 1634
Delta R.T. -0.025 min
Lab File: BN038083.D
Acq: 14 Oct 2025 01:19

Tgt Ion:284 Resp: 4887
Ion Ratio Lower Upper
284 100
142 39.1 30.9 46.3
249 30.2 23.9 35.9









#23
 Atrazine
 Concen: 0.303 ng
 RT: 16.664 min Scan# 1646
 Delta R.T. -0.025 min
 Lab File: BN038083.D
 Acq: 14 Oct 2025 01:19




Instrument :
 BNA_N
 ClientSampleId :
 SSTDCCC0.4


Tgt Ion:200 Resp: 2436
 Ion Ratio Lower Upper
 200 100
 173 29.6 18.3 27.5#
 215 45.4 38.6 58.0



#24
 Pentachlorophenol
 Concen: 0.420 ng
 RT: 16.863 min Scan# 1662
 Delta R.T. -0.012 min
 Lab File: BN038083.D
 Acq: 14 Oct 2025 01:19

Tgt Ion:266 Resp: 1846
 Ion Ratio Lower Upper
 266 100
 264 59.0 50.9 76.3
 268 64.0 54.3 81.5









#25
Phenanthrene
Concen: 0.358 ng
RT: 17.248 min Scan# 1700
Delta R.T. -0.012 min
Lab File: BN038083.D
Acq: 14 Oct 2025 01:19




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion:178 Resp: 19133
Ion Ratio Lower Upper
178 100
176 19.3 15.7 23.5
179 15.1 12.3 18.5



#26
Anthracene
Concen: 0.345 ng
RT: 17.335 min Scan# 1700
Delta R.T. -0.025 min
Lab File: BN038083.D
Acq: 14 Oct 2025 01:19

Tgt Ion:178 Resp: 15982
Ion Ratio Lower Upper
178 100
176 18.6 15.0 22.4
179 15.1 12.3 18.5





#27

Fluoranthene-d10

Concen: 0.404 ng

RT: 19.234 min Scan# 1950

Delta R.T. -0.019 min

Lab File: BN038083.D

Acq: 14 Oct 2025 01:19

Instrument :

BNA_N

ClientSampleId :

SSTDCCC0.4





Tgt Ion:212 Resp: 16484

Ion Ratio Lower Upper

212 100

106 16.2 12.6 19.0

104 9.4 7.4 11.0



#28

Fluoranthene

Concen: 0.352 ng

RT: 19.266 min Scan# 1957

Delta R.T. -0.019 min

Lab File: BN038083.D

Acq: 14 Oct 2025 01:19




Tgt Ion:202 Resp: 20674


Ion Ratio Lower Upper

202 100

101 12.1 9.3 13.9

203 17.2 13.6 20.4





#29

Chrysene-d12

Concen: 0.400 ng

RT: 21.384 min Scan# 21

Delta R.T. -0.018 min

Lab File: BN038083.D

Acq: 14 Oct 2025 01:19

Instrument :

BNA_N

ClientSampleId :

SSTDCCC0.4





Tgt Ion:240 Resp: 13579

Ion Ratio Lower Upper

240 100

120 20.2 11.4 17.2#

236 29.4 23.0 34.6



#30

Pyrene

Concen: 0.386 ng

RT: 19.629 min Scan# 2035

Delta R.T. -0.019 min

Lab File: BN038083.D

Acq: 14 Oct 2025 01:19




Tgt Ion:202 Resp: 20709


Ion Ratio Lower Upper

202 100

200 21.0 16.7 25.1

203 18.1 14.2 21.4









#31
 Terphenyl-d14
 Concen: 0.356 ng
 RT: 19.838 min Scan# 2083
 Delta R.T. -0.014 min
 Lab File: BN038083.D
 Acq: 14 Oct 2025 01:19




Instrument :
 BNA_N
 ClientSampleId :
 SSTDCCC0.4


Tgt Ion:244 Resp: 9629
 Ion Ratio Lower Upper
 244 100
 212 9.5 7.6 11.4
 122 17.2 13.9 20.9



#32
 Benzo(a)anthracene
 Concen: 0.368 ng
 RT: 21.375 min Scan# 2269
 Delta R.T. -0.009 min
 Lab File: BN038083.D
 Acq: 14 Oct 2025 01:19

Tgt Ion:228 Resp: 17445
 Ion Ratio Lower Upper
 228 100
 226 27.3 22.6 33.8
 229 19.6 15.6 23.4





#33

Chrysene

Concen: 0.391 ng

RT: 21.420 min Scan# 21

Delta R.T. -0.018 min

Lab File: BN038083.D

Acq: 14 Oct 2025 01:19

Instrument :

BNA_N

ClientSampleId :

SSTDCCC0.4





Tgt Ion:228 Resp: 20081

Ion Ratio Lower Upper

228 100

226 30.6 24.6 37.0

229 19.3 15.6 23.4



#34

Bis(2-ethylhexyl)phthalate

Concen: 0.447 ng

RT: 21.304 min Scan# 2261

Delta R.T. -0.018 min

Lab File: BN038083.D

Acq: 14 Oct 2025 01:19




Tgt Ion:149 Resp: 8392


Ion Ratio Lower Upper

149 100

167 26.4 20.4 30.6

279 3.2 2.4 3.6









#35
 Perylene-d12
 Concen: 0.400 ng
 RT: 23.709 min Scan# 21
 Delta R.T. -0.018 min
 Lab File: BN038083.D
 Acq: 14 Oct 2025 01:19




Instrument :
 BNA_N
 ClientSampleId :
 SSTDCCC0.4


Tgt Ion:264 Resp: 13727
 Ion Ratio Lower Upper
 264 100
 260 26.5 21.5 32.3
 265 86.3 53.8 80.6#



#36
 Indeno(1,2,3-cd)pyrene
 Concen: 0.424 ng
 RT: 26.072 min Scan# 3638
 Delta R.T. -0.035 min
 Lab File: BN038083.D
 Acq: 14 Oct 2025 01:19


Tgt Ion:276 Resp: 26580
 Ion Ratio Lower Upper
 276 100
 138 28.5 21.4 32.0
 277 23.8 20.3 30.5







#37
Benzo(b)fluoranthene
Concen: 0.383 ng
RT: 23.049 min Scan# 2604
Delta R.T. 0.023 min
Lab File: BN038083.D
Acq: 14 Oct 2025 01:19

Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4







Tgt Ion:252 Resp: 22684
Ion Ratio Lower Upper
252 100
253 25.7 19.4 29.0
125 19.3 13.0 19.4



#38
Benzo(k)fluoranthene
Concen: 0.381 ng
RT: 23.049 min Scan# 2604
Delta R.T. -0.023 min
Lab File: BN038083.D
Acq: 14 Oct 2025 01:19


Tgt Ion:252 Resp: 22684
Ion Ratio Lower Upper
252 100
253 25.7 19.5 29.3
125 19.3 13.0 19.4








#39
Benzo(a)pyrene
Concen: 0.393 ng
RT: 23.607 min Scan# 21
Delta R.T. -0.018 min
Lab File: BN038083.D
Acq: 14 Oct 2025 01:19

Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4







Tgt Ion:252 Resp: 18395
Ion Ratio Lower Upper
252 100
253 28.4 20.6 30.8
125 35.4 16.7 25.1#



#40
Dibenzo(a,h)anthracene
Concen: 0.398 ng
RT: 26.092 min Scan# 3645
Delta R.T. -0.026 min
Lab File: BN038083.D
Acq: 14 Oct 2025 01:19


Tgt Ion:278 Resp: 19499
Ion Ratio Lower Upper
278 100
139 20.3 15.4 23.2
279 26.9 20.8 31.2







#41
Benzo(g,h,i)perylene
Concen: 0.440 ng
RT: 26.788 min Scan# 3883
Delta R.T. -0.041 min
Lab File: BN038083.D
Acq: 14 Oct 2025 01:19

Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4



Tgt Ion:276 Resp: 22628
Ion Ratio Lower Upper
276 100
277 24.3 20.2 30.4
138 25.7 19.8 29.8



Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN101325\
 Data File : BN038088.D
 Acq On : 14 Oct 2025 04:56
 Operator : RC/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SSTDCCC0.4EC

Quant Time: Oct 14 05:20:56 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN092325.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Sep 24 11:00:01 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.797	152	6905	0.400	ng	-0.03
7) Naphthalene-d8	10.594	136	18816	0.400	ng	#-0.03
13) Acenaphthene-d10	14.452	164	9558	0.400	ng	-0.02
19) Phenanthrene-d10	17.198	188	17543	0.400	ng	#-0.02
29) Chrysene-d12	21.384	240	14165	0.400	ng	#-0.02
35) Perylene-d12	23.706	264	13844	0.400	ng	#-0.02
System Monitoring Compounds						
4) 2-Fluorophenol	5.363	112	6075	0.386	ng	-0.02
5) Phenol-d6	6.952	99	6734	0.325	ng	-0.02
8) Nitrobenzene-d5	8.950	82	4804	0.335	ng	-0.02
11) 2-Methylnaphthalene-d10	12.192	152	10106	0.387	ng	-0.03
14) 2,4,6-Tribromophenol	15.945	330	1142	0.315	ng	-0.02
15) 2-Fluorobiphenyl	13.073	172	15070	0.385	ng	-0.02
27) Fluoranthene-d10	19.239	212	17173	0.389	ng	-0.01
31) Terphenyl-d14	19.833	244	10595	0.376	ng	-0.02
Target Compounds						
						Qvalue
2) 1,4-Dioxane	3.283	88	3409	0.486	ng	94
3) n-Nitrosodimethylamine	3.586	42	3947	0.423	ng	# 88
6) bis(2-Chloroethyl)ether	7.219	93	7442	0.387	ng	99
9) Naphthalene	10.648	128	20288	0.391	ng	99
10) Hexachlorobutadiene	10.947	225	3539	0.400	ng	# 97
12) 2-Methylnaphthalene	12.268	142	12369	0.365	ng	99
16) Acenaphthylene	14.174	152	16454	0.379	ng	100
17) Acenaphthene	14.516	154	11586	0.375	ng	99
18) Fluorene	15.499	166	13064	0.349	ng	100
20) 4,6-Dinitro-2-methylph...	15.585	198	690	0.374	ng	# 26
21) 4-Bromophenyl-phenylether	16.391	248	3981	0.348	ng	# 83
22) Hexachlorobenzene	16.515	284	5224	0.377	ng	100
23) Atrazine	16.664	200	2825	0.324	ng	# 90
24) Pentachlorophenol	16.863	266	1613	0.339	ng	97
25) Phenanthrene	17.248	178	20782	0.360	ng	100
26) Anthracene	17.335	178	17015	0.340	ng	100
28) Fluoranthene	19.266	202	22691	0.357	ng	99
30) Pyrene	19.629	202	23294	0.417	ng	100
32) Benzo(a)anthracene	21.375	228	18122	0.366	ng	98
33) Chrysene	21.420	228	21548	0.403	ng	99
34) Bis(2-ethylhexyl)phtha...	21.304	149	7685	0.392	ng	# 99
36) Indeno(1,2,3-cd)pyrene	26.072	276	25477	0.403	ng	98
37) Benzo(b)fluoranthene	23.049	252	22592	0.378	ng	95
38) Benzo(k)fluoranthene	23.049	252	22592	0.376	ng	95
39) Benzo(a)pyrene	23.607	252	18632	0.395	ng	# 87
40) Dibenzo(a,h)anthracene	26.089	278	18798	0.381	ng	97
41) Benzo(g,h,i)perylene	26.788	276	22396	0.432	ng	98


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN101325\
Data File : BN038088.D
Acq On : 14 Oct 2025 04:56
Operator : RC/JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 23 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC

Quant Time: Oct 14 05:20:56 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN092325.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Sep 24 11:00:01 2025
Response via : Initial Calibration









#1
 1,4-Dichlorobenzene-d4
 Concen: 0.400 ng
 RT: 7.797 min Scan# 61
 Delta R.T. -0.028 min
 Lab File: BN038088.D
 Acq: 14 Oct 2025 04:56




Instrument :
 BNA_N
 ClientSampleId :
 SSTDCCC0.4EC


Tgt Ion:152 Resp: 6905
 Ion Ratio Lower Upper
 152 100
 150 154.8 121.0 181.4
 115 62.2 47.4 71.0



#2
 1,4-Dioxane
 Concen: 0.486 ng
 RT: 3.283 min Scan# 28
 Delta R.T. -0.014 min
 Lab File: BN038088.D
 Acq: 14 Oct 2025 04:56

Tgt Ion: 88 Resp: 3409
 Ion Ratio Lower Upper
 88 100
 43 27.4 26.6 39.8
 58 70.3 58.9 88.3









#3
 n-Nitrosodimethylamine
 Concen: 0.423 ng
 RT: 3.586 min Scan# 70
 Delta R.T. -0.014 min
 Lab File: BN038088.D
 Acq: 14 Oct 2025 04:56




Instrument :
 BNA_N
 ClientSampleId :
 SSTDCCC0.4EC


Tgt Ion: 42 Resp: 3947
 Ion Ratio Lower Upper
 42 100
 74 111.5 93.4 140.0
 44 6.5 22.2 33.4#



#4
 2-Fluorophenol
 Concen: 0.386 ng
 RT: 5.363 min Scan# 316
 Delta R.T. -0.022 min
 Lab File: BN038088.D
 Acq: 14 Oct 2025 04:56

Tgt Ion: 112 Resp: 6075
 Ion Ratio Lower Upper
 112 100
 64 57.2 44.6 66.8
 63 30.8 24.9 37.3









#5
Phenol-d6
Concen: 0.325 ng
RT: 6.952 min Scan# 51
Delta R.T. -0.022 min
Lab File: BN038088.D
Acq: 14 Oct 2025 04:56




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC


Tgt Ion: 99 Resp: 6734
Ion Ratio Lower Upper
99 100
42 20.6 17.2 25.8
71 31.6 27.3 40.9



#6
bis(2-Chloroethyl)ether
Concen: 0.387 ng
RT: 7.219 min Scan# 573
Delta R.T. -0.022 min
Lab File: BN038088.D
Acq: 14 Oct 2025 04:56

Tgt Ion: 93 Resp: 7442
Ion Ratio Lower Upper
93 100
63 74.5 60.1 90.1
95 31.9 25.4 38.2









#7
Naphthalene-d8
Concen: 0.400 ng
RT: 10.594 min Scan# 91
Delta R.T. -0.032 min
Lab File: BN038088.D
Acq: 14 Oct 2025 04:56




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC


Tgt Ion:136 Resp: 18816
Ion Ratio Lower Upper
136 100
137 11.3 9.0 13.4
54 9.5 5.8 8.8#
68 6.8 4.3 6.5#



#8
Nitrobenzene-d5
Concen: 0.335 ng
RT: 8.950 min Scan# 784
Delta R.T. -0.021 min
Lab File: BN038088.D
Acq: 14 Oct 2025 04:56

Tgt Ion: 82 Resp: 4804
Ion Ratio Lower Upper
82 100
128 46.7 34.8 52.2
54 51.8 42.2 63.2









#9
Naphthalene
Concen: 0.391 ng
RT: 10.648 min Scan# 945
Delta R.T. -0.021 min
Lab File: BN038088.D
Acq: 14 Oct 2025 04:56




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC


Tgt Ion:	128	Resp:	20288
Ion Ratio	Lower	Upper	
128	100		
129	11.3	9.2	13.8
127	13.6	11.0	16.6



#10
Hexachlorobutadiene
Concen: 0.400 ng
RT: 10.947 min Scan# 971
Delta R.T. -0.032 min
Lab File: BN038088.D
Acq: 14 Oct 2025 04:56

Tgt Ion:	225	Resp:	3539
Ion Ratio	Lower	Upper	
225	100		
223	0.0	0.0	0.0
227	62.1	51.4	77.2









#11
2-Methylnaphthalene-d10
Concen: 0.387 ng
RT: 12.192 min Scan# 1163
Delta R.T. -0.025 min
Lab File: BN038088.D
Acq: 14 Oct 2025 04:56




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC


Tgt Ion:152 Resp: 10106
Ion Ratio Lower Upper
152 100
151 20.9 17.4 26.2



#12
2-Methylnaphthalene
Concen: 0.365 ng
RT: 12.268 min Scan# 1173
Delta R.T. -0.025 min
Lab File: BN038088.D
Acq: 14 Oct 2025 04:56

Tgt Ion:142 Resp: 12369
Ion Ratio Lower Upper
142 100
141 88.9 71.2 106.8
115 35.7 29.7 44.5









#13
Acenaphthene-d10
Concen: 0.400 ng
RT: 14.452 min Scan# 1456
Delta R.T. -0.021 min
Lab File: BN038088.D
Acq: 14 Oct 2025 04:56




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC


Tgt Ion:164 Resp: 9558
Ion Ratio Lower Upper
164 100
162 105.4 84.8 127.2
160 55.4 46.1 69.1



#14
2,4,6-Tribromophenol
Concen: 0.315 ng
RT: 15.945 min Scan# 1588
Delta R.T. -0.024 min
Lab File: BN038088.D
Acq: 14 Oct 2025 04:56

Tgt Ion:330 Resp: 1142
Ion Ratio Lower Upper
330 100
332 97.7 77.2 115.8
141 45.8 37.2 55.8









#15
2-Fluorobiphenyl
Concen: 0.385 ng
RT: 13.073 min Scan# 11
Delta R.T. -0.021 min
Lab File: BN038088.D
Acq: 14 Oct 2025 04:56




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC


Tgt Ion:172 Resp: 15070
Ion Ratio Lower Upper
172 100
171 35.2 28.6 43.0
170 24.3 19.7 29.5



#16
Acenaphthylene
Concen: 0.379 ng
RT: 14.174 min Scan# 1428
Delta R.T. -0.021 min
Lab File: BN038088.D
Acq: 14 Oct 2025 04:56

Tgt Ion:152 Resp: 16454
Ion Ratio Lower Upper
152 100
151 19.5 15.8 23.8
153 13.1 10.5 15.7









#17
Acenaphthene
Concen: 0.375 ng
RT: 14.516 min Scan# 1460
Delta R.T. -0.021 min
Lab File: BN038088.D
Acq: 14 Oct 2025 04:56




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC


Tgt Ion:154 Resp: 11586
Ion Ratio Lower Upper
154 100
153 113.4 90.5 135.7
152 63.0 51.8 77.8



#18
Fluorene
Concen: 0.349 ng
RT: 15.499 min Scan# 1552
Delta R.T. -0.023 min
Lab File: BN038088.D
Acq: 14 Oct 2025 04:56

Tgt Ion:166 Resp: 13064
Ion Ratio Lower Upper
166 100
165 98.7 79.0 118.4
167 13.2 10.6 16.0









#19
Phenanthrene-d10
Concen: 0.400 ng
RT: 17.198 min Scan# 11
Delta R.T. -0.025 min
Lab File: BN038088.D
Acq: 14 Oct 2025 04:56




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC


Tgt Ion:188 Resp: 17543
Ion Ratio Lower Upper
188 100
94 0.0 0.0 0.0
80 13.4 7.6 11.4#



#20
4,6-Dinitro-2-methylphenol
Concen: 0.374 ng
RT: 15.585 min Scan# 1559
Delta R.T. -0.012 min
Lab File: BN038088.D
Acq: 14 Oct 2025 04:56

Tgt Ion:198 Resp: 690
Ion Ratio Lower Upper
198 100
51 165.8 59.1 88.7#
105 67.8 41.3 61.9#









#21
4-Bromophenyl-phenylether
Concen: 0.348 ng
RT: 16.391 min Scan# 1624
Delta R.T. -0.025 min
Lab File: BN038088.D
Acq: 14 Oct 2025 04:56




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC


Tgt Ion:248 Resp: 3981
Ion Ratio Lower Upper
248 100
250 92.5 77.6 116.4
141 89.3 48.8 73.2#



#22
Hexachlorobenzene
Concen: 0.377 ng
RT: 16.515 min Scan# 1634
Delta R.T. -0.025 min
Lab File: BN038088.D
Acq: 14 Oct 2025 04:56

Tgt Ion:284 Resp: 5224
Ion Ratio Lower Upper
284 100
142 38.4 30.9 46.3
249 30.2 23.9 35.9









#23
 Atrazine
 Concen: 0.324 ng
 RT: 16.664 min Scan# 1646
 Delta R.T. -0.025 min
 Lab File: BN038088.D
 Acq: 14 Oct 2025 04:56




Instrument :
 BNA_N
 ClientSampleId :
 SSTDCCC0.4EC


Tgt Ion:200 Resp: 2825
 Ion Ratio Lower Upper
 200 100
 173 33.5 18.3 27.5#
 215 45.1 38.6 58.0



#24
 Pentachlorophenol
 Concen: 0.339 ng
 RT: 16.863 min Scan# 1662
 Delta R.T. -0.012 min
 Lab File: BN038088.D
 Acq: 14 Oct 2025 04:56

Tgt Ion:266 Resp: 1613
 Ion Ratio Lower Upper
 266 100
 264 64.8 50.9 76.3
 268 64.7 54.3 81.5









#25
Phenanthrene
Concen: 0.360 ng
RT: 17.248 min Scan# 1700
Delta R.T. -0.012 min
Lab File: BN038088.D
Acq: 14 Oct 2025 04:56




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC


Tgt Ion:178 Resp: 20782
Ion Ratio Lower Upper
178 100
176 19.4 15.7 23.5
179 15.5 12.3 18.5



#26
Anthracene
Concen: 0.340 ng
RT: 17.335 min Scan# 1700
Delta R.T. -0.025 min
Lab File: BN038088.D
Acq: 14 Oct 2025 04:56

Tgt Ion:178 Resp: 17015
Ion Ratio Lower Upper
178 100
176 18.6 15.0 22.4
179 15.2 12.3 18.5









#27
Fluoranthene-d10
Concen: 0.389 ng
RT: 19.239 min Scan# 1954
Delta R.T. -0.014 min
Lab File: BN038088.D
Acq: 14 Oct 2025 04:56




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC


Tgt Ion:212 Resp: 17173
Ion Ratio Lower Upper
212 100
106 16.4 12.6 19.0
104 9.2 7.4 11.0



#28
Fluoranthene
Concen: 0.357 ng
RT: 19.266 min Scan# 1957
Delta R.T. -0.019 min
Lab File: BN038088.D
Acq: 14 Oct 2025 04:56

Tgt Ion:202 Resp: 22691
Ion Ratio Lower Upper
202 100
101 12.2 9.3 13.9
203 17.2 13.6 20.4





#29

Chrysene-d12

Concen: 0.400 ng

RT: 21.384 min Scan# 21

Delta R.T. -0.018 min

Lab File: BN038088.D

Acq: 14 Oct 2025 04:56

Instrument :

BNA_N

ClientSampleId :

SSTDCCC0.4EC





Tgt Ion:240 Resp: 14165

Ion Ratio Lower Upper

240 100

120 20.2 11.4 17.2#

236 29.2 23.0 34.6



#30

Pyrene

Concen: 0.417 ng

RT: 19.629 min Scan# 2035

Delta R.T. -0.019 min

Lab File: BN038088.D

Acq: 14 Oct 2025 04:56




Tgt Ion:202 Resp: 23294


Ion Ratio Lower Upper

202 100

200 21.0 16.7 25.1

203 17.6 14.2 21.4









#31
 Terphenyl-d14
 Concen: 0.376 ng
 RT: 19.833 min Scan# 2083
 Delta R.T. -0.019 min
 Lab File: BN038088.D
 Acq: 14 Oct 2025 04:56




Instrument :
 BNA_N
 ClientSampleId :
 SSTDCCC0.4EC


Tgt Ion:244 Resp: 10595
 Ion Ratio Lower Upper
 244 100
 212 9.5 7.6 11.4
 122 18.1 13.9 20.9



#32
 Benzo(a)anthracene
 Concen: 0.366 ng
 RT: 21.375 min Scan# 2269
 Delta R.T. -0.009 min
 Lab File: BN038088.D
 Acq: 14 Oct 2025 04:56

Tgt Ion:228 Resp: 18122
 Ion Ratio Lower Upper
 228 100
 226 27.1 22.6 33.8
 229 19.9 15.6 23.4





#33

Chrysene

Concen: 0.403 ng

RT: 21.420 min Scan# 21

Delta R.T. -0.018 min

Lab File: BN038088.D

Acq: 14 Oct 2025 04:56

Instrument :

BNA_N

ClientSampleId :

SSTDCCC0.4EC





Tgt Ion:228 Resp: 21548

Ion Ratio Lower Upper

228 100

226 30.4 24.6 37.0

229 19.1 15.6 23.4



#34

Bis(2-ethylhexyl)phthalate

Concen: 0.392 ng

RT: 21.304 min Scan# 2261

Delta R.T. -0.018 min

Lab File: BN038088.D

Acq: 14 Oct 2025 04:56




Tgt Ion:149 Resp: 7685


Ion Ratio Lower Upper

149 100

167 26.2 20.4 30.6


279 3.6 2.4 3.6#








#35
Perylene-d12
Concen: 0.400 ng
RT: 23.706 min Scan# 21
Delta R.T. -0.020 min
Lab File: BN038088.D
Acq: 14 Oct 2025 04:56

Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC







Tgt Ion:264 Resp: 13844
Ion Ratio Lower Upper
264 100
260 26.7 21.5 32.3
265 86.9 53.8 80.6#



#36
Indeno(1,2,3-cd)pyrene
Concen: 0.403 ng
RT: 26.072 min Scan# 3638
Delta R.T. -0.035 min
Lab File: BN038088.D
Acq: 14 Oct 2025 04:56

Tgt Ion:276 Resp: 25477
Ion Ratio Lower Upper
276 100
138 27.1 21.4 32.0
277 24.1 20.3 30.5









#37
Benzo(b)fluoranthene
Concen: 0.378 ng
RT: 23.049 min Scan# 2604
Delta R.T. 0.023 min
Lab File: BN038088.D
Acq: 14 Oct 2025 04:56




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC


Tgt Ion:252 Resp: 22592
Ion Ratio Lower Upper
252 100
253 26.0 19.4 29.0
125 19.0 13.0 19.4



#38
Benzo(k)fluoranthene
Concen: 0.376 ng
RT: 23.049 min Scan# 2604
Delta R.T. -0.023 min
Lab File: BN038088.D
Acq: 14 Oct 2025 04:56


Tgt Ion:252 Resp: 22592
Ion Ratio Lower Upper
252 100
253 26.0 19.5 29.3
125 19.0 13.0 19.4








#39
Benzo(a)pyrene
Concen: 0.395 ng
RT: 23.607 min Scan# 21
Delta R.T. -0.018 min
Lab File: BN038088.D
Acq: 14 Oct 2025 04:56

Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC







Tgt Ion:252 Resp: 18632
Ion Ratio Lower Upper
252 100
253 28.2 20.6 30.8
125 31.4 16.7 25.1#



#40
Dibenzo(a,h)anthracene
Concen: 0.381 ng
RT: 26.089 min Scan# 3644
Delta R.T. -0.029 min
Lab File: BN038088.D
Acq: 14 Oct 2025 04:56


Tgt Ion:278 Resp: 18798
Ion Ratio Lower Upper
278 100
139 20.7 15.4 23.2
279 27.4 20.8 31.2





#41
 Benzo(g,h,i)perylene
 Concen: 0.432 ng
 RT: 26.788 min Scan# 3883
 Delta R.T. -0.041 min
 Lab File: BN038088.D
 Acq: 14 Oct 2025 04:56

Instrument :
 BNA_N
 ClientSampleId :
 SSTDCCC0.4EC



Tgt Ion:	276	Resp:	22396
Ion	Ratio	Lower	Upper
276	100		
277	24.0	20.2	30.4
138	25.6	19.8	29.8

