

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name:	<u>Alliance</u>	Contract:	<u>TETR06</u>
Lab Code:	<u>ACE</u>	SDG No.:	<u>Q3298</u>
Instrument ID:	<u>MSVOA_X</u>	Calibration Date/Time:	<u>10/07/2025 14:06</u>
Lab File ID:	<u>VX048050.D</u>	Init. Calib. Date(s):	<u>09/16/2025 09/16/2025</u>
Heated Purge: (Y/N)	<u>N</u>	Init. Calib. Time(s):	<u>09:23 12:01</u>
GC Column:	<u>DB-624UI</u>	ID:	<u>0.18</u> (mm)

COMPOUND	RRF	RRF050	MIN RRF	%D	MAX%D
Dichlorodifluoromethane	0.518	0.506		-2.32	50
Chloromethane	0.680	0.564	0.1	-17.06	50
Vinyl Chloride	0.666	0.627		-5.86	50
Bromomethane	0.422	0.375		-11.14	50
Chloroethane	0.445	0.429		-3.6	50
Trichlorofluoromethane	0.989	1.019		3.03	50
1,1,2-Trichlorotrifluoroethane	0.578	0.631		9.17	50
1,1-Dichloroethene	0.594	0.586		-1.35	50
Acetone	0.346	0.328		-5.2	50
Carbon Disulfide	1.626	1.365		-16.05	50
Methyl tert-butyl Ether	2.169	2.270		4.66	50
Methyl Acetate	0.770	0.955		24.03	50
Methylene Chloride	0.732	0.742		1.37	50
trans-1,2-Dichloroethene	0.640	0.629		-1.72	50
1,1-Dichloroethane	1.263	1.317	0.1	4.28	50
Cyclohexane	1.052	0.932		-11.41	50
2-Butanone	0.432	0.490		13.43	50
Carbon Tetrachloride	0.532	0.545		2.44	50
cis-1,2-Dichloroethene	0.783	0.804		2.68	50
Bromochloromethane	0.551	0.569		3.27	50
Chloroform	1.276	1.358		6.43	50
1,1,1-Trichloroethane	1.082	1.127		4.16	50
Methylcyclohexane	0.556	0.519		-6.66	50
Benzene	1.488	1.520		2.15	50
1,2-Dichloroethane	0.566	0.588		3.89	50
Trichloroethene	0.360	0.366		1.67	50
1,2-Dichloropropane	0.381	0.406		6.56	50
Bromodichloromethane	0.572	0.630		10.14	50
4-Methyl-2-Pentanone	0.477	0.578		21.17	50
Toluene	0.917	0.935		1.96	50
t-1,3-Dichloropropene	0.584	0.598		2.4	50
cis-1,3-Dichloropropene	0.624	0.651		4.33	50
1,1,2-Trichloroethane	0.354	0.398		12.43	50
2-Hexanone	0.343	0.406		18.37	50
Dibromochloromethane	0.407	0.462		13.51	50
1,2-Dibromoethane	0.360	0.399		10.83	50
Tetrachloroethene	0.326	0.319		-2.15	50
Chlorobenzene	1.136	1.184	0.3	4.22	50

All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name:	<u>Alliance</u>	Contract:	<u>TETR06</u>
Lab Code:	<u>ACE</u>	SDG No.:	<u>Q3298</u>
Instrument ID:	<u>MSVOA_X</u>	Calibration Date/Time:	<u>10/07/2025 14:06</u>
Lab File ID:	<u>VX048050.D</u>	Init. Calib. Date(s):	<u>09/16/2025 09/16/2025</u>
Heated Purge: (Y/N)	<u>N</u>	Init. Calib. Time(s):	<u>09:23 12:01</u>
GC Column:	<u>DB-624UI</u>	ID:	<u>0.18</u> (mm)

COMPOUND	RRF	RRF050	MIN RRF	%D	MAX%D
Ethyl Benzene	1.960	2.018		2.96	50
m/p-Xylenes	0.729	0.757		3.84	50
o-Xylene	0.706	0.742		5.1	50
Styrene	1.236	1.315		6.39	50
Bromoform	0.293	0.344	0.1	17.41	50
Isopropylbenzene	3.801	3.770		-0.82	50
1,1,2,2-Tetrachloroethane	1.150	1.303	0.3	13.3	50
1,3-Dichlorobenzene	1.739	1.731		-0.46	50
1,4-Dichlorobenzene	1.785	1.762		-1.29	50
1,2-Dichlorobenzene	1.657	1.675		1.09	50
1,2-Dibromo-3-Chloropropane	0.233	0.252		8.15	50
1,2,4-Trichlorobenzene	1.077	1.040		-3.43	50
1,2,3-Trichlorobenzene	1.002	1.030		2.79	50
1,2-Dichloroethane-d4	0.796	0.805		1.13	50
Dibromofluoromethane	0.335	0.357		6.57	50
Toluene-d8	1.160	1.184		2.07	50
4-Bromofluorobenzene	0.440	0.453		2.95	50

All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.