

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX100725\
 Data File : VX048050.D
 Acq On : 07 Oct 2025 14:06
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Oct 08 01:28:36 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091625W.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 17 06:39:58 2025
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	78	0.00
2 T	Dichlorodifluoromethane	0.518	0.506	2.3	79	0.00
3 P	Chloromethane	0.680	0.564	17.1	66	0.00
4 C	Vinyl Chloride	0.666	0.627	5.9#	74	0.00
5 T	Bromomethane	0.422	0.375	11.1	69	0.00
6 T	Chloroethane	0.445	0.429	3.6	75	0.00
7 T	Trichlorofluoromethane	0.989	1.019	-3.0	80	0.00
8 T	Diethyl Ether	0.405	0.401	1.0	76	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.578	0.631	-9.2	84	0.00
10 T	Methyl Iodide	0.902	0.641	28.9#	55	0.00
11 T	Tert butyl alcohol	0.089	0.100	-12.4	88	0.00
12 CM	1,1-Dichloroethene	0.594	0.586	1.3#	76	0.00
13 T	Acrolein	0.082	0.094	-14.6	90	0.00
14 T	Allyl chloride	1.226	1.161	5.3	75	0.00
15 T	Acrylonitrile	0.328	0.393	-19.8	88	0.00
16 T	Acetone	0.346	0.328	5.2	80	0.00
17 T	Carbon Disulfide	1.626	1.365	16.1	68	0.00
18 T	Methyl Acetate	0.770	0.955	-24.0	85	0.00
19 T	Methyl tert-butyl Ether	2.169	2.270	-4.7	79	0.00
20 T	Methylene Chloride	0.732	0.742	-1.4	79	0.00
21 T	trans-1,2-Dichloroethene	0.640	0.629	1.7	75	0.00
22 T	Diisopropyl ether	2.376	2.503	-5.3	79	0.00
23 T	Vinyl Acetate	1.901	2.022	-6.4	79	0.00
24 P	1,1-Dichloroethane	1.263	1.317	-4.3	79	0.00
25 T	2-Butanone	0.432	0.490	-13.4	87	0.00
26 T	2,2-Dichloropropane	1.050	1.014	3.4	74	0.00
27 T	cis-1,2-Dichloroethene	0.783	0.804	-2.7	77	0.00
28 T	Bromochloromethane	0.551	0.569	-3.3	77	0.00
29 T	Tetrahydrofuran	0.260	0.300	-15.4	85	0.00
30 C	Chloroform	1.276	1.358	-6.4#	80	0.00
31 T	Cyclohexane	1.052	0.932	11.4	70	0.00
32 T	1,1,1-Trichloroethane	1.082	1.127	-4.2	78	0.00
33 S	1,2-Dichloroethane-d4	0.796	0.805	-1.1	81	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	77	0.00
35 S	Dibromofluoromethane	0.335	0.357	-6.6	83	0.00
36 T	1,1-Dichloropropene	0.490	0.468	4.5	74	0.00
37 T	Ethyl Acetate	0.525	0.565	-7.6	82	0.00
38 T	Carbon Tetrachloride	0.532	0.545	-2.4	78	0.00
39 T	Methylcyclohexane	0.556	0.519	6.7	71	0.00
40 TM	Benzene	1.488	1.520	-2.2	77	0.00
41 T	Methacrylonitrile	0.271	0.317	-17.0	86	0.00
42 TM	1,2-Dichloroethane	0.566	0.588	-3.9	78	0.00
43 T	Isopropyl Acetate	0.860	0.934	-8.6	80	0.00
44 TM	Trichloroethene	0.360	0.366	-1.7	76	0.00
45 C	1,2-Dichloropropane	0.381	0.406	-6.6#	79	0.00
46 T	Dibromomethane	0.277	0.296	-6.9	79	0.00
47 T	Bromodichloromethane	0.572	0.630	-10.1	80	0.00
48 T	Methyl methacrylate	0.429	0.474	-10.5	81	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX100725\
 Data File : VX048050.D
 Acq On : 07 Oct 2025 14:06
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Oct 08 01:28:36 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091625W.M
 Quant Title : SW846 8260
 QLast Update : Wed Sep 17 06:39:58 2025
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	0.004	0.006	-50.0#	102	0.00
50 S	Toluene-d8	1.160	1.184	-2.1	80	0.00
51 T	4-Methyl-2-Pentanone	0.477	0.578	-21.2	87	0.00
52 CM	Toluene	0.917	0.935	-2.0#	77	0.00
53 T	t-1,3-Dichloropropene	0.584	0.598	-2.4	74	0.00
54 T	cis-1,3-Dichloropropene	0.624	0.651	-4.3	77	0.00
55 T	1,1,2-Trichloroethane	0.354	0.398	-12.4	84	0.00
56 T	Ethyl methacrylate	0.555	0.617	-11.2	79	0.00
57 T	1,3-Dichloropropane	0.607	0.674	-11.0	81	0.00
58 T	2-Chloroethyl Vinyl ether	0.227	0.305	-34.4#	86	0.00
59 T	2-Hexanone	0.343	0.406	-18.4	87	0.00
60 T	Dibromochloromethane	0.407	0.462	-13.5	82	0.00
61 T	1,2-Dibromoethane	0.360	0.399	-10.8	81	0.00
62 S	4-Bromofluorobenzene	0.440	0.453	-3.0	82	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	78	0.00
64 T	Tetrachloroethene	0.326	0.319	2.1	76	0.00
65 PM	Chlorobenzene	1.136	1.184	-4.2	79	0.00
66 T	1,1,1,2-Tetrachloroethane	0.392	0.432	-10.2	82	0.00
67 C	Ethyl Benzene	1.960	2.018	-3.0#	77	0.00
68 T	m/p-Xylenes	0.729	0.757	-3.8	78	0.00
69 T	o-Xylene	0.706	0.742	-5.1	77	0.00
70 T	Styrene	1.236	1.315	-6.4	79	0.00
71 P	Bromoform	0.293	0.344	-17.4	86	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	84	0.00
73 T	Isopropylbenzene	3.801	3.770	0.8	78	0.00
74 T	N-amyl acetate	1.733	1.790	-3.3	81	0.00
75 P	1,1,2,2-Tetrachloroethane	1.150	1.303	-13.3	89	0.00
76 T	1,2,3-Trichloropropane	1.012	0.994	1.8	74	0.00
77 T	Bromobenzene	0.936	0.913	2.5	78	0.00
78 T	n-propylbenzene	4.494	4.475	0.4	78	0.00
79 T	2-Chlorotoluene	2.752	2.730	0.8	79	0.00
80 T	1,3,5-Trimethylbenzene	3.123	3.103	0.6	78	0.00
81 T	trans-1,4-Dichloro-2-butene	0.397	0.386	2.8	78	0.00
82 T	4-Chlorotoluene	3.251	3.222	0.9	79	0.00
83 T	tert-Butylbenzene	3.197	3.208	-0.3	80	0.00
84 T	1,2,4-Trimethylbenzene	3.166	3.155	0.3	79	0.00
85 T	sec-Butylbenzene	3.797	3.829	-0.8	80	0.00
86 T	p-Isopropyltoluene	3.215	3.192	0.7	78	0.00
87 T	1,3-Dichlorobenzene	1.739	1.731	0.5	80	0.00
88 T	1,4-Dichlorobenzene	1.785	1.762	1.3	81	0.00
89 T	n-Butylbenzene	2.975	2.966	0.3	79	0.00
90 T	Hexachloroethane	0.564	0.585	-3.7	83	0.00
91 T	1,2-Dichlorobenzene	1.657	1.675	-1.1	81	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.233	0.252	-8.2	83	0.00
93 T	1,2,4-Trichlorobenzene	1.077	1.040	3.4	76	0.00
94 T	Hexachlorobutadiene	0.366	0.358	2.2	78	0.00
95 T	Naphthalene	3.279	3.408	-3.9	81	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX100725\
Data File : VX048050.D
Acq On : 07 Oct 2025 14:06
Operator : JC/MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_X
LabSampleId :
VSTDCCC050

Quant Time: Oct 08 01:28:36 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X091625W.M
Quant Title : SW846 8260
QLast Update : Wed Sep 17 06:39:58 2025
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	1.002	1.030	-2.8	80	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 6