

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.:	<u>Q3298</u>	Analytical Method:	<u>SW8260-Low</u>
Client:	<u>Tetra Tech NUS, Inc.</u>	Datafile :	<u>VX048040.D</u>

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Low	Limits High	RPD
VX1007WBSD01	Dichlorodifluoromethane	20	19.2	ug/L	96	1		32	152	20
	Chloromethane	20	16.7	ug/L	84	2		50	139	20
	Vinyl chloride	20	18.4	ug/L	92	3		58	137	20
	Bromomethane	20	19.7	ug/L	99	1		53	141	20
	Chloroethane	20	18.5	ug/L	93	7		60	138	20
	Trichlorofluoromethane	20	19.9	ug/L	100	1		65	141	20
	1,1,2-Trichlorotrifluoroethane	20	20.2	ug/L	101	1		70	136	20
	1,1-Dichloroethene	20	19.1	ug/L	96	3		71	131	20
	Acetone	100	88.7	ug/L	89	12		39	160	20
	Carbon disulfide	20	16.5	ug/L	83	2		64	133	20
	Methyl tert-butyl Ether	20	20.4	ug/L	102	4		71	124	20
	Methyl Acetate	20	23.7	ug/L	119	2		56	136	20
	Methylene Chloride	20	20.2	ug/L	101	3		74	124	20
	trans-1,2-Dichloroethene	20	19.3	ug/L	97	0		75	124	20
	1,1-Dichloroethane	20	20.3	ug/L	102	4		77	125	20
	Cyclohexane	20	18.1	ug/L	91	4		71	130	20
	2-Butanone	100	100	ug/L	100	10		56	143	20
	Carbon Tetrachloride	20	20.4	ug/L	102	2		72	136	20
	cis-1,2-Dichloroethene	20	19.8	ug/L	99	9		78	123	20
	Bromochloromethane	20	22.5	ug/L	113	1		78	123	20
	Chloroform	20	21.8	ug/L	109	3		79	124	20
	1,1,1-Trichloroethane	20	21.3	ug/L	106	0		74	131	20
	Methylcyclohexane	20	18.5	ug/L	93	8		72	132	20
	Benzene	20	20.3	ug/L	102	3		79	120	20
	1,2-Dichloroethane	20	21.1	ug/L	106	4		73	128	20
	Trichloroethene	20	20.0	ug/L	100	4		79	123	20
	1,2-Dichloropropane	20	21.7	ug/L	109	1		78	122	20
	Bromodichloromethane	20	22.1	ug/L	111	3		79	125	20
	4-Methyl-2-Pentanone	100	120	ug/L	120	0		67	130	20
	Toluene	20	20.7	ug/L	104	1		80	121	20
	t-1,3-Dichloropropene	20	20.6	ug/L	103	1		73	127	20
	cis-1,3-Dichloropropene	20	20.9	ug/L	104	4		75	124	20
	1,1,2-Trichloroethane	20	22.9	ug/L	115	2		80	119	20
	2-Hexanone	100	110	ug/L	110	0		57	139	20
	Dibromochloromethane	20	23.2	ug/L	116	3		74	126	20
	1,2-Dibromoethane	20	22.4	ug/L	112	2		77	121	20
	Tetrachloroethene	20	19.8	ug/L	99	4		74	129	20
	Chlorobenzene	20	20.5	ug/L	103	0		82	118	20
	Ethyl Benzene	20	19.6	ug/L	98	0		79	121	20
	m/p-Xylenes	40	40.5	ug/L	101	1		80	121	20
	o-Xylene	20	19.8	ug/L	99	1		78	122	20
	Styrene	20	20.4	ug/L	102	2		78	123	20
	Bromoform	20	22.4	ug/L	112	3		66	130	20
	Isopropylbenzene	20	19.2	ug/L	96	3		72	131	20
	1,1,2,2-Tetrachloroethane	20	21.4	ug/L	107	0		71	121	20
	1,3-Dichlorobenzene	20	19.6	ug/L	98	1		80	119	20
	1,4-Dichlorobenzene	20	20.0	ug/L	100	2		79	118	20
	1,2-Dichlorobenzene	20	20.4	ug/L	102	3		80	119	20

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.:	<u>Q3298</u>	Analytical Method:	<u>SW8260-Low</u>
Client:	<u>Tetra Tech NUS, Inc.</u>	Datafile :	<u>VX048040.D</u>

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Low	Limits High	RPD
VX1007WBSD01	1,2-Dibromo-3-Chloropropane	20	20.0	ug/L	100	6		62	128	20
	1,2,4-Trichlorobenzene	20	18.7	ug/L	94	5		69	130	20
	1,2,3-Trichlorobenzene	20	19.9	ug/L	100	6		69	129	20