

Continuing Calibration Report for SequenceID : FC100925AL

Parameter	AreaCount	Conc.	RT_Min	RT_Max	Response Factor	AVGRF	%DEV
-----------	-----------	-------	--------	--------	-----------------	-------	------

File ID : FC069970.D

Aliphatic C9-C12	9108916.000	60.000	3.284	6.578	151815.267	144785.198	-4.856
Aliphatic C12-C16	6796739.000	40.000	6.579	9.979	169918.475	163782.335	-3.747
Aliphatic C16-C21	9958048.000	60.000	9.980	13.346	165967.467	161945.502	-2.484
Aliphatic C21-C28	11596483.000	80.000	13.347	17.010	144956.038	147142.336	1.486
Aliphatic C28-C40	12749847.000	120.000	17.011	21.966	106248.725	107599.833	1.256
Aliphatic EPH	50210033.000	360.000	3.284	21.966	139472.314	137884.728	-1.151