

Continuing Calibration Report for SequenceID : FD100825AR

Parameter	AreaCount	Conc.	RT_Min	RT_Max	Response Factor	AVGRF	%DEV
-----------	-----------	-------	--------	--------	-----------------	-------	------

File ID : FD049828.D

Aromatic C10-C12	4911521.000	40.000	4.394	6.137	122788.025	122789.439	0.001
Aromatic C12-C16	8705027.000	60.000	6.138	8.765	145083.783	143358.701	-1.203
Aromatic C16-C21	11272445.000	80.000	8.766	13.056	140905.563	142171.561	0.890
Aromatic C21-C36	21626869.000	180.000	13.057	18.475	120149.272	123821.788	2.966
Aromatic EPH	46515862.000	360.000	4.394	18.475	129210.728	131040.962	1.397