

Continuing Calibration Report for SequenceID : FD100825AR

Parameter	AreaCount	Conc.	RT_Min	RT_Max	Response Factor	AVGRF	%DEV
-----------	-----------	-------	--------	--------	-----------------	-------	------

File ID : FD049839.D

Aromatic C10-C12	5031231.000	40.000	4.394	6.137	125780.775	122789.439	-2.436
Aromatic C12-C16	9007525.000	60.000	6.138	8.765	150125.417	143358.701	-4.720
Aromatic C16-C21	11947262.000	80.000	8.766	13.056	149340.775	142171.561	-5.043
Aromatic C21-C36	23016679.000	180.000	13.057	18.475	127870.439	123821.788	-3.270
Aromatic EPH	49002697.000	360.000	4.394	18.475	136118.603	131040.962	-3.875