

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name: Alliance Contract: GFEL01
 Lab Code: ACE SDG No.: Q3306
 Instrument ID: MSVOA_L Calibration Date/Time: 10/08/2025 08:54
 Lab File ID: VL043054.D Init. Calib. Date(s): 10/02/2025 10/02/2025
 Heated Purge: (Y/N) N Init. Calib. Time(s): 09:50 13:41
 GC Column: RTX-1 ID: 0.32 (mm)

COMPOUND	Amount	Conc.	%Dev	MAX%D
Naphthalene	10.000	10.39	-3.909998	20