

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL100225\
 Data File : VL043039.D
 Acq On : 02 Oct 2025 14:14
 Operator : SY/MD
 Sample : VSTDICV010
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 ICVVL100225

Manual Integrations
 APPROVED

Reviewed By :Semsettin Yesilyurt 10/03/2025
 Supervised By :Mahesh Dadoda 10/10/2025

Quant Time: Oct 03 08:24:32 2025

Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL100225AIR.M

Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_LFri Aug 2

QLast Update : Fri Oct 03 08:22:04 2025

Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	2.777	49	178544	10.000	ppbv	0.00
33) 1,4-Difluorobenzene	3.939	114	537412	10.000	ppbv	0.00
55) Chlorobenzene-d5	8.866	117	454012	10.000	ppbv	0.00

System Monitoring Compounds

68) 1-Bromo-4-Fluorobenzene 10.364 95 333843 10.207 ppbv 0.00
 Spiked Amount 10.000 Range 65 - 135 Recovery = 102.100%

Target Compounds

						Qvalue
2) Dichlorodifluoromethane	1.496	85	287085	9.540	ppbv	100
3) Chlorodifluoromethane	1.473	51	275618	9.468	ppbv	99
4) Chloromethane	1.528	50	104427	9.501	ppbv	98
5) Vinyl Chloride	1.577	62	98850	9.476	ppbv	97
6) Bromomethane	1.664	94	46326	9.883	ppbv	97
7) Chloroethane	1.700	64	40636	9.336	ppbv	98
8) Dichlorotetrafluoroethane	1.551	85	235901	9.567	ppbv	99
9) Propene	1.480	41	115956	9.879	ppbv	99
10) Heptane	4.956	43	311038	10.779	ppbv	100
11) Trichlorofluoromethane	1.871	101	291376	9.422	ppbv	100
12) 1,1,2-Trichlorotrifluo...	2.133	101	228314	9.374	ppbv	97
13) Ethanol	1.716	45	4643	9.646	ppbv	98
14) Bromoethene	1.777	108	85349	9.774	ppbv	99
15) Acetone	1.829	43	204381	8.176	ppbv	100
16) 1,3-Butadiene	1.606	39	95179	9.234	ppbv	99
17) tert-Butyl alcohol	2.033	59	249385	9.431	ppbv	99
18) 1,1-Dichloroethene	2.026	96	101929	9.547	ppbv	95
19) Isopropyl Alcohol	1.878	45	109403	9.134	ppbv	99
20) Methylene Chloride	2.056	84	89214	8.176	ppbv	97
21) Allyl Chloride	2.088	41	151234	9.443	ppbv	99
22) trans-1,2-Dichloroethene	2.331	96	115000	9.487	ppbv	97
23) Vinyl Acetate	2.454	43	45886	8.646	ppbv	97
24) 1,1-Dichloroethane	2.399	63	217319	9.458	ppbv	98
25) Ethyl Acetate	2.823	43	628433	10.021	ppbv	100
26) Hexane	2.816	57	253815	10.521	ppbv	98
27) Carbon Disulfide	2.143	76	273348	9.598	ppbv	99
28) Methyl tert-Butyl Ether	2.431	73	147450	9.767	ppbv	99
29) Chloroform	2.836	83	388247	9.439	ppbv	100
30) Cyclohexane	3.839	84	216007	10.896	ppbv	98
31) cis-1,2-Dichloroethene	2.709	61	249555	10.063	ppbv	97
32) 1,1,1-Trichloroethane	3.354	97	387043	9.682	ppbv	100
34) 2-Butanone	2.544	43	346463	9.501	ppbv	100
35) Carbon Tetrachloride	3.745	117	399311	9.031	ppbv	96
36) Benzene	3.642	78	514339	10.017	ppbv	98
37) 1,2-Dichloroethane	3.205	62	289339	9.515	ppbv	100
38) Trichloroethene	4.528	130	213727	9.992	ppbv	98
39) 1,2-Dichloropropane	4.292	63	183886	9.662	ppbv	97
40) 1,4-Dioxane	4.558	88	80249	9.831	ppbv #	98
41) Tetrahydrofuran	3.040	42	201471	10.690	ppbv	100
42) Bromodichloromethane	4.467	83	413457	9.629	ppbv	98
43) Methyl Methacrylate	4.855	69	189993	11.147	ppbv	99
44) 2,2,4-Trimethylpentane	4.619	57	847766	10.290	ppbv	99
45) t-1,3-Dichloropropene	6.389	75	164370	10.538	ppbv	97

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL100225\
 Data File : VL043039.D
 Acq On : 02 Oct 2025 14:14
 Operator : SY/MD
 Sample : VSTDICV010
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 ICVVL100225

Manual Integrations
 APPROVED

Reviewed By : Semsettin Yesilyurt 10/03/2025
 Supervised By : Mahesh Dadoda 10/10/2025

Quant Time: Oct 03 08:24:32 2025

Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL100225AIR.M

Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 2

QLast Update : Fri Oct 03 08:22:04 2025

Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
46) cis-1,3-Dichloropropene	5.574	75	232895	10.597	ppbv	98
47) 1,1,2-Trichloroethane	6.558	97	196496	9.282	ppbv	97
48) Dibromochloromethane	7.367	129	336813	9.749	ppbv	99
49) Bromoform	9.468	173	275751	9.557	ppbv	100
50) 4-Methyl-2-Pentanone	5.720	43	484498	10.896	ppbv	100
51) 2-Hexanone	7.412	43	355624	11.236	ppbv	99
52) Tetrachloroethene	8.208	164	182062	9.576	ppbv	99
53) Toluene	6.911	91	572257	11.040	ppbv	100
54) 1,2-Dibromoethane	7.632	107	247276	9.855	ppbv	98
56) 1,1,1,2-Tetrachloroethane	8.908	131	259863	9.390	ppbv	99
57) Chlorobenzene	8.904	112	447850	9.331	ppbv	100
58) Ethyl Benzene	9.338	91	775232	11.022	ppbv	98
59) m/p-Xylene	9.535	91	1265386m	21.330	ppbv	
60) o-Xylene	9.950	91	628486	10.635	ppbv	99
61) Styrene	9.856	104	284955	11.510	ppbv	100
62) Isopropylbenzene	10.526	105	939236	10.353	ppbv	99
63) 1,1,2,2-Tetrachloroethane	9.950	83	360837	8.511	ppbv	99
64) n-propylbenzene	10.976	120	240769	10.375	ppbv	98
65) tert-Butylbenzene	11.520	119	814322	10.086	ppbv	100
66) Benzyl Chloride	11.604	91	26124	8.466	ppbv	99
67) sec-Butylbenzene	11.756	105	1122743	9.965	ppbv	99
69) p-Isopropyltoluene	11.914	119	937600	10.230	ppbv	100
70) n-Butylbenzene	12.270	91	913299	10.349	ppbv	100
71) 2-Chlorotoluene	10.888	91	673074	10.251	ppbv	100
72) 4-Ethyltoluene	11.112	105	744145	10.579	ppbv	100
73) 1,3,5-Trimethylbenzene	11.193	105	621377	10.212	ppbv	100
74) 1,2,4-Trimethylbenzene	11.526	105	684740	9.971	ppbv	93
75) 1,3-Dichlorobenzene	11.594	146	409217	9.161	ppbv	98
76) 1,4-Dichlorobenzene	11.659	146	398965	9.368	ppbv	100
77) 1,2-Dichlorobenzene	11.934	146	384233	8.636	ppbv	99
78) Hexachloro-1,3-Butadiene	13.869	225	289545	9.340	ppbv	99
79) Naphthalene	13.500	128	552627	9.523	ppbv	99
80) Naphthalene,2-methyl-	14.445	142	223063	8.323	ppbv	98
81) 1,2,4-Trichlorobenzene	13.429	180	317372	11.990	ppbv	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

