

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance Contract: GFEL01
 Lab Code: ACE SDG NO.: Q3306
 Lab File ID: VL043054.D Date Analyzed: 10/08/2025
 Instrument ID: MSVOA_L Time Analyzed: 08:54
 GC Column: RTX-1 ID: 0.32 (mm) Heated Purge: (Y/N) N

	IS4 AREA #	RT #				
12 HOUR STD	0	0				
UPPER LIMIT	0					
LOWER LIMIT	0					
EPA SAMPLE NO.						
IA1	0	0.00				
IA1DUP	0	0.00				
OA1	0	0.00				
SV1	0	0.00				
SV2	0	0.00				
VL1008ABL01	0	0.00				
VL1008ABS01	0	0.00				

IS4 =

AREA UPPER LIMIT = +40% of internal standard area
 AREA LOWER LIMIT = -40% of internal standard area
 RT UPPER LIMIT = +0.33 minutes of internal standard RT
 RT LOWER LIMIT = -0.33 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.