


LB137478

7471B

INSTRUMENT ID: CV1

μ Abs.:

21700



Linear

A= 0.0000e+000

B= 4.5497e-004

C= -4.7261e-002

Rho= 0.9987720

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.005	-0.005	93	0.000	93				
0.2	0.200	0.211	0.011	567	0.0 %	567				
2.5	2.500	2.559	0.059	5728	0.0 %	5728				
5.0	5.000	4.772	-0.228	10592	0.0 %	10592				
7.5	7.500	7.838	0.338	17332	0.0 %	17332				
10.0	10.000	9.826	-0.174	21700	0.0 %	21700				

LB137478 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 09 Oct 2025 13:26:31

Type	Sample ID	Extended ID	Conc.	μ Abs.	Units	Type	Std Conc	Method	Date
S	0.0 - 1	50	-	93	PPB	Std	0.0000	7471B	09 Oct 2025 13:33:48
S	0.2 - 1	50.2	-	567	PPB	Std	0.2000	7471B	09 Oct 2025 13:36:05
S	2.5 - 1	52.5	-	5728	PPB	Std	2.5000	7471B	09 Oct 2025 13:38:22
S	5.0 - 1	55	-	10592	PPB	Std	5.0000	7471B	09 Oct 2025 13:40:39
S	7.5 - 1	53.5	-	17332	PPB	Std	7.5000	7471B	09 Oct 2025 13:52:29
S	10.0 - 1	510	-	21700	PPB	Std	10.0000	7471B	09 Oct 2025 14:09:41
U	ICV65 - 1	ICV65	3.9455	8776	PPB	SMPL	-7471B		09 Oct 2025 14:15:23
U	ICB65 - 1	ICB65	-0.0168	67	PPB	SMPL	-7471B		09 Oct 2025 14:20:25
U	CCV10 - 1	CCV10	5.0293	11158	PPB	SMPL	-7471B		09 Oct 2025 14:22:40
U	CCB10 - 1	CCB10	-0.0632	-35	PPB	SMPL	-7471B		09 Oct 2025 14:27:51
U	CRA - 1	CRA	0.1848	510	PPB	SMPL	-7471B		09 Oct 2025 14:30:06
U	HighStd - 1	HighStd	9.6940	21411	PPB	SMPL	-7471B		09 Oct 2025 14:32:21
U	ChkStd - 1	ChkStd	7.3828	16331	PPB	SMPL	-7471B		09 Oct 2025 14:34:36
U	PB170045BL - 1	PBS	-0.0250	49	PPB	SMPL	-7471B		09 Oct 2025 14:40:22
U	PB170045BS - 1	LCSS	4.7922	10637	PPB	SMPL	-7471B		09 Oct 2025 14:42:38
U	Q3310-01 - 1	SS-01-0-1	40.6895	89538	PPB	SMPL	-7471B		09 Oct 2025 14:44:54
U	Q3310-02 - 1	SS-01-1-2	25.2602	55625	PPB	SMPL	-7471B		09 Oct 2025 14:47:27
U	Q3310-03 - 1	SS-01-2-3	26.9896	59426	PPB	SMPL	-7471B		09 Oct 2025 14:50:50
U	Q3313-01 - 1	ETGI-360	0.9523	2197	PPB	SMPL	-7471B		09 Oct 2025 14:54:03
U	Q3315-03 - 1	72-11989	1.2016	2745	PPB	SMPL	-7471B		09 Oct 2025 14:57:08
U	CCV11 - 1	CCV11	4.7617	10570	PPB	SMPL	-7471B		09 Oct 2025 14:59:26
U	CCB11 - 1	CCB11	-0.0855	-84	PPB	SMPL	-7471B		09 Oct 2025 15:04:47
U	Q3315-03DUP - 1	72-11989DUP	1.0510	2414	PPB	SMPL	-7471B		09 Oct 2025 15:07:02
U	Q3315-03MS - 1	72-11989MS	5.4369	12054	PPB	SMPL	-7471B		09 Oct 2025 15:09:17
U	Q3315-03MSD - 1	72-11989MSD	7.1799	15885	PPB	SMPL	-7471B		09 Oct 2025 15:11:35
U	Q3318-01 - 1	VNJ-207	1.5347	3477	PPB	SMPL	-7471B		09 Oct 2025 15:14:13
U	Q3319-01 - 1	OK-01-100825	0.3481	869	PPB	SMPL	-7471B		09 Oct 2025 15:16:50
U	Q3315-03LX5 - 1		0.2935	749	PPB	SMPL	-7471B		09 Oct 2025 15:21:55
U	Q3315-03A - 1		5.2808	11711	PPB	SMPL	-7471B		09 Oct 2025 15:27:13
U	Q3310-01DLX10 - 1		3.3595	7488	PPB	SMPL	-7471B		09 Oct 2025 15:29:31
U	Q3310-02DLX5 - 1		4.9610	11008	PPB	SMPL	-7471B		09 Oct 2025 15:35:22
U	Q3310-03DLX5 - 1		4.8559	10777	PPB	SMPL	-7471B		09 Oct 2025 15:37:39
U	CCV12 - 1	CCV12	4.7272	10494	PPB	SMPL	-7471B		09 Oct 2025 15:40:10
U	CCB12 - 1	CCB12	-0.0718	-54	PPB	SMPL	-7471B		09 Oct 2025 15:45:43