

Instrument ID: CV1

Daily Analysis Runlog For Sequence/QC Batch ID # LB137478

Review By	mohan	Review On	10/10/2025 9:38:07 AM
Supervise By	jaswal	Supervise On	10/10/2025 11:01:35 AM

STD. NAME	STD REF.#
ICAL Standard	MP87494,MP87496,MP87497,MP87498,MP87499,MP87500
ICV Standard	MP87501
CCV Standard	MP87503
ICSA Standard	
CRI Standard	MP87505
LCS Standard	
Chk Standard	MP87502,MP87504,MP87506,MP87509

Sr#	SampleId	ClientID	QcType	Date	Comment	Operator	Status
1	S0	S0	CAL1	10/09/25 13:33		mohan	OK
2	S0.2	S0.2	CAL2	10/09/25 13:36		mohan	OK
3	S2.5	S2.5	CAL3	10/09/25 13:38		mohan	OK
4	S5	S5	CAL4	10/09/25 13:40		mohan	OK
5	S7.5	S7.5	CAL5	10/09/25 13:52		mohan	OK
6	S10	S10	CAL6	10/09/25 14:09		mohan	OK
7	ICV65	ICV65	ICV	10/09/25 14:15		mohan	OK
8	ICB65	ICB65	ICB	10/09/25 14:20		mohan	OK
9	CCV10	CCV10	CCV	10/09/25 14:22		mohan	OK
10	CCB10	CCB10	CCB	10/09/25 14:27		mohan	OK
11	CRA	CRA	CRDL	10/09/25 14:30		mohan	OK
12	HighStd	HighStd	HIGH STD	10/09/25 14:32		mohan	OK
13	ChkStd	ChkStd	SAM	10/09/25 14:34		mohan	OK
14	PB170045BL	PB170045BL	MB	10/09/25 14:40		mohan	OK
15	PB170045BS	PB170045BS	LCS	10/09/25 14:42		mohan	OK
16	Q3310-01	SS-01-0-1	SAM	10/09/25 14:44	Hg high	mohan	Dilution
17	Q3310-02	SS-01-1-2	SAM	10/09/25 14:47	Hg high	mohan	Dilution
18	Q3310-03	SS-01-2-3	SAM	10/09/25 14:50	Hg high	mohan	Dilution

Instrument ID: CV1

Daily Analysis Runlog For Sequence/QC Batch ID # LB137478

Review By	mohan	Review On	10/10/2025 9:38:07 AM
Supervise By	jaswal	Supervise On	10/10/2025 11:01:35 AM
STD. NAME	STD REF.#		
ICAL Standard	MP87494,MP87496,MP87497,MP87498,MP87499,MP87500		
ICV Standard	MP87501		
CCV Standard	MP87503		
ICSA Standard			
CRI Standard	MP87505		
LCS Standard			
Chk Standard	MP87502,MP87504,MP87506,MP87509		

19	Q3313-01	ETGI-360	SAM	10/09/25 14:54		mohan	OK
20	Q3315-03	72-11989	SAM	10/09/25 14:57		mohan	OK
21	CCV11	CCV11	CCV	10/09/25 14:59		mohan	OK
22	CCB11	CCB11	CCB	10/09/25 15:04		mohan	OK
23	Q3315-03DUP	72-11989DUP	DUP	10/09/25 15:07		mohan	OK
24	Q3315-03MS	72-11989MS	MS	10/09/25 15:09		mohan	OK
25	Q3315-03MSD	72-11989MSD	MSD	10/09/25 15:11		mohan	OK
26	Q3318-01	VNJ-207	SAM	10/09/25 15:14		mohan	OK
27	Q3319-01	OK-01-100825	SAM	10/09/25 15:16		mohan	OK
28	Q3315-03L	72-11989L	SD	10/09/25 15:21		mohan	OK
29	Q3315-03A	72-11989A	PS	10/09/25 15:27		mohan	OK
30	Q3310-01DL	SS-01-0-1DL	SAM	10/09/25 15:29	10X For Hg	mohan	Confirms
31	Q3310-02DL	SS-01-1-2DL	SAM	10/09/25 15:35	5X For Hg	mohan	Confirms
32	Q3310-03DL	SS-01-2-3DL	SAM	10/09/25 15:37	5X For Hg	mohan	Confirms
33	CCV12	CCV12	CCV	10/09/25 15:40		mohan	OK
34	CCB12	CCB12	CCB	10/09/25 15:45		mohan	OK

Instrument ID: P4

Daily Analysis Runlog For Sequence/QC Batch ID # LB137500

Review By	Review On
Supervise By	Supervise On

STD. NAME	STD REF.#
ICAL Standard ICV Standard CCV Standard ICSA Standard CRI Standard LCS Standard Chk Standard	

Sr#	SampleId	ClientID	QcType	Date	Comment	Operator	Status
1	S0	S0	CAL1	10/10/25 11:27		Jaswal	OK
2	S1	S1	CAL2	10/10/25 11:31		Jaswal	OK
3	S2	S2	CAL3	10/10/25 11:36		Jaswal	OK
4	S3	S3	CAL4	10/10/25 11:40		Jaswal	OK
5	S4	S4	CAL5	10/10/25 11:44		Jaswal	OK
6	S5	S5	CAL6	10/10/25 11:48		Jaswal	OK
7	ICV01	ICV01	ICV	10/10/25 11:52		Jaswal	OK
8	LLICV01	LLICV01	LLICV	10/10/25 12:00		Jaswal	OK
9	ICB01	ICB01	ICB	10/10/25 12:11		Jaswal	OK
10	CRI01	CRI01	CRDL	10/10/25 12:15		Jaswal	OK
11	ICSA01	ICSA01	ICSA	10/10/25 12:29		Jaswal	OK
12	ICSAB01	ICSAB01	ICSAB	10/10/25 12:33		Jaswal	OK
13	ICSADL	ICSADL	ICSA	10/10/25 12:37		Jaswal	OK
14	ICSABDL	ICSABDL	ICSAB	10/10/25 12:41		Jaswal	OK
15	CCV01	CCV01	CCV	10/10/25 12:52		Jaswal	OK
16	CCB01	CCB01	CCB	10/10/25 12:57		Jaswal	OK
17	LR-2	LR-2	HIGH STD	10/10/25 13:07		Jaswal	OK
18	LR-3	LR-3	HIGH STD	10/10/25 13:11		Jaswal	OK

Instrument ID: P4

Daily Analysis Runlog For Sequence/QC Batch ID # LB137500

Review By	Review On
Supervise By	Supervise On

STD. NAME	STD REF.#
ICAL Standard ICV Standard CCV Standard ICSA Standard CRI Standard LCS Standard Chk Standard	

19	LR-1	LR-1	HIGH STD	10/10/25 13:20		Jaswal	OK
20	PB170047BL	PB170047BL	MB	10/10/25 13:25		Jaswal	OK
21	PB170047BS	PB170047BS	LCS	10/10/25 13:29		Jaswal	OK
22	PB170027TB	PB170027TB	MB	10/10/25 13:33		Jaswal	OK
23	Q3272-08	R7	SAM	10/10/25 13:38		Jaswal	OK
24	Q3315-02	314	SAM	10/10/25 13:42		Jaswal	OK
25	Q3315-02DUP	314DUP	DUP	10/10/25 13:46		Jaswal	OK
26	Q3315-02L	314L	SD	10/10/25 13:51		Jaswal	OK
27	CCV02	CCV02	CCV	10/10/25 14:04		Jaswal	OK
28	CCB02	CCB02	CCB	10/10/25 14:10		Jaswal	OK
29	Q3315-02MS	314MS	MS	10/10/25 14:15		Jaswal	OK
30	Q3315-02MSD	314MSD	MSD	10/10/25 14:19		Jaswal	OK
31	Q3315-02A	314A	PS	10/10/25 14:23		Jaswal	OK
32	PB170008BL	PB170008BL	MB	10/10/25 14:27		Jaswal	OK
33	PB170008BS	PB170008BS	LCS	10/10/25 14:31		Jaswal	OK
34	PB170058BL	PB170058BL	MB	10/10/25 14:35		Jaswal	OK
35	Q3310-01	SS-01-0-1	SAM	10/10/25 14:44		Jaswal	OK
36	PB170058BS	PB170058BS	LCS	10/10/25 14:48		Jaswal	OK
37	Q3310-02	SS-01-1-2	SAM	10/10/25 14:52		Jaswal	OK
38	Q3310-03	SS-01-2-3	SAM	10/10/25 14:56	Ag oversaturated	Jaswal	Dilution

Instrument ID: P4

Daily Analysis Runlog For Sequence/QC Batch ID # LB137500

Review By	Review On
Supervise By	Supervise On

STD. NAME	STD REF.#
ICAL Standard ICV Standard CCV Standard ICSA Standard CRI Standard LCS Standard Chk Standard	

39	CCV03	CCV03	CCV	10/10/25 15:37		Jaswal	OK
40	CCB03	CCB03	CCB	10/10/25 15:41		Jaswal	OK
41	Q3321-01	P001-DS-01	SAM	10/10/25 15:45	Cu,Fe high & Ag oversaturated	Jaswal	Dilution
42	Q3322-01	OR-02-100925	SAM	10/10/25 15:50		Jaswal	OK
43	Q3322-01DUP	OR-02-100925DUP	DUP	10/10/25 15:54		Jaswal	OK
44	Q3322-01L	OR-02-100925L	SD	10/10/25 15:58		Jaswal	OK
45	Q3322-01MS	OR-02-100925MS	MS	10/10/25 16:02		Jaswal	OK
46	Q3322-01MSD	OR-02-100925MSD	MSD	10/10/25 16:06		Jaswal	OK
47	Q3322-01A	OR-02-100925A	PS	10/10/25 16:09		Jaswal	OK
48	Q3321-01DL	P001-DS-01DL	SAM	10/10/25 16:13	5x for Cu,Fe,Ag	Jaswal	Confirms
49	CCV04	CCV04	CCV	10/10/25 16:29		Jaswal	OK
50	CCB04	CCB04	CCB	10/10/25 16:33		Jaswal	OK
51	Q3310-01DL	SS-01-0-1DL	SAM	10/10/25 16:49	NOT USE	Jaswal	Not Ok
52	Q3310-02DL	SS-01-1-2DL	SAM	10/10/25 16:53	NOT USE	Jaswal	Not Ok
53	Q3310-03DL	SS-01-2-3DL	SAM	10/10/25 16:57	2x for Ag	Jaswal	Confirms
54	CCV05	CCV05	CCV	10/10/25 17:04		Jaswal	OK
55	CCB05	CCB05	CCB	10/10/25 17:08		Jaswal	OK