

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance

Contract: LANG01

Lab Code: ACE

SDG NO.: Q3310

Continuing Calib Date: 10/10/2025

Initial Calibration Date(s): 09/23/2025

09/23/2025

Continuing Calib Time: 18:55

Initial Calibration Time(s): 08:45

17:14

GC Column: ZB-MR1 ID: 0.32 (mm)

COMPOUND		CCAL RT	AVG RT	RT WINDOW		DIFF RT
				FROM	TO	
Aroclor-1016-1	(1)	5.81	5.81	5.71	5.91	0.01
Aroclor-1016-2	(2)	5.83	5.83	5.73	5.93	0.00
Aroclor-1016-3	(3)	5.89	5.89	5.79	5.99	0.00
Aroclor-1016-4	(4)	5.99	5.99	5.89	6.09	0.00
Aroclor-1016-5	(5)	6.28	6.28	6.18	6.38	0.00
Aroclor-1260-1	(1)	7.39	7.40	7.30	7.50	0.01
Aroclor-1260-2	(2)	7.65	7.65	7.55	7.75	0.00
Aroclor-1260-3	(3)	8.01	8.01	7.91	8.11	0.00
Aroclor-1260-4	(4)	8.23	8.24	8.14	8.34	0.01
Aroclor-1260-5	(5)	8.56	8.56	8.46	8.66	0.00
Tetrachloro-m-xylene		4.65	4.66	4.56	4.76	0.01
Decachlorobiphenyl		10.43	10.43	10.33	10.53	0.00

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance

Contract: LANG01

Lab Code: ACE

SDG NO.: Q3310

Continuing Calib Date: 10/10/2025

Initial Calibration Date(s): 09/23/2025

09/23/2025

Continuing Calib Time: 18:55

Initial Calibration Time(s): 08:45

17:14

GC Column: ZB-MR2 ID: 0.32 (mm)

COMPOUND		CCAL RT	AVG RT	RT WINDOW FROM TO		DIFF RT
Aroclor-1016-1	(1)	4.89	4.89	4.79	4.99	0.00
Aroclor-1016-2	(2)	4.95	4.95	4.85	5.05	0.00
Aroclor-1016-3	(3)	5.07	5.07	4.97	5.17	0.00
Aroclor-1016-4	(4)	5.11	5.11	5.01	5.21	0.00
Aroclor-1016-5	(5)	5.32	5.33	5.23	5.43	0.01
Aroclor-1260-1	(1)	6.35	6.36	6.26	6.46	0.01
Aroclor-1260-2	(2)	6.54	6.54	6.44	6.64	0.01
Aroclor-1260-3	(3)	6.69	6.70	6.60	6.80	0.01
Aroclor-1260-4	(4)	7.16	7.17	7.07	7.27	0.01
Aroclor-1260-5	(5)	7.40	7.40	7.30	7.50	0.00
Tetrachloro-m-xylene		3.79	3.80	3.70	3.90	0.01
Decachlorobiphenyl		8.79	8.80	8.70	8.90	0.01