

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** LANG01
Lab Code: ACE **SDG NO.:** Q3310
GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 09/23/2025 09/23/2025

Client Sample No.: CCAL03 **Date Analyzed:** 10/15/2025
Lab Sample No.: AR1660CCC500 **Data File :** PP075786.D **Time Analyzed:** 09:46

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	5.806	5.708	5.908	519.890	500.000	4.0
Aroclor-1016-2	5.828	5.729	5.929	529.600	500.000	5.9
Aroclor-1016-3	5.891	5.792	5.992	503.430	500.000	0.7
Aroclor-1016-4	5.988	5.889	6.089	537.410	500.000	7.5
Aroclor-1016-5	6.280	6.181	6.381	557.250	500.000	11.5
Aroclor-1260-1	7.397	7.299	7.499	516.730	500.000	3.3
Aroclor-1260-2	7.649	7.551	7.751	508.790	500.000	1.8
Aroclor-1260-3	8.008	7.909	8.109	552.740	500.000	10.5
Aroclor-1260-4	8.236	8.137	8.337	572.880	500.000	14.6
Aroclor-1260-5	8.561	8.464	8.664	561.840	500.000	12.4
Decachlorobiphenyl	10.431	10.334	10.534	43.700	50.000	-12.6
Tetrachloro-m-xylene	4.654	4.555	4.755	47.940	50.000	-4.1

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** LANG01
Lab Code: ACE **SDG NO.:** Q3310
GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 09/23/2025 09/23/2025

Client Sample No.: CCAL03 **Date Analyzed:** 10/15/2025
Lab Sample No.: AR1660CCC500 **Data File :** PP075786.D **Time Analyzed:** 09:46

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.886	4.793	4.993	525.300	500.000	5.1
Aroclor-1016-2	4.943	4.851	5.051	548.930	500.000	9.8
Aroclor-1016-3	5.064	4.971	5.171	463.710	500.000	-7.3
Aroclor-1016-4	5.104	5.011	5.211	526.750	500.000	5.4
Aroclor-1016-5	5.317	5.225	5.425	517.290	500.000	3.5
Aroclor-1260-1	6.346	6.255	6.455	496.140	500.000	-0.8
Aroclor-1260-2	6.533	6.441	6.641	494.070	500.000	-1.2
Aroclor-1260-3	6.686	6.595	6.795	491.330	500.000	-1.7
Aroclor-1260-4	7.155	7.065	7.265	454.800	500.000	-9.0
Aroclor-1260-5	7.394	7.304	7.504	474.360	500.000	-5.1
Decachlorobiphenyl	8.791	8.703	8.903	46.510	50.000	-7.0
Tetrachloro-m-xylene	3.790	3.696	3.896	47.550	50.000	-4.9