

Surrogate Summary

SDG No.: Q3310

Client: Langan Engineering and Environmental

Analytical Method: 8082A

Lab Sample ID	Client ID	Parameter	Column	Spike	Result	Recovery(%)	Qual	Limits(%)	
								Low	High
I.BLK-PP075246.D	PIBLK-PP075246.D	Tetrachloro-m-xyl	1	20	19.9	100		60	140
		Decachlorobiphen	1	20	19.4	97		60	140
		Tetrachloro-m-xyl	2	20	17.9	90		60	140
		Decachlorobiphen	2	20	20.0	100		60	140
I.BLK-PP075708.D	PIBLK-PP075708.D	Tetrachloro-m-xyl	1	20	20.1	100		60	140
		Decachlorobiphen	1	20	20.8	104		60	140
		Tetrachloro-m-xyl	2	20	16.4	82		60	140
		Decachlorobiphen	2	20	20.6	103		60	140
PB170053BL	PB170053BL	Tetrachloro-m-xyl	1	20	24.7	123		32	144
		Decachlorobiphen	1	20	24.5	123		32	175
		Tetrachloro-m-xyl	2	20	23.4	117		32	144
		Decachlorobiphen	2	20	25.3	126		32	175
Q3310-01	SS-01-0-1	Tetrachloro-m-xyl	1	20	18.5	92		32	144
		Decachlorobiphen	1	20	15.1	76		32	175
		Tetrachloro-m-xyl	2	20	20.9	105		32	144
		Decachlorobiphen	2	20	14.2	71		32	175
Q3310-02	SS-01-1-2	Tetrachloro-m-xyl	1	20	18.0	90		32	144
		Decachlorobiphen	1	20	15.8	79		32	175
		Tetrachloro-m-xyl	2	20	17.9	89		32	144
		Decachlorobiphen	2	20	13.6	68		32	175
Q3310-02MS	SS-01-1-2MS	Tetrachloro-m-xyl	1	20	20.8	104		32	144
		Decachlorobiphen	1	20	16.8	84		32	175
		Tetrachloro-m-xyl	2	20	19.3	97		32	144
		Decachlorobiphen	2	20	14.9	75		32	175
Q3310-02MSD	SS-01-1-2MSD	Tetrachloro-m-xyl	1	20	20.6	103		32	144
		Decachlorobiphen	1	20	17.0	85		32	175
		Tetrachloro-m-xyl	2	20	19.8	99		32	144
		Decachlorobiphen	2	20	16.1	81		32	175
Q3310-03	SS-01-2-3	Tetrachloro-m-xyl	1	20	18.9	95		32	144
		Decachlorobiphen	1	20	14.9	74		32	175
		Tetrachloro-m-xyl	2	20	18.7	93		32	144
		Decachlorobiphen	2	20	12.5	63		32	175
I.BLK-PP075723.D	PIBLK-PP075723.D	Tetrachloro-m-xyl	1	20	19.9	100		60	140
		Decachlorobiphen	1	20	20.3	101		60	140
		Tetrachloro-m-xyl	2	20	18.7	94		60	140
		Decachlorobiphen	2	20	20.8	104		60	140
I.BLK-PP075790.D	PIBLK-PP075790.D	Tetrachloro-m-xyl	1	20	21.9	110		60	140
		Decachlorobiphen	1	20	21.0	105		60	140
		Tetrachloro-m-xyl	2	20	16.4	82		60	140
		Decachlorobiphen	2	20	22.1	110		60	140
PB170053BS	PB170053BS	Tetrachloro-m-xyl	1	20	15.9	80		32	144

Surrogate Summary

SDG No.: Q3310

Client: Langan Engineering and Environmental

Analytical Method: 8082A

Lab Sample ID	Client ID	Parameter	Column	Spike	Result	Recovery(%)	Qual	Limits(%)	
								Low	High
PB170053BS	PB170053BS	Decachlorobiphen	1	20	14.5	73		32	175
		Tetrachloro-m-xyl	2	20	12.7	64		32	144
		Decachlorobiphen	2	20	14.8	74		32	175
I.BLK-PP075805.D	PIBLK-PP075805.D	Tetrachloro-m-xyl	1	20	21.2	106		60	140
		Decachlorobiphen	1	20	20.1	101		60	140
		Tetrachloro-m-xyl	2	20	18.2	91		60	140
		Decachlorobiphen	2	20	21.2	106		60	140