



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: Q3310

Analytical Method: 8082A

Client: Langan Engineering and Environmental S

Datafile : PP075791.D

Lab Sample ID	Parameter	Spike	Result	Units	Rec	RPD	Qual	RPD	Limits		RPD
								Qual	Low	High	
PB170053BS (Column 1)	AR1016	166.6	130	ug/kg	78				71	120	
	AR1260	166.6	127	ug/kg	76				65	130	
PB170053BS (Column 2)	AR1016	166.6	121	ug/kg	73				71	120	
	AR1260	166.6	110	ug/kg	66				65	130	