

Report of Analysis

Client:	Langan Engineering and Environmental Services, Inc		Date Collected:	10/07/25	
Project:	Con Edison Richmond Terrace		Date Received:	10/08/25	
Client Sample ID:	SS-01-2-3		SDG No.:	Q3310	
Lab Sample ID:	Q3310-03		Matrix:	SOIL	
Analytical Method:	8260D		% Solid:	75.1	
Sample Wt/Vol:	8.17	Units: g	Final Vol:	5000	uL
Soil Aliquot Vol:		uL	Test:	VOC-TCLVOA-10	
GC Column:	RXI-624	ID : 0.25	Level :	LOW	
Prep Method :					

File ID/Qc Batch:	Dilution:	Date Analyzed	Prep Batch ID
VY023418.D	1	10/09/25 13:21	VY100925

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units(Dry Weight)
TARGETS						
75-71-8	Dichlorodifluoromethane	0.93	U	0.93	4.10	ug/Kg
74-87-3	Chloromethane	0.93	U	0.93	4.10	ug/Kg
75-01-4	Vinyl Chloride	0.64	U	0.64	4.10	ug/Kg
74-83-9	Bromomethane	0.87	U	0.87	4.10	ug/Kg
75-00-3	Chloroethane	1.00	U	1.00	4.10	ug/Kg
75-69-4	Trichlorofluoromethane	0.99	U	0.99	4.10	ug/Kg
76-13-1	1,1,2-Trichlorotrifluoroethane	0.86	U	0.86	4.10	ug/Kg
75-35-4	1,1-Dichloroethene	0.81	U	0.81	4.10	ug/Kg
67-64-1	Acetone	10.2	J	3.90	20.4	ug/Kg
75-15-0	Carbon Disulfide	0.86	U	0.86	4.10	ug/Kg
1634-04-4	Methyl tert-butyl Ether	0.59	U	0.59	4.10	ug/Kg
79-20-9	Methyl Acetate	1.30	U	1.30	4.10	ug/Kg
75-09-2	Methylene Chloride	3.00	J	2.90	8.10	ug/Kg
156-60-5	trans-1,2-Dichloroethene	0.70	U	0.70	4.10	ug/Kg
75-34-3	1,1-Dichloroethane	0.65	U	0.65	4.10	ug/Kg
110-82-7	Cyclohexane	0.64	U	0.64	4.10	ug/Kg
78-93-3	2-Butanone	5.30	U	5.30	20.4	ug/Kg
56-23-5	Carbon Tetrachloride	0.79	U	0.79	4.10	ug/Kg
156-59-2	cis-1,2-Dichloroethene	0.61	U	0.61	4.10	ug/Kg
74-97-5	Bromochloromethane	0.94	U	0.94	4.10	ug/Kg
67-66-3	Chloroform	0.68	U	0.68	4.10	ug/Kg
71-55-6	1,1,1-Trichloroethane	0.76	U	0.76	4.10	ug/Kg
108-87-2	Methylcyclohexane	0.74	U	0.74	4.10	ug/Kg
71-43-2	Benzene	0.64	U	0.64	4.10	ug/Kg
107-06-2	1,2-Dichloroethane	0.64	U	0.64	4.10	ug/Kg
79-01-6	Trichloroethene	0.66	U	0.66	4.10	ug/Kg
78-87-5	1,2-Dichloropropane	0.74	U	0.74	4.10	ug/Kg
75-27-4	Bromodichloromethane	0.64	U	0.64	4.10	ug/Kg
108-10-1	4-Methyl-2-Pentanone	2.90	U	2.90	20.4	ug/Kg
108-88-3	Toluene	0.64	U	0.64	4.10	ug/Kg

Report of Analysis

Client:	Langan Engineering and Environmental Services, Inc		Date Collected:	10/07/25	
Project:	Con Edison Richmond Terrace		Date Received:	10/08/25	
Client Sample ID:	SS-01-2-3		SDG No.:	Q3310	
Lab Sample ID:	Q3310-03		Matrix:	SOIL	
Analytical Method:	8260D		% Solid:	75.1	
Sample Wt/Vol:	8.17	Units: g	Final Vol:	5000	uL
Soil Aliquot Vol:		uL	Test:	VOC-TCLVOA-10	
GC Column:	RXI-624	ID : 0.25	Level :	LOW	
Prep Method :					

File ID/Qc Batch:	Dilution:	Date Analyzed	Prep Batch ID
VY023418.D	1	10/09/25 13:21	VY100925

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units(Dry Weight)
10061-02-6	t-1,3-Dichloropropene	0.53	U	0.53	4.10	ug/Kg
10061-01-5	cis-1,3-Dichloropropene	0.51	U	0.51	4.10	ug/Kg
79-00-5	1,1,2-Trichloroethane	0.75	U	0.75	4.10	ug/Kg
591-78-6	2-Hexanone	3.00	U	3.00	20.4	ug/Kg
124-48-1	Dibromochloromethane	0.71	U	0.71	4.10	ug/Kg
106-93-4	1,2-Dibromoethane	0.72	U	0.72	4.10	ug/Kg
127-18-4	Tetrachloroethene	0.86	U	0.86	4.10	ug/Kg
108-90-7	Chlorobenzene	0.74	U	0.74	4.10	ug/Kg
100-41-4	Ethyl Benzene	0.55	U	0.55	4.10	ug/Kg
179601-23-1	m/p-Xylenes	1.00	U	1.00	8.10	ug/Kg
95-47-6	o-Xylene	0.67	U	0.67	4.10	ug/Kg
100-42-5	Styrene	0.58	U	0.58	4.10	ug/Kg
75-25-2	Bromoform	0.70	U	0.70	4.10	ug/Kg
98-82-8	Isopropylbenzene	0.64	U	0.64	4.10	ug/Kg
79-34-5	1,1,2,2-Tetrachloroethane	0.99	U	0.99	4.10	ug/Kg
541-73-1	1,3-Dichlorobenzene	1.40	U	1.40	4.10	ug/Kg
106-46-7	1,4-Dichlorobenzene	1.30	U	1.30	4.10	ug/Kg
95-50-1	1,2-Dichlorobenzene	1.20	U	1.20	4.10	ug/Kg
96-12-8	1,2-Dibromo-3-Chloropropane	1.50	U	1.50	4.10	ug/Kg
120-82-1	1,2,4-Trichlorobenzene	2.40	U	2.40	4.10	ug/Kg
87-61-6	1,2,3-Trichlorobenzene	2.60	U	2.60	4.10	ug/Kg
SURROGATES						
17060-07-0	1,2-Dichloroethane-d4	59.8		63 - 155	120%	SPK: 50
1868-53-7	Dibromofluoromethane	54.9		70 - 134	110%	SPK: 50
2037-26-5	Toluene-d8	52.0		74 - 123	104%	SPK: 50
460-00-4	4-Bromofluorobenzene	42.4		17 - 146	85%	SPK: 50
INTERNAL STANDARDS						
363-72-4	Pentafluorobenzene	579000	7.713			
540-36-3	1,4-Difluorobenzene	1100000	8.616			
3114-55-4	Chlorobenzene-d5	964000	11.414			
3855-82-1	1,4-Dichlorobenzene-d4	311000	13.346			

Report of Analysis

Client:	Langan Engineering and Environmental Services, Inc		Date Collected:	10/07/25
Project:	Con Edison Richmond Terrace		Date Received:	10/08/25
Client Sample ID:	SS-01-2-3		SDG No.:	Q3310
Lab Sample ID:	Q3310-03		Matrix:	SOIL
Analytical Method:	8260D		% Solid:	75.1
Sample Wt/Vol:	8.17	Units: g	Final Vol:	5000 uL
Soil Aliquot Vol:		uL	Test:	VOC-TCLVOA-10
GC Column:	RXI-624	ID : 0.25	Level :	LOW
Prep Method :				

File ID/Qc Batch:	Dilution:	Date Analyzed	Prep Batch ID
VY023418.D	1	10/09/25 13:21	VY100925

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units
------------	-----------	-------	-----------	-----	------------	-------

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products