

Instrument ID: MSVOA_Y

Daily Analysis Runlog For Sequence/QC Batch ID # VY100725

Review By	Semsettin Yesilyurt	Review On	10/10/2025 3:00:22 AM
Supervise By	Maresh Dadoda	Supervise On	10/10/2025 4:38:02 AM
SubDirectory	VY100725	HP Acquire Method	MSVOA_Y HP Processing Method 82y100725s.m

STD. NAME	STD REF.#
Tune/Reschk Initial Calibration Stds	VP135766 VP135767,VP135768,VP135769,VP135770,VP135771,VP135772
CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	VP133934 VP135773

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	BFB	BFB	VY023383.D	07 Oct 2025 08:43		SY/MD	Ok
2	VSTDICC005	VSTDICC005	VY023384.D	07 Oct 2025 09:17		SY/MD	Ok,M
3	VSTDICC010	VSTDICC010	VY023385.D	07 Oct 2025 10:02		SY/MD	Ok,M
4	VSTDICC020	VSTDICC020	VY023386.D	07 Oct 2025 11:24		SY/MD	Ok,M
5	VSTDICCC050	VSTDICCC050	VY023387.D	07 Oct 2025 11:47		SY/MD	Ok,M
6	VSTDICC100	VSTDICC100	VY023388.D	07 Oct 2025 12:09		SY/MD	Ok,M
7	VSTDICC150	VSTDICC150	VY023389.D	07 Oct 2025 12:32		SY/MD	Ok,M
8	VIBLK	VIBLK	VY023390.D	07 Oct 2025 13:02		SY/MD	Ok
9	VSTDICV050	ICVVY100725	VY023391.D	07 Oct 2025 14:12		SY/MD	Ok,M

M : Manual Integration

Instrument ID: MSVOA_Y

Daily Analysis Runlog For Sequence/QC Batch ID # VY100925

Review By	Maresh Dadoda	Review On	10/10/2025 4:18:26 PM
Supervise By	Semsettin Yesilyurt	Supervise On	10/10/2025 4:19:06 PM
SubDirectory	VY100925	HP Acquire Method	MSVOA_Y HP Processing Method 82y100924s.m

STD. NAME	STD REF.#
Tune/Reschk Initial Calibration Stds	VP135799
CCC Internal Standard/PEM ICV/I.BLK Surrogate Standard MS/MSD Standard LCS Standard	VP135800,VP135801 VP133934

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	BFB	BFB	VY023406.D	09 Oct 2025 08:16		SY/MD	Ok
2	VSTDCCC050	VSTDCCC050	VY023407.D	09 Oct 2025 08:47		SY/MD	Ok,M
3	VY1009SBL01	VY1009SBL01	VY023408.D	09 Oct 2025 09:16		SY/MD	Ok
4	VY1009SBS01	VY1009SBS01	VY023409.D	09 Oct 2025 09:45		SY/MD	Ok,M
5	VY1009SBSD01	VY1009SBSD01	VY023410.D	09 Oct 2025 10:08		SY/MD	Ok,M
6	Q3315-01	314	VY023411.D	09 Oct 2025 10:36	vial-A Internal Standard fail;Need MeOH	SY/MD	Dilution
7	Q3315-03	72-11989	VY023412.D	09 Oct 2025 11:00	vial-A E flag of previous sample	SY/MD	ReRun
8	Q3313-01	ETGI-360	VY023413.D	09 Oct 2025 11:23	vial-A Internal Standard fail;Confirm concentration	SY/MD	Not Ok
9	Q3318-01	VNJ-207	VY023414.D	09 Oct 2025 11:47	vial-A	SY/MD	Ok
10	Q3319-01	OK-01-100825	VY023415.D	09 Oct 2025 12:10	vial-A Internal Standard fail	SY/MD	ReRun
11	Q3310-01	SS-01-0-1	VY023416.D	09 Oct 2025 12:34	vial-A	SY/MD	Ok
12	Q3310-02	SS-01-1-2	VY023417.D	09 Oct 2025 12:57	vial-A	SY/MD	Ok
13	Q3310-03	SS-01-2-3	VY023418.D	09 Oct 2025 13:21	vial-A	SY/MD	Ok
14	Q3321-01	P001-DS-01	VY023419.D	09 Oct 2025 13:44	vial-A	SY/MD	Ok
15	Q3322-01	OR-02-100925	VY023420.D	09 Oct 2025 14:36	vial-A	SY/MD	Ok
16	VIBLK	VIBLK	VY023421.D	09 Oct 2025 14:59		SY/MD	Ok
17	VSTDCCC050	VSTDCCC050EC	VY023422.D	09 Oct 2025 15:22		SY/MD	Ok,M

M : Manual Integration