

Surrogate Summary

SDG No.: **Q3310**

Client: **Langan Engineering and Environmental Services, Inc**

Analytical Method: **SW8260D**

Lab Sample ID	Client ID	Parameter	Spike	Result	Recovery (%)	Qual	Limits (%)	
							Low	High
Q3310-01	SS-01-0-1	1,2-Dichloroethane-d4	50	63.5	127		63	155
		Dibromofluoromethane	50	55.2	110		70	134
		Toluene-d8	50	50.3	101		74	123
		4-Bromofluorobenzene	50	36.4	73		17	146
Q3310-02	SS-01-1-2	1,2-Dichloroethane-d4	50	60.1	120		63	155
		Dibromofluoromethane	50	53.4	107		70	134
		Toluene-d8	50	51.2	102		74	123
		4-Bromofluorobenzene	50	40.5	81		17	146
Q3310-03	SS-01-2-3	1,2-Dichloroethane-d4	50	59.8	120		63	155
		Dibromofluoromethane	50	54.9	110		70	134
		Toluene-d8	50	52.0	104		74	123
		4-Bromofluorobenzene	50	42.4	85		17	146
VY1009SBL01	VY1009SBL01	1,2-Dichloroethane-d4	50	63.4	127		63	155
		Dibromofluoromethane	50	54.5	109		70	134
		Toluene-d8	50	52.1	104		74	123
		4-Bromofluorobenzene	50	49.8	100		17	146
VY1009SBS01	VY1009SBS01	1,2-Dichloroethane-d4	50	45.8	92		63	155
		Dibromofluoromethane	50	47.9	96		70	134
		Toluene-d8	50	47.9	96		74	123
		4-Bromofluorobenzene	50	46.0	92		17	146
VY1009SBSD01	VY1009SBSD01	1,2-Dichloroethane-d4	50	47.2	94		63	155
		Dibromofluoromethane	50	49.5	99		70	134
		Toluene-d8	50	50.4	101		74	123
		4-Bromofluorobenzene	50	47.7	95		17	146