

VOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
BROMOFLUOROBENZENE (BFB)

Lab Name: Alliance

Contract: LANG01

Lab Code: ACE

SDG NO.: Q3310

Lab File ID: VY023406.D

BFB Injection Date: 10/09/2025

Instrument ID: MSVOA_Y

BFB Injection Time: 08:16

GC Column: RXI-624 ID: 0.25 (mm)

Heated Purge: Y/N Y

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
50	15.0 - 40.0% of mass 95	16.8
75	30.0 - 60.0% of mass 95	49.1
95	Base Peak, 100% relative abundance	100
96	5.0 - 9.0% of mass 95	6.8
173	Less than 2.0% of mass 174	1.1 (1.1) 1
174	50.0 - 100.0% of mass 95	95
175	5.0 - 9.0% of mass 174	6.9 (7.3) 1
176	95.0 - 101.0% of mass 174	91.2 (95.9) 1
177	5.0 - 9.0% of mass 176	6.3 (6.9) 2

1-Value is % mass 174

2-Value is % mass 176

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

CLIENT ID	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
VSTDCCC050	VSTDCCC050	VY023407.D	10/09/2025	08:47
VY1009SBL01	VY1009SBL01	VY023408.D	10/09/2025	09:16
VY1009SBS01	VY1009SBS01	VY023409.D	10/09/2025	09:45
VY1009SBSD01	VY1009SBSD01	VY023410.D	10/09/2025	10:08
SS-01-0-1	Q3310-01	VY023416.D	10/09/2025	12:34
SS-01-1-2	Q3310-02	VY023417.D	10/09/2025	12:57
SS-01-2-3	Q3310-03	VY023418.D	10/09/2025	13:21