

Report of Analysis

Client: Core Environmental Consultants and Services, Inc.
Project: The Wills Building
Client Sample ID: VX1017WBL01
Lab Sample ID: VX1017WBL01
Analytical Method: 8260D
Sample Wt/Vol: 5 mL

Level : LOW
Final Vol: 5000 uL

Date Collected:
Date Received:
SDG No.: Q3311
Matrix: Water
% Solid: 0
Test: VOC-TCLVOA-10

CAS Number	Parameter	Conc.	Qua.	DF	MDL	LOQ / CRQL	Units	Date Ana.	BatchID
TARGETS									
75-71-8	Dichlorodifluoromethane	1.00	U	1	0.22	1.00	ug/L	10/17/25 13:15	VX101725
74-87-3	Chloromethane	1.00	U	1	0.32	1.00	ug/L	10/17/25 13:15	VX101725
75-01-4	Vinyl Chloride	1.00	U	1	0.26	1.00	ug/L	10/17/25 13:15	VX101725
74-83-9	Bromomethane	5.00	U	1	1.40	5.00	ug/L	10/17/25 13:15	VX101725
75-00-3	Chloroethane	1.00	U	1	0.47	1.00	ug/L	10/17/25 13:15	VX101725
75-69-4	Trichlorofluoromethane	1.00	U	1	0.33	1.00	ug/L	10/17/25 13:15	VX101725
76-13-1	1,1,2-Trichlorotrifluoroethane	1.00	U	1	0.25	1.00	ug/L	10/17/25 13:15	VX101725
75-35-4	1,1-Dichloroethene	1.00	U	1	0.23	1.00	ug/L	10/17/25 13:15	VX101725
67-64-1	Acetone	5.00	U	1	1.50	5.00	ug/L	10/17/25 13:15	VX101725
75-15-0	Carbon Disulfide	1.00	U	1	0.21	1.00	ug/L	10/17/25 13:15	VX101725
1634-04-4	Methyl tert-butyl Ether	1.00	U	1	0.16	1.00	ug/L	10/17/25 13:15	VX101725
79-20-9	Methyl Acetate	1.00	U	1	0.27	1.00	ug/L	10/17/25 13:15	VX101725
75-09-2	Methylene Chloride	1.00	U	1	0.28	1.00	ug/L	10/17/25 13:15	VX101725
156-60-5	trans-1,2-Dichloroethene	1.00	U	1	0.23	1.00	ug/L	10/17/25 13:15	VX101725
75-34-3	1,1-Dichloroethane	1.00	U	1	0.23	1.00	ug/L	10/17/25 13:15	VX101725
110-82-7	Cyclohexane	5.00	U	1	1.50	5.00	ug/L	10/17/25 13:15	VX101725
78-93-3	2-Butanone	5.00	U	1	0.98	5.00	ug/L	10/17/25 13:15	VX101725
56-23-5	Carbon Tetrachloride	1.00	U	1	0.25	1.00	ug/L	10/17/25 13:15	VX101725
156-59-2	cis-1,2-Dichloroethene	1.00	U	1	0.19	1.00	ug/L	10/17/25 13:15	VX101725
74-97-5	Bromochloromethane	1.00	U	1	0.22	1.00	ug/L	10/17/25 13:15	VX101725
67-66-3	Chloroform	1.00	U	1	0.25	1.00	ug/L	10/17/25 13:15	VX101725
71-55-6	1,1,1-Trichloroethane	1.00	U	1	0.20	1.00	ug/L	10/17/25 13:15	VX101725
108-87-2	Methylcyclohexane	1.00	U	1	0.16	1.00	ug/L	10/17/25 13:15	VX101725
71-43-2	Benzene	1.00	U	1	0.15	1.00	ug/L	10/17/25 13:15	VX101725
107-06-2	1,2-Dichloroethane	1.00	U	1	0.22	1.00	ug/L	10/17/25 13:15	VX101725
79-01-6	Trichloroethene	1.00	U	1	0.090	1.00	ug/L	10/17/25 13:15	VX101725
78-87-5	1,2-Dichloropropane	1.00	U	1	0.20	1.00	ug/L	10/17/25 13:15	VX101725
75-27-4	Bromodichloromethane	1.00	U	1	0.22	1.00	ug/L	10/17/25 13:15	VX101725
108-10-1	4-Methyl-2-Pentanone	5.00	U	1	0.68	5.00	ug/L	10/17/25 13:15	VX101725
108-88-3	Toluene	1.00	U	1	0.14	1.00	ug/L	10/17/25 13:15	VX101725
10061-02-6	t-1,3-Dichloropropene	1.00	U	1	0.17	1.00	ug/L	10/17/25 13:15	VX101725
10061-01-5	cis-1,3-Dichloropropene	1.00	U	1	0.16	1.00	ug/L	10/17/25 13:15	VX101725
79-00-5	1,1,2-Trichloroethane	1.00	U	1	0.21	1.00	ug/L	10/17/25 13:15	VX101725
591-78-6	2-Hexanone	5.00	U	1	0.89	5.00	ug/L	10/17/25 13:15	VX101725
124-48-1	Dibromochloromethane	1.00	U	1	0.18	1.00	ug/L	10/17/25 13:15	VX101725
106-93-4	1,2-Dibromoethane	1.00	U	1	0.15	1.00	ug/L	10/17/25 13:15	VX101725
127-18-4	Tetrachloroethene	1.00	U	1	0.23	1.00	ug/L	10/17/25 13:15	VX101725
108-90-7	Chlorobenzene	1.00	U	1	0.12	1.00	ug/L	10/17/25 13:15	VX101725
100-41-4	Ethyl Benzene	1.00	U	1	0.13	1.00	ug/L	10/17/25 13:15	VX101725
179601-23-1	m/p-Xylenes	2.00	U	1	0.24	2.00	ug/L	10/17/25 13:15	VX101725
95-47-6	o-Xylene	1.00	U	1	0.12	1.00	ug/L	10/17/25 13:15	VX101725
100-42-5	Styrene	1.00	U	1	0.15	1.00	ug/L	10/17/25 13:15	VX101725
75-25-2	Bromoform	1.00	U	1	0.19	1.00	ug/L	10/17/25 13:15	VX101725

Report of Analysis

Client: Core Environmental Consultants and Services, Inc.
Project: The Wills Building
Client Sample ID: VX1017WBL01
Lab Sample ID: VX1017WBL01
Analytical Method: 8260D
Sample Wt/Vol: 5 mL

Level : LOW
Final Vol: 5000 uL

Date Collected:
Date Received:
SDG No.: Q3311
Matrix: Water
% Solid: 0
Test: VOC-TCLVOA-10

CAS Number	Parameter	Conc.	Qua.	DF	MDL	LOQ / CRQL	Units	Date Ana.	BatchID
98-82-8	Isopropylbenzene	1.00	U	1	0.12	1.00	ug/L	10/17/25 13:15	VX101725
79-34-5	1,1,2,2-Tetrachloroethane	1.00	U	1	0.26	1.00	ug/L	10/17/25 13:15	VX101725
541-73-1	1,3-Dichlorobenzene	1.00	U	1	0.16	1.00	ug/L	10/17/25 13:15	VX101725
106-46-7	1,4-Dichlorobenzene	1.00	U	1	0.19	1.00	ug/L	10/17/25 13:15	VX101725
95-50-1	1,2-Dichlorobenzene	1.00	U	1	0.16	1.00	ug/L	10/17/25 13:15	VX101725
96-12-8	1,2-Dibromo-3-Chloropropane	1.00	U	1	0.53	1.00	ug/L	10/17/25 13:15	VX101725
120-82-1	1,2,4-Trichlorobenzene	1.00	U	1	0.20	1.00	ug/L	10/17/25 13:15	VX101725
87-61-6	1,2,3-Trichlorobenzene	1.00	U	1	0.20	1.00	ug/L	10/17/25 13:15	VX101725
SURROGATES									
17060-07-0	1,2-Dichloroethane-d4	53.1			74 - 125	106%	SPK: 50		
1868-53-7	Dibromofluoromethane	51.2			75 - 124	102%	SPK: 50		
2037-26-5	Toluene-d8	50.0			86 - 113	100%	SPK: 50		
460-00-4	4-Bromofluorobenzene	59.1			77 - 121	118%	SPK: 50		
INTERNAL STANDARDS									
		Area Count							
363-72-4	Pentafluorobenzene	129000							
540-36-3	1,4-Difluorobenzene	259000							
3114-55-4	Chlorobenzene-d5	275000							
3855-82-1	1,4-Dichlorobenzene-d4	139000							

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products