

NEW JERSEY LAB ID#:20012 : NEW YORK LAB ID#: 11376

GC/MS VOA CONFORMANCE/NON-CONFORMANCE SUMMARY

CHEMTECH PROJECT NUMBER: VX101725

SequenceID : VX101725 NA NO YES

1. Chromatograms Labeled/Compounds Identified. (Field samples and Method Blanks) ✓

2. GC/MS Tuning Specifications. BFB Meet Criteria ✓
(NOTE THAT THERE ARE DIFFERENT CRITERIA FOR NY ASP CLP, CLP AND NJ)

3. GC/MS Tuning Frequency - Performed every 24 hours for 600 series and 12 hours for 8000 series ✓

4. GC/MS Calibration - Initial Calibration performed before sample analysis and continuing calibration performed within 24 hours of sample analysis for 600 series and 12 hours for 8000 series ✓

5. GC/MS Calibration Requirements ✓

5a. Initial Calibration Meet Criteria ✓
If not met, list those compounds and their recoveries which fall outside the acceptable ranges.
The %RSD is greater than 20% in the Initial Calibration method (82X101725W.M) for 2-Chloroethyl Vinyl ether passing on Linear Regression.

5b. Continuous Calibration(CCC) Meet Criteria ✓
If not met, list those compounds and their recoveries which fall outside the acceptable ranges.

6. Blank Contamination - If yes, list compounds and concentrations in each blank: ✓

7. Surrogate Recoveries Meet Criteria ✓
If not met, list those compounds and their recoveries which fall outside the acceptable ranges.

8a. Matrix Spike/Matrix Spike Duplicate Recoveries Meet Criteria ✓
If not met, list those compounds and their recoveries which fall outside the acceptable range.

8b. Blank Spike/Blank spike Duplicate Recoveries Meet Criteria ✓
If not met, list those compounds and their recoveries which fall outside the acceptable range.

9. Internal Standard Area/Retention Time Shift Meet Criteria ✓

Comments:

nishadh

Analyst

10/18/2025

Date



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

NEW JERSEY LAB ID#:20012 : NEW YORK LAB ID#: 11376

GC/MS VOA CONFORMANCE/NON-CONFORMANCE SUMMARY

CHEMTECH PROJECT NUMBER: VX101725

SequenceID : VX101725

NA NO YES

10. Analysis Holding Time Met

✓

If not met, list number of days exceeded for each sample:

ADDITIONAL COMMENTS:

nishadh

Analyst

10/18/2025

Date



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

NEW JERSEY LAB ID#:20012 : NEW YORK LAB ID#: 11376

GC/MS VOA CONFORMANCE/NON-CONFORMANCE SUMMARY

CHEMTECH PROJECT NUMBER: VX101725

SequenceID : VX101725 NA NO YES

nishadh
Analyst

10/18/2025
Date