

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance Contract: CORE02
 Lab Code: ACE SDG NO.: Q3311
 Lab File ID: VX048100.D Date Analyzed: 10/10/2025
 Instrument ID: MSVOA_X Time Analyzed: 10:45
 GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: (Y/N) N

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	124915	5.53	216127	6.74	194988	10.04
UPPER LIMIT	249830	6.025	432254	7.238	389976	10.537
LOWER LIMIT	62457.5	5.025	108064	6.238	97494	9.537
EPA SAMPLE NO.						
MW-218DL	107553	5.54	220883	6.75	232642	10.04
MW-1SDL	127982	5.54	267787	6.75	282940	10.04
MW-1DDL	149568	5.54	313800	6.75	328751	10.04
MW-101SDL	127815	5.54	272296	6.75	294793	10.04
C-MW-3DL	130759	5.54	269905	6.75	284581	10.04
MW-6BA	106376	5.54	218035	6.75	223748	10.04
VX1010WBL01	163530	5.54	330351	6.75	333490	10.04
VX1010WBS01	134840	5.53	228857	6.74	211483	10.04

IS1 = Pentafluorobenzene
 IS2 = 1,4-Difluorobenzene
 IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.