

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance Contract: CORE02
 Lab Code: ACE SDG NO.: Q3311
 Lab File ID: VX048201.D Date Analyzed: 10/17/2025
 Instrument ID: MSVOA_X Time Analyzed: 09:55
 GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: (Y/N) N

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	144905	5.54	233738	6.75	208719	10.04
UPPER LIMIT	289810	6.044	467476	7.245	417438	10.537
LOWER LIMIT	72452.5	5.044	116869	6.245	104360	9.537
EPA SAMPLE NO.						
MW-214DDL	83059	5.54	168187	6.75	177290	10.04
VX1017WBL01	129436	5.54	259480	6.75	275267	10.04
VX1017WBS01	157010	5.54	248908	6.75	238012	10.04

IS1 = Pentafluorobenzene
 IS2 = 1,4-Difluorobenzene
 IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.