

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance Contract: CORE02
 Lab Code: ACE SDG NO.: Q3311
 Lab File ID: VX048072.D Date Analyzed: 10/09/2025
 Instrument ID: MSVOA_X Time Analyzed: 09:06
 GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: (Y/N) N

	IS4 AREA #	RT #				
12 HOUR STD	103010	12				
UPPER LIMIT	206020	12.5				
LOWER LIMIT	51505	11.5				
EPA SAMPLE NO.						
MW-218	109726	12.01				
MW-1S	132444	12.01				
MW-1D	142615	12.01				
MW-2S	152199	12.01				
MW-101S	132248	12.01				
C-MW-3	104628	12.01				
VX1009WBL01	160166	12.01				
VX1009WBS02	106240	12.01				

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.