

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance Contract: CORE02
 Lab Code: ACE SDG NO.: Q3311
 Lab File ID: VX048100.D Date Analyzed: 10/10/2025
 Instrument ID: MSVOA_X Time Analyzed: 10:45
 GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: (Y/N) N

	IS4 AREA #	RT #				
12 HOUR STD	96783	12.00€				
UPPER LIMIT	193566	12.50€				
LOWER LIMIT	48391.5	11.50€				
EPA SAMPLE NO.						
MW-218DL	113160	12.01				
MW-1SDL	143470	12.01				
MW-1DDL	162002	12.01				
MW-101SDL	148707	12.01				
C-MW-3DL	145743	12.01				
MW-6BA	109569	12.01				
VX1010WBL01	161186	12.01				
VX1010WBS01	106925	12.01				

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.