

ALLIANCE 284 Sheffield Street, Mountainside New Jersey 07092

NEW JERSEY LAB ID#: 20012; NEW YORK LAB ID#: 11376

GC/MS VOA CONFORMANCE/NON-CONFORMANCE SUMMARY

ORDER ID: Q3312

MATRIX: Water

METHOD: 8260D

| | | NA | NO | YES |
|-----|---|----|----|-----|
| 1. | Chromatograms Labeled/Compounds Identified. (Field samples and Method Blanks) | | | ✓ |
| 2. | GC/MS Tuning Specifications BFB Meet Criteria (NOTE THAT THERE ARE DIFFERENT CRITERIA FOR NY ASP CLP, CLP AND NJ) | | | ✓ |
| 3. | GC/MS Tuning Frequency - Performed every 24 hours for 600 series and 12 hours for 8000 Series. | | | ✓ |
| 4. | GC/MS Calibration - Initial Calibration performed before sample analysis and continuing calibration performed within 24 hours of sample analysis for 600 series and 12 hours for 8000 series. | | | ✓ |
| 5. | GC/MS Calibration Requirements. The Initial Calibration met the requirements. The Continuous Calibration met the requirements. | | | ✓ |
| 6. | Blank Contamination - If yes, list compounds and concentrations in each blank: | | ✓ | |
| 7. | Surrogate Recoveries Meet Criteria If not met, list those compounds and their recoveries which fall outside the acceptable ranges. | | | ✓ |
| 8. | Matrix Spike/Matrix Spike Duplicate Recoveries Meet Criteria If not met, list those compounds and their recoveries which fall outside the acceptable range. | | | ✓ |
| 9. | Internal Standard Area/Retention Time Shift Meet Criteria Comments: | | | ✓ |
| 10. | Analysis Holding Time Met If not met, list number of days exceeded for each sample: | | | ✓ |

ADDITIONAL COMMENTS:

Samples for MS/MSD for VOC analysis were not provided with this set of samples. The Blank Spike Duplicate is reported with the data.

ALLIANCE 284 Sheffield Street, Mountainside New Jersey 07092

NEW JERSEY LAB ID#: 20012; NEW YORK LAB ID#: 11376

GC/MS VOA CONFORMANCE/NON-CONFORMANCE SUMMARY (CONTINUED)

NA NO YES

Trip Blank was not provided with this set of samples.

QA REVIEW

Date