

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name:	<u>Alliance</u>	Contract:	<u>CORE02</u>
Lab Code:	<u>ACE</u>	SDG No.:	<u>Q3312</u>
Instrument ID:	<u>MSVOA_X</u>	Calibration Date/Time:	<u>10/14/2025 10:02</u>
Lab File ID:	<u>VX048153.D</u>	Init. Calib. Date(s):	<u>09/16/2025 09/16/2025</u>
Heated Purge: (Y/N)	<u>N</u>	Init. Calib. Time(s):	<u>09:23 12:01</u>
GC Column:	<u>DB-624UI</u>	ID:	<u>0.18</u> (mm)

COMPOUND	RRF	RRF050	MIN RRF	%D	MAX%D
Benzene	1.488	1.679		12.84	20
Toluene	0.917	1.028		12.1	20
Ethyl Benzene	1.960	2.162		10.31	20
m/p-Xylenes	0.729	0.814		11.66	20
o-Xylene	0.706	0.754		6.8	20
1,2-Dichloroethane-d4	0.796	0.845		6.16	20
Dibromofluoromethane	0.335	0.391		16.72	20
Toluene-d8	1.160	1.287		10.95	20
4-Bromofluorobenzene	0.440	0.491		11.59	20

All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.