

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.:	<u>Q3312</u>	Analytical Method:	<u>SW8260-Low</u>
Client:	<u>Core Environmental Consultants and</u>	Datafile :	<u>VX048156.D</u>

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Low	Limits High	RPD
VX1014WBS01	Benzene	20	19.6	ug/L	98			82	109	
	Toluene	20	20.0	ug/L	100			82	110	
	Ethyl Benzene	20	19.2	ug/L	96			83	109	
	m/p-Xylenes	40	39.5	ug/L	99			82	110	
	o-Xylene	20	19.2	ug/L	96			83	109	