

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.:	<u>Q3312</u>	Analytical Method:	<u>SW8260-Low</u>
Client:	<u>Core Environmental Consultants and</u>	Datafile :	<u>VX048157.D</u>

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Low	Limits High	RPD
VX1014WBSD01	Benzene	20	20.3	ug/L	102	4		82	109	15
	Toluene	20	20.8	ug/L	104	4		82	110	16
	Ethyl Benzene	20	19.2	ug/L	96	0		83	109	16
	m/p-Xylenes	40	39.9	ug/L	100	1		82	110	15
	o-Xylene	20	19.3	ug/L	97	1		83	109	20