

DATA PACKAGE

VOLATILE ORGANICS

PROJECT NAME : 200 RESERVOIR ST, TRENTON NJ

GFE LLC

58 Nokomis Ave

Lake Hiawatha, NJ - 07034

Phone No: 646-542-3465

ORDER ID : Q3320

ATTENTION : Frank Galdun



Laboratory Certification ID # 20012



1) Signature Page	3
2) Case Narrative	4
2.1) TO-15- Case Narrative	4
3) Qualifier Page	6
4) QA Checklist	7
5) TO-15 Data	8
6) Shipping Document	71
6.1) CHAIN OF CUSTODY	72
6.2) Lab Certificate	85
6.3) Internal COC	86

1
2
3
4
5
6

Cover Page

Order ID : Q3320

Project ID : 200 Reservoir St, Trenton NJ

Client : GFE LLC

Lab Sample Number

Q3320-01
Q3320-02
Q3320-03
Q3320-04
Q3320-05
Q3320-06
Q3320-07
Q3320-08
Q3320-09
Q3320-10
Q3320-11
Q3320-12
Q3320-13

Client Sample Number

IA1
SV1
IA2
SV2
IA3
SV3
IA4
SV4
IA5
SV5
IA6
SV6
OA1

I certify that the data package is in compliance with the terms and conditions of the contract, both technically and for completeness, for other than the conditions detailed above. Release of the data contained in this hard copy data package has been authorized by the laboratory manager or his designee, as verified by the following signature.

Signature : _____

Date: 10/13/2025

NYDOH CERTIFICATION NO - 11376

NJDEP CERTIFICATION NO - 20012

CASE NARRATIVE

GFE LLC

Project Name: 200 Reservoir St, Trenton NJ

Project # N/A

Order ID # Q3320

Test Name: TO-15

A. Number of Samples and Date of Receipt:

13 Air samples were received on 10/08/2025.

B. Parameters

According to the Chain of Custody document, the following analyses were requested: TO-15. This data package contains results for TO-15.

C. Analytical Techniques:

The analysis performed on instrument MSVOA_L were done using GC column RTX-1, which is 60 meters, 0.32 mm id, 1.0 um df, Restek Cat. #10157. The Trap was supplied by Entech, glass bead and Tenax , Entech 7100A Preconcentrator. The analysis of TO-15 was based on method TO-15.

D. QA/ QC Samples:

The Holding Times were met for all analysis.

The Surrogate recoveries were met for all analysis.

The Internal Standards Areas were met for all analysis.

The Retention Times were met for all analysis.

The RPD for {Q3320-01DUP} with File ID: VL043084.D met criteria except for tert-Butyl Alcohol[200%], Toluene[200%] due to difference in results of original and DUP.

The Blank Spike met requirements for all compounds.

The Blank analysis did not indicate the presence of lab contamination.

The %RSD is greater than 30% in the Initial Calibration method (VL100225AIR.M) for Naphthalene passing on Quadratic Regression.

The Continuous Calibration File ID VL043054.D met the requirements except for 1,2,4-Trichlorobenzene is failing high but no positive hit in associate sample therefore no corrective action taken.

The Tuning criteria met requirements.

Due to potential high concentration of target analytes, Samples SV1, SV2, SV3, SV4, SV5 and SV6 were initially diluted.



284 Sheffield Street, Mountainside, NJ 07092
Phone: 908 789 8900 Fax: 908 789 8922

2
2.1

Samples SV1, SV2, SV4 and SV6 were diluted due to high concentrations.

E. Additional Comments:

Alliance Technical Group, LLC - Newark holds certification for the analytes "Bromodichloromethane, Chloroethane and Heptane". Currently, we are not eligible to report them, as we have not successfully completed the required Proficiency Testing (PT) study. We are actively engaged in the PT study process to gain reporting eligibility.

F. Manual Integration Comments:

Please refer to the Manual integration Report included with the Run Logs for information on the manual integrations performed.

I certify that the data package is in compliance with the terms and conditions of the contract, both technically and for completeness, for other than the conditions detailed above. The laboratory manager or his designee, as verified by the following signature has authorized release of the data contained in this hard copy data package.

Signature_____

DATA REPORTING QUALIFIERS- ORGANIC

For reporting results, the following “Results Qualifiers” are used:

Value	If the result is a value greater than or equal to the detection limit, report the value
U	Indicates the compound was analyzed for but was not detected. Report the minimum detection limit for the sample with the U, i.e. “10 U”. This is not necessarily the instrument detection limit attainable for this particular sample based on any concentration or dilution that may have been required.
ND	Indicates the analyte was analyzed for, but not detected
J	Indicates an estimated value. This flag is used: (1) When estimating a concentration for a tentatively identified compound (library search hits, where a 1:1 response is assumed.) (2) When the mass spectral data indicated the identification, however the result was less than the specified detection limit greater than zero. If the detection limit was 10ug/L and a concentration of 3 ug/L was calculated report as 3 J. This flag is used when similar situation arise on any organic parameter i.e. Pest, PCB and others.
B	Indicates the analyte was found in the blank as well as the sample report as “12 B”.
E	Indicates the analyte ‘s concentration exceeds the calibrated range of the instrument for that specific analysis.
D	This flag identifies all compounds identified in an analysis at a secondary dilution factor.
P	This flag is used for Pesticide/PCB target analyte when there is >25% difference for detected concentrations between the two GC columns. The lower of the two values is reported on Form 1 and flagged with a “P”.
N	This flag indicates presumptive evidence of a compound. This is only used for tentatively identified compounds (TICs), where the identification is based on a mass spectral library search. It applies to all TIC results. For generic characterization of a TIC, such as chlorinated hydrocarbon, the flag is not used.
A	This flag indicates that a Tentatively Identified Compound is a suspected aldol-condensation product.
Q	Indicates the LCS did not meet the control limits requirements

APPENDIX A

QA REVIEW GENERAL DOCUMENTATION

Project #: Q3320

Completed

For thorough review, the report must have the following:

GENERAL:

Are all original paperwork present (chain of custody, record of communication,airbill, sample management lab chronicle, login page)

✓

Check chain-of-custody for proper relinquish/return of samples

✓

Is the chain of custody signed and complete

✓

Check internal chain-of-custody for proper relinquish/return of samples /sample extracts

✓

Collect information for each project id from server. Were all requirements followed

✓

COVER PAGE:

Do numbers of samples correspond to the number of samples in the Chain of Custody on login page

✓

Do lab numbers and client Ids on cover page agree with the Chain of Custody

✓

CHAIN OF CUSTODY:

Do requested analyses on Chain of Custody agree with form I results

✓

Do requested analyses on Chain of Custody agree with the log-in page

✓

Were the correct method log-in for analysis according to the Analytical Request and Chain of Custody

✓

Were the samples received within hold time

✓

Were any problems found with the samples at arrival recorded in the Sample Management Laboratory Chronicle

✓

ANALYTICAL:

Was method requirement followed?

✓

Was client requirement followed?

✓

Does the case narrative summarize all QC failure?

✓

All runlogs and manual integration are reviewed for requirements

✓

All manual calculations and /or hand notations verified

✓

QA Review Signature: SOHIL JODHANI

Date: 10/13/2025

Hit Summary Sheet
SW-846

SDG No.: Q3320
Client: GFE LLC

Sample ID	Client ID	Matrix	Parameter	Concentration	C	MDL	RDL	Units
Client ID:	IA1							
Q3320-01	IA1	Air	Dichlorodifluoromethane	2.23	J	0.54	2.47	ug/m3
Q3320-01	IA1	Air	Chloromethane	1.36		0.21	1.03	ug/m3
Q3320-01	IA1	Air	Trichlorofluoromethane	1.29	J	0.96	2.81	ug/m3
Q3320-01	IA1	Air	tert-Butyl alcohol	0.55	J	0.49	1.52	ug/m3
Q3320-01	IA1	Air	Acetone	16.6		0.43	1.19	ug/m3
Q3320-01	IA1	Air	Methylene Chloride	3.02		0.80	1.74	ug/m3
Q3320-01	IA1	Air	2-Butanone	1.92		0.18	1.47	ug/m3
Q3320-01	IA1	Air	Carbon Tetrachloride	0.57		0.19	0.19	ug/m3
Q3320-01	IA1	Air	Chloroform	0.98	J	0.49	2.44	ug/m3
Q3320-01	IA1	Air	Benzene	0.35	J	0.26	1.60	ug/m3
Q3320-01	IA1	Air	Trichloroethene	0.38		0.11	0.16	ug/m3
Q3320-01	IA1	Air	4-Methyl-2-Pentanone	0.90	J	0.41	2.05	ug/m3
Q3320-01	IA1	Air	Toluene	0.60	J	0.60	1.88	ug/m3
Q3320-01	IA1	Air	Tetrachloroethene	0.41		0.14	0.20	ug/m3
Q3320-01	IA1	Air	1,1,2,2-Tetrachloroethane	0.34		0.14	0.21	ug/m3
Q3320-01	IA1	Air	Hexane	1.80		0.56	1.76	ug/m3
			Total Voc :		33.3			
			Total Concentration:		33.3			
Client ID:	SV1							
Q3320-02	SV1	Air	Tetrahydrofuran	6.49	J	2.36	14.8	ug/m3
Q3320-02	SV1	Air	Trichlorofluoromethane	364		9.55	28.1	ug/m3
Q3320-02	SV1	Air	Heptane	16.8	J	6.97	20.5	ug/m3
Q3320-02	SV1	Air	Acetone	144		4.28	11.9	ug/m3
Q3320-02	SV1	Air	Carbon Disulfide	23.0		2.49	15.6	ug/m3
Q3320-02	SV1	Air	Methylene Chloride	12.5	J	7.99	17.4	ug/m3
Q3320-02	SV1	Air	2-Butanone	14.8		1.65	14.8	ug/m3
Q3320-02	SV1	Air	Chloroform	26.9		4.69	24.4	ug/m3
Q3320-02	SV1	Air	1,1,1-Trichloroethane	25.6		0.87	1.64	ug/m3
Q3320-02	SV1	Air	Benzene	4.15	J	2.52	16.0	ug/m3
Q3320-02	SV1	Air	Trichloroethene	7520	E	1.29	1.61	ug/m3
Q3320-02	SV1	Air	Toluene	27.1		6.03	18.8	ug/m3
Q3320-02	SV1	Air	Tetrachloroethene	5700	E	1.02	2.03	ug/m3
Q3320-02	SV1	Air	o-Xylene	10.9	J	9.12	21.7	ug/m3
Q3320-02	SV1	Air	Hexane	99.4		5.64	17.6	ug/m3
			Total Voc :		14000			
			Total Concentration:		14000			

Hit Summary Sheet SW-846

SDG No.: Q3320

Client: GFE LLC

Sample ID	Client ID	Matrix	Parameter	Concentration	C	MDL	RDL	Units
Client ID:	SV1DL							
Q3320-02DL	SV1DL	Air	Trichlorofluoromethane	427	JD	191	562	ug/m3
Q3320-02DL	SV1DL	Air	Trichloroethene	8600	D	25.8	32.3	ug/m3
Q3320-02DL	SV1DL	Air	Tetrachloroethene	5220	D	20.3	40.7	ug/m3
			Total Voc :	14200				
			Total Concentration:	14200				
Client ID:	IA2							
Q3320-03	IA2	Air	Dichlorodifluoromethane	2.08	J	0.54	2.47	ug/m3
Q3320-03	IA2	Air	Chloromethane	0.89	J	0.21	1.03	ug/m3
Q3320-03	IA2	Air	Trichlorofluoromethane	1.18	J	0.96	2.81	ug/m3
Q3320-03	IA2	Air	Acetone	7.36		0.43	1.19	ug/m3
Q3320-03	IA2	Air	Methylene Chloride	2.54		0.80	1.74	ug/m3
Q3320-03	IA2	Air	2-Butanone	0.65	J	0.18	1.47	ug/m3
Q3320-03	IA2	Air	Carbon Tetrachloride	0.44		0.19	0.19	ug/m3
Q3320-03	IA2	Air	Chloroform	0.68	J	0.49	2.44	ug/m3
Q3320-03	IA2	Air	Trichloroethene	0.27		0.11	0.16	ug/m3
Q3320-03	IA2	Air	Naphthalene	1.31		0.050	0.52	ug/m3
Q3320-03	IA2	Air	Hexane	1.16	J	0.56	1.76	ug/m3
			Total Voc :	18.6				
			Total Concentration:	18.6				
Client ID:	SV2							
Q3320-04	SV2	Air	Trichlorofluoromethane	66.3		9.55	28.1	ug/m3
Q3320-04	SV2	Air	Acetone	168		4.28	11.9	ug/m3
Q3320-04	SV2	Air	Carbon Disulfide	4.98	J	2.49	15.6	ug/m3
Q3320-04	SV2	Air	Methylene Chloride	10.8	J	7.99	17.4	ug/m3
Q3320-04	SV2	Air	2-Butanone	7.96	J	1.65	14.8	ug/m3
Q3320-04	SV2	Air	1,1,1-Trichloroethane	8.18		0.87	1.64	ug/m3
Q3320-04	SV2	Air	Trichloroethene	155		1.29	1.61	ug/m3
Q3320-04	SV2	Air	Toluene	12.8	J	6.03	18.8	ug/m3
Q3320-04	SV2	Air	Tetrachloroethene	2170	E	1.02	2.03	ug/m3
Q3320-04	SV2	Air	Hexane	33.5		5.64	17.6	ug/m3
			Total Voc :	2640				
			Total Concentration:	2640				
Client ID:	SV2DL							
Q3320-04DL	SV2DL	Air	Trichlorofluoromethane	66.3	JD	38.2	112	ug/m3
Q3320-04DL	SV2DL	Air	Acetone	167	D	17.1	47.5	ug/m3
Q3320-04DL	SV2DL	Air	Methylene Chloride	33.7	JD	32.0	69.5	ug/m3
Q3320-04DL	SV2DL	Air	1,1,1-Trichloroethane	9.28	D	3.49	6.55	ug/m3

Hit Summary Sheet SW-846

SDG No.: Q3320

Client: GFE LLC

Sample ID	Client ID	Matrix	Parameter	Concentration	C	MDL	RDL	Units
Q3320-04DL	SV2DL	Air	Trichloroethene	153	D	5.16	6.45	ug/m3
Q3320-04DL	SV2DL	Air	Tetrachloroethene	2030	D	4.07	8.14	ug/m3
Q3320-04DL	SV2DL	Air	Hexane	31.4	JD	22.6	70.5	ug/m3
Total Voc :				2490				
Total Concentration:				2490				
Client ID:	IA3							
Q3320-05	IA3	Air	Dichlorodifluoromethane	2.62		0.54	2.47	ug/m3
Q3320-05	IA3	Air	Chloromethane	1.07		0.21	1.03	ug/m3
Q3320-05	IA3	Air	Trichlorofluoromethane	1.46	J	0.96	2.81	ug/m3
Q3320-05	IA3	Air	Acetone	9.50		0.43	1.19	ug/m3
Q3320-05	IA3	Air	Methylene Chloride	6.60		0.80	1.74	ug/m3
Q3320-05	IA3	Air	2-Butanone	0.59	J	0.18	1.47	ug/m3
Q3320-05	IA3	Air	Carbon Tetrachloride	0.57		0.19	0.19	ug/m3
Q3320-05	IA3	Air	Chloroform	0.49	J	0.49	2.44	ug/m3
Q3320-05	IA3	Air	Trichloroethene	0.21		0.11	0.16	ug/m3
Q3320-05	IA3	Air	Toluene	0.68	J	0.60	1.88	ug/m3
Q3320-05	IA3	Air	Tetrachloroethene	0.27		0.14	0.20	ug/m3
Q3320-05	IA3	Air	Hexane	4.23		0.56	1.76	ug/m3
Total Voc :				28.3				
Total Concentration:				28.3				
Client ID:	SV3							
Q3320-06	SV3	Air	Tetrahydrofuran	2.45	J	1.18	7.37	ug/m3
Q3320-06	SV3	Air	Trichlorofluoromethane	5.34	J	4.78	14.1	ug/m3
Q3320-06	SV3	Air	Heptane	9.02	J	3.48	10.3	ug/m3
Q3320-06	SV3	Air	Acetone	138		2.14	5.94	ug/m3
Q3320-06	SV3	Air	Methylene Chloride	4.52	J	4.17	8.69	ug/m3
Q3320-06	SV3	Air	2-Butanone	3.24	J	0.83	7.37	ug/m3
Q3320-06	SV3	Air	Trichloroethene	2.53		0.64	0.81	ug/m3
Q3320-06	SV3	Air	Toluene	11.3		3.01	9.42	ug/m3
Q3320-06	SV3	Air	Tetrachloroethene	3.05		0.54	1.02	ug/m3
Q3320-06	SV3	Air	o-Xylene	4.78	J	4.78	10.9	ug/m3
Q3320-06	SV3	Air	Hexane	60.6		2.82	8.81	ug/m3
Total Voc :				245				
Total Concentration:				245				
Client ID:	IA4							
Q3320-07	IA4	Air	Dichlorodifluoromethane	2.52		0.54	2.47	ug/m3
Q3320-07	IA4	Air	Chloromethane	1.03		0.21	1.03	ug/m3
Q3320-07	IA4	Air	Trichlorofluoromethane	1.46	J	0.96	2.81	ug/m3

Hit Summary Sheet
SW-846

SDG No.: Q3320

Client: GFE LLC

Sample ID	Client ID	Matrix	Parameter	Concentration	C	MDL	RDL	Units
Q3320-07	IA4	Air	Acetone	9.74		0.43	1.19	ug/m3
Q3320-07	IA4	Air	Methylene Chloride	2.15		0.80	1.74	ug/m3
Q3320-07	IA4	Air	2-Butanone	0.65	J	0.18	1.47	ug/m3
Q3320-07	IA4	Air	Carbon Tetrachloride	0.57		0.19	0.19	ug/m3
Q3320-07	IA4	Air	Chloroform	0.54	J	0.49	2.44	ug/m3
Q3320-07	IA4	Air	Trichloroethene	0.27		0.11	0.16	ug/m3
Q3320-07	IA4	Air	Hexane	1.30	J	0.56	1.76	ug/m3
Total Voc :				20.2				
Total Concentration:				20.2				
Client ID:	SV4							
Q3320-08	SV4	Air	Tetrahydrofuran	3.83	J	1.89	11.8	ug/m3
Q3320-08	SV4	Air	Heptane	11.9	J	5.74	16.4	ug/m3
Q3320-08	SV4	Air	Acetone	499	E	3.33	9.50	ug/m3
Q3320-08	SV4	Air	Carbon Disulfide	3.43	J	1.99	12.5	ug/m3
Q3320-08	SV4	Air	Methylene Chloride	11.1	J	6.25	13.9	ug/m3
Q3320-08	SV4	Air	2-Butanone	12.1		1.33	11.8	ug/m3
Q3320-08	SV4	Air	Trichloroethene	4.94		1.02	1.29	ug/m3
Q3320-08	SV4	Air	Toluene	17.0		4.90	15.1	ug/m3
Q3320-08	SV4	Air	Tetrachloroethene	10.8		0.81	1.63	ug/m3
Q3320-08	SV4	Air	o-Xylene	9.56	J	7.38	17.4	ug/m3
Q3320-08	SV4	Air	Hexane	90.6		4.58	14.1	ug/m3
Total Voc :				674				
Total Concentration:				674				
Client ID:	SV4DL							
Q3320-08DL	SV4DL	Air	Acetone	546	D	17.1	47.5	ug/m3
Q3320-08DL	SV4DL	Air	Methylene Chloride	33.7	JD	32.0	69.5	ug/m3
Q3320-08DL	SV4DL	Air	2-Butanone	13.9	JD	6.49	59.0	ug/m3
Q3320-08DL	SV4DL	Air	Tetrachloroethene	12.2	D	4.07	8.14	ug/m3
Q3320-08DL	SV4DL	Air	Hexane	59.6	JD	22.6	70.5	ug/m3
Total Voc :				666				
Total Concentration:				666				
Client ID:	IA5							
Q3320-09	IA5	Air	Dichlorodifluoromethane	2.47		0.54	2.47	ug/m3
Q3320-09	IA5	Air	Chloromethane	1.01	J	0.21	1.03	ug/m3
Q3320-09	IA5	Air	Trichlorofluoromethane	1.35	J	0.96	2.81	ug/m3
Q3320-09	IA5	Air	Acetone	24.0		0.43	1.19	ug/m3
Q3320-09	IA5	Air	Methylene Chloride	2.33		0.80	1.74	ug/m3
Q3320-09	IA5	Air	2-Butanone	1.06	J	0.18	1.47	ug/m3

Hit Summary Sheet
SW-846

SDG No.: Q3320

Client: GFE LLC

Sample ID	Client ID	Matrix	Parameter	Concentration	C	MDL	RDL	Units
Q3320-09	IA5	Air	Carbon Tetrachloride	0.50		0.19	0.19	ug/m3
Q3320-09	IA5	Air	Chloroform	1.56	J	0.49	2.44	ug/m3
Q3320-09	IA5	Air	Toluene	1.13	J	0.60	1.88	ug/m3
Q3320-09	IA5	Air	Tetrachloroethene	0.34		0.14	0.20	ug/m3
Q3320-09	IA5	Air	Hexane	1.52	J	0.56	1.76	ug/m3
Total Voc :				37.3				
Total Concentration:				37.3				
Client ID:	SV5							
Q3320-10	SV5	Air	tert-Butyl alcohol	5.15	J	3.94	12.1	ug/m3
Q3320-10	SV5	Air	Acetone	174		3.33	9.50	ug/m3
Q3320-10	SV5	Air	Methylene Chloride	7.30	J	6.25	13.9	ug/m3
Q3320-10	SV5	Air	2-Butanone	2.51	J	1.33	11.8	ug/m3
Q3320-10	SV5	Air	Trichloroethene	5.37		1.02	1.29	ug/m3
Q3320-10	SV5	Air	Toluene	8.67	J	4.90	15.1	ug/m3
Q3320-10	SV5	Air	Tetrachloroethene	42.0		0.81	1.63	ug/m3
Q3320-10	SV5	Air	Hexane	28.2		4.58	14.1	ug/m3
Total Voc :				273				
Total Concentration:				273				
Client ID:	IA6							
Q3320-11	IA6	Air	Dichlorodifluoromethane	2.08	J	0.54	2.47	ug/m3
Q3320-11	IA6	Air	Chloromethane	0.95	J	0.21	1.03	ug/m3
Q3320-11	IA6	Air	Trichlorofluoromethane	1.12	J	0.96	2.81	ug/m3
Q3320-11	IA6	Air	Acetone	20.7		0.43	1.19	ug/m3
Q3320-11	IA6	Air	Methylene Chloride	1.84		0.80	1.74	ug/m3
Q3320-11	IA6	Air	2-Butanone	0.53	J	0.18	1.47	ug/m3
Q3320-11	IA6	Air	Carbon Tetrachloride	0.44		0.19	0.19	ug/m3
Q3320-11	IA6	Air	Chloroform	1.07	J	0.49	2.44	ug/m3
Q3320-11	IA6	Air	Toluene	1.09	J	0.60	1.88	ug/m3
Q3320-11	IA6	Air	Naphthalene	1.31		0.050	0.52	ug/m3
Q3320-11	IA6	Air	Hexane	0.99	J	0.56	1.76	ug/m3
Total Voc :				32.1				
Total Concentration:				32.1				
Client ID:	SV6							
Q3320-12	SV6	Air	tert-Butyl alcohol	9.70		2.43	7.58	ug/m3
Q3320-12	SV6	Air	Acetone	194	E	2.14	5.94	ug/m3
Q3320-12	SV6	Air	Methylene Chloride	4.17	J	4.17	8.69	ug/m3
Q3320-12	SV6	Air	2-Butanone	2.77	J	0.83	7.37	ug/m3
Q3320-12	SV6	Air	Benzene	1.69	J	1.28	7.99	ug/m3

Hit Summary Sheet
SW-846

SDG No.: Q3320

Client: GFE LLC

Sample ID	Client ID	Matrix	Parameter	Concentration	C	MDL	RDL	Units
Q3320-12	SV6	Air	Trichloroethene	9.14		0.64	0.81	ug/m3
Q3320-12	SV6	Air	Toluene	6.41	J	3.01	9.42	ug/m3
Q3320-12	SV6	Air	Tetrachloroethene	57.0		0.54	1.02	ug/m3
Q3320-12	SV6	Air	Ethyl Benzene	43.9		4.13	10.9	ug/m3
Q3320-12	SV6	Air	m/p-Xylene	87.7		9.12	21.7	ug/m3
Q3320-12	SV6	Air	o-Xylene	16.5		4.78	10.9	ug/m3
Q3320-12	SV6	Air	Hexane	20.1		2.82	8.81	ug/m3
Total Voc :				453				
Total Concentration:				453				
Client ID:	SV6DL							
Q3320-12DL	SV6DL	Air	tert-Butyl alcohol	9.70	JD	4.85	15.2	ug/m3
Q3320-12DL	SV6DL	Air	Acetone	187	D	4.28	11.9	ug/m3
Q3320-12DL	SV6DL	Air	Trichloroethene	9.67	D	1.29	1.61	ug/m3
Q3320-12DL	SV6DL	Air	Toluene	6.03	JD	6.03	18.8	ug/m3
Q3320-12DL	SV6DL	Air	Tetrachloroethene	52.9	D	1.02	2.03	ug/m3
Q3320-12DL	SV6DL	Air	Ethyl Benzene	33.9	D	8.25	21.7	ug/m3
Q3320-12DL	SV6DL	Air	m/p-Xylene	67.8	D	17.8	43.4	ug/m3
Q3320-12DL	SV6DL	Air	o-Xylene	13.9	JD	9.12	21.7	ug/m3
Q3320-12DL	SV6DL	Air	Hexane	17.6	D	5.64	17.6	ug/m3
Total Voc :				399				
Total Concentration:				399				
Client ID:	OA1							
Q3320-13	OA1	Air	Dichlorodifluoromethane	2.32	J	0.54	2.47	ug/m3
Q3320-13	OA1	Air	Chloromethane	0.91	J	0.21	1.03	ug/m3
Q3320-13	OA1	Air	Trichlorofluoromethane	1.29	J	0.96	2.81	ug/m3
Q3320-13	OA1	Air	Acetone	8.08		0.43	1.19	ug/m3
Q3320-13	OA1	Air	Methylene Chloride	2.43		0.80	1.74	ug/m3
Q3320-13	OA1	Air	2-Butanone	0.41	J	0.18	1.47	ug/m3
Q3320-13	OA1	Air	Carbon Tetrachloride	0.50		0.19	0.19	ug/m3
Q3320-13	OA1	Air	Hexane	1.20	J	0.56	1.76	ug/m3
Total Voc :				17.1				
Total Concentration:				17.1				



SAMPLE DATA

Report of Analysis

Client:	GFE LLC	Date Collected:	10/08/25
Project:	200 Reservoir St, Trenton NJ	Date Received:	10/08/25
Client Sample ID:	IA1	SDG No.:	Q3320
Lab Sample ID:	Q3320-01	Matrix:	Air
Analytical Method:	TO-15	Test:	TO-15
Sample Wt/Vol:	400	Units:	mL

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VL043083.D	1		10/09/25 12:09	VL100925

CAS Number	Parameter	Conc. ppbv	Conc. ug/M3	Qualifier	MDL ug/m3	LOQ / CRQL ug/m3
TARGETS						
75-71-8	Dichlorodifluoromethane	0.45	2.23	J	0.54	2.47
74-87-3	Chloromethane	0.66	1.36		0.21	1.03
75-01-4	Vinyl Chloride	0.030	0.080	U	0.080	0.080
74-83-9	Bromomethane	0.080	0.31	U	0.31	1.94
75-00-3	Chloroethane	0.15	0.40	U	0.40	1.32
109-99-9	Tetrahydrofuran	0.080	0.24	U	0.24	1.47
75-69-4	Trichlorofluoromethane	0.23	1.29	J	0.96	2.81
76-13-1	1,1,2-Trichlorotrifluoroethane	0.14	1.07	U	1.07	3.83
76-14-2	Dichlorotetrafluoroethane	0.15	1.05	U	1.05	3.49
75-65-0	tert-Butyl alcohol	0.18	0.55	J	0.49	1.52
142-82-5	Heptane	0.17	0.70	U	0.70	2.05
75-35-4	1,1-Dichloroethene	0.15	0.59	U	0.59	1.98
67-64-1	Acetone	7.00	16.6		0.43	1.19
75-15-0	Carbon Disulfide	0.080	0.25	U	0.25	1.56
1634-04-4	Methyl tert-Butyl Ether	0.23	0.83	U	0.83	1.80
75-09-2	Methylene Chloride	0.87	3.02		0.80	1.74
156-60-5	trans-1,2-Dichloroethene	0.12	0.48	U	0.48	1.98
75-34-3	1,1-Dichloroethane	0.13	0.53	U	0.53	2.02
110-82-7	Cyclohexane	0.22	0.76	U	0.76	1.72
78-93-3	2-Butanone	0.65	1.92		0.18	1.47
56-23-5	Carbon Tetrachloride	0.090	0.57		0.19	0.19
156-59-2	cis-1,2-Dichloroethene	0.10	0.40	U	0.40	1.98
67-66-3	Chloroform	0.20	0.98	J	0.49	2.44
71-55-6	1,1,1-Trichloroethane	0.020	0.11	U	0.11	0.16
540-84-1	2,2,4-Trimethylpentane	0.14	0.65	U	0.65	2.34
71-43-2	Benzene	0.11	0.35	J	0.26	1.60
107-06-2	1,2-Dichloroethane	0.090	0.36	U	0.36	2.02
79-01-6	Trichloroethene	0.070	0.38		0.11	0.16
78-87-5	1,2-Dichloropropane	0.13	0.60	U	0.60	2.31
75-27-4	Bromodichloromethane	0.060	0.40	U	0.40	3.35
108-10-1	4-Methyl-2-Pentanone	0.22	0.90	J	0.41	2.05
108-88-3	Toluene	0.16	0.60	J	0.60	1.88
10061-02-6	t-1,3-Dichloropropene	0.15	0.68	U	0.68	2.27
10061-01-5	cis-1,3-Dichloropropene	0.11	0.50	U	0.50	2.27
79-00-5	1,1,2-Trichloroethane	0.080	0.44	U	0.44	2.73

Report of Analysis

Client:	GFE LLC	Date Collected:	10/08/25
Project:	200 Reservoir St, Trenton NJ	Date Received:	10/08/25
Client Sample ID:	IA1	SDG No.:	Q3320
Lab Sample ID:	Q3320-01	Matrix:	Air
Analytical Method:	TO-15	Test:	TO-15
Sample Wt/Vol:	400	Units:	mL

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VL043083.D	1		10/09/25 12:09	VL100925

CAS Number	Parameter	Conc. ppbv	Conc. ug/M3	Qualifier	MDL ug/m3	LOQ / CRQL ug/m3
124-48-1	Dibromochloromethane	0.090	0.77	U	0.77	4.26
106-93-4	1,2-Dibromoethane	0.090	0.69	U	0.69	0.77
127-18-4	Tetrachloroethene	0.060	0.41		0.14	0.20
108-90-7	Chlorobenzene	0.080	0.37	U	0.37	2.30
100-41-4	Ethyl Benzene	0.19	0.83	U	0.83	2.17
179601-23-1	m/p-Xylene	0.41	1.78	U	1.78	4.34
95-47-6	o-Xylene	0.21	0.91	U	0.91	2.17
100-42-5	Styrene	0.16	0.68	U	0.68	2.13
75-25-2	Bromoform	0.050	0.52	U	0.52	5.17
79-34-5	1,1,2,2-Tetrachloroethane	0.050	0.34		0.14	0.21
95-49-8	2-Chlorotoluene	0.17	0.88	U	0.88	2.59
108-67-8	1,3,5-Trimethylbenzene	0.18	0.88	U	0.88	2.46
95-63-6	1,2,4-Trimethylbenzene	0.18	0.88	U	0.88	2.46
541-73-1	1,3-Dichlorobenzene	0.050	0.30	U	0.30	3.01
106-46-7	1,4-Dichlorobenzene	0.13	0.78	U	0.78	3.01
95-50-1	1,2-Dichlorobenzene	0.13	0.78	U	0.78	3.01
120-82-1	1,2,4-Trichlorobenzene	0.21	1.56	U	1.56	3.71
87-68-3	Hexachloro-1,3-Butadiene	0.13	1.39	U	1.39	5.33
106-99-0	1,3-Butadiene	0.050	0.11	U	0.11	1.11
91-20-3	Naphthalene	0.010	0.050	U	0.050	0.52
622-96-8	4-Ethyltoluene	0.21	1.03	U	1.03	2.46
110-54-3	Hexane	0.51	1.80		0.56	1.76
107-05-1	Allyl Chloride	0.11	0.34	U	0.34	1.57
123-91-1	1,4-Dioxane	0.22	0.79	U	0.79	1.80
80-62-6	Methyl Methacrylate	0.14	0.57	U	0.57	2.05
SURROGATES						
460-00-4	1-Bromo-4-Fluorobenzene	10.7			65 - 135	107% SPK: 10
INTERNAL STANDARDS						
74-97-5	Bromochloromethane	182000		2.774		
540-36-3	1,4-Difluorobenzene	494000		3.939		
3114-55-4	Chlorobenzene-d5	418000		8.862		

Report of Analysis

Client:	GFE LLC	Date Collected:	10/08/25
Project:	200 Reservoir St, Trenton NJ	Date Received:	10/08/25
Client Sample ID:	IA1	SDG No.:	Q3320
Lab Sample ID:	Q3320-01	Matrix:	Air
Analytical Method:	TO-15	Test:	TO-15
Sample Wt/Vol:	400 Units: mL		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VL043083.D	1		10/09/25 12:09	VL100925

CAS Number	Parameter	Conc. ppbv	Conc. ug/M3	Qualifier	MDL	LOQ / CRQL
------------	-----------	---------------	----------------	-----------	-----	------------

U = Not Detected

RL = Reporting Limit

MDL = Method Detection Limit

E = Value Exceeds Calibration Range

D = Dilution

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

Q = indicates LCS control criteria did not meet requirements

Report of Analysis

Client:	GFE LLC	Date Collected:	10/08/25
Project:	200 Reservoir St, Trenton NJ	Date Received:	10/08/25
Client Sample ID:	SV1	SDG No.:	Q3320
Lab Sample ID:	Q3320-02	Matrix:	Air
Analytical Method:	TO-15	Test:	TO-15
Sample Wt/Vol:	400	Units:	mL

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VL043089.D	10		10/09/25 15:31	VL100925

CAS Number	Parameter	Conc. ppbv	Conc. ug/M3	Qualifier	MDL ug/m3	LOQ / CRQL ug/m3
TARGETS						
75-71-8	Dichlorodifluoromethane	1.10	5.44	U	5.44	24.7
74-87-3	Chloromethane	0.98	2.02	U	2.02	10.3
75-01-4	Vinyl Chloride	0.25	0.64	U	0.64	0.77
74-83-9	Bromomethane	0.77	2.99	U	2.99	19.4
75-00-3	Chloroethane	1.50	3.96	U	3.96	13.2
109-99-9	Tetrahydrofuran	2.20	6.49	J	2.36	14.8
75-69-4	Trichlorofluoromethane	64.7	364		9.55	28.1
76-13-1	1,1,2-Trichlorotrifluoroethane	1.40	10.7	U	10.7	38.3
76-14-2	Dichlorotetrafluoroethane	1.50	10.5	U	10.5	35.0
75-65-0	tert-Butyl alcohol	1.60	4.85	U	4.85	15.2
142-82-5	Heptane	4.10	16.8	J	6.97	20.5
75-35-4	1,1-Dichloroethene	1.50	5.95	U	5.95	19.8
67-64-1	Acetone	60.5	144		4.28	11.9
75-15-0	Carbon Disulfide	7.40	23.0		2.49	15.6
1634-04-4	Methyl tert-Butyl Ether	2.30	8.29	U	8.29	18.0
75-09-2	Methylene Chloride	3.60	12.5	J	7.99	17.4
156-60-5	trans-1,2-Dichloroethene	1.20	4.76	U	4.76	19.8
75-34-3	1,1-Dichloroethane	1.30	5.26	U	5.26	20.2
110-82-7	Cyclohexane	2.20	7.57	U	7.57	17.2
78-93-3	2-Butanone	5.00	14.8		1.65	14.8
56-23-5	Carbon Tetrachloride	0.25	1.57	U	1.57	1.89
156-59-2	cis-1,2-Dichloroethene	0.99	3.93	U	3.93	19.8
67-66-3	Chloroform	5.50	26.9		4.69	24.4
71-55-6	1,1,1-Trichloroethane	4.70	25.6		0.87	1.64
540-84-1	2,2,4-Trimethylpentane	1.40	6.54	U	6.54	23.4
71-43-2	Benzene	1.30	4.15	J	2.52	16.0
107-06-2	1,2-Dichloroethane	0.93	3.76	U	3.76	20.2
79-01-6	Trichloroethene	1400	7520	E	1.29	1.61
78-87-5	1,2-Dichloropropane	1.30	6.01	U	6.01	23.1
75-27-4	Bromodichloromethane	0.63	4.22	U	4.22	33.5
108-10-1	4-Methyl-2-Pentanone	0.97	3.98	U	3.98	20.5
108-88-3	Toluene	7.20	27.1		6.03	18.8
10061-02-6	t-1,3-Dichloropropene	1.50	6.81	U	6.81	22.7
10061-01-5	cis-1,3-Dichloropropene	1.10	4.99	U	4.99	22.7
79-00-5	1,1,2-Trichloroethane	0.81	4.42	U	4.42	27.3

Report of Analysis

Client:	GFE LLC	Date Collected:	10/08/25
Project:	200 Reservoir St, Trenton NJ	Date Received:	10/08/25
Client Sample ID:	SV1	SDG No.:	Q3320
Lab Sample ID:	Q3320-02	Matrix:	Air
Analytical Method:	TO-15	Test:	TO-15
Sample Wt/Vol:	400 Units: mL		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VL043089.D	10		10/09/25 15:31	VL100925

CAS Number	Parameter	Conc. ppbv	Conc. ug/M3	Qualifier	MDL ug/m3	LOQ / CRQL ug/m3
124-48-1	Dibromochloromethane	0.85	7.24	U	7.24	42.6
106-93-4	1,2-Dibromoethane	0.85	6.53	U	6.53	7.69
127-18-4	Tetrachloroethene	840	5700	E	1.02	2.03
108-90-7	Chlorobenzene	0.77	3.55	U	3.55	23.0
100-41-4	Ethyl Benzene	1.90	8.25	U	8.25	21.7
179601-23-1	m/p-Xylene	4.10	17.8	U	17.8	43.4
95-47-6	o-Xylene	2.50	10.9	J	9.12	21.7
100-42-5	Styrene	1.60	6.81	U	6.81	21.3
75-25-2	Bromoform	0.48	4.96	U	4.96	51.7
79-34-5	1,1,2,2-Tetrachloroethane	0.18	1.24	U	1.24	2.06
95-49-8	2-Chlorotoluene	1.70	8.80	U	8.80	25.9
108-67-8	1,3,5-Trimethylbenzene	1.80	8.85	U	8.85	24.6
95-63-6	1,2,4-Trimethylbenzene	1.80	8.85	U	8.85	24.6
541-73-1	1,3-Dichlorobenzene	0.54	3.25	U	3.25	30.1
106-46-7	1,4-Dichlorobenzene	1.30	7.82	U	7.82	30.1
95-50-1	1,2-Dichlorobenzene	1.30	7.82	U	7.82	30.1
120-82-1	1,2,4-Trichlorobenzene	2.10	15.6	U	15.6	37.1
87-68-3	Hexachloro-1,3-Butadiene	1.30	13.9	U	13.9	53.3
106-99-0	1,3-Butadiene	0.51	1.13	U	1.13	11.1
91-20-3	Naphthalene	0.13	0.68	U	0.68	5.24
622-96-8	4-Ethyltoluene	2.10	10.3	U	10.3	24.6
110-54-3	Hexane	28.2	99.4		5.64	17.6
107-05-1	Allyl Chloride	1.10	3.44	U	3.44	15.7
123-91-1	1,4-Dioxane	2.20	7.93	U	7.93	18.0
80-62-6	Methyl Methacrylate	1.40	5.73	U	5.73	20.5
SURROGATES						
460-00-4	1-Bromo-4-Fluorobenzene	11.1			65 - 135	111% SPK: 10
INTERNAL STANDARDS						
74-97-5	Bromochloromethane	192000		2.774		
540-36-3	1,4-Difluorobenzene	544000		3.936		
3114-55-4	Chlorobenzene-d5	536000		8.862		

Report of Analysis

Client:	GFE LLC	Date Collected:	10/08/25
Project:	200 Reservoir St, Trenton NJ	Date Received:	10/08/25
Client Sample ID:	SV1	SDG No.:	Q3320
Lab Sample ID:	Q3320-02	Matrix:	Air
Analytical Method:	TO-15	Test:	TO-15
Sample Wt/Vol:	400 Units: mL		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VL043089.D	10		10/09/25 15:31	VL100925

CAS Number	Parameter	Conc. ppbv	Conc. ug/M3	Qualifier	MDL	LOQ / CRQL
------------	-----------	---------------	----------------	-----------	-----	------------

U = Not Detected

RL = Reporting Limit

MDL = Method Detection Limit

E = Value Exceeds Calibration Range

D = Dilution

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

Q = indicates LCS control criteria did not meet requirements

Report of Analysis

Client:	GFE LLC	Date Collected:	10/08/25
Project:	200 Reservoir St, Trenton NJ	Date Received:	10/08/25
Client Sample ID:	SV1DL	SDG No.:	Q3320
Lab Sample ID:	Q3320-02DL	Matrix:	Air
Analytical Method:	TO-15	Test:	TO-15
Sample Wt/Vol:	400	Units:	mL

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VL043102.D	200		10/10/25 07:46	VL100925

CAS Number	Parameter	Conc. ppbv	Conc. ug/M3	Qualifier	MDL ug/m3	LOQ / CRQL ug/m3
TARGETS						
75-71-8	Dichlorodifluoromethane	22.0	109	UD	109	494
74-87-3	Chloromethane	19.6	40.5	UD	40.5	207
75-01-4	Vinyl Chloride	5.00	12.8	UD	12.8	15.3
74-83-9	Bromomethane	15.4	59.8	UD	59.8	388
75-00-3	Chloroethane	30.0	79.2	UD	79.2	264
109-99-9	Tetrahydrofuran	16.0	47.2	UD	47.2	295
75-69-4	Trichlorofluoromethane	76.0	427	JD	191	562
76-13-1	1,1,2-Trichlorotrifluoroethane	28.0	215	UD	215	766
76-14-2	Dichlorotetrafluoroethane	30.0	210	UD	210	699
75-65-0	tert-Butyl alcohol	32.0	97.0	UD	97.0	303
142-82-5	Heptane	34.0	139	UD	139	410
75-35-4	1,1-Dichloroethene	30.0	119	UD	119	396
67-64-1	Acetone	36.0	85.5	UD	85.5	238
75-15-0	Carbon Disulfide	16.0	49.8	UD	49.8	311
1634-04-4	Methyl tert-Butyl Ether	46.0	166	UD	166	361
75-09-2	Methylene Chloride	46.0	160	UD	160	347
156-60-5	trans-1,2-Dichloroethene	24.0	95.2	UD	95.2	396
75-34-3	1,1-Dichloroethane	26.0	105	UD	105	405
110-82-7	Cyclohexane	44.0	151	UD	151	344
78-93-3	2-Butanone	11.2	33.0	UD	33.0	295
56-23-5	Carbon Tetrachloride	5.00	31.4	UD	31.4	37.7
156-59-2	cis-1,2-Dichloroethene	19.8	78.5	UD	78.5	396
67-66-3	Chloroform	19.2	93.8	UD	93.8	488
71-55-6	1,1,1-Trichloroethane	3.20	17.5	UD	17.5	32.7
540-84-1	2,2,4-Trimethylpentane	28.0	131	UD	131	467
71-43-2	Benzene	15.8	50.5	UD	50.5	319
107-06-2	1,2-Dichloroethane	18.6	75.3	UD	75.3	405
79-01-6	Trichloroethene	1600	8600	D	25.8	32.3
78-87-5	1,2-Dichloropropane	26.0	120	UD	120	462
75-27-4	Bromodichloromethane	12.6	84.4	UD	84.4	670
108-10-1	4-Methyl-2-Pentanone	19.4	79.5	UD	79.5	410
108-88-3	Toluene	32.0	121	UD	121	377
10061-02-6	t-1,3-Dichloropropene	30.0	136	UD	136	454
10061-01-5	cis-1,3-Dichloropropene	22.0	99.9	UD	99.9	454
79-00-5	1,1,2-Trichloroethane	16.2	88.4	UD	88.4	546

Report of Analysis

Client:	GFE LLC	Date Collected:	10/08/25
Project:	200 Reservoir St, Trenton NJ	Date Received:	10/08/25
Client Sample ID:	SV1DL	SDG No.:	Q3320
Lab Sample ID:	Q3320-02DL	Matrix:	Air
Analytical Method:	TO-15	Test:	TO-15
Sample Wt/Vol:	400 Units: mL		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VL043102.D	200		10/10/25 07:46	VL100925

CAS Number	Parameter	Conc. ppbv	Conc. ug/M3	Qualifier	MDL ug/m3	LOQ / CRQL ug/m3
124-48-1	Dibromochloromethane	17.0	145	UD	145	852
106-93-4	1,2-Dibromoethane	17.0	131	UD	131	154
127-18-4	Tetrachloroethene	770	5220	D	20.3	40.7
108-90-7	Chlorobenzene	15.4	70.9	UD	70.9	461
100-41-4	Ethyl Benzene	38.0	165	UD	165	434
179601-23-1	m/p-Xylene	82.0	356	UD	356	869
95-47-6	o-Xylene	42.0	182	UD	182	434
100-42-5	Styrene	32.0	136	UD	136	426
75-25-2	Bromoform	9.60	99.3	UD	99.3	1030
79-34-5	1,1,2,2-Tetrachloroethane	3.60	24.7	UD	24.7	41.2
95-49-8	2-Chlorotoluene	34.0	176	UD	176	518
108-67-8	1,3,5-Trimethylbenzene	36.0	177	UD	177	492
95-63-6	1,2,4-Trimethylbenzene	36.0	177	UD	177	492
541-73-1	1,3-Dichlorobenzene	10.8	64.9	UD	64.9	601
106-46-7	1,4-Dichlorobenzene	26.0	156	UD	156	601
95-50-1	1,2-Dichlorobenzene	26.0	156	UD	156	601
120-82-1	1,2,4-Trichlorobenzene	42.0	312	UD	312	742
87-68-3	Hexachloro-1,3-Butadiene	26.0	277	UD	277	1070
106-99-0	1,3-Butadiene	10.2	22.6	UD	22.6	221
91-20-3	Naphthalene	2.60	13.6	UD	13.6	105
622-96-8	4-Ethyltoluene	42.0	206	UD	206	492
110-54-3	Hexane	32.0	113	UD	113	352
107-05-1	Allyl Chloride	22.0	68.9	UD	68.9	313
123-91-1	1,4-Dioxane	44.0	159	UD	159	360
80-62-6	Methyl Methacrylate	28.0	115	UD	115	409
SURROGATES						
460-00-4	1-Bromo-4-Fluorobenzene	11.8			65 - 135	118% SPK: 10
INTERNAL STANDARDS						
74-97-5	Bromochloromethane	155000		2.771		
540-36-3	1,4-Difluorobenzene	402000		3.936		
3114-55-4	Chlorobenzene-d5	368000		8.859		

Report of Analysis

Client:	GFE LLC	Date Collected:	10/08/25
Project:	200 Reservoir St, Trenton NJ	Date Received:	10/08/25
Client Sample ID:	SV1DL	SDG No.:	Q3320
Lab Sample ID:	Q3320-02DL	Matrix:	Air
Analytical Method:	TO-15	Test:	TO-15
Sample Wt/Vol:	400 Units: mL		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VL043102.D	200		10/10/25 07:46	VL100925

CAS Number	Parameter	Conc. ppbv	Conc. ug/M3	Qualifier	MDL	LOQ / CRQL
------------	-----------	---------------	----------------	-----------	-----	------------

U = Not Detected
 RL = Reporting Limit
 MDL = Method Detection Limit
 E = Value Exceeds Calibration Range
 D = Dilution

J = Estimated Value
 B = Analyte Found in Associated Method Blank
 N = Presumptive Evidence of a Compound
 * = Values outside of QC limits
 Q = indicates LCS control criteria did not meet requirements

Report of Analysis

Client:	GFE LLC	Date Collected:	10/08/25
Project:	200 Reservoir St, Trenton NJ	Date Received:	10/08/25
Client Sample ID:	IA2	SDG No.:	Q3320
Lab Sample ID:	Q3320-03	Matrix:	Air
Analytical Method:	TO-15	Test:	TO-15
Sample Wt/Vol:	400	Units:	mL

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VL043085.D	1		10/09/25 13:18	VL100925

CAS Number	Parameter	Conc. ppbv	Conc. ug/M3	Qualifier	MDL ug/m3	LOQ / CRQL ug/m3
TARGETS						
75-71-8	Dichlorodifluoromethane	0.42	2.08	J	0.54	2.47
74-87-3	Chloromethane	0.43	0.89	J	0.21	1.03
75-01-4	Vinyl Chloride	0.030	0.080	U	0.080	0.080
74-83-9	Bromomethane	0.080	0.31	U	0.31	1.94
75-00-3	Chloroethane	0.15	0.40	U	0.40	1.32
109-99-9	Tetrahydrofuran	0.080	0.24	U	0.24	1.47
75-69-4	Trichlorofluoromethane	0.21	1.18	J	0.96	2.81
76-13-1	1,1,2-Trichlorotrifluoroethane	0.14	1.07	U	1.07	3.83
76-14-2	Dichlorotetrafluoroethane	0.15	1.05	U	1.05	3.49
75-65-0	tert-Butyl alcohol	0.16	0.49	U	0.49	1.52
142-82-5	Heptane	0.17	0.70	U	0.70	2.05
75-35-4	1,1-Dichloroethene	0.15	0.59	U	0.59	1.98
67-64-1	Acetone	3.10	7.36		0.43	1.19
75-15-0	Carbon Disulfide	0.080	0.25	U	0.25	1.56
1634-04-4	Methyl tert-Butyl Ether	0.23	0.83	U	0.83	1.80
75-09-2	Methylene Chloride	0.73	2.54		0.80	1.74
156-60-5	trans-1,2-Dichloroethene	0.12	0.48	U	0.48	1.98
75-34-3	1,1-Dichloroethane	0.13	0.53	U	0.53	2.02
110-82-7	Cyclohexane	0.22	0.76	U	0.76	1.72
78-93-3	2-Butanone	0.22	0.65	J	0.18	1.47
56-23-5	Carbon Tetrachloride	0.070	0.44		0.19	0.19
156-59-2	cis-1,2-Dichloroethene	0.10	0.40	U	0.40	1.98
67-66-3	Chloroform	0.14	0.68	J	0.49	2.44
71-55-6	1,1,1-Trichloroethane	0.020	0.11	U	0.11	0.16
540-84-1	2,2,4-Trimethylpentane	0.14	0.65	U	0.65	2.34
71-43-2	Benzene	0.080	0.26	U	0.26	1.60
107-06-2	1,2-Dichloroethane	0.090	0.36	U	0.36	2.02
79-01-6	Trichloroethene	0.050	0.27		0.11	0.16
78-87-5	1,2-Dichloropropane	0.13	0.60	U	0.60	2.31
75-27-4	Bromodichloromethane	0.060	0.40	U	0.40	3.35
108-10-1	4-Methyl-2-Pentanone	0.10	0.41	U	0.41	2.05
108-88-3	Toluene	0.16	0.60	U	0.60	1.88
10061-02-6	t-1,3-Dichloropropene	0.15	0.68	U	0.68	2.27
10061-01-5	cis-1,3-Dichloropropene	0.11	0.50	U	0.50	2.27
79-00-5	1,1,2-Trichloroethane	0.080	0.44	U	0.44	2.73

Report of Analysis

Client:	GFE LLC	Date Collected:	10/08/25
Project:	200 Reservoir St, Trenton NJ	Date Received:	10/08/25
Client Sample ID:	IA2	SDG No.:	Q3320
Lab Sample ID:	Q3320-03	Matrix:	Air
Analytical Method:	TO-15	Test:	TO-15
Sample Wt/Vol:	400 Units: mL		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VL043085.D	1		10/09/25 13:18	VL100925

CAS Number	Parameter	Conc. ppbv	Conc. ug/M3	Qualifier	MDL ug/m3	LOQ / CRQL ug/m3
124-48-1	Dibromochloromethane	0.090	0.77	U	0.77	4.26
106-93-4	1,2-Dibromoethane	0.090	0.69	U	0.69	0.77
127-18-4	Tetrachloroethene	0.020	0.14	U	0.14	0.20
108-90-7	Chlorobenzene	0.080	0.37	U	0.37	2.30
100-41-4	Ethyl Benzene	0.19	0.83	U	0.83	2.17
179601-23-1	m/p-Xylene	0.41	1.78	U	1.78	4.34
95-47-6	o-Xylene	0.21	0.91	U	0.91	2.17
100-42-5	Styrene	0.16	0.68	U	0.68	2.13
75-25-2	Bromoform	0.050	0.52	U	0.52	5.17
79-34-5	1,1,2,2-Tetrachloroethane	0.020	0.14	U	0.14	0.21
95-49-8	2-Chlorotoluene	0.17	0.88	U	0.88	2.59
108-67-8	1,3,5-Trimethylbenzene	0.18	0.88	U	0.88	2.46
95-63-6	1,2,4-Trimethylbenzene	0.18	0.88	U	0.88	2.46
541-73-1	1,3-Dichlorobenzene	0.050	0.30	U	0.30	3.01
106-46-7	1,4-Dichlorobenzene	0.13	0.78	U	0.78	3.01
95-50-1	1,2-Dichlorobenzene	0.13	0.78	U	0.78	3.01
120-82-1	1,2,4-Trichlorobenzene	0.21	1.56	U	1.56	3.71
87-68-3	Hexachloro-1,3-Butadiene	0.13	1.39	U	1.39	5.33
106-99-0	1,3-Butadiene	0.050	0.11	U	0.11	1.11
91-20-3	Naphthalene	0.25	1.31		0.050	0.52
622-96-8	4-Ethyltoluene	0.21	1.03	U	1.03	2.46
110-54-3	Hexane	0.33	1.16	J	0.56	1.76
107-05-1	Allyl Chloride	0.11	0.34	U	0.34	1.57
123-91-1	1,4-Dioxane	0.22	0.79	U	0.79	1.80
80-62-6	Methyl Methacrylate	0.14	0.57	U	0.57	2.05
SURROGATES						
460-00-4	1-Bromo-4-Fluorobenzene	10.6			65 - 135	106% SPK: 10
INTERNAL STANDARDS						
74-97-5	Bromochloromethane	200000		2.78		
540-36-3	1,4-Difluorobenzene	529000		3.946		
3114-55-4	Chlorobenzene-d5	446000		8.869		

Report of Analysis

Client:	GFE LLC	Date Collected:	10/08/25
Project:	200 Reservoir St, Trenton NJ	Date Received:	10/08/25
Client Sample ID:	IA2	SDG No.:	Q3320
Lab Sample ID:	Q3320-03	Matrix:	Air
Analytical Method:	TO-15	Test:	TO-15
Sample Wt/Vol:	400	Units:	mL

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VL043085.D	1		10/09/25 13:18	VL100925

CAS Number	Parameter	Conc. ppbv	Conc. ug/M3	Qualifier	MDL	LOQ / CRQL
------------	-----------	---------------	----------------	-----------	-----	------------

U = Not Detected
 RL = Reporting Limit
 MDL = Method Detection Limit
 E = Value Exceeds Calibration Range
 D = Dilution

J = Estimated Value
 B = Analyte Found in Associated Method Blank
 N = Presumptive Evidence of a Compound
 * = Values outside of QC limits
 Q = indicates LCS control criteria did not meet requirements

Report of Analysis

Client:	GFE LLC	Date Collected:	10/08/25
Project:	200 Reservoir St, Trenton NJ	Date Received:	10/08/25
Client Sample ID:	SV2	SDG No.:	Q3320
Lab Sample ID:	Q3320-04	Matrix:	Air
Analytical Method:	TO-15	Test:	TO-15
Sample Wt/Vol:	400	Units:	mL

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VL043087.D	10		10/09/25 14:25	VL100925

CAS Number	Parameter	Conc. ppbv	Conc. ug/M3	Qualifier	MDL ug/m3	LOQ / CRQL ug/m3
TARGETS						
75-71-8	Dichlorodifluoromethane	1.10	5.44	U	5.44	24.7
74-87-3	Chloromethane	0.98	2.02	U	2.02	10.3
75-01-4	Vinyl Chloride	0.25	0.64	U	0.64	0.77
74-83-9	Bromomethane	0.77	2.99	U	2.99	19.4
75-00-3	Chloroethane	1.50	3.96	U	3.96	13.2
109-99-9	Tetrahydrofuran	0.80	2.36	U	2.36	14.8
75-69-4	Trichlorofluoromethane	11.8	66.3		9.55	28.1
76-13-1	1,1,2-Trichlorotrifluoroethane	1.40	10.7	U	10.7	38.3
76-14-2	Dichlorotetrafluoroethane	1.50	10.5	U	10.5	35.0
75-65-0	tert-Butyl alcohol	1.60	4.85	U	4.85	15.2
142-82-5	Heptane	1.70	6.97	U	6.97	20.5
75-35-4	1,1-Dichloroethene	1.50	5.95	U	5.95	19.8
67-64-1	Acetone	70.7	168		4.28	11.9
75-15-0	Carbon Disulfide	1.60	4.98	J	2.49	15.6
1634-04-4	Methyl tert-Butyl Ether	2.30	8.29	U	8.29	18.0
75-09-2	Methylene Chloride	3.10	10.8	J	7.99	17.4
156-60-5	trans-1,2-Dichloroethene	1.20	4.76	U	4.76	19.8
75-34-3	1,1-Dichloroethane	1.30	5.26	U	5.26	20.2
110-82-7	Cyclohexane	2.20	7.57	U	7.57	17.2
78-93-3	2-Butanone	2.70	7.96	J	1.65	14.8
56-23-5	Carbon Tetrachloride	0.25	1.57	U	1.57	1.89
156-59-2	cis-1,2-Dichloroethene	0.99	3.93	U	3.93	19.8
67-66-3	Chloroform	0.96	4.69	U	4.69	24.4
71-55-6	1,1,1-Trichloroethane	1.50	8.18		0.87	1.64
540-84-1	2,2,4-Trimethylpentane	1.40	6.54	U	6.54	23.4
71-43-2	Benzene	0.79	2.52	U	2.52	16.0
107-06-2	1,2-Dichloroethane	0.93	3.76	U	3.76	20.2
79-01-6	Trichloroethene	28.8	155		1.29	1.61
78-87-5	1,2-Dichloropropane	1.30	6.01	U	6.01	23.1
75-27-4	Bromodichloromethane	0.63	4.22	U	4.22	33.5
108-10-1	4-Methyl-2-Pentanone	0.97	3.98	U	3.98	20.5
108-88-3	Toluene	3.40	12.8	J	6.03	18.8
10061-02-6	t-1,3-Dichloropropene	1.50	6.81	U	6.81	22.7
10061-01-5	cis-1,3-Dichloropropene	1.10	4.99	U	4.99	22.7
79-00-5	1,1,2-Trichloroethane	0.81	4.42	U	4.42	27.3

Report of Analysis

Client:	GFE LLC	Date Collected:	10/08/25
Project:	200 Reservoir St, Trenton NJ	Date Received:	10/08/25
Client Sample ID:	SV2	SDG No.:	Q3320
Lab Sample ID:	Q3320-04	Matrix:	Air
Analytical Method:	TO-15	Test:	TO-15
Sample Wt/Vol:	400 Units: mL		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VL043087.D	10		10/09/25 14:25	VL100925

CAS Number	Parameter	Conc. ppbv	Conc. ug/M3	Qualifier	MDL ug/m3	LOQ / CRQL ug/m3
124-48-1	Dibromochloromethane	0.85	7.24	U	7.24	42.6
106-93-4	1,2-Dibromoethane	0.85	6.53	U	6.53	7.69
127-18-4	Tetrachloroethene	320	2170	E	1.02	2.03
108-90-7	Chlorobenzene	0.77	3.55	U	3.55	23.0
100-41-4	Ethyl Benzene	1.90	8.25	U	8.25	21.7
179601-23-1	m/p-Xylene	4.10	17.8	U	17.8	43.4
95-47-6	o-Xylene	2.10	9.12	U	9.12	21.7
100-42-5	Styrene	1.60	6.81	U	6.81	21.3
75-25-2	Bromoform	0.48	4.96	U	4.96	51.7
79-34-5	1,1,2,2-Tetrachloroethane	0.18	1.24	U	1.24	2.06
95-49-8	2-Chlorotoluene	1.70	8.80	U	8.80	25.9
108-67-8	1,3,5-Trimethylbenzene	1.80	8.85	U	8.85	24.6
95-63-6	1,2,4-Trimethylbenzene	1.80	8.85	U	8.85	24.6
541-73-1	1,3-Dichlorobenzene	0.54	3.25	U	3.25	30.1
106-46-7	1,4-Dichlorobenzene	1.30	7.82	U	7.82	30.1
95-50-1	1,2-Dichlorobenzene	1.30	7.82	U	7.82	30.1
120-82-1	1,2,4-Trichlorobenzene	2.10	15.6	U	15.6	37.1
87-68-3	Hexachloro-1,3-Butadiene	1.30	13.9	U	13.9	53.3
106-99-0	1,3-Butadiene	0.51	1.13	U	1.13	11.1
91-20-3	Naphthalene	0.13	0.68	U	0.68	5.24
622-96-8	4-Ethyltoluene	2.10	10.3	U	10.3	24.6
110-54-3	Hexane	9.50	33.5		5.64	17.6
107-05-1	Allyl Chloride	1.10	3.44	U	3.44	15.7
123-91-1	1,4-Dioxane	2.20	7.93	U	7.93	18.0
80-62-6	Methyl Methacrylate	1.40	5.73	U	5.73	20.5
SURROGATES						
460-00-4	1-Bromo-4-Fluorobenzene	11.8			65 - 135	118% SPK: 10
INTERNAL STANDARDS						
74-97-5	Bromochloromethane	189000		2.777		
540-36-3	1,4-Difluorobenzene	495000		3.943		
3114-55-4	Chlorobenzene-d5	493000		8.862		

Report of Analysis

Client:	GFE LLC	Date Collected:	10/08/25
Project:	200 Reservoir St, Trenton NJ	Date Received:	10/08/25
Client Sample ID:	SV2	SDG No.:	Q3320
Lab Sample ID:	Q3320-04	Matrix:	Air
Analytical Method:	TO-15	Test:	TO-15
Sample Wt/Vol:	400 Units: mL		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VL043087.D	10		10/09/25 14:25	VL100925

CAS Number	Parameter	Conc. ppbv	Conc. ug/M3	Qualifier	MDL	LOQ / CRQL
------------	-----------	---------------	----------------	-----------	-----	------------

U = Not Detected

RL = Reporting Limit

MDL = Method Detection Limit

E = Value Exceeds Calibration Range

D = Dilution

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

Q = indicates LCS control criteria did not meet requirements

Report of Analysis

Client:	GFE LLC	Date Collected:	10/08/25
Project:	200 Reservoir St, Trenton NJ	Date Received:	10/08/25
Client Sample ID:	SV2DL	SDG No.:	Q3320
Lab Sample ID:	Q3320-04DL	Matrix:	Air
Analytical Method:	TO-15	Test:	TO-15
Sample Wt/Vol:	400	Units:	mL

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VL043093.D	40		10/09/25 18:20	VL100925

CAS Number	Parameter	Conc. ppbv	Conc. ug/M3	Qualifier	MDL ug/m3	LOQ / CRQL ug/m3
TARGETS						
75-71-8	Dichlorodifluoromethane	4.40	21.8	UD	21.8	98.9
74-87-3	Chloromethane	3.90	8.05	UD	8.05	41.3
75-01-4	Vinyl Chloride	1.00	2.56	UD	2.56	3.07
74-83-9	Bromomethane	3.10	12.0	UD	12.0	77.7
75-00-3	Chloroethane	6.00	15.8	UD	15.8	52.8
109-99-9	Tetrahydrofuran	3.20	9.44	UD	9.44	59.0
75-69-4	Trichlorofluoromethane	11.8	66.3	JD	38.2	112
76-13-1	1,1,2-Trichlorotrifluoroethane	5.60	42.9	UD	42.9	153
76-14-2	Dichlorotetrafluoroethane	6.00	41.9	UD	41.9	140
75-65-0	tert-Butyl alcohol	6.40	19.4	UD	19.4	60.6
142-82-5	Heptane	6.80	27.9	UD	27.9	82.0
75-35-4	1,1-Dichloroethene	6.00	23.8	UD	23.8	79.3
67-64-1	Acetone	70.2	167	D	17.1	47.5
75-15-0	Carbon Disulfide	3.20	9.97	UD	9.97	62.3
1634-04-4	Methyl tert-Butyl Ether	9.20	33.2	UD	33.2	72.1
75-09-2	Methylene Chloride	9.70	33.7	JD	32.0	69.5
156-60-5	trans-1,2-Dichloroethene	4.80	19.0	UD	19.0	79.3
75-34-3	1,1-Dichloroethane	5.20	21.1	UD	21.1	81.0
110-82-7	Cyclohexane	8.80	30.3	UD	30.3	68.8
78-93-3	2-Butanone	2.20	6.49	UD	6.49	59.0
56-23-5	Carbon Tetrachloride	1.00	6.29	UD	6.29	7.55
156-59-2	cis-1,2-Dichloroethene	4.00	15.9	UD	15.9	79.3
67-66-3	Chloroform	3.80	18.6	UD	18.6	97.7
71-55-6	1,1,1-Trichloroethane	1.70	9.28	D	3.49	6.55
540-84-1	2,2,4-Trimethylpentane	5.60	26.2	UD	26.2	93.4
71-43-2	Benzene	3.20	10.2	UD	10.2	63.9
107-06-2	1,2-Dichloroethane	3.70	15.0	UD	15.0	81.0
79-01-6	Trichloroethene	28.5	153	D	5.16	6.45
78-87-5	1,2-Dichloropropane	5.20	24.0	UD	24.0	92.4
75-27-4	Bromodichloromethane	2.50	16.8	UD	16.8	134
108-10-1	4-Methyl-2-Pentanone	3.90	16.0	UD	16.0	82.0
108-88-3	Toluene	6.40	24.1	UD	24.1	75.4
10061-02-6	t-1,3-Dichloropropene	6.00	27.2	UD	27.2	90.8
10061-01-5	cis-1,3-Dichloropropene	4.40	20.0	UD	20.0	90.8
79-00-5	1,1,2-Trichloroethane	3.20	17.5	UD	17.5	109

Report of Analysis

Client:	GFE LLC	Date Collected:	10/08/25
Project:	200 Reservoir St, Trenton NJ	Date Received:	10/08/25
Client Sample ID:	SV2DL	SDG No.:	Q3320
Lab Sample ID:	Q3320-04DL	Matrix:	Air
Analytical Method:	TO-15	Test:	TO-15
Sample Wt/Vol:	400 Units: mL		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VL043093.D	40		10/09/25 18:20	VL100925

CAS Number	Parameter	Conc. ppbv	Conc. ug/M3	Qualifier	MDL ug/m3	LOQ / CRQL ug/m3
124-48-1	Dibromochloromethane	3.40	29.0	UD	29.0	170
106-93-4	1,2-Dibromoethane	3.40	26.1	UD	26.1	30.7
127-18-4	Tetrachloroethene	300	2030	D	4.07	8.14
108-90-7	Chlorobenzene	3.10	14.3	UD	14.3	92.1
100-41-4	Ethyl Benzene	7.60	33.0	UD	33.0	86.9
179601-23-1	m/p-Xylene	16.4	71.2	UD	71.2	174
95-47-6	o-Xylene	8.40	36.5	UD	36.5	86.9
100-42-5	Styrene	6.40	27.3	UD	27.3	85.2
75-25-2	Bromoform	1.90	19.6	UD	19.6	207
79-34-5	1,1,2,2-Tetrachloroethane	0.72	4.94	UD	4.94	8.24
95-49-8	2-Chlorotoluene	6.80	35.2	UD	35.2	104
108-67-8	1,3,5-Trimethylbenzene	7.20	35.4	UD	35.4	98.3
95-63-6	1,2,4-Trimethylbenzene	7.20	35.4	UD	35.4	98.3
541-73-1	1,3-Dichlorobenzene	2.20	13.2	UD	13.2	120
106-46-7	1,4-Dichlorobenzene	5.20	31.3	UD	31.3	120
95-50-1	1,2-Dichlorobenzene	5.20	31.3	UD	31.3	120
120-82-1	1,2,4-Trichlorobenzene	8.40	62.4	UD	62.4	148
87-68-3	Hexachloro-1,3-Butadiene	5.20	55.5	UD	55.5	213
106-99-0	1,3-Butadiene	2.00	4.42	UD	4.42	44.3
91-20-3	Naphthalene	0.52	2.73	UD	2.73	21.0
622-96-8	4-Ethyltoluene	8.40	41.3	UD	41.3	98.3
110-54-3	Hexane	8.90	31.4	JD	22.6	70.5
107-05-1	Allyl Chloride	4.40	13.8	UD	13.8	62.6
123-91-1	1,4-Dioxane	8.80	31.7	UD	31.7	72.1
80-62-6	Methyl Methacrylate	5.60	22.9	UD	22.9	81.9
SURROGATES						
460-00-4	1-Bromo-4-Fluorobenzene	11.8			65 - 135	118% SPK: 10
INTERNAL STANDARDS						
74-97-5	Bromochloromethane	170000		2.777		
540-36-3	1,4-Difluorobenzene	455000		3.942		
3114-55-4	Chlorobenzene-d5	411000		8.865		

Report of Analysis

Client:	GFE LLC	Date Collected:	10/08/25
Project:	200 Reservoir St, Trenton NJ	Date Received:	10/08/25
Client Sample ID:	SV2DL	SDG No.:	Q3320
Lab Sample ID:	Q3320-04DL	Matrix:	Air
Analytical Method:	TO-15	Test:	TO-15
Sample Wt/Vol:	400 Units: mL		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VL043093.D	40		10/09/25 18:20	VL100925

CAS Number	Parameter	Conc. ppbv	Conc. ug/M3	Qualifier	MDL	LOQ / CRQL
------------	-----------	---------------	----------------	-----------	-----	------------

U = Not Detected

RL = Reporting Limit

MDL = Method Detection Limit

E = Value Exceeds Calibration Range

D = Dilution

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

Q = indicates LCS control criteria did not meet requirements

Report of Analysis

Client:	GFE LLC	Date Collected:	10/08/25
Project:	200 Reservoir St, Trenton NJ	Date Received:	10/08/25
Client Sample ID:	IA3	SDG No.:	Q3320
Lab Sample ID:	Q3320-05	Matrix:	Air
Analytical Method:	TO-15	Test:	TO-15
Sample Wt/Vol:	400	Units:	mL

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VL043074.D	1		10/09/25 00:22	VL100825

CAS Number	Parameter	Conc. ppbv	Conc. ug/M3	Qualifier	MDL ug/m3	LOQ / CRQL ug/m3
TARGETS						
75-71-8	Dichlorodifluoromethane	0.53	2.62		0.54	2.47
74-87-3	Chloromethane	0.52	1.07		0.21	1.03
75-01-4	Vinyl Chloride	0.030	0.080	U	0.080	0.080
74-83-9	Bromomethane	0.080	0.31	U	0.31	1.94
75-00-3	Chloroethane	0.15	0.40	U	0.40	1.32
109-99-9	Tetrahydrofuran	0.080	0.24	U	0.24	1.47
75-69-4	Trichlorofluoromethane	0.26	1.46	J	0.96	2.81
76-13-1	1,1,2-Trichlorotrifluoroethane	0.14	1.07	U	1.07	3.83
76-14-2	Dichlorotetrafluoroethane	0.15	1.05	U	1.05	3.49
75-65-0	tert-Butyl alcohol	0.16	0.49	U	0.49	1.52
142-82-5	Heptane	0.17	0.70	U	0.70	2.05
75-35-4	1,1-Dichloroethene	0.15	0.59	U	0.59	1.98
67-64-1	Acetone	4.00	9.50		0.43	1.19
75-15-0	Carbon Disulfide	0.080	0.25	U	0.25	1.56
1634-04-4	Methyl tert-Butyl Ether	0.23	0.83	U	0.83	1.80
75-09-2	Methylene Chloride	1.90	6.60		0.80	1.74
156-60-5	trans-1,2-Dichloroethene	0.12	0.48	U	0.48	1.98
75-34-3	1,1-Dichloroethane	0.13	0.53	U	0.53	2.02
110-82-7	Cyclohexane	0.22	0.76	U	0.76	1.72
78-93-3	2-Butanone	0.20	0.59	J	0.18	1.47
56-23-5	Carbon Tetrachloride	0.090	0.57		0.19	0.19
156-59-2	cis-1,2-Dichloroethene	0.10	0.40	U	0.40	1.98
67-66-3	Chloroform	0.10	0.49	J	0.49	2.44
71-55-6	1,1,1-Trichloroethane	0.020	0.11	U	0.11	0.16
540-84-1	2,2,4-Trimethylpentane	0.14	0.65	U	0.65	2.34
71-43-2	Benzene	0.080	0.26	U	0.26	1.60
107-06-2	1,2-Dichloroethane	0.090	0.36	U	0.36	2.02
79-01-6	Trichloroethene	0.040	0.21		0.11	0.16
78-87-5	1,2-Dichloropropane	0.13	0.60	U	0.60	2.31
75-27-4	Bromodichloromethane	0.060	0.40	U	0.40	3.35
108-10-1	4-Methyl-2-Pentanone	0.10	0.41	U	0.41	2.05
108-88-3	Toluene	0.18	0.68	J	0.60	1.88
10061-02-6	t-1,3-Dichloropropene	0.15	0.68	U	0.68	2.27
10061-01-5	cis-1,3-Dichloropropene	0.11	0.50	U	0.50	2.27
79-00-5	1,1,2-Trichloroethane	0.080	0.44	U	0.44	2.73

Report of Analysis

Client:	GFE LLC	Date Collected:	10/08/25
Project:	200 Reservoir St, Trenton NJ	Date Received:	10/08/25
Client Sample ID:	IA3	SDG No.:	Q3320
Lab Sample ID:	Q3320-05	Matrix:	Air
Analytical Method:	TO-15	Test:	TO-15
Sample Wt/Vol:	400	Units:	mL

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VL043074.D	1		10/09/25 00:22	VL100825

CAS Number	Parameter	Conc. ppbv	Conc. ug/M3	Qualifier	MDL ug/m3	LOQ / CRQL ug/m3
124-48-1	Dibromochloromethane	0.090	0.77	U	0.77	4.26
106-93-4	1,2-Dibromoethane	0.090	0.69	U	0.69	0.77
127-18-4	Tetrachloroethene	0.040	0.27		0.14	0.20
108-90-7	Chlorobenzene	0.080	0.37	U	0.37	2.30
100-41-4	Ethyl Benzene	0.19	0.83	U	0.83	2.17
179601-23-1	m/p-Xylene	0.41	1.78	U	1.78	4.34
95-47-6	o-Xylene	0.21	0.91	U	0.91	2.17
100-42-5	Styrene	0.16	0.68	U	0.68	2.13
75-25-2	Bromoform	0.050	0.52	U	0.52	5.17
79-34-5	1,1,2,2-Tetrachloroethane	0.020	0.14	U	0.14	0.21
95-49-8	2-Chlorotoluene	0.17	0.88	U	0.88	2.59
108-67-8	1,3,5-Trimethylbenzene	0.18	0.88	U	0.88	2.46
95-63-6	1,2,4-Trimethylbenzene	0.18	0.88	U	0.88	2.46
541-73-1	1,3-Dichlorobenzene	0.050	0.30	U	0.30	3.01
106-46-7	1,4-Dichlorobenzene	0.13	0.78	U	0.78	3.01
95-50-1	1,2-Dichlorobenzene	0.13	0.78	U	0.78	3.01
120-82-1	1,2,4-Trichlorobenzene	0.21	1.56	U	1.56	3.71
87-68-3	Hexachloro-1,3-Butadiene	0.13	1.39	U	1.39	5.33
106-99-0	1,3-Butadiene	0.050	0.11	U	0.11	1.11
91-20-3	Naphthalene	0.010	0.050	U	0.050	0.52
622-96-8	4-Ethyltoluene	0.21	1.03	U	1.03	2.46
110-54-3	Hexane	1.20	4.23		0.56	1.76
107-05-1	Allyl Chloride	0.11	0.34	U	0.34	1.57
123-91-1	1,4-Dioxane	0.22	0.79	U	0.79	1.80
80-62-6	Methyl Methacrylate	0.14	0.57	U	0.57	2.05
SURROGATES						
460-00-4	1-Bromo-4-Fluorobenzene	9.70			65 - 135	97% SPK: 10
INTERNAL STANDARDS						
74-97-5	Bromochloromethane	167000		2.781		
540-36-3	1,4-Difluorobenzene	427000		3.949		
3114-55-4	Chlorobenzene-d5	339000		8.872		

Report of Analysis

Client:	GFE LLC	Date Collected:	10/08/25
Project:	200 Reservoir St, Trenton NJ	Date Received:	10/08/25
Client Sample ID:	IA3	SDG No.:	Q3320
Lab Sample ID:	Q3320-05	Matrix:	Air
Analytical Method:	TO-15	Test:	TO-15
Sample Wt/Vol:	400 Units: mL		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VL043074.D	1		10/09/25 00:22	VL100825

CAS Number	Parameter	Conc. ppbv	Conc. ug/M3	Qualifier	MDL	LOQ / CRQL
------------	-----------	---------------	----------------	-----------	-----	------------

U = Not Detected

RL = Reporting Limit

MDL = Method Detection Limit

E = Value Exceeds Calibration Range

D = Dilution

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

Q = indicates LCS control criteria did not meet requirements

Report of Analysis

Client:	GFE LLC	Date Collected:	10/08/25
Project:	200 Reservoir St, Trenton NJ	Date Received:	10/08/25
Client Sample ID:	SV3	SDG No.:	Q3320
Lab Sample ID:	Q3320-06	Matrix:	Air
Analytical Method:	TO-15	Test:	TO-15
Sample Wt/Vol:	400	Units:	mL

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VL043094.D	5		10/09/25 18:54	VL100925

CAS Number	Parameter	Conc. ppbv	Conc. ug/M3	Qualifier	MDL ug/m3	LOQ / CRQL ug/m3
TARGETS						
75-71-8	Dichlorodifluoromethane	0.55	2.72	U	2.72	12.4
74-87-3	Chloromethane	0.49	1.01	U	1.01	5.16
75-01-4	Vinyl Chloride	0.13	0.33	U	0.33	0.38
74-83-9	Bromomethane	0.39	1.51	U	1.51	9.71
75-00-3	Chloroethane	0.75	1.98	U	1.98	6.60
109-99-9	Tetrahydrofuran	0.83	2.45	J	1.18	7.37
75-69-4	Trichlorofluoromethane	0.95	5.34	J	4.78	14.1
76-13-1	1,1,2-Trichlorotrifluoroethane	0.70	5.37	U	5.37	19.2
76-14-2	Dichlorotetrafluoroethane	0.75	5.24	U	5.24	17.5
75-65-0	tert-Butyl alcohol	0.80	2.43	U	2.43	7.58
142-82-5	Heptane	2.20	9.02	J	3.48	10.3
75-35-4	1,1-Dichloroethene	0.75	2.97	U	2.97	9.91
67-64-1	Acetone	58.0	138		2.14	5.94
75-15-0	Carbon Disulfide	0.40	1.25	U	1.25	7.79
1634-04-4	Methyl tert-Butyl Ether	1.20	4.33	U	4.33	9.01
75-09-2	Methylene Chloride	1.30	4.52	J	4.17	8.69
156-60-5	trans-1,2-Dichloroethene	0.60	2.38	U	2.38	9.91
75-34-3	1,1-Dichloroethane	0.65	2.63	U	2.63	10.1
110-82-7	Cyclohexane	1.10	3.79	U	3.79	8.61
78-93-3	2-Butanone	1.10	3.24	J	0.83	7.37
56-23-5	Carbon Tetrachloride	0.13	0.82	U	0.82	0.94
156-59-2	cis-1,2-Dichloroethene	0.50	1.98	U	1.98	9.91
67-66-3	Chloroform	0.48	2.34	U	2.34	12.2
71-55-6	1,1,1-Trichloroethane	0.080	0.44	U	0.44	0.82
540-84-1	2,2,4-Trimethylpentane	0.70	3.27	U	3.27	11.7
71-43-2	Benzene	0.40	1.28	U	1.28	7.99
107-06-2	1,2-Dichloroethane	0.47	1.90	U	1.90	10.1
79-01-6	Trichloroethene	0.47	2.53		0.64	0.81
78-87-5	1,2-Dichloropropane	0.65	3.00	U	3.00	11.6
75-27-4	Bromodichloromethane	0.32	2.14	U	2.14	16.8
108-10-1	4-Methyl-2-Pentanone	0.49	2.01	U	2.01	10.3
108-88-3	Toluene	3.00	11.3		3.01	9.42
10061-02-6	t-1,3-Dichloropropene	0.75	3.40	U	3.40	11.3
10061-01-5	cis-1,3-Dichloropropene	0.55	2.50	U	2.50	11.3
79-00-5	1,1,2-Trichloroethane	0.41	2.24	U	2.24	13.6

Report of Analysis

Client:	GFE LLC	Date Collected:	10/08/25
Project:	200 Reservoir St, Trenton NJ	Date Received:	10/08/25
Client Sample ID:	SV3	SDG No.:	Q3320
Lab Sample ID:	Q3320-06	Matrix:	Air
Analytical Method:	TO-15	Test:	TO-15
Sample Wt/Vol:	400	Units:	mL

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VL043094.D	5		10/09/25 18:54	VL100925

CAS Number	Parameter	Conc. ppbv	Conc. ug/M3	Qualifier	MDL ug/m3	LOQ / CRQL ug/m3
124-48-1	Dibromochloromethane	0.43	3.66	U	3.66	21.3
106-93-4	1,2-Dibromoethane	0.43	3.30	U	3.30	3.84
127-18-4	Tetrachloroethene	0.45	3.05		0.54	1.02
108-90-7	Chlorobenzene	0.39	1.80	U	1.80	11.5
100-41-4	Ethyl Benzene	0.95	4.13	U	4.13	10.9
179601-23-1	m/p-Xylene	2.10	9.12	U	9.12	21.7
95-47-6	o-Xylene	1.10	4.78	J	4.78	10.9
100-42-5	Styrene	0.80	3.41	U	3.41	10.6
75-25-2	Bromoform	0.24	2.48	U	2.48	25.9
79-34-5	1,1,2,2-Tetrachloroethane	0.090	0.62	U	0.62	1.03
95-49-8	2-Chlorotoluene	0.85	4.40	U	4.40	12.9
108-67-8	1,3,5-Trimethylbenzene	0.90	4.42	U	4.42	12.3
95-63-6	1,2,4-Trimethylbenzene	0.90	4.42	U	4.42	12.3
541-73-1	1,3-Dichlorobenzene	0.27	1.62	U	1.62	15.0
106-46-7	1,4-Dichlorobenzene	0.65	3.91	U	3.91	15.0
95-50-1	1,2-Dichlorobenzene	0.65	3.91	U	3.91	15.0
120-82-1	1,2,4-Trichlorobenzene	1.10	8.17	U	8.17	18.6
87-68-3	Hexachloro-1,3-Butadiene	0.65	6.93	U	6.93	26.7
106-99-0	1,3-Butadiene	0.26	0.58	U	0.58	5.53
91-20-3	Naphthalene	0.070	0.37	U	0.37	2.62
622-96-8	4-Ethyltoluene	1.10	5.41	U	5.41	12.3
110-54-3	Hexane	17.2	60.6		2.82	8.81
107-05-1	Allyl Chloride	0.55	1.72	U	1.72	7.83
123-91-1	1,4-Dioxane	1.10	3.96	U	3.96	9.01
80-62-6	Methyl Methacrylate	0.70	2.87	U	2.87	10.2
SURROGATES						
460-00-4	1-Bromo-4-Fluorobenzene	11.2			65 - 135	112% SPK: 10
INTERNAL STANDARDS						
74-97-5	Bromochloromethane	192000		2.774		
540-36-3	1,4-Difluorobenzene	503000		3.942		
3114-55-4	Chlorobenzene-d5	519000		8.862		

Report of Analysis

Client:	GFE LLC	Date Collected:	10/08/25
Project:	200 Reservoir St, Trenton NJ	Date Received:	10/08/25
Client Sample ID:	SV3	SDG No.:	Q3320
Lab Sample ID:	Q3320-06	Matrix:	Air
Analytical Method:	TO-15	Test:	TO-15
Sample Wt/Vol:	400 Units: mL		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VL043094.D	5		10/09/25 18:54	VL100925

CAS Number	Parameter	Conc. ppbv	Conc. ug/M3	Qualifier	MDL	LOQ / CRQL
------------	-----------	---------------	----------------	-----------	-----	------------

U = Not Detected

RL = Reporting Limit

MDL = Method Detection Limit

E = Value Exceeds Calibration Range

D = Dilution

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

Q = indicates LCS control criteria did not meet requirements

Report of Analysis

Client:	GFE LLC	Date Collected:	10/08/25
Project:	200 Reservoir St, Trenton NJ	Date Received:	10/08/25
Client Sample ID:	IA4	SDG No.:	Q3320
Lab Sample ID:	Q3320-07	Matrix:	Air
Analytical Method:	TO-15	Test:	TO-15
Sample Wt/Vol:	400	Units:	mL

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VL043076.D	1		10/09/25 01:28	VL100825

CAS Number	Parameter	Conc. ppbv	Conc. ug/M3	Qualifier	MDL ug/m3	LOQ / CRQL ug/m3
TARGETS						
75-71-8	Dichlorodifluoromethane	0.51	2.52		0.54	2.47
74-87-3	Chloromethane	0.50	1.03		0.21	1.03
75-01-4	Vinyl Chloride	0.030	0.080	U	0.080	0.080
74-83-9	Bromomethane	0.080	0.31	U	0.31	1.94
75-00-3	Chloroethane	0.15	0.40	U	0.40	1.32
109-99-9	Tetrahydrofuran	0.080	0.24	U	0.24	1.47
75-69-4	Trichlorofluoromethane	0.26	1.46	J	0.96	2.81
76-13-1	1,1,2-Trichlorotrifluoroethane	0.14	1.07	U	1.07	3.83
76-14-2	Dichlorotetrafluoroethane	0.15	1.05	U	1.05	3.49
75-65-0	tert-Butyl alcohol	0.16	0.49	U	0.49	1.52
142-82-5	Heptane	0.17	0.70	U	0.70	2.05
75-35-4	1,1-Dichloroethene	0.15	0.59	U	0.59	1.98
67-64-1	Acetone	4.10	9.74		0.43	1.19
75-15-0	Carbon Disulfide	0.080	0.25	U	0.25	1.56
1634-04-4	Methyl tert-Butyl Ether	0.23	0.83	U	0.83	1.80
75-09-2	Methylene Chloride	0.62	2.15		0.80	1.74
156-60-5	trans-1,2-Dichloroethene	0.12	0.48	U	0.48	1.98
75-34-3	1,1-Dichloroethane	0.13	0.53	U	0.53	2.02
110-82-7	Cyclohexane	0.22	0.76	U	0.76	1.72
78-93-3	2-Butanone	0.22	0.65	J	0.18	1.47
56-23-5	Carbon Tetrachloride	0.090	0.57		0.19	0.19
156-59-2	cis-1,2-Dichloroethene	0.10	0.40	U	0.40	1.98
67-66-3	Chloroform	0.11	0.54	J	0.49	2.44
71-55-6	1,1,1-Trichloroethane	0.020	0.11	U	0.11	0.16
540-84-1	2,2,4-Trimethylpentane	0.14	0.65	U	0.65	2.34
71-43-2	Benzene	0.080	0.26	U	0.26	1.60
107-06-2	1,2-Dichloroethane	0.090	0.36	U	0.36	2.02
79-01-6	Trichloroethene	0.050	0.27		0.11	0.16
78-87-5	1,2-Dichloropropane	0.13	0.60	U	0.60	2.31
75-27-4	Bromodichloromethane	0.060	0.40	U	0.40	3.35
108-10-1	4-Methyl-2-Pentanone	0.10	0.41	U	0.41	2.05
108-88-3	Toluene	0.16	0.60	U	0.60	1.88
10061-02-6	t-1,3-Dichloropropene	0.15	0.68	U	0.68	2.27
10061-01-5	cis-1,3-Dichloropropene	0.11	0.50	U	0.50	2.27
79-00-5	1,1,2-Trichloroethane	0.080	0.44	U	0.44	2.73

Report of Analysis

Client:	GFE LLC	Date Collected:	10/08/25
Project:	200 Reservoir St, Trenton NJ	Date Received:	10/08/25
Client Sample ID:	IA4	SDG No.:	Q3320
Lab Sample ID:	Q3320-07	Matrix:	Air
Analytical Method:	TO-15	Test:	TO-15
Sample Wt/Vol:	400	Units:	mL

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VL043076.D	1		10/09/25 01:28	VL100825

CAS Number	Parameter	Conc. ppbv	Conc. ug/M3	Qualifier	MDL ug/m3	LOQ / CRQL ug/m3
124-48-1	Dibromochloromethane	0.090	0.77	U	0.77	4.26
106-93-4	1,2-Dibromoethane	0.090	0.69	U	0.69	0.77
127-18-4	Tetrachloroethene	0.020	0.14	U	0.14	0.20
108-90-7	Chlorobenzene	0.080	0.37	U	0.37	2.30
100-41-4	Ethyl Benzene	0.19	0.83	U	0.83	2.17
179601-23-1	m/p-Xylene	0.41	1.78	U	1.78	4.34
95-47-6	o-Xylene	0.21	0.91	U	0.91	2.17
100-42-5	Styrene	0.16	0.68	U	0.68	2.13
75-25-2	Bromoform	0.050	0.52	U	0.52	5.17
79-34-5	1,1,2,2-Tetrachloroethane	0.020	0.14	U	0.14	0.21
95-49-8	2-Chlorotoluene	0.17	0.88	U	0.88	2.59
108-67-8	1,3,5-Trimethylbenzene	0.18	0.88	U	0.88	2.46
95-63-6	1,2,4-Trimethylbenzene	0.18	0.88	U	0.88	2.46
541-73-1	1,3-Dichlorobenzene	0.050	0.30	U	0.30	3.01
106-46-7	1,4-Dichlorobenzene	0.13	0.78	U	0.78	3.01
95-50-1	1,2-Dichlorobenzene	0.13	0.78	U	0.78	3.01
120-82-1	1,2,4-Trichlorobenzene	0.21	1.56	U	1.56	3.71
87-68-3	Hexachloro-1,3-Butadiene	0.13	1.39	U	1.39	5.33
106-99-0	1,3-Butadiene	0.050	0.11	U	0.11	1.11
91-20-3	Naphthalene	0.010	0.050	U	0.050	0.52
622-96-8	4-Ethyltoluene	0.21	1.03	U	1.03	2.46
110-54-3	Hexane	0.37	1.30	J	0.56	1.76
107-05-1	Allyl Chloride	0.11	0.34	U	0.34	1.57
123-91-1	1,4-Dioxane	0.22	0.79	U	0.79	1.80
80-62-6	Methyl Methacrylate	0.14	0.57	U	0.57	2.05
SURROGATES						
460-00-4	1-Bromo-4-Fluorobenzene	9.90			65 - 135	99% SPK: 10
INTERNAL STANDARDS						
74-97-5	Bromochloromethane	166000		2.781		
540-36-3	1,4-Difluorobenzene	421000		3.949		
3114-55-4	Chlorobenzene-d5	326000		8.872		

Report of Analysis

Client:	GFE LLC	Date Collected:	10/08/25
Project:	200 Reservoir St, Trenton NJ	Date Received:	10/08/25
Client Sample ID:	IA4	SDG No.:	Q3320
Lab Sample ID:	Q3320-07	Matrix:	Air
Analytical Method:	TO-15	Test:	TO-15
Sample Wt/Vol:	400 Units: mL		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VL043076.D	1		10/09/25 01:28	VL100825

CAS Number	Parameter	Conc. ppbv	Conc. ug/M3	Qualifier	MDL	LOQ / CRQL
------------	-----------	---------------	----------------	-----------	-----	------------

U = Not Detected

RL = Reporting Limit

MDL = Method Detection Limit

E = Value Exceeds Calibration Range

D = Dilution

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

Q = indicates LCS control criteria did not meet requirements

Report of Analysis

Client:	GFE LLC	Date Collected:	10/08/25
Project:	200 Reservoir St, Trenton NJ	Date Received:	10/08/25
Client Sample ID:	SV4	SDG No.:	Q3320
Lab Sample ID:	Q3320-08	Matrix:	Air
Analytical Method:	TO-15	Test:	TO-15
Sample Wt/Vol:	400	Units:	mL

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VL043095.D	8		10/09/25 19:29	VL100925

CAS Number	Parameter	Conc. ppbv	Conc. ug/M3	Qualifier	MDL ug/m3	LOQ / CRQL ug/m3
TARGETS						
75-71-8	Dichlorodifluoromethane	0.88	4.35	U	4.35	19.8
74-87-3	Chloromethane	0.78	1.61	U	1.61	8.26
75-01-4	Vinyl Chloride	0.20	0.51	U	0.51	0.61
74-83-9	Bromomethane	0.62	2.41	U	2.41	15.5
75-00-3	Chloroethane	1.20	3.17	U	3.17	10.6
109-99-9	Tetrahydrofuran	1.30	3.83	J	1.89	11.8
75-69-4	Trichlorofluoromethane	1.40	7.87	U	7.87	22.5
76-13-1	1,1,2-Trichlorotrifluoroethane	1.10	8.43	U	8.43	30.7
76-14-2	Dichlorotetrafluoroethane	1.20	8.39	U	8.39	28.0
75-65-0	tert-Butyl alcohol	1.30	3.94	U	3.94	12.1
142-82-5	Heptane	2.90	11.9	J	5.74	16.4
75-35-4	1,1-Dichloroethene	1.20	4.76	U	4.76	15.9
67-64-1	Acetone	210	499	E	3.33	9.50
75-15-0	Carbon Disulfide	1.10	3.43	J	1.99	12.5
1634-04-4	Methyl tert-Butyl Ether	1.80	6.49	U	6.49	14.4
75-09-2	Methylene Chloride	3.20	11.1	J	6.25	13.9
156-60-5	trans-1,2-Dichloroethene	0.96	3.81	U	3.81	15.9
75-34-3	1,1-Dichloroethane	1.00	4.05	U	4.05	16.2
110-82-7	Cyclohexane	1.80	6.20	U	6.20	13.8
78-93-3	2-Butanone	4.10	12.1		1.33	11.8
56-23-5	Carbon Tetrachloride	0.20	1.26	U	1.26	1.51
156-59-2	cis-1,2-Dichloroethene	0.79	3.13	U	3.13	15.9
67-66-3	Chloroform	0.77	3.76	U	3.76	19.5
71-55-6	1,1,1-Trichloroethane	0.13	0.71	U	0.71	1.31
540-84-1	2,2,4-Trimethylpentane	1.10	5.14	U	5.14	18.7
71-43-2	Benzene	0.63	2.01	U	2.01	12.8
107-06-2	1,2-Dichloroethane	0.74	3.00	U	3.00	16.2
79-01-6	Trichloroethene	0.92	4.94		1.02	1.29
78-87-5	1,2-Dichloropropane	1.00	4.62	U	4.62	18.5
75-27-4	Bromodichloromethane	0.50	3.35	U	3.35	26.8
108-10-1	4-Methyl-2-Pentanone	0.78	3.20	U	3.20	16.4
108-88-3	Toluene	4.50	17.0		4.90	15.1
10061-02-6	t-1,3-Dichloropropene	1.20	5.45	U	5.45	18.2
10061-01-5	cis-1,3-Dichloropropene	0.88	4.00	U	4.00	18.2
79-00-5	1,1,2-Trichloroethane	0.65	3.55	U	3.55	21.8

Report of Analysis

Client:	GFE LLC	Date Collected:	10/08/25
Project:	200 Reservoir St, Trenton NJ	Date Received:	10/08/25
Client Sample ID:	SV4	SDG No.:	Q3320
Lab Sample ID:	Q3320-08	Matrix:	Air
Analytical Method:	TO-15	Test:	TO-15
Sample Wt/Vol:	400	Units:	mL

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VL043095.D	8		10/09/25 19:29	VL100925

CAS Number	Parameter	Conc. ppbv	Conc. ug/M3	Qualifier	MDL ug/m3	LOQ / CRQL ug/m3
124-48-1	Dibromochloromethane	0.68	5.79	U	5.79	34.1
106-93-4	1,2-Dibromoethane	0.68	5.23	U	5.23	6.15
127-18-4	Tetrachloroethene	1.60	10.8		0.81	1.63
108-90-7	Chlorobenzene	0.62	2.86	U	2.86	18.4
100-41-4	Ethyl Benzene	1.50	6.52	U	6.52	17.4
179601-23-1	m/p-Xylene	3.30	14.3	U	14.3	34.8
95-47-6	o-Xylene	2.20	9.56	J	7.38	17.4
100-42-5	Styrene	1.30	5.53	U	5.53	17.0
75-25-2	Bromoform	0.38	3.93	U	3.93	41.4
79-34-5	1,1,2,2-Tetrachloroethane	0.14	0.96	U	0.96	1.65
95-49-8	2-Chlorotoluene	1.40	7.25	U	7.25	20.7
108-67-8	1,3,5-Trimethylbenzene	1.40	6.88	U	6.88	19.7
95-63-6	1,2,4-Trimethylbenzene	1.40	6.88	U	6.88	19.7
541-73-1	1,3-Dichlorobenzene	0.43	2.59	U	2.59	24.1
106-46-7	1,4-Dichlorobenzene	1.00	6.01	U	6.01	24.1
95-50-1	1,2-Dichlorobenzene	1.00	6.01	U	6.01	24.1
120-82-1	1,2,4-Trichlorobenzene	1.70	12.6	U	12.6	29.7
87-68-3	Hexachloro-1,3-Butadiene	1.00	10.7	U	10.7	42.7
106-99-0	1,3-Butadiene	0.41	0.91	U	0.91	8.85
91-20-3	Naphthalene	0.10	0.52	U	0.52	4.19
622-96-8	4-Ethyltoluene	1.70	8.36	U	8.36	19.7
110-54-3	Hexane	25.7	90.6		4.58	14.1
107-05-1	Allyl Chloride	0.88	2.75	U	2.75	12.5
123-91-1	1,4-Dioxane	1.80	6.49	U	6.49	14.4
80-62-6	Methyl Methacrylate	1.10	4.50	U	4.50	16.4
SURROGATES						
460-00-4	1-Bromo-4-Fluorobenzene	11.4			65 - 135	114% SPK: 10
INTERNAL STANDARDS						
74-97-5	Bromochloromethane	173000		2.774		
540-36-3	1,4-Difluorobenzene	465000		3.939		
3114-55-4	Chlorobenzene-d5	485000		8.859		

Report of Analysis

Client:	GFE LLC	Date Collected:	10/08/25
Project:	200 Reservoir St, Trenton NJ	Date Received:	10/08/25
Client Sample ID:	SV4	SDG No.:	Q3320
Lab Sample ID:	Q3320-08	Matrix:	Air
Analytical Method:	TO-15	Test:	TO-15
Sample Wt/Vol:	400 Units: mL		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VL043095.D	8		10/09/25 19:29	VL100925

CAS Number	Parameter	Conc. ppbv	Conc. ug/M3	Qualifier	MDL	LOQ / CRQL
------------	-----------	---------------	----------------	-----------	-----	------------

U = Not Detected

RL = Reporting Limit

MDL = Method Detection Limit

E = Value Exceeds Calibration Range

D = Dilution

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

Q = indicates LCS control criteria did not meet requirements

Report of Analysis

Client:	GFE LLC	Date Collected:	10/08/25
Project:	200 Reservoir St, Trenton NJ	Date Received:	10/08/25
Client Sample ID:	SV4DL	SDG No.:	Q3320
Lab Sample ID:	Q3320-08DL	Matrix:	Air
Analytical Method:	TO-15	Test:	TO-15
Sample Wt/Vol:	400	Units:	mL

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VL043101.D	40		10/09/25 22:58	VL100925

CAS Number	Parameter	Conc. ppbv	Conc. ug/M3	Qualifier	MDL ug/m3	LOQ / CRQL ug/m3
TARGETS						
75-71-8	Dichlorodifluoromethane	4.40	21.8	UD	21.8	98.9
74-87-3	Chloromethane	3.90	8.05	UD	8.05	41.3
75-01-4	Vinyl Chloride	1.00	2.56	UD	2.56	3.07
74-83-9	Bromomethane	3.10	12.0	UD	12.0	77.7
75-00-3	Chloroethane	6.00	15.8	UD	15.8	52.8
109-99-9	Tetrahydrofuran	3.20	9.44	UD	9.44	59.0
75-69-4	Trichlorofluoromethane	6.80	38.2	UD	38.2	112
76-13-1	1,1,2-Trichlorotrifluoroethane	5.60	42.9	UD	42.9	153
76-14-2	Dichlorotetrafluoroethane	6.00	41.9	UD	41.9	140
75-65-0	tert-Butyl alcohol	6.40	19.4	UD	19.4	60.6
142-82-5	Heptane	6.80	27.9	UD	27.9	82.0
75-35-4	1,1-Dichloroethene	6.00	23.8	UD	23.8	79.3
67-64-1	Acetone	230	546	D	17.1	47.5
75-15-0	Carbon Disulfide	3.20	9.97	UD	9.97	62.3
1634-04-4	Methyl tert-Butyl Ether	9.20	33.2	UD	33.2	72.1
75-09-2	Methylene Chloride	9.70	33.7	JD	32.0	69.5
156-60-5	trans-1,2-Dichloroethene	4.80	19.0	UD	19.0	79.3
75-34-3	1,1-Dichloroethane	5.20	21.1	UD	21.1	81.0
110-82-7	Cyclohexane	8.80	30.3	UD	30.3	68.8
78-93-3	2-Butanone	4.70	13.9	JD	6.49	59.0
56-23-5	Carbon Tetrachloride	1.00	6.29	UD	6.29	7.55
156-59-2	cis-1,2-Dichloroethene	4.00	15.9	UD	15.9	79.3
67-66-3	Chloroform	3.80	18.6	UD	18.6	97.7
71-55-6	1,1,1-Trichloroethane	0.64	3.49	UD	3.49	6.55
540-84-1	2,2,4-Trimethylpentane	5.60	26.2	UD	26.2	93.4
71-43-2	Benzene	3.20	10.2	UD	10.2	63.9
107-06-2	1,2-Dichloroethane	3.70	15.0	UD	15.0	81.0
79-01-6	Trichloroethene	0.96	5.16	UD	5.16	6.45
78-87-5	1,2-Dichloropropane	5.20	24.0	UD	24.0	92.4
75-27-4	Bromodichloromethane	2.50	16.8	UD	16.8	134
108-10-1	4-Methyl-2-Pentanone	3.90	16.0	UD	16.0	82.0
108-88-3	Toluene	6.40	24.1	UD	24.1	75.4
10061-02-6	t-1,3-Dichloropropene	6.00	27.2	UD	27.2	90.8
10061-01-5	cis-1,3-Dichloropropene	4.40	20.0	UD	20.0	90.8
79-00-5	1,1,2-Trichloroethane	3.20	17.5	UD	17.5	109

Report of Analysis

Client:	GFE LLC	Date Collected:	10/08/25
Project:	200 Reservoir St, Trenton NJ	Date Received:	10/08/25
Client Sample ID:	SV4DL	SDG No.:	Q3320
Lab Sample ID:	Q3320-08DL	Matrix:	Air
Analytical Method:	TO-15	Test:	TO-15
Sample Wt/Vol:	400	Units:	mL

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VL043101.D	40		10/09/25 22:58	VL100925

CAS Number	Parameter	Conc. ppbv	Conc. ug/M3	Qualifier	MDL ug/m3	LOQ / CRQL ug/m3
124-48-1	Dibromochloromethane	3.40	29.0	UD	29.0	170
106-93-4	1,2-Dibromoethane	3.40	26.1	UD	26.1	30.7
127-18-4	Tetrachloroethene	1.80	12.2	D	4.07	8.14
108-90-7	Chlorobenzene	3.10	14.3	UD	14.3	92.1
100-41-4	Ethyl Benzene	7.60	33.0	UD	33.0	86.9
179601-23-1	m/p-Xylene	16.4	71.2	UD	71.2	174
95-47-6	o-Xylene	8.40	36.5	UD	36.5	86.9
100-42-5	Styrene	6.40	27.3	UD	27.3	85.2
75-25-2	Bromoform	1.90	19.6	UD	19.6	207
79-34-5	1,1,2,2-Tetrachloroethane	0.72	4.94	UD	4.94	8.24
95-49-8	2-Chlorotoluene	6.80	35.2	UD	35.2	104
108-67-8	1,3,5-Trimethylbenzene	7.20	35.4	UD	35.4	98.3
95-63-6	1,2,4-Trimethylbenzene	7.20	35.4	UD	35.4	98.3
541-73-1	1,3-Dichlorobenzene	2.20	13.2	UD	13.2	120
106-46-7	1,4-Dichlorobenzene	5.20	31.3	UD	31.3	120
95-50-1	1,2-Dichlorobenzene	5.20	31.3	UD	31.3	120
120-82-1	1,2,4-Trichlorobenzene	8.40	62.4	UD	62.4	148
87-68-3	Hexachloro-1,3-Butadiene	5.20	55.5	UD	55.5	213
106-99-0	1,3-Butadiene	2.00	4.42	UD	4.42	44.3
91-20-3	Naphthalene	0.52	2.73	UD	2.73	21.0
622-96-8	4-Ethyltoluene	8.40	41.3	UD	41.3	98.3
110-54-3	Hexane	16.9	59.6	JD	22.6	70.5
107-05-1	Allyl Chloride	4.40	13.8	UD	13.8	62.6
123-91-1	1,4-Dioxane	8.80	31.7	UD	31.7	72.1
80-62-6	Methyl Methacrylate	5.60	22.9	UD	22.9	81.9
SURROGATES						
460-00-4	1-Bromo-4-Fluorobenzene	12.4			65 - 135	124% SPK: 10
INTERNAL STANDARDS						
74-97-5	Bromochloromethane	170000		2.774		
540-36-3	1,4-Difluorobenzene	428000		3.939		
3114-55-4	Chlorobenzene-d5	411000		8.862		

Report of Analysis

Client:	GFE LLC	Date Collected:	10/08/25
Project:	200 Reservoir St, Trenton NJ	Date Received:	10/08/25
Client Sample ID:	SV4DL	SDG No.:	Q3320
Lab Sample ID:	Q3320-08DL	Matrix:	Air
Analytical Method:	TO-15	Test:	TO-15
Sample Wt/Vol:	400	Units:	mL

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VL043101.D	40		10/09/25 22:58	VL100925

CAS Number	Parameter	Conc. ppbv	Conc. ug/M3	Qualifier	MDL	LOQ / CRQL
------------	-----------	---------------	----------------	-----------	-----	------------

U = Not Detected

RL = Reporting Limit

MDL = Method Detection Limit

E = Value Exceeds Calibration Range

D = Dilution

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

Q = indicates LCS control criteria did not meet requirements

Report of Analysis

Client:	GFE LLC	Date Collected:	10/08/25
Project:	200 Reservoir St, Trenton NJ	Date Received:	10/08/25
Client Sample ID:	IA5	SDG No.:	Q3320
Lab Sample ID:	Q3320-09	Matrix:	Air
Analytical Method:	TO-15	Test:	TO-15
Sample Wt/Vol:	400	Units:	mL

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VL043072.D	1		10/08/25 23:16	VL100825

CAS Number	Parameter	Conc. ppbv	Conc. ug/M3	Qualifier	MDL ug/m3	LOQ / CRQL ug/m3
TARGETS						
75-71-8	Dichlorodifluoromethane	0.50	2.47		0.54	2.47
74-87-3	Chloromethane	0.49	1.01	J	0.21	1.03
75-01-4	Vinyl Chloride	0.030	0.080	U	0.080	0.080
74-83-9	Bromomethane	0.080	0.31	U	0.31	1.94
75-00-3	Chloroethane	0.15	0.40	U	0.40	1.32
109-99-9	Tetrahydrofuran	0.080	0.24	U	0.24	1.47
75-69-4	Trichlorofluoromethane	0.24	1.35	J	0.96	2.81
76-13-1	1,1,2-Trichlorotrifluoroethane	0.14	1.07	U	1.07	3.83
76-14-2	Dichlorotetrafluoroethane	0.15	1.05	U	1.05	3.49
75-65-0	tert-Butyl alcohol	0.16	0.49	U	0.49	1.52
142-82-5	Heptane	0.17	0.70	U	0.70	2.05
75-35-4	1,1-Dichloroethene	0.15	0.59	U	0.59	1.98
67-64-1	Acetone	10.1	24.0		0.43	1.19
75-15-0	Carbon Disulfide	0.080	0.25	U	0.25	1.56
1634-04-4	Methyl tert-Butyl Ether	0.23	0.83	U	0.83	1.80
75-09-2	Methylene Chloride	0.67	2.33		0.80	1.74
156-60-5	trans-1,2-Dichloroethene	0.12	0.48	U	0.48	1.98
75-34-3	1,1-Dichloroethane	0.13	0.53	U	0.53	2.02
110-82-7	Cyclohexane	0.22	0.76	U	0.76	1.72
78-93-3	2-Butanone	0.36	1.06	J	0.18	1.47
56-23-5	Carbon Tetrachloride	0.080	0.50		0.19	0.19
156-59-2	cis-1,2-Dichloroethene	0.10	0.40	U	0.40	1.98
67-66-3	Chloroform	0.32	1.56	J	0.49	2.44
71-55-6	1,1,1-Trichloroethane	0.020	0.11	U	0.11	0.16
540-84-1	2,2,4-Trimethylpentane	0.14	0.65	U	0.65	2.34
71-43-2	Benzene	0.080	0.26	U	0.26	1.60
107-06-2	1,2-Dichloroethane	0.090	0.36	U	0.36	2.02
79-01-6	Trichloroethene	0.020	0.11	U	0.11	0.16
78-87-5	1,2-Dichloropropane	0.13	0.60	U	0.60	2.31
75-27-4	Bromodichloromethane	0.060	0.40	U	0.40	3.35
108-10-1	4-Methyl-2-Pentanone	0.10	0.41	U	0.41	2.05
108-88-3	Toluene	0.30	1.13	J	0.60	1.88
10061-02-6	t-1,3-Dichloropropene	0.15	0.68	U	0.68	2.27
10061-01-5	cis-1,3-Dichloropropene	0.11	0.50	U	0.50	2.27
79-00-5	1,1,2-Trichloroethane	0.080	0.44	U	0.44	2.73

Report of Analysis

Client:	GFE LLC	Date Collected:	10/08/25
Project:	200 Reservoir St, Trenton NJ	Date Received:	10/08/25
Client Sample ID:	IA5	SDG No.:	Q3320
Lab Sample ID:	Q3320-09	Matrix:	Air
Analytical Method:	TO-15	Test:	TO-15
Sample Wt/Vol:	400	Units:	mL

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VL043072.D	1		10/08/25 23:16	VL100825

CAS Number	Parameter	Conc. ppbv	Conc. ug/M3	Qualifier	MDL ug/m3	LOQ / CRQL ug/m3
124-48-1	Dibromochloromethane	0.090	0.77	U	0.77	4.26
106-93-4	1,2-Dibromoethane	0.090	0.69	U	0.69	0.77
127-18-4	Tetrachloroethene	0.050	0.34		0.14	0.20
108-90-7	Chlorobenzene	0.080	0.37	U	0.37	2.30
100-41-4	Ethyl Benzene	0.19	0.83	U	0.83	2.17
179601-23-1	m/p-Xylene	0.41	1.78	U	1.78	4.34
95-47-6	o-Xylene	0.21	0.91	U	0.91	2.17
100-42-5	Styrene	0.16	0.68	U	0.68	2.13
75-25-2	Bromoform	0.050	0.52	U	0.52	5.17
79-34-5	1,1,2,2-Tetrachloroethane	0.020	0.14	U	0.14	0.21
95-49-8	2-Chlorotoluene	0.17	0.88	U	0.88	2.59
108-67-8	1,3,5-Trimethylbenzene	0.18	0.88	U	0.88	2.46
95-63-6	1,2,4-Trimethylbenzene	0.18	0.88	U	0.88	2.46
541-73-1	1,3-Dichlorobenzene	0.050	0.30	U	0.30	3.01
106-46-7	1,4-Dichlorobenzene	0.13	0.78	U	0.78	3.01
95-50-1	1,2-Dichlorobenzene	0.13	0.78	U	0.78	3.01
120-82-1	1,2,4-Trichlorobenzene	0.21	1.56	U	1.56	3.71
87-68-3	Hexachloro-1,3-Butadiene	0.13	1.39	U	1.39	5.33
106-99-0	1,3-Butadiene	0.050	0.11	U	0.11	1.11
91-20-3	Naphthalene	0.010	0.050	U	0.050	0.52
622-96-8	4-Ethyltoluene	0.21	1.03	U	1.03	2.46
110-54-3	Hexane	0.43	1.52	J	0.56	1.76
107-05-1	Allyl Chloride	0.11	0.34	U	0.34	1.57
123-91-1	1,4-Dioxane	0.22	0.79	U	0.79	1.80
80-62-6	Methyl Methacrylate	0.14	0.57	U	0.57	2.05
SURROGATES						
460-00-4	1-Bromo-4-Fluorobenzene	10.3			65 - 135	103% SPK: 10
INTERNAL STANDARDS						
74-97-5	Bromochloromethane	167000		2.784		
540-36-3	1,4-Difluorobenzene	430000		3.952		
3114-55-4	Chlorobenzene-d5	343000		8.875		

Report of Analysis

Client:	GFE LLC	Date Collected:	10/08/25
Project:	200 Reservoir St, Trenton NJ	Date Received:	10/08/25
Client Sample ID:	IA5	SDG No.:	Q3320
Lab Sample ID:	Q3320-09	Matrix:	Air
Analytical Method:	TO-15	Test:	TO-15
Sample Wt/Vol:	400 Units: mL		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VL043072.D	1		10/08/25 23:16	VL100825

CAS Number	Parameter	Conc. ppbv	Conc. ug/M3	Qualifier	MDL	LOQ / CRQL
------------	-----------	---------------	----------------	-----------	-----	------------

U = Not Detected

RL = Reporting Limit

MDL = Method Detection Limit

E = Value Exceeds Calibration Range

D = Dilution

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

Q = indicates LCS control criteria did not meet requirements

Report of Analysis

Client:	GFE LLC	Date Collected:	10/08/25
Project:	200 Reservoir St, Trenton NJ	Date Received:	10/08/25
Client Sample ID:	SV5	SDG No.:	Q3320
Lab Sample ID:	Q3320-10	Matrix:	Air
Analytical Method:	TO-15	Test:	TO-15
Sample Wt/Vol:	400	Units:	mL

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VL043096.D	8		10/09/25 20:04	VL100925

CAS Number	Parameter	Conc. ppbv	Conc. ug/M3	Qualifier	MDL ug/m3	LOQ / CRQL ug/m3
TARGETS						
75-71-8	Dichlorodifluoromethane	0.88	4.35	U	4.35	19.8
74-87-3	Chloromethane	0.78	1.61	U	1.61	8.26
75-01-4	Vinyl Chloride	0.20	0.51	U	0.51	0.61
74-83-9	Bromomethane	0.62	2.41	U	2.41	15.5
75-00-3	Chloroethane	1.20	3.17	U	3.17	10.6
109-99-9	Tetrahydrofuran	0.64	1.89	U	1.89	11.8
75-69-4	Trichlorofluoromethane	1.40	7.87	U	7.87	22.5
76-13-1	1,1,2-Trichlorotrifluoroethane	1.10	8.43	U	8.43	30.7
76-14-2	Dichlorotetrafluoroethane	1.20	8.39	U	8.39	28.0
75-65-0	tert-Butyl alcohol	1.70	5.15	J	3.94	12.1
142-82-5	Heptane	1.40	5.74	U	5.74	16.4
75-35-4	1,1-Dichloroethene	1.20	4.76	U	4.76	15.9
67-64-1	Acetone	73.1	174		3.33	9.50
75-15-0	Carbon Disulfide	0.64	1.99	U	1.99	12.5
1634-04-4	Methyl tert-Butyl Ether	1.80	6.49	U	6.49	14.4
75-09-2	Methylene Chloride	2.10	7.30	J	6.25	13.9
156-60-5	trans-1,2-Dichloroethene	0.96	3.81	U	3.81	15.9
75-34-3	1,1-Dichloroethane	1.00	4.05	U	4.05	16.2
110-82-7	Cyclohexane	1.80	6.20	U	6.20	13.8
78-93-3	2-Butanone	0.85	2.51	J	1.33	11.8
56-23-5	Carbon Tetrachloride	0.20	1.26	U	1.26	1.51
156-59-2	cis-1,2-Dichloroethene	0.79	3.13	U	3.13	15.9
67-66-3	Chloroform	0.77	3.76	U	3.76	19.5
71-55-6	1,1,1-Trichloroethane	0.13	0.71	U	0.71	1.31
540-84-1	2,2,4-Trimethylpentane	1.10	5.14	U	5.14	18.7
71-43-2	Benzene	0.63	2.01	U	2.01	12.8
107-06-2	1,2-Dichloroethane	0.74	3.00	U	3.00	16.2
79-01-6	Trichloroethene	1.00	5.37		1.02	1.29
78-87-5	1,2-Dichloropropane	1.00	4.62	U	4.62	18.5
75-27-4	Bromodichloromethane	0.50	3.35	U	3.35	26.8
108-10-1	4-Methyl-2-Pentanone	0.78	3.20	U	3.20	16.4
108-88-3	Toluene	2.30	8.67	J	4.90	15.1
10061-02-6	t-1,3-Dichloropropene	1.20	5.45	U	5.45	18.2
10061-01-5	cis-1,3-Dichloropropene	0.88	4.00	U	4.00	18.2
79-00-5	1,1,2-Trichloroethane	0.65	3.55	U	3.55	21.8

Report of Analysis

Client:	GFE LLC	Date Collected:	10/08/25
Project:	200 Reservoir St, Trenton NJ	Date Received:	10/08/25
Client Sample ID:	SV5	SDG No.:	Q3320
Lab Sample ID:	Q3320-10	Matrix:	Air
Analytical Method:	TO-15	Test:	TO-15
Sample Wt/Vol:	400	Units:	mL

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VL043096.D	8		10/09/25 20:04	VL100925

CAS Number	Parameter	Conc. ppbv	Conc. ug/M3	Qualifier	MDL ug/m3	LOQ / CRQL ug/m3
124-48-1	Dibromochloromethane	0.68	5.79	U	5.79	34.1
106-93-4	1,2-Dibromoethane	0.68	5.23	U	5.23	6.15
127-18-4	Tetrachloroethene	6.20	42.0		0.81	1.63
108-90-7	Chlorobenzene	0.62	2.86	U	2.86	18.4
100-41-4	Ethyl Benzene	1.50	6.52	U	6.52	17.4
179601-23-1	m/p-Xylene	3.30	14.3	U	14.3	34.8
95-47-6	o-Xylene	1.70	7.38	U	7.38	17.4
100-42-5	Styrene	1.30	5.53	U	5.53	17.0
75-25-2	Bromoform	0.38	3.93	U	3.93	41.4
79-34-5	1,1,2,2-Tetrachloroethane	0.14	0.96	U	0.96	1.65
95-49-8	2-Chlorotoluene	1.40	7.25	U	7.25	20.7
108-67-8	1,3,5-Trimethylbenzene	1.40	6.88	U	6.88	19.7
95-63-6	1,2,4-Trimethylbenzene	1.40	6.88	U	6.88	19.7
541-73-1	1,3-Dichlorobenzene	0.43	2.59	U	2.59	24.1
106-46-7	1,4-Dichlorobenzene	1.00	6.01	U	6.01	24.1
95-50-1	1,2-Dichlorobenzene	1.00	6.01	U	6.01	24.1
120-82-1	1,2,4-Trichlorobenzene	1.70	12.6	U	12.6	29.7
87-68-3	Hexachloro-1,3-Butadiene	1.00	10.7	U	10.7	42.7
106-99-0	1,3-Butadiene	0.41	0.91	U	0.91	8.85
91-20-3	Naphthalene	0.10	0.52	U	0.52	4.19
622-96-8	4-Ethyltoluene	1.70	8.36	U	8.36	19.7
110-54-3	Hexane	8.00	28.2		4.58	14.1
107-05-1	Allyl Chloride	0.88	2.75	U	2.75	12.5
123-91-1	1,4-Dioxane	1.80	6.49	U	6.49	14.4
80-62-6	Methyl Methacrylate	1.10	4.50	U	4.50	16.4
SURROGATES						
460-00-4	1-Bromo-4-Fluorobenzene	11.6			65 - 135	116% SPK: 10
INTERNAL STANDARDS						
74-97-5	Bromochloromethane	196000		2.777		
540-36-3	1,4-Difluorobenzene	526000		3.939		
3114-55-4	Chlorobenzene-d5	487000		8.862		

Report of Analysis

Client:	GFE LLC	Date Collected:	10/08/25
Project:	200 Reservoir St, Trenton NJ	Date Received:	10/08/25
Client Sample ID:	SV5	SDG No.:	Q3320
Lab Sample ID:	Q3320-10	Matrix:	Air
Analytical Method:	TO-15	Test:	TO-15
Sample Wt/Vol:	400 Units: mL		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VL043096.D	8		10/09/25 20:04	VL100925

CAS Number	Parameter	Conc. ppbv	Conc. ug/M3	Qualifier	MDL	LOQ / CRQL
------------	-----------	---------------	----------------	-----------	-----	------------

U = Not Detected

RL = Reporting Limit

MDL = Method Detection Limit

E = Value Exceeds Calibration Range

D = Dilution

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

Q = indicates LCS control criteria did not meet requirements

Report of Analysis

Client:	GFE LLC	Date Collected:	10/08/25
Project:	200 Reservoir St, Trenton NJ	Date Received:	10/08/25
Client Sample ID:	IA6	SDG No.:	Q3320
Lab Sample ID:	Q3320-11	Matrix:	Air
Analytical Method:	TO-15	Test:	TO-15
Sample Wt/Vol:	400	Units:	mL

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VL043086.D	1		10/09/25 13:52	VL100925

CAS Number	Parameter	Conc. ppbv	Conc. ug/M3	Qualifier	MDL ug/m3	LOQ / CRQL ug/m3
TARGETS						
75-71-8	Dichlorodifluoromethane	0.42	2.08	J	0.54	2.47
74-87-3	Chloromethane	0.46	0.95	J	0.21	1.03
75-01-4	Vinyl Chloride	0.030	0.080	U	0.080	0.080
74-83-9	Bromomethane	0.080	0.31	U	0.31	1.94
75-00-3	Chloroethane	0.15	0.40	U	0.40	1.32
109-99-9	Tetrahydrofuran	0.080	0.24	U	0.24	1.47
75-69-4	Trichlorofluoromethane	0.20	1.12	J	0.96	2.81
76-13-1	1,1,2-Trichlorotrifluoroethane	0.14	1.07	U	1.07	3.83
76-14-2	Dichlorotetrafluoroethane	0.15	1.05	U	1.05	3.49
75-65-0	tert-Butyl alcohol	0.16	0.49	U	0.49	1.52
142-82-5	Heptane	0.17	0.70	U	0.70	2.05
75-35-4	1,1-Dichloroethene	0.15	0.59	U	0.59	1.98
67-64-1	Acetone	8.70	20.7		0.43	1.19
75-15-0	Carbon Disulfide	0.080	0.25	U	0.25	1.56
1634-04-4	Methyl tert-Butyl Ether	0.23	0.83	U	0.83	1.80
75-09-2	Methylene Chloride	0.53	1.84		0.80	1.74
156-60-5	trans-1,2-Dichloroethene	0.12	0.48	U	0.48	1.98
75-34-3	1,1-Dichloroethane	0.13	0.53	U	0.53	2.02
110-82-7	Cyclohexane	0.22	0.76	U	0.76	1.72
78-93-3	2-Butanone	0.18	0.53	J	0.18	1.47
56-23-5	Carbon Tetrachloride	0.070	0.44		0.19	0.19
156-59-2	cis-1,2-Dichloroethene	0.10	0.40	U	0.40	1.98
67-66-3	Chloroform	0.22	1.07	J	0.49	2.44
71-55-6	1,1,1-Trichloroethane	0.020	0.11	U	0.11	0.16
540-84-1	2,2,4-Trimethylpentane	0.14	0.65	U	0.65	2.34
71-43-2	Benzene	0.080	0.26	U	0.26	1.60
107-06-2	1,2-Dichloroethane	0.090	0.36	U	0.36	2.02
79-01-6	Trichloroethene	0.020	0.11	U	0.11	0.16
78-87-5	1,2-Dichloropropane	0.13	0.60	U	0.60	2.31
75-27-4	Bromodichloromethane	0.060	0.40	U	0.40	3.35
108-10-1	4-Methyl-2-Pentanone	0.10	0.41	U	0.41	2.05
108-88-3	Toluene	0.29	1.09	J	0.60	1.88
10061-02-6	t-1,3-Dichloropropene	0.15	0.68	U	0.68	2.27
10061-01-5	cis-1,3-Dichloropropene	0.11	0.50	U	0.50	2.27
79-00-5	1,1,2-Trichloroethane	0.080	0.44	U	0.44	2.73

Report of Analysis

Client:	GFE LLC	Date Collected:	10/08/25
Project:	200 Reservoir St, Trenton NJ	Date Received:	10/08/25
Client Sample ID:	IA6	SDG No.:	Q3320
Lab Sample ID:	Q3320-11	Matrix:	Air
Analytical Method:	TO-15	Test:	TO-15
Sample Wt/Vol:	400 Units: mL		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VL043086.D	1		10/09/25 13:52	VL100925

CAS Number	Parameter	Conc. ppbv	Conc. ug/M3	Qualifier	MDL ug/m3	LOQ / CRQL ug/m3
124-48-1	Dibromochloromethane	0.090	0.77	U	0.77	4.26
106-93-4	1,2-Dibromoethane	0.090	0.69	U	0.69	0.77
127-18-4	Tetrachloroethene	0.020	0.14	U	0.14	0.20
108-90-7	Chlorobenzene	0.080	0.37	U	0.37	2.30
100-41-4	Ethyl Benzene	0.19	0.83	U	0.83	2.17
179601-23-1	m/p-Xylene	0.41	1.78	U	1.78	4.34
95-47-6	o-Xylene	0.21	0.91	U	0.91	2.17
100-42-5	Styrene	0.16	0.68	U	0.68	2.13
75-25-2	Bromoform	0.050	0.52	U	0.52	5.17
79-34-5	1,1,2,2-Tetrachloroethane	0.020	0.14	U	0.14	0.21
95-49-8	2-Chlorotoluene	0.17	0.88	U	0.88	2.59
108-67-8	1,3,5-Trimethylbenzene	0.18	0.88	U	0.88	2.46
95-63-6	1,2,4-Trimethylbenzene	0.18	0.88	U	0.88	2.46
541-73-1	1,3-Dichlorobenzene	0.050	0.30	U	0.30	3.01
106-46-7	1,4-Dichlorobenzene	0.13	0.78	U	0.78	3.01
95-50-1	1,2-Dichlorobenzene	0.13	0.78	U	0.78	3.01
120-82-1	1,2,4-Trichlorobenzene	0.21	1.56	U	1.56	3.71
87-68-3	Hexachloro-1,3-Butadiene	0.13	1.39	U	1.39	5.33
106-99-0	1,3-Butadiene	0.050	0.11	U	0.11	1.11
91-20-3	Naphthalene	0.25	1.31		0.050	0.52
622-96-8	4-Ethyltoluene	0.21	1.03	U	1.03	2.46
110-54-3	Hexane	0.28	0.99	J	0.56	1.76
107-05-1	Allyl Chloride	0.11	0.34	U	0.34	1.57
123-91-1	1,4-Dioxane	0.22	0.79	U	0.79	1.80
80-62-6	Methyl Methacrylate	0.14	0.57	U	0.57	2.05
SURROGATES						
460-00-4	1-Bromo-4-Fluorobenzene	10.8			65 - 135	108% SPK: 10
INTERNAL STANDARDS						
74-97-5	Bromochloromethane	194000		2.777		
540-36-3	1,4-Difluorobenzene	514000		3.943		
3114-55-4	Chlorobenzene-d5	434000		8.866		

Report of Analysis

Client:	GFE LLC	Date Collected:	10/08/25
Project:	200 Reservoir St, Trenton NJ	Date Received:	10/08/25
Client Sample ID:	IA6	SDG No.:	Q3320
Lab Sample ID:	Q3320-11	Matrix:	Air
Analytical Method:	TO-15	Test:	TO-15
Sample Wt/Vol:	400 Units: mL		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VL043086.D	1		10/09/25 13:52	VL100925

CAS Number	Parameter	Conc. ppbv	Conc. ug/M3	Qualifier	MDL	LOQ / CRQL
------------	-----------	---------------	----------------	-----------	-----	------------

U = Not Detected

RL = Reporting Limit

MDL = Method Detection Limit

E = Value Exceeds Calibration Range

D = Dilution

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

Q = indicates LCS control criteria did not meet requirements

Report of Analysis

Client:	GFE LLC	Date Collected:	10/08/25
Project:	200 Reservoir St, Trenton NJ	Date Received:	10/08/25
Client Sample ID:	SV6	SDG No.:	Q3320
Lab Sample ID:	Q3320-12	Matrix:	Air
Analytical Method:	TO-15	Test:	TO-15
Sample Wt/Vol:	400	Units:	mL

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VL043097.D	5		10/09/25 20:39	VL100925

CAS Number	Parameter	Conc. ppbv	Conc. ug/M3	Qualifier	MDL ug/m3	LOQ / CRQL ug/m3
TARGETS						
75-71-8	Dichlorodifluoromethane	0.55	2.72	U	2.72	12.4
74-87-3	Chloromethane	0.49	1.01	U	1.01	5.16
75-01-4	Vinyl Chloride	0.13	0.33	U	0.33	0.38
74-83-9	Bromomethane	0.39	1.51	U	1.51	9.71
75-00-3	Chloroethane	0.75	1.98	U	1.98	6.60
109-99-9	Tetrahydrofuran	0.40	1.18	U	1.18	7.37
75-69-4	Trichlorofluoromethane	0.85	4.78	U	4.78	14.1
76-13-1	1,1,2-Trichlorotrifluoroethane	0.70	5.37	U	5.37	19.2
76-14-2	Dichlorotetrafluoroethane	0.75	5.24	U	5.24	17.5
75-65-0	tert-Butyl alcohol	3.20	9.70		2.43	7.58
142-82-5	Heptane	0.85	3.48	U	3.48	10.3
75-35-4	1,1-Dichloroethene	0.75	2.97	U	2.97	9.91
67-64-1	Acetone	81.7	194	E	2.14	5.94
75-15-0	Carbon Disulfide	0.40	1.25	U	1.25	7.79
1634-04-4	Methyl tert-Butyl Ether	1.20	4.33	U	4.33	9.01
75-09-2	Methylene Chloride	1.20	4.17	J	4.17	8.69
156-60-5	trans-1,2-Dichloroethene	0.60	2.38	U	2.38	9.91
75-34-3	1,1-Dichloroethane	0.65	2.63	U	2.63	10.1
110-82-7	Cyclohexane	1.10	3.79	U	3.79	8.61
78-93-3	2-Butanone	0.94	2.77	J	0.83	7.37
56-23-5	Carbon Tetrachloride	0.13	0.82	U	0.82	0.94
156-59-2	cis-1,2-Dichloroethene	0.50	1.98	U	1.98	9.91
67-66-3	Chloroform	0.48	2.34	U	2.34	12.2
71-55-6	1,1,1-Trichloroethane	0.080	0.44	U	0.44	0.82
540-84-1	2,2,4-Trimethylpentane	0.70	3.27	U	3.27	11.7
71-43-2	Benzene	0.53	1.69	J	1.28	7.99
107-06-2	1,2-Dichloroethane	0.47	1.90	U	1.90	10.1
79-01-6	Trichloroethene	1.70	9.14		0.64	0.81
78-87-5	1,2-Dichloropropane	0.65	3.00	U	3.00	11.6
75-27-4	Bromodichloromethane	0.32	2.14	U	2.14	16.8
108-10-1	4-Methyl-2-Pentanone	0.49	2.01	U	2.01	10.3
108-88-3	Toluene	1.70	6.41	J	3.01	9.42
10061-02-6	t-1,3-Dichloropropene	0.75	3.40	U	3.40	11.3
10061-01-5	cis-1,3-Dichloropropene	0.55	2.50	U	2.50	11.3
79-00-5	1,1,2-Trichloroethane	0.41	2.24	U	2.24	13.6

Report of Analysis

Client:	GFE LLC	Date Collected:	10/08/25
Project:	200 Reservoir St, Trenton NJ	Date Received:	10/08/25
Client Sample ID:	SV6	SDG No.:	Q3320
Lab Sample ID:	Q3320-12	Matrix:	Air
Analytical Method:	TO-15	Test:	TO-15
Sample Wt/Vol:	400 Units: mL		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VL043097.D	5		10/09/25 20:39	VL100925

CAS Number	Parameter	Conc. ppbv	Conc. ug/M3	Qualifier	MDL ug/m3	LOQ / CRQL ug/m3
124-48-1	Dibromochloromethane	0.43	3.66	U	3.66	21.3
106-93-4	1,2-Dibromoethane	0.43	3.30	U	3.30	3.84
127-18-4	Tetrachloroethene	8.40	57.0		0.54	1.02
108-90-7	Chlorobenzene	0.39	1.80	U	1.80	11.5
100-41-4	Ethyl Benzene	10.1	43.9		4.13	10.9
179601-23-1	m/p-Xylene	20.2	87.7		9.12	21.7
95-47-6	o-Xylene	3.80	16.5		4.78	10.9
100-42-5	Styrene	0.80	3.41	U	3.41	10.6
75-25-2	Bromoform	0.24	2.48	U	2.48	25.9
79-34-5	1,1,2,2-Tetrachloroethane	0.090	0.62	U	0.62	1.03
95-49-8	2-Chlorotoluene	0.85	4.40	U	4.40	12.9
108-67-8	1,3,5-Trimethylbenzene	0.90	4.42	U	4.42	12.3
95-63-6	1,2,4-Trimethylbenzene	0.90	4.42	U	4.42	12.3
541-73-1	1,3-Dichlorobenzene	0.27	1.62	U	1.62	15.0
106-46-7	1,4-Dichlorobenzene	0.65	3.91	U	3.91	15.0
95-50-1	1,2-Dichlorobenzene	0.65	3.91	U	3.91	15.0
120-82-1	1,2,4-Trichlorobenzene	1.10	8.17	U	8.17	18.6
87-68-3	Hexachloro-1,3-Butadiene	0.65	6.93	U	6.93	26.7
106-99-0	1,3-Butadiene	0.26	0.58	U	0.58	5.53
91-20-3	Naphthalene	0.070	0.37	U	0.37	2.62
622-96-8	4-Ethyltoluene	1.10	5.41	U	5.41	12.3
110-54-3	Hexane	5.70	20.1		2.82	8.81
107-05-1	Allyl Chloride	0.55	1.72	U	1.72	7.83
123-91-1	1,4-Dioxane	1.10	3.96	U	3.96	9.01
80-62-6	Methyl Methacrylate	0.70	2.87	U	2.87	10.2
SURROGATES						
460-00-4	1-Bromo-4-Fluorobenzene	11.8			65 - 135	118% SPK: 10
INTERNAL STANDARDS						
74-97-5	Bromochloromethane	179000		2.777		
540-36-3	1,4-Difluorobenzene	517000		3.939		
3114-55-4	Chlorobenzene-d5	480000		8.862		

Report of Analysis

Client:	GFE LLC	Date Collected:	10/08/25
Project:	200 Reservoir St, Trenton NJ	Date Received:	10/08/25
Client Sample ID:	SV6	SDG No.:	Q3320
Lab Sample ID:	Q3320-12	Matrix:	Air
Analytical Method:	TO-15	Test:	TO-15
Sample Wt/Vol:	400 Units: mL		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VL043097.D	5		10/09/25 20:39	VL100925

CAS Number	Parameter	Conc. ppbv	Conc. ug/M3	Qualifier	MDL	LOQ / CRQL
------------	-----------	---------------	----------------	-----------	-----	------------

U = Not Detected

RL = Reporting Limit

MDL = Method Detection Limit

E = Value Exceeds Calibration Range

D = Dilution

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

Q = indicates LCS control criteria did not meet requirements

Report of Analysis

Client:	GFE LLC	Date Collected:	10/08/25
Project:	200 Reservoir St, Trenton NJ	Date Received:	10/08/25
Client Sample ID:	SV6DL	SDG No.:	Q3320
Lab Sample ID:	Q3320-12DL	Matrix:	Air
Analytical Method:	TO-15	Test:	TO-15
Sample Wt/Vol:	400	Units:	mL

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VL043092.D	10		10/09/25 17:17	VL100925

CAS Number	Parameter	Conc. ppbv	Conc. ug/M3	Qualifier	MDL ug/m3	LOQ / CRQL ug/m3
TARGETS						
75-71-8	Dichlorodifluoromethane	1.10	5.44	UD	5.44	24.7
74-87-3	Chloromethane	0.98	2.02	UD	2.02	10.3
75-01-4	Vinyl Chloride	0.25	0.64	UD	0.64	0.77
74-83-9	Bromomethane	0.77	2.99	UD	2.99	19.4
75-00-3	Chloroethane	1.50	3.96	UD	3.96	13.2
109-99-9	Tetrahydrofuran	0.80	2.36	UD	2.36	14.8
75-69-4	Trichlorofluoromethane	1.70	9.55	UD	9.55	28.1
76-13-1	1,1,2-Trichlorotrifluoroethane	1.40	10.7	UD	10.7	38.3
76-14-2	Dichlorotetrafluoroethane	1.50	10.5	UD	10.5	35.0
75-65-0	tert-Butyl alcohol	3.20	9.70	JD	4.85	15.2
142-82-5	Heptane	1.70	6.97	UD	6.97	20.5
75-35-4	1,1-Dichloroethene	1.50	5.95	UD	5.95	19.8
67-64-1	Acetone	78.8	187	D	4.28	11.9
75-15-0	Carbon Disulfide	0.80	2.49	UD	2.49	15.6
1634-04-4	Methyl tert-Butyl Ether	2.30	8.29	UD	8.29	18.0
75-09-2	Methylene Chloride	2.30	7.99	UD	7.99	17.4
156-60-5	trans-1,2-Dichloroethene	1.20	4.76	UD	4.76	19.8
75-34-3	1,1-Dichloroethane	1.30	5.26	UD	5.26	20.2
110-82-7	Cyclohexane	2.20	7.57	UD	7.57	17.2
78-93-3	2-Butanone	0.56	1.65	UD	1.65	14.8
56-23-5	Carbon Tetrachloride	0.25	1.57	UD	1.57	1.89
156-59-2	cis-1,2-Dichloroethene	0.99	3.93	UD	3.93	19.8
67-66-3	Chloroform	0.96	4.69	UD	4.69	24.4
71-55-6	1,1,1-Trichloroethane	0.16	0.87	UD	0.87	1.64
540-84-1	2,2,4-Trimethylpentane	1.40	6.54	UD	6.54	23.4
71-43-2	Benzene	0.79	2.52	UD	2.52	16.0
107-06-2	1,2-Dichloroethane	0.93	3.76	UD	3.76	20.2
79-01-6	Trichloroethene	1.80	9.67	D	1.29	1.61
78-87-5	1,2-Dichloropropane	1.30	6.01	UD	6.01	23.1
75-27-4	Bromodichloromethane	0.63	4.22	UD	4.22	33.5
108-10-1	4-Methyl-2-Pentanone	0.97	3.98	UD	3.98	20.5
108-88-3	Toluene	1.60	6.03	JD	6.03	18.8
10061-02-6	t-1,3-Dichloropropene	1.50	6.81	UD	6.81	22.7
10061-01-5	cis-1,3-Dichloropropene	1.10	4.99	UD	4.99	22.7
79-00-5	1,1,2-Trichloroethane	0.81	4.42	UD	4.42	27.3

Report of Analysis

Client:	GFE LLC	Date Collected:	10/08/25
Project:	200 Reservoir St, Trenton NJ	Date Received:	10/08/25
Client Sample ID:	SV6DL	SDG No.:	Q3320
Lab Sample ID:	Q3320-12DL	Matrix:	Air
Analytical Method:	TO-15	Test:	TO-15
Sample Wt/Vol:	400 Units: mL		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VL043092.D	10		10/09/25 17:17	VL100925

CAS Number	Parameter	Conc. ppbv	Conc. ug/M3	Qualifier	MDL ug/m3	LOQ / CRQL ug/m3
124-48-1	Dibromochloromethane	0.85	7.24	UD	7.24	42.6
106-93-4	1,2-Dibromoethane	0.85	6.53	UD	6.53	7.69
127-18-4	Tetrachloroethene	7.80	52.9	D	1.02	2.03
108-90-7	Chlorobenzene	0.77	3.55	UD	3.55	23.0
100-41-4	Ethyl Benzene	7.80	33.9	D	8.25	21.7
179601-23-1	m/p-Xylene	15.6	67.8	D	17.8	43.4
95-47-6	o-Xylene	3.20	13.9	JD	9.12	21.7
100-42-5	Styrene	1.60	6.81	UD	6.81	21.3
75-25-2	Bromoform	0.48	4.96	UD	4.96	51.7
79-34-5	1,1,2,2-Tetrachloroethane	0.18	1.24	UD	1.24	2.06
95-49-8	2-Chlorotoluene	1.70	8.80	UD	8.80	25.9
108-67-8	1,3,5-Trimethylbenzene	1.80	8.85	UD	8.85	24.6
95-63-6	1,2,4-Trimethylbenzene	1.80	8.85	UD	8.85	24.6
541-73-1	1,3-Dichlorobenzene	0.54	3.25	UD	3.25	30.1
106-46-7	1,4-Dichlorobenzene	1.30	7.82	UD	7.82	30.1
95-50-1	1,2-Dichlorobenzene	1.30	7.82	UD	7.82	30.1
120-82-1	1,2,4-Trichlorobenzene	2.10	15.6	UD	15.6	37.1
87-68-3	Hexachloro-1,3-Butadiene	1.30	13.9	UD	13.9	53.3
106-99-0	1,3-Butadiene	0.51	1.13	UD	1.13	11.1
91-20-3	Naphthalene	0.13	0.68	UD	0.68	5.24
622-96-8	4-Ethyltoluene	2.10	10.3	UD	10.3	24.6
110-54-3	Hexane	5.00	17.6	D	5.64	17.6
107-05-1	Allyl Chloride	1.10	3.44	UD	3.44	15.7
123-91-1	1,4-Dioxane	2.20	7.93	UD	7.93	18.0
80-62-6	Methyl Methacrylate	1.40	5.73	UD	5.73	20.5
SURROGATES						
460-00-4	1-Bromo-4-Fluorobenzene	11.6			65 - 135	116% SPK: 10
INTERNAL STANDARDS						
74-97-5	Bromochloromethane	192000		2.774		
540-36-3	1,4-Difluorobenzene	513000		3.936		
3114-55-4	Chlorobenzene-d5	466000		8.862		

Report of Analysis

Client:	GFE LLC	Date Collected:	10/08/25
Project:	200 Reservoir St, Trenton NJ	Date Received:	10/08/25
Client Sample ID:	SV6DL	SDG No.:	Q3320
Lab Sample ID:	Q3320-12DL	Matrix:	Air
Analytical Method:	TO-15	Test:	TO-15
Sample Wt/Vol:	400 Units: mL		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VL043092.D	10		10/09/25 17:17	VL100925

CAS Number	Parameter	Conc. ppbv	Conc. ug/M3	Qualifier	MDL	LOQ / CRQL
------------	-----------	---------------	----------------	-----------	-----	------------

U = Not Detected

RL = Reporting Limit

MDL = Method Detection Limit

E = Value Exceeds Calibration Range

D = Dilution

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

Q = indicates LCS control criteria did not meet requirements

Report of Analysis

Client:	GFE LLC	Date Collected:	10/08/25
Project:	200 Reservoir St, Trenton NJ	Date Received:	10/08/25
Client Sample ID:	OA1	SDG No.:	Q3320
Lab Sample ID:	Q3320-13	Matrix:	Air
Analytical Method:	TO-15	Test:	TO-15
Sample Wt/Vol:	400	Units:	mL

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VL043070.D	1		10/08/25 22:10	VL100825

CAS Number	Parameter	Conc. ppbv	Conc. ug/M3	Qualifier	MDL ug/m3	LOQ / CRQL ug/m3
TARGETS						
75-71-8	Dichlorodifluoromethane	0.47	2.32	J	0.54	2.47
74-87-3	Chloromethane	0.44	0.91	J	0.21	1.03
75-01-4	Vinyl Chloride	0.030	0.080	U	0.080	0.080
74-83-9	Bromomethane	0.080	0.31	U	0.31	1.94
75-00-3	Chloroethane	0.15	0.40	U	0.40	1.32
109-99-9	Tetrahydrofuran	0.080	0.24	U	0.24	1.47
75-69-4	Trichlorofluoromethane	0.23	1.29	J	0.96	2.81
76-13-1	1,1,2-Trichlorotrifluoroethane	0.14	1.07	U	1.07	3.83
76-14-2	Dichlorotetrafluoroethane	0.15	1.05	U	1.05	3.49
75-65-0	tert-Butyl alcohol	0.16	0.49	U	0.49	1.52
142-82-5	Heptane	0.17	0.70	U	0.70	2.05
75-35-4	1,1-Dichloroethene	0.15	0.59	U	0.59	1.98
67-64-1	Acetone	3.40	8.08		0.43	1.19
75-15-0	Carbon Disulfide	0.080	0.25	U	0.25	1.56
1634-04-4	Methyl tert-Butyl Ether	0.23	0.83	U	0.83	1.80
75-09-2	Methylene Chloride	0.70	2.43		0.80	1.74
156-60-5	trans-1,2-Dichloroethene	0.12	0.48	U	0.48	1.98
75-34-3	1,1-Dichloroethane	0.13	0.53	U	0.53	2.02
110-82-7	Cyclohexane	0.22	0.76	U	0.76	1.72
78-93-3	2-Butanone	0.14	0.41	J	0.18	1.47
56-23-5	Carbon Tetrachloride	0.080	0.50		0.19	0.19
156-59-2	cis-1,2-Dichloroethene	0.10	0.40	U	0.40	1.98
67-66-3	Chloroform	0.10	0.49	U	0.49	2.44
71-55-6	1,1,1-Trichloroethane	0.020	0.11	U	0.11	0.16
540-84-1	2,2,4-Trimethylpentane	0.14	0.65	U	0.65	2.34
71-43-2	Benzene	0.080	0.26	U	0.26	1.60
107-06-2	1,2-Dichloroethane	0.090	0.36	U	0.36	2.02
79-01-6	Trichloroethene	0.020	0.11	U	0.11	0.16
78-87-5	1,2-Dichloropropane	0.13	0.60	U	0.60	2.31
75-27-4	Bromodichloromethane	0.060	0.40	U	0.40	3.35
108-10-1	4-Methyl-2-Pentanone	0.10	0.41	U	0.41	2.05
108-88-3	Toluene	0.16	0.60	U	0.60	1.88
10061-02-6	t-1,3-Dichloropropene	0.15	0.68	U	0.68	2.27
10061-01-5	cis-1,3-Dichloropropene	0.11	0.50	U	0.50	2.27
79-00-5	1,1,2-Trichloroethane	0.080	0.44	U	0.44	2.73

Report of Analysis

Client:	GFE LLC	Date Collected:	10/08/25
Project:	200 Reservoir St, Trenton NJ	Date Received:	10/08/25
Client Sample ID:	OA1	SDG No.:	Q3320
Lab Sample ID:	Q3320-13	Matrix:	Air
Analytical Method:	TO-15	Test:	TO-15
Sample Wt/Vol:	400 Units: mL		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VL043070.D	1		10/08/25 22:10	VL100825

CAS Number	Parameter	Conc. ppbv	Conc. ug/M3	Qualifier	MDL ug/m3	LOQ / CRQL ug/m3
124-48-1	Dibromochloromethane	0.090	0.77	U	0.77	4.26
106-93-4	1,2-Dibromoethane	0.090	0.69	U	0.69	0.77
127-18-4	Tetrachloroethene	0.020	0.14	U	0.14	0.20
108-90-7	Chlorobenzene	0.080	0.37	U	0.37	2.30
100-41-4	Ethyl Benzene	0.19	0.83	U	0.83	2.17
179601-23-1	m/p-Xylene	0.41	1.78	U	1.78	4.34
95-47-6	o-Xylene	0.21	0.91	U	0.91	2.17
100-42-5	Styrene	0.16	0.68	U	0.68	2.13
75-25-2	Bromoform	0.050	0.52	U	0.52	5.17
79-34-5	1,1,2,2-Tetrachloroethane	0.020	0.14	U	0.14	0.21
95-49-8	2-Chlorotoluene	0.17	0.88	U	0.88	2.59
108-67-8	1,3,5-Trimethylbenzene	0.18	0.88	U	0.88	2.46
95-63-6	1,2,4-Trimethylbenzene	0.18	0.88	U	0.88	2.46
541-73-1	1,3-Dichlorobenzene	0.050	0.30	U	0.30	3.01
106-46-7	1,4-Dichlorobenzene	0.13	0.78	U	0.78	3.01
95-50-1	1,2-Dichlorobenzene	0.13	0.78	U	0.78	3.01
120-82-1	1,2,4-Trichlorobenzene	0.21	1.56	U	1.56	3.71
87-68-3	Hexachloro-1,3-Butadiene	0.13	1.39	U	1.39	5.33
106-99-0	1,3-Butadiene	0.050	0.11	U	0.11	1.11
91-20-3	Naphthalene	0.010	0.050	U	0.050	0.52
622-96-8	4-Ethyltoluene	0.21	1.03	U	1.03	2.46
110-54-3	Hexane	0.34	1.20	J	0.56	1.76
107-05-1	Allyl Chloride	0.11	0.34	U	0.34	1.57
123-91-1	1,4-Dioxane	0.22	0.79	U	0.79	1.80
80-62-6	Methyl Methacrylate	0.14	0.57	U	0.57	2.05
SURROGATES						
460-00-4	1-Bromo-4-Fluorobenzene	9.80			65 - 135	98% SPK: 10
INTERNAL STANDARDS						
74-97-5	Bromochloromethane	172000		2.784		
540-36-3	1,4-Difluorobenzene	451000		3.955		
3114-55-4	Chlorobenzene-d5	353000		8.878		

Report of Analysis

Client:	GFE LLC	Date Collected:	10/08/25
Project:	200 Reservoir St, Trenton NJ	Date Received:	10/08/25
Client Sample ID:	OA1	SDG No.:	Q3320
Lab Sample ID:	Q3320-13	Matrix:	Air
Analytical Method:	TO-15	Test:	TO-15
Sample Wt/Vol:	400 Units: mL		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VL043070.D	1		10/08/25 22:10	VL100825

CAS Number	Parameter	Conc. ppbv	Conc. ug/M3	Qualifier	MDL	LOQ / CRQL
------------	-----------	---------------	----------------	-----------	-----	------------

U = Not Detected

RL = Reporting Limit

MDL = Method Detection Limit

E = Value Exceeds Calibration Range

D = Dilution

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

Q = indicates LCS control criteria did not meet requirements

Report of Analysis

Client:	GFE LLC	Date Collected:	
Project:	200 Reservoir St, Trenton NJ	Date Received:	
Client Sample ID:	IA1DUP	SDG No.:	Q3320
Lab Sample ID:	Q3320-01DUP	Matrix:	Air
Analytical Method:	TO-15	Test:	TO-15
Sample Wt/Vol:	400 Units: mL		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VL043084.D	1		10/09/25 12:44	VL100925

CAS Number	Parameter	Conc. ppbv	Conc. ug/M3	Qualifier	MDL ug/m3	LOQ / CRQL ug/m3
TARGETS						
75-71-8	Dichlorodifluoromethane	0.42	2.08	J	0.54	2.47
74-87-3	Chloromethane	0.65	1.34		0.21	1.03
75-01-4	Vinyl Chloride	0.030	0.080	U	0.080	0.080
74-83-9	Bromomethane	0.080	0.31	U	0.31	1.94
75-00-3	Chloroethane	0.15	0.40	U	0.40	1.32
109-99-9	Tetrahydrofuran	0.080	0.24	U	0.24	1.47
75-69-4	Trichlorofluoromethane	0.21	1.18	J	0.96	2.81
76-13-1	1,1,2-Trichlorotrifluoroethane	0.14	1.07	U	1.07	3.83
76-14-2	Dichlorotetrafluoroethane	0.15	1.05	U	1.05	3.49
75-65-0	tert-Butyl alcohol	0.16	0.49	U	0.49	1.52
142-82-5	Heptane	0.17	0.70	U	0.70	2.05
75-35-4	1,1-Dichloroethene	0.15	0.59	U	0.59	1.98
67-64-1	Acetone	6.70	15.9		0.43	1.19
75-15-0	Carbon Disulfide	0.080	0.25	U	0.25	1.56
1634-04-4	Methyl tert-Butyl Ether	0.23	0.83	U	0.83	1.80
75-09-2	Methylene Chloride	0.88	3.06		0.80	1.74
156-60-5	trans-1,2-Dichloroethene	0.12	0.48	U	0.48	1.98
75-34-3	1,1-Dichloroethane	0.13	0.53	U	0.53	2.02
110-82-7	Cyclohexane	0.22	0.76	U	0.76	1.72
78-93-3	2-Butanone	0.64	1.89		0.18	1.47
56-23-5	Carbon Tetrachloride	0.090	0.57		0.19	0.19
156-59-2	cis-1,2-Dichloroethene	0.10	0.40	U	0.40	1.98
67-66-3	Chloroform	0.20	0.98	J	0.49	2.44
71-55-6	1,1,1-Trichloroethane	0.020	0.11	U	0.11	0.16
540-84-1	2,2,4-Trimethylpentane	0.14	0.65	U	0.65	2.34
71-43-2	Benzene	0.10	0.32	J	0.26	1.60
107-06-2	1,2-Dichloroethane	0.090	0.36	U	0.36	2.02
79-01-6	Trichloroethene	0.070	0.38		0.11	0.16
78-87-5	1,2-Dichloropropane	0.13	0.60	U	0.60	2.31
75-27-4	Bromodichloromethane	0.060	0.40	U	0.40	3.35
108-10-1	4-Methyl-2-Pentanone	0.22	0.90	J	0.41	2.05
108-88-3	Toluene	0.16	0.60	U	0.60	1.88
10061-02-6	t-1,3-Dichloropropene	0.15	0.68	U	0.68	2.27
10061-01-5	cis-1,3-Dichloropropene	0.11	0.50	U	0.50	2.27
79-00-5	1,1,2-Trichloroethane	0.080	0.44	U	0.44	2.73

Report of Analysis

Client:	GFE LLC	Date Collected:	
Project:	200 Reservoir St, Trenton NJ	Date Received:	
Client Sample ID:	IA1DUP	SDG No.:	Q3320
Lab Sample ID:	Q3320-01DUP	Matrix:	Air
Analytical Method:	TO-15	Test:	TO-15
Sample Wt/Vol:	400	Units:	mL

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VL043084.D	1		10/09/25 12:44	VL100925

CAS Number	Parameter	Conc. ppbv	Conc. ug/M3	Qualifier	MDL ug/m3	LOQ / CRQL ug/m3
124-48-1	Dibromochloromethane	0.090	0.77	U	0.77	4.26
106-93-4	1,2-Dibromoethane	0.090	0.69	U	0.69	0.77
127-18-4	Tetrachloroethene	0.050	0.34		0.14	0.20
108-90-7	Chlorobenzene	0.080	0.37	U	0.37	2.30
100-41-4	Ethyl Benzene	0.19	0.83	U	0.83	2.17
179601-23-1	m/p-Xylene	0.41	1.78	U	1.78	4.34
95-47-6	o-Xylene	0.21	0.91	U	0.91	2.17
100-42-5	Styrene	0.16	0.68	U	0.68	2.13
75-25-2	Bromoform	0.050	0.52	U	0.52	5.17
79-34-5	1,1,2,2-Tetrachloroethane	0.050	0.34		0.14	0.21
95-49-8	2-Chlorotoluene	0.17	0.88	U	0.88	2.59
108-67-8	1,3,5-Trimethylbenzene	0.18	0.88	U	0.88	2.46
95-63-6	1,2,4-Trimethylbenzene	0.18	0.88	U	0.88	2.46
541-73-1	1,3-Dichlorobenzene	0.050	0.30	U	0.30	3.01
106-46-7	1,4-Dichlorobenzene	0.13	0.78	U	0.78	3.01
95-50-1	1,2-Dichlorobenzene	0.13	0.78	U	0.78	3.01
120-82-1	1,2,4-Trichlorobenzene	0.21	1.56	U	1.56	3.71
87-68-3	Hexachloro-1,3-Butadiene	0.13	1.39	U	1.39	5.33
106-99-0	1,3-Butadiene	0.050	0.11	U	0.11	1.11
91-20-3	Naphthalene	0.010	0.050	U	0.050	0.52
622-96-8	4-Ethyltoluene	0.21	1.03	U	1.03	2.46
110-54-3	Hexane	0.49	1.73	J	0.56	1.76
107-05-1	Allyl Chloride	0.11	0.34	U	0.34	1.57
123-91-1	1,4-Dioxane	0.22	0.79	U	0.79	1.80
80-62-6	Methyl Methacrylate	0.14	0.57	U	0.57	2.05
SURROGATES						
460-00-4	1-Bromo-4-Fluorobenzene	10.7			65 - 135	107% SPK: 10
INTERNAL STANDARDS						
74-97-5	Bromochloromethane	196000		2.777		
540-36-3	1,4-Difluorobenzene	518000		3.943		
3114-55-4	Chlorobenzene-d5	435000		8.866		

Report of Analysis

Client:	GFE LLC	Date Collected:	
Project:	200 Reservoir St, Trenton NJ	Date Received:	
Client Sample ID:	IA1DUP	SDG No.:	Q3320
Lab Sample ID:	Q3320-01DUP	Matrix:	Air
Analytical Method:	TO-15	Test:	TO-15
Sample Wt/Vol:	400 Units: mL		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VL043084.D	1		10/09/25 12:44	VL100925

CAS Number	Parameter	Conc. ppbv	Conc. ug/M3	Qualifier	MDL	LOQ / CRQL
------------	-----------	---------------	----------------	-----------	-----	------------

U = Not Detected

RL = Reporting Limit

MDL = Method Detection Limit

E = Value Exceeds Calibration Range

D = Dilution

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

Q = indicates LCS control criteria did not meet requirements

LAB CHRONICLE

OrderID:	Q3320	OrderDate:	10/8/2025 3:14:00 PM
Client:	GFE LLC	Project:	200 Reservoir St, Trenton NJ
Contact:	Frank Galdun	Location:	D31,VOA Lab


LabID	ClientID	Matrix	Test	Method	Sample Date	Prep Date	Anal Date	Received
Q3320-01	IA1	Air	TO-15	TO-15	10/08/25		10/09/25	10/08/25
Q3320-02	SV1	Air	TO-15	TO-15	10/08/25		10/09/25	10/08/25
Q3320-02DL	SV1DL	Air	TO-15	TO-15	10/08/25		10/10/25	10/08/25
Q3320-03	IA2	Air	TO-15	TO-15	10/08/25		10/09/25	10/08/25
Q3320-04	SV2	Air	TO-15	TO-15	10/08/25		10/09/25	10/08/25
Q3320-04DL	SV2DL	Air	TO-15	TO-15	10/08/25		10/09/25	10/08/25
Q3320-05	IA3	Air	TO-15	TO-15	10/08/25		10/09/25	10/08/25
Q3320-06	SV3	Air	TO-15	TO-15	10/08/25		10/09/25	10/08/25
Q3320-07	IA4	Air	TO-15	TO-15	10/08/25		10/09/25	10/08/25
Q3320-08	SV4	Air	TO-15	TO-15	10/08/25		10/09/25	10/08/25
Q3320-08DL	SV4DL	Air	TO-15	TO-15	10/08/25		10/09/25	10/08/25
Q3320-09	IA5	Air			10/08/25			10/08/25

LAB CHRONICLE

			TO-15	TO-15		10/08/25	
Q3320-10	SV5	Air			10/08/25		10/08/25
			TO-15	TO-15		10/09/25	
Q3320-11	IA6	Air			10/08/25		10/08/25
			TO-15	TO-15		10/09/25	
Q3320-12	SV6	Air			10/08/25		10/08/25
			TO-15	TO-15		10/09/25	
Q3320-12DL	SV6DL	Air			10/08/25		10/08/25
			TO-15	TO-15		10/09/25	
Q3320-13	OA1	Air			10/08/25		10/08/25
			TO-15	TO-15		10/08/25	



SHIPPING DOCUMENTS

Client Contact Information				Bottle Order ID : B2509079				Courier : FGALDUN				1 of 13 COCs																	
Client ID : GFEL01				Project ID : All Projects				Sampler Name(s) : FRANK GALDUN				Analysis		Matrix															
Customer Name : GFE LLC Address : 58 Nokomis Ave				Project Manager : Frank galdun				AIR ANALYSIS CHAIN-OF-CUSTODY Batch Certified																					
				Phone Number : 646-542-3465																									
				Fax Number : 973-334-1692																									
				Site Details ZOO RESERVOIR ST TRENTON NJ																									
City : Lake Hiawatha				Analysis Turnaround Time 7 DAYS				Data Package Type : RESULTS ONLY																					
State : NJ				Standard : 10 business days OR																									
Zip Code : 07034				Rush (Specify): 7 Days				EDD Type : PDF																					
Sample Identification	Sample Date(s)	Time Start (24 hr Clock)	Time Stop (24 hr Clock)	Can Vacuum in Field ("Hg) (Start)	Can Vacuum in Field ("Hg) (Stop)**	Interior Temp. (F) (Start)	Interior Temp. (F) (Stop)	Out going Can Pressure ("Hg)(Lab)	In coming Can Pressure ("Hg)(Lab)	Flow Reg. ID	Can ID		Flow Controller Readout (ml/min)	Can Cert ID	TO-15	Indoor/Ambient Air	Soil Gas												
IAI	10/8/25	8:36	10:36	OVER 30	7	70	70	-30	-4.3	10616	10053	6 L	50	VL042942.D	1	1													
Temperature (Fahrenheit) <table border="1"> <tr> <td></td> <td>Ambient</td> <td>Maximum</td> <td>Minimum</td> </tr> <tr> <td>Start</td> <td></td> <td></td> <td></td> </tr> <tr> <td>Stop</td> <td></td> <td></td> <td></td> </tr> </table>											Ambient	Maximum	Minimum	Start				Stop				GC/MS Analyst Signature (TO-15) 							
	Ambient	Maximum	Minimum																										
Start																													
Stop																													
Pressure (Inches of Hg) <table border="1"> <tr> <td></td> <td>Ambient</td> <td>Maximum</td> <td>Minimum</td> </tr> <tr> <td>Start</td> <td></td> <td></td> <td></td> </tr> <tr> <td>Stop</td> <td></td> <td></td> <td></td> </tr> </table>											Ambient	Maximum	Minimum	Start				Stop				** Submittal of this COC indicates approval of the analysis based on existing conditions. Please follow the instructions on the back of this COC.							
	Ambient	Maximum	Minimum																										
Start																													
Stop																													
Special Instructions/QC Requirements & Comments :																													
Suspected Contamination: High Medium Low PID Readings: 0.0																													
Sampling site (State):																													
Quick Connector required : NO																													
Canisters Shipped by: 				Date/Time: 10/07/25				Canisters Received by: 				Date/Time: 10/8/25 1340																	
Samples Relinquished by: 				Date/Time: 10/8/25				Received by:				Date/Time:																	
Relinquished by:				Date/Time:				Received by:				Date/Time:																	
B2509079 - 1																													

Client Contact Information				Bottle Order ID : B2509079				Courier : FGALDUN				2 of 13 COCs							
Client ID : GFEL01 Project ID : All Projects				Sampler Name(s) : FRANK GALDUN				Analysis				Matrix							
Customer Name : GFE LLC Address : 58 Nokomis Ave				Project Manager : Frank galdun				AIR ANALYSIS CHAIN-OF-CUSTODY Batch Certified											
				Phone Number : 646-542-3465															
				Fax Number : 973-334-1692															
City : Lake Hiawatha				Site Details ZOO RESERVOIR ST IRENTON NJ												Indoor/Ambinet Air Soil Gas			
State : NJ				Analysis Turnaround Time 7 DAYS															
Zip Code : 07034				Standard : 10 business days OR															
Country :				Rush (Specify): 7 Days				Data Package Type :				EDD Type :							
Sample Identification	Sample Date(s)	Time Start (24 hr Clock)	Time Stop (24 hr Clock)	Can Vacuum in Field ("Hg) (Start)	Can Vacuum in Field ("Hg) (Stop)**	Interior Temp. (F) (Start)	Interior Temp. (F) (Stop)	Out going Can Pressure ("Hg)(Lab)	In coming Can Pressure ("Hg)(Lab)	Flow Reg. ID	Can ID		Flow Controller Readout (ml/min)	Can Cert ID	TO-15				
SV1	10/8/25	8:59	10:39	30	2	70	70	-30	-5.9	10767	10332	6 L	50	VL042828.D	1				1
Temperature (Fahrenheit)										GC/MS Analyst Signature (TO-15) 									
		Ambient	Maximum	Minimum															
Start																			
Stop																			
Pressure (Inches of Hg)										** Submittal of this COC indicates approval of the analysis based on existing conditions. Please follow the instructions on the back of this COC.									
		Ambient	Maximum	Minimum															
Start																			
Stop																			
Special Instructions/QC Requirements & Comments :																			
Suspected Contamination: High Medium Low PID Readings: 2.0																			
Sampling site (State):																			
Quick Connector required : No																			
Canisters Shipped by: FRANK					Date/Time: 10/8/25					Canisters Received by: QR					Date/Time: 10/8/25 1340				
Samples Relinquished by: FRANK					Date/Time: 10/8/25					Received by:					Date/Time:				
Relinquished by:					Date/Time:					Received by:					Date/Time:				

B2509079 - 4

Client Contact Information				Bottle Order ID : B2509079				Courier : F GALDUN				3 of 13 COCs							
Client ID : GFEL01 Project ID : All Projects				Sampler Name(s) FRANK GALDUN				Analysis				Matrix							
Customer Name : GFE LLC Address : 58 Nokomis Ave City : Lake Hiawatha State : NJ Zip Code : 07034 Country :				Project Manager : Frank galdun				AIR ANALYSIS CHAIN-OF-CUSTODY Batch Certified											
				Phone Number : 646-542-3465															
				Fax Number : 973-334-1692															
				Site Details: 200 RESERVOIR ST TRENTON, NJ															
Analysis Turnaround Time 1 DAY				Standard : 10 Business days OR				Data Package Type : RESULTS ONLY											
Rush (Specify): 7 Days				EDD Type : PDE															
Sample Identification	Sample Date(s)	Time Start (24 hr Clock)	Time Stop (24 hr Clock)	Can Vacuum in Field ("Hg) (Start)	Can Vacuum in Field ("Hg) (Stop)**	Interior Temp. (F) (Start)	Interior Temp. (F) (Stop)	Out going Can Pressure ("Hg)(Lab)	In coming Can Pressure ("Hg)(Lab)	Flow Reg. ID	Can ID		Flow Controller Readout (ml/min)	Can Cert ID	TO-15	Indoor Ambient Air	Soil Gas		
IAZ	10/8/25	8:49	10:49	OVER 30	S	70	70	-30	-5.1	10109	10602	6 L	50	VL042942.D	1	1			
Temperature (Fahrenheit)										GC/MS Analyst Signature (TO-15) 									
		Ambient	Maximum	Minimum															
Start																			
Stop																			
Pressure (Inches of Hg)										** Submittal of this COC indicates approval of the analysis based on existing conditions. Please follow the instructions on the back of this COC.									
		Ambient	Maximum	Minimum															
Start																			
Stop																			
Special Instructions/QC Requirements & Comments :																			
Suspected Contamination: High Medium Low PID Readings: 0.0																			
Sampling site (State):																			
Quick Connector required : No																			
Canisters Shipped by: Sam				Date/Time: 10/6/25				Canisters Received by: cl				Date/Time: 10/8/25 1340				B2509079 - 10			
Samples Relinquished by: FRANK				Date/Time: 10/8/25				Received by:				Date/Time:							
Relinquished by:				Date/Time:				Received by:				Date/Time:							

Client Contact Information				Bottle Order ID : B2509079				Courier : F GALDUN				4 of 13 COCs					
Client ID : GFEL01				Project ID : All Projects				Sampler Name(s) FRANK GALDUN				Analysis		Matrix			
Customer Name : GFE LLC Address : 58 Nokomis Ave City : Lake Hiawatha State : NJ Zip Code : 07034 Country :				Project Manager : Frank galdun				AIR ANALYSIS CHAIN-OF-CUSTODY Batch Certified									
				Phone Number : 646-542-3465													
				Fax Number : 973-334-1692													
				Site Details: 200 RESERVOIR ST TRENTON, NJ													
Analysis Turnaround Time 7 DAY				Standard : 10 business days OR				Data Package Type : RESULTS ONLY									
Rush (Specify): 7 Days				EDD Type : PDF													
Sample Identification	Sample Date(s)	Time Start (24 hr Clock)	Time Stop (24 hr Clock)	Can Vacuum in Field ("Hg) (Start)	Can Vacuum in Field ("Hg) (Stop)**	Interior Temp. (F) (Start)	Interior Temp. (F) (Stop)	Out going Can Pressure ("Hg)(Lab)	In coming Can Pressure ("Hg)(Lab)	Flow Reg. ID	Can ID		Flow Controller Readout (ml/min)	Can Cert ID	TO-15	Indoor/Ambient Air	Soil Gas
SVZ	10/8/25	8:54	10:54	30	30	70	70	-30	-5.1	10185	10445	6 L	50	VL042828.D	1		
8:54 10:54 over 20 Temperature (Fahrenheit)										GC/MS Analyst Signature (TO-15) 							
		Ambient	Maximum	Minimum													
Start																	
Stop																	
Pressure (Inches of Hg)										** Submittal of this COC indicates approval of the analysis based on existing conditions. Please follow the instructions on the back of this COC.							
		Ambient	Maximum	Minimum													
Start																	
Stop																	
Special Instructions/QC Requirements & Comments :																	
Suspected Contamination: High Medium Low PID Readings: 0.0																	
Sampling site (State):																	
Quick Connector required : No																	
Canisters Shipped by: Seam				Date/Time: 10/8/25				Canisters Received by: CR				Date/Time: 10/8/25 1340					
Samples Relinquished by: SM				Date/Time: 10/8/25				Received by:				Date/Time:					
Relinquished by:				Date/Time:				Received by:				Date/Time:					

B2509079 - 7

Client Contact Information				Bottle Order ID : B2509079				Courier : FGALDUN				5 of 13 COCs																	
Client ID : GFEL01				Project ID : All Projects				Sampler Name(s) : FRANK GALDUN				Analysis		Matrix															
Customer Name : GFE LLC Address : 58 Nokomis Ave				Project Manager : Frank galdun				AIR ANALYSIS CHAIN-OF-CUSTODY Batch Certified																					
				Phone Number : 646-542-3465																									
				Fax Number : 973-334-1692																									
				Site Details: 200 RESERVOIR ST TRENTON NJ																									
City : Lake Hiawatha				Analysis Turnaround Time 7 DAY				Data Package Type : RESULTS ONLY																					
State : NJ				Standard : 10 business days OR				EDD Type : PDF																					
Zip Code : 07034				Rush (Specify): 7 Days																									
Country :																													
Sample Identification	Sample Date(s)	Time Start (24 hr Clock)	Time Stop (24 hr Clock)	Can Vacuum in Field ("Hg) (Start)	Can Vacuum in Field ("Hg) (Stop)**	Interior Temp. (F) (Start)	Interior Temp. (F) (Stop)	Out going Can Pressure ("Hg)(Lab)	In coming Can Pressure ("Hg)(Lab)	Flow Reg. ID	Can ID		Flow Controller Readout (ml/min)	Can Cert ID	TO-15	Indoor/ Ambient Air	Soil Gas												
IA3	10/25/12	11:12	30	5	70	70	-30	-5.1	10579	10285	6 L	50	VL042907.D	1															
Temperature (Fahrenheit) <table border="1"> <tr> <td></td> <td>Ambient</td> <td>Maximum</td> <td>Minimum</td> </tr> <tr> <td>Start</td> <td></td> <td></td> <td></td> </tr> <tr> <td>Stop</td> <td></td> <td></td> <td></td> </tr> </table>											Ambient	Maximum	Minimum	Start				Stop				GC/MS Analyst Signature (TO-15) 							
	Ambient	Maximum	Minimum																										
Start																													
Stop																													
Pressure (Inches of Hg) <table border="1"> <tr> <td></td> <td>Ambient</td> <td>Maximum</td> <td>Minimum</td> </tr> <tr> <td>Start</td> <td></td> <td></td> <td></td> </tr> <tr> <td>Stop</td> <td></td> <td></td> <td></td> </tr> </table>											Ambient	Maximum	Minimum	Start				Stop				** Submittal of this COC indicates approval of the analysis based on existing conditions. Please follow the instructions on the back of this COC.							
	Ambient	Maximum	Minimum																										
Start																													
Stop																													
Special Instructions/QC Requirements & Comments :																													
Suspected Contamination: High Medium LOW PID Readings: 0.0																													
Sampling site (State):																													
Quick Connector required : No																													
Canisters Shipped by: Sam				Date/Time: 10/07/12				Canisters Received by: CR				Date/Time: 10/8/12 1340																	
Samples Relinquished by: FRANK				Date/Time: 10/8/12				Received by:				Date/Time:																	
Relinquished by:				Date/Time:				Received by:				Date/Time:																	

B2509079 - 2

Client Contact Information				Bottle Order ID : B2509079				Courier : F. Galdun				6 of 13 COCs					
Client ID : GFEL01 Project ID : All Projects				Sampler Name(s) : FRANK GALDUN				Analysis				Matrix					
Customer Name : GFE LLC Address : 58 Nokomis Ave City : Lake Hiawatha State : NJ Zip Code : 07034 Country :				Project Manager : Frank galdun				AIR ANALYSIS CHAIN-OF-CUSTODY Batch Certified									
				Phone Number : 646-542-3465													
				Fax Number : 973-334-1692													
				Site Details: 200 RESERVOIR ST TRENTON, NJ													
Analysis Turnaround Time : 7 DAY				Standard : 20 Business days OR				Data Package Type :									
Rush (Specify): 7 Days				EDD Type :													
Sample Identification	Sample Date(s)	Time Start (24 hr Clock)	Time Stop (24 hr Clock)	Can Vacuum in Field ("Hg) (Start)	Can Vacuum in Field ("Hg) (Stop)**	Interior Temp. (F) (Start)	Interior Temp. (F) (Stop)	Out going Can Pressure ("Hg)(Lab)	In coming Can Pressure ("Hg)(Lab)	Flow Reg. ID	Can ID		Flow Controller Readout (ml/min)	Can Cert ID	TO-15	Indoor/Ambient Air	Soil Gas
SV3	10/8/25	9:16	11:16	20	2	70	70	-30	-3.9	10613	10443	6 L	50	VL042942.D			1
Temperature (Fahrenheit)										GC/MS Analyst Signature (TO-15) 							
		Ambient	Maximum	Minimum													
Start																	
Stop																	
Pressure (Inches of Hg)										** Submittal of this COC indicates approval of the analysis based on existing conditions. Please follow the instructions on the back of this COC.							
		Ambient	Maximum	Minimum													
Start																	
Stop																	
Special Instructions/QC Requirements & Comments :																	
Suspected Contamination: High Medium <u>Low</u> PID Readings: 0.0																	
Sampling site (State):																	
Quick Connector required : NO																	
Canisters Shipped by:				Date/Time: 10/8/25				Canisters Received by:				Date/Time: 10/8/25 1340					
Samples Relinquished by:				Date/Time: 10/8/25				Received by:				Date/Time:					
Relinquished by:				Date/Time:				Received by:				Date/Time:					

B2509079 - 6


Client Contact Information				Bottle Order ID : B2509079				Courier : FGALDUN				7 of 13 COCs																
Client ID : GFEL01 Project ID : All Projects				Sampler Name(s) FRANK GALDUN				Analysis				Matrix																
Customer Name : GFE LLC Address : 58 Nokomis Ave City : Lake Hiawatha State : NJ Zip Code : 07034 Country :				Project Manager : Frank galdun				AIR ANALYSIS CHAIN-OF-CUSTODY Individual Certified																				
				Phone Number : 646-542-3465																								
				Fax Number : 973-334-1692																								
				Site Details: 200 RESERVOIR ST. TRENTON, NJ																								
Analysis Turnaround Time 7 DAY				Standard : 10 business days OR				Data Package Type : RESULTS ONLY																				
Rush (Specify): 7 Days				EDD Type : PDI																								
Sample Identification	Sample Date(s)	Time Start (24 hr Clock)	Time Stop (24 hr Clock)	Can Vacuum in Field ("Hg) (Start)	Can Vacuum in Field ("Hg) (Stop)**	Interior Temp. (F) (Start)	Interior Temp. (F) (Stop)	Out going Can Pressure ("Hg)(Lab)	In coming Can Pressure ("Hg)(Lab)	Flow Reg. ID	Can ID	Flow Controller Readout (ml/min)	Can Cert ID	TO-15	Indoor/ Ambient Air	Soil Gas												
IA4	10/8/25	11:20	11:30	4	70	70	-30	-43	10226	10311	6 L	50	VL042703.D	1	1													
Temperature (Fahrenheit) <table border="1" style="width:100%; border-collapse: collapse;"> <tr> <td></td> <td>Ambient</td> <td>Maximum</td> <td>Minimum</td> </tr> <tr> <td>Start</td> <td></td> <td></td> <td></td> </tr> <tr> <td>Stop</td> <td></td> <td></td> <td></td> </tr> </table>											Ambient	Maximum	Minimum	Start				Stop				GC/MS Analyst Signature (TO-15) [Signature]						
	Ambient	Maximum	Minimum																									
Start																												
Stop																												
Pressure (Inches of Hg) <table border="1" style="width:100%; border-collapse: collapse;"> <tr> <td></td> <td>Ambient</td> <td>Maximum</td> <td>Minimum</td> </tr> <tr> <td>Start</td> <td></td> <td></td> <td></td> </tr> <tr> <td>Stop</td> <td></td> <td></td> <td></td> </tr> </table>											Ambient	Maximum	Minimum	Start				Stop				** Submittal of this COC indicates approval of the analysis based on existing conditions. Please follow the instructions on the back of this COC.						
	Ambient	Maximum	Minimum																									
Start																												
Stop																												
Special Instructions/QC Requirements & Comments :																												
Suspected Contamination: High Medium Low PID Readings: 0.0																												
Sampling site (State):																												
Quick Connector required : NO																												
Canisters Shipped by: Sam				Date/Time: 10/8/25				Canisters Received by: ch				Date/Time: 10/8/25 1340																
Samples Relinquished by: FRANK				Date/Time: 10/8/25				Received by:				Date/Time:																
Relinquished by:				Date/Time:				Received by:				Date/Time:																

B2509079 - 3

Client Contact Information				Bottle Order ID : B2509079				Courier : F GALDUN				8 of 13 COCs					
Client ID : GFEL01				Project ID : All Projects				Sampler Name(s) : FRANK GALDUN				Analysis		Matrix			
Customer Name : GFE LLC Address : 58 Nokomis Ave				Project Manager : Frank galdun				AIR ANALYSIS CHAIN-OF-CUSTODY Batch Certified									
				Phone Number : 646-542-3465													
				Fax Number : 973-334-1692													
				Site Details: 200 RESERVOIR ST IRENTON, NJ													
City : Lake Hiawatha				Analysis Turnaround Time 7 DAY				Data Package Type RESULTS ONLY									
State : NJ				Standard : 10 business days OR				EDD Type : PDF									
Zip Code : 07034				Rush (Specify): 7 Days													
Country :																	
Sample Identification	Sample Date(s)	Time Start (24 hr Clock)	Time Stop (24 hr Clock)	Can Vacuum in Field ("Hg) (Start)	Can Vacuum in Field ("Hg) (Stop)**	Interior Temp. (F) (Start)	Interior Temp. (F) (Stop)	Out going Can Pressure ("Hg)(Lab)	In coming Can Pressure ("Hg)(Lab)	Flow Reg. ID	Can ID		Flow Controller Readout (ml/min)	Can Cert ID	TO-15	Indoor/Ambient Air	Soil Gas
SVH	10/18/25	9:26	11:26	COVER	30	70	70	-30	-5.5	10525	10588	6 L	50	VL042702.D	1		1
Temperature (Fahrenheit)										GC/MS Analyst Signature (TO-15) Sut							
		Ambient	Maximum	Minimum													
Start																	
Stop																	
Pressure (Inches of Hg)										** Submittal of this COC indicates approval of the analysis based on existing conditions. Please follow the instructions on the back of this COC.							
		Ambient	Maximum	Minimum													
Start																	
Stop																	
Special Instructions/QC Requirements & Comments :																	
Suspected Contamination: High Medium Low PID Readings: 20																	
Sampling site (State):																	
Quick Connector required : NO																	
Canisters Shipped by: Sam				Date/Time: 10/17/25				Canisters Received by: al				Date/Time: 10/18/25 1340				B2509079 - 8	
Samples Relinquished by: Sam				Date/Time:				Received by:				Date/Time:					
Relinquished by:				Date/Time:				Received by:				Date/Time:					

Client Contact Information				Bottle Order ID : B2509079				Courier : F Galdun				9 of 13 COCs					
Client ID : GFEL01 Project ID : All Projects				Sampler Name(s) FRANK GALDUN				Analysis				Matrix					
Customer Name : GFE LLC Address : 58 Nokomis Ave City : Lake Hiawatha State : NJ Zip Code : 07034 Country :				Project Manager : Frank galdun				AIR ANALYSIS CHAIN-OF-CUSTODY Batch Certified									
				Phone Number : 646-542-3465													
				Fax Number : 973-334-1692													
				Site Details: 200 RESERVOIR ST TRENTON, NJ													
Analysis Turnaround Time 7 DAY				Standard : 10 business days OR				Data Package Type RESULTS ONLY									
Rush (Specify): 7 Days				EDD Type : PDF													
Sample Identification	Sample Date(s)	Time Start (24 hr Clock)	Time Stop (24 hr Clock)	Can Vacuum in Field ("Hg) (Start)	Can Vacuum in Field ("Hg) (Stop)**	Interior Temp. (F) (Start)	Interior Temp. (F) (Stop)	Out going Can Pressure ("Hg)(Lab)	In coming Can Pressure ("Hg)(Lab)	Flow Reg. ID	Can ID		Flow Controller Readout (ml/min)	Can Cert ID	TO-15	Indoor Ambient Air	Soil Gas
IAS	10/6/25 9:36	11:36	1:30	5.5	7.0	70	70	-30	-5.1	10550	10601	6 L	50	VL042907.D	1	1	
Temperature (Fahrenheit)										GC/MS Analyst Signature (TO-15) [Signature]							
		Ambient	Maximum	Minimum													
Start																	
Stop																	
Pressure (Inches of Hg)										** Submittal of this COC indicates approval of the analysis based on existing conditions. Please follow the instructions on the back of this COC.							
		Ambient	Maximum	Minimum													
Start																	
Stop																	
Special Instructions/QC Requirements & Comments :																	
Suspected Contamination: High Medium Low PID Readings: 0.0																	
Sampling site (State):																	
Quick Connector required : NO																	
Canisters Shipped by: SEM				Date/Time: 10/6/25				Canisters Received by: CR				Date/Time: 10/6/25 1340					
Samples Relinquished by: SEM				Date/Time: 10/8/25				Received by:				Date/Time:					
Relinquished by:				Date/Time:				Received by:				Date/Time:					
B2509079 - 9																	

Client Contact Information				Bottle Order ID : B2509079				Courier : F. GALDUN				ID of 13 COCs										
Client ID : GFEL01				Project ID : All Projects				Sampler Name(s) : FRANK GALDUN				Analysis		Matrix								
Customer Name : GFE LLC Address : 58 Nokomis Ave				Project Manager : Frank galdun				AIR ANALYSIS CHAIN-OF-CUSTODY Batch Certified														
				Phone Number : 646-542-3465																		
				Fax Number : 973-334-1692																		
				Site Details: 200 RESERVOIR ST. TRENTON, NJ																		
City : Lake Hiawatha				Analysis Turnaround Time 7 DAY				Data Package Type : RESULTS ONLY														
State : NJ				Standard : 10 business days OR																		
Zip Code : 07034				Rush (Specify): 7 Days				EDD Type : PDF														
Sample Identification		Sample Date(s)	Time Start (24 hr Clock)	Time Stop (24 hr Clock)	Can Vacuum in Field ("Hg) (Start)	Can Vacuum in Field ("Hg) (Stop)**	Interior Temp. (F) (Start)	Interior Temp. (F) (Stop)	Out going Can Pressure ("Hg)(Lab)	In coming Can Pressure ("Hg)(Lab)	Flow Reg. ID	Can ID		Flow Controller Readout (ml/min)	Can Cert ID	TO-15	Indoor/Ambient Air	Soil Gas				
SVS		10/8/25	9:44	11:34	30	30	70	70	-30	-5.1	10648	10604	6 L	50	VL042907.D	1		1				
Temperature (Fahrenheit)											GC/MS Analyst Signature (TO-15) 											
		Ambient		Maximum		Minimum																
Start																						
Stop																						
Pressure (Inches of Hg)											** Submittal of this COC indicates approval of the analysis based on existing conditions. Please follow the instructions on the back of this COC.											
		Ambient		Maximum		Minimum																
Start																						
Stop																						
Special Instructions/QC Requirements & Comments :																						
Suspected Contamination: High Medium Low PID Readings: 0.0																						
Sampling site (State):																						
Quick Connector required : No																						
Canisters Shipped by: Sam				Date/Time: 10/8/25				Canisters Received by: Chen				Date/Time: 10/8/25 1340				B2509079 - 11						
Samples Relinquished by: FRANK				Date/Time: 10/8/25				Received by:				Date/Time:										
Relinquished by:				Date/Time:				Received by:				Date/Time:										

Client Contact Information				Bottle Order ID : B2509079				Courier: FGaldun				11 of 13 COCs							
Client ID : GFEL01 Project ID : All Projects				Sampler Name(s): Frank Galdun				Analysis				Matrix							
Customer Name : GFE LLC Address : 58 Nokomis Ave City : Lake Hiawatha State : NJ Zip Code : 07034 Country :				Project Manager : Frank galdun				AIR ANALYSIS CHAIN-OF-CUSTODY Batch Certified				Indoor Ambient Air Soil Gas							
				Phone Number : 646-542-3465															
				Fax Number : 973-334-1692															
				Site Details: 200 Reservoir St TRENTON NJ															
Analysis Turnaround Time 7 DAY				Standard : 10 Business Days OR				Data Package Type: Results Only											
Rush (Specify): 7 Days				EDD Type : PDF															
Sample Identification	Sample Date(s)	Time Start (24 hr Clock)	Time Stop (24 hr Clock)	Can Vacuum in Field ("Hg) (Start)	Can Vacuum in Field ("Hg) (Stop)**	Interior Temp. (F) (Start)	Interior Temp. (F) (Stop)	Out going Can Pressure ("Hg)(Lab)	In coming Can Pressure ("Hg)(Lab)	Flow Reg. ID	Can ID		Flow Controller Readout (ml/min)	Can Cert ID	TO-15				
IA6	10/8/25	7:48	11:48	30	29	70	70	-30	-5.1	10617	10607	6 L	50	VL042907.D	1				
Temperature (Fahrenheit)										GC/MS Analyst Signature (TO-15) 									
		Ambient	Maximum	Minimum															
Start																			
Stop																			
Pressure (Inches of Hg)										** Submittal of this COC indicates approval of the analysis based on existing conditions. Please follow the instructions on the back of this COC.									
		Ambient	Maximum	Minimum															
Start																			
Stop																			
Special Instructions/QC Requirements & Comments :																			
Suspected Contamination: High Medium Low PID Readings: 0.0																			
Sampling site (State):																			
Quick Connector required : No																			
Canisters Shipped by: Sam				Date/Time: 10/8/25				Canisters Received by: CR				Date/Time: 10/8/25 1340				B2509079 - 12			
Samples Relinquished by: Frank				Date/Time: 10/8/25				Received by:				Date/Time:							
Relinquished by:				Date/Time:				Received by:				Date/Time:							

Client Contact Information				Bottle Order ID : B2509079				Courier : F. Galdun				12 of 13 COCs				
Client ID : GFEL01 Project ID : All Projects				Sampler Name(s) : Frank Galdun				Analysis				Matrix				
Customer Name : GFE LLC Address : 58 Nokomis Ave City : Lake Hiawatha State : NJ Zip Code : 07034 Country :				Project Manager : Frank galdun				AIR ANALYSIS CHAIN-OF-CUSTODY Batch Certified								
				Phone Number : 646-542-3465												
				Fax Number : 973-334-1692												
				Site Details: 300 Reservoir St TRENTON NJ												
Analysis Turnaround Time 7 DAY				Standard : 10 business days OR				Data Package Type : Results only								
Rush (Specify): 7 Days				EDD Type : RDE												
Sample Identification	Sample Date(s)	Time Start (24 hr Clock)	Time Stop (24 hr Clock)	Can Vacuum in Field ("Hg) (Start)	Can Vacuum in Field ("Hg) (Stop)**	Interior Temp. (F) (Start)	Interior Temp. (F) (Stop)	Out going Can Pressure ("Hg)(Lab)	In coming Can Pressure ("Hg)(Lab)	Flow Reg. ID	Can ID	Flow Controller Readout (ml/min)	Can Cert ID	TO-15	Indoor/Ambient Air	Soil Gas
SV6	10/18/15	9:58	11:58	6.5	70	70	-30	-5.1	10528	10333	6 L	50	VL042907.D	1		1
Temperature (Fahrenheit)										GC/MS Analyst Signature (TO-15) Sust						
		Ambient	Maximum	Minimum												
Start																
Stop																
Pressure (Inches of Hg)										** Submittal of this COC indicates approval of the analysis based on existing conditions. Please follow the instructions on the back of this COC.						
		Ambient	Maximum	Minimum												
Start																
Stop																
Special Instructions/QC Requirements & Comments :																
Suspected Contamination: High Medium Low PID Readings: 0.7																
Sampling site (State):																
Quick Connector required : NO																
Canisters Shipped by: Sam				Date/Time: 10/18/15				Canisters Received by: AR				Date/Time: 10/18/15 1340				B2509079 - 5
Samples Relinquished by: FDN				Date/Time: 10/18/15				Received by:				Date/Time:				
Relinquished by:				Date/Time:				Received by:				Date/Time:				

Client Contact Information				Bottle Order ID : B2509079				Courier : FGALDUN				13 of 13 COCs																	
Client ID : GFEL01				Project ID : All Projects				Sampler Name(s) : FRANK GALDUN				Analysis		Matrix															
Customer Name : GFE LLC Address : 58 Nokomis Ave City : Lake Hiawatha State : NJ Zip Code : 07034 Country :				Project Manager : Frank galdun				AIR ANALYSIS CHAIN-OF-CUSTODY Batch Certified																					
				Phone Number : 646-542-3465																									
				Fax Number : 973-334-1692																									
				Site Details: ZOO RESERVOIR ST TRENTON, NJ																									
				Analysis Turnaround Time 7 DAY				Data Package Type : RESULTS ONLY PDF																					
				Standard : 10 business days OR				EDD Type :																					
Rush (Specify): 7 Days																													
Sample Identification	Sample Date(s)	Time Start (24 hr Clock)	Time Stop (24 hr Clock)	Can Vacuum in Field ("Hg) (Start)	Can Vacuum in Field ("Hg) (Stop)**	Interior Temp. (F) (Start)	Interior Temp. (F) (Stop)	Out going Can Pressure ("Hg)(Lab)	In coming Can Pressure ("Hg)(Lab)	Flow Reg. ID	Can ID		Flow Controller Readout (ml/min)	Can Cert ID	TO-15	Indoor/Ambient Air	Soil Gas												
OA1	10/8/25	10:05	12:05	over 30	5.5			-30	-4.3	10707	10609	6 L	50	VL042907.D	1	1													
Temperature (Fahrenheit) <table border="1"> <tr> <td></td> <td>Ambient</td> <td>Maximum</td> <td>Minimum</td> </tr> <tr> <td>Start</td> <td>70</td> <td></td> <td></td> </tr> <tr> <td>Stop</td> <td>67</td> <td></td> <td></td> </tr> </table>											Ambient	Maximum	Minimum	Start	70			Stop	67			GC/MS Analyst Signature (TO-15) [Signature]							
	Ambient	Maximum	Minimum																										
Start	70																												
Stop	67																												
Pressure (Inches of Hg) <table border="1"> <tr> <td></td> <td>Ambient</td> <td>Maximum</td> <td>Minimum</td> </tr> <tr> <td>Start</td> <td></td> <td></td> <td></td> </tr> <tr> <td>Stop</td> <td></td> <td></td> <td></td> </tr> </table>											Ambient	Maximum	Minimum	Start				Stop				** Submittal of this COC indicates approval of the analysis based on existing conditions. Please follow the instructions on the back of this COC.							
	Ambient	Maximum	Minimum																										
Start																													
Stop																													
Special Instructions/QC Requirements & Comments :																													
Suspected Contamination: High Medium Low PID Readings: 0.0																													
Sampling site (State):																													
Quick Connector required : NO																													
Canisters Shipped by: [Signature]				Date/Time: 10/8/25				Canisters Received by: [Signature]				Date/Time: 10/8/25 1340																	
Samples Relinquished by: [Signature]				Date/Time: 10/8/25				Received by:				Date/Time:																	
Relinquished by:				Date/Time:				Received by:				Date/Time:																	

B2509079 - 13

Laboratory Certification

Certified By	License No.
Connecticut	PH-0830
DOD ELAP (ANAB)	L2219
Maine	2024021
Maryland	296
New Hampshire	255425
New Jersey	20012
New York	11376
Pennsylvania	68-00548
Soil Permit	525-24-234-08441
Texas	TX-C25-00189
Virginia	460312

Internal Chain of Custody

Instructions: Use 1 form for each 20 samples of aliquot

Laboratory Person Breaking Field Seal on Sample Shuttle & Accepting Responsibility for Sample

Laboratory: Chemtech

Location: 284 Sheffield Street, Mountainside, NJ 7092

NORGE

Title: Sample Custodian

Field Sample Seal No. Q3320


Date Broken: 10/8/2025

Military Time Seal Broken: 13:40:00

Case No.: 200 Reservoir St, Trenton

Analytical Parameter/Fraction: WOCMS Group2

Sample No.	Aliquot/Extract No.	Sample No.	Aliquot/Extract No.
Q3320-01	IA1	Q3320-11	IA6
Q3320-02	SV1	Q3320-12	SV6
Q3320-03	IA2	Q3320-13	OA1
Q3320-04	SV2		
Q3320-05	IA3		
Q3320-06	SV3		
Q3320-07	IA4		
Q3320-08	SV4		
Q3320-09	IA5		
Q3320-10	SV5		

Date	Time	Relinquished By	Received By	Purpose of Change of Custody
10/9	0850	Signature 	Signature	
		Printed Name <u>George N.</u>	Printed Name <u>George N.</u>	
		Signature 	Signature	
		Printed Name	Printed Name	
		Signature	Signature	
		Printed Name	Printed Name	
		Signature	Signature	
		Printed Name	Printed Name	
		Signature	Signature	
		Printed Name	Printed Name	
		Signature	Signature	
		Printed Name	Printed Name	
		Signature	Signature	
		Printed Name	Printed Name	

Distribution: White - Original (Sent With Report)

Yellow - Contractor Archive

Pink - Sample Custodian - Interim Copy