

Surrogate Summary

SDG No.: Q3320

Client: GFE LLC

Analytical Method: SWTO-15

Lab Sample ID	Client ID	Parameter	Spike	Result	Recovery (%)	Qual	Limits (%)	
							Low	High
Q3320-01	IA1	1-Bromo-4-Fluorobenzene	10	10.7	107		65	135
Q3320-01DUP	IA1DUP	1-Bromo-4-Fluorobenzene	10	10.7	107		65	135
Q3320-02	SV1	1-Bromo-4-Fluorobenzene	10	11.1	111		65	135
Q3320-02DL	SV1DL	1-Bromo-4-Fluorobenzene	10	11.8	118		65	135
Q3320-03	IA2	1-Bromo-4-Fluorobenzene	10	10.6	106		65	135
Q3320-04	SV2	1-Bromo-4-Fluorobenzene	10	11.8	118		65	135
Q3320-04DL	SV2DL	1-Bromo-4-Fluorobenzene	10	11.8	118		65	135
Q3320-05	IA3	1-Bromo-4-Fluorobenzene	10	9.74	97		65	135
Q3320-06	SV3	1-Bromo-4-Fluorobenzene	10	11.2	112		65	135
Q3320-07	IA4	1-Bromo-4-Fluorobenzene	10	9.88	99		65	135
Q3320-08	SV4	1-Bromo-4-Fluorobenzene	10	11.4	114		65	135
Q3320-08DL	SV4DL	1-Bromo-4-Fluorobenzene	10	12.4	124		65	135
Q3320-09	IA5	1-Bromo-4-Fluorobenzene	10	10.3	103		65	135
Q3320-10	SV5	1-Bromo-4-Fluorobenzene	10	11.6	116		65	135
Q3320-11	IA6	1-Bromo-4-Fluorobenzene	10	10.8	108		65	135
Q3320-12	SV6	1-Bromo-4-Fluorobenzene	10	11.8	118		65	135
Q3320-12DL	SV6DL	1-Bromo-4-Fluorobenzene	10	11.6	116		65	135
Q3320-13	OA1	1-Bromo-4-Fluorobenzene	10	9.80	98		65	135
VL1008ABL01	VL1008ABL01	1-Bromo-4-Fluorobenzene	10	9.46	95		65	135
VL1008ABS01	VL1008ABS01	1-Bromo-4-Fluorobenzene	10	10.3	103		65	135
VL1009ABL01	VL1009ABL01	1-Bromo-4-Fluorobenzene	10	10.3	103		65	135
VL1009ABS01	VL1009ABS01	1-Bromo-4-Fluorobenzene	10	11.4	114		65	135