

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance Contract: GFEL01
 Lab Code: ACE SDG NO.: Q3320
 Lab File ID: VL043079.D Date Analyzed: 10/09/2025
 Instrument ID: MSVOA_L Time Analyzed: 09:29
 GC Column: RTX-1 ID: 0.32 (mm) Heated Purge: (Y/N) N

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	197992	2.78	568897	3.94	498408	8.87
UPPER LIMIT	277189	3.11	796456	4.27	697771	9.20
LOWER LIMIT	118795	2.45	341338	3.61	299045	8.54
EPA SAMPLE NO.						
IA1	181836	2.77	493969	3.94	418241	8.86
IA1DUP	196118	2.78	517619	3.94	434677	8.87
SV1	191613	2.77	543620	3.94	535675	8.86
SV1DL	154952	2.77	402038	3.94	367716	8.86
IA2	200129	2.78	528770	3.95	446168	8.87
SV2	189452	2.78	494714	3.94	492557	8.86
SV2DL	169568	2.78	455409	3.94	411145	8.87
SV3	191981	2.77	503256	3.94	519202	8.86
SV4	172723	2.77	465309	3.94	484936	8.86
SV4DL	170025	2.77	427891	3.94	411008	8.86
SV5	196019	2.78	525980	3.94	487210	8.86
IA6	194363	2.78	513990	3.94	433895	8.87
SV6	179161	2.78	516837	3.94	480307	8.86
SV6DL	192317	2.77	513474	3.94	466022	8.86
VL1009ABL01	178561	2.78	511863	3.94	429488	8.87
VL1009ABS01	181390	2.78	514089	3.94	451201	8.87

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: Alliance Contract: GFEL01
 Lab Code: ACE SDG NO.: Q3320
 Lab File ID: VL043079.D Date Analyzed: 10/09/2025
 Instrument ID: MSVOA_L Time Analyzed: 09:29
 GC Column: RTX-1 ID: 0.32 (mm) Heated Purge: (Y/N) N

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	197992	2.78	568897	3.94	498408	8.87
UPPER LIMIT	277189	3.11	796456	4.27	697771	9.20
LOWER LIMIT	118795	2.45	341338	3.61	299045	8.54
EPA SAMPLE NO.						

IS1 = Bromochloromethane
 IS2 = 1,4-Difluorobenzene
 IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +40% of internal standard area
 AREA LOWER LIMIT = -40% of internal standard area
 RT UPPER LIMIT = +0.33 minutes of internal standard RT
 RT LOWER LIMIT = -0.33 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.