

CASE NARRATIVE

Weston Solutions, Inc.

Project Name: RFP 918

Project # N/A

Order ID # Q3321

Test Name: TCLP Pesticide

A. Number of Samples and Date of Receipt:

2 Solid samples were received on 10/09/2025.

B. Parameters

According to the Chain of Custody document, the following analyses were requested: TCLP Pesticide. This data package contains results for TCLP Pesticide.

C. Analytical Techniques:

The analysis was performed on instrument ECD_D. The front column is ZB-MR1 which is 30 meters, 0.32 mm ID, 0.5 um df,; Catalog # 7HM-G016-17. The rear column is ZB-MR2 which is 30 meters, 0.32 mm ID, 0.25 um df, Catalog #: 7HMG017- 11. The analysis of TCLP Pesticides was based on method 8081B and extraction was done based on method 3510 and TCLP extraction method was 1311.

D. QA/ QC Samples:

The Holding Times were met for all analysis.

The Surrogate recoveries were met for all analysis.

The Retention Times were met for all analysis.

The MS recoveries met the requirements for all compounds.

The MSD recoveries met the requirements for all compounds.

The RPD were met for all analysis.

The Blank Spike met requirements for all compounds.

The Blank analysis did not indicate the presence of lab contamination.

The Initial Calibration met the requirements.

The Continuous Calibration met the requirements.

E. Additional Comments:

F. Calculation for water sample:

$$\text{Concentration ug/L} = \frac{(A_x) (V_t) (DF) (GPC)}{(CF) (V_o) (V_i)}$$

Where,

Ax = Response (peak area or height) of the compound to be measured.



284 Sheffield Street, Mountainside, NJ 07092
Phone: 908 789 8900 Fax: 908 789 8922

CF = Mean Calibration Factor from the initial calibration (area/ng).

Vo = Volume of water extracted in mL

Vt = Volume of the concentrated extract in uL

Vi = Volume of extract injected (uL).

GPC = $\frac{V_{in}}{V_{out}}$ = GPC factor (If no GPC is performed, GPC=1)

Vin = Volume of extract loaded onto GPC column.

Vout = Volume of extract collected after GPC cleanup

DF = Dilution Factor

G. Manual Integration Comments:

Please refer to the Manual integration Report included with the Run Logs for information on the manual integrations performed.

I certify that the data package is in compliance with the terms and conditions of the contract, both technically and for completeness, for other than the conditions detailed above. The laboratory manager or his designee, as verified by the following signature has authorized release of the data contained in this hard copy data package.

Signature_____