

Initial and Continuing Calibration Verification

Client: Weston Solutions, Inc.

SDG No.: Q3321

Project: RFP 918

RunNo.: LB137516

Analyte		Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID: Cyanide	ICV1	mg/L	0.096	0.099	97	90-110	10/14/2025
Sample ID: Cyanide	CCV1	mg/L	0.24	0.25	96	90-110	10/14/2025
Sample ID: Cyanide	CCV2	mg/L	0.24	0.25	96	90-110	10/14/2025