



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Weston Solutions, Inc.

SDG No.: Q3321

Contract: ROYF02

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
ICB70	Mercury	0.076	+/-0.2	U	0.20	CV	10/14/2025	11:12	LB137514

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Weston Solutions, Inc.

SDG No.: Q3321

Contract: ROYF02

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB26	Mercury	0.076	+/-0.2	U	0.20	CV	10/14/2025	11:22	LB137514
CCB27	Mercury	0.076	+/-0.2	U	0.20	CV	10/14/2025	11:55	LB137514
CCB28	Mercury	0.076	+/-0.2	U	0.20	CV	10/14/2025	12:25	LB137514
CCB29	Mercury	0.076	+/-0.2	U	0.20	CV	10/14/2025	12:41	LB137514

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Weston Solutions, Inc.

SDG No.: Q3321

Contract: ROYF02

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
ICB01	Arsenic	5.12	+/-10	U	20.0	P	10/13/2025	12:55	LB137518
	Barium	14.6	+/-50	U	100	P	10/13/2025	12:55	LB137518
	Cadmium	0.50	+/-3	U	6.00	P	10/13/2025	12:55	LB137518
	Chromium	2.12	+/-5	U	10.0	P	10/13/2025	12:55	LB137518
	Lead	2.30	+/-6	U	12.0	P	10/13/2025	12:55	LB137518
	Selenium	9.64	+/-10	U	20.0	P	10/13/2025	12:55	LB137518
	Silver	1.62	+/-5	U	10.0	P	10/13/2025	12:55	LB137518

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Weston Solutions, Inc.

SDG No.: Q3321

Contract: ROYF02

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB01	Arsenic	5.12	+/-10	U	20.0	P	10/13/2025	13:25	LB137518
	Barium	14.6	+/-50	U	100	P	10/13/2025	13:25	LB137518
	Cadmium	0.50	+/-3	U	6.00	P	10/13/2025	13:25	LB137518
	Chromium	2.12	+/-5	U	10.0	P	10/13/2025	13:25	LB137518
	Lead	2.30	+/-6	U	12.0	P	10/13/2025	13:25	LB137518
	Selenium	9.64	+/-10	U	20.0	P	10/13/2025	13:25	LB137518
	Silver	1.62	+/-5	U	10.0	P	10/13/2025	13:25	LB137518
CCB02	Arsenic	5.12	+/-10	U	20.0	P	10/13/2025	14:22	LB137518
	Barium	14.6	+/-50	U	100	P	10/13/2025	14:22	LB137518
	Cadmium	0.50	+/-3	U	6.00	P	10/13/2025	14:22	LB137518
	Chromium	2.12	+/-5	U	10.0	P	10/13/2025	14:22	LB137518
	Lead	2.30	+/-6	U	12.0	P	10/13/2025	14:22	LB137518
	Selenium	9.64	+/-10	U	20.0	P	10/13/2025	14:22	LB137518
	Silver	1.62	+/-5	U	10.0	P	10/13/2025	14:22	LB137518
CCB03	Arsenic	5.12	+/-10	U	20.0	P	10/13/2025	16:08	LB137518
	Barium	14.6	+/-50	U	100	P	10/13/2025	16:08	LB137518
	Cadmium	0.50	+/-3	U	6.00	P	10/13/2025	16:08	LB137518
	Chromium	2.12	+/-5	U	10.0	P	10/13/2025	16:08	LB137518
	Lead	2.30	+/-6	U	12.0	P	10/13/2025	16:08	LB137518
	Selenium	9.64	+/-10	U	20.0	P	10/13/2025	16:08	LB137518
	Silver	1.62	+/-5	U	10.0	P	10/13/2025	16:08	LB137518
CCB04	Arsenic	5.12	+/-10	U	20.0	P	10/13/2025	17:06	LB137518
	Barium	14.6	+/-50	U	100	P	10/13/2025	17:06	LB137518
	Cadmium	0.50	+/-3	U	6.00	P	10/13/2025	17:06	LB137518
	Chromium	2.12	+/-5	U	10.0	P	10/13/2025	17:06	LB137518
	Lead	2.30	+/-6	U	12.0	P	10/13/2025	17:06	LB137518
	Selenium	9.64	+/-10	U	20.0	P	10/13/2025	17:06	LB137518
	Silver	1.62	+/-5	U	10.0	P	10/13/2025	17:06	LB137518
CCB05	Arsenic	5.12	+/-10	U	20.0	P	10/13/2025	18:15	LB137518
	Barium	14.6	+/-50	U	100	P	10/13/2025	18:15	LB137518
	Cadmium	0.50	+/-3	U	6.00	P	10/13/2025	18:15	LB137518
	Chromium	2.12	+/-5	U	10.0	P	10/13/2025	18:15	LB137518
	Lead	2.30	+/-6	U	12.0	P	10/13/2025	18:15	LB137518
	Selenium	9.64	+/-10	U	20.0	P	10/13/2025	18:15	LB137518
	Silver	1.62	+/-5	U	10.0	P	10/13/2025	18:15	LB137518
CCB06	Arsenic	5.12	+/-10	U	20.0	P	10/13/2025	19:01	LB137518
	Barium	14.6	+/-50	U	100	P	10/13/2025	19:01	LB137518
	Cadmium	0.50	+/-3	U	6.00	P	10/13/2025	19:01	LB137518
	Chromium	2.12	+/-5	U	10.0	P	10/13/2025	19:01	LB137518

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Weston Solutions, Inc.

SDG No.: Q3321

Contract: ROYF02

Lab Code: ACE

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB06	Lead	2.30	+/-6	U	12.0	P	10/13/2025	19:01	LB137518
	Selenium	9.64	+/-10	U	20.0	P	10/13/2025	19:01	LB137518
	Silver	1.62	+/-5	U	10.0	P	10/13/2025	19:01	LB137518
CCB07	Arsenic	5.12	+/-10	U	20.0	P	10/13/2025	19:50	LB137518
	Barium	14.6	+/-50	U	100	P	10/13/2025	19:50	LB137518
	Cadmium	0.50	+/-3	U	6.00	P	10/13/2025	19:50	LB137518
	Chromium	2.12	+/-5	U	10.0	P	10/13/2025	19:50	LB137518
	Lead	2.30	+/-6	U	12.0	P	10/13/2025	19:50	LB137518
	Selenium	9.64	+/-10	U	20.0	P	10/13/2025	19:50	LB137518
	Silver	1.62	+/-5	U	10.0	P	10/13/2025	19:50	LB137518