

Metals
- 4 -
INTERFERENCE CHECK SAMPLE

Client: <u>Weston Solutions, Inc.</u>	SDG No.: <u>Q3321</u>
Contract: <u>ROYF02</u>	Lab Code: <u>ACE</u>
ICS Source: <u>EPA</u>	Instrument ID: <u>P4</u>

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Low Limit (ug/L)	High Limit (ug/L)	Analysis Date	Analysis Time	Run Number
ICSA01	Arsenic	12.5			-20	20	10/13/2025	13:04	LB137518
	Barium	6.08	6.0	101	-94	106	10/13/2025	13:04	LB137518
	Cadmium	1.68	1.0	168	-5	7	10/13/2025	13:04	LB137518
	Chromium	47.3	52.0	91	42	62	10/13/2025	13:04	LB137518
	Lead	-2.87			-12	12	10/13/2025	13:04	LB137518
	Selenium	13.3			-20	20	10/13/2025	13:04	LB137518
	Silver	-1.68			-10	10	10/13/2025	13:04	LB137518
ICSAB01	Arsenic	115	104	111	88.4	120	10/13/2025	13:08	LB137518
	Barium	495	537	92	437	637	10/13/2025	13:08	LB137518
	Cadmium	1000	972	103	826	1120	10/13/2025	13:08	LB137518
	Chromium	558	542	103	460	624	10/13/2025	13:08	LB137518
	Lead	45.6	49.0	93	37	61	10/13/2025	13:08	LB137518
	Selenium	59.2	46.0	129	26	66	10/13/2025	13:08	LB137518
	Silver	211	201	105	170	232	10/13/2025	13:08	LB137518