

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance
Contract: ROYF02
Lab Code: ACE
SDG NO.: Q3321
GC Column: ZB-MR1
ID: 0.32 (mm)
 Initi. Calib. Date(s): 09/23/2025 09/23/2025

Client Sample No.: CCAL01
Date Analyzed: 10/10/2025
Lab Sample No.: AR1660CCC500
Data File : PP075704.D
Time Analyzed: 08:51

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	5.799	5.708	5.908	521.150	500.000	4.2
Aroclor-1016-2	5.820	5.729	5.929	524.570	500.000	4.9
Aroclor-1016-3	5.883	5.792	5.992	506.450	500.000	1.3
Aroclor-1016-4	5.980	5.889	6.089	547.300	500.000	9.5
Aroclor-1016-5	6.272	6.181	6.381	560.740	500.000	12.1
Aroclor-1260-1	7.389	7.299	7.499	513.460	500.000	2.7
Aroclor-1260-2	7.641	7.551	7.751	504.300	500.000	0.9
Aroclor-1260-3	8.000	7.909	8.109	542.520	500.000	8.5
Aroclor-1260-4	8.227	8.137	8.337	558.320	500.000	11.7
Aroclor-1260-5	8.553	8.464	8.664	543.040	500.000	8.6
Decachlorobiphenyl	10.419	10.334	10.534	52.610	50.000	5.2
Tetrachloro-m-xylene	4.647	4.555	4.755	55.450	50.000	10.9

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** ROYF02
Lab Code: ACE **SDG NO.:** Q3321
GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 09/23/2025 09/23/2025

Client Sample No.: CCAL01 **Date Analyzed:** 10/10/2025
Lab Sample No.: AR1660CCC500 **Data File :** PP075704.D **Time Analyzed:** 08:51

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.884	4.793	4.993	600.150	500.000	20.0
Aroclor-1016-2	4.940	4.851	5.051	574.760	500.000	15.0
Aroclor-1016-3	5.062	4.971	5.171	539.830	500.000	8.0
Aroclor-1016-4	5.102	5.011	5.211	569.380	500.000	13.9
Aroclor-1016-5	5.316	5.225	5.425	543.520	500.000	8.7
Aroclor-1260-1	6.344	6.255	6.455	547.260	500.000	9.5
Aroclor-1260-2	6.530	6.441	6.641	559.980	500.000	12.0
Aroclor-1260-3	6.684	6.595	6.795	540.760	500.000	8.2
Aroclor-1260-4	7.152	7.065	7.265	487.630	500.000	-2.5
Aroclor-1260-5	7.392	7.304	7.504	503.720	500.000	0.7
Decachlorobiphenyl	8.788	8.703	8.903	58.990	50.000	18.0
Tetrachloro-m-xylene	3.789	3.696	3.896	61.290	50.000	22.6