

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** ROYF02
Lab Code: ACE **SDG NO.:** Q3321
GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 09/23/2025 09/23/2025

Client Sample No.: CCAL02 **Date Analyzed:** 10/10/2025
Lab Sample No.: AR1660CCC500 **Data File :** PP075719.D **Time Analyzed:** 15:57

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	5.802	5.708	5.908	487.160	500.000	-2.6
Aroclor-1016-2	5.823	5.729	5.929	496.470	500.000	-0.7
Aroclor-1016-3	5.887	5.792	5.992	475.660	500.000	-4.9
Aroclor-1016-4	5.983	5.889	6.089	515.150	500.000	3.0
Aroclor-1016-5	6.276	6.181	6.381	522.770	500.000	4.6
Aroclor-1260-1	7.393	7.299	7.499	487.010	500.000	-2.6
Aroclor-1260-2	7.646	7.551	7.751	482.090	500.000	-3.6
Aroclor-1260-3	8.003	7.909	8.109	516.070	500.000	3.2
Aroclor-1260-4	8.231	8.137	8.337	539.600	500.000	7.9
Aroclor-1260-5	8.557	8.464	8.664	524.180	500.000	4.8
Decachlorobiphenyl	10.423	10.334	10.534	49.780	50.000	-0.4
Tetrachloro-m-xylene	4.650	4.555	4.755	52.690	50.000	5.4

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** ROYF02
Lab Code: ACE **SDG NO.:** Q3321
GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 09/23/2025 09/23/2025

Client Sample No.: CCAL02 **Date Analyzed:** 10/10/2025
Lab Sample No.: AR1660CCC500 **Data File :** PP075719.D **Time Analyzed:** 15:57

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.887	4.793	4.993	545.000	500.000	9.0
Aroclor-1016-2	4.945	4.851	5.051	567.840	500.000	13.6
Aroclor-1016-3	5.065	4.971	5.171	504.620	500.000	0.9
Aroclor-1016-4	5.105	5.011	5.211	549.820	500.000	10.0
Aroclor-1016-5	5.319	5.225	5.425	522.850	500.000	4.6
Aroclor-1260-1	6.347	6.255	6.455	498.140	500.000	-0.4
Aroclor-1260-2	6.534	6.441	6.641	499.880	500.000	0.0
Aroclor-1260-3	6.687	6.595	6.795	483.000	500.000	-3.4
Aroclor-1260-4	7.155	7.065	7.265	455.280	500.000	-8.9
Aroclor-1260-5	7.396	7.304	7.504	464.890	500.000	-7.0
Decachlorobiphenyl	8.791	8.703	8.903	56.430	50.000	12.9
Tetrachloro-m-xylene	3.793	3.696	3.896	56.550	50.000	13.1