

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** ROYF02
Lab Code: ACE **SDG NO.:** Q3321
GC Column: ZB-MR1 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 09/23/2025 09/23/2025

Client Sample No.: CCAL03 **Date Analyzed:** 10/10/2025
Lab Sample No.: AR1660CCC500 **Data File :** PP075726.D **Time Analyzed:** 18:55

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	5.805	5.708	5.908	500.590	500.000	0.1
Aroclor-1016-2	5.826	5.729	5.929	505.510	500.000	1.1
Aroclor-1016-3	5.889	5.792	5.992	486.290	500.000	-2.7
Aroclor-1016-4	5.986	5.889	6.089	524.170	500.000	4.8
Aroclor-1016-5	6.278	6.181	6.381	526.640	500.000	5.3
Aroclor-1260-1	7.394	7.299	7.499	491.160	500.000	-1.8
Aroclor-1260-2	7.647	7.551	7.751	482.040	500.000	-3.6
Aroclor-1260-3	8.005	7.909	8.109	522.530	500.000	4.5
Aroclor-1260-4	8.233	8.137	8.337	537.420	500.000	7.5
Aroclor-1260-5	8.558	8.464	8.664	527.050	500.000	5.4
Decachlorobiphenyl	10.425	10.334	10.534	50.070	50.000	0.1
Tetrachloro-m-xylene	4.653	4.555	4.755	53.860	50.000	7.7

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** ROYF02
Lab Code: ACE **SDG NO.:** Q3321
GC Column: ZB-MR2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 09/23/2025 09/23/2025

Client Sample No.: CCAL03 **Date Analyzed:** 10/10/2025
Lab Sample No.: AR1660CCC500 **Data File :** PP075726.D **Time Analyzed:** 18:55

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.889	4.793	4.993	559.430	500.000	11.9
Aroclor-1016-2	4.946	4.851	5.051	582.130	500.000	16.4
Aroclor-1016-3	5.065	4.971	5.171	545.950	500.000	9.2
Aroclor-1016-4	5.106	5.011	5.211	554.840	500.000	11.0
Aroclor-1016-5	5.321	5.225	5.425	530.750	500.000	6.2
Aroclor-1260-1	6.348	6.255	6.455	485.170	500.000	-3.0
Aroclor-1260-2	6.535	6.441	6.641	495.840	500.000	-0.8
Aroclor-1260-3	6.688	6.595	6.795	487.670	500.000	-2.5
Aroclor-1260-4	7.157	7.065	7.265	441.990	500.000	-11.6
Aroclor-1260-5	7.396	7.304	7.504	471.200	500.000	-5.8
Decachlorobiphenyl	8.792	8.703	8.903	57.300	50.000	14.6
Tetrachloro-m-xylene	3.793	3.696	3.896	58.270	50.000	16.5