

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** ROYF02
Lab Code: ACE **SDG NO.:** Q3321
GC Column: RTX-CLP **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 10/13/2025 10/13/2025

Client Sample No.: CCAL01 **Date Analyzed:** 10/14/2025
Lab Sample No.: HSTDCCC750 **Data File :** PS032014.D **Time Analyzed:** 21:32

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
2,4,5-TP(Silvex)	9.278	9.180	9.380	703.230	712.500	-1.3
2,4-D	8.408	8.312	8.512	715.770	705.000	1.5
2,4-DCAA	7.280	7.182	7.382	736.750	750.000	-1.8

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** ROYF02
Lab Code: ACE **SDG NO.:** Q3321
GC Column: RTX-CLP2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 10/13/2025 10/13/2025

Client Sample No.: CCAL01 **Date Analyzed:** 10/14/2025
Lab Sample No.: HSTDCCC750 **Data File :** PS032014.D **Time Analyzed:** 21:32

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
2,4,5-TP(Silvex)	9.891	9.792	9.992	709.630	712.500	-0.4
2,4-D	9.000	8.902	9.102	720.290	705.000	2.2
2,4-DCAA	7.742	7.644	7.844	791.240	750.000	5.5