

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** ROYF02
Lab Code: ACE **SDG NO.:** Q3321
GC Column: RTX-CLP **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 10/13/2025 10/13/2025

Client Sample No.: CCAL02 **Date Analyzed:** 10/15/2025
Lab Sample No.: HSTDCCC750 **Data File :** PS032022.D **Time Analyzed:** 01:09

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
2,4,5-TP(Silvex)	9.280	9.180	9.380	710.560	712.500	-0.3
2,4-D	8.411	8.312	8.512	630.440	705.000	-10.6
2,4-DCAA	7.282	7.182	7.382	743.540	750.000	-0.9

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance
Contract: ROYF02
Lab Code: ACE
SDG NO.: Q3321
GC Column: RTX-CLP2
ID: 0.32 (mm)
 Initi. Calib. Date(s): 10/13/2025 10/13/2025

Client Sample No.: CCAL02
Date Analyzed: 10/15/2025
Lab Sample No.: HSTDCCC750
Data File : PS032022.D
Time Analyzed: 01:09

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
2,4,5-TP(Silvex)	9.892	9.792	9.992	705.040	712.500	-1.0
2,4-D	9.003	8.902	9.102	684.950	705.000	-2.8
2,4-DCAA	7.744	7.644	7.844	763.040	750.000	1.7