

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** ROYF02
Lab Code: ACE **SDG NO.:** Q3321
GC Column: RTX-CLP **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 10/13/2025 10/13/2025

Client Sample No.: CCAL03 **Date Analyzed:** 10/16/2025
Lab Sample No.: HSTDCCC750 **Data File :** PS032055.D **Time Analyzed:** 09:59

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
2,4,5-TP(Silvex)	9.279	9.180	9.380	706.490	712.500	-0.8
2,4-D	8.412	8.312	8.512	692.590	705.000	-1.8
2,4-DCAA	7.281	7.182	7.382	845.810	750.000	12.8

CALIBRATION VERIFICATION SUMMARY

Lab Name: Alliance **Contract:** ROYF02
Lab Code: ACE **SDG NO.:** Q3321
GC Column: RTX-CLP2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 10/13/2025 10/13/2025

Client Sample No.: CCAL03 **Date Analyzed:** 10/16/2025
Lab Sample No.: HSTDCCC750 **Data File :** PS032055.D **Time Analyzed:** 09:59

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
2,4,5-TP(Silvex)	9.892	9.792	9.992	710.900	712.500	-0.2
2,4-D	9.002	8.902	9.102	692.080	705.000	-1.8
2,4-DCAA	7.744	7.644	7.844	763.370	750.000	1.8